Operational Qualification Procedure

© Grifols — DG-AT-TSM-000166 1.12.0

Device

Instrument	Serial Number	Code	Installation Date

Qualification Date

Customer

Name	Site	Department	Telephone

Email

Service Engineer

Name	Department	Telephone	Email

Software Versions

Checked/done	Action	Comments
☑	Check the available version and update the instrument to the latest version of the following software: • Erytra application • Technical Service software • Diagnostic application	

Operating System Version

Checked/done	Action	Comments
☑	Check the available version and update the instrument to the latest version of the following operating system: • OS	

Firmware Versions

Checked/done	Action	Comments
☑	Check the available version and update the instrument to the latest version of the following firmware: • CEN-SRM • INC-CUP • CAT-CM • PIP • FLU-CTRL59 • RD-IL421	

Visual Inspection

Action	Results	Comments
Mechanical components electronic cabinet	PASS	
Mechanical components	PASS	

Functional Inspection: Incubators

Action	Results	Comments
Go to tab Incubators\Incubator 1 in TSS. In Regulate Actions section, select MODE RT. Wait until temperature is stable. With an external thermometer and card probe, check that the temperature is within the 24 ºC ± 3 ºC range. Follow steps 1 to 6 of the following procedure.	PASS	
Go to tab Incubators\Incubator 2 in TSS. In Regulate Actions section, select MODE RT. Wait until temperature is stable. With an external thermometer and card probe, check that the temperature is within the 24 ºC ± 3 ºC range. Follow steps 1 to 6 of the following procedure.	PASS	
Go to tab Incubators\Incubator 3 in TSS. In Regulate Actions section, select MODE RT. Wait until temperature is stable. With an external thermometer and card probe, check that the temperature is within the 24 ºC ± 3 ºC range. Follow steps 1 to 6 of the following procedure.	PASS	
Go to tab Incubators\Incubator 4 in TSS. In Regulate Actions section, select MODE RT. Wait until temperature is stable. With an external thermometer and card probe, check that the temperature is within the 24 ºC ± 3 ºC range. Follow steps 1 to 6 of the following procedure.	PASS	
Go to tab Incubators\Incubator 1 in TSS. In Regulate Actions section, select MODE 37. Wait until temperature is stable. With an external thermometer and card probe, check that the temperature is within the 37 ºC ± 1 ºC range. Follow steps 7 to 8 of the following procedure.	PASS	
Go to tab Incubators\Incubator 2 in TSS. Regulate Actions section, select MODE 37. Wait until temperature is stable. With an external thermometer and card probe, check that the temperature is within the 37 ºC ± 1 ºC range. Follow steps 7 to 8 of the following procedure.	PASS	
Go to tab Incubators\Incubator 3 in TSS. Regulate Actions section, select MODE 37. Wait until temperature is stable. With an external thermometer and card probe, check that the temperature is within the 37 ºC ± 1 ºC range. Follow steps 7 to 8 of the following procedure.	PASS	
Go to tab Incubators\Incubator 4 in TSS. Regulate Actions section, select MODE 37. Wait until temperature is stable. With an external thermometer and card probe, check that the temperature is within the 37 ºC ± 1 ºC range. Follow steps 7 to 8 of the following procedure.	PASS	

Instrument Reader Reference Verification

Action	Results	Comments
Perform the Reference check routine to make sure the X, Y, and Z axes are true. Find information of the routine in the following section.	PASS	

Functional Tests: Diagnostic Application

Action

Test the Diagnostic application.

Initialization

Action	Results	Comments
Initialization test	PASS	

Cards Management

Action	Results	Comments
Robot movements	PASS	
BC reading and presence detection	PASS	
Card transport card drawers	PASS	
Card transport centrifuges	PASS	
Card transport incubators	PASS	

Pipetting

Action	Results	Comments
Robot movements	PASS	
Presence detection and BC reading	PASS	
Level detection	PASS	
Pipetting precision and accuracy	PASS	
CPD and PCA level detector	PASS	
Impact detection	PASS	

Fluidics

Action	Results	Comments
Drawers and Load cells	PASS	
Pressure circuits	PASS	
Vacuum circuits	PASS	
Prime, rinse and empty	PASS	

Reader

Action	Results	Comments
Initialization	PASS	

Incubators

Action	Results	Comments

Movements	PASS	
Incubation	PASS	

Centrifuges

Action	Results	Comments
Movements	PASS	
Centrifugation	PASS	

Samples Drawers

Action	Results	Comments
Drawers	PASS	

Reagents Drawers

Action	Results	Comments
Drawers and Stirrers	PASS	

Cards Drawers

Action	Results	Comments
Drawers	PASS	

Dilution Cup

Action	Results	Comments
Stirring	PASS	

Miscellaneous

Action	Results	Comments
Door, keyboard, LEDs and buzzer	PASS	

Functional Inspections: Probe Dispensation Position Verification

Action	Results	Comments
Visually check the cards used in Diagnostic Application test Pipetting precision and accuracy. Holes in card's film must be centered in X axis and shifted about 1.5 mm to the rear part of the instrument in Y axis (distance from the center of the well to the upper border of the hole).	PASS	

Functional Checks

Action	Results	Comments
If LIS is configured to process the 'Units' field of the ASTM message (R-5 of Result record) or LIS is not used: User/operator executes 4 tests of the enabled profiles in Erytra application. Otherwise, if LIS is used but not configured to process the 'Units' field of the ASTM message (not recommended configuration): User/operator executes one test for each enabled profile in the Erytra application. Check that for negative results the size of the button is homogeneous between wells.	PASS	

If the results are sent to LIS, check that LIS receives them successfully and that the results match the ones sent by the instrument.	PASS	

Comments

Date

Service engineer signature

Customer signature

© Grifols — DG-AT-TSM-000166 1.12.0