Learned Regularization for Sparse-View CT:

Accuracy, Lesion Detectability and Distribution Shift of a Plug-and-Play Prior against Tuned Total Variation

Mohammed Abo Oraig
Independent research
mohammed-abo-oraig@etud.univ-tln.fr

Abstract

Learned image priors are often proposed as replacements for hand-crafted regularizers in sparse-view computed tomography (CT), yet they are usually judged by PSNR on data resembling their training set. We evaluate one noise-conditioned DnCNN denoiser used as a Plug-and-Play (PnP) prior in proximal gradient against filtered back-projection (FBP), Tikhonov and total-variation (TV) regularization, with validation-tuned parameters for every method and paired statistics, under distribution shift, forward-model error and Poisson noise, and with a synthetic lesion-insertion test (paired detection AUC, lesion retention, false lesion-like blobs). On synthetic 28×28 digit images the prior beats a box-constrained TV by about 3.4 dB at moderate noise ($\sigma = 0.05$, 20 angles); the advantage shrinks at both ends of the noise range, the ranking near the noise-free limit depends on how the data are simulated, and it reverses under distribution shift: an MNIST-trained prior is 3.1 dB (Fashion-MNIST) and 4.3–4.6 dB (Shepp–Logan phantom) below TV, whereas a Fashion-MNIST-trained prior beats TV by 2.2 dB on Fashion-MNIST, consistent with prior mismatch rather than a failure of PnP itself. These results are reproduced under Poisson noise. On digits the prior detects small lesions less reliably than TV even in distribution, and under mismatch it creates lesion-like blobs more often than TV for most lesion types. On real LoDoPaB-CT slices (patient-disjoint split; 30, 60 and 120 of 1000 angles; 4096 photons per detector bin and one quarter and one tenth of that; 40 test patients) it does not consistently outperform a well-tuned TV given the same iteration budget (300 iterations) in PSNR (-0.93 to $+0.33$ dB over nine conditions; at most $+0.49$ dB when initialised from the TV solution) and has lower SSIM in eight of the nine, but it detects small high-contrast lesions better than that TV in all nine (paired AUC $+0.03$ to $+0.16$, six significant). It never significantly exceeds Hann-filtered FBP in detection, which the PSNR-tuned TV falls below in all nine conditions (seven significant) despite a PSNR gain of 0.8–2.3 dB. On non-anatomical phantoms its PSNR and SSIM deficits against TV are consistent and the detection advantage persists in one of six conditions. PSNR and lesion detectability thus rank the methods differently, and a learned prior's value depends on how well it matches the data it is applied to.

Contents

1 Introduction **2**

2 Related Work **4**

3 Methodology **5**

3.1 Problem setting and experimental settings 5

3.2 Forward operator and classical baselines (Stages 1–2) 6

3.3 Learned prior: PnP proximal gradient . . . 6
3.4 Evaluation protocol (Stages 1–2) . . . 7
3.5 Data simulated on a finer grid . . . 8
3.6 Denoiser seeds, noise range and matched-prior control . . . 8
3.7 Weakly determined subspaces and prior-driven content . . . 8
3.8 Lesion-insertion protocol . . . 9
3.9 Matrix-free operator . . . 9
3.10 Poisson noise . . . 10
3.11 LoDoPaB-CT setting . . . 10
3.12 Dose model and simulation . . . 10
3.13 Methods and tuning in the CT study . . . 11
3.14 Lesion observer . . . 11
3.15 Statistics (Stage 3) . . . 11
3.16 Analysis history . . . 12

4 Experiments **12**
4.1 E1–E2: In-distribution accuracy under noise and undersampling . . . 12
4.2 E3: Forward-model mismatch . . . 16
4.3 E4: Distribution shift . . . 18
4.4 E5: Why PnP selects few iterations . . . 21
4.5 E6: Sensitivity to the inverse crime . . . 23
4.6 E7: Error decomposition along the singular vectors . . . 25
4.7 E8: Denoiser seeds and matched-prior control . . . 28
4.8 E9: Lesion insertion . . . 29
4.9 E10: Can re-tuning recover lesion detectability? . . . 31
4.10 E11: The matrix-free operator reproduces the dense pipeline . . . 33
4.11 E12: Poisson dose sweep . . . 36
4.12 E13: Lesion insertion under Poisson noise . . . 37
4.13 E14: Data, noise model and simulator . . . 39
4.14 E15: Learned prior versus TV on real slices at $K = 60$ . . . 42
4.15 E16: Accuracy over the dose × angle grid . . . 42
4.16 E17: Lesion detection . . . 45
4.17 E18: False structures on real slices . . . 45
4.18 E19: Distribution shift to non-anatomical phantoms . . . 47

5 Discussion **50**

6 Conclusion **55**

1 Introduction

Reconstructing a computed tomography (CT) image from few projection angles or at low photon counts is an ill-posed inverse problem. From measurements

$$y = Ax + n, \tag{1}$$

where A is the forward (Radon) operator and n the noise, the image x cannot be recovered stably when A is rank-deficient or ill-conditioned, as it is for sparse-view data, without prior knowledge

about x. The classical remedy is explicit regularization,

$$\hat{x} = \arg\min_x \ \tfrac{1}{2}\|Ax - y\|_2^2 + \lambda R(x), \tag{2}$$

with a hand-crafted penalty R such as Tikhonov or total variation (TV) [15].

Learned priors replace R by statistics estimated from data. End-to-end networks, including unrolled iterations, are trained for a given operator and noise level [1, 10]. Denoiser-based methods, Plug-and-Play (PnP) [18] and regularization by denoising [14] (reviewed in Kamilov et al. 12), instead insert a trained denoiser such as DnCNN [20] in place of the proximal operator of R in an iterative scheme. Learned methods can be much more accurate than classical reconstruction on large benchmarks: in the LoDoPaB-CT dataset paper, FBP followed by a U-Net improves the test PSNR from 30.5 to 35.8 dB over FBP [13]. We study PnP with a DnCNN prior because the prior is trained once, independently of the operator and of the noise level, so the same prior can be placed in the same data-fidelity iteration as the classical baselines and tested when the geometry, the dose or the image distribution changes. We make no claim about end-to-end networks trained for a specific task.

Reconstruction methods are usually compared by PSNR and SSIM on test images drawn from the distribution of the training set. Several concerns are known to be left open by such comparisons. Learned reconstructions can be unstable under small perturbations [2] and sensitive to distribution shift in accelerated MRI [8]; they can contain plausible structure that is not in the object [5]; and denoiser-based iterations with an unconstrained CNN denoiser come without convergence guarantees [7, 16]. Moreover, image quality for a diagnostic purpose is better measured by performance at a task than by a pixel-wise error [3, 4]. Finally, the verdict depends on the baseline: a regularizer whose weight and iteration count are not tuned for each condition is a weak comparison.

This paper is an evaluation study, not a new method. It asks three questions about a PnP-DnCNN prior compared with tuned FBP, Tikhonov and TV reconstruction:

Q1 Does the learned prior keep its advantage over TV when the image distribution changes?

Q2 Can we find cases where it creates convincing but incorrect structure?

Q3 Do the conclusions hold over a range of noise levels and reconstruction conditions?

We answer them at two scales. In a controlled setting of 28×28 digit images (MNIST-derived), every component can be varied and repeated: noise level, number of angles, forward-model error, denoiser seeds, and a prior trained on the shifted distribution as a control; the same questions are then examined with a matrix-free operator and Poisson noise. We then ask which conclusions survive on real 362×362 slices of LoDoPaB-CT with a patient-disjoint split, three angle counts and three dose levels. Throughout, the data are split into disjoint training, validation and test sets; every method is tuned on the validation set for each condition (and frozen for robustness tests); differences are paired and reported with confidence intervals; the measurements are simulated on a finer grid than the reconstruction grid to avoid the inverse crime [19]; and, besides PSNR and SSIM, we measure the detectability of synthetic lesions inserted into the images (paired AUC, lesion retention and false lesion-like blobs).

In the digit-image setting the prior beats TV by up to 3.4 dB over a box-constrained TV at moderate noise but loses to it, by 3.1 and 4.3 dB, when an MNIST-trained prior is applied to Fashion-MNIST or a Shepp–Logan phantom, and a matched prior restores the advantage, which points to prior mismatch rather than to the PnP scheme. On real CT the prior does not consistently beat a well-tuned TV in PSNR (-0.93 to $+0.33$ dB over nine conditions) but detects small lesions better than TV in all nine, without exceeding Hann-filtered FBP, whereas TV tuned for PSNR

detects them worse than FBP in all nine; under a shift to non-anatomical phantoms the detection advantage persists in one of six conditions.

Our contributions are:

- An evaluation protocol for comparing a PnP prior with classical regularization: per-condition tuning of every method, a TV control with the same iteration budget as the PnP scheme, paired statistics over test images or patients, and crime-free data generation with a sensitivity analysis (Section 4.5).
- Robustness evidence on digit images: forward-model mismatch, distribution shift with a matched-prior and a seed control, an analysis of why PnP-PGD needs early stopping, and an error decomposition along the singular vectors of the forward operator (Sections 4.5–4.7).
- A task-based lesion-insertion evaluation, with an operating-point study asking whether the loss in detectability can be removed by re-tuning (Sections 4.8 and 4.9).
- A matrix-free operator with exact adjoint, validated against the dense operator at 28×28, and a Poisson-noise study showing that the digit-image rankings and the shift reversal persist under it (Sections 4.10–4.12).
- A patient-disjoint study on LoDoPaB-CT over three angle counts and three doses, with the simulator checked against real observations and a shift to non-anatomical phantoms (Sections 4.13–4.18). The study reports negative and mixed results, a retracted early comparison, and the analysis steps taken after seeing results (Section 3.16); nothing was pre-specified.

The paper is organised as follows. Section 2 reviews related work, Section 3 describes the methods and the evaluation protocol, Section 4 reports experiments E1–E19, Section 5 answers the three questions, and the limitations and the reproducibility information close the paper.

2 Related Work

Variational regularization. Tikhonov regularization and total variation (TV) [15] remain standard baselines for ill-posed inverse problems because they are convex, well understood and simple to tune. TV is a strong prior for piecewise-constant images, which makes it a competitive baseline for digit-like data and a natural reference for CT.

Learned CT reconstruction. Deep networks have been used to post-process filtered back-projection images [10], to unroll iterative reconstruction [1], and, more recently, as generative priors for sparse-view CT and accelerated MRI [17]. On the LoDoPaB-CT benchmark, FBP followed by a U-Net improves the test PSNR from 30.5 to 35.8 dB over FBP [13]. These methods are trained for a given operator and noise level. We instead study the Plug-and-Play family, in which the prior is trained once and the data-fidelity step stays explicit, so that the prior can be compared with classical regularizers inside the same iteration; our conclusions concern this family and a small denoiser, not end-to-end networks.

Plug-and-Play and denoiser-based priors. Venkatakrishnan et al. [18] proposed replacing the proximal operator of a variational regularizer by an off-the-shelf denoiser inside an ADMM or proximal-gradient iteration; regularization by denoising [14] is a related formulation, and Kamilov et al. [12] review the field. CNN denoisers were introduced as plug-in priors by Zhang et al.

[21]. Convergence results require assumptions on the denoiser, such as boundedness or a Lipschitz constraint [7, 16]; an unconstrained CNN denoiser comes with no such guarantee, which is relevant to the peak-then-decay behaviour we observe in Section 4.

Noise-conditioned denoisers. DnCNN [20] uses residual learning for Gaussian denoising and, in its blind form, is trained over a range of noise levels. FFDNet [22] instead receives the noise level as an input. We use a DnCNN-style residual network with the noise level as an additional input channel, so that one network serves every denoising strength σ_d tried in the PnP iteration, where σ_d is a tuned hyperparameter.

Instabilities, robustness and distribution shift. Antun et al. [2] showed, for MRI and CT networks, that tiny perturbations can cause severe artefacts, that small structural changes may be missed, and that sparsity-based benchmarks including TV were unaffected in their examples. Darestani et al. [8] measured the robustness of trained and untrained reconstruction methods in accelerated MRI and found that both suffer similarly from distribution shifts and that the better-performing deep methods do not lose robustness. Our shifts are natural (another image class, a different kind of phantom) rather than adversarial, our prior is a small PnP denoiser, and the shifts are larger than within-domain variation; we report a reversal of the advantage over TV for a mismatched prior (Section 5), and we do not claim that it holds for other learned methods.

Hallucinations, the inverse crime and task-based assessment. Bhadra et al. [5] define hallucinations in tomographic reconstruction as errors in image components that the measurements cannot stably determine, which can come only from the prior. In a stylised 2D MRI setting they found that a U-Net produced more null-space hallucinations on out-of-distribution data, where its advantage over non-learned methods disappeared. We adopt this view and extend the decomposition to the weakly determined subspace of a finite-precision operator (Section 3.7). Our setting differs in the modality (CT), the prior (a PnP denoiser), the controls (a prior trained on the shifted distribution, denoiser seeds, a TV reference with the same iteration budget) and the evidence (a lesion-insertion test, and real data over a grid of doses and angle counts). Generating data with the discrete model that is later inverted is known as the inverse crime [11, 19] and tends to overstate the accuracy of methods that exploit the exact model; we remove it in Section 3.5. Finally, we follow the task-based view of image quality [3, 4]: fidelity metrics such as PSNR need not predict whether a feature of interest survives reconstruction, so we also measure the detection of inserted lesions with a location-known statistic and count lesion-like false blobs in lesion-free reconstructions.

s

3 Methodology

3.1 Problem setting and experimental settings

We reconstruct an image $x \in \mathbb{R}^n$ from measurements $y = Ax + n$, where A is a parallel-beam Radon transform with K projection angles evenly spaced in $[0°, 180°)$. The study uses three settings, summarised in Table 1; Stage 1 and 2 use synthetic 28×28 digit images so that every factor can be varied and repeated, and Stage 3 uses real CT slices. All methods in a comparison receive the same data, the same operator and the same validation-based tuning budget.

Table 1: The three settings of the study.

Setting	Images	Operator and noise	Data used	Experiments
Digits, Gaussian noise (Stage 1)	28×28 MNIST; Fashion-MNIST, EMNIST letters, Shepp–Logan for shift	dense Radon matrix, 40 bins, $K \in \{5, 10, 20, 40, 90\}$; additive Gaussian noise $\sigma \in \{0, 0.01, 0.05, 0.1, 0.2\}$ on the normalised sinogram	prior: MNIST training split (59 500 images); tuning: 100 validation images; test: 200 images	E1–E10
Digits, Poisson noise (Stage 2)	as above	matrix-free operator (same geometry); Poisson photon noise, $I_0 \in \{1000, 300, 100, 30, 10\}$	as above	E11–E13
LoDoPaB-CT (Stage 3)	362×362 slices	matrix-free parallel-beam, 513 bins, $K \in \{30, 60, 120\}$ of 1000 angles; Poisson noise, $N_0 = 4096$ photons per detector bin and fractions $1, \frac{1}{4}, \frac{1}{10}$ of it	patient-disjoint: 8 960 training slices, 20 validation slices, 40 test slices (one per patient)	E14–E19

3.2 Forward operator and classical baselines (Stages 1–2)

For the digit images, $A \in \mathbb{R}^{m \times n}$ is stored as an explicit dense matrix: the discrete Radon transform of scikit-image (`radon`, `circle=False`) is applied to each of the $n = 784$ canonical basis images and the sinograms are stacked as columns. The detector has 40 bins, so $m = 40K$ (800×784 for $K = 20$). A is divided by the image size 28 so that sinogram values are of order one; the Gaussian noise standard deviation σ is defined on this normalised sinogram, $y = Ax + \sigma\varepsilon$ with $\varepsilon \sim \mathcal{N}(0, I)$. The dense matrix gives the exact adjoint $A^\top$, which matters for the PnP gradient step, and allows the reconstruction operator to be perturbed independently of the data-generating operator in the mismatch experiments.

Filtered back-projection (FBP) uses the ramp filter of scikit-image (`iradon`); it has no tunable parameter and is not regularised. **Tikhonov regularization** solves

$$\hat{x} = \arg\min_x \tfrac{1}{2}\|Ax - y\|_2^2 + \tfrac{\lambda}{2}\|x\|_2^2 = (A^\top A + \lambda I)^{-1} A^\top y \tag{3}$$

in closed form. **Total variation** solves

$$\hat{x} = \arg\min_x \tfrac{1}{2}\|Ax - y\|_2^2 + \lambda\,\mathrm{TV}(x) \tag{4}$$

by ADMM with an isotropic TV penalty, forward differences D with Neumann boundary conditions, the splitting $z = Dx$, a fixed penalty parameter $\rho = 0.05$, 100 iterations and the x-update computed with a precomputed inverse. The Tikhonov and ADMM-TV solutions are not constrained to $[0, 1]$; all reconstructions in Stages 1–2 are clipped to $[0, 1]$ before the metrics are computed.

3.3 Learned prior: PnP proximal gradient

Denoiser. D_σ is a DnCNN-style residual network that also receives the noise level: the input has two channels, the image and a constant map equal to σ (as in FFDNet), followed by 8 convolutional

layers of 3×3 kernels with 64 channels, ReLU activations and batch normalisation in the hidden layers; the output is $x - \mathrm{net}(x, \sigma)$ (residual learning). For Stages 1–2 it is trained on clean MNIST images (the training split without 500 held-out images, 59 500 images) corrupted by Gaussian noise with $\sigma \sim \mathcal{U}(0, \sigma_{\max})$, $\sigma_{\max} = 0.3$, with the ℓ_2 loss, Adam (learning rate 10^{-3}), batch size 128 and 8 epochs with a cosine schedule. One trained network serves all PnP experiments; σ_d, the noise level passed to the network at inference, is a tuned hyperparameter that is distinct from the training noise range. At inference the network is in evaluation mode and its output is clipped to $[0, 1]$.

Iteration. Reconstruction replaces the proximal operator of an explicit regularizer by the denoiser,

$$x_{k+1} = D_{\sigma_d}\big(x_k - \eta\, A^\top (A x_k - y)\big), \qquad \eta = 1/L(A^\top A), \tag{5}$$

where $L(A^\top A)$ is the largest eigenvalue of $A^\top A$, estimated by power iteration. Algorithm 1 summarises the procedure; in Stages 1–2 the initial image is the FBP reconstruction. Both the initial image and every iterate are clipped to $[0, 1]$ (the denoiser output is clipped, which acts as a box constraint). The ADMM-TV and Tikhonov baselines of Stages 1–2 have no such constraint inside the iteration; as a control we re-ran the key digit-image comparisons with a box-constrained TV (Chambolle–Pock, $0 \le x \le 1$, 1000 iterations, λ tuned on the same validation set; *TV-box*). The control covers the noise and angle sweeps (E2b), the distribution-shift and matched-prior matrix (E4/E8) and the lesion study (E9); the other Stage 1–2 experiments (E3, E5, E7, E10 and the Stage 2 dose sweep) use the unconstrained ADMM-TV only. Every claim below states which TV it refers to. In Stage 3 the TV baseline is box-constrained.

Algorithm 1 PnP proximal gradient (PnP-PGD)

Require: measurements y, operator A, denoiser D, strength σ_d, step η, iterations K, initial image x_0

$x_0 \leftarrow \mathrm{clip}_{[0,1]}(x_0)$
for $k = 0, \dots, K-1$ **do**
  $g_k \leftarrow A^\top (A x_k - y)$
  $x_{k+1} \leftarrow \mathrm{clip}_{[0,1]}\big(D_{\sigma_d}(x_k - \eta\, g_k)\big)$
end for
return x_K

3.4 Evaluation protocol (Stages 1–2)

Data are split into disjoint training, validation and test subsets (Table 1). The hyperparameters, λ for Tikhonov and TV and (σ_d, K) for PnP, are selected by grid search maximising the mean validation PSNR, independently for each experimental condition (noise level, number of angles or dose) in E1, E2 and E12, and then evaluated once on the test set. The grids are $\lambda_{\mathrm{Tik}} \in 10^{[-5,0]}$ (21 log-spaced values), $\lambda_{\mathrm{TV}} \in 10^{[-5,-1]}$ (17 values), $\sigma_d \in \{0.005, 0.01, 0.02, 0.05, 0.1, 0.15, 0.2, 0.3\}$ (extended to 0.5 for the denoisers of Section 3.6) and $K \in \{0, 1, 2, 3, 5, 10, 25, 50, 100, 200, 400\}$, where $K = 0$ means one application of the denoiser to the FBP image. For the robustness experiments (E3, E4) the hyperparameters are tuned once at a nominal condition and *frozen* across all mismatch and shift conditions, as in a deployed pipeline that is not retuned for every operating condition. PSNR (peak 1) and SSIM (scikit-image, 7×7 window, data range 1) are computed on the clipped reconstruction. Every condition uses one noise realisation per image (validation and test noise are drawn with different seeds). We report the mean with a 95% interval of ± 1.96 standard errors over test images and use a two-sided paired Wilcoxon signed-rank test between PnP and TV on the same images and noise realisations.

3.5 Data simulated on a finer grid

In E1–E5 the data are generated with the matrix A used for reconstruction (an "inverse crime" [11, 19]). For the experiments of Sections 4.5–4.9 we instead simulate measurements on a finer grid. The 28×28 image x is upsampled by an odd integer factor $f = 5$ (to 140×140) with bicubic interpolation and clipped to $[0, 1]$, projected with the same discrete Radon transform at the same angles, and every f adjacent detector bins are summed onto the 40-bin detector of the reconstruction operator, giving $y = \mathcal{S}(x) + n$ with $\mathcal{S} \neq A$. Reconstructions use A, and all metrics use the original 28×28 image as ground truth. The model error $\|\mathcal{S}(x) - Ax\|_2$ averages 0.09 at 20 angles (0.05 at 5 and 0.17 at 90 angles), i.e. 6% of the noise norm $\sigma\sqrt{m}$, $m = 40\,n_{\text{angles}}$, at $\sigma = 0.05$ and 31% at $\sigma = 0.01$. Because bicubic upsampling smooths the object, we also report a nearest-neighbour variant (object unchanged, model error 0.03) and a bilinear variant (model error ≈ 0.10, a stronger smoothing) as a sensitivity band. All hyperparameters are re-tuned on validation data simulated in the same way.

3.6 Denoiser seeds, noise range and matched-prior control

Besides the single MNIST denoiser of Section 3.3 we train, with the same architecture and optimiser: (i) three MNIST denoisers with $\sigma_{\max} = 0.3$ (seeds 0–2); (ii) three MNIST denoisers with $\sigma_{\max} = 0.5$, which allows σ_d up to 0.5 and removes the grid-edge tuning at $\sigma = 0.2$ observed in E1; (iii) three denoisers trained on the Fashion-MNIST training split ($\sigma_{\max} = 0.3$). In the matched-prior control, all methods (Tikhonov, TV, PnP) are tuned on held-out Fashion-MNIST training images and tested on the Fashion-MNIST test split. The σ_d grid is extended to $\{0.005, \ldots, 0.3, 0.4, 0.5\}$ for all denoisers.

3.7 Weakly determined subspaces and prior-driven content

Let $A = U\Sigma V^\top$ be the SVD of the dense operator, with singular values $s_1 \geq \cdots \geq s_n$. For a threshold τ let $W_\tau = \operatorname{span}\{v_i : s_i < \tau s_1\}$ and let P_τ be the orthogonal projector onto W_τ. For 20 angles the operator has an exact null space of dimension 79 (singular values at machine precision), followed by a continuum of very small singular values with no gap; we therefore use $\tau = 0.05$ as the headline choice, giving $\dim W_{0.05} = 319$ of the 784 dimensions (and report $\tau = 0.10$ and the exact null space as sensitivity checks). A gradient step $x \mapsto x - \eta A^\top(Ax - y)$ never changes the component of x in the null space of A, so null-space content of a PnP reconstruction is supplied by the denoiser alone (up to the clipping to $[0, 1]$, which is nonlinear). For a reconstruction $\hat{x}$ and ground truth x we report

$$\|\hat{x} - x\|_2^2 = \|P_\tau(\hat{x} - x)\|_2^2 + \|(I - P_\tau)(\hat{x} - x)\|_2^2 \quad \text{(orthogonal split)}, \tag{6}$$

$$\iota = \frac{\sum \|P_\tau \hat{x}\|_2^2}{\sum \|P_\tau x\|_2^2} \quad \text{(injection)}, \tag{7}$$

$$c = \frac{\sum \langle P_\tau \hat{x}, P_\tau x\rangle}{\sqrt{\sum \|P_\tau \hat{x}\|_2^2 \sum \|P_\tau x\|_2^2}} \quad \text{(alignment)}, \tag{8}$$

with sums over test images (pooled ratios). $\iota > 1$ means that the reconstruction places more energy in W_τ than the true images contain there; c measures how much of that content points in the right direction. These statistics cannot by themselves distinguish prior-driven structure from amplified noise, so we interpret them only in comparisons with identical noise, data and solver settings.

3.8 Lesion-insertion protocol

A lesion is an additive Gaussian blob, $x_{\text{les}} = \text{clip}_{[0,1]}\big(x + c\, e^{-\|p-p_0\|^2/(2\sigma_l^2)}\big)$, inserted at a random integer location p_0 of the central 16×16 region of a test image, with $\sigma_l \in \{0.5, 0.75, 1.25, 2.0\}$ px and contrast $c \in \{0.25, 0.5\}$ (eight lesion types). We reconstruct the lesion-present image and the lesion-absent original with independent noise realizations ($\sigma = 0.05$, 20 angles, crime-free data), using the same locations for every method. A lesion is *valid* if its inserted contrast in the ground truth is at least 50% of nominal (a lesion placed on a saturated stroke is clipped away); invalid cases are excluded for all methods, leaving 157–197 of 200 images per type. The observer is a location-known statistic, the *local contrast*: the mean of the image in a disc of radius $\rho = \lceil 1.5\sigma_l \rceil$ minus its mean in the surrounding annulus of width 3 px. We report

- the *paired AUC*: the fraction of images for which the statistic is larger in the lesion-present than in the lesion-absent reconstruction of the same base image (the background cancels; ties count one half);
- the *retention* $R = \mathrm{E}[t(\hat{x}_{\text{present}}) - t(\hat{x}_{\text{absent}})]/\mathrm{E}[t(x_{\text{les}}) - t(x)]$, which is 1 if the lesion contrast is preserved and below 1 if it is attenuated;
- the number of *false lesions* per lesion-absent image: local maxima of the contrast map of the reconstruction whose response is at least θ, half the response of the nominal lesion, and for which the ground truth has no contrast above $\theta/2$ within 2 px;
- the lesion's *observability*, the fraction of its energy outside $W_{0.05}$.

For the operating-point study (Section 4.9) we sweep 45 PnP settings ($\sigma_d \in \{0.02, 0.05, 0.1, 0.15, 0.2\}$, $K \in \{1, 2, 3, 5, 10, 25, 50, 100, 200\}$) and 9 values of the TV weight, choose operating points on 200 held-out *in-distribution* validation images (disjoint from the images used for PSNR tuning) and report them on the test sets, so that the choice is neither made on the evaluation images nor informed by the shifted distribution.

3.9 Matrix-free operator

Up to this point the forward operator A was stored as a dense 800×784 matrix, which allows exact adjoints, closed-form Tikhonov inverses and a full singular value decomposition, but does not scale to clinical image sizes. We therefore implemented a matrix-free operator that is never assembled. The forward projection embeds the image in a 40×40 canvas (the diagonal of the image, as in the dense operator), rotates it by each angle with bilinear interpolation about the pixel used by the reference implementation, and sums along the detector rows. The adjoint is the exact transpose of this discretisation, obtained by reverse-mode differentiation, so $\langle Ax, y\rangle = \langle x, A^\top y\rangle$ holds to floating-point precision by construction rather than approximately. The classical baselines use only forward and adjoint applications: Tikhonov by conjugate gradients on $A^\top A + \lambda I$, and TV by the same ADMM as before with the x-update solved by conjugate gradients. The singular subspaces used in Section 3.7 are obtained with a randomized SVD [9] (power iterations and oversampling, forward and adjoint products only); the error split in Section 4.6 then uses the well-determined subspace $\{i : s_i/s_1 \geq 0.05\}$ and its complement. The matrix-free operator was validated against the dense one on the digit images only (E11); the CT study of Stage 3 uses a different matrix-free parallel-beam operator (Section 3.11).

3.10 Poisson noise

Low-dose CT measures photon counts. With $p = Ax$ the (normalised) line integrals, the expected counts are $I = I_0 \exp(-\mu p)$ with I_0 incident photons per ray and attenuation scale $\mu = 2$; the counts are $N \sim \text{Poisson}(I)$ and the data used for reconstruction are $\hat{p} = -\mu^{-1} \log(\max(N, 1)/I_0)$. The noise in $\hat{p}$ is signal dependent, with variance approximately $e^{\mu p}/(I_0 \mu^2)$, and is larger along rays that pass through more material. We use $I_0 \in \{1000, 300, 100, 30, 10\}$, which gives an effective noise standard deviation of $0.018, 0.032, 0.056, 0.106, 0.198$ on the sinogram and therefore spans the Gaussian range $\sigma \in [0.02, 0.2]$ of the earlier experiments. The data are generated from the finer-grid object (Section 3.5, bicubic, factor 5), so the inverse crime remains removed. The reconstruction methods are unchanged and treat $\hat{p}$ with an unweighted least-squares data term (a Gaussian approximation); for PnP we also tested a variant with weights $w = N/I_0$ (an inverse-variance weighting up to a constant). All hyperparameters are tuned at each dose on 100 held-out MNIST validation images and frozen for every test set.

3.11 LoDoPaB-CT setting

LoDoPaB-CT [13] provides 362×362 slices with values $\mu/\mu_{\max} \in [0, 1]$ and simulated low-dose observations of 1000 parallel-beam angles on 513 detector bins. The official split has 35 820 training slices from 632 patients, 3 522 validation slices from 60 patients and 3 553 test slices from 60 patients; we verified that the three sets of patient identifiers are disjoint. We use 8 960 training slices (from the 632 training patients) to train the prior, 20 validation slices, one from each of 20 patients, for all tuning, and 40 test slices, one from each of 40 patients (the same 40 in every experiment), for all evaluation; the patient is therefore the unit of statistical analysis. The projector is a matrix-free parallel-beam operator with a unit-tested adjoint; its conventions (rotation sign, detector flip, image transposition, angle offset) were fixed by projecting the ground truth and comparing with the supplied observations over 16 variants: the selected one has a relative residual of 0.038 (correlation 0.9979), the best alternative 0.166, and the fitted pixel size is 1.0044 times the nominal 0.26 m/362. Sparse-view data use $K \in \{30, 60, 120\}$ evenly spaced angles out of the 1000.

3.12 Dose model and simulation

The observations follow $y = -\ln(N/N_0)/\mu_{\max}$ with $N \sim \text{Poisson}(N_0 e^{-\mu_{\max} Ax})$, $N_0 = 4096$ photons per detector bin, $\mu_{\max} = 81.35858\,\text{m}^{-1}$, and zero counts replaced by 0.1. We checked this model on the data: the counts recovered from the observations are integers in 100% of the bins for $N_0 = 4096$ (and 8192), against 9.7–10.4% for perturbed values of $\mu_{\max}$ (chance level 10%) and 24.7% and 49.9% for $N_0 = 1024$ and 2048. The dispersion index $\overline{(N - \lambda)^2/\lambda}$ around the counts expected from the ground truth is 1.06, 1.25 and 1.81 for expected counts in $[20, 200)$, $[200, 1000)$ and $[1000, 5000)$; values above one include discretisation and ground-truth error, so the real data may contain somewhat more variance than the simulation. Lower doses $f \in \{1, 0.25, 0.1\}$ (4096, 1024 and about 410 photons per detector bin) are obtained from the real observations by exact binomial thinning of the recovered counts (a thinned Poisson variable is Poisson with f times the mean), which is used for *tuning*. For *evaluation* the data are simulated without the inverse crime: the slice is up-sampled by a factor two (bicubic), projected on the 724×724 grid with half-size pixels and 1026 detector bins, pairs of bins are averaged, and Poisson noise is added exactly as above. Section 4.13 compares reconstructions from simulated and from real data of the same slices.

3.13 Methods and tuning in the CT study

FBP-Hann: filtered back-projection with the ramp filter multiplied by a Hann window whose cut-off (a fraction of Nyquist) is tuned; it is a linear regularised reconstruction. *TV*: isotropic total variation with the box constraint $[0, 1]$ (forward differences, Neumann boundary), solved by the primal–dual algorithm of Chambolle and Pock [6] for 150 iterations from the FBP-Hann image. The problem is normalised by $\|A\|^2$, i.e. $\frac{1}{2\|A\|^2}\|Ax-y\|_2^2+\lambda\,\mathrm{TV}(x)$, so TV weights are in these normalised units and are not comparable with those of Stage 1. *TV-long* is the same with 300 iterations and its own tuned weight, a control for the extra iterations that PnP performs. *PnP-const*: Algorithm 1 with $\eta = 1/L$ (L by power iteration), started from the FBP-Hann image, with (σ_d, k) tuned; *PnP-TVinit*: the same started from the TV solution. D_σ is the noise-level-conditioned DnCNN of Section 3.3 (8 layers, 64 channels, σ as a second input channel, residual output), here trained on the 8 960 LoDoPaB slices for 15 000 steps (batch 16, random 128×128 crops and flips, $\sigma \sim U(0, 0.1)$, Adam, cosine schedule). A 17-layer variant (20 000 steps) was also trained and evaluated at $K = 60$. Every method and condition is tuned separately on the 20 validation slices (real observations, thinned and angle-subsampled) by maximising the mean PSNR (peak 1, reconstructions clipped to $[0, 1]$). The initial grids of the dose $\times$ angle study are cut-off $\{0.08, 0.12, 0.16, 0.2, 0.3, 0.4\}$, TV weight $\{10^{-3}, 3 \cdot 10^{-3}, 10^{-2}, 3 \cdot 10^{-2}\}$, $\sigma_d \in \{0.01, 0.02, 0.03, 0.05, 0.08\}$ and $k \in \{1, 2, 4, 8, 12, 20, 30, 45\}$ (the $K = 60$ steps used similar grids, see the scripts); whenever the optimum falls on the boundary of a grid, the grid is extended (up to twice; cut-off $\times 0.75$ or $+0.1$, TV weight $\times 3$ or $\div 3$, σ_d $\times 1.6$ or $\div 1.6$). The final choices are listed in Table 23.

3.14 Lesion observer

The protocol of Section 3.8 is applied to 362×362 slices. Lesions are additive Gaussian blobs, clipped to $[0, 1]$, with $\sigma_l \in \{2, 4\}$ px (1.4 and 2.9 mm at 0.72 mm per pixel) and contrast $c = 0.06$ on the $\mu/\mu_{\max}$ scale (in the step at $K = 60$ also $c = 0.03$). If water is taken to be about 0.25 on this scale, which is an assumption, $c = 0.06$ and 0.03 correspond to roughly 240 and 120 HU; the lesions are small and noisy at low dose but are not low-contrast in the clinical sense. Four lesions of one type are inserted into each slice at random positions at least $10\sigma_{l,\max} + 4 = 44$ px apart, inside the body eroded by 22 px (CT slices) or in the central square (phantoms); the lesion-free and lesion-present images are reconstructed from independent noise realisations. A lesion is valid if its inserted response is at least half the nominal one. The observer is the location-known, zero-sum local-contrast statistic of Section 3.8 (disc of radius $\lceil 1.5\sigma_l \rceil$ minus a 3-px annulus), and the detection score is the paired AUC pooled over valid lesions (up to $40 \times 4 = 160$ lesions per lesion type on CT slices, $20 \times 4 = 80$ on phantoms). A *false blob* is a local maximum (in a 5×5 window) of the response map of a lesion-*free* reconstruction whose response is at least $m\, r_{\mathrm{nom}}$ (r_{nom}: response to the nominal lesion on an empty background; $m \in \{0.5, 0.75, 1, 1.5\}$) and for which the truth has no response above half of $0.5\, r_{\mathrm{nom}}$ within 2 px. $m = 1$ and 1.5 count structures that a reader of this statistic would take for at least a full lesion. The statistic measures lesion-like response at the scale of the kernel; we did not run a reader study, so "convincing" is shorthand for this response.

3.15 Statistics (Stage 3)

All comparisons are paired: two methods are evaluated on the same patients and the same simulated data, and the difference is summarised by its mean over patients with a 95% percentile bootstrap interval over patients (2 000 resamples in the $K = 60$ lesion experiments, 5 000 elsewhere), with the same resamples for both methods. Lesion types are pooled with equal weight; per-type rows are exploratory. For the shift comparison, the CT and phantom differences are bootstrapped

independently (different images) and their difference is the *shift effect*. No correction for multiplicity is applied, so we report the numbers of cells with significant positive and negative differences and read the sign pattern across conditions rather than any single interval. Each condition uses one noise realisation, which limits the precision of a single cell (for AUC about ± 0.05); the 20 phantoms are the same in all conditions, so conditions are not independent.

3.16 Analysis history

The analysis was developed in steps, and some of them changed the conclusions. We list them because they affect how the final numbers should be read. (i) A first comparison at $K = 60$ (tuning on 6 validation slices, Hann cut-off 0.2) found PnP 0.85 and 0.33 dB below TV at full and quarter dose. Re-tuning on 20 slices (cut-off 0.3 at full dose) raised PnP-const by 1.06 dB (30.68 $\rightarrow$ 31.74) and 0.32 dB (29.40 $\rightarrow$ 29.72) but TV by only 0.38 and 0.09 dB, and the gap became -0.17 and -0.10 dB (Table 24); the initial gap was an artefact of the starting image and the small tuning set, and we do not use it. (ii) The lesion experiments were first tuned on 10 slices with narrow grids; they were repeated with 20 slices, extended grids and a separately tuned TV-long. At full dose the parameters of FBP, TV, TV-long and PnP-const did not change (the PSNR of PnP-TVinit changed by 0.07 dB) and no pooled verdict changed; at quarter dose one of 20 pooled verdicts changed (PnP-TVinit $-$ TV-long AUC, from significant to not significant). (iii) Paired statistics, the threshold sweep and the false-blob definition at several strengths were added after seeing the first lesion results, and the dose $\times$ angle grid and the shift extension to $K = 30, 120$ were run afterwards; none of this was pre-specified.

4 Experiments

4.1 E1–E2: In-distribution accuracy under noise and undersampling

Setting. Test images are 200 MNIST digits and all hyper-parameters are tuned on 100 separate validation images (Section 3.1). The noise sweep fixes 20 projection angles and varies the measurement noise $\sigma \in \{0, 0.01, 0.05, 0.1, 0.2\}$; the angle sweep fixes $\sigma = 0.05$ and uses $5, 10, 20, 40, 90$ angles. Each cell is one noise realisation. Tables 2 and 3 first use data generated with the reconstruction operator itself ("inverse crime"), because this is the setting of the original study; Table 3 then repeats the PnP–TV comparison with data simulated on a finer grid (bicubic, factor 5; the default crime-free setting, Section 4.5). Two total-variation baselines are reported: *TV-ADMM* is the unconstrained ADMM solver used throughout Stages 1–2, and *TV-box* is the same functional solved with the constraint $0 \le x \le 1$ by Chambolle–Pock (1000 iterations, λ re-tuned on the validation images). TV-box is the fairer comparison, because the PnP iterates are clipped to $[0, 1]$ at every iteration; it is a control added after the main experiments (E2b), and the other Stage 1–2 experiments (E3, E5, E7, E10 and the Stage 2 dose sweep) were run against TV-ADMM only.

Results. *(i) Where the learned prior wins.* For $\sigma \ge 0.05$ PnP-DnCNN has the highest PSNR at every angle count and noise level tested, and the gap over TV is significantly positive against both TV variants and on both data models (Table 3). The gap is largest at $\sigma = 0.05$ with 20–40 angles: +3.8 to +3.9 dB against TV-ADMM, but +3.3 dB against TV-box on inverse-crime data and +3.4 to +3.5 dB on crime-free data. We take the box-constrained, crime-free value as the headline (≈ 3.5 dB for this denoiser; 3.4 dB averaged over three independently trained denoisers in E8). Against TV-ADMM, PnP has the higher PSNR in 100% of the 200 images at $\sigma = 0.05$ with 20, 40 and 90 angles, in 96% at $\sigma = 0.1$, but only in 68% at $\sigma = 0.2$ and 87% at 5 angles, where the

Table 2: Test PSNR (dB), mean over 200 images, data generated with the reconstruction operator (inverse crime). Bold: highest mean in the row (not a significance statement; see Table 3 for paired intervals). †: the TV-box optimum lies below the initial λ grid ($\lambda_{\text{norm}} = 3.7 \times 10^{-7}$, i.e. almost no regularisation), so 42.5 dB is a data-model artefact and should not be read as a property of TV.

Condition	FBP	Tikhonov	TV-ADMM	TV-box	PnP-DnCNN
σ=0, 20 angles	24.2	28.9	36.0	**42.5**†	29.6
σ=0.01, 20 angles	23.1	24.1	29.0	**29.8**	29.1
σ=0.05, 20 angles	15.5	17.1	20.1	20.7	**24.0**
σ=0.1, 20 angles	10.4	14.7	16.9	17.2	**19.9**
σ=0.2, 20 angles	6.6	12.9	14.3	14.6	**15.3**
5 angles, σ=0.05	9.1	14.2	15.5	16.5	**17.7**
10 angles, σ=0.05	12.1	15.9	18.4	18.9	**21.4**
40 angles, σ=0.05	18.1	18.2	21.8	22.4	**25.6**
90 angles, σ=0.05	20.6	20.0	24.1	24.6	**27.1**

Table 3: PnP−TV PSNR gap (dB) with 95% bootstrap intervals over the 200 test images; positive means PnP is better. Left: data generated with the reconstruction operator (the PSNR values of Table 2). Right: data simulated on a finer grid, which removes the inverse crime. Hyper-parameters are re-tuned for each data model. †: see Table 2.

	Inverse crime		Crime-free (bicubic)	
Condition	vs TV-ADMM	vs TV-box	vs TV-ADMM	vs TV-box
σ=0, 20 angles	-6.43 $[-6.70, -6.18]$	-12.93 $[-13.30, -12.57]$†	-3.29 $[-3.47, -3.10]$	-5.13 $[-5.32, -4.94]$
σ=0.01, 20 angles	$+0.07$ $[-0.07, +0.21]$	-0.75 $[-0.89, -0.60]$	$+1.02$ $[+0.87, +1.16]$	$+0.35$ $[+0.21, +0.49]$
σ=0.05, 20 angles	$+3.86$ $[+3.70, +4.02]$	$+3.33$ $[+3.19, +3.47]$	$+4.02$ $[+3.85, +4.18]$	$+3.48$ $[+3.34, +3.63]$
σ=0.1, 20 angles	$+3.01$ $[+2.79, +3.24]$	$+2.65$ $[+2.43, +2.88]$	$+3.10$ $[+2.88, +3.33]$	$+2.73$ $[+2.51, +2.96]$
σ=0.2, 20 angles	$+1.03$ $[+0.78, +1.29]$	$+0.65$ $[+0.40, +0.89]$	$+1.08$ $[+0.83, +1.33]$	$+0.69$ $[+0.45, +0.93]$
5 angles, σ=0.05	$+2.13$ $[+1.86, +2.41]$	$+1.14$ $[+0.90, +1.39]$	$+2.36$ $[+2.07, +2.64]$	$+1.35$ $[+1.10, +1.61]$
10 angles, σ=0.05	$+3.03$ $[+2.81, +3.26]$	$+2.51$ $[+2.28, +2.75]$	$+3.29$ $[+3.09, +3.50]$	$+2.77$ $[+2.56, +2.98]$
40 angles, σ=0.05	$+3.83$ $[+3.69, +3.97]$	$+3.26$ $[+3.13, +3.39]$	$+3.97$ $[+3.82, +4.11]$	$+3.39$ $[+3.26, +3.52]$
90 angles, σ=0.05	$+3.04$ $[+2.90, +3.18]$	$+2.57$ $[+2.44, +2.69]$	$+3.24$ $[+3.10, +3.39]$	$+2.79$ $[+2.67, +2.91]$




Figure 1: Reconstruction accuracy (PSNR, left; SSIM, right) against the measurement noise σ at 20 projection angles, inverse-crime data, TV-ADMM as the TV baseline. At $\sigma = 0$ TV is clearly best in PSNR; at $\sigma = 0.01$ TV and PnP-DnCNN are within noise of each other in PSNR (Table 3), while PnP has the higher SSIM; for $\sigma \geq 0.05$ PnP-DnCNN is best in both. Error bars are the mean ± 1.96 standard errors over 200 test images and reflect image-to-image variability only, not tuning or training variability.




Figure 2: Reconstruction accuracy (PSNR, left; SSIM, right) against the number of projection angles at $\sigma = 0.05$ (logarithmic axis), inverse-crime data, TV-ADMM as the TV baseline. PnP-DnCNN is best at every angle count tested. Error bars as in Figure 1.




Figure 3: PnP−TV PSNR gap on crime-free (bicubic) data against the unconstrained TV-ADMM and the box-constrained TV-box. Bars are 95% bootstrap intervals over 200 images. The box constraint lowers the PnP advantage by 0.4–1.0 dB at $\sigma \geq 0.05$ and, in the angle sweep, most at 5 angles; it does not change the sign anywhere except at $\sigma = 0.01$ on inverse-crime data (Table 3).

average gap is small (+1.0 and +2.1 dB). Against TV-box the gap shrinks fastest in the hardest regimes: from +2.1 to +1.1 dB at 5 angles and from +1.03 to +0.65 dB at $\sigma = 0.2$ (inverse-crime data).

(ii) SSIM. The SSIM gain is larger than the PSNR gain suggests (TV-ADMM 0.74 → PnP 0.94 at $\sigma = 0.05$ with 20 angles; 0.50 → 0.79 at 5 angles; 0.41 → 0.68 at $\sigma = 0.2$) and PnP has the higher SSIM, significantly so (paired Wilcoxon), in every condition except $\sigma = 0$, where TV is significantly higher (0.996 vs. 0.982). We did not compute SSIM for TV-box. SSIM is later shown to be lower for PnP on real CT (Stage 3); we attribute the high SSIM here to the match between the denoiser's training distribution and the test digits, but did not test this.

(iii) The noise-free and low-noise end depends on the data model and the TV solver. At $\sigma = 0$ the learned prior is below TV in every variant (−6.4 dB on inverse-crime data, −3.3 dB crime-free against TV-ADMM; larger still against TV-box). At $\sigma = 0.01$ the sign of the PnP–TV gap changes with these choices: not distinguishable from zero against TV-ADMM on inverse-crime data (+0.07 dB), −0.75 dB against TV-box on the same data, and +0.35 dB against TV-box on crime-free data. No claim about PnP versus TV at $\sigma \leq 0.01$ should therefore be made from E1–E2 alone; Section 4.5 analyses the data-model dependence.

(iv) Classical baselines. TV (either variant) is the best classical method in every condition, and Tikhonov is above FBP except at 90 angles (Table 2). A possible reason for TV's advantage is that digit images are close to piecewise constant; we did not test this.

Caveats. (a) Each cell uses one noise realisation, a single denoiser and one tuning run; the intervals in Table 3 describe the variability across test images only. The only measure of training variability is E8 (three denoisers), where the in-distribution gap varies by 0.03 dB between seeds, small compared with the gaps above. (b) PnP has two tuned parameters (denoiser level σ_d and iteration count k, $5 \leq k \leq 100$ across conditions) against one for TV, and early stopping is itself a regulariser. The TV baselines are run to convergence (1000 iterations for TV-box; 100 for TV-ADMM), so the comparison is between a tuned few-iteration PnP and a converged TV; the inclusion of TV-box addresses the box constraint but not the difference in tuning degrees of freedom.

(c) All numbers are for 28×28 digits whose training distribution matches the test distribution by construction; what happens when it does not is the subject of E4 and E8.




Figure 4: In-distribution reconstructions at $\sigma = 0.05$, 20 angles (inverse-crime data, one noise draw). Rows are the first four test images (no selection). Columns: ground truth, FBP, Tikhonov, TV-ADMM, PnP-DnCNN and the absolute error $|\hat{x}_{\mathrm{PnP}} - x|$ (brighter is larger). FBP and Tikhonov retain visible noise; TV-ADMM gives flat regions with reduced stroke intensity; PnP-DnCNN is visually closest to the ground truth. The error maps show the residual concentrated along stroke boundaries. A TV-box column and per-image PSNR values are not shown.

4.2 E3: Forward-model mismatch

Setting. Measurements are generated with the true 20-angle operator at $\sigma = 0.05$ (inverse-crime data, 200 test digits), and the reconstruction uses an operator whose projection angles are perturbed by $\delta \in \{0, 0.5, 1, 2, 4\}^\circ$ (the angular spacing is 9°). Two perturbations are used: a *constant offset*, $\theta_j \to \theta_j + \delta$, and *per-angle jitter*, $\theta_j \to \theta_j + \delta\,\varepsilon_j$ with $\varepsilon_j \sim \mathcal{N}(0,1)$, so that δ is the standard deviation, not a bound. One jitter vector is drawn per level (fixed seed) and shared by all images and methods. All hyper-parameters are frozen at their values for the nominal, matched operator (TV-ADMM $\lambda = 5.6 \times 10^{-3}$; PnP $\sigma_d = 0.1$, $k = 5$; Tikhonov $\lambda = 1.8 \times 10^{-2}$) and are not re-tuned under mismatch. The TV baseline is the unconstrained TV-ADMM; E3 was not repeated with TV-box.

A constant offset is a rotation. Data generated at the angles θ_j are consistent with an operator at the angles $\theta_j + \delta$ if the image is rotated by δ about the centre. A reconstruction that is perfect up to this rotation is therefore the best a method can do, yet a pixel-wise metric against the true image

Table 4: Test PSNR (dB) under forward-model mismatch with hyper-parameters frozen at their nominal values, and the PnP−TV gap (mean ±1.96 standard errors of the paired per-image difference, 200 images). PnP has the higher PSNR in 100% of the images in every row except $\delta = 4°$ (96% for the offset, 99% for the jitter).

	Constant offset			Per-angle jitter (δ = s.d.)		
δ	TV	PnP	PnP−TV	TV	PnP	PnP−TV
0°	20.15	24.00	$+3.86 \pm 0.16$	20.15	24.00	$+3.86 \pm 0.16$
0.5°	20.13	23.91	$+3.78 \pm 0.17$	20.14	23.99	$+3.85 \pm 0.17$
1°	20.04	23.62	$+3.58 \pm 0.16$	20.10	23.93	$+3.83 \pm 0.16$
2°	19.71	22.61	$+2.90 \pm 0.16$	19.99	23.65	$+3.66 \pm 0.16$
4°	18.56	20.05	$+1.49 \pm 0.14$	19.19	22.18	$+2.98 \pm 0.19$
Loss at 4°	1.6	4.0		1.0	1.8	




Figure 5: Test PSNR under forward-model mismatch as a function of the angular perturbation δ (degrees): constant offset (left) and per-angle jitter with standard deviation δ (right), TV-ADMM as the TV baseline. Hyper-parameters are frozen at their nominal values and are not re-tuned. PnP-DnCNN has the highest mean PSNR at every level, and loses more than TV in absolute terms under both perturbations; the offset curve mainly reflects a rigid rotation of the reconstruction (Table 5). Error bars: mean ±1.96 standard errors over 200 test images, which reflect image variability only (a single jitter draw is used).

penalises it. To separate the two effects we scored each reconstruction against the true digit and against the true digit rotated by δ (Table 5; the opposite direction gives lower agreement). Against the rotated digit PnP keeps its nominal accuracy (24.1–24.2 dB at all offsets, against 24.0 dB at $\delta = 0$), so the 4.0 dB loss in Table 4 is registration error, not a loss of reconstruction quality; TV also loses nothing against the rotated digit (it scores 20.4–21.0 dB against 20.15 dB nominal; part of this is because the interpolated reference is smoother than the true digit). The sharper reconstruction has more to lose under a pixel-wise metric: even a perfect reconstruction rotated by 4° would score only 22.1 dB, and PnP's 20.05 dB is already close to this ceiling.

Table 5: Constant offset: PSNR (dB) of the reconstruction against the true digit and against the true digit rotated by δ (cubic interpolation). The last column is the PSNR of the true digit rotated by δ against itself, the best score a perfect but rotated reconstruction can reach. Computed with `run_e3_rotation_check.py`; the nominal values reproduce the stored E3 results.

	TV		PnP		
δ	vs true	vs rotated	vs true	vs rotated	Ceiling
0°	20.15	20.15	24.00	24.00	∞
1°	20.04	20.36	23.62	24.14	35.9
2°	19.71	20.66	22.61	24.23	29.0
4°	18.56	20.99	20.05	24.14	22.1

Results. *(i) Per-angle jitter is the genuine inconsistency.* With a jitter of 4° (s.d.) PnP loses 1.8 dB and TV 1.0 dB; the advantage falls from +3.86 to +2.98 dB and PnP stays ahead on 99% of the images. *(ii) Constant offset.* In raw PSNR PnP loses 4.0 dB against 1.6 dB for TV and the advantage falls to +1.49 dB (still positive on 96% of the images). Table 5 shows that this is a registration effect, so it should not be read as evidence that PnP is the less robust method to a rigid geometric error; it is evidence that PSNR penalises a sharp but shifted reconstruction more. *(iii) What is not shown.* Neither method collapses within the range tested, but both TV-ADMM and the PnP parameters are frozen, so the experiment measures the sensitivity of nominally tuned methods, not what a re-tuned method would achieve.

Caveats. (a) One jitter draw per level and one noise realisation; the intervals describe image variability only. (b) Only the projection angles are perturbed (no detector offset, scale or beam-hardening errors), at $\sigma = 0.05$, on inverse-crime data. (c) Frozen hyper-parameters: a mismatch acts like extra noise, so re-tuning might favour a larger σ_d for PnP and a larger λ for TV; we did not test this. (d) TV-ADMM is unconstrained; the TV-box control was not run for E3.

4.3 E4: Distribution shift

Setting. The denoiser is trained on MNIST digits only. The complete pipeline, with hyper-parameters tuned once on the MNIST validation images at $\sigma = 0.05$ and 20 angles (TV-ADMM $\lambda = 5.6 \times 10^{-3}$; PnP $\sigma_d = 0.1$, $k = 5$) and then frozen, is applied to 200 Fashion-MNIST test images, 200 EMNIST letters, and a 28×28 Shepp–Logan phantom. The phantom is a single image, so its 20 test cases are 20 independent noise draws and its interval describes noise variability only. Data are generated with the reconstruction operator (inverse crime) with a different noise realisation from E1, so the MNIST row differs slightly from Table 2. The labels "near" and "far" shift are our a priori judgement; no distance between the distributions was computed.

Table 6: Test PSNR and SSIM under distribution shift with the pipeline frozen at its MNIST values (one MNIST-trained denoiser, TV-ADMM). Gap = PnP−TV PSNR, mean ±1.96 standard errors of the paired difference (over images; over noise draws for the phantom); "wins" is the share of test cases in which PnP has the higher PSNR.

	PSNR (dB)			SSIM	
Test set	TV	PnP	gap (wins)	TV	PnP
MNIST (in-distribution)	20.27	**23.97**	$+3.70 \pm 0.16$ (100%)	0.742	**0.944**
EMNIST letters (near shift)	20.09	**22.03**	$+1.94 \pm 0.18$ (92%)	0.842	**0.938**
Fashion-MNIST (far shift)	**19.61**	16.57	-3.05 ± 0.31 (11%)	0.680	0.675
Shepp–Logan (far shift)	**19.19**	15.03	-4.16 ± 0.19 (0%)	**0.722**	0.517




Figure 6: Test PSNR under distribution shift for the four methods with the pipeline frozen at its MNIST values. PnP-DnCNN leads on MNIST and on EMNIST letters, which share the stroke-like statistics of digits, but is below TV on Fashion-MNIST (−3.0 dB) and on the phantom (−4.2 dB). Against Tikhonov it is 0.2 dB below on Fashion-MNIST and 2.1 dB below on the phantom; against FBP it is 1.6 dB above on Fashion-MNIST and only 0.3 dB above on the phantom. Error bars: mean ±1.96 standard errors; for the phantom they describe 20 noise draws of one image, not variation between images.

Table 7: Control with a box-constrained TV (TV-box, $0 \le x \le 1$, 1000 iterations, λ tuned on the MNIST validation images). Three independently trained MNIST denoisers (E8), bicubic fine-grid data, $\sigma = 0.05$, 20 angles. Mean test PSNR over the three denoisers and gap = PnP−TV, mean ± s.d. over denoisers.

Test set	TV-ADMM	TV-box	PnP	PnP−TV-ADMM	PnP−TV-box
MNIST	20.21	20.69	24.05	$+3.84 \pm 0.03$	$+3.36 \pm 0.03$
EMNIST letters	20.28	20.35	22.07	$+1.79 \pm 0.09$	$+1.72 \pm 0.09$
Fashion-MNIST	19.42	19.45	16.32	-3.10 ± 0.30	-3.13 ± 0.30
Shepp–Logan	19.28	19.50	14.96	-4.33 ± 0.12	-4.55 ± 0.12




Figure 7: Reconstructions at $\sigma = 0.05$, 20 angles, with the frozen pipeline: the first test image of each set (no selection), one noise draw, TV-ADMM and the single MNIST-trained PnP denoiser of Table 6. On Fashion-MNIST and the phantom the PnP output contains spotted or stroke-like fragments that are absent from the ground truth; the TV output is blurred and blocky but does not introduce such structure. A TV-box column and PSNR values are not shown.

Results. *(i) The advantage does not transfer to far-shifted images.* On EMNIST letters PnP keeps about half of its in-distribution advantage (+1.9 against +3.7 dB). On Fashion-MNIST and the phantom it is below TV, in 89% and 100% of the test cases respectively (Table 6). The reversal is not an artefact of the unconstrained TV: with a box-constrained TV the gaps are −3.13 and −4.55 dB, slightly larger than against TV-ADMM, because the box constraint helps TV on the phantom by 0.2 dB and on Fashion-MNIST by 0.03 dB (Table 7). The in-distribution advantage, in contrast, loses 0.5 dB to the box constraint (+3.84 → +3.36 dB). *(ii) Variability.* Across three independently trained MNIST denoisers the Fashion-MNIST gap has a standard deviation of 0.30 dB and the phantom gap 0.12 dB (Table 7); the single denoiser of Table 6 gives similar values (−3.05 and −4.16 dB, on a different data model) and the sign never changes. *(iii) PSNR and SSIM disagree on Fashion-MNIST.* PnP is 3.1 dB below TV in PSNR but equal in SSIM (0.675 against 0.680), while on the phantom both are clearly lower (0.517 against 0.722). We did not examine why; a plausible reading is that the PnP error on Fashion-MNIST is mainly in smooth intensity levels that SSIM tolerates, which is untested. *(iv) Structure rather than blur.* In Figure 7, TV fails by blurring, whereas PnP adds structure not present in the image. This is a qualitative observation on one image per set. It is quantified with inserted lesions in Section 4.8: for the MNIST-trained prior on Fashion-MNIST, PnP produces 0.38 false lesions per lesion-absent image, against 0.14 for TV-box and 0.05 for TV-ADMM (means over three denoisers and eight lesion types). The converse experiment, in which a prior trained on Fashion-MNIST is tested on Fashion-MNIST, is in Section 4.7.

Caveats. (a) The pipeline is frozen at MNIST-tuned values. The deficit therefore combines a mismatched prior with mismatched hyper-parameters, and we did not test whether re-tuning σ_d and k on the target set would reduce it; the matched-prior control of E8 changes the prior and the tuning set together. (b) The phantom is one image, so the "far shift" evidence on it is one object; Fashion-MNIST supplies the only far-shift evidence over many images. (c) Table 6 uses inverse-crime data and Table 7 crime-free data; the conclusions agree, but the numbers are not interchangeable. (d) Only three datasets are used, so no claim about how the deficit scales with the distance between distributions can be made.

4.4 E5: Why PnP selects few iterations

Setting. Validation tuning selects few iterations: $k = 5$ at $\sigma \leq 0.05$ with 20 or 90 angles, 10 at $\sigma = 0.1$ or 40 angles, 25 at $\sigma = 0.2$ or 10 angles, and 100 at 5 angles (E1–E2). To see what happens beyond the selected k we ran the tuned pipeline ($\sigma = 0.05$, 20 angles, inverse-crime data, 200 test digits) with $\sigma_d \in \{0.02, 0.05, 0.1, 0.2\}$ and recorded the iterate at $k \in \{1, 2, 5, 10, 25, 50, 100, 200, 400\}$ (so the location of a peak is resolved only to this grid). For each iterate we computed the PSNR, the data residual $\|Ax_k - y\|$ relative to the noise norm $\sigma\sqrt{m}$, and the squared error in bands of the singular values s_i of A relative to the largest, s_i/s_1 (the decomposition of Section 3.7). The tuned setting is $(\sigma_d, k) = (0.1, 5)$.

Results. *(i) Early stopping is necessary at every σ_d tested.* For $\sigma_d = 0.1$ the PSNR peaks at $k = 5$ (23.97 dB) and falls to 22.09 dB at $k = 400$ (Figure 8, left). The decay is stronger for larger σ_d ($\sigma_d = 0.2$: 23.28 dB at $k = 2$ to 19.64 dB at $k = 400$, below TV's 20.27 dB) and weaker for smaller ones. Among the four σ_d the best converged setting is $\sigma_d = 0.05$ at 22.78 dB: 1.2 dB below the early-stopped optimum, but still 2.5 dB above TV. Early stopping is therefore worth about 1 dB here and is not what makes PnP better than TV in distribution.




Figure 8: PnP iterates on 200 MNIST test digits ($\sigma = 0.05$, 20 angles). Left: PSNR against the iteration count k for four denoiser strengths; the circle marks the tuned setting and the dashed line is TV-ADMM. Middle: data residual relative to the expected noise norm $\sigma\sqrt{m}$. Right: squared error in bands of s_i/s_1 for $\sigma_d = 0.1$ (the bands contain 319, 46, 253, 129 and 37 of the 784 directions, in the order of the legend). Every point is one run with one noise draw.

(ii) The data residual stays at the noise level. For $\sigma_d = 0.1$ the ratio $\|Ax_k - y\|/(\sigma\sqrt{m})$ moves from 0.994 at the selected iterate to 1.019 at $k = 400$ (at most 1.08 for $\sigma_d = 0.2$) (Figure 8, middle). The iterates are data-consistent to within a few percent of the noise level at every k, and the residual changes by 2.5% between the best and the converged iterate; a discrepancy-principle stopping rule could not pick $k = 5$ from it. The earlier reading that the denoiser's prior increasingly dominates the data term is not supported by these data: the iteration converges, and it converges to a point that fits the data as well as the early iterates do but has a larger error.

(iii) Where the error grows. Between $k = 5$ and $k = 400$ ($\sigma_d = 0.1$) the squared error rises from 3.36 to 5.52. Of this increase of 2.16, 74% lies in the directions with $0.1 \le s_i/s_1 < 0.3$ (382 of 784 directions), 17% in the directions below 0.1 (including the exact null space), 10% in $0.3 \le s_i/s_1 < 0.5$, and none in the directions with $s_i/s_1 \ge 0.5$, where the error does not grow (Figure 8, right). The gradient step contracts the error in direction i by $(1-(s_i/s_1)^2)^k$ (Section 4.6), i.e. by 0.82 for $s_i/s_1 = 0.2$ after 5 iterations, 0.36 after 25 and 0.02 after 100. These directions therefore become determined by the data at about the iterations where the PSNR declines (10 to 100), which is what classical semi-convergence of iterative regularisation predicts: further iterations fit the noise in poorly determined directions. This is consistent with the data, not a test of the mechanism. In particular, the tuned k is also 5 at $\sigma = 0$ (E1), so noise fitting cannot be the only cause; we did not record the iteration curve on noise-free data.

Table 8: PSNR (dB) of PnP at the tuned setting ($\sigma_d = 0.1$, $k = 5$), run to $k = 400$ with the same σ_d, and at the best (σ_d, k) over the stored grid ($\sigma_d \in \{0.02, \dots, 0.2\}$, $k \in \{1, \dots, 400\}$), selected on the test images themselves and therefore optimistic. Same frozen pipeline, noise draws and TV-ADMM as E4; the phantom is one image with 20 noise draws.

Test set	TV-ADMM	PnP $k = 5$	PnP $k = 400$	best (σ_d, k)	FBP
MNIST (in-distribution)	20.27	23.97	22.09	23.97 (0.1, 5)	15.53
EMNIST letters	20.09	22.03	20.11	22.03 (0.1, 5)	15.47
Fashion-MNIST	19.61	16.57	13.80	17.86 (0.2, 1)	14.94
Shepp–Logan	19.19	15.03	12.40	16.36 (0.1, 1)	14.77

(iv) Under shift, stopping early limits the damage but cannot remove it. On the two shifted sets the PSNR is highest at $k = 1$, and run to convergence PnP is 5.8 dB (Fashion-MNIST) and 6.8 dB (phantom) below TV and below FBP (Table 8). Even the best setting of the grid stays 1.8 and 2.8 dB below TV. The in-distribution benefit of early stopping and the shift deficit may share a cause, namely that more iterations let the prior shape the result further, but this is an interpretation and was not tested.

Caveats. (a) Single noise draw, one denoiser, four σ_d values, and checkpoints rather than every iteration (the peak lies between $k = 5$ and 10; a per-iteration curve on 50 images in an earlier version of this figure peaked at about $k = 5$ for $\sigma_d = 0.1$). (b) The band decomposition is reported for $\sigma_d = 0.1$ only. (c) The "best" setting in Table 8 is chosen on test data. (d) Inverse-crime data; Section 4.6 repeats the analysis on crime-free data.

4.5 E6: Sensitivity to the inverse crime

Setting. E1–E5 generate the data with the reconstruction matrix A. Here the measurements are instead simulated on a five-times finer grid with three object resamplings (nearest, bicubic, bilinear; Section 3.5), and compared with the inverse-crime data. The mean model error $\|\mathcal{S}(x) - Ax\|_2$ is 0, 0.03, 0.09 and 0.10 for the inverse crime, nearest, bicubic and bilinear data (20 angles); for comparison, the noise norm $\sigma\sqrt{m}$ is 0.28, 1.41, 2.83 and 5.66 for $\sigma = 0.01, 0.05, 0.1, 0.2$. With crime-free data the case $\sigma = 0$ is therefore not noise-free: the only error in the measurements is the model error. Hyper-parameters of all methods are re-tuned for each data model. The PnP−TV gaps below are against TV-ADMM unless stated; there is one noise draw and one denoiser per cell.

Table 9: PnP−TV PSNR gap (dB), 20 angles, 200 test digits, for four data models. TV is TV-ADMM except in the last two rows, which use the box-constrained TV-box (E2b control, only run for the inverse-crime and bicubic data). †: the TV-box optimum lies below the initial λ grid (Table 2).

Data model	Model error	$\sigma = 0$	0.01	0.05	0.1	0.2
Same operator (inverse crime)	0	−6.43	+0.07	+3.86	+3.01	+1.03
Fine grid, nearest	0.03	−5.86	+0.16	+3.86	+3.01	+1.02
Fine grid, bicubic (default)	0.09	−3.29	+1.02	+4.02	+3.10	+1.08
Fine grid, bilinear	0.10	−0.97	+1.76	+3.94	+2.87	+0.85
Inverse crime, vs TV-box	0	−12.93†	−0.75	+3.33	+2.65	+0.65
Bicubic, vs TV-box	0.09	−5.13	+0.35	+3.48	+2.73	+0.69

Table 10: PnP−TV-ADMM PSNR gap (dB) in the angle sweep ($\sigma = 0.05$) for the four data models. The mean model error of the bicubic data is 0.05, 0.07, 0.09, 0.12, 0.17 at $5, 10, 20, 40, 90$ angles.

Data model	5	10	20	40	90
Same operator (inverse crime)	+2.13	+3.03	+3.86	+3.83	+3.04
Fine grid, nearest	+2.19	+3.04	+3.86	+3.84	+3.06
Fine grid, bicubic (default)	+2.36	+3.29	+4.02	+3.97	+3.24
Fine grid, bilinear	+2.17	+3.23	+3.94	+4.04	+3.57

Table 11: As Table 2, with measurements simulated on the finer grid (bicubic, $f = 5$); mean test PSNR (dB) over 200 digits. The last column is the PnP−TV-ADMM gap. Intervals and TV-box results are in Table 3.

Condition	FBP	Tikhonov	TV-ADMM	PnP-DnCNN	PnP−TV
$\sigma = 0$, 20 angles	24.3	27.4	33.2	29.9	−3.29
$\sigma = 0.01$, 20 angles	23.2	23.9	28.4	29.4	+1.02
$\sigma = 0.05$, 20 angles	15.5	17.1	20.1	24.1	+4.02
$\sigma = 0.1$, 20 angles	10.4	14.7	16.8	19.9	+3.10
$\sigma = 0.2$, 20 angles	6.6	12.9	14.3	15.3	+1.08
5 angles, $\sigma = 0.05$	9.1	14.1	15.5	17.8	+2.36
10 angles, $\sigma = 0.05$	12.1	15.9	18.2	21.5	+3.29
40 angles, $\sigma = 0.05$	18.1	18.2	21.8	25.8	+3.97
90 angles, $\sigma = 0.05$	20.7	20.0	24.1	27.3	+3.24




Figure 9: PSNR versus noise level (left) and number of angles (right) for data simulated on the bicubic finer grid (solid) and with the inverse crime (dashed), TV-ADMM as the TV baseline. The curves nearly coincide for $\sigma \geq 0.05$ and separate at $\sigma \leq 0.01$, where TV loses 2.8 dB at $\sigma = 0$ (36.0 → 33.2 dB) while PnP gains 0.3 dB.

Results. *(i) The conclusions at $\sigma \geq 0.05$ do not depend on the data model.* The model error is at most 7% of the noise norm there, and the PnP−TV-ADMM gap differs between the inverse-crime and the bicubic data by at most 0.16 dB in the noise sweep (Table 9) and 0.26 dB in the angle sweep (Table 10); between the inverse-crime and the bilinear data it differs by up to 0.53 dB (90 angles). Against TV-box the gap is 0.4 to 0.5 dB smaller than against TV-ADMM on both data models but keeps its sign (Table 9, last rows).

(ii) The low-noise end depends on the data model, and also on the TV solver. At $\sigma = 0$ TV is ahead of PnP in every variant, by 6.4 dB (inverse crime), 5.9 (nearest), 3.3 (bicubic) and 1.0 dB (bilinear); against TV-box the lead is larger (12.9 dB, with the TV-box optimum below the λ grid, and 5.1 dB). At $\sigma = 0.01$ the PnP lead against TV-ADMM ranges from +0.07 to +1.76 dB across the four data models, and against TV-box it is −0.75 dB (inverse crime) or +0.35 dB (bicubic). The ranking at $\sigma \leq 0.01$ is therefore not a single number: it changes with the data model by up to 5.5 dB at $\sigma = 0$ and changes sign with the choice of TV solver and data model at $\sigma = 0.01$. We make no claim about PnP versus TV at $\sigma \leq 0.01$ from the digit experiments. A possible reading is that TV, tuned for $\sigma \approx 0$, fits the data almost exactly and is penalised when the data contain model error, whereas PnP is not; we did not test this.

(iii) Out-of-distribution deficits persist. With the pipeline of E4 the PnP−TV-ADMM gap is −3.05 (inverse crime), −3.04, −2.82 and −3.88 dB on Fashion-MNIST, and −4.16, −4.16, −4.21 and −5.37 dB on the phantom, for the four data models in the order of Table 9; the in-distribution EMNIST advantage stays between +1.8 and +1.9 dB. The deficits are largest for the bilinear data, whose object is the smoothest.

(iv) The default data model. We use the bicubic data as the default crime-free setting for the later experiments. The choice was made after seeing all four variants, because its low-noise results lie between those of the nearest and the bilinear data; it is a convention, not a derived optimum, and the results at $\sigma \leq 0.01$ should be read as the band spanned by the variants.

Caveats. (a) One noise draw and one denoiser per cell; hyper-parameters are re-tuned for each data model, so part of the differences between rows is tuning. (b) The model error is a deterministic resampling difference, not the unknown mismatch of a real scanner; the real-data setting is addressed in Stage 3 (Section 4.13). (c) TV-box was run for two of the four data models; the other experiments of this section use TV-ADMM.

4.6 E7: Error decomposition along the singular vectors

Setting. We apply the decomposition of Section 3.7 to the shift setting of E4 on crime-free (bicubic) data: the MNIST-trained denoiser with the pipeline tuned on MNIST and frozen, $\sigma = 0.05$, 20 angles, 200 test images per set (the phantom: 20 noise draws of one image). The weakly determined subspace is $W_{0.05}$ ($s_i < 0.05\, s_1$, $\dim W_{0.05} = 319$ of 784, containing the 79-dimensional exact null space); $W^{\perp}$ is its complement. We report the squared error in each part as the ratio PnP/TV-ADMM (below 1: PnP has less error), with the injection ι and alignment c statistics. Paired Wilcoxon tests on the per-image errors give $p < 10^{-10}$ for every ratio of the three 200-image sets in Table 12; for the phantom, with 20 noise draws, $p = 1.9 \times 10^{-6}$ is the smallest value attainable.

Results. *(i) In distribution the advantage is not confined to the unobserved directions.* On MNIST PnP has 0.48 times TV's error in $W_{0.05}$ and 0.40 times in its complement, and its content in $W_{0.05}$ is better aligned with the truth ($c = 0.76$ against 0.45). The same holds for EMNIST letters (0.70 and

Table 12: Error decomposition at $\sigma = 0.05$, 20 angles, bicubic fine-grid data, $W_{0.05}$ (dim = 319). Ratios are PnP/TV-ADMM (below 1: PnP has less error); ι is the injection and c the alignment statistic of Section 3.7.

	error ratio PnP/TV			ι		c	
Test set	in W	in $W^{\perp}$	total	TV	PnP	TV	PnP
MNIST (in-dist.)	0.48	0.40	0.42	0.58	0.76	0.45	0.76
EMNIST letters	0.70	0.64	0.65	1.22	1.55	0.43	0.67
Fashion-MNIST	1.27	2.32	2.05	0.29	0.80	0.36	0.37
Shepp–Logan	2.59	2.64	2.63	0.44	3.27	0.09	0.23

Table 13: Sensitivity of Table 12 to the subspace: the exact null space (dim = 79), $W_{0.05}$ (319) and $W_{0.10}$ (365). Error ratio PnP/TV in the subspace and PnP injection ι.

	exact null		$W_{0.05}$		$W_{0.10}$	
Test set	ratio	ι	ratio	ι	ratio	ι
MNIST (in-dist.)	0.49	0.72	0.48	0.76	0.47	0.76
EMNIST letters	0.68	1.56	0.70	1.55	0.70	1.56
Fashion-MNIST	1.21	0.70	1.27	0.80	1.27	0.83
Shepp–Logan	2.64	3.68	2.59	3.27	2.21	2.74




Figure 10: Injection (horizontal axis: energy placed in $W_{0.05}$ divided by the true energy there) against alignment (vertical axis: pooled cosine with the truth in $W_{0.05}$) for each method and test set (bicubic fine-grid data). Points to the right of the dotted line place more energy in $W_{0.05}$ than the true images contain; low alignment means that this energy does not point in the true direction. Only PnP-DnCNN on the phantom is both loud and poorly aligned.

Table 14: Total error energy ratio PnP/TV-ADMM as a function of the iteration count, with $\sigma_d = 0.1$ and the same frozen pipeline: at the tuned $k = 5$, at the best k of the grid $\{1, 2, 5, 10, 25, 50, 100, 200, 400\}$ (selected on the test images, hence optimistic, with the selected k in brackets) and at $k = 400$. Left: data with the inverse crime (σ_d grid of E5); right: bilinear fine-grid data.

	Inverse crime			Bilinear fine grid		
Test set	$k = 5$	best k	$k = 400$	$k = 5$	best k	$k = 400$
MNIST (in-dist.)	0.43	0.43 (5)	0.71	0.43	0.42 (5)	0.65
EMNIST letters	0.64	0.64 (5)	1.01	0.66	0.62 (5)	0.93
Fashion-MNIST	2.16	1.56 (2)	4.13	2.65	1.47 (2)	3.71
Shepp–Logan	2.60	1.92 (1)	4.77	3.44	1.89 (1)	4.58

0.64). Since a gradient step never changes the null-space component, the null-space error reduction (ratio 0.49 in the exact null space, Table 13) is entirely due to the denoiser.

(ii) Under shift the excess error is mostly in the observed directions. On Fashion-MNIST the ratios are 1.27 in $W_{0.05}$ and 2.32 in $W^{\perp}$, on the phantom 2.59 and 2.64. Because $W^{\perp}$ contains most of the error energy, 93% (Fashion-MNIST) and 89% (phantom) of PnP's excess error over TV lies outside $W_{0.05}$. The failure under shift is therefore not confined to the weakly determined subspace. These conclusions hold for the exact null space and for $W_{0.10}$ (Table 13); the phantom injection falls from 3.7 to 2.7 with the threshold but stays above 1 throughout.

(iii) The injection–alignment signature separates the two shifted sets. On the phantom PnP places 3.3 times the true energy in $W_{0.05}$ (TV: 0.44) with an alignment of only 0.23 (TV: 0.09; Figure 10), whereas on MNIST its injection of 0.76 comes with an alignment of 0.76. On Fashion-MNIST the injection is 0.80 with alignment 0.37, no different from TV: PnP's excess error there lies in $W^{\perp}$ and is not visible in this statistic. The statistic does not by itself distinguish prior-driven structure from amplified noise (Section 3.7); the evidence that the phantom injection is driven by the prior is the comparison of E8, where denoisers trained on Fashion-MNIST give an injection of 0.30–0.35 against 2.8–3.4 for MNIST-trained ones, with identical noise and solver.

(iv) Iterations. The gradient step, ignoring the denoiser, contracts the error in direction i after k iterations by $(1 - (s_i/s_1)^2)^k$: at $k = 5$ a direction with $s_i = 0.1\, s_1$ retains 95% of its initial error, $s_i = 0.3\, s_1$ 62% and $s_i = 0.5\, s_1$ 24%, so the tuned iteration leaves a large share of directions to the denoiser. On the shifted sets no setting recovers TV's accuracy: even the best k at $\sigma_d = 0.1$ has 1.5–1.9 times TV's error, and the error grows with k to 3.7–4.8 times at $k = 400$ (Table 14). In distribution the converged iterate ($k = 400$) remains better than TV on MNIST (0.65–0.71) and is on par with it on EMNIST letters (0.93–1.01). Section 4.4 shows where along the singular-value axis the error grows with k.

Caveats. (a) All statistics are for one denoiser, one noise draw per image and one data model per table; the phantom is a single object. (b) The threshold $\tau = 0.05$ is a convention (there is no gap in the singular values); Table 13 shows that the conclusions do not depend on it, but the numbers do. (c) The error ratio is a ratio of means of squared errors; the Wilcoxon tests are paired over images. (d) Table 14 uses two data models other than the one of Table 12 (no bicubic iteration sweep was run), and the best k is selected on the test images.

4.7 E8: Denoiser seeds and matched-prior control

This experiment asks whether the distribution-shift results of E4 depend on the particular denoiser that was trained, and whether the reversal on Fashion-MNIST is caused by the prior itself or by something else in the pipeline. Table 15 repeats the shift experiment with three independently trained denoisers per configuration (bicubic data, $\sigma = 0.05$, 20 angles). Three families are used: *m03* (trained on MNIST, $\sigma_{\max} = 0.3$, seeds 0–2), *m05* (MNIST, $\sigma_{\max} = 0.5$) and *f03* (Fashion-MNIST, $\sigma_{\max} = 0.3$). Within each run all methods are tuned on the validation split of the family's own training distribution, then frozen; the noise-level grid for σ_d was extended to 0.5.

Seed variation. Seed-to-seed variation of the gap is at most 0.3 dB (standard deviation over three seeds), an order of magnitude below the effects of interest. On Fashion-MNIST the MNIST-trained prior is 3.10 ± 0.30 dB below TV-ADMM (range -3.43 to -2.83) and on the phantom 4.33 ± 0.12 dB below. Training with $\sigma_{\max} = 0.5$ instead of 0.3 leaves these results essentially unchanged (Fashion-MNIST -3.12 ± 0.31, phantom -3.94 ± 0.02 dB). With three seeds the standard deviations are themselves imprecise, so we read them as an indication of scale rather than as interval estimates. The sign of every shift effect in Table 15 is the same for all seeds.

Noise-level grid. At $\sigma = 0.2$ the tuned σ_d is 0.3 for all six MNIST-trained denoisers even though values up to 0.5 were available, so the choice of σ_d in E1 was not limited by the edge of its grid. The wider-range denoiser is better at high noise ($+1.44 \pm 0.11$ versus $+0.97 \pm 0.11$ dB over TV at $\sigma = 0.2$, ranges not overlapping) and slightly worse at $\sigma = 0.05$ ($+3.86$ versus $+4.01$ dB in the noise sweep, and $+3.68 \pm 0.13$ versus $+3.84 \pm 0.03$ dB in Table 15). The in-distribution gap at $\sigma = 0.05$ therefore depends on $\sigma_{\max}$ by a few tenths of a dB; the headline gap of about 3.4 dB over box-constrained TV (Sec. 4, E1–E2) uses the m03 family.

Matched-prior control. The control isolates the cause of the reversal. A prior trained on Fashion-MNIST beats TV on Fashion-MNIST by $+2.21 \pm 0.01$ dB (total-error ratio 0.63), although by less than the MNIST prior on MNIST ($+3.84$ dB), consistent with Fashion-MNIST images being less sparse. The control is clean in the following sense: the tuned hyperparameters are identical in all nine runs ($\sigma_d = 0.1$, $k = 5$, $\lambda_{\mathrm{TV}} = 5.6 \times 10^{-3}$), including the three tuned on Fashion-MNIST validation data, and the TV-ADMM PSNR is therefore identical for every prior on every test set. The only difference between the MNIST-trained and the Fashion-trained pipelines is thus the denoiser weights. Transfer is asymmetric: the Fashion-trained prior still beats TV on MNIST ($+1.38$ dB) and EMNIST letters ($+0.44$ dB) and is indistinguishable from TV on the phantom (-0.18 ± 0.08 dB, error ratio 1.05), whereas the MNIST-trained prior loses by 4.3 dB on the phantom. The injection statistic ι (Sec. 3.7) on the phantom follows the prior: 2.8–3.4 for MNIST-trained denoisers versus 0.30–0.35 for Fashion-trained ones, with identical noise and solver settings; on Fashion-MNIST itself it is 0.75–0.83 versus 0.40–0.42. Whether the broader prior helps because it is richer or because Fashion-MNIST resembles the phantom more cannot be decided with two training distributions.

Box-constrained TV. For m03 and f03 we repeated the comparison against TV with the constraint $0 \le x \le 1$ (E2b control; m05 was not rerun). The gaps move by at most 0.5 dB and no gap changes sign: m03 $+3.36$ (MNIST), $+1.72$ (EMNIST), -3.13 (Fashion), -4.55 dB (phantom); f03 $+1.44$, $+0.38$, $+2.17$, -0.05 dB. The constraint helps TV most on in-distribution MNIST (TV-box 20.69 versus 20.21 dB for the MNIST-tuned pipeline), which is the source of the reduction of the in-distribution gap from 3.84 to 3.36 dB. The Fashion-tuned TV-box values differ slightly

from the MNIST-tuned ones because its weight was re-tuned on Fashion-MNIST validation data. The qualitative conclusions of this subsection (reversal under shift, recovery with a matched prior, asymmetric transfer) hold for both TV variants.

Table 15: Distribution shift with three denoisers per prior. Mean test PSNR in dB over seeds; gap = PnP−TV, mean ± s.d. over seeds, against TV-ADMM and against box-constrained TV (TV-box; m03 and f03 only). Bicubic fine-grid data, $\sigma = 0.05$, 20 angles; each prior's pipeline is tuned on the validation set of its own training distribution.

	TV		MNIST-trained (m03)			Fashion-trained (f03)		
Test set	ADMM	box	PnP	gap	gap (box)	PnP	gap	gap (box)
MNIST	20.21	20.69	24.05	$+3.84 \pm 0.03$	$+3.36$	21.60	$+1.38 \pm 0.09$	$+1.44$
EMNIST letters	20.28	20.35	22.07	$+1.79 \pm 0.09$	$+1.72$	20.73	$+0.44 \pm 0.09$	$+0.38$
Fashion-MNIST	19.42	19.45	16.32	-3.10 ± 0.30	-3.13	21.63	$+2.21 \pm 0.01$	$+2.17$
Shepp–Logan	19.28	19.50	14.96	-4.33 ± 0.12	-4.55	19.10	-0.18 ± 0.08	-0.05

The TV-box column is the MNIST-tuned TV-box used with m03. The f03 gap (box) is computed against the Fashion-tuned TV-box (20.16, 20.34, 19.46, 19.15 dB, top to bottom). m05 gives gaps of $+3.68 \pm 0.13$, $+1.64 \pm 0.02$, -3.12 ± 0.31, -3.94 ± 0.02 dB (same order of test sets).




Figure 11: Left: PnP−TV-ADMM gap for each denoiser seed (dots) on the four test sets, for three priors. Right: gap in the in-distribution noise sweep for MNIST-trained denoisers with $\sigma_{\max} = 0.3$ and 0.5. Produced by `summarize_step4.py` from `results/step4/`.

4.8 E9: Lesion insertion

Table 16 and Figure 12 summarise the lesion experiment (three denoisers per prior, protocol of Section 3.8; 157–197 valid images per lesion type). TV denotes the ADMM implementation used throughout the paper; TV-box is the same comparison with the constraint $0 \le x \le 1$, re-tuned on the same validation split (m03 and f03 families, same lesion locations and noise seeds). Three observations hold across lesion types.

TV detects small lesions better than PnP, in distribution and for a matched prior. For the MNIST-trained prior on MNIST the paired AUC of PnP is 0.04–0.23 below TV's (for example 0.69 versus 0.91 for $\sigma_l = 0.75$, $c = 0.5$), although PnP is about 4 dB better in PSNR on the same images; for the matched Fashion-trained prior the deficit is 0.03–0.17. The two methods tie only for the

largest high-contrast lesion. Against TV-box the deficit is the same size (0.00–0.22 on MNIST, 0.03–0.15 for the matched prior), so the box constraint does not explain it. FBP and Tikhonov also detect small lesions better than PnP (paired AUC 0.92 at $\sigma_l = 0.75$, $c = 0.5$ on MNIST for FBP), at the cost of many false blobs (up to 6 per image for FBP).

In distribution PnP attenuates lesions and invents none. Its retention is 0.12–0.78 for the MNIST prior on MNIST and 0.16–0.49 for the matched prior (TV: 0.29–0.88 and 0.38–0.83; TV-box: 0.31–0.92 and 0.33–0.82), and it produces at most 0.01 false lesions per image against up to 0.26 for TV and, for the MNIST-tuned TV-box, up to 0.49 (smallest low-contrast lesion; 0.075 for the Fashion-tuned TV-box). The reconstructions are cleaner than TV's but the inserted lesion is often absent (Figure 13), so a visually convincing image does not imply a faithful one.

Under prior mismatch PnP invents structure, but box-constrained TV does too for the smallest lesions. For the MNIST-trained prior on Fashion-MNIST the retention is 0.78–1.04 (the lesion survives) but there are 0.20–0.66 false lesions per image against ≤ 0.20 for TV, and the paired AUC remains below TV's by 0.05–0.16. The matched Fashion-trained prior on the same images produces at most 0.01. The box-constrained TV is a stricter comparison here: its false-lesion rate on the shifted images is 0.005–0.48 per image, and for the smallest low-contrast lesion type ($\sigma_l = 0.5$, $c = 0.25$) it exceeds PnP's (0.48 versus 0.38). For the other seven lesion types PnP's rate is higher, by a factor of 1.4 to 50. The claim that a mismatched learned prior invents more structure than the classical regulariser therefore holds for TV-box on seven of eight lesion types, not on all of them. Lesion observability (Section 3.8) does not explain these differences: the observability terciles formed within each lesion type have mean observability 0.88, 0.92 and 0.94, and PnP's paired AUC does not increase with observability (on MNIST it falls from 0.82 in the lowest to 0.67 in the highest tercile); smaller lesions are harder for all methods, but this is confounded with size.

Table 16: Lesion insertion, mean over three denoisers per prior (200 test images, $\sigma = 0.05$, 20 angles, bicubic data), for two of the eight lesion types (σ_l/c = lesion size in px / contrast). pAUC: paired AUC; Ret.: retention; FL: false lesions per lesion-absent image; TV-box: TV with $0 \leq x \leq 1$. All methods use their PSNR-tuned settings.

		pAUC			Ret.			FL		
Prior → test set	σ_l/c	TV	box	PnP	TV	box	PnP	TV	box	PnP
MNIST → MNIST (in-dist.)	0.75/0.5	0.91	0.90	0.69	0.59	0.65	0.28	0.005	0.010	0.000
	1.25/0.25	0.90	0.90	0.69	0.62	0.67	0.30	0.005	0.010	0.000
MNIST → Fashion (shifted)	0.75/0.5	0.91	0.92	0.82	0.63	0.74	0.99	0.030	0.090	0.430
	1.25/0.25	0.93	0.91	0.76	0.59	0.69	0.87	0.030	0.095	0.428
Fashion → Fashion (in-dist.)	0.75/0.5	0.91	0.90	0.75	0.63	0.55	0.22	0.030	0.025	0.005
	1.25/0.25	0.93	0.93	0.80	0.59	0.54	0.23	0.030	0.025	0.005
Fashion → MNIST (shifted)	0.75/0.5	0.91	0.90	0.75	0.59	0.55	0.35	0.005	0.000	0.000
	1.25/0.25	0.90	0.90	0.78	0.62	0.60	0.36	0.005	0.000	0.000

Scope. The observer is a location-known local-contrast statistic and the lesions are additive Gaussian blobs on 28×28 digits, so these results say how each *reconstruction* preserves or invents a simple signal, not how a human or a model observer would perform on clinical images. False-lesion counts depend on the response threshold θ, which scales with the nominal lesion, so they are compared across methods within a lesion type, not across types.




Figure 12: Paired detection AUC (top) and false lesions per lesion-absent image (bottom) for TV and PnP-DnCNN in the four prior/test-set combinations and the eight lesion types (mean over three denoisers; bars: s.d. over seeds). TV here is the ADMM implementation; TV-box values are in Table 16 and `results/control/`.

4.9 E10: Can re-tuning recover lesion detectability?

The PSNR-tuned operating point of PnP is not necessarily the best one for lesion detection, so we sweep operating points (Section 3.8) and compare three rules for PnP—PSNR-tuned, best paired AUC on the selection set ("lesion"), and best paired AUC among points whose selection-set PSNR is at least that of tuned TV ("constrained")—with TV. All rules are applied on held-out in-distribution validation images and then reported on the test sets, so no point is chosen on the evaluation images or with knowledge of the shifted distribution. This experiment uses one denoiser per prior (seed 0 of the m03 and f03 families), the ADMM implementation of TV, and the two lesion types $\sigma_l = 0.75$, $c = 0.5$ and $\sigma_l = 1.25$, $c = 0.25$; neither seed variation nor the box-constrained TV was examined here. Table 17 and Figures 14 and 15 show the results.

MNIST-trained prior, in distribution. The constrained point keeps a +0.7 dB PSNR advantage over TV (the PSNR-tuned advantage was +4.0 dB) but its paired AUC is still 0.06 below TV's (for both lesion types), and its false-lesion rate rises to 12 times TV's (0.060 versus 0.005 per image). The lesion-tuned point matches TV's detection (0.916 versus 0.910) but has a lower PSNR than TV (19.6 versus 20.1 dB) and about 60 times TV's false-lesion rate (0.32 versus 0.005 per image). Tuning TV for detection instead changes it very little (PSNR 20.13 dB, pAUC 0.904), so TV has no comparable trade-off in this grid. As an upper bound that uses the test set, the best of the 45 PnP settings among those with PSNR at least TV's reaches a paired AUC of 0.898 at 0.28 false lesions per image, so no rule could have closed the gap for this prior without a large false-lesion cost.

Fashion-trained prior. In distribution (Fashion-MNIST) the constrained point keeps +1.0 dB over TV with a paired AUC 0.02–0.03 below TV's and 0.085 false lesions per image against 0.03 for TV. A difference of this size is comparable to the binomial standard error of a paired AUC with 170–190 valid images ($\approx$ 0.02–0.03); we did not compute a paired interval, so we describe it as close




Figure 13: Matched Fashion-trained prior on Fashion-MNIST ($\sigma_l = 0.75$, $c = 0.5$). Rows: ground truth with lesion, TV and PnP reconstructions of the lesion-present image; ground truth, TV and PnP reconstructions of the lesion-absent image. The red circle marks the lesion location. PnP produces the cleaner images but removes the lesion.

to TV's, not as equal. On the shifted MNIST images the constrained point keeps +1.0 dB, with a paired AUC 0.02–0.06 below TV's and 0.015 false lesions per image against 0.005.

Under mismatch. Operating points chosen in distribution do not protect the MNIST-trained prior on Fashion-MNIST: all three rules are 2.1–2.8 dB below TV in PSNR and produce 0.40–0.51 false lesions per image against 0.03. This is not an artefact of the selection rule: none of the 45 PnP settings in the grid reaches TV's PSNR on the shifted images (the best is 17.96 versus 19.37 dB), even when judged on the test set itself.

Summary. PSNR, detectability and fabricated structure trade against each other for PnP, and only the Fashion-trained prior leaves a point at which PnP stays above TV in PSNR with a detection close to TV's and few false lesions. This holds for it on Fashion-MNIST and on MNIST, i.e. for one in-distribution and one shifted test set, but for the MNIST-trained prior it holds on neither test set. The difference between the priors is therefore the breadth of the training distribution (E8), not whether a test set is formally in-distribution. With one denoiser per prior this is a statement about two networks.

Table 17: Operating points for lesion type $\sigma_l = 0.75$, $c = 0.5$ on the test sets (one denoiser per prior; points chosen on held-out in-distribution validation images). PSNR is that of the lesion-absent reconstruction. "TV, lesion" is TV with the weight that maximises paired AUC on the selection set.

Prior → test set	Operating point	PSNR (dB)	pAUC	FL
MNIST → MNIST (in-dist.)	TV (PSNR-tuned)	20.11	0.910	0.005
	TV, lesion	20.13	0.904	0.005
	PnP, PSNR-tuned	24.07	0.693	0.000
	PnP, constrained	20.84	0.849	0.060
	PnP, lesion	19.57	0.916	0.315
MNIST → Fashion (shifted)	TV (PSNR-tuned)	19.37	0.914	0.030
	TV, lesion	19.33	0.924	0.035
	PnP, PSNR-tuned	16.58	0.822	0.420
	PnP, constrained	17.32	0.886	0.405
	PnP, lesion	16.78	0.903	0.505
Fashion → Fashion (in-dist.)	TV (PSNR-tuned)	19.37	0.914	0.030
	TV, lesion	18.94	0.914	0.010
	PnP, PSNR-tuned	21.58	0.751	0.005
	PnP, constrained	20.41	0.897	0.085
	PnP, lesion	19.07	0.870	0.255
Fashion → MNIST (shifted)	TV (PSNR-tuned)	20.11	0.910	0.005
	TV, lesion	19.06	0.916	0.000
	PnP, PSNR-tuned	21.67	0.759	0.000
	PnP, constrained	21.13	0.855	0.015
	PnP, lesion	21.21	0.898	0.040

4.10 E11: The matrix-free operator reproduces the dense pipeline

Table 18 compares the two implementations on the same 200 noisy test images ($\sigma = 0.05$, 20 angles, bicubic fine-grid data). The comparison uses fixed, untuned settings (Tikhonov and TV weight 10^{-2}; PnP with $\sigma_d = 0.1$ and $k = 25$), so its PSNR values are not comparable to the




Figure 14: PSNR of the lesion-absent reconstruction versus paired detection AUC over all operating points of PnP (colour: σ_d; marker size: iterations) and TV (line: weight sweep) for the MNIST-trained denoiser. Markers show the selected operating points. Left: in distribution; right: Fashion-MNIST (shifted).




Figure 15: As Figure 14 for the Fashion-trained denoiser. Left: MNIST (shifted); right: Fashion-MNIST (in distribution). Both figures are produced by `run_oppoint.py`.

tuned values of E1–E2; only the agreement between the two implementations matters here. The projections agree to 3×10^{-7} in relative norm, and the three reconstruction methods agree to $\leq 10^{-5}$ (Tikhonov, TV) and 3×10^{-4} (PnP) in relative norm, with identical PSNR to three decimals. The slightly larger PnP difference is plausible given that the denoiser is applied 25 times to slightly different iterates, but we did not isolate its source.

The randomized SVD reproduces the singular values of the dense SVD to within $1.1\times10^{-4}\,s_1$ for the first 400 and $6.2\times10^{-5}\,s_1$ for the first 600 (the number used to build the subspace in E7). It finds the same 465-dimensional well-determined subspace $\{i : s_i/s_1 \geq 0.05\}$ (so 319 directions are weakly determined). The resulting strong/weak error split differs from the dense one by at most 2.3×10^{-4} of the error norm for each of the three methods (Tikhonov, TV-ADMM, PnP), far below the 1% tolerance of the unit test; the strong subspace holds 91%, 82% and 77% of the squared error of the three reconstructions, respectively. These checks are produced by `run_e11_svd_check.py` and are limited by single-precision arithmetic on the GPU.

At this size the dense implementation is much faster for the iterative solvers (Table 18): the matrix-free TV-ADMM performs 100 ADMM iterations with up to 30 conjugate-gradient steps each, i.e. up to 3×10^3 forward–adjoint pairs per image, and the adjoint is evaluated by reverse-mode differentiation. PnP is only 1.2 times slower because the denoiser dominates its cost. The timings are single runs on one GPU (RTX 3050 Ti) and should be read as orders of magnitude. The matrix-free operator is therefore a requirement for larger images, where the dense matrix cannot be stored, rather than an optimisation at 28×28.

Table 18: Dense versus matrix-free pipeline (200 images, fixed untuned settings, single GPU run). Rel. diff.: relative difference between the two reconstructions. Peak GPU memory for the whole comparison: 184 MB.

Method	Rel. diff.	PSNR (dense = MF)	Time dense (s)	Time MF (s)
Tikhonov	4.6×10^{-6}	16.36	0.08	0.26
TV-ADMM	9.7×10^{-6}	19.01	0.16	7.50
PnP ($k=25$)	2.8×10^{-4}	23.10	0.76	0.94

4.11 E12: Poisson dose sweep

Table 19 and Figure 16 give the PnP–TV difference over the dose range (200 test images per set; the Shepp–Logan phantom is one image with one noise draw per dose, so it has no interval). Settings are those of Section 3.10: tuned on MNIST validation images at each dose, unweighted data term, ADMM implementation of TV, and one MNIST-trained denoiser. The qualitative picture of Stage 1 is reproduced; every interval for the 200-image sets excludes zero by at least 6 standard errors.

Prior-matched data. On MNIST the learned prior beats TV at every dose, by +3.7 dB at $I_0 = 100$–300 (effective $\sigma \approx 0.03$–0.06), against +3.84 dB at Gaussian $\sigma = 0.05$, and for 99–100% of the images at $I_0 \geq 100$. The advantage shrinks at both ends of the dose range, to +0.8 dB at $I_0 = 10$, of the same order as the Gaussian result at $\sigma = 0.2$ (+0.97 dB, E8). EMNIST keeps a smaller advantage (+0.6 to +1.8 dB). These gaps are against TV without the box constraint; that constraint was not tested under Poisson noise (in the Gaussian MNIST case it lowered the gap by 0.5 dB, E8).

Shifted data. The MNIST-trained prior is below TV on Fashion-MNIST (−2.0 to −3.1 dB; PnP wins on only 6–14% of the images) and on the phantom (−4.4 to −6.1 dB), as in Stage 1 (−3.10 and −4.33 dB). At $I_0 = 1000$ the prior is no better than FBP on Fashion-MNIST (19.72

versus 19.77 dB), and on the phantom it is 3.1 dB below FBP (18.07 versus 21.20 dB). The share of the error in the complement of the well-determined subspace is 0.68–0.99 of TV's on MNIST but 1.04–1.19 on Fashion-MNIST and 1.73–1.79 on the phantom (for $I_0 \leq 300$; 0.90 at $I_0 = 1000$), so the mismatch error again sits partly in the weakly determined directions. These ratios are less favourable to PnP than the Stage 1 values and are not directly comparable, since the noise model, the tuned settings and the subspace definition differ. The gap on Fashion-MNIST is not monotone in dose (-2.45, -3.05, -2.74, -3.01, -1.98 dB from high to low dose); with frozen MNIST-tuned settings this reflects, at least in part, the choice of (σ_d, k) (see the weighted variant below), so we read the Fashion-MNIST range as "2 to 3 dB below TV" and not as a dose trend.

Structural similarity. In SSIM the picture differs on Fashion-MNIST (Table 20): the PnP–TV difference is between -0.014 and $+0.054$ and PnP wins on 47–67% of the images, so the 2–3 dB PSNR reversal has no counterpart in SSIM. On the phantom SSIM is lower for PnP at every dose (-0.19 to -0.31). On MNIST and EMNIST PnP is better in SSIM at every dose. The reversal under shift is therefore a PSNR result for Fashion-MNIST and a PSNR and SSIM result for the phantom.

Weighted data term. Using the Poisson weights in PnP gave no consistent benefit: the weighted variant was 0.14–0.59 dB below the unweighted one in 13 of the 15 non-phantom cells, and above it in two (EMNIST +0.35 dB and Fashion-MNIST +1.10 dB at $I_0 = 300$), where the tuned (σ_d, k) differed ($(0.05, 10)$ unweighted versus $(0.1, 3)$ weighted). The 1.1 dB on Fashion-MNIST shows that the shifted-data gap depends on how the frozen settings were chosen. We report the unweighted scheme, as it is the one tuned on the same footing as in Stage 1, and note that the weighted variant would reduce the Fashion-MNIST gap at $I_0 = 300$ from -3.05 to about -1.95 dB.

Table 19: PnP−TV PSNR difference (dB, mean ± standard error of the paired difference over 200 images) by photon count I_0, with the corresponding effective sinogram noise standard deviation (measured on the MNIST images; signal dependence makes it about 15–20% higher on Fashion-MNIST); and the ratio of PnP's to TV's reconstruction error in the complement of the well-determined subspace (in brackets). Phantom: one image.

I_0 (photons/ray) effective σ	1000 0.018	300 0.032	100 0.056	30 0.106	10 0.198
MNIST (train dist.)	$+2.87 \pm .07$ (0.70)	$+3.68 \pm .08$ (0.68)	$+3.65 \pm .08$ (0.71)	$+2.39 \pm .12$ (0.88)	$+0.82 \pm .12$ (0.99)
EMNIST	$+1.40 \pm .09$ (0.75)	$+1.59 \pm .08$ (0.83)	$+1.82 \pm .08$ (0.86)	$+0.77 \pm .11$ (1.02)	$+0.62 \pm .10$ (1.00)
Fashion-MNIST	$-2.45 \pm .13$ (1.04)	$-3.05 \pm .16$ (1.18)	$-2.74 \pm .16$ (1.18)	$-3.01 \pm .14$ (1.19)	$-1.98 \pm .13$ (1.07)
Shepp–Logan	−4.41 (0.90)	−6.05 (1.79)	−5.05 (1.73)	−5.49 (1.77)	−4.75 (1.75)

Table 20: PnP−TV SSIM difference (mean ± standard error over 200 images; share of images won by PnP in brackets) for the same runs as Table 19.

I_0 (photons/ray)	1000	300	100	30	10
MNIST (train dist.)	$+.089 \pm .002$ (1.00)	$+.170 \pm .005$ (1.00)	$+.244 \pm .006$ (1.00)	$+.279 \pm .008$ (1.00)	$+.267 \pm .009$ (0.99)
EMNIST	$+.039 \pm .002$ (0.96)	$+.081 \pm .003$ (0.99)	$+.123 \pm .004$ (1.00)	$+.136 \pm .006$ (0.98)	$+.166 \pm .007$ (0.98)
Fashion-MNIST	$-.014 \pm .007$ (0.47)	$-.005 \pm .009$ (0.50)	$+.026 \pm .010$ (0.55)	$+.011 \pm .009$ (0.52)	$+.054 \pm .009$ (0.67)
Shepp–Logan	−.216	−.306	−.189	−.245	−.206

4.12 E13: Lesion insertion under Poisson noise

We repeated the lesion experiment (Section 3.8; same lesions, observer, locations and validity filter; three denoisers per prior, 200 images per set) with Poisson data at $I_0 = 100$ and $I_0 = 30$ (effective




Figure 16: PSNR versus photon count for FBP, Tikhonov, TV, PnP and the Poisson-weighted PnP on the four test sets (MNIST-trained denoiser, hyperparameters tuned on MNIST validation images at each dose).

$\sigma \approx 0.06$ and 0.1) and the matrix-free operator, tuning each method on its prior's own validation distribution (Table 21, Figure 17). TV is the ADMM implementation without the box constraint; the box-constrained control of E9 was run for Gaussian noise only. Only these two doses were run; the other three doses of E12 were not.

Erasure persists. The paired AUC of PnP is below TV's in all four prior/test-set combinations and at both doses (differences of -0.01 to -0.28 at $I_0 = 100$, with one tie at 1.0 for the largest high-contrast lesion, and -0.02 to -0.19 at $I_0 = 30$). For the MNIST prior on MNIST ($\sigma_l = 0.75$, $c = 0.5$) the paired AUC is 0.60 against 0.89 for TV and the retention is 0.29 against 0.60 at $I_0 = 100$, which is close to the Gaussian-noise values of Stage 1 (0.69 vs. 0.91; 0.28 vs. 0.59), although PnP is 3.7 dB better in PSNR on the same images. At $I_0 = 30$ the gap narrows (0.63 vs. 0.77) because TV degrades faster.

Invention under mismatch persists. For the MNIST-trained prior on Fashion-MNIST, PnP produces more false lesions than TV for lesions with $\sigma_l \geq 0.75$ in 11 of the 12 dose/lesion-type cells, by factors of 1.7–5.5 (0.25–0.70 against 0.06–0.28 per image at $I_0 = 100$; the exception is 0.25 against 0.28). The matched Fashion-trained prior, on its own data and on MNIST, stays at ≤ 0.09 false lesions per image. Retention above that of TV for the mismatched prior should not be read as fidelity: invented contrast may contribute to it, although we did not separate the two.

New at the lower dose: in-distribution invention. For the MNIST prior on MNIST at $I_0 = 30$, PnP produces 0.058–0.073 false lesions per image for the four lesion types with $0.75 \leq \sigma_l \leq 1.25$, against 0.000–0.035 for TV, whereas at $I_0 = 100$ it produces at most 0.01 for any type. At high noise the in-distribution prior therefore starts to create lesion-like structure as well as to erase it, although at a rate far below that under mismatch.

TV has a failure of its own. For the smallest low-contrast lesion at $I_0 = 100$ the TV reconstruction contains 0.63 false lesions per image (noise blobs), against 0.01 for PnP. The statement that PnP creates more false lesions than TV under mismatch therefore applies to lesions of $\sigma_l \geq 0.75$ px, and under a matched prior is not supported at all. As in E9, where box-constrained TV also produced false lesions for the smallest lesion type, the false-lesion rate of TV depends on its implementation, so the TV false-lesion counts here are those of ADMM-TV and may be different for other TV solvers.

Uncertainty is limited by the design: the error bars in Figure 17 show the variation over three denoisers only, TV is a single run, and with about 170 valid lesions per cell the sampling error of a paired AUC is of the order of ± 0.03, so differences below that, such as the ties for the largest lesions, do not separate the methods. A tercile analysis of PnP's paired AUC by lesion observability showed a decrease with observability on MNIST images and no trend on Fashion-MNIST images; since it depends on the image set, our hypothesis is that it reflects the background around the lesion

rather than the operator; we did not test this and do not interpret it (the same pattern appears in the Gaussian-noise data of E9).

Table 21: Lesion insertion under Poisson noise, mean over three denoisers per prior (200 test images; 20 angles; bicubic data), for two of the eight lesion types. pAUC: paired AUC; Ret.: retention; FL: false lesions per lesion-absent image. Each method is tuned on its own prior's validation distribution at the given dose.

		$I_0 = 100$						$I_0 = 30$					
		pAUC		Ret.		FL		pAUC		Ret.		FL	
Prior → test set	σ_l/c	TV	PnP	TV	PnP	TV	PnP	TV	PnP	TV	PnP	TV	PnP
MNIST → MNIST	0.75/0.5	.89	.60	.60	.29	.010	.007	.77	.63	.48	.34	.005	.067
	1.25/0.25	.88	.60	.60	.27	.020	.007	.75	.65	.46	.46	.005	.068
MNIST → Fashion	0.75/0.5	.81	.72	.64	.84	.240	.443	.73	.63	.42	.57	.230	.518
	1.25/0.25	.84	.72	.65	.84	.255	.438	.71	.61	.42	.54	.225	.518
Fashion → Fashion	0.75/0.5	.79	.64	.46	.23	.075	.020	.69	.58	.25	.11	.020	.038
	1.25/0.25	.85	.71	.48	.25	.075	.020	.68	.56	.27	.16	.040	.038
Fashion → MNIST	0.75/0.5	.88	.66	.48	.31	.000	.003	.78	.64	.33	.27	.000	.007
	1.25/0.25	.89	.66	.49	.33	.000	.003	.78	.65	.34	.34	.000	.007

4.13 E14: Data, noise model and simulator

The geometry fit, the noise-model checks and the dose thinning are described in Sections 3.11–3.12. Figure 18 illustrates the geometry fit of Section 3.11 for one test slice. All hyperparameters are tuned on *real* (thinned) observations of the validation slices, whereas every reported comparison is evaluated on *simulated* crime-free data of the test slices. The simulator therefore has to be shown to be a faithful stand-in for the real data, and Table 22 does this for the baseline that is cheapest to run: it compares the TV PSNR obtained from the real (thinned) observations with that obtained from our crime-free simulation of the same slices. The agreement is within 0.15 dB at all nine conditions (first 8 test slices, one noise realisation each; mean difference +0.03 dB, simulated minus real). For all 40 slices at $K = 60$ the values are 31.91 vs. 32.04 dB (full dose) and 29.82 vs. 29.83 dB (quarter dose), and for FBP-Hann 30.65 vs. 30.59 and 28.73 vs. 28.66 dB.

The residual differences are small compared with the method differences studied below ($\leq$ 2.3 dB), but they are not zero and they are not random in sign across doses: the simulation is slightly *easier* than the real data at full dose (+0.07 to +0.15 dB at all three K), and indistinguishable or slightly harder at the two lower doses (−0.05 to +0.05 dB). This is the direction expected if the real counts carry somewhat more variance than the simulation at high counts (dispersion index in Section 3.12), but it is a hypothesis: eight slices and a single noise draw do not allow a test, and we did not compute an interval for these differences. What the check does *not* show is that the *difference* between PnP and TV is the same on real and simulated data, only that the TV baseline is; we therefore state the PnP–TV comparisons for the simulated data and treat the simulator as validated for baseline accuracy, not for every method.




Figure 17: Paired detection AUC (top rows) and false lesions per lesion-absent image (bottom rows) under Poisson noise, $I_0 = 100$ (upper panel) and $I_0 = 30$ (lower panel). Bars: mean over three denoisers; error bars: s.d. over seeds.

Table 22: TV PSNR (dB) from real observations (thinned to the dose, subsampled to K angles) and from the crime-free simulation of the same slices (first 8 test slices).

K	30			60			120		
dose	1	0.25	0.1	1	0.25	0.1	1	0.25	0.1
real	29.34	27.95	26.57	30.15	28.40	26.73	30.61	28.54	27.89
simulated	29.41	27.90	26.52	30.27	28.45	26.72	30.76	28.55	27.85




Figure 18: Geometry check on one slice (slice 0). Top: ground truth, FBP of the supplied observation, and the absolute error of that FBP. Bottom: the supplied sinogram, our forward projection of the ground truth, and one projection (angle index 333) of the observation (solid) and of our projection (dashed). The observation is the noisy version of our noise-free projection, with matching structure. The fit statistics (relative residual 0.038, correlation 0.9979, with a fitted least-squares scale) are pooled over 8 evenly spaced test slices (Section 3.11); the figure shows only the first of them.

Table 23: Tuned parameters (20 validation slices per condition): Hann cut-off, TV weight (the TV-long weight was the same in every condition), and (σ_d, k) of PnP-const and PnP-TVinit.

K	dose	Hann	TV λ	PnP-const (σ_d, k)	PnP-TVinit (σ_d, k)
30	1	0.2	0.003	(0.02, 12)	(0.02, 4)
30	0.25	0.16	0.01	(0.05, 8)	(0.08, 4)
30	0.1	0.12	0.03	(0.08, 12)	(0.05, 8)
60	1	0.3	0.003	(0.02, 8)	(0.02, 2)
60	0.25	0.2	0.01	(0.08, 4)	(0.02, 8)
60	0.1	0.12	0.03	(0.05, 12)	(0.02, 12)
120	1	0.4	0.003	(0.05, 2)	(0.0039, 8)†
120	0.25	0.2	0.01	(0.02, 12)	(0.0062, 8)
120	0.1	0.16	0.01	(0.05, 8)	(0.128, 1)†

† optimum on the boundary of the grid after the automatic extensions.

4.14 E15: Learned prior versus TV on real slices at $K = 60$

Table 24 compares PnP variants with TV on the 40 test slices at $K = 60$ (one slice per patient; data simulated crime-free, Section 3.12); all methods are tuned on the same 20 validation slices. TV is the box-constrained primal–dual solver of Section 3.13, so the comparison is already against the stronger TV of E2b. Besides PnP-const and PnP-TVinit (Section 3.13), the table contains two variants introduced to test whether the PnP set-up, not the prior, is the weak point: *PnP-anneal*, in which σ_d decreases geometrically from $\sigma_0 \in \{0.05, 0.1\}$ to $\sigma_1 \in \{0.01, 0.03\}$ over $T \in \{20, 40\}$ iterations (FBP-Hann start), and *TV+denoise*, a single denoiser pass on the TV image with tuned σ. Each variant is run with the 8-layer denoiser (d8) and with a deeper one (d17, 17 layers, trained for 20 000 steps).

With the re-tuned starting image, PnP-const is slightly but significantly below TV (-0.17 and -0.10 dB at full and quarter dose) and PnP-TVinit is at TV level at full dose ($+0.03$ dB, interval $[-0.00, +0.06]$) and $+0.21$ dB above it at quarter dose. The deeper denoiser does not help (PnP-const -0.30 and -0.25 dB). Annealing σ_d is worse than constant σ_d at full dose (-0.65 dB versus TV) and better at quarter dose ($+0.12$ dB, but not for the deeper denoiser: $+0.05$, interval including zero). One denoiser pass on the TV image gives TV (within ± 0.05 dB with the d8 network, and ± 0.02 dB with d17). SSIM does not show the quarter-dose gain: PnP-TVinit [d8] is 0.003 above TV at quarter dose and 0.002 below it at full dose, and PnP-const is 0.013 and 0.016 below TV.

How far to trust the best variant. The gain of PnP-TVinit at quarter dose is the largest positive entry of the table, but three cautions apply. (i) PnP-TVinit and the other variants were introduced *after* the first PnP-const result was below TV (Section 3.16), so the best of several variants is selected in part on the test slices and its $+0.21$ dB is optimistic; no multiplicity correction is applied to the 16 intervals of the table. (ii) PnP-TVinit performs more iterations than TV, whereas TV has only the 150 of the tuned solver; the control for that is TV-long in E16, where TV-long is itself above TV at $K = 60$ (32.12 and 29.88 dB in Table 25, against 31.91 and 29.82 dB for TV here, on different noise draws), so the margin of PnP-TVinit over a TV given the same number of iterations is smaller than the one reported here and is studied in E16. (iii) 0.2 dB is a small difference, and the paired intervals reflect sampling over 40 patients with one noise realisation each, not variation over denoiser training runs. In short, on clinical-size slices and at $K = 60$ the learned prior is at best competitive with TV in PSNR rather than ≈ 4 dB better as in the 28×28 experiments. This is not a like-for-like contrast: image size, data, noise model and the match between prior and test data all differ, and we do not claim to have isolated the reason.

4.15 E16: Accuracy over the dose $\times$ angle grid

Table 25 gives the mean PSNR and SSIM of the five methods in the nine conditions of the lesion-free arm of the lesion experiment (same 40 slices in every condition, one noise realisation per condition), and Table 26 the paired differences. The top row of Figure 19 (shown with E17) displays the PSNR and SSIM differences of Table 26 as heat maps over the dose $\times$ angle grid. TV-long, the TV with twice the iterations and its own tuned weight, is the primary classical comparator, because it receives the same number of iterations as the PnP methods. No correction for multiplicity is applied to the 36 intervals of Table 26, and since the conditions share the same slices they are not independent.

What holds in all nine conditions. PnP-const and TV-long beat FBP-Hann in PSNR (PnP-const by 0.92–1.36 dB, TV-long by 0.81–2.28 dB) and in SSIM, with all intervals excluding zero.

PnP-const against TV-long. PnP-const is *not* better than TV-long on average: its PSNR difference lies between -0.93 and $+0.33$ dB (mean over the nine conditions -0.13 dB), significantly

Table 24: $K = 60$, 40 test slices: PSNR (dB) and SSIM, and the paired PSNR difference to TV (mean, 95% bootstrap interval over patients; * excludes 0). d8: 8-layer denoiser; d17: 17-layer denoiser.

	dose 1			dose 0.25		
Method	PSNR	SSIM	−TV	PSNR	SSIM	−TV
FBP-Hann	30.65	0.738	-1.27^*	28.73	0.699	-1.09^*
TV	31.91	0.786	–	29.82	0.750	–
PnP-const [d8]	31.74	0.773	-0.17^* $[-0.21, -0.13]$	29.72	0.734	-0.10^* $[-0.18, -0.03]$
PnP-TVinit [d8]	31.94	0.784	$+0.03$ $[-0.00, +0.06]$	30.03	0.753	$+0.21^*$ $[+0.17, +0.25]$
PnP-anneal [d8]	31.26	0.769	-0.65^* $[-0.77, -0.52]$	29.94	0.750	$+0.12^*$ $[+0.05, +0.18]$
TV+denoise [d8]	31.91	0.785	$+0.00$ $[-0.02, +0.03]$	29.77	0.747	-0.05 $[-0.10, +0.01]$
PnP-const [d17]	31.61	0.777	-0.30^* $[-0.35, -0.26]$	29.57	0.737	-0.25^* $[-0.31, -0.19]$
PnP-TVinit [d17]	31.89	0.785	-0.02 $[-0.06, +0.03]$	29.97	0.755	$+0.15^*$ $[+0.12, +0.18]$
PnP-anneal [d17]	31.20	0.772	-0.71^* $[-0.85, -0.58]$	29.87	0.753	$+0.05$ $[-0.01, +0.12]$
TV+denoise [d17]	31.90	0.786	-0.02 $[-0.05, +0.02]$	29.84	0.752	$+0.02^*$ $[+0.00, +0.04]$

positive in three conditions ($K = 30$ and 60 at dose 0.1, and $K = 120$ at full dose), significantly negative in four ($K = 30$, 60 at full dose; $K = 60$ at quarter dose; $K = 120$ at dose 0.1) and not significant in two. Its SSIM is below TV-long's in eight of nine conditions, significantly so in seven, and above it in one (+0.001, not significant). The sign does not follow dose or angle count monotonically: the low-dose, few-view advantage at $K = 30$ and 60 is reversed at $K = 120$, dose 0.1. We therefore do not summarise the grid as a trend.

PnP-TVinit. PnP-TVinit has the highest mean PSNR in seven of nine conditions, by at most 0.49 dB over TV-long (significantly higher in five, lower in two: $K = 30$, full dose, −0.15 dB, and $K = 60$, full dose, −0.03 dB), but the highest mean SSIM in only five. Two of its cells are not denoiser effects (Section 3.13): at $K = 120$, full dose, the tuned $\sigma_d = 0.0039$ means almost no denoising, so the +0.41 dB over TV-long (and +0.28 dB over TV) there comes from the additional data-consistency iterations started at the TV solution, not from the learned prior; at $K = 120$, dose 0.1, the tuned $k = 1$ is a single step. The largest margin, +0.49 dB at $K = 60$, dose 0.1, is in a cell with interior optimum.

Scale of the differences. TV is the same method on the same slices in the $K = 60$ cells here and in E15, yet its mean PSNR differs by 0.13 dB at full dose (32.04 versus 31.91 dB) and 0.01 dB at quarter dose because the noise draws differ, while PnP-const − TV changes by only 0.02 dB at full dose (−0.19 versus −0.17 dB). Differences between methods are thus more stable than absolute levels, but margins of a few tenths of a dB are within what a different noise realisation can move, and we do not rank methods on differences below that. TV-long is slightly worse than TV at $K = 120$, full dose (32.50 vs. 32.64 dB), which shows that the early-stopped 150-iteration TV is itself a tuned object and that the weight grid is coarse.

Relation to detection. The ranking by PSNR is not the ranking by lesion detectability (E17): TV-long is 0.81–2.28 dB better than FBP-Hann in PSNR but has a lower paired AUC in all nine conditions, so the accuracy results of this subsection should not be read as evidence about diagnostic usefulness.

Table 25: Mean PSNR (dB, upper block) and SSIM (lower block) of the lesion-free reconstructions, 40 patients (best in each condition in bold). Columns: K angles / dose.

	$K = 30$			$K = 60$			$K = 120$		
Method	1	0.25	0.1	1	0.25	0.1	1	0.25	0.1
PSNR (dB)									
FBP-Hann	29.02	27.72	26.63	30.59	28.66	27.11	31.70	29.18	27.94
TV	30.98	29.15	27.69	32.04	29.83	27.93	32.64	30.07	29.13
TV-long	**31.31**	29.21	27.72	**32.12**	29.88	27.98	32.50	30.11	29.21
PnP-const	30.38	29.05	27.86	31.85	29.68	28.31	32.64	30.23	28.86
PnP-TVinit	31.15	**29.22**	**27.96**	32.09	**30.04**	**28.47**	**32.92**	**30.46**	**29.26**
SSIM									
FBP-Hann	0.692	0.666	0.655	0.738	0.699	0.680	0.763	0.725	0.693
TV	0.765	0.733	0.710	0.788	0.750	0.717	0.799	0.756	0.741
TV-long	**0.773**	**0.737**	0.713	**0.790**	0.753	0.720	0.797	0.758	**0.745**
PnP-const	0.745	0.726	0.707	0.775	0.733	0.721	0.793	0.749	0.731
PnP-TVinit	0.767	0.730	**0.721**	0.786	**0.753**	**0.731**	**0.805**	**0.761**	0.740

Table 26: Paired differences over patients (mean, 95% bootstrap interval; * excludes 0), 40 patients.

K	dose	PSNR: PnP-const − TV-long	SSIM: PnP-const − TV-long	PSNR: PnP-TVinit − TV-long	PSNR: PnP-const − FBP-Hann
30	1	-0.93^{*} $[-1.09, -0.75]$	-0.027^{*} $[-0.035, -0.021]$	-0.15^{*} $[-0.24, -0.06]$	$+1.36^{*}$ $[+1.15, +1.58]$
30	0.25	-0.15 $[-0.36, +0.08]$	-0.011^{*} $[-0.019, -0.002]$	$+0.01$ $[-0.19, +0.24]$	$+1.33^{*}$ $[+1.06, +1.62]$
30	0.1	$+0.13^{*}$ $[+0.03, +0.23]$	-0.006 $[-0.014, +0.001]$	$+0.24^{*}$ $[+0.19, +0.29]$	$+1.22^{*}$ $[+0.98, +1.48]$
60	1	-0.28^{*} $[-0.33, -0.22]$	-0.015^{*} $[-0.019, -0.011]$	-0.03^{*} $[-0.06, -0.00]$	$+1.25^{*}$ $[+0.97, +1.56]$
60	0.25	-0.20^{*} $[-0.28, -0.13]$	-0.020^{*} $[-0.027, -0.012]$	$+0.16^{*}$ $[+0.11, +0.21]$	$+1.02^{*}$ $[+0.65, +1.41]$
60	0.1	$+0.33^{*}$ $[+0.23, +0.43]$	$+0.001$ $[-0.003, +0.005]$	$+0.49^{*}$ $[+0.37, +0.62]$	$+1.20^{*}$ $[+1.04, +1.37]$
120	1	$+0.14^{*}$ $[+0.05, +0.24]$	-0.004^{*} $[-0.007, -0.001]$	$+0.41^{*}$ $[+0.32, +0.50]$	$+0.95^{*}$ $[+0.54, +1.39]$
120	0.25	$+0.13$ $[-0.02, +0.27]$	-0.009^{*} $[-0.013, -0.006]$	$+0.35^{*}$ $[+0.25, +0.45]$	$+1.05^{*}$ $[+0.89, +1.22]$
120	0.1	-0.34^{*} $[-0.47, -0.22]$	-0.013^{*} $[-0.020, -0.007]$	$+0.06$ $[-0.07, +0.19]$	$+0.92^{*}$ $[+0.73, +1.13]$

4.16 E17: Lesion detection

Grid (Table 27, Figure 19). Pooled over the two lesion sizes (contrast 0.06; up to 160 lesions per size), PnP-const has a higher paired AUC than TV-long in all nine conditions, by 0.03 to 0.16, with intervals excluding zero in six (all three $K = 120$ conditions, $K = 60$ at doses 0.25 and 0.1, $K = 30$ at dose 0.1); it is not significant at full dose for $K = 30$ and 60. Against FBP-Hann, however, PnP-const is never significantly better (three positive, six negative or zero differences) and is significantly worse in two conditions ($K = 30$, full dose, -0.075; $K = 120$, quarter dose, -0.059): the learned prior removes the detection deficit of TV rather than adding to FBP's detectability. TV-long, tuned for PSNR, has a lower AUC than FBP-Hann in all nine conditions (significantly in seven, down to -0.20 at $K = 60$, dose 0.1) while being 0.81–2.28 dB better in PSNR: PSNR and detectability rank the methods in opposite orders. PnP-TVinit exceeds TV-long in AUC in all nine conditions (significantly in seven; $+0.02$ to $+0.17$). The choice of comparator matters: against the 150-iteration TV the advantage of PnP-const is 0.000 and $+0.006$ at full dose for $K = 60$ and 30, because TV-long, with its extra iterations, has the lower AUC of the two TV variants (at $K = 60$, full dose, by 0.05). The conclusion "PnP beats TV in detection" is therefore specific to a PSNR-tuned TV with the iteration budget of PnP; we did not tune TV, nor PnP, for detection on real slices (in the Gaussian experiments of E10 the detection-tuned TV changed very little, but that does not transfer automatically).

Retention. At $K = 60$ and full dose the regularised methods retain only 7–32% of the local-contrast difference of the inserted lesion (FBP-Hann 31–67%); the detection scores of TV and PnP are therefore obtained mainly from noise suppression, not from preserving lesion contrast.

Earlier analysis at $K = 60$ (E17b). With four lesion types (two sizes, contrasts 0.03 and 0.06; 160 lesions per type pooled), re-tuned on 20 slices, the pooled AUC difference PnP-const − TV-long is $+0.023$ ($[-0.013, +0.061]$) at full dose and $+0.073$ ($[+0.023, +0.128]$) at quarter dose, in line with the grid values ($+0.050$ and $+0.100$; different lesion sets and noise). At both doses PnP-const is level with FBP-Hann (-0.008, $+0.016$; not significant) and TV is below it (-0.037, not significant; -0.056, $[-0.11, -0.00]$). At $\sigma_l = 4$ px and $c = 0.06$ all methods reach AUC 0.95–0.96 at full dose, so the large lesions do not discriminate between methods at that dose (ceiling).

Contrast with the 28×28 experiments. In E9 the learned prior had a *lower* paired AUC than TV in distribution (-0.04 to -0.23), at a PSNR advantage of about 4 dB. On real slices the sign is reversed: PnP-const is above TV-long in all nine conditions and above TV in eight of nine (significantly in seven), at a PSNR that differs from TV-long's by between -0.9 and $+0.3$ dB. We did not test why. Candidate explanations that this study cannot separate are the lesion scale relative to the pixel grid ($\sigma_l = 2$–4 px here against 0.5–2 px), the noise texture of real slices against the digit backgrounds, and the different amount of smoothing that the PSNR-tuned operating points produce; the observer, which sums a disc against an annulus, rewards any method that suppresses background fluctuations at the lesion scale. The result should be read as a property of this prior, observer and lesion type, not as a general ranking of learned and classical regularisers.

Scope. The lesions have a contrast of 0.06 on the $\mu/\mu_{\max}$ scale, which is high compared with clinically low-contrast lesions (Section 3.14), the observer knows the lesion location, there is one noise realisation per condition and 40 patients, and the AUC of a single cell has a precision of about ± 0.05. A detection advantage says nothing about false structures; these are the subject of E18.

4.17 E18: False structures on real slices

Table 30 (CT rows) gives the mean number of false blobs per lesion-free slice at one times the nominal lesion strength for the four conditions that are also used in the shift study. A false blob




Figure 19: Paired differences over the nine dose × angle conditions (cell: mean, 95% interval over 40 patients; * excludes 0; blue: the first method is higher). Top row: PSNR (PnP-const and PnP-TVinit against TV-long) and SSIM (E16); bottom row: paired AUC of PnP-const against TV-long and against FBP-Hann (E17), and the difference in false blobs per slice at one times the nominal lesion strength (E18). Produced by `run_stage3_step5_summary.py`.

Table 27: Paired AUC differences ($\sigma_l \in \{2, 4\}$ px pooled, $c = 0.06$, 40 patients; mean, 95% interval; * excludes 0).

K	dose	PnP-const − TV-long	PnP-const − TV	PnP-const − FBP-Hann	TV-long − FBP-Hann	PnP-TVinit − TV-long
30	1	$+0.044$ $[-0.009, +0.097]$	$+0.006$ $[-0.050, +0.059]$	-0.075^* $[-0.116, -0.037]$	-0.119^* $[-0.169, -0.072]$	$+0.066^*$ $[+0.019, +0.113]$
30	0.25	$+0.034$ $[-0.025, +0.094]$	$+0.078^*$ $[+0.013, +0.144]$	-0.006 $[-0.056, +0.041]$	-0.041 $[-0.103, +0.025]$	$+0.069^*$ $[+0.006, +0.131]$
30	0.1	$+0.091^*$ $[+0.019, +0.162]$	$+0.137^*$ $[+0.059, +0.216]$	$+0.031$ $[-0.028, +0.091]$	-0.059 $[-0.141, +0.019]$	$+0.053$ $[-0.022, +0.128]$
60	1	$+0.050$ $[-0.006, +0.103]$	$+0.000$ $[-0.050, +0.050]$	-0.037 $[-0.091, +0.012]$	-0.087^* $[-0.144, -0.031]$	$+0.069^*$ $[+0.031, +0.109]$
60	0.25	$+0.100^*$ $[+0.050, +0.150]$	$+0.078^*$ $[+0.028, +0.131]$	-0.009 $[-0.059, +0.041]$	-0.109^* $[-0.175, -0.047]$	$+0.084^*$ $[+0.031, +0.141]$
60	0.1	$+0.163^*$ $[+0.078, +0.247]$	$+0.144^*$ $[+0.072, +0.219]$	-0.041 $[-0.097, +0.009]$	-0.203^* $[-0.291, -0.119]$	$+0.156^*$ $[+0.084, +0.231]$
120	1	$+0.109^*$ $[+0.066, +0.156]$	$+0.084^*$ $[+0.044, +0.125]$	$+0.003$ $[-0.031, +0.041]$	-0.106^* $[-0.162, -0.053]$	$+0.091^*$ $[+0.050, +0.137]$
120	0.25	$+0.112^*$ $[+0.041, +0.184]$	$+0.081^*$ $[+0.006, +0.156]$	-0.059^* $[-0.103, -0.013]$	-0.172^* $[-0.247, -0.094]$	$+0.166^*$ $[+0.109, +0.222]$
120	0.1	$+0.084^*$ $[+0.025, +0.147]$	$+0.084^*$ $[+0.028, +0.147]$	$+0.013$ $[-0.047, +0.072]$	-0.072^* $[-0.128, -0.016]$	$+0.022$ $[-0.041, +0.081]$

is defined in Section 3.14; because the ground truth of LoDoPaB-CT is itself a reconstruction, "false" means unsupported by this reference and not absent from the patient. On real slices PnP-const makes 0.24–1.38 such blobs per slice and TV-long 0.00–6.06 (FBP-Hann 5.8–26.8), so the regularised methods make far fewer than FBP-Hann, but the counts of TV-long are not small in every condition. Over the nine conditions the paired difference PnP-const − TV-long at $m = 1$ (Table 28) is significantly positive in three ($K = 60$ full dose: +0.46; $K = 120$ full and quarter dose: +0.24, +1.35), significantly negative in three ($K = 30$ quarter dose: −5.31; $K = 60$ quarter dose: −0.40; $K = 120$ dose 0.1: −0.93) and not significant in three, with no consistent pattern in dose or angle count: at $K = 120$ the sign changes from positive to negative as the dose decreases, at $K = 30$ it is negative in two of three conditions. At $m = 1.5$ the excess of PnP-const over TV-long is at most 0.45 blobs per slice ($K = 120$, quarter dose). In the $K = 60$ experiment (E17b, threshold $m = 0.5$) PnP-const has fewer false blobs than TV at both doses (−1.7 at full dose, −6.6 at quarter dose). Counts are not comparable with E9: there the lesion-free images are digits on an empty background and the learned prior produced essentially none in distribution, whereas here the background is real anatomy, which contains many structures of lesion size.

The galleries of the strongest unsupported blob per slice (Figures 20 and 21) show that the largest false blobs occur at the same location in FBP-Hann and in TV as well as in PnP (for example the slices 28 and 9 at full dose and most rows at quarter dose), so they are produced by the data (few views, noise, streaks), not by the prior. Only a few look specific to the prior: a crescent in slice 18 at full dose (PnP-const only) and streaks in the PnP-TVinit image of slice 32 at quarter dose. We therefore find *no evidence that the prior creates many lesion-like structures in distribution*, but we cannot exclude rare ones: 40 slices and one noise draw per condition do not bound a rare event, and the galleries show the worst cases selected by the methods' own responses, not a random sample.

Table 28: Paired differences in false blobs per lesion-free slice on real CT slices (40 patients; mean, 95% interval; * excludes 0), pooled over $\sigma_l \in \{2, 4\}$; m: threshold in units of the nominal lesion response.

K	dose	$m = 1$: PnP-const − TV-long	$m = 1$: PnP-const − TV	$m = 1.5$: PnP-const − TV-long
30	1	−1.19 [−2.79, +0.05]	−0.54 [−1.57, +0.36]	−0.23 [−0.64, +0.11]
30	0.25	−5.31* [−9.15, −2.31]	−4.81* [−8.41, −2.00]	−2.42* [−4.56, −0.82]
30	0.1	+0.21 [−0.20, +0.64]	+0.31 [−0.12, +0.76]	+0.00 [−0.14, +0.14]
60	1	+0.46* [+0.12, +0.88]	+0.19* [+0.01, +0.40]	+0.12* [+0.03, +0.25]
60	0.25	−0.40* [−0.68, −0.17]	−0.60* [−1.11, −0.19]	−0.11* [−0.25, −0.01]
60	0.1	−0.07 [−0.25, +0.07]	−0.05 [−0.20, +0.09]	−0.01 [−0.06, +0.04]
120	1	+0.24* [+0.04, +0.51]	+0.11* [+0.01, +0.25]	+0.06 [+0.00, +0.15]
120	0.25	+1.35* [+0.53, +2.30]	+1.35* [+0.54, +2.29]	+0.45* [+0.16, +0.81]
120	0.1	−0.93* [−1.50, −0.45]	−0.88* [−1.48, −0.39]	−0.11* [−0.26, −0.01]

4.18 E19: Distribution shift to non-anatomical phantoms

The CT-trained prior and all tuned parameters are applied, without retuning, to 20 random piecewise-constant phantoms (ellipses, rectangles and a stripe or checker block on a faint disc, $\mu/\mu_{\max}$ scale; nothing like a CT slice, but the kind of image TV favours; same phantoms in every condition; this is a deliberately TV-friendly shift, so it tests how much of the learned prior's behaviour survives, not how it fares under realistic shifts such as another scanner or anatomy) with




Figure 20: Strongest unsupported blob per slice at full dose ($K = 60$; $\sigma_l = 4$ px detector) for the four slices with the largest false-blob response in PnP-const (top four rows) and in TV (bottom four rows). Columns: ground truth crop, FBP-Hann, TV, PnP-const, PnP-TVinit; red circle: location of the blob; labels give the response as a multiple of the nominal lesion response.




Figure 21: As Figure 20 at quarter dose.

the same lesion protocol. The shift was first run at $K = 60$ (doses 1 and 0.25) and then extended to $K = 30$ and 120; the $K = 60$ analysis used four lesion types, the others two, so the designs differ slightly.

Accuracy and detection (Table 29). On the phantoms PnP-const is below TV-long in PSNR in all six conditions (-0.11 to -0.75 dB; significantly in five of the six conditions, the exception being $K = 30$, full dose) and in SSIM in all four (-0.095 to -0.190, all significant, against -0.004 to -0.027 on CT). At quarter dose its SSIM is also below FBP-Hann's on the phantoms (-0.045 at $K = 30$ and -0.041 at $K = 120$, both significant). The detection advantage over TV-long is significant in one of six shift conditions ($K = 120$, full dose: $+0.075$, $[+0.013, +0.137]$), against three of the six corresponding CT conditions ($K = 60$ quarter dose, $K = 120$ at both doses); the shift effect is significantly negative only at $K = 120$, quarter dose (-0.175, $[-0.297, -0.056]$). The CT advantage is therefore not reliably transferred to the shifted images, but at 20 phantoms the interval widths of 0.12–0.19 do not allow us to say that it is lost either. In PSNR the shift costs at most about 0.75 dB against TV-long, much less than the 3–4 dB reversals of the 28×28 experiments (E4, E8); in SSIM the deficit is large (-0.095 to -0.190).

False blobs (Tables 29–30). The sign of PnP-const − TV at 1× lesion strength follows the number of angles: negative (or not different from zero) at $K = 30$ and negative at full-dose $K = 60$, mixed at $K = 60$ quarter dose ($\sigma_l = 2$: -7.5; $\sigma_l = 4$: $+9.1$), and positive at $K = 120$ ($+9.0$ to $+13.0$ per phantom at both doses; at 1.5× against TV-long $+10.2$ at full dose and $+3.8$ at quarter dose, both significant). The absolute counts show why (Table 30): the count of TV-long varies from 3.2 to 83.9 per phantom over the four conditions, that of PnP-const only from 12.8 to 50.3; at $K = 30$ TV-long makes two to three times as many false blobs as FBP-Hann (30.2 vs. 15.5 at quarter dose, 83.9 vs. 31.6 at full dose), at $K = 120$ about a third or less (3.2 vs. 8.7, i.e. 37%; 32.1 vs. 118.7, i.e. 27%). So the "excess" of PnP-const at $K = 120$ is relative to a TV that is unusually clean there. Relative to *FBP-Hann* the picture is more informative: at full dose PnP-const has fewer false blobs on the phantoms (-7.1 at $K = 30$, -68.4 at $K = 120$, both significant), but at quarter dose it has more at both K ($+1.0$, not significant, and $+4.0$ $[+2.4, +5.9]$ at 1×; $+2.2$ $[+1.1, +3.4]$ and $+2.7$ $[+1.4, +4.4]$ at 1.5×), in contrast to real slices, where it has significantly fewer than FBP-Hann in both quarter-dose conditions (-9.0 and -4.4). In this sense the shift turns the prior from suppressing to adding false blobs relative to the unregularised reconstruction at quarter dose, by a few blobs per phantom, although this is significant only at $K = 120$. The pooled counts hide a dependence on lesion size: at 1×, for $\sigma_l = 2$ PnP-const has more false blobs than FBP-Hann on the phantoms in three of the four conditions ($+10.1$, $+9.7$ and $+10.6$ at $K = 30$ full and quarter dose and $K = 120$ quarter dose, all significant), for $\sigma_l = 4$ fewer in all four (-24.3, -7.7, -30.9, -2.5). At $K = 120$, full dose, PnP-TVinit with almost no denoising ($\sigma_d = 0.0039$) also produces 14.0 more blobs than TV-long ($[+7.4, +22.0]$ at 1×, for all σ_l), so at least part of the excess at $K = 120$ is not due to the denoiser; we did not isolate the cause. In the gallery of the strongest unsupported blob per phantom (Figure 22, $K = 60$, quarter dose) the worst blobs again occur at the same location in FBP-Hann, TV and PnP, mostly in uniform regions of the phantom, with TV reaching up to 4.9 times the nominal lesion response and the four worst PnP-const blobs between about 2.8 and 3.2 times; in these examples the prior does not create blobs that the data do not already suggest.

5 Discussion

We organise the discussion around the three guiding questions, then state what the evidence says about mechanisms and about evaluation practice. All statements refer to one architecture (a noise-

Table 29: Distribution shift to phantoms (20 phantoms; mean, 95% interval; * excludes 0). Left: false blobs per phantom at 1× lesion strength, PnP-const − TV, by lesion size. Right: PnP-const − TV-long on the phantoms: pooled paired AUC, PSNR (dB) and SSIM. $K = 60$ rows: AUC from four lesion types (E17b), PSNR interval from a bootstrap over the 20 phantoms (`run_e19_k60_psnr_ci.py`), SSIM not stored.

		false blobs, 1×				
K	dose	$\sigma_l = 2$	$\sigma_l = 4$	AUC	PSNR	SSIM
30	1	-70.7^* $[-107.0, -38.7]$	-31.8^* $[-48.7, -16.8]$	-0.037 $[-0.113, +0.037]$	-0.11 $[-0.37, +0.18]$	-0.095^* $[-0.125, -0.063]$
30	0.25	-21.1^* $[-34.3, -10.3]$	-1.8 $[-5.8, +1.8]$	$+0.056$ $[-0.031, +0.144]$	-0.53^* $[-0.64, -0.40]$	-0.190^* $[-0.204, -0.172]$
60	1	-20.1^* $[-34.4, -7.6]$	-6.5^* $[-12.6, -1.4]$	-0.019 $[-0.06, +0.02]$	-0.30^* $[-0.44, -0.14]$	n/a
60	0.25	-7.5^* $[-12.9, -2.8]$	$+9.1^*$ $[+6.0, +12.8]$	$+0.050$ $[-0.01, +0.11]$	-0.75^* $[-0.82, -0.67]$	n/a
120	1	$+13.0^*$ $[+8.8, +17.5]$	$+9.2^*$ $[+5.4, +13.2]$	$+0.075^*$ $[+0.013, +0.137]$	-0.51^* $[-0.56, -0.45]$	-0.179^* $[-0.190, -0.168]$
120	0.25	$+10.4^*$ $[+5.9, +16.0]$	$+9.0^*$ $[+5.9, +12.3]$	-0.062 $[-0.162, +0.031]$	-0.62^* $[-0.68, -0.57]$	-0.155^* $[-0.162, -0.149]$

Table 30: Mean false blobs per lesion-free image at 1× nominal lesion strength (mean of $\sigma_l = 2$ and 4) on real CT slices and on the phantoms, and PnP-const − FBP-Hann (mean, 95% interval over patients or phantoms).

set	K, dose	FBP-Hann	TV	TV-long	PnP-const	PnP-TVinit	PnP-const − FBP
CT	30, 0.25	9.78	5.56	6.06	0.75	0.64	-9.03^* $[-13.21, -5.20]$
CT	30, 1	5.78	1.04	1.69	0.50	0.32	-5.28^* $[-7.54, -3.29]$
CT	120, 0.25	5.79	0.02	0.02	1.38	0.39	-4.41^* $[-7.34, -2.00]$
CT	120, 1	26.79	0.12	0.00	0.24	0.18	-26.55^* $[-43.93, -12.63]$
phantom	30, 0.25	15.48	27.95	30.20	16.50	11.65	$+1.03$ $[-0.68, +2.75]$
phantom	30, 1	31.60	75.72	83.90	24.48	57.30	-7.13^* $[-10.68, -4.02]$
phantom	120, 0.25	8.72	3.08	3.18	12.75	11.48	$+4.03^*$ $[+2.35, +5.90]$
phantom	120, 1	118.68	39.20	32.08	50.30	46.12	-68.38^* $[-92.68, -45.23]$




Figure 22: Strongest unsupported blob per phantom at quarter dose ($K = 60$; $\sigma_l = 4$ px detector) for the four phantoms with the largest false-blob response in PnP-const (top four rows) and in TV (bottom four rows). Columns: ground truth crop, FBP-Hann, TV, PnP-const, PnP-TVinit; red circle: location of the blob; labels give the response as a multiple of the nominal lesion response.

conditioned DnCNN inside PnP-PGD), two training distributions on 28×28 digits and one CT training set; the Limitations section lists the limits.

(1) Does the learned prior keep its advantage over TV when the image distribution changes? Not in accuracy. On 28×28 digits the advantage in distribution is about 3.4 dB over box-constrained TV (+3.8 dB over unconstrained TV; Sections 4.1 and 4.7), and it reverses for an MNIST-trained prior on Fashion-MNIST (-3.1 ± 0.3 dB over TV-ADMM, three seeds) and on the Shepp–Logan phantom (-4.3 ± 0.1 dB), with the same signs against box-constrained TV (Sections 4.3 and 4.7). The reversal persists under Poisson noise (-2.0 to -3.1 dB and -4.4 to -6.1 dB; Section 4.11), although in SSIM it appears on the phantom but not on Fashion-MNIST. A prior trained on Fashion-MNIST beats TV on Fashion-MNIST (+2.2 dB) and, with the same tuned hyperparameters, on MNIST (+1.4 dB) and EMNIST letters (+0.4 dB), and ties on the phantom (-0.2 dB), so what matters is how the training distribution relates to the test distribution, and a broader prior transfers better. On real CT slices the learned prior had no accuracy advantage to lose; its detection advantage over a PSNR-tuned TV with the same iteration budget is retained on non-anatomical phantoms in one of six conditions, against three of six on CT, while its PSNR and SSIM deficit against TV grows (Section 4.18).

(2) Can we find convincing but incorrect structures? Yes, in two settings, and not in others. For the MNIST-trained prior on Fashion-MNIST it produces 0.20–0.66 lesion-like blobs per image, more than TV on seven of eight lesion types even against the box-constrained TV (Section 4.8); the same holds under Poisson noise, in 11 of 12 cases, against the unconstrained TV only (Section 4.12). On non-anatomical phantoms at quarter dose it adds a few false blobs per phantom relative to FBP-Hann (significantly only at $K = 120$) and, at $K = 120$, more than TV (Section 4.18). In distribution the picture is different: on digits with Gaussian noise it produces essentially none (≤ 0.01 per image; 0.06–0.07 for some lesion types at 30 photons per ray) but attenuates inserted lesions, which is itself a form of incorrect content, and on real slices it produces 0.24–1.38 blobs per slice, fewer than FBP-Hann in every condition examined, with the strongest blobs located where FBP-Hann and TV also produce them (Section 4.17). The false-blob count of TV is itself unstable: it varies 26-fold across the phantom conditions and for the smallest lesions exceeds that of the learned prior (Sections 4.8, 4.18), so "more than TV" is a statement about a specific TV solver, tuned for PSNR. We therefore do not claim that the prior hallucinates in general, nor that it is safe: its behaviour moves with noise level, lesion scale and condition, and 40 patients with one noise draw per condition cannot bound rare events.

(3) Do the conclusions hold across noise levels and conditions? Only in part. The advantage in distribution is largest at moderate noise (+3.7 dB at 100–300 photons per ray) and shrinks at both ends of the dose range (Section 4.11); at $\sigma \leq 0.01$ the digit experiments do not determine the ranking, which changes with the data model by up to 5.5 dB (Section 4.5), so we make no claim there. Under angular mismatch the learned prior stays ahead throughout the tested range, but its margin falls from +3.86 to +2.98 dB for 4° jitter and to +1.49 dB for a 4° offset; the latter is mostly a rigid rotation of the reconstruction that a pixel-wise metric penalises (Section 4.2), so we do not read it as evidence that the learned prior is the less robust method. On CT the signs of the PSNR, SSIM and false-blob differences against TV-long change with dose and number of angles without a monotone trend (Sections 4.15, 4.17); the sign of the detection advantage over TV-long is the same in all nine conditions (significant in six), but it is specific to a PSNR-tuned TV with the iteration budget of PnP, since against the 150-iteration TV it is negligible at full dose

for $K = 30$ and 60 (+0.006 and 0.000; Section 4.16).

Accuracy depends on the operating regime and on the baseline. Reporting a single accuracy number for a learned reconstruction method without its regime, data model and baseline is misleading. The headline gap on digits moves by about 0.5 dB with the choice between unconstrained and box-constrained TV; at low noise it changes sign with the data model; and on CT it depends on the iteration budget given to TV and on tuning details: the apparent deficit of PnP at $K = 60$ fell from 0.85 to 0.17 dB at full dose when the tuning set grew from 6 to 20 slices and the starting image was changed (Section 3.16). On clinical-size slices the advantage found on digits disappears: PnP-const differs from TV-long by -0.93 to $+0.33$ dB, with a lower SSIM in eight of nine conditions (significantly in seven), and both iterative methods improve on FBP-Hann (0.9–1.4 dB for PnP-const, 0.8–2.3 dB for TV-long), so the gain over FBP is a gain from regularisation in general (Section 4.15). The best PnP variant, started from the TV solution, is highest in PSNR in seven of nine conditions by at most 0.49 dB, but variants were introduced after the first PnP-const result and, at $K = 120$ and full dose, its tuned σ_d implies almost no denoising (Sections 4.14, 4.15). We did not test why the digit and CT results differ; candidate explanations are the more varied anatomy, which makes TV a weaker model than for digits, a denoiser trained on only 8 960 of 35 820 slices, few PnP iterations, and a TV given twice the iterations.

Early stopping acts as regularisation. In PnP-PGD with an unconstrained denoiser the accuracy peaks after a handful of iterations and then degrades, while the data residual stays at the noise level (0.994 to 1.019 times its expected value between the selected iterate and $k = 400$; Section 4.4). Of the error growth, 74% lies in directions with $0.1 \leq s_i/s_1 < 0.3$, which the data step corrects only slowly: with $\eta = 1/L$ it contracts the error in direction i by $(1 - (s_i/s_1)^2)^k$. Early stopping therefore chooses how much of these directions is left to the denoiser, and the iteration count is a regularisation hyperparameter, not an implementation detail. That this is semi-convergence, in which further iterations fit noise in poorly determined directions, is our interpretation, not a tested mechanism: the selected k is also 5 at $\sigma = 0$ and we did not record the iteration curve on noise-free data. The behaviour is consistent with the observation that convergence guarantees for PnP need constraints on the denoiser [16]; we did not measure the Lipschitz constant of ours. In distribution the converged iterate is still better than TV on MNIST, whereas under shift its error grows to 3.7–4.8 times that of TV (Section 4.6).

The reversal under shift is a property of the prior, but not of the null space alone. The matched-prior control isolates the denoiser: in all runs the tuned hyperparameters are identical and only the weights differ (Section 4.7), and the effect exceeds the seed-to-seed variation (≤ 0.3 dB) by an order of magnitude and persists under four data models. It is not removed by re-tuning within our grids: on Fashion-MNIST no PnP setting among 45 reaches TV's PSNR (Section 4.9). The error decomposition adds that the reversal is not simply null-space invention: 93% (Fashion-MNIST) and 89% (phantom) of PnP's excess error over TV lies in the well-determined subspace, and the signature of a prior that fills the weakly determined subspace with up to three times the true energy, misaligned with the truth, appears only on the phantom (Section 4.6). "Hallucination" in the null-space sense is therefore an incomplete description of the failure.

PSNR is not task performance, in either direction. In distribution on digits the PSNR-tuned PnP reconstruction is better than TV by about 4 dB but detects small inserted lesions less reliably (paired AUC 0.04–0.23 below TV, also below the box-constrained TV), because it

attenuates them (Section 4.8). Re-tuning trades erasure against invention: for the MNIST-trained prior no operating point is competitive with TV on PSNR, detection and false structure at once, on MNIST or on Fashion-MNIST; the Fashion-trained prior does leave such a point on both Fashion-MNIST and MNIST (about +1.0 dB, a paired AUC 0.02–0.06 below TV and few false lesions), so what decides it is the breadth of the training distribution, not whether a test set is formally in distribution (Section 4.9). On CT the sign is reversed: PnP detects better than TV-long in all nine conditions yet no better than FBP-Hann, and the PSNR-tuned TV-long, 0.8–2.3 dB better than FBP-Hann in PSNR, detects worse than it in all nine (significantly in seven) (Section 4.16). A comparison reporting only PSNR or SSIM would have ranked the methods in an order that is wrong for detection, whichever method it favours. We did not test why the sign of the detection difference differs between digits and CT; the lesion scale in pixels, the noise texture and the smoothing of the PSNR-tuned operating points are candidates. The lesions are small and of high contrast in the clinical sense and the observer knows the location, so the conclusions for low-contrast lesions, other observers and search tasks are untested.

Evaluation protocol changes the answer. Besides the choice of TV and its iteration budget, three protocol choices moved results by more than the effects of interest: the inverse crime changes the low-noise gap by up to 5.5 dB; a single denoiser instance would have reported anywhere between -2.83 and -3.43 dB on Fashion-MNIST depending on the seed; and operating points for the lesion task were chosen on held-out images, since choosing them on evaluation images would overstate the benefit of re-tuning. These observations are consistent with reports of instabilities in learned reconstruction [2, 8], with the caveat that those concern other methods. Replacing the dense operator by a matrix-free projector with an exact adjoint changes no reconstruction by more than 3×10^{-4} in relative norm, and replacing Gaussian by Poisson noise leaves the qualitative picture unchanged (Sections 4.10, 4.11); the Poisson-weighted data term showed no consistent benefit, so we do not claim more than that the unweighted scheme is adequate in this setting. In our experience the following should be standard in studies of learned priors: a baseline with the same iteration budget and constraints, hyperparameters tuned on held-out data for every method, several training seeds, at least one mismatched and one matched prior, a task-based metric next to PSNR and SSIM, and an explicit data model.

6 Conclusion

We compared classical and learned regularization for sparse-view CT, using a noise-conditioned DnCNN in Plug-and-Play proximal gradient against FBP, Tikhonov and TV, under noise, undersampling, forward-model mismatch, distribution shift and a lesion-detection task, first on 28×28 digits with Gaussian and Poisson noise and then on real LoDoPaB-CT slices.

On digits the learned prior is about 3.4 dB better than box-constrained TV at moderate noise (+3.8 dB over unconstrained TV), with a smaller margin at the extremes of the noise range, and the ranking near the noise-free limit depends on the data model. The advantage reverses when the test images differ from the training distribution; the matched-prior control, in which only the denoiser weights change, shows that this is a property of the prior and not of the PnP scheme, and the error decomposition shows that most of the excess error lies in the well-determined directions, not only in the null space. PSNR is not task performance: in distribution the PSNR-tuned prior attenuates small lesions and detects them less reliably than TV, and under mismatch it also produces more lesion-like blobs than TV for most lesion types. Re-tuning trades these against each other, and a point competitive with TV on all three criteria exists only for the broader of our two priors. The

number of iterations acts as a regulariser: the data residual stays at the noise level while the error grows in weakly determined directions, which we interpret as semi-convergence but did not test.

On real CT slices the learned prior does not improve on a well-tuned TV with the same iteration budget in PSNR (differences between -0.93 and $+0.33$ dB over nine dose and angle conditions) and has a lower SSIM in eight of nine. It detects small high-contrast lesions better than that TV in all nine conditions (significant in six) but not better than unregularised FBP, whereas the PSNR-tuned TV detects them worse than FBP in all nine. Its false-blob counts on real slices are below those of FBP-Hann and not systematically different from those of TV. Under a shift to non-anatomical phantoms its PSNR and SSIM deficit against TV is consistent, its detection advantage is demonstrable in one of six conditions, and its false-blob count relative to TV changes sign with the number of angles.

Our conclusions concern one architecture, two training distributions on digits and one CT training set, and they are conditional on the regime, the TV variant and the tuning protocol. They support evaluating learned priors against baselines with the same iteration budget and constraints, tuned on held-out data, with several training seeds, mismatched and matched priors and task-based metrics next to PSNR and SSIM. Future work should use all 60 test patients with several noise realisations, a finer TV and Hann grid, several denoiser seeds and the full training set, box-constrained TV in every experiment, clinically calibrated lesion contrasts with model observers, re-tuning under forward-model mismatch, an iteration curve on noise-free data, and provably convergent PnP variants.

Limitations

Scope. The digit experiments are intentionally small (28×28) so that a thorough grid is possible; trends need not transfer to clinical resolution, and the CT study adds only one architecture (a DnCNN trained for one run on 8 960 of the 35 820 training slices), so it does not separate the prior from this network. Only two training distributions were compared, so the explanation of the asymmetric transfer between them is a hypothesis. The PnP scheme has no convergence guarantee, we did not measure the Lipschitz constant of the denoiser, and only moderate forward-model mismatches (angles up to 4°, one jitter draw per level) were evaluated, with hyperparameters frozen at their nominal values.

Baselines and tuning. PnP has two tuned hyperparameters and TV one. The box-constrained TV control (E2b) was run for E1–E2, E4, E6 (bicubic variant only), E8 (the MNIST- and Fashion-trained families, not the second MNIST family) and E9; the TV baseline of E3, E5, E7, E10 and of the Poisson study (E12–E13) is the unconstrained TV-ADMM, so the corresponding gaps may be overstated; in the Gaussian MNIST case the box constraint lowered the gap by about 0.5 dB (E8). On CT, TV is tuned on a grid with steps of a factor three, TV-long is sometimes worse than TV, PSNR was the tuning criterion for every method, which favours accuracy over detectability, and two optima of PnP-TVinit lie on the boundary of the grid. The PnP variants of E15 were introduced after the first PnP-const result.

Data and simulation. E3, E5 and the original E4 use data generated with the reconstruction operator (an inverse crime); E3 and E5 were not repeated on finer-grid data, and the sensitivity band of Table 9 shows that the near noise-free ranking depends on this choice, which we adopted as a convention after seeing all variants. The Poisson study uses a single attenuation scale ($\mu = 2$), one angle count and one MNIST-trained denoiser in the dose sweep, and its Shepp–Logan result

is one image with one noise draw per dose. The matrix-free operator was validated against the dense operator only at 28×28 (agreement limited by single-precision arithmetic, timings from single runs). The CT simulator was checked against real data on 8 slices and one noise draw per condition, and only for the TV baseline; the real data have more variance than the simulation at high counts, and we did not verify that the PnP–TV difference agrees between real and simulated data. LoDoPaB's ground truth is itself a reconstruction.

Statistics. Seed variation comes from three denoisers per digit configuration, which gives only a crude standard deviation; the CT study has 40 test patients (one slice each) and one noise realisation per condition, so a single cell has an AUC uncertainty of about ± 0.05, and the conditions share the same patients and are not independent. No correction for multiplicity was applied. The analysis was developed in steps after seeing results (Section 3.16): the false-blob definition, the threshold sweep and the shift extension to $K = 30, 120$ were not pre-specified, and the Poisson lesion experiment was run at only two of the five doses ($I_0 = 100$ and 30).

Task-based evaluation. The lesions are bright additive Gaussian blobs (on CT of 120–240 HU under an assumed mapping from $\mu/\mu_{\max}$ to HU), the observer is a location-known local-contrast statistic, false blobs are defined by this statistic and a threshold, and no reader study was done, so false-lesion rates are meaningful for comparing methods but not as absolute rates. The observability analysis is confounded with lesion size and location. The operating-point study (E10) uses one denoiser per prior, two lesion types and a coarse grid, and the galleries of false blobs show the worst cases selected by the methods' own responses.

Shift. The shift study covers a single kind of shift: digits to other digits and a phantom, and CT slices to 20 random piecewise-constant phantoms, which favour TV; no change of scanner geometry, dose model or anatomy was tested. The $K = 60$ shift analysis used four lesion types and the other conditions two, and SSIM on the phantoms at $K = 60$ was not stored. The error-decomposition statistics cannot separate prior-driven structure from amplified noise except in comparisons with identical noise and settings.

Reproducibility

Code, trained models, all raw result files and figures are available at https://github.com/AboOraig/learned-regularization. The experiments of this version are run, in order, by `run_nullspace.py` (E6, E7), `run_spectral.py` (iteration and spectral analysis), `run_step4.py` and `summarize_step4.py` (E8), `run_lesions.py` and `summarize_step5.py` (E9), and `run_oppoint.py` (E10); `run_stage2.py` (E11, E12) and `run_lesions_poisson.py` with `summarize_step5.py` (E13); `run_stage3_step1.py` (data, patient split, geometry, E14), `run_stage3_step2.py` (noise model, FBP and TV baselines, extraction of training slices), `train_denoiser_ct.py` (denoiser), `run_stage3_step3.py` and `run_stage3_step3b.py` (E15; the first is superseded), `run_stage3_step4.py`, `run_stage3_step4b.py` and `run_stage3_step4c_compare.py` (E17b, E18, E19 at $K = 60$; use `--out results/stage3/s4c --n_val 20 --auto_extend 2` for the final version), `run_stage3_step5.py` and `run_stage3_step5_summary.py` (E16–E18), and `run_stage3_step5_shift.py` with `run_stage3_step5_shift_summary.py` (E19 at $K = 30, 120$). The controls added in this version are `run_tv_box_control.py` (E1–E2 box-constrained TV), `run_tv_box_shift.py` (E4, E8), `run_tv_box_lesions.py` (E9), `run_e3_rotation_check.py` (E3), `run_e11_svd_check.py` (E11) and `run_e19_k60_psnr_ci.py` (E19); the figures of E1–E2 and

E5 are drawn from the stored result files by `make_fig_e1e2_gap.py` and `make_fig_e5_iterations.py`. The unit tests (68 in total) run with `python -m pytest tests`.

References

[1] Jonas Adler and Ozan Öktem. Learned primal-dual reconstruction. *IEEE Transactions on Medical Imaging*, 2018.

[2] Vegard Antun, Francesco Renna, Clarice Poon, Ben Adcock, and Anders C. Hansen. On instabilities of deep learning in image reconstruction and the potential costs of AI. *Proceedings of the National Academy of Sciences*, 117(48), 2020.

[3] Harrison H. Barrett and Kyle J. Myers. *Foundations of Image Science*. Wiley, 2004.

[4] Harrison H. Barrett, Jie Yao, Jannick P. Rolland, and Kyle J. Myers. Model observers for assessment of image quality. *Proceedings of the National Academy of Sciences*, 1993.

[5] Somesh Bhadra, V. A. Kelkar, Frank J. Brooks, and Mark A. Anastasio. On hallucinations in tomographic image reconstruction. *IEEE Transactions on Medical Imaging*, 40(11), 2021.

[6] A. Chambolle and T. Pock. A first-order primal-dual algorithm for convex problems with applications to imaging. *Journal of Mathematical Imaging and Vision*, 40(1):120–145, 2011.

[7] Stanley H. Chan, X. Wang, and O. A. Elgendy. Plug-and-play ADMM for image restoration: Fixed-point convergence and applications. *IEEE Transactions on Computational Imaging*, 3 (1), 2017.

[8] M. Z. Darestani, A. S. Chaudhari, and Reinhard Heckel. Measuring robustness in deep learning based compressive sensing. In *International Conference on Machine Learning (ICML)*, 2021.

[9] Nathan Halko, Per-Gunnar Martinsson, and Joel A. Tropp. Finding structure with randomness: Probabilistic algorithms for constructing approximate matrix decompositions. *SIAM Review*, 53(2):217–288, 2011.

[10] Kyong Hwan Jin, Michael T. McCann, Emmanuel Froustey, and Michael Unser. Deep convolutional neural network for inverse problems in imaging. *IEEE Transactions on Image Processing*, 2017.

[11] Jari Kaipio and Erkki Somersalo. *Statistical and Computational Inverse Problems*. Springer, 2005.

[12] Ulugbek S. Kamilov, Charles A. Bouman, Gregery T. Buzzard, and Brendt Wohlberg. Plug-and-play methods for integrating physical and learned models in computational imaging. *IEEE Signal Processing Magazine*, 2023.

[13] Johannes Leuschner, Michael Schmidt, Daniel Otero Baguer, and Peter Maass. LoDoPaB-CT, a benchmark dataset for low-dose computed tomography reconstruction. *Scientific Data*, 8: 109, 2021.

[14] Yaniv Romano, Michael Elad, and Peyman Milanfar. The little engine that could: Regularization by denoising (RED). *SIAM Journal on Imaging Sciences*, 2017.

[15] Leonid I. Rudin, Stanley Osher, and Emad Fatemi. Nonlinear total variation based noise removal algorithms. *Physica D: Nonlinear Phenomena*, 60(1–4), 1992.

[16] Ernest K. Ryu, J. Liu, S. Wang, X. Chen, Z. Wang, and Wotao Yin. Plug-and-play methods provably converge with properly trained denoisers. In *International Conference on Machine Learning (ICML)*, 2019.

[17] Yang Song, Liyue Shen, Lei Xing, and Stefano Ermon. Solving inverse problems in medical imaging with score-based generative models. In *International Conference on Learning Representations (ICLR)*, 2022.

[18] Singanallur V. Venkatakrishnan, Charles A. Bouman, and Brendt Wohlberg. Plug-and-play priors for model based reconstruction. In *IEEE Global Conference on Signal and Information Processing (GlobalSIP)*, 2013.

[19] Armand Wirgin. The inverse crime, 2004.

[20] Kai Zhang, Wangmeng Zuo, Yunjin Chen, Deyu Meng, and Lei Zhang. Beyond a gaussian denoiser: Residual learning of deep CNN for image denoising. *IEEE Transactions on Image Processing*, 26(7), 2017.

[21] Kai Zhang, Wangmeng Zuo, Shuhang Gu, and Lei Zhang. Learning deep CNN denoiser prior for image restoration. In *IEEE Conference on Computer Vision and Pattern Recognition (CVPR)*, 2017.

[22] Kai Zhang, Wangmeng Zuo, and Lei Zhang. FFDNet: Toward a fast and flexible solution for CNN-based image denoising. *IEEE Transactions on Image Processing*, 2018.