

SAFETY

The device is intended for supervised rehabilitation research and is not a diagnostic device.
Stop use if skin irritation or pain occurs.

CALIBRATION

Calibrate the IMU while stationary for 10 seconds. Zero the force-sensitive resistor with no load before each session.