

METHODS

The prototype used a six-axis IMU mounted on the shank and sampled motion at 100 Hz. A force-sensitive resistor under the heel recorded foot contact.

RESULTS

Ten healthy volunteers completed the walking protocol. Mean step-count error was 3.2 percent, and measured battery life was 6 hours.