


BP3L Function

Note: Api call must be serial call, that is, after the end of a Api, you can call the next Api, otherwise it will appear Api function failure.




Note: The above 3 functions are parallel relations, there is no order.

BP3L Function A – Online Measurement Function

Note: 1. Api call must be serial call, that is, after the end of a Api, you can call the next Api, otherwise it will appear Api function failure.


2. In the BP3L measurement process, only to receive the stop measurement function (A-2), other Api failure.



A-1: `commandStartMeasureWithUser: clientID:`

A-2: `stopBPMeasureErrorBlock: errorBlock:`


BP3L Function B –Query Functions



B-1: commandFunction: errorBlock:

B-2: commandEnergy: errorBlock:

BP3L Function C –Disconnect Device



C-1: `commandDisconnectDevice`