

2025L 2389

Format No. : CI05

DATE:

15-Sep-2025



SHRETA LAB TECH

CALIBRATION CERTIFICATE



CC-2403

SHEET:

1 OF 2

ULR-CC240325000076001F

CERTIFICATE NUMBER : SLT/25/09/742/011
 CUSTOMER NAME : M/S.KSB MIL CONTROLS LIMITED, KERALA
 CUSTOMER REFERENCE NUMBER : ONSITE
 DESCRIPTION OF THE ITEM : DIGITAL STOP WATCH
 CALIBRATION PROCEDURE NUMBER : SHRETA-WI-ET-06
 CALIBRATION FORMAT NO. : SHRETA/FT-25-05
 CONDITION OF RECEIVED ITEM : GOOD
 CALIBRATED AT : IN LAB
 MAKE/IDENTIFICATION NUMBER : RACER / M509
 DATE OF CALIBRATION : Monday, September 15, 2025
 DATE OF ISSUE : Monday, September 15, 2025
 CALIBRATION NEXT DUE DATE : Monday, September 14, 2026
 PARAMETER CALIBRATED : HOUR/MINUTE/SECONDS.
 RANGE : 0 TO 24 Hrs
 UNCERTAINTY OF MEASUREMENT : $\pm 0.61s$ TO $7.7s$
 (EXPANDED UNCERTAINTY IS REPORTED AT
 95% CONFIDENCE LEVEL WITH COVERAGE FACTOR $k=2$)
 REFERENCE STANDARD : IS 10996 : 1984
 TEMPERATURE DURING CALIBRATION : $25^{\circ}\text{C} \pm 4^{\circ}\text{C}$
 HUMIDITY DURING CALIBRATION : 30 TO 75% RH
 DESCRIPTION : DIGITAL TIME CALIBRATOR
 MAKE : GLAXO INFRASTRUTURE
 IDENTIFICATION NUMBER : ESLT-011
 TRACEABILITY : MSIR
 CERTIFICATE NO. : MSIR/410725/536-010
 CALIBRATION DATE : 23.07.2025
 NEXT CALIBRATION DUE : 22.07.2026
 ALL ABOVE MENTIONED MASTERS ARE TRACEABLE TO NATIONAL STANDARDS

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY:

M. VINCENT
Technical Manager



AUTHORISED SIGNATORY

R. GLEETUS
Quality Manager

Mobile : 94863 33444, 94888 77777

e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641 009.

DATE : 15-Sep-2025

SHEET : 2 OF 2



SHRETA LAB TECH

CALIBRATION CERTIFICATE



CUSTOMER NAME : M/S.KSB MIL CONTROLS LIMITED, KERALA
 CERTIFICATE NUMBER : SLT/25/09/742/011
 MAKE/IDENTIFICATION NUMBER : RACER / M509

ALL VALUES ARE IN Hrs Min Sec

Sl.No	UUC READINGS IN Hrs./Min./Sec.	MASTER READING in Hrs./Min./Sec.	ERROR OBSERVED IN ($\pm$ Hrs./Min./Sec.)
1	0.00'.00"	0.00'.00"	0.00'.00"
2	0.10'.00"	0.10'.00"	0.00'.00"
3	0.20'.00"	0.20'.00"	0.00'.00"
4	0.30'.00"	0.30'.00"	0.00'.00"
5	0.40'.00"	0.40'.00"	0.00'.00"
6	0.50'.00"	0.50'.00"	0.00'.00"
7	1.00'.00"	1.00'.00"	0.00'.00"
8	2.00'.00"	2.00'.00"	0.00'.00"
9	4.00'.00"	4.00'.01"	0.00'.01"
10	8.00'.00"	8.00'.01"	0.00'.01"
11	12.00'.00"	12.00'.01"	0.00'.01"
12	16.00'.00"	16.00'.02"	0.00'.02"
13	20.00'.00"	20.00'.02"	0.00'.02"
14	24.00'.00"	24.00'.02"	0.00'.02"

STATUS OF CALIBRATION : ACCEPTED

PERMISSIBLE ERROR AS PER MANUFACTURER STANDARD 5 Sec

NOTE 1 : NEXT CALIBRATION DUE AS PER CUSTOMER'S PLAN AND REQUIREMENT.

NOTE 2 : THE REPORTED RESULTS ARE VALID ONLY FOR THE CONDITION OF THE RECEIVED DIGITAL TIMER AT THE TIME OF AND UNDER THE STATED CONDITION OF CALIBRATION.

NOTE 3 : NO CORRECTION OR ADJUSTMENT IS MADE BEFORE OR AFTER CALIBRATION AS FOUND AS DATA IS FOUND IN THE REPORT.

-End-

Verified and Found OK

RANJITH. K.K
For KSB MIL Controls Ltd.

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY :

M.VINCENT
Technical Manager



AUTHORISED SIGNATORY

R. GLEETUS
Quality Manager

Mobile : 94863 33444, 94888 77777

e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641 009.



SHRETA LAB TECH

CALIBRATION CERTIFICATE

DATE : 15-Sep-2025

SHEET : 1 OF 3

CERTIFICATE NUMBER : SLT/25/09/460/001
CUSTOMER NAME : M/S.KSB MIL CONTROLS LIMITED, KERALA-680 741.
CUSTOMER REFERENCE NUMBER : ONSITE CALIBRATION
IDENTIFICATION NUMBER : E 144
MAKE : PC - PROGETTI
MODEL : SKA 100
RANGE : FLOW - 3,20 & 60 LPM. PRESSURE - 1200 bar
DATE OF CALIBRATION : Monday, September 15, 2025
DATE OF ISSUE : Monday, September 15, 2025
CALIBRATION OF NEXT DUE DATE : Monday, September 14, 2026
DESCRIPTION OF INSTRUMENT : LEAK TESTING MACHINE

DETAILS OF MASTERS USED FOR CALIBRATION

SL.NO	MASTER DESCRIPTION	MASTER CALIBRATED AT	CALI.CC.NO.	NEXT DUE DATE
1	DIGITAL PRESSURE CALIBRATOR	ESLT006	SLT/24/06/606/003	24.06.2026
2	STOP WATCH	ESLT011	MSIR/410724/437-003	21.03.2026
3	STANDARD WEIGHT BOX	ESLT-112	TSC/23-24/319/1	30.01.2026

ABOVE MASTER IS TRACEABLE TO NATIONAL & INTERNATIONAL STANDARDS

ENVIRONMENTAL CONDITIONS : AMBIENT.

UNCERTAINTY OF MEASUREMENT AT : ± 0.4 % FLOW AND 0.2% PRESSURE

95% CONFIDENCE LEVEL

REFERENCE STANDARD : MANUFACTURER SPECIFICATION

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY :

CHECKED BY :

APPROVED BY :

Mobile : 94863 33444, 94888 77777

e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641 009.



SHRETA LAB TECH

CALIBRATION CERTIFICATE

DATE: 15-Sep-2025

SHEET: 2 OF 3

CUSTOMER NAME : M/S.KSB MIL CONTROLS LIMITED, KERALA-680 741.

CERTIFICATE NUMBER : SLT/25/09/460/001.

IDENTIFICATION NUMBER : E 144

FLOW METER CALIBRATION**A. FLOW METER OF 3 LPM**

THEORETICAL FLOW IN MACHINE	STOP WATCH STARTING READING (Sec)	STOP WATCH ENDING READING (Sec)	TOTAL TIME (Sec)	WEIGHED OUTPUT (Kg)	CALCULAT ED FLOW IN LITERS	ERROR OBSERVED LITERS	ERROR in %
1.000 Liters	0	20	20	0.99	0.98	-0.020	-2.0
2.000 Liters	0	40	40	1.99	1.97	-0.030	-1.5
3.000 Liters	0	60	60	3.02	3.03	0.030	1.0

B. FLOW METER OF 20 LPM

THEORETICAL FLOW IN MACHINE	STOP WATCH STARTING READING (Sec)	STOP WATCH ENDING READING (Sec)	TOTAL TIME (Sec)	WEIGHED OUTPUT (Kg)	CALCULAT ED FLOW IN LITERS	ERROR OBSERVED LITERS	ERROR in %
6.666 Liters	0	20	20	6.666	6.662	-0.004	-0.06
13.333 Liters	0	40	40	13.325	13.322	-0.011	-0.08
19.998 Liters	0	60	60	19.998	19.982	-0.016	-0.08

C. FLOW METER OF 60 LPM

THEORETICAL FLOW IN MACHINE	STOP WATCH STARTING READING (Sec)	STOP WATCH ENDING READING (Sec)	TOTAL TIME (Sec)	WEIGHED OUTPUT (Kg)	CALCULAT ED FLOW IN LITERS	ERROR OBSERVED LITERS	ERROR in %
20.000 Liters	0	20	20	20.03	20.04	0.040	0.200
40.000 Liters	0	40	40	40.02	40.03	0.030	0.075
60.000 Liters	0	60	60	60.04	60.06	0.060	0.100

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY :

CHECKED BY :

APPROVED BY :

Mobile : 94863 33444, 94888 77777

e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641 009.

15-Sep-2025



SHRETA LAB TECH

CALIBRATION CERTIFICATE

DATE :

3 OF 3

SHEET :

CUSTOMER NAME : M/S.KSB MIL CONTROLS LIMITED, KERALA-680 741.
 CERTIFICATE NUMBER : SLT/25/09/460/001
 IDENTIFICATION NUMBER : E 144

ALL READINGS ARE IN bar

SL.NO	INDICATED READING	MASTER READING				MAXIMUM ERROR OBSERVED	OBSERVED ERROR IN % FSD
		INCREASING	ERROR OBSERVED	DECREASING	ERROR OBSERVED		
1	0.00	0.00	0.00	0.00	0.00	0.00	0.00
2	250.00	250.44	-0.44	250.40	-0.40	-0.44	-0.06
3	500.00	500.66	-0.66	500.62	-0.62	-0.66	-0.09
4	750.00	750.76	-0.76	750.71	-0.71	-0.76	-0.10
5	900.00	900.86	-0.86	900.86	-0.86	-0.86	-0.11

STATUS OF CALIBRATION : TO BE DECIDED BY CUSTOMER

NOTE 1 : NEXT CALIBRATION DUE AS PER CUSTOMER'S PLAN AND REQUIREMENT.**NOTE 2 :** THE REPORTED RESULTS ARE VALID ONLY FOR THE CONDITION OF THE RECEIVED LEAK TESTING MACHINE AT THE TIME OF AND UNDER THE STATED CONDITION OF CALIBRATION.**NOTE 3 :** NO CORRECTION OR ADJUSTMENT IS MADE BEFORE OR AFTER CALIBRATION AS FOUND AS DATA IS FOUND IN THE REPORT.

-- END OF THE REPORT --

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY :

CHECKED BY :

APPROVED BY :

Mobile : 94863 33444, 94888 77777

e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641 009.

BATCH TEST CERTIFICATE

Date : 23/04/2024

Product : SKL – SP1

Batch No : 24D27

Mfd. in : APRIL 2024

We hereby certify that when tested at time of manufacture, the above material:

1. Meets the requirements of the following specifications:

- A. ASME Boiler and Pressure Vessel Code, 2004, 2007, 2010, 2013, 2015, 2017, 2019 and 2021 Edition, Section V, Nondestructive Examination, including 2005, 2006, 2008, 2009b, and 2011a Addenda, Article 6 Paragraph T-641 and Article 24 as applicable.
- B. ASME Boiler and Pressure Vessel Code, 1995, 1998 and 2001 Edition, Section V Nondestructive Examination, including 1999, 2000, 2002 and 2003 Addenda, Article 6 Paragraph T-640 and Article 24 as applicable.
- C. ASME Boiler and Pressure Vessel Code, 1986, 1989 and 1992 Edition, Section V, Nondestructive Examination, Article 6 including 1992 Addenda, Paragraph T-625, 1993 Addenda Paragraph T-640 and Article 24 as applicable.
- D. ASTM E-165-92, ASTM E-165-94, ASTM E-165-95, ASTM E-165-02, ASTM E-165-09, ASTM E-165/E-165M-12, ASTM E-165/E-165M-18, Paragraph 7.1,
- E. ASTM E 1417, Paragraph 5.1.

2. Meets the requirements of **NUCLEAR POWER CORPORATION OF INDIA** for materials for Dye penetrant Inspection process.

The following test results were obtained:

- | | | | |
|---------------------------------|---|-----|--------|
| 1. Sulphur | : | <25 | ppm |
| 2. Total Chlorine + Fluorine | : | <25 | ppm |
| 3. Cleaner residue (see note 2) | : | NA | g/100g |

Penetrant Tests:

Flash Point (PMCC)	:	>93 °C
Viscosity, (cs. Nominal)	:	3.16 cST @100° F
Water Tolerance	:	NA
Removability,	:	PASS

3. We further certify that this material does not contain mercury as a basic element and no mercury bearing equipment was used in its manufacture.

Bhaskar Veeravalli

Quality Assurance

Notes:

- 1. Batch numbers are printed on the base of aerosol cans and bulk containers
- 2. The above certification gives the results obtained at the time of manufacture. Age and use may alter the properties of any material

Form No: F/QC/008/14 Rev: 072023



ITW India Private Limited, Plot No.51, 52, 207 & 208, Phase II, IDA, TSIC, Pashamylaram, Sangareddy Dist.- 502 307, Telangana State, India
Telephone: +91 8455 224710, Facsimile: +91 8455 224709, Email: magnafluxinfo@magnafluxindia.com, Website: www.magnaflux.in

Regd. Office: ITW India Private Limited, Level 1, Lotus Plaza, 732/1, Mehrauli Gurgaon Road, Sector - 14, Gurgaon - 122 001, Haryana, India.
Telephone : +91 - 124 - 4284381, Facsimile : +91 - 124 - 4284389, E-mail : infoindia@itw.com

Corporate Identity Number (CIN) : U32301HR1979PTC038643

BATCH TEST CERTIFICATE

Date : 25/06/2025

Product : SKD-S2

Batch No : 25F08

Mfd. in : JUNE 2025

We hereby certify that when tested at time of manufacture, the above material:

1. Meets the requirements of the following specifications:

- A. ASME Boiler and Pressure Vessel Code 2004, 2007, 2010, 2013, 2015, 2017, 2019, 2021 and 2023 Edition, Section V, Nondestructive Examination, including 2005, 2006, 2008, 2009b, and 2011a Addenda, Article 6 Paragraph T-641 and Article 24 as applicable.
- B. ASME Boiler and Pressure Vessel Code, 1995, 1998 and 2001 Edition, Section V Nondestructive Examination, including 1999, 2000, 2002 and 2003 Addenda, Article 6 Paragraph T-640 and Article 24 as applicable.
- C. ASME Boiler and Pressure Vessel Code, 1986, 1989 and 1992 Edition, Section V, Nondestructive Examination, Article 6 including 1992 Addenda, Paragraph T-625, 1993 Addenda Paragraph T-640 and Article 24 as applicable.
- D. ASTM-165/E-165M-23, Paragraph 7.1,

2. Meets the requirements of **NUCLEAR POWER CORPORATION OF INDIA** for materials for Dye penetrant Inspection process.

The following test results were obtained:

- | | | | |
|---------------------------------|---|-----|--------|
| 1. Sulphur | : | <25 | ppm |
| 2. Total Chlorine + Fluorine | : | <25 | ppm |
| 3. Cleaner residue (see note 2) | : | NA | g/100g |

3. Meets the requirements of EN ISO 3452-2, EN 571-1 and ASTM E 1417 Paragraph 5.1

4. It is hereby certified that the above listed inspection material and batch number meets the requirements of AMS 2644H and is approved by the U.S. Air Force and listed on QPL-AMS-2644.

When tested according to paragraph 4.3.2, Sampling Plan A, the following test results were obtained:

• 4.2.2.1 Penetrant Tests:

Flash Point (PMCC), 3.3.3	NA ° F
Viscosity, 3.3.4 (cs. Nominal)	NA @100° F
Fluorescent Brightness, 3.3.8.3.2 (FP-4PE Standard)	NA %
Water Tolerance (Method A, B, C, D), 3.3.8.5	NA
Removability, 3.3.8.6	NA

• 4.2.2.1 Emulsifier Tests:

Flash Point (PMCC), 3.3.3	NA ° F
Viscosity, 3.3.4 (cs. Nominal)	NA cs@100° F
Water Content (Method B Only), 3.3.9.6	NA %

• 4.2.2.3 Developer Tests:

Developer Fluorescence, 3.3.10.2	PASS
Developer Removability, 3.3.10.4	PASS
Redispersibility, 3.3.10.5	PASS

• 3.3.11.4 Remover Tests:

Penetrant Removal, 4.4.11.2	NA
-----------------------------	----

5. We further certify that this material does not contain mercury as a basic element and no mercury bearing equipment was used in its manufacture.

Bhaskar Veeravalli

Quality Assurance

Notes:

1. Batch numbers are printed on the base of aerosol cans and bulk containers
2. The above certification gives the results obtained at the time of manufacture. Age and use may alter the properties of any material

Page 2 of 2

Form No: F/QC/012/14 Rev: 07 /2023



ITW India Private Limited, Plot No.51, 52, 207 & 208, Phase II, IDA, TSIIIC, Pashamylaram, Sangareddy Dist.- 502 307, Telangana State, India
Telephone: +91 8455 224710, Facsimile: +91 8455 224709, Email: magnafluxinfo@magnafluxindia.com, Website: www.magnaflux.in

Regd. Office: ITW India Private Limited, Level 1, Lotus Plaza, 732/1, Mehrauli Gurgaon Road, Sector - 14, Gurgaon - 122 001, Haryana, India.
Telephone : +91 - 124 - 4284381, Facsimile : +91 - 124 - 4284389, E-mail : infoindia@itw.com

Corporate Identity Number (CIN) : U32301HR1979PTC038643

BATCH TEST CERTIFICATE

Date : 06/12/2025

Product : SKC-I Solvent Cleaner / Remover

Batch No : 25M01

Mfd. in : DECEMBER 2025

We hereby certify that when tested at time of manufacture, the above material:

1. Meets the requirements of the following specifications:

- A. ASME Boiler and Pressure Vessel Code, 2004, 2007, 2010, 2013, 2015, 2017, 2019, 2021 and 2023 Edition, Section V, Nondestructive Examination, including 2005, 2006, 2008, 2009b, and 2011a Addenda, Article 6 Paragraph T-641 and Article 24 as applicable.
- B. ASME Boiler and Pressure Vessel Code, 1995, 1998 and 2001 Edition, Section V Nondestructive Examination, including 1999, 2000, 2002 and 2003 Addenda, Article 6 Paragraph T-640 and Article 24 as applicable.
- C. ASME Boiler and Pressure Vessel Code, 1986, 1989 and 1992 Edition, Section V, Nondestructive Examination, Article 6 including 1992 Addenda, Paragraph T-625, 1993 Addenda Paragraph T-640 and Article 24 as applicable.
- D. ASTM-165/E-165M-23, Paragraph 7.1,
- E. ASTM E 1417-21, Paragraph 5.1.

2. Meets the requirements of **NUCLEAR POWER CORPORATION OF INDIA** for materials for Dye penetrant Inspection process.

3. Meets the requirements of EN ISO 3452-2.

The following test results were obtained:

- | | | | |
|---------------------------------|---|--------|--------|
| 1. Sulphur | : | <25 | ppm |
| 2. Total Chlorine + Fluorine | : | <25 | ppm |
| 3. Cleaner residue (see note 2) | : | <0.001 | g/100g |
4. We further certify that this material does not contain mercury as a basic element and no mercury bearing equipment was used in its manufacture.

5. It is hereby certified that the above listed inspection material and batch number meets the requirements of AMS 2644H and is approved by the U.S. Air Force and listed on QPL-AMS-2644.

When tested according to paragraph 4.3.2, Sampling Plan A, the following test results were obtained:

• 4.2.2.1 Penetrant Tests:

Flash Point (PMCC), 3.3.3	NA _____ ° F
Viscosity, 3.3.4	NA _____ cs@100° F
Fluorescent Brightness, 3.3.8.3.2 (FP-4PE Standard)	NA _____
Water Tolerance (Method A, B, C, D), 3.3.8.5	NA _____
Removability, 3.3.8.6	NA _____

• 4.2.2.1 Emulsifier Tests:


Flash Point (PMCC), 3.3.3	NA _____ ° F
Viscosity, 3.3.4 (cs. Nominal)	NA _____ cs@100° F
Water Content (Method B Only), 3.3.9.6	NA _____ %

• 4.2.2.3 Developer Tests:

Developer Fluorescence, 3.3.10.2	NA _____
Developer Removability, 3.3.10.4	NA _____
Redispersibility, 3.3.10.5	NA _____

• 3.3.11.4 Remover Tests:

Penetrant Removal, 4.4.11.2	PASS _____
-----------------------------	------------



Quality Assurance

Notes:

1. Our batch number appears on the bottom of all aerosol cans and bulk containers.
2. The above certification gives the results obtained at the time of manufacture. Age and use may alter the properties of any material.

EN CERTIFICATION

6. We hereby certify that the Penetrant inspection material Type SKC-I Cleaner, Batch No 25M01 manufactured in DECEMBER 2025 meets the requirements of EN ISO 3452-2 with the following results:

Individual Property	Section	Requirement	Result
Appearance	6.1	Clear	Clear
Sensitivity (30 mic panel)	6.2	Sensitivity Level (1(<75%) or 2 (>=75%)) Penetrant : SKL – SP2 Developer: SKD-S2	Level 2
Density	6.3	0.684 – 0.756 @ 20 deg C	0.74
Residue on evaporation / Solid content	6.13	< 5 mg residue	0.1 mg
Flash point	6.5	NA	NA
Corrosive properties (Mg)	6.11	No evidence of staining, pitting and corrosion	PASS

Flashpoint is not required as per ISO3452-2 : 2006 6.5.1 flash point measurement only required for material with nominal flash of 20 deg C and 110 deg C. SKC-I flash point is below 20 deg C.

Testing in accordance with 5.4.2 Table 3 , as applicable.

Bhaskar Veeravalli

Quality Assurance