



SHRETA LAB TECH

CALIBRATION CERTIFICATE

DATE : 13-May-2026

SHEET : 1 OF 3

CERTIFICATE NUMBER : SLT/26/05/240/001

CUSTOMER NAME : M/S.KSB MIL CONTROLS LIMITED, KERALA-680 741.

CUSTOMER REFERENCE NUMBER : ONSITE CALIBRATION

IDENTIFICATION NUMBER : E 144

MAKE : PC - PROGETTI

MODEL : SKA 100

RANGE : FLOW - 3,20 & 60 LPM. PRESSURE - 1200 bar ,
TIME : 0 - 16 min

DATE OF CALIBRATION : Wednesday, May 13, 2026

DATE OF ISSUE : Wednesday, May 13, 2026

CALIBRATION OF NEXT DUE DATE : Wednesday, May 12, 2027

DESCRIPTION OF INSTRUMENT : LEAK TESTING MACHINE

DETAILS OF MASTERS USED FOR CALIBRATION

SL.NO	MASTER DESCRIPTION & RANGE / L.C	ID NUMBER	CALIBRATION AT (NABL LAB ID)	CALI.CC.NO.	NEXT DUE DATE
1	DIGITAL PRESSURE GAUGE RANGE : 2500 bar / L.C : 0.01 bar	ESLT064	CC3220	2025/PSL/243-001	28.09.2026
2	STOP WATCH RANGE : 86400 Sec / L.C : 0.01 Sec	ESLT011	CC3687	MSIR/410725/536- 010	22.07.2026
3	WEIGHING MACHINE / RANGE : 0 -100 Kg, L.C : 0.001 Kg	ESLT-112	CC3687	MSIR/410526/160- 001	13.01.2027

ABOVE MASTER IS TRACEABLE TO NATIONAL & INTERNATIONAL STANDARDS

ENVIRONMENTAL CONDITIONS : AMBIENT.

UNCERTAINTY OF MEASUREMENT AT : $\pm 0.4\%$ FLOW AND 0.2% PRESSURE / TIME : ± 21 Sec
95% CONFIDENCE LEVELREFERENCE STANDARD : ASTM E4:2016, ISO7500-1:2018
(AT 95.45% CONFIDENCE LEVEL WITH COVERAGE FACTOR $k=2$)

REFERENCE PROCEDURE NUMBER : SLT/145 Date:15.09.2023

ADARSH M S
Incharge, O&A&NOC
17/5/26

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY :

CHECKED BY :

APPROVED BY :

Mobile 94863 33444, 94888 77777
e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641 009.



SHRETA LAB TECH

CALIBRATION CERTIFICATE

DATE : 13-May-2026

SHEET : 2 OF 3

CUSTOMER NAME : M/S.KSB MIL CONTROLS LIMITED, KERALA-680 741.

CERTIFICATE NUMBER : SLT/26/05/240/001

IDENTIFICATION NUMBER : E 144

FLOW METER CALIBRATION**A. FLOW METER OF 3 LPM**

THEORETICAL FLOW IN MACHINE	STOP WATCH STARTING READING (Sec)	STOP WATCH ENDING READING (Sec)	TOTAL TIME (Sec)	WEIGHED OUTPUT (Kg)	CALCULAT ED FLOW IN LITERS	ERROR OBSERVED LITERS	ERROR in %
0.000 Liters	0	0	0	0.00	0.00	0.000	0.0
1.000 Liters	0	20	20	0.98	0.98	-0.020	-2.0
2.000 Liters	0	40	40	1.98	1.98	-0.020	-1.0
3.000 Liters	0	60	60	3.01	3.01	0.010	0.3

B. FLOW METER OF 20 LPM

THEORETICAL FLOW IN MACHINE	STOP WATCH STARTING READING (Sec)	STOP WATCH ENDING READING (Sec)	TOTAL TIME (Sec)	WEIGHED OUTPUT (Kg)	CALCULAT ED FLOW IN LITERS	ERROR OBSERVED LITERS	ERROR in %
0.000 Liters	0	0	0	0.00	0.00	0.000	0.0
6.666 Liters	0	20	20	6.667	6.667	0.001	0.02
13.333 Liters	0	40	40	13.327	13.327	-0.006	-0.05
19.998 Liters	0	60	60	19.999	19.999	0.001	0.01

C. FLOW METER OF 60 LPM

THEORETICAL FLOW IN MACHINE	STOP WATCH STARTING READING (Sec)	STOP WATCH ENDING READING (Sec)	TOTAL TIME (Sec)	WEIGHED OUTPUT (Kg)	CALCULAT ED FLOW IN LITERS	ERROR OBSERVED LITERS	ERROR in %
0.000 Liters	0	0	0	0.00	0.00	0.000	0.0
20.000 Liters	0	20	20	20.02	20.02	0.020	0.100
40.000 Liters	0	40	40	40.03	40.03	0.030	0.075
60.000 Liters	0	60	60	60.03	60.03	0.030	0.050

ADARSH M S
Interfered DBO ADNOC
17/9/26

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY :

CHECKED BY :

APPROVED BY :

Mobile : 94863 33444, 94888 77777
e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641 009.



SHRETA LAB TECH

CALIBRATION CERTIFICATE

DATE: 13-May-2026

SHEET: 3 OF 3

CUSTOMER NAME : M/S.KSB MIL CONTROLS LIMITED, KERALA-680 741.
 CERTIFICATE NUMBER : SLT/26/05/240/001
 IDENTIFICATION NUMBER : E 144

ALL READINGS ARE IN bar

SL.NO	INDICATED	MASTER READING				OBSERVED
	READING	INCREASING	ERROR OBSERVED	DECREASING	ERROR OBSERVED	MAXIMUM ERROR OBSERVED
1	0.00	0.00	0.00	0.00	0.00	0.00
2	250.00	250.46	-0.46	250.45	-0.45	-0.46
3	500.00	500.65	-0.65	500.62	-0.62	-0.65
4	750.00	750.75	-0.75	750.72	-0.72	-0.75
5	900.00	900.85	-0.85	900.82	-0.82	-0.85
6	1000.00	1000.89	-0.89	1000.85	-0.85	-0.89
7	1200.00	1200.96	-0.96	1200.96	-0.96	-0.96

ALL VALUES ARE IN MINUTE/SECONDS

RANGE	UUC TIME READINGS		MASTER		ERROR
	IN	(Min) (Sec)	Mints	Seconds	
0-16 Min		0.00		0.00	-00.00
		1.00		1.01	-00.01
		2.00		2.01	-00.01
		3.00		3.01	-00.01
		4.00		4.02	-00.02
		5.00		5.03	-00.03
		6.00		6.03	-00.03
		7.00		7.03	-00.03
		8.00		8.04	-00.04
		9.00		9.04	-00.04
		10.00		10.04	-00.04
		15.00		15.05	-00.05
		16.00		16.05	-00.05

STATUS OF CALIBRATION : TO BE DECIDED BY CUSTOMER

NOTE 1 : NEXT CALIBRATION DUE AS PER CUSTOMER'S PLAN AND REQUIREMENT.

NOTE 2 : THE REPORTED RESULTS ARE VALID ONLY FOR THE CONDITION OF THE RECEIVED LEAK TESTING MACHINE AT THE TIME OF AND UNDER THE STATED CONDITION OF CALIBRATION.

NOTE 3 : NO CORRECTION OR ADJUSTMENT IS MADE BEFORE OR AFTER CALIBRATION AS FOUND AS DATA IS FOUND IN THE REPORT.

-- END OF THE REPORT --

Verified and Found OK

 RANITH K.K.
 For KSB MIL Controls Ltd.

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY :

CHECKED BY :

APPROVED BY :

 Mobile 94863 33444, 94888 77777
 e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641 009.



SHRETA LAB TECH

CALIBRATION CERTIFICATE

DATE : 17-Apr-2026

SHEET :
1 OF 2

CERTIFICATE NUMBER : SLT/26/04/480/021

CUSTOMER NAME : M/S. KSB MIL CONTROLS LIMITED ,KERALA.

MAKE/IDENTIFICATION NUMBER : MIL-QC-1319

CALIBRATION PROCEDURE REF NO : SLT/77

DATE OF CALIBRATION : Friday, April 17, 2026

DATE OF ISSUE : Friday, April 17, 2026

CALIBRATION NEXT DUE DATE : Friday, October 16, 2026

DESCRIPTION OF INSTRUMENT : 5 Meter , LC = 1 mm MEASURING TAPE

DETAILS OF MASTERS USED FOR CALIBRATION

SL.NO	MASTER DESCRIPTION	MASTER ID.NO.	MASTER CALIBRATED AT	CALI.CC.NO.	CALI DATE	NEXT DUE DATE
1	VIDEO MEASURING MACHINE	ESLT003	SLT	SLT/25/06/606/003	21.06.2025	20.06.2026

ALL ABOVE MENTIONED MASTERS ARE TRACEABLE TO NATIONAL STANDARDS

ENVIRONMENTAL CONDITIONS : TEMPERATURE 20° ±1°C, HUMIDITY 50 ±10% RH.

UNCERTAINTY OF MEASUREMENT : 0.5 mm

(AT 95% CONFIDENCE LEVEL WITH
COVERAGE FACTOR k=2)

REFERENCE STANDARD : IS 1269 PART-II

⑫

A DARTH MS
Interch OB O ADNOC
17/4/26



17/04/2026

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY :

CHECKED BY :

APPROVED BY :

Mobile : 94863 33444, 94888 77777
e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641 009.

17-Apr-2026



SHRETA LAB TECH

CALIBRATION CERTIFICATE

DATE :

2 OF 2

SHEET :

CUSTOMER NAME : M/S. KSB MIL CONTROLS LIMITED , KERALA.

CERTIFICATE NUMBER : SLT/26/04/480/021

MAKE/IDENTIFICATION NUMBER : MIL-QC-1319

CALIBRATION OF MEASURING TAPE

(ALL VALUES ARE IN cm)

SL.NO.	ACTUAL SIZE	MEASURED SIZE	ERROR
1	10.00	10.00	0.00
2	20.00	20.01	0.01
3	30.00	30.01	0.01
4	40.00	40.01	0.01
5	50.00	50.01	0.01
6	60.00	60.01	0.01
7	70.00	70.01	0.01
8	80.00	80.01	0.01
9	90.00	90.01	0.01
10	100.00	100.01	0.01
11	110.00	110.01	0.01
12	120.00	120.01	0.01
13	130.00	130.01	0.01
14	140.00	140.01	0.01
15	150.00	150.02	0.02
16	160.00	160.01	0.01
17	170.00	170.01	0.01
18	180.00	180.02	0.02
19	190.00	190.03	0.03
20	200.00	200.01	0.01
21	210.00	210.02	0.02
22	220.00	220.02	0.02
23	230.00	230.03	0.03
24	240.00	240.02	0.02
25	250.00	250.02	0.02

(ALL VALUES ARE IN cm)

SL.NO.	ACTUAL SIZE	MEASURED SIZE	ERROR
26	260.00	260.01	0.01
27	270.00	270.01	0.01
28	280.00	280.01	0.01
29	290.00	290.02	0.02
30	300.00	300.01	0.01
31	310.00	310.02	0.02
32	320.00	320.02	0.02
33	330.00	330.01	0.01
34	340.00	340.02	0.02
35	350.00	350.00	0.00
36	360.00	360.03	0.03
37	370.00	370.01	0.01
38	380.00	380.02	0.02
39	390.00	390.01	0.01
40	400.00	400.01	0.01
41	410.00	410.03	0.03
42	420.00	420.02	0.02
43	430.00	430.02	0.02
44	440.00	440.02	0.02
45	450.00	450.03	0.03
46	460.00	460.00	0.00
47	470.00	470.01	0.01
48	480.00	480.03	0.03
49	490.00	490.02	0.02
50	500.00	500.03	0.03

ALLOWABLE ERROR ± 0.05 cm

STATUS OF CALIBRATION : ACCEPTED

NOTE 1 : NEXT CALIBRATION DUE AS PER CUSTOMER'S PLAN AND REQUIREMENT.

NOTE 2 : THE REPORTED RESULTS ARE VALID ONLY FOR THE CONDITION OF THE RECEIVED MEASURING TAPE AT THE TIME OF AND UNDER THE STATED CONDITION OF CALIBRATION.

NOTE 3 : NO CORRECTION OR ADJUSTMENT IS MADE BEFORE OR AFTER CALIBRATION AND FOUND A DATA IS FOUND IN THE REPORT.

RANITH K.K.

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced or used in any form without the written approval of SLT

CALIBRATED BY :

CHECKED BY :

APPROVED BY :

Mobile : 94863 33444, 94888 77777

e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641 009.

2026L566

Format No. : CI05

DATE : 28-Feb-2026	 SHRETA LAB TECH CALIBRATION CERTIFICATE	 CC - 2403
SHEET : 1 OF 2		

CERTIFICATE NUMBER : SLT/26/02/415/008
 CUSTOMER NAME : M/S.KSE MIL CONTROLS LIMITED, KERALA
 CUSTOMER REFERENCE NUMBER : ONSITE
 DESCRIPTION OF THE ITEM : DIGITAL STOP WATCH
 CALIBRATION PROCEDURE NUMBER : SHRETA-WI-BT-06
 CALIBRATION FORMAT NO. : SHRETA/IT-25-05
 CONDITION OF RECEIVED ITEM : GOOD
 CALIBRATED AT : IN LAB
 MAKE/IDENTIFICATION NUMBER : RACER / MQC599 / -
 DATE OF CALIBRATION : Saturday, February 28, 2026
 DATE OF ISSUE : Saturday, February 28, 2026
 CALIBRATION NEXT DUE DATE : Saturday, February 27, 2027
 PARAMETER CALIBRATED : HOUR/MINUTE/SECONDS.
 RANGE : 0 TO 24 Hrs
 UNCERTAINTY OF MEASUREMENT : 0.61s TO 7.7s
 REFERENCE STANDARD : IS 10996 : 1984
 TEMPERATURE DURING CALIBRATION : 25°C ±4°C
 HUMIDITY DURING CALIBRATION : 30 TO 75% RH

MASTER DETAILS:

DESCRIPTION : DIGITAL TIME CALIBRATOR
 MAKE : GLAXO INFRASTRUTURE
 IDENTIFICATION NUMBER : ESLT-011
 TRACEABILITY : MSIR
 CERTIFICATE NO. : MSIR/410725/536-010
 CALIBRATION DATE : 23.07.2025
 NEXT CALIBRATION DUE : 22.07.2026
 ALL ABOVE MENTIONED MASTERS ARE TRACEABLE TO NATIONAL STANDARDS

(R)
 ADARSH MS
 Intertech OBO ADJOC
 Adarsh
 19/12/22



Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY :  M.VINCENT Technical Manager	 	AUTHORISED SIGNATORY  R. GLEETUS Quality Manager
Mobile : 94863 33444, 94888 77777 e-mail : shrlab@rediffmail.com	SHRETA LAB TECH No.44, Bharathi Road, Ramnagar, Coimbatore - 641009.	

DATE : 25-Feb-2026

SHEET : 2 OF 2



SHRETA LAB TECH

CALIBRATION CERTIFICATE



CC - 2403

CUSTOMER NAME : M/S.KSB MIL CONTROLS LIMITED, KERALA
 CERTIFICATE NUMBER : SLI/25/02/415/008
 MAKE/IDENTIFICATION NUMBER : RACER / MQC599 / -

ALL VALUES ARE IN Hrs Min Sec

Sl.No	UQC READINGS IN Hrs./Min./Sec.	MASTER READING in Hrs./Min./Sec.	ERROR OBSERVED IN (Hrs./Min./Sec.)
1	0.00'.00"	0.00'.00"	0.00'.00"
2	0.10'.00"	0.10'.00"	0.00'.00"
3	0.20'.00"	0.20'.00"	0.00'.00"
4	0.30'.00"	0.30'.00"	0.00'.00"
5	0.40'.00"	0.40'.00"	0.00'.00"
6	0.50'.00"	0.50'.00"	0.00'.00"
7	1.00'.00"	1.00'.00"	0.00'.00"
8	2.00'.00"	2.00'.00"	0.00'.00"
9	4.00'.00"	4.00'.01"	0.00'.01"
10	8.00'.00"	8.00'.01"	0.00'.01"
11	12.00'.00"	12.00'.01"	0.00'.01"
12	16.00'.00"	16.00'.02"	0.00'.02"
13	20.00'.00"	20.00'.02"	0.00'.02"
14	24.00'.00"	24.00'.02"	0.00'.02"

STATUS OF CALIBRATION : ACCEPTED

PERMISSIBLE ERROR AS PER MANUFACTURER STANDARD 5 Sec

NOTE 1 : NEXT CALIBRATION DUE AS PER CUSTOMER'S PLAN AND REQUIREMENT.

NOTE 2 : THE REPORTED RESULTS ARE VALID ONLY FOR THE CONDITION OF THE RECEIVED DIGITAL TIMER AT THE TIME OF AND UNDER THE STATED CONDITION OF CALIBRATION.

NOTE 3 : NO CORRECTION OR ADJUSTMENT IS MADE BEFORE OR AFTER CALIBRATION AS FOUND AS DATA IS FOUND IN THE REPORT.

-End-

Verified and Found OK

 RANJITH K.K
 For KSB MIL Controls Ltd.


Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY :

 M.VINCENT
 Technical Manager


AUTHORISED SIGNATORY

 R. G. SETHUS
 Quality Manager

 Mobile : 94863 33444, 94888 77777
 e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641009.