



SHRETA LAB TECH

CALIBRATION CERTIFICATE

Format No. : CI05

17-Apr-2026

DATE :

2 OF 2

SHEET :

CUSTOMER NAME

: M/S. KSB MIL CONTROLS LIMITED , KERALA.

CERTIFICATE NUMBER

: SLT/26/04/480/021

MAKE/IDENTIFICATION NUMBER

: MIL-QC-1319

CALIBRATION OF MEASURING TAPE

(ALL VALUES ARE IN cm)

SL.NO.	ACTUAL SIZE	MEASURED SIZE	ERROR
1	10.00	10.00	0.00
2	20.00	20.01	0.01
3	30.00	30.01	0.01
4	40.00	40.01	0.01
5	50.00	50.01	0.01
6	60.00	60.01	0.01
7	70.00	70.01	0.01
8	80.00	80.01	0.01
9	90.00	90.01	0.01
10	100.00	100.01	0.01
11	110.00	110.01	0.01
12	120.00	120.01	0.01
13	130.00	130.01	0.01
14	140.00	140.01	0.01
15	150.00	150.02	0.02
16	160.00	160.01	0.01
17	170.00	170.01	0.01
18	180.00	180.02	0.02
19	190.00	190.03	0.03
20	200.00	200.01	0.01
21	210.00	210.02	0.02
22	220.00	220.02	0.02
23	230.00	230.03	0.03
24	240.00	240.02	0.02
25	250.00	250.02	0.02

ALLOWABLE ERROR ± 0.05 cm

STATUS OF CALIBRATION : ACCEPTED

NOTE 1 : NEXT CALIBRATION DUE AS PER CUSTOMER'S PLAN AND REQUIREMENT.

NOTE 2 : THE REPORTED RESULTS ARE VALID ONLY FOR THE CONDITION OF THE RECEIVED

MEASURING TAPE AT THE TIME OF AND UNDER THE STATED CONDITION OF CALIBRATION.

NOTE 3 : NO CORRECTION OR ADJUSTMENT IS MADE BEFORE OR AFTER CALIBRATION AND FOUND A

DATA IS FOUND IN THE REPORT.

RANITH K.K

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced without the written approval of SLT

CALIBRATED BY :

CHECKED BY :

APPROVED BY :

Mobile : 94863 33444, 94888 77777


e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641 009.

DATE: 28-Feb-2026

SHEET: 1 OF 2



SHRETA LAB TECH

CALIBRATION CERTIFICATE



ULR-CC240326000008978E

CC - 2403

CERTIFICATE NUMBER : SLT/26/02/415/008
 CUSTOMER NAME : M/S.KSB MIL CONTROLS LIMITED, KERALA
 CUSTOMER REFERENCE NUMBER : ONSITE
 DESCRIPTION OF THE ITEM : DIGITAL STOP WATCH
 CALIBRATION PROCEDURE NUMBER : SHRETA-WI-ET-06
 CALIBRATION FORMAT NO. : SHRETA/ET-25-05
 CONDITION OF RECEIVED ITEM : GOOD
 CALIBRATED AT : IN LAB
 MAKE/IDENTIFICATION NUMBER : RACER / MQC599 / -
 DATE OF CALIBRATION : Saturday, February 28, 2026
 DATE OF ISSUE : Saturday, February 28, 2026
 CALIBRATION NEXT DUE DATE : Saturday, February 27, 2027
 PARAMETER CALIBRATED : HOUR/MINUTE/SECONDS.
 RANGE : 0 TO 24 Hrs
 UNCERTAINTY OF MEASUREMENT : 0.61s TO 7.7s
 REFERENCE STANDARD : IS 10396 : 1984
 TEMPERATURE DURING CALIBRATION : 25°C $\pm$ 4°C
 HUMIDITY DURING CALIBRATION : 30 TO 75% RH
MASTER DETAILS:
 DESCRIPTION : DIGITAL TIME CALIBRATOR
 MAKE : GLAXO INFRASTRUCTURE
 IDENTIFICATION NUMBER : ESLT-011
 TRACEABILITY : MSIR
 CERTIFICATE NO. : MSIR/410725/536-010
 CALIBRATION DATE : 23.07.2025
 NEXT CALIBRATION DUE : 22.07.2026
 ALL ABOVE MENTIONED MASTERS ARE TRACEABLE TO NATIONAL STANDARDS

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY:

M.VINCENT

Technical Manager



AUTHORISED SIGNATORY

R. GLEATUS

Quality Manager

Mobile : 94863 33444, 94888 77777
 e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641009.

20261883

DATE: 17-Apr-2026

SLT

SHRETA LABTECH

CALIBRATION CERTIFICATE

SHEET:

1 OF 2

CERTIFICATE NUMBER

: SLT/26/04/480/021

CUSTOMER NAME

: M/S. KSB MIL CONTROLS LIMITED , KERALA.

MAKE/IDENTIFICATION NUMBER

: MIL-QC-1319

CALIBRATION PROCEDURE REF NO

: SLT/77

DATE OF CALIBRATION

: Friday, April 17, 2026

DATE OF ISSUE

: Friday, April 17, 2026

CALIBRATION NEXT DUE DATE

: Friday, October 16, 2026

DESCRIPTION OF INSTRUMENT

: 5 Meter , LC = 1 mm MEASURING TAPE

DETAILS OF MASTERS USED FOR CALIBRATION

SL.NO	MASTER DESCRIPTION	MASTER ID.NO.	MASTER CALIBRATED AT	CALI CC.NO.	CALI DATE	NEXT DUE DATE
1	VIDEO MEASURING MACHINE	ESLT003	SLT	SLT/25/06/606/003	21.06.2025	20.06.2026

ALL ABOVE MENTIONED MASTERS ARE TRACEABLE TO NATIONAL STANDARDS

ENVIRONMENTAL CONDITIONS

: TEMPERATURE 20° ±1°C, HUMIDITY 50 ±10% RH.

UNCERTAINTY OF MEASUREMENT

: 0.5 mm

(AT 95% CONFIDENCE LEVEL WITH

COVERAGE FACTOR k=2)

REFERENCE STANDARD

: IS 1269 PART-II

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY:

CHECKED BY:

APPROVED BY:


Mobile
e-mail94863 33444, 94888 77777
: shrlab@rediffmail.com

SHRETA LABTECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641 009.

DATE: 28-Feb-2026

SHEET: 2 OF 2



SHRETA LAB TECH

CALIBRATION CERTIFICATE



CC - 2403

CUSTOMER NAME : M/S.KSB MIL CONTROLS LIMITED, KERALA
 CERTIFICATE NUMBER : SLT/26/02/415/008
 MAKE/IDENTIFICATION NUMBER : RACER / MQC599 / -

ALL VALUES ARE IN Hrs Min Sec

Sl.No	UUC READINGS IN Hrs./Min./Sec.	MASTER READING in Hrs./Min./Sec.	ERROR OBSERVED IN ($\pm$ Hrs./Min./Sec.)
1	0.00'.00"	0.00'.00"	0.00'.00"
2	0.10'.00"	0.10'.00"	0.00'.00"
3	0.20'.00"	0.20'.00"	0.00'.00"
4	0.30'.00"	0.30'.00"	0.00'.00"
5	0.40'.00"	0.40'.00"	0.00'.00"
6	0.50'.00"	0.50'.00"	0.00'.00"
7	1.00'.00"	1.00'.00"	0.00'.00"
8	2.00'.00"	2.00'.00"	0.00'.00"
9	4.00'.00"	4.00'.01"	0.00'.01"
10	8.00'.00"	8.00'.01"	0.00'.01"
11	12.00'.00"	12.00'.01"	0.00'.01"
12	16.00'.00"	16.00'.02"	0.00'.02"
13	20.00'.00"	20.00'.02"	0.00'.02"
14	24.00'.00"	24.00'.02"	0.00'.02"

STATUS OF CALIBRATION : ACCEPTED

PERMISSIBLE ERROR AS PER MANUFACTURER STANDARD 5 Sec

NOTE 1 : NEXT CALIBRATION DUE AS PER CUSTOMER'S PLAN AND REQUIREMENT.

NOTE 2 : THE REPORTED RESULTS ARE VALID ONLY FOR THE CONDITION OF THE RECEIVED DIGITAL TIMER AT THE TIME OF AND UNDER THE STATED CONDITION OF CALIBRATION.

NOTE 3 : NO CORRECTION OR ADJUSTMENT IS MADE BEFORE OR AFTER CALIBRATION

AS FOUND AS DATA IS FOUND IN THE REPORT.

Verified and Found OK

-End-

RANJITH.K.K
 For KSB MIL Controls Ltd.

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY:

M. VINCENT
 Technical Manager



AUTHORISED SIGNATORY

R. G. SETHUS
 Quality Manager

Mobile : 94863 33444, 94868 77777
 e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641009.