

2025L 2389

Format No. : CI05

DATE:

15-Sep-2025



SHRETA LAB TECH

CALIBRATION CERTIFICATE



CC-2403

SHEET:

1 OF 2

ULR-CC240325000076001F

CERTIFICATE NUMBER : SLT/25/09/742/011
 CUSTOMER NAME : M/S.KSB MIL CONTROLS LIMITED, KERALA
 CUSTOMER REFERENCE NUMBER : ONSITE
 DESCRIPTION OF THE ITEM : DIGITAL STOP WATCH
 CALIBRATION PROCEDURE NUMBER : SHRETA-WI-ET-06
 CALIBRATION FORMAT NO. : SHRETA/FT-25-05
 CONDITION OF RECEIVED ITEM : GOOD
 CALIBRATED AT : IN LAB
 MAKE/IDENTIFICATION NUMBER : RACER / M509
 DATE OF CALIBRATION : Monday, September 15, 2025
 DATE OF ISSUE : Monday, September 15, 2025
 CALIBRATION NEXT DUE DATE : Monday, September 14, 2026
 PARAMETER CALIBRATED : HOUR/MINUTE/SECONDS.
 RANGE : 0 TO 24 Hrs
 UNCERTAINTY OF MEASUREMENT : $\pm 0.61s$ TO $7.7s$
 (EXPANDED UNCERTAINTY IS REPORTED AT
 95% CONFIDENCE LEVEL WITH COVERAGE FACTOR $k=2$)
 REFERENCE STANDARD : IS 10996 : 1984
 TEMPERATURE DURING CALIBRATION : $25^{\circ}\text{C} \pm 4^{\circ}\text{C}$
 HUMIDITY DURING CALIBRATION : 30 TO 75% RH
 DESCRIPTION : DIGITAL TIME CALIBRATOR
 MAKE : GLAXO INFRASTRUTURE
 IDENTIFICATION NUMBER : ESLT-011
 TRACEABILITY : MSIR
 CERTIFICATE NO. : MSIR/410725/536-010
 CALIBRATION DATE : 23.07.2025
 NEXT CALIBRATION DUE : 22.07.2026
 ALL ABOVE MENTIONED MASTERS ARE TRACEABLE TO NATIONAL STANDARDS

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY:

M. VINCENT
Technical Manager



AUTHORISED SIGNATORY

R. GLEETUS
Quality Manager

Mobile : 94863 33444, 94888 77777

e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641 009.

DATE : 15-Sep-2025

SHEET : 2 OF 2



SHRETA LAB TECH

CALIBRATION CERTIFICATE



CUSTOMER NAME : M/S.KSB MIL CONTROLS LIMITED, KERALA
 CERTIFICATE NUMBER : SLT/25/09/742/011
 MAKE/IDENTIFICATION NUMBER : RACER / M509

ALL VALUES ARE IN Hrs Min Sec

Sl.No	UUC READINGS IN Hrs./Min./Sec.	MASTER READING in Hrs./Min./Sec.	ERROR OBSERVED IN (±Hrs./Min./Sec.)
1	0.00'.00"	0.00'.00"	0.00'.00"
2	0.10'.00"	0.10'.00"	0.00'.00"
3	0.20'.00"	0.20'.00"	0.00'.00"
4	0.30'.00"	0.30'.00"	0.00'.00"
5	0.40'.00"	0.40'.00"	0.00'.00"
6	0.50'.00"	0.50'.00"	0.00'.00"
7	1.00'.00"	1.00'.00"	0.00'.00"
8	2.00'.00"	2.00'.00"	0.00'.00"
9	4.00'.00"	4.00'.01"	0.00'.01"
10	8.00'.00"	8.00'.01"	0.00'.01"
11	12.00'.00"	12.00'.01"	0.00'.01"
12	16.00'.00"	16.00'.02"	0.00'.02"
13	20.00'.00"	20.00'.02"	0.00'.02"
14	24.00'.00"	24.00'.02"	0.00'.02"

STATUS OF CALIBRATION : ACCEPTED

PERMISSIBLE ERROR AS PER MANUFACTURER STANDARD 5 Sec

NOTE 1 : NEXT CALIBRATION DUE AS PER CUSTOMER'S PLAN AND REQUIREMENT.

NOTE 2 : THE REPORTED RESULTS ARE VALID ONLY FOR THE CONDITION OF THE RECEIVED DIGITAL TIMER AT THE TIME OF AND UNDER THE STATED CONDITION OF CALIBRATION.

NOTE 3 : NO CORRECTION OR ADJUSTMENT IS MADE BEFORE OR AFTER CALIBRATION AS FOUND AS DATA IS FOUND IN THE REPORT.

-End-

Verified and Found OK

RANJITH. K.K
For KSB MIL Controls Ltd.

Results presented in this report relates only to the item mentioned & calibrated at SLT & shall not be reproduced except in full without the written approval of SLT

CALIBRATED BY :

M.VINCENT
Technical Manager



AUTHORISED SIGNATORY

R. GLEETUS
Quality Manager

Mobile : 94863 33444, 94888 77777

e-mail : shrlab@rediffmail.com

SHRETA LAB TECH

No.44, Bharathi Road, Ramnagar, Coimbatore - 641 009.