EBS-010 communcation Protocol

*** Com Port Setting**

Baudrate	4800
Data Bits	8
Stop Bits	1
Parity	None Parity

*** Communication Description**

1. All commands are capitalized.
2. The format of the command is as follows.

ex) HEX code

↗

$START 0D 0A

↘

ASCⅡcode

Breathlayzer EBS-010 to COMPUTER

Definition	Command		Note
	ASCⅡ	HEX	
Ready	$END	0x0D 0x0A	send command for 2 sec interval.
Warming up	$WAIT	0x0D 0x0A	send command for 1 sec interval.
Calibration period	$CALIBRATION	0x0D 0x0A	send command once at beginning of warm up. 1,000 tests is calibration period.
Standby	$STANBY	0x0D 0x0A	send command for 1 sec interval.
Sampling	$BREATH	0x0D 0x0A	correct breath sampling.
TRIGGER	$TRIGGER	0x0D 0x0A	breath sampling activated.
Result	$RESULT,0.000-OK	0x0D 0x0A	"OK" or "HIGH" command. based on pre-set limit(eg. If you set preset limit 0.30g/L and test result shows over 0.30g/L, command is "HIGH", if result is lower than 0.3g/L, command is "OK")
Recall	$U/B,L/000,H/000,T/0000	0x0D 0x0A	U: measurenent unit (B:%BAC,G:g/L,M:Mg/L), L: LOW result, H: HIGH result, T:testing number (0000~9,999)
FLOW ERROR	$FLOW,ERR	0x0D 0x0A	Incorrect breath sampling

COMPUTER to Breathalyzer EBS-010

Definition	Command		Note
	ASCⅡ	HEX	
START	$START	0x0D 0x0A	Operating command. (Works only on state "$END")
Setting Recall	$RECALL	0x0D 0x0A	Check the set value. (Works only on state "$END")
Device OFF	$RESET	0x0D 0x0A	Make breathalyzer turn OFF (Works only on state "$STANBY")