Field Trial: Evaluation and Acceptance Form

On-site testing of G-one at a pharmacy or medical retail establishment, by its staff, its pharmacist and its customers

This form records an on-site trial of **G-one**, a wrist strap and Android app that watches movement and vital signs, explains them in plain language, and texts an emergency contact after a fall. Fill in **what was actually done and seen**. Where a capability was not tried, tick *Not tested* or leave the row blank. The **developer's reference** column on page 2 gives the G-one team's own laboratory results on public datasets, for comparison only; it is not a result of this trial.

A ESTABLISHMENT DETAILS

Name of establishment _____

Type ☐ Pharmacy / chemist ☐ Medical store ☐ Surgical & medical equipment retailer ☐ Other: _____

Drug licence / registration no. _____ GSTIN (optional) _____

Address _____

City / town _____ District _____ State _____ PIN _____

Contact person _____ Designation _____

Phone _____ Email _____

Pharmacist-in-charge _____ Pharmacy Council registration no. _____

B TRIAL DETAILS

Trial date: from _____ to _____ Start time _____ End time _____

G-one app version _____ Phone model(s) used _____

Strap ID / serial no. _____ Strap firmware version _____

G-one evaluator(s) present _____

Where on the premises _____ Total hours of testing _____

C WHO TOOK PART

Group	Number of people	How they took part (e.g. wore the strap, ran the app, watched a demonstration)
Establishment staff		
Pharmacist(s)		
Customers / walk-in volunteers		
Others (state)		

☐ Every participant was told what the trial is and agreed before taking part.

☐ Customers' names and health readings were not kept after the trial, or kept only with their written consent.

☐ Emergencies were simulated (staged falls, demo mode); no real emergency contact was messaged without that person's agreement.

☐ No participant was given medical advice on the basis of G-one readings.

D CAPABILITIES TESTED

Reference figures: G-one lab results, not this trial

For each capability: tick **Tested** or **Not tested**, write how many times it was tried and how many times it behaved as described, and note anything seen. A staged fall means a safe, deliberate jolt or a replayed recording, never a real fall.

#	Capability	Developer's reference (lab, not this trial)	Tested	Not tested	Times tried	Times it worked	What was seen / remarks
1	**Live readings from the strap over Bluetooth (movement, skin temperature)**	Designed for 2 readings a second	☐	☐			
2	**Fall alert from the strap (impact above 5 g)**	Rule on lab recordings: 50.7% of falls caught; 89.6% of daily activities with no alarm	☐	☐			
3	**Fall model, replayed in the app's Model Lab**	96.2% of falls caught, 95.9% of daily activities with no alarm (people held out); 64.8% / 98.1% on a dataset it never saw	☐	☐			
4	**"I'm OK" cancel window before the SOS text**	20 s by default (10, 20 or 30 s in Settings)	☐	☐			
5	**SOS text message to the emergency contact**	Sent after the window; never sent for a model-only or city-weather alert	☐	☐			
6	**Heart rate in context model**	2.24 false alerts per 24 h vs 51.58 for a fixed rule (PAMAP2). Strap pulse sensor not yet active: shown with recorded data	☐	☐			
7	**Sleep or awake model**	91.9% of 30-s periods right; Cohen's kappa 0.46 vs 0.37 for a movement threshold	☐	☐			
8	**On-phone AI assistant (works offline)**	Rewords alerts in plain language; does not decide them	☐	☐			
9	**Official disaster alerts for the district (NDMA SACHET)**	District matched on the phone; no location is sent	☐	☐			
10	**City weather in the heat rules, and India AQI**	City-based warnings capped at moderate; never an SOS	☐	☐			
11	**Health report PDF and NFC emergency card**	-	☐	☐			
12	**Model speed on this phone (Model Lab)**	9-27 microseconds per run on the CPU of a Samsung SM-S721B	☐	☐			

E RATINGS BY GROUP

1 = poor, 3 = acceptable, 5 = excellent; leave blank if not asked

Statement	Staff (average)	Pharmacist	Customers (average)
Easy to put on and set up			
Readings and alerts were easy to understand			
Fall alert and the "I'm OK" cancel worked as shown			
Assistant explanations were clear and in plain language			
Comfortable to wear for the time tested			
Would suggest it for customers who live alone or are elderly			

F OBSERVATIONS

Problems, failures or false alarms seen (what happened, when, how often)

Comments and suggestions from staff, the pharmacist and customers

G OUTCOME OF THE TRIAL

- ☐ G-one performed as described above and is acceptable for continued use or trial at this establishment.
- ☐ Acceptable, with the conditions or changes written below.
- ☐ Not acceptable at this stage (reasons below).

Conditions, changes needed or reasons

H DECLARATION AND SIGNATURES

We confirm that the trial described in this form took place at our premises on the dates stated, and that the results, ratings and remarks recorded here are what we and the participants observed. We understand that G-one is a research prototype: it is **not a certified medical device**, it does not diagnose or treat any condition, and this form is a record of a user trial, **not** a medical, regulatory or CDSCO approval.

For the establishment (authorised signatory)

Name

Designation

Date

Place

Signature

ESTABLISHMENT STAMP / SEAL

Stamp across the box edge

Pharmacist-in-charge

Name

Registration no.

Date

Signature

G-one evaluator

Name

Role

Date

Signature