

Test Report Title


SPECIMEN / PATIENT INFORMATION

Patient Name:	Patient ID:
Study Date: 20180614	Modality: OT
Institution:	Description:

EXAMINATION PARAMETERS

Manufacturer:	Model:
Station:	Series:
Image Size: 1000 x 1000	Pixel Spacing: 1.000 / 1.000 mm
Bits Allocated: 16	Bits Stored: 16
Window Center: 39104	Window Width: 11117

RADIOGRAPHIC IMAGE



IQI WIRE PAIR ANALYSIS

Wire Pair	Resolution (%)	Status
D1	96.6	Resolved
D2	96.3	Resolved
D3	94.6	Resolved
D4	91.3	Resolved
D5	84.1	Resolved
D6	58.4	Resolved
D7	44.4	Resolved
D8	42.8	Resolved
D9	18.1	Unresolved
D10	96.7	Resolved

Duplex Wire Line Profile

