

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\Sam-8chcoil\localizer_50V_newcoil

TA: 0:27 PAT: Off Voxel size: 1.2x1.1x3.0 mm Rel. SNR: 1.00 SIEMENS: gre

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	5
Dist. factor	20 %
Position	L0.0 A8.8 F17.5
Orientation	Sagittal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 2	
Slices	5
Dist. factor	20 %
Position	L0.0 P68.8 H0.7
Orientation	Coronal
Phase enc. dir.	R >> L
Rotation	0.00 deg
Slice group 3	
Slices	5
Dist. factor	20 %
Position	L0.0 P68.8 H0.7
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	280 mm
FoV phase	100.0 %
Slice thickness	3.0 mm
TR	10.0 ms
TE	3.00 ms
Averages	1
Concatenations	15
Filter	None
Coil elements	B1-8

Contrast

TD	0 ms
MTC	Off
Magn. preparation	None
Flip angle	10 deg
Fat suppr.	None
Water suppr.	None
SWI	Off
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	256
-----------------	-----

Phase resolution	90 %
Phase partial Fourier	6/8
Interpolation	On
PAT mode	None
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Sequential
Series	Interleaved
Saturation mode	Standard
Special sat.	None
Table position	H
Table position	0 mm
Inline Composing	Off
Tim CT mode	Off

System

B1	On
B2	On
B3	On
B4	On
B5	On
B6	On
B7	On
B8	On
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	On
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Off
Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	50.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	350 mm
A >> P	263 mm
F >> H	350 mm

Physio

1st Signal/Mode	None
Segments	1
Tagging	None
Dark blood	Off
Resp. control	Off

Inline

Subtract	Off
Liver registration	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On
<hr/>	
Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
<hr/>	
MapIt	None
Contrasts	1

Sequence

Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Allowed
Bandwidth	320 Hz/Px
Flow comp.	No
<hr/>	
RF pulse type	Normal
Gradient mode	Whisper
Excitation	Slice-sel.
RF spoiling	On

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\Sam-8chcoil\AV_ep2d_bold_sd1ipat2mb2_pt5mm_visLoc_8

TA: 5:30 PAT: 2 Voxel size: 0.5x0.5x0.5 mm Rel. SNR: 1.00 USER: AV_ep2d_bold_sd_20140727

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	60
Dist. factor	0 %
Position	L8.8 P72.2 F11.7
Orientation	Coronal
Phase enc. dir.	F >> H
Rotation	90.00 deg
Phase oversampling	0 %
FoV read	90 mm
FoV phase	88.9 %
Slice thickness	0.50 mm
TR	3000 ms
TE	26.0 ms
Multi-band accel. factor	2
Filter	None
Coil elements	B1-8

Contrast

MTC	Off
Magn. preparation	None
Flip angle	80 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	102
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	180
Phase resolution	100 %
Phase partial Fourier	5/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	36
Reference scan mode	Segmented
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Special sat. None

Table position	H
Table position	0 mm
Inline Composing	Off

System

B1	On
B2	On
B3	On
B4	On
B5	On
B6	On
B7	On
B8	On
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	On
Assume Silicone	Off
! Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L8.8 P72.2 F11.7
Orientation	Coronal
Rotation	90.00 deg
R >> L	90 mm
F >> H	80 mm
A >> P	30 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
Bandwidth	816 Hz/Px
Flow comp.	No
Free echo spacing	Off
Echo spacing	1.43 ms
SIR accel. factor	1
EPI factor	160
Gradient mode	Normal
RF spoiling	Off
Excite pulse duration	5820 us
Slice multiplier	1
Multi-band PE shift	0 1/FoV
zBlip scheme	0
MB kernel size	0
MB knockout band	0
No. of interleaved TEs	0
RF pulse shape	1
EPI noise scans	0
EPI full reference scan	0
Single-band images	On
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Readout slice trace	Off
Disable ramp sampling	Off
PF omits higher k-space	Off
Online multi-band recon.	Online
FFT scale factor	0.02
Send B1 shim trigger	Never
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\Sam-8chcoil\AV_ep2d_bold_sd1ipat2mb2_pt5mm_visLoc_8

TA: 5:30 PAT: 2 Voxel size: 0.5x0.5x0.5 mm Rel. SNR: 1.00 USER: AV_ep2d_bold_sd_20140727

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	60
Dist. factor	0 %
Position	L8.8 P72.2 F11.7
Orientation	Coronal
Phase enc. dir.	F >> H
Rotation	90.00 deg
Phase oversampling	0 %
FoV read	90 mm
FoV phase	88.9 %
Slice thickness	0.50 mm
TR	3000 ms
TE	26.0 ms
Multi-band accel. factor	2
Filter	None
Coil elements	B1-8

Contrast

MTC	Off
Magn. preparation	None
Flip angle	80 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	102
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	180
Phase resolution	100 %
Phase partial Fourier	5/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	36
Reference scan mode	Segmented
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Special sat. None

Table position	H
Table position	0 mm
Inline Composing	Off

System

B1	On
B2	On
B3	On
B4	On
B5	On
B6	On
B7	On
B8	On
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	On
Assume Silicone	Off
! Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L8.8 P72.2 F11.7
Orientation	Coronal
Rotation	90.00 deg
R >> L	90 mm
F >> H	80 mm
A >> P	30 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
Bandwidth	816 Hz/Px
Flow comp.	No
Free echo spacing	Off
Echo spacing	1.43 ms
SIR accel. factor	1
EPI factor	160
Gradient mode	Normal
RF spoiling	Off
Excite pulse duration	5820 us
Slice multiplier	1
Multi-band PE shift	0 1/FoV
zBlip scheme	0
MB kernel size	0
MB knockout band	0
No. of interleaved TEs	0
RF pulse shape	1
EPI noise scans	0
EPI full reference scan	0
Single-band images	On
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Readout slice trace	Off
Disable ramp sampling	Off
PF omits higher k-space	Off
Online multi-band recon.	Online
FFT scale factor	0.02
Send B1 shim trigger	Never
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\Sam-8chcoil\GE_p8mm_SB1IPAT3_pf6_te23_tr3000_sat_bowties_102i

TA: 5:29 PAT: 3 Voxel size: 0.8x0.8x0.8 mm Rel. SNR: 1.00 USER: AV_ep2d_bold_sd_20140727

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	43
Dist. factor	0 %
Position	L2.1 A10.0 F9.9
Orientation	T > C29.9
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	180 mm
FoV phase	77.7 %
Slice thickness	0.80 mm
TR	3000 ms
TE	23.4 ms
Multi-band accel. factor	1
Filter	None
Coil elements	B4;M2,3;T1

Contrast

MTC	Off
Magn. preparation	None
Flip angle	60 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	102
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	224
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	48
Reference scan mode	GRE
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Sat. region 1

Thickness	50 mm
Position	L10.9 A51.7 F73.1
Orientation	T > C-33.5 > S4.0
Special sat.	None

Table position	H
Table position	0 mm
Inline Composing	Off

System

T1	On
M2	On
B4	On
M3	On
V32	Off

Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default

Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	On
Assume Silicone	Off
! Ref. amplitude 1H	230.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L2.1 A10.0 F9.9
Orientation	T > C29.9
Rotation	0.00 deg
R >> L	180 mm
A >> P	140 mm
F >> H	35 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	12
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
Bandwidth	1174 Hz/Px
Flow comp.	No
Free echo spacing	Off
Echo spacing	1 ms
<hr/>	
SIR accel. factor	1
EPI factor	174
Gradient mode	Normal
RF spoiling	Off
<hr/>	
Excite pulse duration	3640 us
Slice multiplier	1
Fake MB factor for SB	1
No. of interleaved TEs	0
RF pulse shape	1
EPI noise scans	0
EPI full reference scan	0
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Readout slice trace	Off
Disable ramp sampling	Off
PF omits higher k-space	Off
FFT scale factor	0.10
GRE iPAT ref. FA	12.0 deg
Send B1 shim trigger	Never
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\Sam-8chcoil\GE_p8mm_SB1IPAT3_pf6_te23_tr3000_sat_bowties_102i

TA: 5:29 PAT: 3 Voxel size: 0.8x0.8x0.8 mm Rel. SNR: 1.00 USER: AV_ep2d_bold_sd_20140727

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	43
Dist. factor	0 %
Position	L2.1 A10.0 F9.9
Orientation	T > C29.9
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	180 mm
FoV phase	77.7 %
Slice thickness	0.80 mm
TR	3000 ms
TE	23.4 ms
Multi-band accel. factor	1
Filter	None
Coil elements	B4;M2,3;T1

Contrast

MTC	Off
Magn. preparation	None
Flip angle	60 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	102
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	224
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	48
Reference scan mode	GRE
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Sat. region 1

Thickness	50 mm
Position	L10.9 A51.7 F73.1
Orientation	T > C-33.5 > S4.0
Special sat.	None

Table position	H
Table position	0 mm
Inline Composing	Off

System

T1	On
M2	On
B4	On
M3	On
V32	Off

Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default

Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	On
Assume Silicone	Off
! Ref. amplitude 1H	230.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L2.1 A10.0 F9.9
Orientation	T > C29.9
Rotation	0.00 deg
R >> L	180 mm
A >> P	140 mm
F >> H	35 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	12
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
Bandwidth	1174 Hz/Px
Flow comp.	No
Free echo spacing	Off
Echo spacing	1 ms
<hr/>	
SIR accel. factor	1
EPI factor	174
Gradient mode	Normal
RF spoiling	Off
<hr/>	
Excite pulse duration	3640 us
Slice multiplier	1
Fake MB factor for SB	1
No. of interleaved TEs	0
RF pulse shape	1
EPI noise scans	0
EPI full reference scan	0
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Readout slice trace	Off
Disable ramp sampling	Off
PF omits higher k-space	Off
FFT scale factor	0.10
GRE iPAT ref. FA	12.0 deg
Send B1 shim trigger	Never
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\Sam-8chcoil\BP_ep3D_bold_multiecho_new_SH_CAIPi_noICE

TA: 4:50 PAT: 4 Voxel size: 0.9x0.9x0.9 mm Rel. SNR: 1.00 USER: BP_ep3D_bold_multiecho_SH

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	80
FoV read	198 mm
FoV phase	100.0 %
Slice thickness	0.90 mm
TR	50 ms
TE	20 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	B4;M2,3;T1

Contrast

MTC	Off
Flip angle	13 deg
Fat suppr.	None
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	92
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	220
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	6/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	4
Ref. lines PE	96
Accel. factor 3D	1
Ref. lines 3D	32
Reference scan mode	Separate
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	On

Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved
Special sat.	None
Table position	H
Table position	0 mm
Inline Composing	Off

System

T1	On
M2	On
B4	On
M3	On
V32	Off
Positioning mode	REF
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	200.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	198 mm
A >> P	198 mm
F >> H	72 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
Dimension	3D
Reordering	Linear
Contrasts	1
Bandwidth	1262 Hz/Px
Free echo spacing	Off
Echo spacing	0.9 ms
EPI factor	220
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slab-sel.
RF spoiling	On
use Ernst angle	Off
Maxwell Correction	Off
log physio files	Off

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

FFT scale	1.00
z shim	0.00 mT/m*ms
RF duration	2560 us
RF BWTP	5.2
EFFECTIVE TR	3000 ms
PatPartitions	60
EPI phase correction	local
PAT refscan mode	segm LIN->PAR
use CAIPI	On
CAIPI shift kz	0
CAIPI shift ky	1
dummy prepscan time	3 s
silent gap	0.000 s

Table of contents

\\USER

Feinberglab
Jen

Sam-8chcoil

localizer_50V_newcoil

AV_ep2d_bold_sd1ipat2mb2_pt5mm_visLoc_8

AV_ep2d_bold_sd1ipat2mb2_pt5mm_visLoc_8

---32ch coil comparison----

GE_p8mm_SB1IPAT3_pf6_te23_tr3000_sat_bowties_102i

GE_p8mm_SB1IPAT3_pf6_te23_tr3000_sat_bowties_102i

BP_ep3D_bold_multiecho_new_SH_CAIPI_nolCE