

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\gonogo frontal pilot\localizer_200V_nova

TA: 0:27 PAT: Off Voxel size: 1.2x1.1x3.0 mm Rel. SNR: 1.00 SIEMENS: gre

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	5
Dist. factor	500 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 2	
Slices	5
Dist. factor	20 %
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
Rotation	0.00 deg
Slice group 3	
Slices	5
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	280 mm
FoV phase	100.0 %
Slice thickness	3.0 mm
TR	10.0 ms
TE	3.00 ms
Averages	1
Concatenations	15
Filter	None
Coil elements	B4;M2,3;T1

Contrast

TD	0 ms
MTC	Off
Magn. preparation	None
Flip angle	10 deg
Fat suppr.	None
Water suppr.	None
SWI	Off
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	256
-----------------	-----

Phase resolution	90 %
Phase partial Fourier	6/8
Interpolation	On
PAT mode	None
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Sequential
Series	Interleaved
Saturation mode	Standard
Special sat.	None
Table position	H
Table position	0 mm
Inline Composing	Off
Tim CT mode	Off

System

T1	On
M2	On
B4	On
M3	On
V32	Off
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	On
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Off
Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	200.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	350 mm
A >> P	263 mm
F >> H	350 mm

Physio

1st Signal/Mode	None
Segments	1
Tagging	None
Dark blood	Off
Resp. control	Off

Inline

Subtract	Off
----------	-----

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Liver registration	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On
<hr/>	
Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
<hr/>	
MapIt	None
Contrasts	1

Sequence

Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Allowed
Bandwidth	320 Hz/Px
Flow comp.	No
<hr/>	
RF pulse type	Normal
Gradient mode	Whisper
Excitation	Slice-sel.
RF spoiling	On

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\gonogo frontal pilot\b1map_200V_32

TA: 2:10

Voxel size: 3.9x3.9x5.0 mm

Rel. SNR: 1.00

USER: b1map_658

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	12
Dist. factor	100 %
Position	R0.7 A30.3 F0.6
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
FoV read	250 mm
FoV phase	100.0 %
Slice thickness	5 mm
TR	1938 ms
TE 1	14 ms
TE 2	14 ms
Averages	1
Filter	None
Coil elements	B4;M2,3;T1

Contrast

Flip angle 1	90 deg
Flip angle 2	120 deg
Flip angle 3	60 deg
Flip angle 4	135 deg
Flip angle 5	45 deg
Measurements	1

Resolution

Base resolution	64
Phase resolution	100 %
Raw filter	Off

Geometry

Series	Interleaved
Navigator 1	
Position	L0.0 P35.8 F18.2
Orientation	Transversal
Rotation	0.00 deg
Base size phase	50 mm
Base size read	50 mm
Thickness	50 mm
Table position	H
Table position	0 mm
Inline Composing	Off

System

T1	On
M2	On
B4	On

M3 On
V32 Off

Positioning mode FIX
MSMA S - C - T
Sagittal R >> L
Coronal A >> P
Transversal F >> H
Save uncombined Off
Coil Combine Mode Adaptive Combine
AutoAlign ---
Auto Coil Select Default

Shim mode Tune up
Adjust with body coil Off
Confirm freq. adjustment Off
Assume Silicone Off
! Ref. amplitude 1H 200.000 V
Adjustment Tolerance Auto
Adjust volume
Position Isocenter
Orientation Transversal
Rotation 0.00 deg
R >> L 350 mm
A >> P 263 mm
F >> H 350 mm

Composing

Sequence

Contrasts 2
Bandwidth 260.416667 Hz/Px
T1 Compensation Mean T1
Mean T1 1000.0 ms
Angles 1
Amplitude Weighting Linear
Scale Bar Enabled
Raw Data Disabled

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\gonogo frontal pilot\mp2rage_1mm_TR4000

TA: 5:26 PAT: 3 Voxel size: 1.0x1.0x1.0 mm Rel. SNR: 1.00 USER: mp2rage_wip602B

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	L1.9 A29.6 F31.5
Orientation	Sagittal
Phase enc. dir.	H >> F
Rotation	90.00 deg
Phase oversampling	0 %
Slice oversampling	11.1 %
Slices per slab	144
FoV read	200 mm
FoV phase	90.6 %
Slice thickness	1.00 mm
TR	4000 ms
TE	3.23 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	B4;M2,3;T1

Contrast

Magn. preparation	Non-sel. IR
TI 1	1000 ms
TI 2	3200 ms
Flip angle 1	4 deg
Flip angle 2	4 deg
Fat suppr.	Water excit. fast
Water suppr.	None
2nd Inversion-Contrast	On
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	192
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	6/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	36
Accel. factor 3D	1
Reference scan mode	Integrated

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Single shot
Series	Interleaved
Table position	H
Table position	0 mm
Inline Composing	Off

System

T1	On
M2	On
B4	On
M3	On
V32	Off
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	240.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	L1.9 A24.9 F9.3
! Orientation	Sagittal
! Rotation	0.00 deg
! F >> H	108 mm
! A >> P	160 mm
! R >> L	127 mm

Physio

1st Signal/Mode	None
Dark blood	Off
Resp. control	Off

Inline

Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Sequence

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Introduction	On
Dimension	3D
Elliptical scanning	Off
Asymmetric echo	Off
Contrasts	1
Bandwidth	200 Hz/Px
Flow comp.	Slice
Echo spacing	7.8 ms
<hr/>	
RF pulse type	Fast
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On
<hr/>	
FFT Scale Factor	200 %
Line/Partition Swap	Off
Homodyne Phase Filter	Off
Flat Image	On
T1 Map	On
Division Image	Off
ExtInvPulseOn	On
OffResFreqInv	0
Invflipangle	970

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\gonogo frontal pilot\GE_1mm_MB2IPAT3_pf5_302i_gonogo_test_tr1000
 TA: 5:19 PAT: 3 Voxel size: 1.0x1.0x1.0 mm Rel. SNR: 1.00 USER: AV_ep2d_bold_sd_20140727

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	30
Dist. factor	0 %
Position	R1.7 A23.6 H11.6
Orientation	T > C-2.6
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	200 mm
FoV phase	93.0 %
Slice thickness	1.00 mm
TR	1000 ms
TE	22.8 ms
Multi-band accel. factor	2
Filter	None
Coil elements	B4;M2,3;T1

Contrast

MTC	Off
Magn. preparation	None
Flip angle	50 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	302
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	200
Phase resolution	100 %
Phase partial Fourier	5/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	48
Reference scan mode	GRE
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Special sat. None

Table position	H
Table position	0 mm
Inline Composing	Off

System

T1	On
M2	On
B4	On
M3	On
V32	Off
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	On
Assume Silicone	Off
! Ref. amplitude 1H	230.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	R1.7 A23.6 H11.6
Orientation	T > C-2.6
Rotation	0.00 deg
R >> L	200 mm
A >> P	186 mm
F >> H	30 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
Bandwidth	1250 Hz/Px
Flow comp.	No
Free echo spacing	Off
Echo spacing	1 ms
<hr/>	
SIR accel. factor	1
EPI factor	186
Gradient mode	Normal
RF spoiling	Off
<hr/>	
Excite pulse duration	3640 us
Slice multiplier	1
Multi-band PE shift	0 1/FoV
zBlip scheme	0
MB kernel size	0
MB knockout band	0
No. of interleaved TEs	0
RF pulse shape	1
EPI noise scans	0
EPI full reference scan	0
Single-band images	On
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Readout slice trace	Off
Disable ramp sampling	Off
PF omits higher k-space	Off
Online multi-band recon.	Online
FFT scale factor	0.10
GRE iPAT ref. FA	12.0 deg
Send B1 shim trigger	Never
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\gonogo frontal pilot\GE_1mm_MB2IPAT3_pf5_151i_gonogo_test_tr2000

TA: 5:32 PAT: 3 Voxel size: 1.0x1.0x1.0 mm Rel. SNR: 1.00 USER: AV_ep2d_bold_sd_20140727

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	60
Dist. factor	0 %
Position	R1.7 A23.6 H11.6
Orientation	T > C-2.6
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	200 mm
FoV phase	93.0 %
Slice thickness	1.00 mm
TR	2000 ms
TE	22.8 ms
Multi-band accel. factor	2
Filter	None
Coil elements	B4;M2,3;T1

Contrast

MTC	Off
Magn. preparation	None
Flip angle	70 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	151
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	200
Phase resolution	100 %
Phase partial Fourier	5/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	48
Reference scan mode	GRE
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Special sat. None

Table position	H
Table position	0 mm
Inline Composing	Off

System

T1	On
M2	On
B4	On
M3	On
V32	Off
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	On
Assume Silicone	Off
! Ref. amplitude 1H	230.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	R1.7 A23.6 H11.6
Orientation	T > C-2.6
Rotation	0.00 deg
R >> L	200 mm
A >> P	186 mm
F >> H	60 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
Bandwidth	1250 Hz/Px
Flow comp.	No
Free echo spacing	Off
Echo spacing	1 ms
<hr/>	
SIR accel. factor	1
EPI factor	186
Gradient mode	Normal
RF spoiling	Off
<hr/>	
Excite pulse duration	3640 us
Slice multiplier	1
Multi-band PE shift	0 1/FoV
zBlip scheme	0
MB kernel size	0
MB knockout band	0
No. of interleaved TEs	0
RF pulse shape	1
EPI noise scans	0
EPI full reference scan	0
Single-band images	On
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Readout slice trace	Off
Disable ramp sampling	Off
PF omits higher k-space	Off
Online multi-band recon.	Online
FFT scale factor	0.10
GRE iPAT ref. FA	12.0 deg
Send B1 shim trigger	Never
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\gonogo frontal pilot\GE_p8mm_MB2IPAT3_pf5_151i_gonogo_tr2000

TA: 5:34 PAT: 3 Voxel size: 0.8x0.8x0.8 mm Rel. SNR: 1.00 USER: AV_ep2d_bold_sd_20140727

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	60
Dist. factor	0 %
Position	L1.4 A31.7 F4.9
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	160 mm
FoV phase	91.0 %
Slice thickness	0.80 mm
TR	2000 ms
TE	22.8 ms
Multi-band accel. factor	2
Filter	None
Coil elements	B4;M2,3;T1

Contrast

MTC	Off
Magn. preparation	None
Flip angle	70 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	151
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	200
Phase resolution	100 %
Phase partial Fourier	5/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	48
Reference scan mode	GRE
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Special sat. None

Table position	H
Table position	0 mm
Inline Composing	Off

System

T1	On
M2	On
B4	On
M3	On
V32	Off

Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default

Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	On
Assume Silicone	Off
! Ref. amplitude 1H	230.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L1.4 A31.7 F4.9
Orientation	Transversal
Rotation	0.00 deg
R >> L	160 mm
A >> P	146 mm
F >> H	48 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
Bandwidth	1250 Hz/Px
Flow comp.	No
Free echo spacing	Off
Echo spacing	1.01 ms
<hr/>	
SIR accel. factor	1
EPI factor	182
Gradient mode	Normal
RF spoiling	Off
<hr/>	
Excite pulse duration	3640 us
Slice multiplier	1
Multi-band PE shift	0 1/FoV
zBlip scheme	0
MB kernel size	0
MB knockout band	0
No. of interleaved TEs	0
RF pulse shape	1
EPI noise scans	0
EPI full reference scan	0
Single-band images	On
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Readout slice trace	Off
Disable ramp sampling	Off
PF omits higher k-space	Off
Online multi-band recon.	Online
FFT scale factor	0.10
GRE iPAT ref. FA	12.0 deg
Send B1 shim trigger	Never
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\gonogo frontal pilot\GE_p8mm_MB2IPAT3_pf5_302i_gonogo_tr1000

TA: 2:49 PAT: 3 Voxel size: 0.8x0.8x0.8 mm Rel. SNR: 1.00 USER: AV_ep2d_bold_sd_20140727

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	30
Dist. factor	0 %
Position	L1.4 A31.7 F4.9
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	160 mm
FoV phase	91.0 %
Slice thickness	0.80 mm
TR	1000 ms
TE	22.8 ms
Multi-band accel. factor	2
Filter	None
Coil elements	B4;M2,3;T1

Contrast

MTC	Off
Magn. preparation	None
Flip angle	50 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	151
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	200
Phase resolution	100 %
Phase partial Fourier	5/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	48
Reference scan mode	GRE
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Special sat. None

Table position	H
Table position	0 mm
Inline Composing	Off

System

T1	On
M2	On
B4	On
M3	On
V32	Off

Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default

Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	On
Assume Silicone	Off
! Ref. amplitude 1H	230.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L1.4 A31.7 F4.9
Orientation	Transversal
Rotation	0.00 deg
R >> L	160 mm
A >> P	146 mm
F >> H	25 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
Bandwidth	1250 Hz/Px
Flow comp.	No
Free echo spacing	Off
Echo spacing	1.01 ms
<hr/>	
SIR accel. factor	1
EPI factor	182
Gradient mode	Normal
RF spoiling	Off
<hr/>	
Excite pulse duration	3640 us
Slice multiplier	1
Multi-band PE shift	0 1/FoV
zBlip scheme	0
MB kernel size	0
MB knockout band	0
No. of interleaved TEs	0
RF pulse shape	1
EPI noise scans	0
EPI full reference scan	0
Single-band images	On
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Readout slice trace	Off
Disable ramp sampling	Off
PF omits higher k-space	Off
Online multi-band recon.	Online
FFT scale factor	0.10
GRE iPAT ref. FA	12.0 deg
Send B1 shim trigger	Never
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\gonogo frontal pilot\b1map_200V_TR1000_nova

TA: 1:09

Voxel size: 3.9x3.9x5.0 mm

Rel. SNR: 1.00

USER: b1map_658

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	10
Dist. factor	100 %
Position	R0.7 A36.4 H11.5
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
FoV read	250 mm
FoV phase	100.0 %
Slice thickness	5 mm
TR	1000 ms
TE 1	14 ms
TE 2	14 ms
Averages	1
Filter	None
Coil elements	B4;M2,3;T1

Contrast

Flip angle 1	90 deg
Flip angle 2	120 deg
Flip angle 3	60 deg
Flip angle 4	135 deg
Flip angle 5	45 deg
Measurements	1

Resolution

Base resolution	64
Phase resolution	100 %
Raw filter	Off

Geometry

Series	Interleaved
Navigator 1	
Position	R2.0 P26.3 F10.8
Orientation	Transversal
Rotation	0.00 deg
Base size phase	50 mm
Base size read	119 mm
Thickness	50 mm
Table position	H
Table position	0 mm
Inline Composing	Off

System

T1	On
M2	On
B4	On

M3 On
V32 Off

Positioning mode FIX
MSMA S - C - T
Sagittal R >> L
Coronal A >> P
Transversal F >> H
Save uncombined Off
Coil Combine Mode Adaptive Combine
AutoAlign ---
Auto Coil Select Default

Shim mode Tune up
Adjust with body coil Off
Confirm freq. adjustment Off
Assume Silicone Off
! Ref. amplitude 1H 200.000 V
Adjustment Tolerance Auto
Adjust volume
Position Isocenter
Orientation Transversal
Rotation 0.00 deg
R >> L 350 mm
A >> P 263 mm
F >> H 350 mm

Composing

Sequence

Contrasts 2
Bandwidth 260.416667 Hz/Px
T1 Compensation Mean T1
Mean T1 1000.0 ms
Angles 1
Amplitude Weighting Linear
Scale Bar Enabled
Raw Data Disabled

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\gonogo frontal pilot\GE_p8mm_MB2IPAT3_pf6_151i_gonogo_tr2000

TA: 5:33 PAT: 3 Voxel size: 0.8x0.8x0.8 mm Rel. SNR: 1.00 USER: AV_ep2d_bold_sd_20140727

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	58
Dist. factor	0 %
Position	L1.4 A31.7 F4.9
Orientation	Transversal
Phase enc. dir.	R >> L
Rotation	90.00 deg
Phase oversampling	0 %
FoV read	160 mm
FoV phase	91.0 %
Slice thickness	0.80 mm
TR	2000 ms
TE	23.6 ms
Multi-band accel. factor	2
Filter	None
Coil elements	B4;M2,3;T1

Contrast

MTC	Off
Magn. preparation	None
Flip angle	70 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	151
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	200
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	48
Reference scan mode	GRE
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Special sat. None

Table position	H
Table position	0 mm
Inline Composing	Off

System

T1	On
M2	On
B4	On
M3	On
V32	Off

Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default

Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	On
Assume Silicone	Off
! Ref. amplitude 1H	230.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L1.4 A31.7 F4.9
Orientation	Transversal
Rotation	90.00 deg
A >> P	160 mm
R >> L	146 mm
F >> H	47 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
Bandwidth	1250 Hz/Px
Flow comp.	No
Free echo spacing	Off
Echo spacing	1.01 ms
SIR accel. factor	1
EPI factor	182
Gradient mode	Normal
RF spoiling	Off
Excite pulse duration	3640 us
Slice multiplier	1
Multi-band PE shift	0 1/FoV
zBlip scheme	0
MB kernel size	0
MB knockout band	0
No. of interleaved TEs	0
RF pulse shape	1
EPI noise scans	0
EPI full reference scan	0
Single-band images	On
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Readout slice trace	Off
Disable ramp sampling	Off
PF omits higher k-space	Off
Online multi-band recon.	Online
FFT scale factor	0.10
GRE iPAT ref. FA	12.0 deg
Send B1 shim trigger	Never
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\gonogo frontal pilot\GE_p8mm_MB2IPAT3_pf6_202i_gonogo_tr1500

TA: 5:28 PAT: 3 Voxel size: 0.8x0.8x0.8 mm Rel. SNR: 1.00 USER: AV_ep2d_bold_sd_20140727

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	44
Dist. factor	0 %
Position	L1.4 A31.7 F4.9
Orientation	Transversal
Phase enc. dir.	R >> L
Rotation	90.00 deg
Phase oversampling	0 %
FoV read	160 mm
FoV phase	91.0 %
Slice thickness	0.80 mm
TR	1500 ms
TE	23.6 ms
Multi-band accel. factor	2
Filter	None
Coil elements	B4;M2,3;T1

Contrast

MTC	Off
Magn. preparation	None
Flip angle	60 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	202
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	200
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	48
Reference scan mode	GRE
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Special sat. None

Table position	H
Table position	0 mm
Inline Composing	Off

System

T1	On
M2	On
B4	On
M3	On
V32	Off

Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default

Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	On
Assume Silicone	Off
! Ref. amplitude 1H	230.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L1.4 A31.7 F4.9
Orientation	Transversal
Rotation	90.00 deg
A >> P	160 mm
R >> L	146 mm
F >> H	36 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
Bandwidth	1250 Hz/Px
Flow comp.	No
Free echo spacing	Off
Echo spacing	1.01 ms
<hr/>	
SIR accel. factor	1
EPI factor	182
Gradient mode	Normal
RF spoiling	Off
<hr/>	
Excite pulse duration	3640 us
Slice multiplier	1
Multi-band PE shift	0 1/FoV
zBlip scheme	0
MB kernel size	0
MB knockout band	0
No. of interleaved TEs	0
RF pulse shape	1
EPI noise scans	0
EPI full reference scan	0
Single-band images	On
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Readout slice trace	Off
Disable ramp sampling	Off
PF omits higher k-space	Off
Online multi-band recon.	Online
FFT scale factor	0.10
GRE iPAT ref. FA	12.0 deg
Send B1 shim trigger	Never
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\gonogo frontal pilot\GE_p8mm_MB2IPAT3_pf6_302i_gonogo_tr1000

TA: 5:19 PAT: 3 Voxel size: 0.8x0.8x0.8 mm Rel. SNR: 1.00 USER: AV_ep2d_bold_sd_20140727

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	28
Dist. factor	0 %
Position	L1.4 A31.7 F4.9
Orientation	Transversal
Phase enc. dir.	R >> L
Rotation	90.00 deg
Phase oversampling	0 %
FoV read	160 mm
FoV phase	91.0 %
Slice thickness	0.80 mm
TR	1000 ms
TE	23.8 ms
Multi-band accel. factor	2
Filter	None
Coil elements	B4;M2,3;T1

Contrast

MTC	Off
Magn. preparation	None
Flip angle	50 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	302
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	200
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	48
Reference scan mode	GRE
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Special sat. None

Table position	H
Table position	0 mm
Inline Composing	Off

System

T1	On
M2	On
B4	On
M3	On
V32	Off

Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default

Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	On
Assume Silicone	Off
! Ref. amplitude 1H	230.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L1.4 A31.7 F4.9
Orientation	Transversal
Rotation	90.00 deg
A >> P	160 mm
R >> L	146 mm
F >> H	23 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
Bandwidth	1250 Hz/Px
Flow comp.	No
Free echo spacing	Off
Echo spacing	1.01 ms
<hr/>	
SIR accel. factor	1
EPI factor	182
Gradient mode	Normal
RF spoiling	Off
<hr/>	
Excite pulse duration	3640 us
Slice multiplier	1
Multi-band PE shift	0 1/FoV
zBlip scheme	0
MB kernel size	0
MB knockout band	0
No. of interleaved TEs	0
RF pulse shape	1
EPI noise scans	0
EPI full reference scan	0
Single-band images	On
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Readout slice trace	Off
Disable ramp sampling	Off
PF omits higher k-space	Off
Online multi-band recon.	Online
FFT scale factor	0.10
GRE iPAT ref. FA	12.0 deg
Send B1 shim trigger	Never
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\gonogo frontal pilot\GE_1mm_MB2IPAT3_pf6_151i_gonogo_test_tr2000

TA: 5:30 PAT: 3 Voxel size: 1.0x1.0x1.5 mm Rel. SNR: 1.00 USER: AV_ep2d_bold_sd_20140727

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	58
Dist. factor	0 %
Position	R1.7 A23.6 H11.6
Orientation	T > C-2.6
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	200 mm
FoV phase	93.0 %
Slice thickness	1.50 mm
TR	2000 ms
TE	24.0 ms
Multi-band accel. factor	2
Filter	None
Coil elements	B4;M2,3;T1

Contrast

MTC	Off
Magn. preparation	None
Flip angle	70 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	151
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	200
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	48
Reference scan mode	GRE
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Special sat. None

Table position	H
Table position	0 mm
Inline Composing	Off

System

T1	On
M2	On
B4	On
M3	On
V32	Off
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	On
Assume Silicone	Off
! Ref. amplitude 1H	230.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	R1.7 A23.6 H11.6
Orientation	T > C-2.6
Rotation	0.00 deg
R >> L	200 mm
A >> P	186 mm
F >> H	87 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active

Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
Bandwidth	1250 Hz/Px
Flow comp.	No
Free echo spacing	Off
Echo spacing	1 ms
<hr/>	
SIR accel. factor	1
EPI factor	186
Gradient mode	Normal
RF spoiling	Off
<hr/>	
Excite pulse duration	3640 us
Slice multiplier	1
Multi-band PE shift	0 1/FoV
zBlip scheme	0
MB kernel size	0
MB knockout band	0
No. of interleaved TEs	0
RF pulse shape	1
EPI noise scans	0
EPI full reference scan	0
Single-band images	On
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Readout slice trace	Off
Disable ramp sampling	Off
PF omits higher k-space	Off
Online multi-band recon.	Online
FFT scale factor	0.10
GRE iPAT ref. FA	12.0 deg
Send B1 shim trigger	Never
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Jen\gonogo frontal pilot\GE_1p4mm_MB2IPAT3_pf6_tr2255_te23_68SL_320i_resting

TA: 12:24 PAT: 3 Voxel size: 1.4x1.4x1.4 mm Rel. SNR: 1.00 USER: AV_ep2d_bold_sd_20140727

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	68
Dist. factor	0 %
Position	L1.4 A12.2 F0.9
Orientation	T > C-5.4
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	280 mm
FoV phase	91.0 %
Slice thickness	1.40 mm
TR	2255 ms
TE	23.0 ms
Multi-band accel. factor	2
Filter	None
Coil elements	B4;M2,3;T1

Contrast

MTC	Off
Magn. preparation	None
Flip angle	89 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	320
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	200
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	48
Reference scan mode	Segmented
Distortion Corr.	Off
Prescan Normalize	Off
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Special sat. None

Table position	H
Table position	0 mm
Inline Composing	Off

System

T1	On
M2	On
B4	On
M3	On
V32	Off

Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default

Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	On
Assume Silicone	Off
! Ref. amplitude 1H	200.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L1.4 A12.2 F0.9
Orientation	T > C-5.4
Rotation	0.00 deg
R >> L	280 mm
A >> P	255 mm
F >> H	96 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	12
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
Bandwidth	1250 Hz/Px
Flow comp.	No

Free echo spacing	Off
Echo spacing	1 ms
SIR accel. factor	1
EPI factor	182
Gradient mode	Normal
RF spoiling	Off
Excite pulse duration	3640 us
Slice multiplier	1
Multi-band PE shift	0 1/FoV
zBlip scheme	0
MB kernel size	0
MB knockout band	0
No. of interleaved TEs	0
RF pulse shape	1
EPI noise scans	0
EPI full reference scan	0
Single-band images	On
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Readout slice trace	Off
Disable ramp sampling	Off
PF omits higher k-space	Off
Online multi-band recon.	Online
FFT scale factor	0.10
Send B1 shim trigger	Never
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

Table of contents

\\USER

Feinberglab
Jen

gonogo frontal pilot

localizer_200V_nova

b1map_200V_32

mp2rage_1mm_TR4000

GE_1mm_MB2IPAT3_pf5_302i_gonogo_test_tr1000

GE_1mm_MB2IPAT3_pf5_151i_gonogo_test_tr2000

GE_p8mm_MB2IPAT3_pf5_151i_gonogo_tr2000

GE_p8mm_MB2IPAT3_pf5_302i_gonogo_tr1000

b1map_200V_TR1000_nova

GE_p8mm_MB2IPAT3_pf6_151i_gonogo_tr2000

GE_p8mm_MB2IPAT3_pf6_202i_gonogo_tr1500

GE_p8mm_MB2IPAT3_pf6_302i_gonogo_tr1000

07112019

GE_1mm_MB2IPAT3_pf6_151i_gonogo_test_tr2000

GE_1p4mm_MB2IPAT3_pf6_tr2255_te23_68SL_320i_resting