\\USER\Testing\Testing\VAARC_Phantom_Tune_up_20210326\localizer

TA: 0:27 PM: FIX Voxel size: 0.5×0.5×3.0 mmPAT: Off Rel. SNR: 1.00 : qfl

Properties

Prio recon	On
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	5
Dist. factor	200 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	5
Dist. factor	200 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Slices	5
Dist. factor	200 %
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	---
Phase oversampling	0 %
FoV read	280 mm
FoV phase	100.0 %
Slice thickness	3.0 mm
TR	10.0 ms
TE	3.00 ms
Averages	1
Concatenations	15
Filter	Elliptical filter
Coil elements	TC

Contrast - Common

TR	10.0 ms
TE	3.00 ms
TD	0 ms
MTC	Off
Magn. preparation	None
Flip angle	10 deg
Fat suppr.	None
Water suppr.	None
SWI	Off

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Contrast - Dynamic

Multiple series	Each measurement

Resolution - Common

FoV read	280 mm
FoV phase	100.0 %
Slice thickness	3.0 mm
Base resolution	256
Phase resolution	90 %
Phase partial Fourier	6/8
Interpolation	On

Resolution - iPAT

PAT mode	None

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	5
Dist. factor	200 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	5
Dist. factor	200 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Slices	5
Dist. factor	200 %
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
FoV read	280 mm
FoV phase	100.0 %
Slice thickness	3.0 mm
TR	10.0 ms
Multi-slice mode	Sequential
Series	Interleaved
Concatenations	15

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Position	Isocenter
Orientation	Transversal

Geometry - AutoAlign

Phase enc. dir.	A >> P
Slice group	3
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

Geometry - Tim CT

Tim CT mode	Off
Slices	5
Slice thickness	3.0 mm
Dist. factor	200 %
FoV read	280 mm
FoV phase	100.0 %
Segments	1

System - Miscellaneous

Positioning mode	FIX
Table position	F
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.211069 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000

System - Tx/Rx

Reset	Off
! Ref. amplitude 1H	230.000 V

Physio - Signal1

1st Signal/Mode	None
TR	10.0 ms
Concatenations	15
Segments	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	280 mm
FoV phase	100.0 %
Phase resolution	90 %

Physio - PACE

Resp. control	Off
Concatenations	15

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Distortion Corr.	Off

Sequence - Part 1

Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Allowed
Contrasts	1
Flow comp.	No
Multi-slice mode	Sequential
Bandwidth	320 Hz/Px

Sequence - Part 2

Segments	1
Acoustic noise reduction	Active
RF pulse type	Normal
Gradient mode	Normal
Excitation	Slice-sel.
RF spoiling	On

Sequence - Assistant

Mode	Off

\\USER\Testing\Testing\VAARC_Phantom_Tune_up_20210326\b1map_150V	
TA: 0:16 PM: FIX Voxel size: 4.0×4.0×4.0 mmPAT: Off Rel. SNR: 1.00 : tfl	

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	25
Dist. factor	100 %
Position	L1.3 A6.1 F16.2 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	7270.0 ms
TE	1.72 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	TC

Contrast - Common

TR	7270.0 ms
TE	1.72 ms
Magn. preparation	None
Flip angle	10 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None

Resolution - Filter Image

Image Filter	Off

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	25
Dist. factor	100 %
Position	L1.3 A6.1 F16.2 mm
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	7270.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	L1.3 A6.1 F16.2 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	L1.3 A6.1 F16.2
L	1.3 mm
A	6.1 mm
F	16.2 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.211069 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	150.000 V

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off

Sequence - Part 1

Introduction	Off
Dimension	2D
Asymmetric echo	Allowed
Flow comp.	No
Multi-slice mode	Interleaved
Echo spacing	4.5 ms
Bandwidth	490 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Normal
Excitation	Slice-sel.
RF spoiling	On
Turbo factor	64

Sequence - Assistant

Mode	Off

\\USER\Testing\Testing\VAARC_Phantom_Tune_up_20210326\b1map_230V

TA: 0:16 PM: FIX Voxel size: 4.0×4.0×4.0 mmPAT: Off Rel. SNR: 1.00 : tfl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	25
Dist. factor	100 %
Position	L1.3 A6.1 F16.2 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	7270.0 ms
TE	1.72 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	TC

Contrast - Common

TR	7270.0 ms
TE	1.72 ms
Magn. preparation	None
Flip angle	10 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None

Resolution - Filter Image

Image Filter	Off

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	25
Dist. factor	100 %
Position	L1.3 A6.1 F16.2 mm
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	7270.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	L1.3 A6.1 F16.2 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	L1.3 A6.1 F16.2
L	1.3 mm
A	6.1 mm
F	16.2 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.211069 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	230.000 V

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off

Sequence - Part 1

Introduction	Off
Dimension	2D
Asymmetric echo	Allowed
Flow comp.	No
Multi-slice mode	Interleaved
Echo spacing	4.5 ms
Bandwidth	490 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Normal
Excitation	Slice-sel.
RF spoiling	On
Turbo factor	64

Sequence - Assistant

Mode	Off

\\USER\Testing\Testing\VAARC_Phantom_Tune_up_20210326\ep2d_bold_1mm_reshim
TA: 1:37 PM: FIX Voxel size: 1.0×1.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	40
Dist. factor	0 %
Position	L2.7 A6.1 F12.8 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	130 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	4640 ms
TE	40.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	TC

Contrast - Common

TR	4640 ms
TE	40.0 ms
MTC	Off
Flip angle exc	90 deg
Flip angle fat sat	110 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	20
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	130 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
Base resolution	130
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

Accel. mode	None

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	40
Dist. factor	0 %
Position	L2.7 A6.1 F12.8 mm
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	130 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	4640 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	L2.7 A6.1 F12.8 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	L2.7 A6.1 F12.8
L	2.7 mm
A	6.1 mm
F	12.8 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L2.7 A6.1 F12.8 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	130 mm
R >> L	130 mm
F >> H	80 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.211069 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	4640 ms
Concatenations	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off
Measurements	20
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.8 ms
Bandwidth	1326 Hz/Px

Sequence - Part 2

EPI factor	130
RF pulse type	Fast
Gradient mode	Fast