SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\Feinberglab\Tanja\T1\ep2d_se

		Inline Composing	Off
operties			OII
Prio Recon	Off	System	
Before measurement		T1	On O#
After measurement		M2	Off
Load to viewer	On O"	B4	Off
Inline movie	Off	M3	Off
Auto store images	On	V32	Off
Load to stamp segments	Off	Positioning mode	REF
Load images to graphic	Off	MSMA	S - C - T
segments		Sagittal	R >> L
Auto open inline display	Off	Coronal	A >> P
Start measurement without	On	Transversal	F >> H
further preparation		Save uncombined	Off
Wait for user to start	Off		
Start measurements	single	Coil Combine Mode	Sum of Squares
outin o		Auto Ocil Colort	D-fIt
outine		Auto Coil Select	Default
Slice group 1		Shim mode	Standard
Slices	1	Adjust with body coil	Off
Dist. factor	50 %	Confirm freq. adjustment	Off
Position	Isocenter .	Assume Silicone	Off
Orientation	Transversal	? Ref. amplitude 1H	0.000 V
Phase enc. dir.	A >> P	Adjustment Tolerance	Auto
Rotation	0.00 deg	Adjust volume	Adio
Phase oversampling	0 %	Position	Isocenter
FoV read	500 mm	Orientation	Transversal
FoV phase	100.0 %	Rotation	
Slice thickness	5.0 mm		0.00 deg
TR	154 ms	R >> L	500 mm
TE	97 ms	A >> P	500 mm
Averages	1	F >> H	5 mm
Concatenations	1	Physio	
Filter	None	1st Signal/Mode	None
Coil elements	T1	1	110110
		Inline	
ontrast		Subtract	Off
MTC	Off	Std-Dev-Sag	Off
Magn. preparation	None	Std-Dev-Cor	Off
Fat suppr.	Fat sat.	Std-Dev-Tra	Off
Avoraging mode	Long term	Std-Dev-Time	Off
Averaging mode Reconstruction	Magnitude	MIP-Sag	Off
	-	MIP-Cor	Off
Measurements	1 0 ms	MIP-Tra	Off
Delay in TR	0 ms	MIP-Time	Off
Multiple series	Off	Save original images	On
esolution			
Base resolution	64	Sequence	
Phase resolution	100 %	Introduction	Off
Phase partial Fourier	Off	Bandwidth	752 Hz/Px
Interpolation	Off	Free echo spacing	Off
		Echo spacing	1.4 ms
PAT mode	None		
Distortion Corr.	Off	EPI factor	64
Prescan Normalize	Off	RF pulse type	Normal
Raw filter	Off	Gradient mode	Fast
Elliptical filter	Off	•	
Hamming	Off		
Taililling	OII		
eometry			
Multi-slice mode	Interleaved		
Series	Interleaved		
Special sat.	None		
Table position	 Н		
Table position	0 mm		
	V 111111		
rabio position		1/4	

	\\USER\Feinberglab\Tanja\T	1\SpinEchoFieldMap_PA	
TA: 1:07 PAT: 2	2 Voxel size: 1.6×1.6×1.6 mm	Rel. SNR: 1.00 USER	R: cmrr_mbep2d_se
Properties		Series	Interleaved
Prio Recon	Off	Special sat.	None
Before measurement			
After measurement		Table position	Н
Load to viewer	On	Table position	0 mm
Inline movie	Off	Inline Composing	Off
Auto store images	On	System	
Load to stamp segments	Off	T1	On
Load images to graphic	Off	M2	On
segments		B4	On
Auto open inline display	Off	M3	On
Start measurement without	On	V32	Off
further preparation			
Wait for user to start	Off	Positioning mode	FIX
Start measurements	single	MSMA	S - C - T
ļ	9	Sagittal	R >> L
Routine		Coronal	A >> P
Slice group 1		Transversal	F >> H
Slices	85	Coil Combine Mode	Sum of Squares
Dist. factor	0 %	AutoAlign	
Position	L0.0 P13.2 H9.5	Auto Coil Select	Default
Orientation	T > C-15.6	Shim mode	Standard
Phase enc. dir.	A >> P	Adjust with body coil	Off
Rotation	0.00 deg	Confirm freq. adjustment	Off
Phase oversampling	0 %	Assume Silicone	Off
FoV read	208 mm	? Ref. amplitude 1H	0.000 V
FoV phase	100.0 %	Adjustment Tolerance	Auto
Slice thickness	1.60 mm	Adjust volume	Auto
TR	7130 ms	Position	L0.0 P13.2 H9.5
TE	60.0 ms	Orientation	T > C-15.6
Multi-band accel. factor	1	Rotation	0.00 deg
Filter	None	R >> L	208 mm
Coil elements	B4;M2,3;T1	A >> P	208 mm
Contrast		F >> H	136 mm
MTC	Off	l	100 11111
Magn. preparation	None	Physio	
Flip angle	90 deg	1st Signal/Mode	None
Refocus flip angle	180 deg	BOLD	
Fat suppr.	None	GLM Statistics	Off
Grad. rev. fat suppr.	Disabled	Dynamic t-maps	Off
Averaging mode	Long term	Starting ignore meas	0
Reconstruction	Magnitude	Ignore after transition	0
Measurements	3	Model transition states	On
Delay in TR	0 ms	Temp. highpass filter	On
Multiple series	Off	Threshold	4.00
		Paradigm size	20
Resolution		Meas[1]	Baseline
Base resolution	130	Meas[2]	Baseline
Phase resolution	100 %	Meas[3]	Baseline
Phase partial Fourier	7/8	Meas[4]	Baseline
Interpolation	Off	Meas[5]	Baseline
PAT mode	GRAPPA	Meas[6]	Baseline
Accel. factor PE	2	Meas[7]	Baseline
Ref. lines PE	66	Meas[8]	Baseline
Reference scan mode	GRE	Meas[9]	Baseline
		Meas[10]	Baseline
Distortion Corr.	Off	Meas[11]	Active
Prescan Normalize	Off	Meas[12]	Active
Raw filter	On	Meas[13]	Active
Elliptical filter	Off	Meas[14]	Active
Hamming	Off	Meas[15]	Active
Geometry		Meas[16]	Active
		Meas[17]	Active

Meas[17]

Multi-slice mode

Interleaved

Active

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Meas[18] Meas[19] Meas[20] Motion correction Spatial filter	Active Active Active Off Off
Sequence	
Introduction	Off

_	•	
	Introduction	Off
	Contrasts	1
	Bandwidth	1924 Hz/Px
	Free echo spacing	Off
	Echo spacing	0.64 ms
	EPI factor	130
	Gradient mode	Normal
	Excite pulse duration	4480 us
	Refocus pulse duration	10240 us
	Slice multiplier	1
	SENSE1 coil combine	Off
	Invert RO/PE polarity	On
	PF omits higher k-space	Off
	Force equal slice timing	Off
	FFT scale factor	1.00
	GRE iPAT ref. FA	12.0 deg
	Physio recording	Off
	Triggering scheme	Standard

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

Table of contents	

\\USER

Feinberglab
Tanja
T1
ep2d_se
SpinEchoFieldMap_PA