SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\AMRIT\Jen\4d_flow\localizer

TA: 0:13 PAT: Off Voxel size: 1.1×1.0×7.0 mm Rel. SNR: 1.00 SIEMENS: gre

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 2	
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 3	
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	250 mm
FoV phase	100.0 %
Slice thickness	7.0 mm
TR	8.6 ms
TE	4.00 ms
Averages	2
Concatenations	3
Filter	Prescan Normalize, Elliptical filter
Coil elements	HEA;HEP

Contrast

TD	0 ms
MTC	Off
Magn. preparation	None
Flip angle	20 deg
Fat suppr.	None
Water suppr.	None
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	256
Phase resolution	90 %
Phase partial Fourier	Off
Interpolation	On
PAT mode	None
Matrix Coil Mode	Auto (CP)
Image Filter	Off
Distortion Corr.	Off
Unfiltered images	Off
Prescan Normalize	On
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	On
Mode	Inplane

Geometry

Multi-slice mode	Sequential
Series	Interleaved
Saturation mode	Standard
Special sat.	None
Tim CT mode	Off

System

Body	Off
HEP	On
HEA	On
Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	---
Auto Coil Select	Default
Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	350 mm
A >> P	263 mm
F >> H	350 mm

Physio

1st Signal/Mode	None
Segments	1
Dark blood	Off
Resp. control	Off

Inline

Subtract	Off
Liver registration	Off
Std-Dev-Sag	Off

Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On
Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off

Sequence

Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Allowed
Contrasts	1
Bandwidth	320 Hz/Px
Flow comp.	No
Allowed delay	0 s
RF pulse type	Normal
Gradient mode	Normal
Excitation	Slice-sel.
RF spoiling	On

\\USER\AMRIT\Jen\4d_flow\fl_fq_mb1f1_gre3D_seg2_up

TA: 0:30 PAT: 2 Voxel size: 1.6×1.6×3.0 mm Rel. SNR: 1.00 USER: fl_fq_mb_gre_3D_seg

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	100 %
Position	L0.0 P0.0 H36.0
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	12
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
TR	74.85 ms
TE	5.85 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HEA;HEP

Contrast

Flip angle	12 deg
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	128
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Ref. lines 3D	12
Matrix Coil Mode	Auto (Triple)
Reference scan mode	Separate
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Sequential
Series	Ascending
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	350 mm
A >> P	263 mm
F >> H	350 mm

Physio

1st Signal/Mode	None
Segments	2

Angio

Flow mode	Single vel.
Encodings	3
Velocity enc.	90 cm/s
Direction 1	Through plane
Direction 2	A >> P
Direction 3	R >> L
Rephased images	On
Magnitude images	On
Magnitude sum	Off
Phase images	On
Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Sequence

Introduction	On
Dimension	3D

Elliptical scanning	Off
Asymmetric echo	Off
Contrasts	1
Bandwidth	260 Hz/Px
Flow comp.	No
RF pulse type	Normal
Gradient mode	Fast
RF spoiling	On
MB Number	1
FOV Shift	1

\\USER\AMRIT\Jen\4d_flow\fl_fq_mb1f1_gre3D_seg2_down

TA: 0:30 PAT: 2 Voxel size: 1.6×1.6×3.0 mm Rel. SNR: 1.00 USER: fl_fq_mb_gre_3D_seg

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	100 %
Position	L0.0 P0.0 F36.0
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	12
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
TR	74.85 ms
TE	5.85 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HEA;HEP

Contrast

Flip angle	12 deg
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	128
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Ref. lines 3D	12
Matrix Coil Mode	Auto (Triple)
Reference scan mode	Separate
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Sequential
Series	Ascending
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	350 mm
A >> P	263 mm
F >> H	350 mm

Physio

1st Signal/Mode	None
Segments	2

Angio

Flow mode	Single vel.
Encodings	3
Velocity enc.	90 cm/s
Direction 1	Through plane
Direction 2	A >> P
Direction 3	R >> L
Rephased images	On
Magnitude images	On
Magnitude sum	Off
Phase images	On
Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Sequence

Introduction	On
Dimension	3D

Elliptical scanning	Off
Asymmetric echo	Off
Contrasts	1
Bandwidth	260 Hz/Px
Flow comp.	No
RF pulse type	Normal
Gradient mode	Fast
RF spoiling	On
MB Number	1
FOV Shift	1

\\USER\AMRIT\Jen\4d_flow\fl_fq_mb2f2_gre_3D_seg2

TA: 1:33 PAT: 2 Voxel size: 1.6×1.6×3.0 mm Rel. SNR: 1.00 USER: fl_fq_mb_gre_3D_seg

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slab group 1	
Slabs	2
Dist. factor	100 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	12
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
TR	75.30 ms
TE	5.91 ms
Averages	1
Concatenations	2
Filter	None
Coil elements	HEA;HEP

Contrast

Flip angle	12 deg
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	128
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Ref. lines 3D	12
Matrix Coil Mode	Auto (Triple)
Reference scan mode	Separate
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Sequential
Series	Ascending
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	350 mm
A >> P	263 mm
F >> H	350 mm

Physio

1st Signal/Mode	Pulse/Trigger
Average cycle	No Signal ms
Captured cycle	-not set-
Acquisition window	240 ms
Trigger pulse	1
Trigger delay	0 ms
Segments	2
Phases	3

Angio

Flow mode	Single vel.
Encodings	3
Velocity enc.	90 cm/s
Direction 1	Through plane
Direction 2	A >> P
Direction 3	R >> L
Rephased images	On
Magnitude images	On
Magnitude sum	Off
Phase images	On
Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off

MIP-Time	Off
Save original images	On

Sequence

Introduction	On
Dimension	3D
Elliptical scanning	Off
Asymmetric echo	Off
Contrasts	1
Bandwidth	260 Hz/Px
Flow comp.	No
RF pulse type	Normal
Gradient mode	Fast
RF spoiling	On
MB Number	2
FOV Shift	2

\\USER\AMRIT\Jen\4d_flow\fl_fq_mb1f1_gre3D_seg2_up

TA: 0:30 PAT: 2 Voxel size: 1.6×1.6×3.0 mm Rel. SNR: 1.00 USER: fl_fq_mb_gre_3D_seg

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	100 %
Position	L0.0 P0.0 H32.4
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	12
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
TR	74.85 ms
TE	5.85 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HEA;HEP

Contrast

Flip angle	12 deg
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	128
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Ref. lines 3D	12
Matrix Coil Mode	Auto (Triple)
Reference scan mode	Separate
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Sequential
Series	Ascending
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	350 mm
A >> P	263 mm
F >> H	350 mm

Physio

1st Signal/Mode	None
Segments	2

Angio

Flow mode	Single vel.
Encodings	3
Velocity enc.	90 cm/s
Direction 1	Through plane
Direction 2	A >> P
Direction 3	R >> L
Rephased images	On
Magnitude images	On
Magnitude sum	Off
Phase images	On
Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Sequence

Introduction	On
Dimension	3D

Elliptical scanning	Off
Asymmetric echo	Off
Contrasts	1
Bandwidth	260 Hz/Px
Flow comp.	No
RF pulse type	Normal
Gradient mode	Fast
RF spoiling	On
MB Number	1
FOV Shift	1

\\USER\AMRIT\Jen\4d_flow\fl_fq_mb1f1_gre3D_seg2_down

TA: 0:30 PAT: 2 Voxel size: 1.6×1.6×3.0 mm Rel. SNR: 1.00 USER: fl_fq_mb_gre_3D_seg

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	100 %
Position	L0.0 P0.0 F32.4
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	12
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
TR	74.85 ms
TE	5.85 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HEA;HEP

Contrast

Flip angle	12 deg
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	128
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Ref. lines 3D	12
Matrix Coil Mode	Auto (Triple)
Reference scan mode	Separate
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Sequential
Series	Ascending
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	350 mm
A >> P	263 mm
F >> H	350 mm

Physio

1st Signal/Mode	None
Segments	2

Angio

Flow mode	Single vel.
Encodings	3
Velocity enc.	90 cm/s
Direction 1	Through plane
Direction 2	A >> P
Direction 3	R >> L
Rephased images	On
Magnitude images	On
Magnitude sum	Off
Phase images	On
Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Sequence

Introduction	On
Dimension	3D

Elliptical scanning	Off
Asymmetric echo	Off
Contrasts	1
Bandwidth	260 Hz/Px
Flow comp.	No
RF pulse type	Normal
Gradient mode	Fast
RF spoiling	On
MB Number	1
FOV Shift	1

\\USER\AMRIT\Jen\4d_flow\fl_fq_mb2f2_gre_3D_seg2

TA: 1:33 PAT: 2 Voxel size: 1.6×1.6×3.0 mm Rel. SNR: 1.00 USER: fl_fq_mb_gre_3D_seg

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slab group 1	
Slabs	2
Dist. factor	80 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	12
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
TR	75.30 ms
TE	5.91 ms
Averages	1
Concatenations	2
Filter	None
Coil elements	HEA;HEP

Contrast

Flip angle	12 deg
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	128
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Ref. lines 3D	12
Matrix Coil Mode	Auto (Triple)
Reference scan mode	Separate
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Sequential
Series	Ascending
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	350 mm
A >> P	263 mm
F >> H	350 mm

Physio

1st Signal/Mode	Pulse/Trigger
Average cycle	No Signal ms
Captured cycle	-not set-
Acquisition window	240 ms
Trigger pulse	1
Trigger delay	0 ms
Segments	2
Phases	3

Angio

Flow mode	Single vel.
Encodings	3
Velocity enc.	90 cm/s
Direction 1	Through plane
Direction 2	A >> P
Direction 3	R >> L
Rephased images	On
Magnitude images	On
Magnitude sum	Off
Phase images	On
Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off

MIP-Time	Off
Save original images	On

Sequence

Introduction	On
Dimension	3D
Elliptical scanning	Off
Asymmetric echo	Off
Contrasts	1
Bandwidth	260 Hz/Px
Flow comp.	No
RF pulse type	Normal
Gradient mode	Fast
RF spoiling	On
MB Number	2
FOV Shift	2