frontiers

# Design Factors for Isometric Robotic Knee Torque Using Braided Pneumatic Actuators

**Ben Bolen** [1,*], **Lindie Burgess** [1], **Connor Morrow** [1], **and Alex Hunt** [1]

[1] *Agile and Adaptive Robotics Laboratory, Portland State University, Department of Mechanical and Materials Engineering, Portland, OR, USA*

Correspondence*:
Ben Bolen
Portland State University
Department of Mechanical and Materials Engineering
PO Box 751
Portland, OR, 97207-0751, USA
bbolen83@gmail.com

## ABSTRACT

Artificial muscles such as braided pneumatic actuators (BPAs) reproduce the force and contraction behavior of skeletal muscle, making them attractive for the legs of musculoskeletal robots. However, the isometric torque a BPA delivers about a joint remains difficult to predict, because rigid-body torque models omit attachment compliance, effective-length offsets, and the contraction lost where the actuator wraps over the joint. In this work, we measured isometric knee torque on pinned and biomimetic knee test stands across actuator resting lengths, isolated error sources with a hybrid torque calculation, and identified correction terms by multiobjective optimization. Our analysis resulted in (1) a corrected torque model whose length offset, effective stiffness, and fictive-strain wrapping loss, expressed through the unloaded $620\,\mathrm{kPa}$ contraction, reduce held-out prediction error on both pinned-knee test families and transfer to the biomimetic knee without refitting; (2) a cross-configuration-consistent stiffness pair selected by a simultaneous design requirement; and (3) redesigned attachment routes that meet or exceed human isometric knee-torque targets with quantified margins, together with the first hardware torque measurements of the redesigned extensor route. The corrected model and redesign pipeline support first-iteration torque design of BPA-actuated robot joints.

**Keywords: Braided Pneumatic Actuator, Artificial Muscle, Isometric Torque, Moment Arm, System Identification, Actuator Routing, Biomimetic Robotics, Knee Joint**

## 1 INTRODUCTION

Musculoskeletal robots built on biological design principles are experimental platforms for motor-control questions that are impractical or unethical to study in human subjects, from spinal reflex loops to the coupling of posture and locomotion (Ijspeert, 2020; Goldsmith et al., 2021; Liu et al., 2018). The same principles also yield machines that copy the anatomy of their biological counterparts: human-mimetic humanoids and anthropomorphic arms with anatomically arranged musculotendon networks (Asano et al., 2017, 2019; Hitzmann et al., 2018), lower-limb robots driven by multifilament muscles (Kurumaya et al., 2016), and platforms that isolate the function of a single muscle such as the gluteus medius (Shin et al., 2018) or select actions during adaptive locomotion (Schilling et al., 2022). Each of these machines asks its

actuators to do what skeletal muscle does: pull as a lightweight tension member whose force depends on excitation and length, routed from origin to insertion across a joint. The joint-level quantity that matters is not force but torque, and torque depends as much on the routed path as on the actuator itself.

No actuator family reproduces muscle exactly, and each alternative trades something away. Geared electric motors dominate position control, but they place mass and reflected inertia at the joint and gain compliance only through added series elasticity (Veale and Xie, 2016). Hydraulics supply high force density at the price of pumps, valves, and rigid supply lines (Suzumori and Faudzi, 2018). Shape-memory alloys respond slowly, and dielectric elastomers deliver low force at impractical drive voltages (Madden et al., 2004). Braided pneumatic actuators (BPAs), also called McKibben artificial muscles, occupy the practical middle ground: a pressurized sleeved bladder that shortens and pulls, producing force that falls with contraction in a muscle-like way while remaining a compliant tension member. Air-muscle quadrupeds have demonstrated robust, adaptive gaits (Aschenbeck et al., 2006), and the Festo fluidic BPAs used in this work are commercially available in the sizes a human-scale leg requires.

For BPAs, the force side of the design problem is well characterized: we have published closed-form equations for the maximum force and the isometric force of $10\,\mathrm{mm}$ and $20\,\mathrm{mm}$ Festo actuators as functions of pressure, contraction, and resting length (Bolen et al., 2026). The torque side has received far less attention. Joint torque is the product of actuator force and the moment arm of the actuator's line of action about the joint axis, and the moment arm is a property of the attachment route that changes with joint angle. Biomechanics treats moment-arm determination as its own measurement problem, from lower-limb muscular anatomy (Brand et al., 1982) and three-dimensional musculotendon estimation (Menegaldo et al., 2004) to curved paths with moving attachment points (Suderman and Vasavada, 2012), because short moment arms amplify path error (Young et al., 2019). The robot hardware adds further sources of error: brackets and fixtures deform under load, artificial tendons and cabling stretch, and at deep flexion the actuator bends over the joint and loses usable contraction. None of these effects appears in a rigid-body torque model built from a characterized actuator and nominal geometry.

Torque modeling for braided actuators has consequently concentrated on the force law and on placement. Chou and Hannaford (1996) derived the static force relation of McKibben muscles from braid geometry and quantified its hysteresis, and later work refined static models and approximations for pneumatic and fluidic muscles (Tondu, 2022; Sárosi, 2012). For the actuator family used here, Martens and Boblan (2017) modeled static force from a geometric and material framework, and Hunt et al. (2017) modeled length effects. At the joint level, Bolen and Hunt (2019) determined artificial-muscle placement for biomimetic humanoid robot legs, and Morrow et al. (2020) optimized placements so that a bipedal robot matches human torque and range-of-motion capabilities. The present tests evaluate errors that remain when these force and routing models are applied to a physical joint. None of them represents fixture and tendon compliance in the load path, carries an effective-length offset between modeled and physical muscle length, or accounts for the contraction lost where the actuator wraps over the joint. To our knowledge, no BPA torque model has been validated across joint configurations, actuator resting lengths, and both antagonistic sides of a joint with a single consistent parameter set.

The present work treats these mispredictions as identifiable design factors, not residual noise. We measured isometric knee torque on two test stands: a robot leg with a pinned revolute knee, and the same leg with a biomimetic knee whose instantaneous center of rotation migrates with flexion like that of the human joint (Steele, 2018; Steele et al., 2017). Each knee was driven by antagonistic flexor and extensor BPAs of $10\,\mathrm{mm}$ and $20\,\mathrm{mm}$ diameter across a range of resting lengths, with and without artificial tendons.

A hybrid torque calculation, which combines measured muscle lengths and moment arms with modeled actuator force, isolated kinematic error from force error before any parameter was fitted.

Our analysis resulted in (1) a corrected torque model whose constant length offset ($\chi_0$), lumped attachment stiffness ($\chi_1$ and $\chi_2$), and fictive-strain wrapping loss ($\chi_3$) reduce held-out prediction error on both pinned-knee test families and carry to the biomimetic knee without refitting; (2) a two-stage multiobjective identification that holds one stiffness pair consistent across both knees and both actuators, validates it on held-out resting lengths and on the biomimetic knee, and selects the adopted set by an external design requirement; and (3) redesigned attachment routes, each driven by two $20\,\mathrm{mm}$ BPAs, that meet or exceed human isometric knee-torque targets from OpenSim musculoskeletal models (Seth et al., 2018; Delp et al., 1990) with a required $5\,\%$ margin at every angle, together with the first hardware torque measurements of the redesigned extensor route. Throughout, the wrapping loss is expressed through fictive strain: the excess contraction the force model must see beyond the actuator's physical contraction, normalized by the unloaded $620\,\mathrm{kPa}$ contraction $\varepsilon_{620}$ of the free actuator. Together, the corrected model and the redesign pipeline let a designer predict isometric joint torque, correct it against hardware, and route the actuators to a human-derived target on the first design iteration.

## 2 METHODS AND MATERIALS

### 2.1 Overview

Isometric knee torque is the product of the force an actuator develops and the moment arm of its line of action about the joint axis, and the force itself is set by the actuator's effective length through the force model. Each link in that chain is a design quantity: the attachment routing fixes the moment arm, the resting length and attachment compliance set the effective contraction, and wrapping over the joint removes usable length. We built two knee test stands, measured isometric torque across each joint's range of motion with $10\,\mathrm{mm}$ and $20\,\mathrm{mm}$ Festo braided pneumatic actuators (BPAs) in flexor and extensor configurations, and compared the measurements against the original rigid-body model, the improved actuator force characterization alone, and the corrected model developed here. A hybrid calculation isolated the error sources, a multiobjective identification produced the correction terms, and the corrected model then drove a redesign of the attachment routes against human isometric torque targets.

### 2.2 Actuator Force Model

The actuator force model is fully described in Bolen et al. (2026) and is reproduced here only as far as the torque model requires. The contraction $\varepsilon$ of a BPA is

$$\varepsilon = \frac{(l_0 - l_\mathrm{m})}{l_0} \tag{1}$$

where $l_0$ is the uninflated resting length measured from hose clamp to hose clamp and $l_\mathrm{m}$ is the current muscle length. The unloaded contraction at $620\,\mathrm{kPa}$ is

$$\varepsilon_{620} = \frac{(l_0 - l_{620})}{l_0} \tag{2}$$

where $l_{620}$ is the actuator length at $620\,\mathrm{kPa}$ with no external load, and the relative contraction is

$$\varepsilon^* = \frac{\varepsilon}{\varepsilon_{620}} \tag{3}$$

The pair $(\varepsilon_{620}, \varepsilon^*)$ is the strain basis of this work: every length-based correction introduced below is expressed in $\varepsilon_{620}$-normalized units. The normalized isometric force is

$$F^*(\varepsilon^*, P^*) = \begin{cases} c_0 \cdot \big(\exp(-c_1 \cdot \varepsilon^*) - 1\big) + P^* \cdot \exp\big(-c_2(\varepsilon^*)^2\big) & \text{where positive} \\ 0 & \text{otherwise,} \end{cases} \tag{4}$$

that is, the normalized force is the value of the bracketed expression where that expression is positive and zero elsewhere, where $P^*$ is pressure normalized by $620\,\mathrm{kPa}$ and the diameter-dependent coefficients $c_0$, $c_1$, and $c_2$ are tabulated in Bolen et al. (2026) (p. 10). The maximum force is

$$F_{\max}(l_0, P) = a_1\, P\, \arctan\big(a_2\, P\, (l_0 - 0.0075)\big) \tag{5}$$

with $l_0$ in meters, a fixed $7.5\,\mathrm{mm}$ end-contact offset, and coefficients tabulated in Bolen et al. (2026) (p. 7); at the $620\,\mathrm{kPa}$ system ceiling, $F_{\max}(l_0, 620\,\mathrm{kPa}) = F_{620}(l_0)$. The isometric force is the composition

$$F(\varepsilon^*, P^*, l_0) = F^*(\varepsilon^*, P^*) \cdot F_{620}(l_0) \tag{6}$$

## 2.3 Robot Leg Model and Knee Test Stands

The robot leg consists of a knee joint, femur, tibia, BPA assemblies, and artificial tendons (Figure 1**A–C**). The artificial bones were 3D printed in Onyx, carbon fiber, and high-strength high-temperature fiberglass, with machined links for the knee mechanism, and the tendons are Shimano bicycle brake cable. The leg geometry and dimensions derive from bone scans of a $6\,\mathrm{ft}$ $(1.83\,\mathrm{m})$ tall person (Bolen and Hunt, 2019; Seth et al., 2018; Delp et al., 1990, 2007).





**Figure 1.** **(A)** CAD solid model assembly of the revolute joint knee in the test stand. **(B)** 1-DoF pinned-joint robot knee in the test apparatus with important components labeled; the leg is shown with 30 degrees of flexion, i.e. $\theta_k = -30^\circ$. **(C)** Picture from isometric torque testing. The setup shows the revolute knee configured for a test of the extensor BPA, with the knee flexed at $\theta_k = -120^\circ$; note the compressed shape of the inflated BPA on the knee at this high degree of flexion.

Two knee joints were tested: a pure revolute joint, and a biomimetic joint whose four-bar linkage translates the instantaneous center of rotation (ICR) as the knee rotates (Steele, 2018; Steele et al., 2017, 2018), moving it posteriorly in $X$ and distally in $Y$ in the femur frame, except briefly proximally near

full flexion, much as the human knee does (Bolen and Hunt, 2019). Reference coordinate frames follow Wu and Cavanagh (1995); Wu et al. (2002), and Figure 2 shows the measured ICR migration. Both joints share the same rotation center in the femur frame at $\theta_k = 0°$, each was driven by an antagonistic flexor and extensor, and $\theta_k$ is negative in flexion throughout this work.





**Figure 2.** ICR for the biomimetic robot knee as a function of knee angle, measured from $\theta_1$ on the tibia. The plot is overlaid on a model of the knee at $\theta_k = 0°$ (solid line) and $-60°$ and $-120°$ (dotted dash lines). White circles represent measured points in $-15°$ increments from $-150°$ to $-15°$, in $-5°$ increments from $-15°$ to $0°$, and then at $2°$, $5°$, and $10°$.

Each stand was tested with $\varnothing 10\,\mathrm{mm}$ Festo BPAs in both configurations, and the biomimetic knee also with a $\varnothing 20\,\mathrm{mm}$ BPA. Each actuator was pinned at its origin on the femur, with the far end either pinned at the tibial insertion or attached there through an artificial tendon. Bolen and Hunt (2019); Morrow et al. (2020) give full details of the modeled robot and its torque calculation.

## 2.4 Equipment and Data Collection

A test jig was built to collect knee torque at set knee angles over the range of motion. The stand frame is made of 80/20 1515 series extruded aluminum. The femur is fixed to the frame while the knee rotates the tibia, with the joint axis co-axial with gravity to eliminate gravitational torque. A force sensor connects the tibia to a swing arm tied to a winch with 3/16-inch Kevlar rope (Quality Nylon Rope) routed over idler pulleys. At each measured $\theta_{\mathrm{k}}$ the winch was locked and the BPA was inflated and then deflated, keeping the measurement as isometric as possible. Because the sensor and its rope load path tie the tibia to the locked winch, the sensor reading is the tension that holds the tibia at the set angle. The measured knee torque at each angle is that tension multiplied by the perpendicular distance from the knee axis to the rope's line of action at its tibial attachment, the measured counterpart of the scalar reduction of Equation (10), with the attachment geometry and the sensor-line angle recorded at each test position.

Force was collected with one of two sensors: a MODERN STEP $300\,\mathrm{kg}$ digital crane scale, or a CALT DYLY-103 $100\,\mathrm{kg}$ S-shaped load cell used with an HX711 load-cell amplifier. Pressure was measured with a Freescale MPX5700 GP $5\,\mathrm{V}$ pressure sensor. Supply pressure was controlled by two regulators in series, a Parker 20R113GC ($0\,\psi$ to $120\,\psi$) and a Husky HDA72200, and air was delivered through a Festo VTUG-10 valve manifold. Pressure and load-cell data were sent to MATLAB through a Sparkfun BlackBoard C (Arduino Uno style) microcontroller. Lengths were measured with a tape measure, with a flexible string marking the axial length of the BPA when its curvature was sufficient, and angles with a Medigauge digital goniometer or MALENOO analog goniometers of $6\,\mathrm{in}$, $8\,\mathrm{in}$ and $12\,\mathrm{in}$ lengths.

## 2.5 Hybrid Torque Calculation

Torque about the knee is the moment of the actuator force about the ICR. Two vectors enter it: the position vector $\mathbf{d}$ from the ICR to the point of force application, and the force vector $\mathbf{F} = \|\mathbf{F}\|\hat{\mathbf{u}}$ along the line of action. The moment arm is the perpendicular distance from the joint axis to the line of action,

$$r = \|\mathbf{d} \times \hat{\mathbf{u}}\| \tag{7}$$

the same quantity computed from the vectors or measured with a tape measure. In the model it follows the projection method of Young et al. (2019): the scalar moment arm about the joint axis $\hat{\mathbf{z}}_k$ (the knee-frame $Z$ axis) is

$$r_k = \mathbf{p}_{proj,xy} \cdot \frac{\mathbf{p}_f \times \hat{\mathbf{z}}_k}{\|\mathbf{p}_f \times \hat{\mathbf{z}}_k\|} \tag{8}$$

where $\mathbf{p}_f$ is the three-dimensional free muscle segment vector, the segment whose endpoints ride on different bodies and whose in-plane component sets the line of action, and $\mathbf{p}_{proj,xy}$ is the position vector from the joint-axis origin to the point of force application, projected onto the plane normal to the joint axis. The ratio term is the in-plane unit normal to the projected line of action, so the product is the perpendicular distance of Equation (7), computed in the tibial body frame.

The hybrid torque calculation localizes disagreement between the model and the measurements. The classical moment is the cross product

$$\mathbf{m} = \mathbf{d} \times \mathbf{F} \tag{9}$$

and its scalar component about the joint axis, written $M_z$, reduces through the moment arm to the ordinary product

$$M_z = r\,F \tag{10}$$

The hybrid method replaces the model-predicted moment arm of Equation (8) and the model-predicted muscle length with measured counterparts: $r$ is measured directly as the perpendicular distance from the knee axis to the line of action, and $F$ follows from the measured $P$, $l_0$, $l_{620}$, and $l_{\mathrm{m}}$ through the force model of Section 2.2, which separates kinematic error from force error. This calculation is also referred to as the back-calculated torque; the two names denote the same quantity and are used interchangeably in this paper.

## 2.6 Improved Torque Model: Length Offset and Bracket Compliance

Measured muscle lengths were consistently shorter than predicted at all joint angles, and the shortfall persisted at $\varepsilon^* = 1$ with $F = 0$, indicating a constant rather than a load-dependent offset. The corrected muscle length is therefore

$$l_{\mathrm{m}} = l_{LMT} - 2\, l_{fitting} - \chi_0 \tag{11}$$


where $l_{LMT}$ is the musculotendon path length from origin to insertion, $l_{fitting}$ is the end-cap length ($25.4\,\mathrm{mm}$ per end), and $\chi_0$ is the constant length offset.

As the actuator develops force, the compliant attachments deform, changing the effective BPA length and, through it, the force. The combined compliance is modeled with Hooke's law in a local bracket frame at each attachment. The two-rotation frame is constructed from the vector joining the bracket reference point to the BPA attachment point, $\mathbf{p}_A$ at the femur-side muscle origin or $\mathbf{p}_B$ at the tibial insertion (Figure 3): a rotation through $\theta_z$ about the fixed space-frame axis $\hat{\mathbf{z}}_s$ aligns an intermediate axis with the sagittal-plane projection of the attachment point, and a rotation through $-\theta_y$ about the current $\hat{\mathbf{y}}$ axis, which points posteriorly in the primary cantilever bending direction, aligns the bracket axis $\hat{\mathbf{x}}_{br}$ with the full three-dimensional point vector. The orientation and homogeneous transform are

$$\mathbf{R}_{sb} = \mathbf{R}_z(\theta_z)\mathbf{R}_y(\theta_y)^T, \qquad \mathbf{T}_{sb} = \begin{bmatrix} \mathbf{R}_{sb} & \mathbf{p}_{br} \\ \mathbf{0}^T & 1 \end{bmatrix}. \tag{12}$$

Figure 4 shows the bracket frame for each test configuration.

**(A)** First rotation: $\hat{x}_1$ targets the sagittal projection




**(B)** Second rotation: $\hat{x}_{br}$ targets $\mathbf{p}_{A/B}$





**Figure 3.** Two-rotation construction of the bracket frame from the attachment point $\mathbf{p}_A$, the femur-side muscle origin, or $\mathbf{p}_B$, the tibial insertion. **(A)** Rotation through $\theta_z$ about the fixed $\hat{\mathbf{z}}_s$ axis aligns the intermediate axis $\hat{\mathbf{x}}_1$ with the point's sagittal-plane projection. **(B)** Rotation through $-\theta_y$ about the current $\hat{\mathbf{y}}_1 = \hat{\mathbf{y}}_{br}$ axis aligns the final bracket axis $\hat{\mathbf{x}}_{br}$ with the full point vector.





**Figure 4.** Bracket frame orientations for: **(A)** flexor configuration on the pinned knee, **(B)** extensor configuration on the pinned knee, **(C)** flexor and extensor configurations on the biomimetic knee with $\varnothing 10\,\mathrm{mm}$ BPAs, and **(D)** flexor configuration on the biomimetic knee with the $\varnothing 20\,\mathrm{mm}$ BPA.

In each local frame the bracket is modeled as three independent linear springs. For the pinned-flexor insertion bracket, the identified stiffness assignment is

$$\mathbf{K}_{br} = \begin{bmatrix} \chi_1 & 0 & 0 \\ 0 & \chi_2 & 0 \\ 0 & 0 & \chi_1 \end{bmatrix} \tag{13}$$

with $\chi_1$ assigned to $x_{br}$ and $z_{br}$, and $\chi_2$ to $y_{br}$. The stiffness pair is shared across configurations, but its axis assignment follows the hardware model: the pinned-flexor origin bracket uses $\mathrm{diag}(\chi_1, \chi_1, \chi_2)$, the pinned extensor and biomimetic flexor use $\mathrm{diag}(\chi_1, \chi_2, \chi_2)$, and the biomimetic extensor uses Equation (13). The flexor evaluator constructs both bracket frames with the first rotation alone; the pinned-extensor evaluator uses the two-rotation construction. A rotation about $y_{br}$ leaves a stiffness matrix unchanged only when its $x_{br}$ and $z_{br}$ stiffnesses are equal. Frame independence therefore applies to Equation (13), not to every assignment.

The local force vector is the force magnitude times its unit direction,

$$\mathbf{F}_{br} = F\hat{\mathbf{u}}, \qquad \hat{\mathbf{u}} = [u_x, u_y, u_z]^T. \tag{14}$$

Hooke's law relates this force to the bracket displacement vector:

$$\mathbf{F}_{br} = \mathbf{K}_{br}\boldsymbol{\eta}_{br}, \qquad \boldsymbol{\eta}_{br} = \mathbf{C}_{br}\mathbf{F}_{br}, \qquad \mathbf{C}_{br} = \mathbf{K}_{br}^{-1}. \tag{15}$$

For the insertion-bracket assignment of Equation (13), this gives

$$\boldsymbol{\eta}_{br} = F \begin{bmatrix} u_x/\chi_1 \\ u_y/\chi_2 \\ u_z/\chi_1 \end{bmatrix}. \tag{16}$$

The displacement need not be parallel to the force. Only its projection along the local line of action contributes to the scalar shortening used in the force balance (Figure 5):

$$\delta_{br} = \hat{\mathbf{u}}^T \boldsymbol{\eta}_{br}. \tag{17}$$

Substitution of Equation (15) gives

$$\delta_{br} = F\,\hat{\mathbf{u}}^T\,\mathbf{K}_{br}^{-1}\,\hat{\mathbf{u}} \tag{18}$$

so the bracket presents the scalar stiffness

$$k_{\hat{u},br} = \frac{F}{\delta_{br}} = \left(\hat{\mathbf{u}}^T\mathbf{K}_{br}^{-1}\hat{\mathbf{u}}\right)^{-1} \tag{19}$$

along the force path. An artificial tendon is modeled as an axial spring of stiffness $k_{ten} = nA_cE/L_{ten}$, with $n = 2$ parallel cable strands (six for the $\varnothing 20\,\mathrm{mm}$ assembly), cable cross-sectional area $A_c = 1.51 \times 10^{-6}\,\mathrm{m}^2$, and modulus $E = 193\,\mathrm{GPa}$. Bracket deflection and tendon stretch both shorten the effective BPA length under the same force, so their scalar compliances add in series:

$$k_{eq} = \left(\frac{1}{k_{\hat{u},br}} + \frac{1}{k_{ten}}\right)^{-1} \tag{20}$$

This reduction assumes quasi-static tensile loading, linear translational springs, and additive displacements along the force path. The direction is held fixed during the scalar solve, then updated with the displaced attachments; rotational compliance and force–moment coupling are not modeled. With a second compliant bracket, the same reduction gives

$$k_{eq}^{-1} = \sum_j \hat{\mathbf{u}}_j^T\mathbf{K}_{br,j}^{-1}\hat{\mathbf{u}}_j + k_{ten}^{-1}, \tag{21}$$

where the tendon term is omitted for a direct attachment. All unit directions and stiffness matrices in a projection are expressed in the same local frame.

The compliance displacement shortens the actuator, so the active length entering the force model is

$$l_m(\delta) = l_m^{(0)} - \delta \tag{22}$$

where $l_m^{(0)}$ is the initial length of Equation (11). Because the force depends on contraction, and the contraction on the compliance displacement, $\delta$ is the root of the implicit balance

$$F_{620}(l_0)\, F^*(\varepsilon^*(\delta), P^*) - k_{eq}\, \delta = 0 \tag{23}$$

solved by bracketed scalar root-finding between zero deflection and the deflection that would bring the actuator to its maximum usable contraction, with $\delta = 0$ returned when the residual shows no sign change across that interval. The solved force gives the vector bracket deflection, which updates the attachment location; the routed length, force direction, moment arm, and torque are recalculated at the deflected geometry. The stiffness terms are effective parameters for the attachment system as a whole, absorbing the printed brackets, the bone fixtures, and the winch cable, and they are identified from torque rather than measured on any component in isolation.

Figure 5 shows how both ends of the actuator path can be compliant: the insertion-side bracket together with a second, origin-side bracket. Each bracket $j$ develops a three-dimensional deflection $\boldsymbol{\eta}_{br,j}$ along its local frame basis, of which only the projection along the corresponding force direction $\hat{\mathbf{u}}_j$ changes the actuator length, and the two projections act in series with the tendon elongation:

$$\begin{aligned} \boldsymbol{\eta}_{br,j} &= \sum_{i=1}^{3} \eta_{i,j} \hat{\mathbf{e}}_i^{(j)}, \\ \delta_{br,j} &= \hat{\mathbf{u}}_j^T \boldsymbol{\eta}_{br,j}, \qquad j \in \{1, 2\}, \\ \delta_{\text{series}} &= \delta_{br,1} + \delta_{\text{tendon}} + \delta_{br,2}. \end{aligned} \tag{24}$$





**Figure 5.** Local bracket displacement $\boldsymbol{\eta}_{br,j}$ and its projection $\delta_{br,j}$ along the force direction $\hat{\mathbf{u}}_j$. The inset shows that an anisotropic bracket can deflect off the force axis. The two projected bracket displacements and the tendon elongation add to the actuator shortening. Displacements are exaggerated for clarity.

## 2.7 Fictive Strain and Wrapping Loss

As the knee flexes, the extensor BPA contacts the anterior joint surface and bends around it (Figure 6). The rigid routed-length calculation does not account for the resulting change in usable contraction. We represent this effect with fictive strain: an empirical increase in the contraction seen by the force model, beyond the actuator's geometric contraction. The correction is phenomenological; it does not resolve wall stresses or the deformation of the braided actuator. A wrapping-loss term $\chi_3$ scales the associated reduction in usable length:

$$\Delta l = \begin{cases} \chi_3 \, R \, |\theta_{\text{wrap}} - \theta_{\text{k}}| \, \zeta^2 & \theta_{\text{k}} \leqslant \theta_{\text{wrap}} \\ 0 & \theta_{\text{k}} \geqslant \theta_{\text{wrap}} \end{cases} \tag{25}$$

where $\Delta l$ is the change in usable length of the BPA, $R$ is the wrapping radius, $|\theta_{\text{wrap}} - \theta_{\text{k}}|$ is the angle through which the BPA is bent once contact begins, and $\theta_{\text{wrap}}$, the knee flexion angle at which the BPA starts to bend over the joint, is determined from solid modeling. The factor $\zeta$ is the additive complement to the relative contraction,

$$\zeta = 1 - \varepsilon^* \tag{26}$$

so the wrapping loss vanishes at full contraction and grows quadratically toward resting length. Combining Equations (26) and (25) with the offset of Equation (11) yields the fully corrected muscle length

$$l_{\text{m}} = l_{LMT} - 2\, l_{fitting} - \chi_0 - \Delta l \tag{27}$$





**Figure 6.** Sagittal schematic of an extensor BPA crossing a flexed robot knee, with its braided body following an anterior contact arc between the proximal attachment and distal tendon. Dashed radii bound the contact arc; $R$ is the BPA-path wrapping radius, and $\alpha = |\theta_{\text{wrap}} - \theta_k|$ is the bend angle after contact begins, expressed in radians in Equation (25). The drawing illustrates the local circular-wrap approximation, not the full multi-contact route. The fitted usable-length loss $\Delta l$ is not an arc-minus-chord length.

## 2.8 Identification Procedure

The correction terms were identified by global optimization against measured isometric torque. For a candidate correction vector $\boldsymbol{\chi}$, $(\chi_0, \chi_1, \chi_2)$ in the flexor stage and $(\chi_0, \chi_3)$ on the locked stiffness pair in the extensor refit, the predicted torque curve of test $i$ is interpolated to its $n_i$ measured knee angles $\theta_{ij}$, giving residuals $e_{ij} = M_{ij} - \hat{M}(\theta_{ij}; \boldsymbol{\chi})$ against the measured torque $M_{ij}$, and the three goodness-of-fit (GoF) components of that test are

$$\text{RMSE}_i = \sqrt{\frac{1}{n_i}\sum_j e_{ij}^2}, \qquad \text{FVU}_i = \frac{n_i - 1}{n_i} \cdot \frac{\sum_j e_{ij}^2}{\sum_j (M_{ij} - \bar{M}_i)^2}, \qquad e_i^{\max} = \max_j |e_{ij}|. \tag{28}$$

RMSE summarizes the overall error magnitude. FVU expresses the residual sum of squares relative to the sum of squared deviations of the measured torque from its mean $\bar{M}_i$, with the factor $(n_i - 1)/n_i$ carrying the

small-sample adjustment of the evaluators. The maximum absolute residual records the largest pointwise error. RMSE and the maximum residual have torque units, whereas FVU is dimensionless; smaller values indicate closer agreement for all three. The solver minimizes the objective vector

$$\mathrm{g}(\boldsymbol{\chi}) = \frac{1}{|\mathcal{T}|} \sum_{i \in \mathcal{T}} \begin{bmatrix} g_{1,i}(\boldsymbol{\chi}) / g_{1,i}^{(0)} \\ g_{2,i}(\boldsymbol{\chi}) / g_{2,i}^{(0)} \\ g_{3,i}(\boldsymbol{\chi}) / g_{3,i}^{(0)} \end{bmatrix}, \tag{29}$$

where $g_{1,i}$, $g_{2,i}$, and $g_{3,i}$ are the RMSE, FVU, and maximum absolute residual of Equation (28), the superscript $(0)$ marks the zero-correction baseline value of the same component, the same kinematics and force model with every correction term disabled, and $\mathcal{T}$ is the training fold; the fold mean of the baseline-normalized components makes the three objectives dimensionless and the training tests commensurate. The search carried no constraints beyond the parameter bounds of each stage, and the returned Pareto fronts were filtered to members scoring at or below the zero-correction baseline on every component of every test in that stage's identification pool.

Together, these measures assess agreement in torque magnitude and variation across the range of motion. Offsets, amplitude differences, and shifts along the knee-angle axis can affect several metrics at once, so the metrics are interpreted jointly rather than assigning each one to a single type of disagreement. Across tests with different torque ranges, the absolute and variance-normalized errors also weight the tests differently. Multiobjective optimization retains these comparisons alongside the largest local error, making the trade-offs visible instead of allowing a favorable average score to conceal a poor fit elsewhere. The resulting Pareto front contains alternative compromises among the objectives, from which the adopted terms are selected using held-out validation.

The solver was an elitist nondominated-sorting genetic algorithm (gamultiobj in MATLAB, RRID:SCR_001622) with adaptive-feasible mutation, scattered crossover at a fraction of 0.8, and a Pareto fraction of 0.35; each member of the returned Pareto front is scored on the reserved validation test, and the adopted solution is the member whose validation goodness of fit lies closest to its training-set fit. Where a family of members validates nearly equally well, as the flexor stiffness pairs do, the external redesign requirement of Section 4.3 selects among them.

The identification proceeded in two stages. The flexor stage fits the two-bracket model $(\chi_0, \chi_1, \chi_2)$ across the four pinned-knee flexor tests, with the $l_0 = 48.5\,\mathrm{cm}$ and $41.0\,\mathrm{cm}$ tests in training and the $l_0 = 45.7\,\mathrm{cm}$ and $40.6\,\mathrm{cm}$ tendon tests held out for validation. It evolved a population of 150 for up to 600 generations, terminating early once the relative change in the spread of the Pareto front fell below $4 \times 10^{-3}$. The extensor stage locks $(\chi_1, \chi_2)$ to a member of the flexor Pareto front, enforcing a single stiffness pair across configurations, and solves only the extensor-specific terms $(\chi_0, \chi_3)$ on a six-test pool of the pinned-extensor campaign, at resting lengths of $40.0\,\mathrm{cm}$ plain and with tendon, $43.6\,\mathrm{cm}$, $43.2\,\mathrm{cm}$ with tendon, $45.7\,\mathrm{cm}$, and $46.5\,\mathrm{cm}$, with the $43.6\,\mathrm{cm}$, $43.2\,\mathrm{cm}$-tendon, and $45.7\,\mathrm{cm}$ tests in training and the $40.0\,\mathrm{cm}$ plain, $40.0\,\mathrm{cm}$-tendon, and $46.5\,\mathrm{cm}$ tests held out for validation, and the extensor offset allowed negative values. The refit used a population of 50 for up to 150 generations at a spread tolerance of $1 \times 10^{-3}$. Both stages searched a normalized space, with length offsets in centimeters and stiffnesses on a base-ten logarithmic scale, so that the genetic search samples the decade-spanning stiffness bounds uniformly. The flexor bounds were $\chi_0 \in [0, 20]\,\mathrm{mm}$, $\chi_1 \in [3 \times 10^4, 10^6]\,\mathrm{N\,m^{-1}}$, and $\chi_2 \in [5 \times 10^3, 2 \times 10^4]\,\mathrm{N\,m^{-1}}$; the

extensor refit searched $\chi_0 \in [-20, +5]$ mm and $\chi_3 \in [0, 1]$ with the stiffness pair locked. The initial flexor population was drawn from $\chi_0 \in [5, 15]$ mm, $\chi_1 \in [5 \times 10^4, 5 \times 10^5]$, and $\chi_2 \in [7 \times 10^3, 1.5 \times 10^4]$ N m$^{-1}$. An initial single-bracket identification, produced with the same framework before the origin-side bracket was added, is reported alongside the results figures; the two-stage procedure superseded it, and its adopted values are given in the Discussion.

## 2.9 Route-Redesign Procedure

With the correction terms fixed, the attachment routes themselves become design variables, and their best values are the ones that make the redesigned routes meet or exceed human knee torque capability. The redesign was carried out on the robot-body model derived from the Gait2392 lower-extremity model (Delp et al., 2007; Seth et al., 2018). The human targets are maximum-isometric torque curves: OpenSim moment arms, the partial derivative of muscle-tendon length with respect to the knee coordinate, $r(\theta_k) = \partial l_{\mathrm{MT}}/\partial\theta_k$, multiplied by each muscle's maximum isometric force (Delp et al., 1990; Hoy et al., 1990; Yamaguchi and Zajac, 1989; Sherman et al., 2013; Millard et al., 2013; Thelen, 2003; Seth et al., 2011). The extensor is graded against vastus medialis and the flexor against biceps femoris short head; each target carries a required margin of 5 % above the human curve, so a route that merely touches the target at any angle is rejected.

The design variables are the BPA origin $\mathbf{p}_1$, the insertion $\mathbf{p}_{\mathrm{end}}$, the resting length, and the tendon length. Because the route prediction (contact activation, wrapping, and the compliance correction) is expensive to evaluate and only moderately smooth, the redesign minimizes a scalar cost assembled from weighted penalty terms:

$$J = w_w\, J_{\text{worst}} + w_t\, J_{\text{torque}} + \textstyle\sum_a J_a + J_{\text{geom}} + J_{\text{len}} \tag{30}$$

where $J_{\text{worst}}$ is the largest fraction by which robot torque falls short of the human target at any angle, weighted at $w_w = 10^5$, and $J_{\text{torque}}$ is the mean-squared shortfall fraction, weighted at $w_t = 10^3$, two orders of magnitude lower so that no single-angle shortfall can be traded for a better average. The $J_a$ are configuration-specific penalties (torque overshoot above the human curve, off-axis torque, wrap release), and $J_{\text{geom}}$ and $J_{\text{len}}$ regularize the endpoints and actuator lengths toward their seed values. Hard constraints, including bone clearance, wrap feasibility, actuator strain and contraction limits, and tendon geometry, reject failed candidates with a prohibitive penalty. The shared actuator plan is two $\varnothing 20$ mm BPAs per route under a $620$ kPa pressure ceiling, with a maximum usable contraction of 0.255 in the force model: the flexor model carries its two actuators on two separately solved routes, the second derived from the first by a mirrored distal attachment, with the two actuator forces summed, and the extensor model scales the force of a single shared route by the actuator count. Adjacent routes are required to maintain $25$ mm of clearance.

Each run began with a sanity evaluation of the seed design, then executed a surrogate-assisted global solver under per-configuration evaluation budgets of 3,000 evaluations for the extensor route and 1,000 for the flexor route, and finally refined the best feasible point with a derivative-free mesh-adaptive pattern search for 15,000 further evaluations. The surrogate budgets cap run time rather than define convergence; the pattern search stops on its budget or early, once its mesh collapses with the constraints satisfied.

The contact-release calculation is given in the Supplementary Material (Section 1.4). Three variants extend the search space. In the cam variant, the tibial-side route points trace a cam profile, so the routing geometry itself is shaped rather than only its endpoints. The pulley variants make the transmission ratio itself a design variable: a reeved tackle divides the force delivered at the insertion by its number of moving line parts $G$ and multiplies the insertion travel per unit of actuator contraction by the same factor, so the

tackle converts surplus actuator force into range of motion, and the surplus is supplied by mounting BPAs in parallel or mirrored pairs, the same two-actuator-per-route arrangement the knee designs of this paper already use. With $n$ actuators in parallel feeding a tackle of $G$ line parts, the transmission's force gain is $n/G$ and its travel gain is $G$; for example, three actuators through a two-part tackle deliver 1.5 times the force of a single actuator at twice the travel. The mono-pulley variant reeves the tendon over a single tackle and serves uniarticular muscles, which cross one joint, and the bi-pulley variant reeves one tackle per joint crossed and serves biarticular and multi-articular muscles; the bi-pulley machinery was applied as a batch over five anatomical muscle sites of the robot-body model's leg. Every pulley run of record operated at unity transmission, the batch sites with one BPA and one line part per crossing and the mono-pulley flexor rig with its two actuators over a single line part, so no tackle altered the force or travel anywhere; the tendon exit followed the moving-force-point routing, fixed to the proximal body with the tendon direction recomputed as the joint rotates, rather than a bowden cable housing.

## 3 RESULTS

With the actuator force model characterized at the actuator level, we applied the same model to predict isometric torque about the knee on both test stands (Section 2.3). In every configuration, the measured torque is compared against three predictions: the original rigid-body torque model, the improved BPA force characterization applied to the rigid kinematics, and the corrected torque model of Section 2.6 at its identified terms. The hybrid torque calculation (Section 2.5) replaces the modeled muscle length and moment arm with their measured counterparts, which isolates kinematic error from force error. The coefficients quoted in this section are the adopted two-stage identification of Section 2.8; the superseded initial single-bracket fit and its values are reported in the Discussion. The two campaigns reserve different validation tests: the flexor fit held out the $l_0 = 45.7\,\mathrm{cm}$ and $40.6\,\mathrm{cm}$ tendon actuators, while the extensor pool holds out the $l_0 = 40.0\,\mathrm{cm}$ plain, $40.0\,\mathrm{cm}$-tendon, and $46.5\,\mathrm{cm}$ actuators.

### 3.1 Flexor on Pin Joint

Figure 7 shows the measured flexor torque on the pinned knee next to the pre-optimization prediction series and the corrected prediction at the adopted terms for the four cross-validation tests. Both pre-optimization force models overpredicted the magnitude of the measured flexor torque (Table 1); each prediction is scored against all measured points at which it is defined. The Hunt equation attains the lower RMSE of the two on every test, but its score covers only the angles at which it is defined: near full extension, where its modeled strain falls below zero, it offers no prediction at all, while the improved BPA characterization remains defined over those same angles with its largest residuals. On these four tests the Hunt equation's valid domain coincides with the common domain on which every pre-optimization prediction is defined, so its per-model and common-domain scores are identical. The four tests are the cross-validation tests of the adopted two-stage identification. The back-calculated (hybrid) torque reproduced the measurements closely, which localizes the misprediction to the modeled kinematics, the muscle length $l_\mathrm{m}$ and the moment arm $r$, rather than to the force model.

The flexor identification of record is the two-bracket fit of Section 2.8 on the four cross-validation tests: the $l_0 = 48.5\,\mathrm{cm}$ and $41.0\,\mathrm{cm}$ tests in training and the $l_0 = 45.7\,\mathrm{cm}$ and $40.6\,\mathrm{cm}$-tendon tests held out. Its adopted terms are a length offset $\chi_0 = 3.94\,\mathrm{mm}$ and effective stiffnesses $\chi_1 = 39.98\,\mathrm{kN\,m^{-1}}$ and $\chi_2 = 14.73\,\mathrm{kN\,m^{-1}}$ (Table 4). At these terms the corrected model follows the measured torque across the range of motion, including the loss of torque near full extension that the rigid-body prediction does not produce (Figure 7). The fitting-set RMSE is $2.165\,\mathrm{N\,m}$ for the $l_0 = 48.5\,\mathrm{cm}$ test and $1.084\,\mathrm{N\,m}$ for the

$l_0 = 41.0\,\mathrm{cm}$ test, and the held-out validation tests give $1.368\,\mathrm{N\,m}$ at $45.7\,\mathrm{cm}$ and $1.611\,\mathrm{N\,m}$ at $40.6\,\mathrm{cm}$ with tendon (Table 2). An initial single-bracket identification, fitted on the $48.5\,\mathrm{cm}$ test with the $45.7\,\mathrm{cm}$ test as validation, preceded the origin-side bracket and was superseded by this fit; its values are reported in the Discussion.





**Figure 7.** Pre-optimization and optimized pinned-knee torque versus knee angle for the four cross-validation flexor tests. Knee angle is negative in flexion, and flexion torque is negative. Gold circles are the hybrid calculation; indigo filled circles are experimentally measured torque; the gray dashed line is the Hunt-equation rigid-body prediction; the light orange dash-dot line is the unoptimized prediction, the improved BPA force characterization applied to the same rigid kinematics; and the magenta line is the corrected prediction at the adopted terms. Panel titles give the nominal resting length; the actual resting lengths $l_0$ are **(A)** 48.5 cm, **(B)** 41.0 cm, **(C)** 45.7 cm, and **(D)** 40.6 cm with tendon. Panels **(A)** and **(B)** are the training tests and panels **(C)** and **(D)** the validation tests of the adopted cross-validation fold; the pre-optimization goodness-of-fit values are given in Table 1 and the corrected-model values in Table 2.

**Table 1.** Pre-optimization pinned-knee flexor torque-prediction goodness of fit per test: the Hunt-equation rigid-body prediction, the unoptimized prediction of the improved BPA characterization, and the back-calculated (hybrid) torque. The first two rows are the training tests and the last two the validation tests of the adopted cross-validation fold; RMSE and maximum absolute residual (Max.) are in $\mathrm{N\,m}$ with FVU the fraction of variance unexplained. Each prediction is scored against all measured points at which it is defined, so the Hunt equation is scored only away from full extension, where its modeled strain is negative and it offers no prediction.

| | | **Hunt equation** | | | **Improved char. alone** | | | **Back-calculated (hybrid)** | | |
|---|---|---|---|---|---|---|---|---|---|---|
| $l_0$ (cm) | $l_{\text{ten}}$ (mm) | **RMSE** | **FVU** | **Max.** | **RMSE** | **FVU** | **Max.** | **RMSE** | **FVU** | **Max.** |
| 48.5 | – | 5.223 | 0.9477 | 9.534 | 7.617 | 2.1278 | 25.531 | 1.635 | 0.0991 | 3.842 |
| 41.0 | – | 6.249 | 2.6768 | 10.510 | 9.694 | 2.1149 | 21.076 | 2.187 | 0.1184 | 4.243 |
| 45.7 | – | 6.048 | 1.1466 | 9.956 | 8.803 | 1.9380 | 26.264 | 1.246 | 0.0409 | 2.275 |
| 40.6 | 35 | 3.956 | 0.6301 | 7.344 | 6.373 | 0.5935 | 16.859 | 2.085 | 0.0755 | 3.073 |

**Table 2.** Optimized pinned-knee flexor torque-prediction goodness of fit at the adopted correction terms (Table 4): the improved BPA characterization alone (the zero-correction prediction, the same kinematics and force model with every correction term disabled) and the corrected model; training and validation tests follow the adopted cross-validation fold. RMSE and maximum absolute residual (Max.) are in $\mathrm{N\,m}$, and FVU denotes the fraction of variance unexplained.

| | | **Improved BPA char. alone** | | | **Corrected model** | | |
|---|---|---|---|---|---|---|---|
| $l_0$ (cm) | $l_{\text{ten}}$ (mm) | **RMSE** | **FVU** | **Max.** | **RMSE** | **FVU** | **Max.** |
| **Training** | | | | | | | |
| 48.5 | – | 7.617 | 2.1278 | 25.531 | 2.165 | 0.1718 | 5.467 |
| 41.0 | – | 9.694 | 2.1149 | 21.076 | 1.084 | 0.0277 | 1.766 |
| Mean | | 8.656 | 2.1213 | 23.303 | 1.624 | 0.0998 | 3.617 |
| **Validation** | | | | | | | |
| 45.7 | – | 8.803 | 1.9380 | 26.264 | 1.368 | 0.0468 | 3.614 |
| 40.6 | 35 | 6.373 | 0.5935 | 16.859 | 1.611 | 0.0427 | 5.891 |
| Mean | | 7.588 | 1.2657 | 21.562 | 1.490 | 0.0447 | 4.752 |

## 3.2 Extensor on Pin Joint

Two features of the pinned-extensor measurements shaped the extensor identification. First, the measured torque crossed zero at a more extended angle than the prediction, indicating a length offset opposite in sign to the flexor offset. Second, on the $l_0 = 45.7\,\mathrm{cm}$ test the hybrid calculation tracked the original prediction closely while the measured torque diverged from both, and the divergence grew with torque as the actuator wrapped over the joint.

The extensor identification of record locks the stiffness pair at the flexor-identified values and solves the two extensor-specific terms: $\chi_0 = -6.38\,\mathrm{mm}$ and $\chi_3 = 0.158$, so the extensor length offset is close

to the flexor offset in magnitude and opposite in sign (Table 4). The refit used the six-test identification pool of Section 2.8, with the $43.6\,\mathrm{cm}$, $43.2\,\mathrm{cm}$-tendon, and $45.7\,\mathrm{cm}$ tests in training and the $40.0\,\mathrm{cm}$ plain, $40.0\,\mathrm{cm}$-tendon, and $46.5\,\mathrm{cm}$ tests held out for validation; Figure 8 shows the six pool tests at the adopted terms. The initial extensor terms, $\chi_0 = -10.9\,\mathrm{mm}$ and $\chi_3 = 0.123$, belong to the superseded single-bracket identification and are reported in the Discussion.





**Figure 8.** Pinned-knee torque versus knee angle for the extensor BPAs: the six identification-pool tests at the adopted correction terms, with the extensor-specific terms $(\chi_0, \chi_3)$ on the shared stiffness pair (Table 4). Knee angle is negative in flexion, and extension torque is positive. Gold circles are the hybrid calculation; magenta-rimmed circles are experimentally measured torque; the dashed light orange line is the unoptimized prediction of the improved BPA characterization; the magenta line is the optimized prediction of the corrected model. Panel titles give the nominal resting length; the actual resting lengths $l_0$ are **(A)** 43.6 cm, **(B)** 43.2 cm with tendon, **(C)** 45.7 cm, **(D)** 40.0 cm, **(E)** 40.0 cm with tendon, and **(F)** 46.5 cm. Panels **(A)**–**(C)** are the training tests and panels **(D)**–**(F)** the held-out validation tests of the adopted cross-validation

**Table 3.** Pinned-knee extensor torque-prediction goodness of fit, at the adopted correction terms with the effective-stiffness pair locked to the flexor-identified values (Table 4): the improved BPA characterization alone (the zero-correction prediction of the text) and the corrected model. RMSE and maximum absolute residual (Max.) are in $\mathrm{N\,m}$, and FVU denotes the fraction of variance unexplained.

| | | **Improved BPA char. alone** | | | **Corrected model** | | |
|---|---|---|---|---|---|---|---|
| $l_0$ (cm) | $l_{\text{ten}}$ (mm) | **RMSE** | **FVU** | **Max.** | **RMSE** | **FVU** | **Max.** |
| **Training** | | | | | | | |
| 43.6 | – | 2.437 | 0.9618 | 4.284 | 0.978 | 0.1634 | 1.818 |
| 43.2 | 22 | 2.878 | 4.7304 | 4.817 | 1.525 | 1.3270 | 2.481 |
| 45.7 | – | 2.178 | 1.0586 | 4.974 | 0.490 | 0.0535 | 1.206 |
| Mean | | 2.498 | 2.2503 | 4.692 | 0.997 | 0.5146 | 1.835 |
| **Validation** | | | | | | | |
| 40.0 | – | 3.092 | 0.5638 | 4.842 | 1.244 | 0.1303 | 2.196 |
| 40.0 | 40 | 3.622 | 6.3919 | 7.845 | 1.350 | 0.8878 | 1.890 |
| 46.5 | – | 0.949 | 0.3152 | 2.272 | 0.699 | 0.1714 | 1.559 |
| Mean | | 2.554 | 2.4236 | 4.986 | 1.098 | 0.3965 | 1.882 |

## 3.3 Extensor on Biomimetic Knee

We tested the extensor configuration on the biomimetic knee with a $\varnothing 10\,\mathrm{mm}$ BPA at $l_0 = 51.8\,\mathrm{cm}$. The measured torque fell well below both the original prediction and the hybrid calculation (Figure 9). Applying the extensor terms identified on the pinned knee improved the prediction: at the adopted two-stage terms, the corrected model reached an RMSE of $1.1702\,\mathrm{N\,m}$, an FVU of 0.6681, and a maximum absolute residual of $2.2088\,\mathrm{N\,m}$, against $3.0322\,\mathrm{N\,m}$, 4.4856, and $5.5870\,\mathrm{N\,m}$ for the original rigid-body prediction. On this configuration the actuator path passes between the femoral condyles close to the instantaneous center of rotation, a geometry that is difficult to represent with the modeled path (Figure 9**A**).





**Figure 9.** Biomimetic 2-DoF robot leg in the extensor test configuration and its measured torque. **(A)** The leg installed in the extensor test configuration, with the BPA passing between the femoral condyles. **(B)** Biomimetic-knee torque versus knee angle for the $\varnothing 10\,\mathrm{mm}$, $l_0 = 51.8\,\mathrm{cm}$ extensor BPA; knee angle is negative in flexion, and extension torque is positive. Indigo filled circles are experimentally measured torque; gold circles are the hybrid calculation; the gray dashed line is the original rigid-body prediction; the magenta line is the optimized prediction of the corrected model using the extensor terms identified on the pinned knee.

## 3.4 Flexor on Biomimetic Knee

Figure 10 shows the biomimetic-knee flexor measurements with the flexor terms carried over unchanged from the pinned-knee identification. Panel **A** presents the $\varnothing 10\,\mathrm{mm}$ actuator at $l_0 = 41.5\,\mathrm{cm}$ with a $21\,\mathrm{mm}$ artificial tendon ($\varnothing 1.5\,\mathrm{mm}$ bicycle brake cable wrapped four times), pressurized to $604\,\mathrm{kPa}$; the corrected model reached an RMSE of $0.3007\,\mathrm{N\,m}$, an FVU of 0.0102, and a maximum absolute residual of $0.4647\,\mathrm{N\,m}$. Panel **B** presents the $\varnothing 20\,\mathrm{mm}$ actuator at $l_0 = 41.5\,\mathrm{cm}$ with a $14\,\mathrm{mm}$ artificial tendon at $620\,\mathrm{kPa}$, for which the corrected model reached an RMSE of $1.8024\,\mathrm{N\,m}$, an FVU of 0.0499, and a maximum absolute residual of $3.5678\,\mathrm{N\,m}$.

The torque of the single $\varnothing 20\,\mathrm{mm}$ flexor actuator of panel **B** did not exceed the magnitude of the maximum human isometric flexion torque (Delp et al., 1990; Seth et al., 2018) between $60^\circ$ and $100^\circ$ of knee flexion; within a $\pm 20\%$ band about the human mean, it still fell short of human-level torque between $78^\circ$ and $103^\circ$ of flexion, the shortfall extending slightly past the upper end of the no-tolerance comparison. Because operating at $620\,\mathrm{kPa}$ across the full desired range of motion would load the printed femur beyond its ultimate strength, we also recorded single points at the reduced pressures $560\,\mathrm{kPa}$, $421\,\mathrm{kPa}$, $325\,\mathrm{kPa}$ and $281\,\mathrm{kPa}$, which show that the desired range of motion is achievable at reduced pressure. This shortfall in flexion torque defines the target that the attachment-route redesign reported in the Discussion addresses.

The redesigned flexor route was then validated physically on the same stand with the two $\varnothing 20\,\mathrm{mm}$ actuators of the design of record at $l_0 = 56.2\,\mathrm{cm}$ (Figure 10**C**). The original rigid-body prediction of this design at $620\,\mathrm{kPa}$ reaches a peak torque magnitude of $190.3\,\mathrm{N\,m}$, while the corrected model, carrying the length offset, attachment compliance, and wrapping loss identified on the pinned knee, reaches $91.9\,\mathrm{N\,m}$.

Fifteen isometric points held from $-119^\circ$ to $5^\circ$ at hung loads of $11.1\,\mathrm{N}$ to $106\,\mathrm{N}$ and gauge pressures of $226.5\,\mathrm{kPa}$ to $638\,\mathrm{kPa}$, with the measured wrench carried to the moving knee center by the same adjoint transform as the identification data; the measured torque magnitude met or exceeded the human target at every point, with margins from $0.42\,\mathrm{N\,m}$ at $-105^\circ$ to $17.0\,\mathrm{N\,m}$ at $-14^\circ$.





**Figure 10.** Biomimetic-knee torque versus knee angle for the flexor BPAs, in a 2×2 layout with the legend in the fourth tile. Knee angle is negative in flexion, and flexion torque is negative. **(A)** 10 mm panel: ∅10 mm, $l_0 = 41.5$ cm actuator with a 21 mm artificial tendon at 604 kPa. **(B)** 20 mm panel: ∅20 mm, $l_0 = 41.5$ cm actuator with a 14 mm artificial tendon at 620 kPa. **(C)** 20 mm panel, the redesigned two-BPA flexor route (two ∅20 mm BPAs, $l_0 = 56.2$ cm) at 620 kPa, a different configuration from panel **(B)** despite the shared title: corrected prediction and measured points against the human target. Indigo filled circles are measured torque; the magenta line is the optimized prediction of the corrected model using the flexor terms identified on the pinned knee; the dashed light orange line is the unoptimized prediction of the improved

# 4 DISCUSSION

## 4.1 Sources of Error in Isometric Torque Generation

The improved BPA force characterization removes part of the torque-prediction error, but only part (Figure 7). The remainder traces to the test system itself. The isometric system is not rigid: the winch cable, the artificial tendons, and the attachment brackets all bend and stretch under load, and the correction family $\chi_0$, $\chi_1$, and $\chi_2$ was created to represent them. Positioning tolerances, imperfect path modeling, and variation between individual actuators contribute as well.

The modeling of extensor paths that wrap around the knee is an approximation with inherent error, and the test-stand geometry amplifies it. When the moment arm is short, small changes to it have a large effect on the resultant torque (Young et al., 2019), which is why moment-arm determination receives so much attention in musculoskeletal modeling (Brand et al., 1982; Menegaldo et al., 2004; Suderman and Vasavada, 2012). We chose to wrap the BPA around the joint rather than add a patella with a deterministically designed location and friction; the resulting path, which passes between the femoral condyles close to the migrating center of rotation of the biomimetic knee (Steele, 2018), resists any straight-line idealization (Figure 9**A**). At high flexion, the shorter actuators were also observed to stretch and compress (Figure 1**C**), and the same loading occasionally slipped the actuator off the anterior bolt head that holds the femur to the knee, displacing the path by up to $\pm 20\,\mathrm{mm}$ in the $Z$ direction and changing both the force vector and the torque.

The pinned extensor tests exposed one further error source, with a signature distinct from compliance. Figure 8 shows a deviation between measured and originally predicted torque that widens as the expected torque grows, yet the hybrid calculation matches or exceeds the prediction except at the highest flexion angles. That combination points to a loss of functional muscle length rather than excess compliance, and it is why the wrapping-loss term $\chi_3$ was created in Equation (25). Bent over the joint, the actuator behaves like a beam, and the wrapped arc stops contributing usable contraction, so the actuator behaves as if it were more contracted than it physically is; this is the fictive strain of Section 2.7, and Section 4.2 quantifies it.

Working from the simple pinned knee to the complex biomimetic one isolated each of these mechanisms in turn. The products of that progression, a predictor for bending and axial compliance, a constant length offset, and a measure of the usable length lost as the actuator bends around a joint, are what make first-iteration torque design tractable for BPA-driven joints.

## 4.2 Two-Stage Re-Identification and the Adopted Correction Terms

The Results quote the adopted two-stage terms. For completeness, the superseded initial single-bracket identification, fitted on the $48.5\,\mathrm{cm}$ flexor test with the $45.7\,\mathrm{cm}$ test as its validation, yielded a length offset $\chi_0 = 11.9\,\mathrm{mm}$ and effective stiffnesses $\chi_1 = 102\,\mathrm{kN\,m^{-1}}$ and $\chi_2 = 10.6\,\mathrm{kN\,m^{-1}}$, with extensor-specific terms $\chi_0 = -10.9\,\mathrm{mm}$ and $\chi_3 = 0.123$ on the locked pair. Re-identifying the correction terms over the full test sets, with the two-bracket flexor model and the locked extensor refit, yields the adopted values: $\chi_0 = 3.94\,\mathrm{mm}$, $\chi_1 = 3.998 \times 10^4\,\mathrm{N\,m^{-1}}$, and $\chi_2 = 1.473 \times 10^4\,\mathrm{N\,m^{-1}}$ for the flexor, and $\chi_0 = -6.38\,\mathrm{mm}$ with $\chi_3 = 0.158$ for the extensor on the same locked stiffness pair (Table 4).

**Table 4.** Adopted correction terms of the two-stage identification; the extensor refit locked the effective-stiffness pair $(\chi_1, \chi_2)$ to the flexor-identified values, so one stiffness pair is shared across both configurations, and a dash marks a term not applied in that configuration. The stiffness terms are effective system stiffnesses that lump the compliance of the brackets, fixtures, and winch cable, and $\chi_3$ is unitless. The wrapping factor $\chi_3$ was identified on the extensor configuration and is applied unchanged in the flexor route redesign, where a route's wrap is active; it was not fitted to the pinned-flexor identification, whose routes carry no wrap, so the flexor configuration takes a dash.

| **Configuration** | $\boldsymbol{\chi_0}$ (m) | $\boldsymbol{\chi_1}$ ($\mathrm{N\,m^{-1}}$) | $\boldsymbol{\chi_2}$ ($\mathrm{N\,m^{-1}}$) | $\boldsymbol{\chi_3}$ |
|---|---|---|---|---|
| Flexor configuration | $3.94 \times 10^{-3}$ | $3.998 \times 10^{4}$ | $1.473 \times 10^{4}$ | – |
| Extensor configuration | $-6.38 \times 10^{-3}$ | $3.998 \times 10^{4}$ | $1.473 \times 10^{4}$ | 0.158 |

Each term has a direct physical reading. The offsets $\chi_0$ correct the effective muscle length by a constant, and the flexor and extensor offsets carry opposite signs, as the initial identification already showed. The stiffness pair is an effective system stiffness rather than the beam stiffness of any single part: it lumps the two attachment brackets together with the fixtures and the winch cable, which is why one pair can serve both configurations once the extensor refit locks it. The wrapping factor $\chi_3$ scales the usable length that fictive strain removes, $\Delta l = \chi_3 \, R \, |\theta_{\text{wrap}} - \theta_{\text{k}}| \, \zeta^2$ with $\zeta = 1 - \varepsilon^*$ (Equations (25) and (26)). Because $\varepsilon^*$ is normalized by the unloaded $620\,\mathrm{kPa}$ contraction $\varepsilon_{620}$ of the free actuator, the loss is expressed in fictive-contraction units: it vanishes as the actuator approaches full contraction, where $\zeta \to 0$, and grows quadratically toward resting length, where the unbent actuator would otherwise produce its largest force. The wrapping radii are fixed by the route geometry rather than fitted: the pinned-extensor identification evaluates two contact arcs of wrapping radii $R = 12\,\mathrm{mm}$ and $R = 40\,\mathrm{mm}$, and the biomimetic configurations use a wrap radius of $25\,\mathrm{mm}$, so $\chi_3 = 0.158$ scales a physical arc of known size.

Tables 2 and 3 quantify what the corrections buy. On the pinned-flexor set, the corrected model reduces the training-mean RMSE, over the $l_0 = 48.5\,\mathrm{cm}$ and $41.0\,\mathrm{cm}$ training tests, from $8.66\,\mathrm{N\,m}$ to $1.62\,\mathrm{N\,m}$ under the zero-correction prediction, the same kinematics and force model with every correction term disabled, and the validation mean over the held-out $l_0 = 45.7\,\mathrm{cm}$ and $40.6\,\mathrm{cm}$ tendon tests improves from $7.59\,\mathrm{N\,m}$ to $1.49\,\mathrm{N\,m}$; it holds the FVU at or below 0.172 on all four tests. On the extensor pool, the corrected model improves the RMSE of every test, reducing the training mean, over the $43.6\,\mathrm{cm}$, $43.2\,\mathrm{cm}$-tendon, and $45.7\,\mathrm{cm}$ tests, from $2.50\,\mathrm{N\,m}$ to $1.00\,\mathrm{N\,m}$, and the validation mean, over the held-out $40.0\,\mathrm{cm}$ plain, $40.0\,\mathrm{cm}$-tendon, and $46.5\,\mathrm{cm}$ tests, from $2.55\,\mathrm{N\,m}$ to $1.10\,\mathrm{N\,m}$; the FVU stays below unity on every test except the $l_0 = 43.2\,\mathrm{cm}$ tendon test. The biomimetic extensor test, a configuration held out of every identification, improves under the adopted two-stage terms as well: the corrected prediction reduces the RMSE from $3.0322\,\mathrm{N\,m}$ to $1.1702\,\mathrm{N\,m}$ and the FVU from 4.4856 to 0.6681 against the zero-correction baseline (Figure 9). Evaluated at the initial single-bracket extensor terms instead, the corrected model reaches $1.5394\,\mathrm{N\,m}$ and FVU 1.1560, so the adopted terms fit this held-out configuration more closely than the configuration-matched initial terms do.

The solver assignment follows the problem. The identification retains the genetic multiobjective solver, an elitist nondominated-sorting genetic algorithm, because the Pareto front it returns is what the holdout-validation procedure selects from. The route redesign is served by the surrogate-plus-pattern-search pipeline of Section 2.9, a surrogate-assisted global search refined by a derivative-free mesh-adaptive pattern search, because a single scalar cost is optimized there and each route evaluation is expensive.

### 4.3 Route Redesign Meeting Human Isometric Torque Targets

The redesign addresses the single-actuator shortfall documented in the Results: one BPA per route does not reach the human flexor envelope through the mid-flexion range. Every redesigned configuration therefore carries two $\varnothing 20\,\mathrm{mm}$ BPAs per route, a $620\,\mathrm{kPa}$ pressure ceiling, and a maximum usable contraction of 0.255 in the force model, and every family was produced by the procedure of Section 2.9 under the adopted correction terms of Table 4. The targets are maximum-isometric torque profiles computed on the Gait2392 lower-extremity model (Delp et al., 2007; Seth et al., 2018), the vastus medialis for the extensor and the biceps femoris short head for the flexor, each enlarged by a required $5\,\%$ overshoot margin so that a route merely touching the target at any angle is rejected, and the redesign itself runs on the robot-body model derived from it. Knee angles are negative in flexion throughout this section, and Supplementary Figure S1 graphs the four redesigned route families against their human-scaled requirements.

The direct extensor route exceeds the human envelope in magnitude at every angle, with zero shortfall points. Across the design range, the predicted two-BPA torque spans $35.45\,\mathrm{N\,m}$ to $109.4\,\mathrm{N\,m}$ against a human-scaled requirement of $20.25\,\mathrm{N\,m}$ to $50.32\,\mathrm{N\,m}$, and the required overshoot is consumed as the thinnest margin exactly as designed, $+5.02$ % at the worst-case angle of $-20.2^{\circ}$. The route respects the wrap-release schedule of the route-elimination rule, releasing its six sequential contacts at knee angles of $-89.8^{\circ}$, $-55.7^{\circ}$, $-47.8^{\circ}$, $-20.2^{\circ}$, $6.1^{\circ}$, and $7.4^{\circ}$, and the optimized design places the femur-side origin at $(40.0, 137.0, 0)$ mm and the tibial insertion at $(34.4, -197.1, 0)$ mm, with a resting length of $827.7\,\mathrm{mm}$ and $36.8\,\mathrm{mm}$ of tendon.





**Figure 11.** Route geometry of the cam extensor design of record, drawn in the femur frame with schematic femur and tibia outlines. Panels **(A)**–**(D)** show four knee poses that track the contact-release cascade: **(A)** $\theta_k = -120°$, full flexion with all contacts active; **(B)** $\theta_k = -91°$, after the first condyle contact has released; **(C)** $\theta_k = -56°$, with one further release; and **(D)** $\theta_k = 10°$, full extension with the cam wrap released; knee angle is negative in flexion. Each panel draws the optimized route and its cam contacts over the femur-cylinder, femur-wall, condyle, and tibia clearance envelopes of the route-elimination rule, with the local bend-radius checks dashed; the square marker is the femur-side origin, the diamond marker is the tibial insertion; the fifth tile carries the legend and the sixth the contact-release schedule.

The extensor route was then redesigned once more by allowing the tibial-side route points to trace a cam profile, and this cam redesign is the current extensor design. The cam increases the torque margin relative to the direct route. On the final geometry, which folds the optimizer winner together with a verified adjustment of the distal route point and tendon length, the minimum torque margin is $+30.83$ % with respect to the required envelope and $+39.77$ % with respect to the plain human target; the unadjusted optimizer solution carried $+31.30$ % and $+42.34$ % at the same checks. The pressure required to meet the envelope spans $269\,\mathrm{kPa}$ to $507\,\mathrm{kPa}$ over the range of motion, comfortably above the $150\,\mathrm{kPa}$ regulated floor of the valve manifold that supplies the hardware, so the remaining pressure is a design resource. The geometric checks give the tibial wrap follows a bend radius of $96.9\,\mathrm{mm}$, the guiding route point clears the femoral-condyle envelope by $43.1\,\mathrm{mm}$ at $10^\circ$ of knee extension, and the moment arm spans $59.4\,\mathrm{mm}$ to $102.8\,\mathrm{mm}$ with a mean of $72.4\,\mathrm{mm}$ (Figure 11).

The flexor redesign uses two staggered BPA routes rather than a single shared route. The first actuator's route is carried directly as design variables, its origin on the femur and its insertion on the tibia, and the second actuator's endpoints are derived from it by an offset-preserving endpoint rule: the second insertion mirrors the first across the mid-sagittal plane, and the second origin is placed so that the second route spans the same origin-to-insertion offset as the first, which keeps the two routes from crossing. Each route is then rebuilt by the route solver, so the second carries its own wrap point and its own tibial and femoral-condyle clearance state rather than inheriting the first's, with condyle clearance enforced in the nonlinear constraints. In the design of record the first route clears the tibial envelope at every angle and never wraps, while the second rides its standoff only within about $5^\circ$ of full extension before the wrap releases. The design meets the flexor target with a minimum margin fraction of 0.0501 against the required 0.05. Its predicted knee-axis torque spans $-81.23\,\mathrm{N\,m}$ to $-3.864\,\mathrm{N\,m}$ against the human-scaled requirement of $-23.16\,\mathrm{N\,m}$ to $-2.461\,\mathrm{N\,m}$ (peak magnitude $23.16\,\mathrm{N\,m}$), again exceeding the envelope at every angle. The flexor prediction applies the adopted correction terms, including the wrapping factor identified on the extensor configuration: where a route's wrap is active, the actuator loses usable length through the same wrapping mechanism, and the extensor-identified $\chi_3$ was applied to it unchanged, a transfer evaluated in the redesigned flexor's isometric tests.

The bi-pulley machinery was then applied as a batch over five anatomical muscle sites of the robot-body model's leg, covering both halves of the pulley doctrine: three uniarticular sites that the mono-pulley variant serves, soleus and tibialis anterior at the ankle and biceps femoris short head at the knee, and two biarticular sites that the bi-pulley variant serves, medial gastrocnemius across the knee and ankle and rectus femoris across the hip and knee. Every run of record operated at unity transmission, one BPA and one tackle line part per crossing, with the tendon exit on the moving-force-point routing; the bowden-housing arrangement is implemented in the machinery but was never exercised. Because the point of a reeved tackle is to exchange actuator force for range of motion, each site was then re-solved in full at three doctrine configurations of the transmission: a parallel pair of BPAs through a two-part tackle (force gain 1, travel gain 2), the same pair through a four-part tackle (force gain 0.5, travel gain 4), and a four-actuator group through a four-part tackle (force gain 1, travel gain 4), all at supply pressures of $300\,\mathrm{kPa}$, $450\,\mathrm{kPa}$ and $620\,\mathrm{kPa}$. The actuator sizes are re-solved rather than carried over, because a route sized for direct drive runs its tendon slack under a multi-part tackle that feeds takeup faster than the span demands; the optimizer re-fits the actuator lengths to the transmission at every configuration. The resulting torque surfaces at the two-actuator, two-part configuration were evaluated over joint angle and supply pressure at each site's evaluation plane: soleus and tibialis anterior over the ankle, biceps femoris short head over the knee, medial gastrocnemius over the knee with the ankle pinned at zero, and rectus femoris over the hip of its recorded route; the margins against the human requirements were likewise evaluated over joint angle

along the transmission ladder at $620\,\mathrm{kPa}$. The human requirements are maximum-isometric torques of the stock Gait2392 model (Delp et al., 2007; Seth et al., 2018), computed as OpenSim moment arms times maximum isometric force ($3549\,\mathrm{N}$ soleus, $905\,\mathrm{N}$ tibialis anterior, $1558\,\mathrm{N}$ medial gastrocnemius, $1169\,\mathrm{N}$ rectus femoris, and $804\,\mathrm{N}$ biceps femoris short head), all five sites on the same OpenSim basis.

The re-solves show the exchange operating in the direction the doctrine intends: at matched net force, the meeting range against the human requirement grows with the tackle's travel gain. The tibialis-anterior route is the strongest site: the single-actuator baseline meets the human dorsiflexor requirement, which spans $24.60\,\mathrm{N\,m}$ to $39.35\,\mathrm{N\,m}$ over the ankle range, across $48.1^\circ$ of the $70^\circ$ ankle range ($46.7^\circ$ contiguous) at $620\,\mathrm{kPa}$ with a peak of $52.0\,\mathrm{N\,m}$, and its meeting range grows from zero at $300\,\mathrm{kPa}$ to $14.6^\circ$ at $450\,\mathrm{kPa}$ and $48.1^\circ$ at $620\,\mathrm{kPa}$; the parallel pair through the two-part tackle holds the peak at $53.0\,\mathrm{N\,m}$ and extends the meeting range to $56.9^\circ$ ($55.4^\circ$ contiguous), and the four-actuator group through the four-part tackle extends it further to $59.8^\circ$ ($58.3^\circ$ contiguous). At the knee, the biceps femoris short head is already served over its whole range by the direct two-actuator flexor redesign of Section 2.9, which exceeds its requirement at every angle of the full $130^\circ$ span at a minimum margin fraction of 0.0501; the transmission study at this site therefore starts from the batch's unity-transmission single-actuator run, which meets its OpenSim-graded requirement at no angle, with eight of twenty-five grid cells infeasible at every pressure. Against that single-actuator baseline, the matched-force pair removes every infeasible cell and meets over $21.7^\circ$ of the span at a peak of $34.1\,\mathrm{N\,m}$, and the four-actuator group reaches $40.6^\circ$; at this site the tackle's contribution is the multiplied insertion travel and the actuator-count scalability it buys, not meeting range that the direct design lacks. The medial gastrocnemius baseline meets over $21.7^\circ$ of the knee sweep with seven of twenty-five cells infeasible; the matched-force pair extends the meeting range to $65.0^\circ$, and at force gain 0.5, with the pair alone feeding the four-part tackle, the design meets over $105.6^\circ$ of the sweep at a peak of $118.0\,\mathrm{N\,m}$, while the four-actuator group at matched force meets over $119.2^\circ$ at a peak of $236.1\,\mathrm{N\,m}$. The rectus femoris route was recorded with its knee crossing rigid, so its margins are evaluated over the hip angle, where the baseline meets over $46.9^\circ$ at $620\,\mathrm{kPa}$ and the matched-force pair more than doubles the meeting range to $99.0^\circ$. The soleus site is the honest limit of the exchange: it never meets its human requirement at any transmission, a peak of $25.3\,\mathrm{N\,m}$ against a required magnitude of $134.08\,\mathrm{N\,m}$ to $171.35\,\mathrm{N\,m}$, a factor of 6.8 short, because its shortfall is in force, and a tackle that divides force cannot add it, only the actuator count can.

Physical validation of the redesigned extensor route is under way on the rebuilt cam route with the $20\,\mathrm{mm}$ actuator, and the first full angle sweep is complete. An initial three-point measurement near full extension on the earlier direct route measured knee torque of $1.163\,\mathrm{N\,m}$ at $10^\circ$, $7.849\,\mathrm{N\,m}$ at $11^\circ$, and $23.437\,\mathrm{N\,m}$ at $10^\circ$, against hung loads of $3.7\,\mathrm{N}$, $25.2\,\mathrm{N}$ and $76\,\mathrm{N}$ at gauge pressures of $388.5\,\mathrm{kPa}$, $388\,\mathrm{kPa}$ and $592\,\mathrm{kPa}$, with the measured wrench carried to the moving knee center by the same adjoint transform as the identification data. On the rebuilt cam route the sweep then recorded nine isometric data measurements from deep flexion to extension: $116\,\mathrm{N}$ at $-119^\circ$ and $256\,\mathrm{kPa}$, $140\,\mathrm{N}$ at $-86.5^\circ$ and $350\,\mathrm{kPa}$, $168\,\mathrm{N}$ at $-61^\circ$ and $425\,\mathrm{kPa}$, $181\,\mathrm{N}$ at $-43^\circ$ and $435\,\mathrm{kPa}$, $181.5\,\mathrm{N}$ at $-30^\circ$ and $420\,\mathrm{kPa}$, $180.2\,\mathrm{N}$ at $-26.2^\circ$ and $437\,\mathrm{kPa}$, $173.4\,\mathrm{N}$ at $-22^\circ$ and $435\,\mathrm{kPa}$, $126\,\mathrm{N}$ at $-1^\circ$ and $488\,\mathrm{kPa}$, and $100\,\mathrm{N}$ at $5^\circ$ and $480\,\mathrm{kPa}$, each with its load-cell angle recorded in the tibia frame. The post-test rig measurement of the reaction-point geometry is complete, and the adjoint reduction of the nine points follows in Figure 12. The corresponding flexor validation is reported with the biomimetic flexor results (Figure 10**C**).

The pressures alone already read against the design. The deep-flexion data measurement at $-119^\circ$ was carried at $256\,\mathrm{kPa}$, at the low end of the cam design's predicted $269\,\mathrm{kPa}$ to $507\,\mathrm{kPa}$ requirement band, and

the near-extension measurement at $-1^\circ$ at $488\,\mathrm{kPa}$ against the $507\,\mathrm{kPa}$ worst case, so the route carried loads across the full range without approaching the $620\,\mathrm{kPa}$ ceiling.

With the reaction-point constants entered, the nine data measurements reduce to knee-extension torques of $32.52\,\mathrm{N\,m}$ at $-119^\circ$, $38.91\,\mathrm{N\,m}$ at $-86.5^\circ$, $47.12\,\mathrm{N\,m}$ at $-61^\circ$, $50.50\,\mathrm{N\,m}$ at $-43^\circ$, $51.17\,\mathrm{N\,m}$ at $-30^\circ$, $50.50\,\mathrm{N\,m}$ at $-26.2^\circ$, $48.71\,\mathrm{N\,m}$ at $-22^\circ$, $35.90\,\mathrm{N\,m}$ at $-1^\circ$, and $29.02\,\mathrm{N\,m}$ at $5^\circ$, with the extensor tab's positive load-cell angles carried through the same adjoint transform as the identification data, the reaction force at the negated tab angle in the tibia frame (Figure 12). Against the corrected prediction of the cam design evaluated at the measured angles and pressures, the data measurements exceed the prediction at eight of the nine angles, by up to $13.75\,\mathrm{N\,m}$ at $-30^\circ$, and fall below it once, by $2.06\,\mathrm{N\,m}$ at $-86.5^\circ$. Against the maximum-isometric vasti target at the measured angles, all nine data measurements meet or exceed the human target, with margins from $0.24\,\mathrm{N\,m}$ at $-86.5^\circ$ to $3.52\,\mathrm{N\,m}$ at $-119^\circ$, at supply pressures of $256\,\mathrm{kPa}$ to $488\,\mathrm{kPa}$; the corrected $620\,\mathrm{kPa}$ prediction exceeds the target at every tested angle.





**Figure 12.** Physical validation of the redesigned extensor route: the cam design of record with two $\varnothing 20\,\mathrm{mm}$ BPAs on the biomimetic knee. Knee angle is negative in flexion, and extension torque is positive. The magenta line is the corrected prediction of the cam route at $620\,\mathrm{kPa}$, and the black dotted line is the maximum-isometric vasti target. Indigo filled circles are the nine measurements; gold circles are the corrected model evaluated at each measured angle and pressure, a pressure-specific force-model prediction, not the hybrid calculation, which would require the measured muscle length and moment arm that this test did not record. Marker area scales with gauge pressure.

The same pipeline was extended from the knee to all 27 muscle-actuators of the robot-body model, and we report that extension as a status rather than a completed result. The first pass reached 19 of the 27 actuators, and its chosen routes showed the penalty-based formulation to be the binding limitation: all but one violated the feasibility constraints, the exception being the biceps femoris long-head pair at a worst-angle margin of $+0.03$ %, indistinguishable from zero. The formulation was then restructured so that the task envelope, the joint-angle window each route must serve, acts as a hard feasible set rather than a

penalty. In the restructured pass, twelve actuators terminated when the route builder failed while assembling the redesigned path segments, an implementation failure of the optimization pipeline rather than a property of the routes, leaving fifteen finalized routes; eight were chosen from constraint-feasible configurations (semimembranosus, biceps femoris long head, tensor fasciae latae, soleus, lateral gastrocnemius, tibialis anterior, iliacus, and psoas), and six of those met the torque requirement at every angle, with worst-angle margins from $+0.25$ % (biceps femoris long head) to $+250.7$ % (tensor fasciae latae). All of these redesign solutions remain optimizer-stage results: the reference model contains no pulley entities, no via points or markers were added to it for these routes, and the redesigned geometry has not yet been propagated into the dynamic simulation models.

## 4.4 Whole-Leg Capability Against Human Torque Surfaces

The capability survey grades each actuator against the same-named single muscle of the Gait2392 model (Delp et al., 2007; Seth et al., 2018) on a $5 \times 5$ grid of two joint coordinates. The hip adduction domain is clipped by its flexion-dependent envelope. A cell meets the requirement when the BPA is taut and its torque has the human torque's sign and at least its magnitude. The coverage metric is the fraction of included grid cells meeting that condition; the aggregate is the unweighted mean of the per-actuator fractions, not a continuous-area integral or a whole-leg torque sum.

The BPA force model is evaluated at $620\,\mathrm{kPa}$ (Bolen et al., 2026). The comparison includes all 27 actuator sites (Supplementary Table S6). The baseline uses the master's-era attachment routing at unity transmission; its eight nominal $40\,\mathrm{mm}$ sites are evaluated with four parallel $20\,\mathrm{mm}$ BPAs, while the other sites use the survey's base configuration. The updated side selects the saved transmission-ladder winner at each site, with actuator length and tendon length re-solved per configuration. It retains the same underlying attachment routing and the original diameter specification, including $40\,\mathrm{mm}$ actuators at those eight sites. The comparison therefore combines transmission, actuator sizing, and the baseline diameter substitution; it does not isolate an effect of attachment placement or compare two all-$20\,\mathrm{mm}$ assemblies.

The mean grid coverage increases from $37.28\,\%$ to $65.19\,\%$, improving at 23 sites and tying at four (Table 5). A separate sign-aware torque-ratio score averages $\min(M_R/M_H, 1)$ over the graded cells, assigning zero to inactive cells. Its mean falls from $-1.93$ to $-3.63$. Negative ratios penalize torque in the wrong direction, so greater coverage can coexist with a worse mean ratio where severe wrong-sign predictions remain. These metrics support different conclusions and should be considered together.

For comparison, the average of the two axis-wise mean coverages is $38.39\,\%$ for the baseline and $66.17\,\%$ for the updated configurations. On a complete rectangular grid, averaging row and column coverages equals the overall cell fraction exactly; it cannot reduce the influence of edge poses there. The two summaries differ only on the four hip-envelope domains with excluded cells. The two metrics and the torque-ratio score are reported together, so their different weightings remain visible.

**Table 5.** Capability across 27 actuator sites against single-muscle human torque grids. Means are taken over sites. The baseline substitutes four parallel $20\,\mathrm{mm}$ BPAs at each nominal $40\,\mathrm{mm}$ site; the transmission-ladder winners retain the original diameter specifications. Negative torque-ratio scores reflect wrong-sign torque, not a torque deficit in newton meters.

| Configuration | Grid coverage (%) | Axis mean (%) | Torque-ratio score |
|---|---|---|---|
| Baseline | 37.28 | 38.39 | $-1.93$ |
| Updated | 65.19 | 66.17 | $-3.63$ |

## 4.5 Design Factors for Isometric Torque

Read as a design problem, isometric torque about a BPA-actuated joint factors into three quantities the designer controls: the moment arm $r$, the normalized contraction $\varepsilon^*$ at which the actuator operates, and the compliance of the path and its fixtures, carried by the $\chi$ terms. Each result in this paper maps onto one corner of that triangle. Short moment arms amplify every error, so stiffness must be budgeted where the arm is short (Young et al., 2019). For a given $l_0$, moment arm and contraction allowance trade against each other, and the cam redesign shows the trade can be favorable: it raises the moment arm from the direct route's $44.9\,\mathrm{mm}$ to $61.2\,\mathrm{mm}$ to a span of $59.4\,\mathrm{mm}$ to $102.8\,\mathrm{mm}$ while improving the torque margin. This increased moment arm comes at a cost to biomimicry. Pressure headroom, finally, is a design resource: the cam route meets its envelope between $269\,\mathrm{kPa}$ and $507\,\mathrm{kPa}$ against a $620\,\mathrm{kPa}$ ceiling.

These factors situate our work among musculoskeletal robot legs. Multifilament muscle routing has produced human-scale lower limbs (Kurumaya et al., 2016), anthropomorphic arms (Hitzmann et al., 2018), and full human-mimetic humanoids (Asano et al., 2019), and tendon-driven bodies have supported investigations from reflexive hopping (Liu et al., 2018) to human-like manipulation (Chen et al., 2019) and cognitive locomotion (Schilling et al., 2022). Linkage-based knees manage the moment-arm problem geometrically (Romero et al., 2020), reviews of dynamic multilegged robots treat actuation routing as a first-order design decision (He and Gao, 2020), and tunable-compliance actuators make fixture compliance a design element rather than an error source (Leibach, 2020). What these designs share with ours is the recognition that the attachment route, not the actuator alone, sets the deliverable torque. What the corrected model adds is a quantitative accounting of the compliance and wrapping losses along that route, so that a route can be designed against a torque target rather than checked after the fact. Morrow et al. (2020) optimized artificial-muscle placement on a bipedal robot with a rigid-body torque model, and the correction terms quantified here are the physics such rigid-body placements leave unmodeled.

The correction family also complements the BPA force-modeling literature. Static models of the actuator in isolation, from the classical McKibben analyses (Chou and Hannaford, 1996) through approximation formulas (Sárosi, 2012) to geometrically grounded models of the Festo muscle (Martens and Boblan, 2017; Tondu, 2022), describe the actuator itself; none of them carries attachment compliance, length offsets, or wrapping losses into a joint-torque prediction. The correction terms used here attach to the force characterization of Bolen et al. (2026) and travel with it across joint configurations under one consistent parameter set.

### 4.6 Limitations and Future Work

The correction terms are identified for this test stand, these knee configurations, and these actuator families. Other joints or actuator types should repeat the two-stage procedure rather than reuse the values. Transfer to other systems requires validation against their measured torque.

Isometric torque testing has yielded important design factors for roboticists seeking to design biomimetic humanoid robots. Isokinetic and isotonic tests would assess behavior outside the isometric conditions evaluated here.

## 5 CONCLUSION

Isometric knee-torque measurements on pinned and biomimetic test stands showed that actuator force characterization alone does not determine delivered joint torque. A hybrid calculation using measured muscle length and moment arm helped separate force-model error from routing and attachment effects. The corrected torque model represents a constant length offset, effective attachment compliance, and an empirical wrapping loss. A two-stage identification shares one stiffness pair across flexor and extensor configurations and reduces prediction error on the held-out tests.

Route redesign under the corrected model produces configurations with two $\varnothing 20\,\mathrm{mm}$ BPAs that meet the selected human-muscle torque targets over their design grids, with the required $5\,\%$ margin. The cam extensor increases predicted torque headroom while reducing the required operating pressure. Physical tests meet or exceed the respective human targets at all 15 measured flexor points and all nine measured extensor points. These measurements validate the tested isometric configurations; dynamic loading and transfer to other joints remain to be evaluated.

## RESOURCE IDENTIFICATION INITIATIVE

- MATLAB (RRID:SCR_001622)
- OpenSim (RRID:SCR_002683)
- Arduino UNO (RRID:SCR_017284)

## CONFLICT OF INTEREST STATEMENT

The authors declare no conflicts of interest. The funders had no role in the design of the study; in the collection, analysis, or interpretation of data; in the writing of the manuscript; or in the decision to publish the results.

## AUTHOR CONTRIBUTIONS

Conceptualization, B.B. and C.M.; Methodology, B.B., C.M., and A.H.; Software, B.B., C.M., and L.B.; Validation, B.B. and A.H.; Formal Analysis, B.B.; Investigation, B.B., C.M., and L.B.; Resources, B.B., L.B., and A.H.; Data Curation, B.B.; Writing – Original Draft Preparation, B.B.; Writing – Review & Editing, B.B., C.M., L.B., and A.H.; Visualization, B.B. and A.H.; Supervision, B.B. and A.H.; Project Administration, B.B. and A.H.; Funding Acquisition, A.H.

## FUNDING

Research for this article was funded by the Department of Mechanical and Materials Engineering at Portland State University, the National Science Foundation (NSF) grant for NeuroNex: Communication, Coordination, and Control in Systems (C3NS) 2015317 and NSF grant 1943483.

## ACKNOWLEDGMENTS

The authors would like to acknowledge the contribution of Alex Steele, who designed the initial biomimetic 4-bar knee linkage we used for the test. He gave us solid models, built a 3D prototype, and let us break said prototype. His prompt and thoughtful responses to our questions about the previous design was very much appreciated. The authors would also like to thank Jasmine Bradley for her help reworking the figures in the results section. As a visual designer, her contributions improved the aesthetic quality and accessibility of the images.

## SUPPLEMENTAL DATA

Supplemental Data includes Supplementary Tables S1 through S6 and Supplementary Figures S1 and S2.

## DATA AVAILABILITY STATEMENT

The data sets are available from the authors upon reasonable request.

## REFERENCES

Asano, Y., Okada, K., and Inaba, M. (2017). Design principles of a human mimetic humanoid: Humanoid platform to study human intelligence and internal body system. *Science Robotics* 2. doi:10.1126/scirobotics.aaq0899

Asano, Y., Okada, K., and Inaba, M. (2019). Musculoskeletal design, control, and application of human mimetic humanoid Kenshiro. *Bioinspiration & Biomimetics* 14, 036011. doi:10.1088/1748-3190/ab03fc

Aschenbeck, K., Kern, N., Bachmann, R., and Quinn, R. (2006). Design of a Quadruped Robot Driven by Air Muscles. In *The First IEEE/RAS-EMBS International Conference on Biomedical Robotics and Biomechatronics, 2006. BioRob 2006.* 875–880. doi:10.1109/BIOROB.2006.1639201

Bolen, B., Elzein, M., Pang, L., and Hunt, A. (2026). Isometric Force Characterization of Braided Pneumatic Actuators. Manuscript submitted for publication

Bolen, B. P. and Hunt, A. J. (2019). Determination of Artificial Muscle Placement for Biomimetic Humanoid Robot Legs. In *Biomimetic and Biohybrid Systems*, eds. U. Martinez-Hernandez, V. Vouloutsi, A. Mura, M. Mangan, M. Asada, T. J. Prescott, and P. F. Verschure (Cham: Springer International Publishing), vol. 11556. 15–26. doi:10.1007/978-3-030-24741-6_2

Brand, R. A., Crowninshield, R. D., Wittstock, C. E., Pedersen, D. R., Clark, C. R., and van Krieken, F. M. (1982). A Model of Lower Extremity Muscular Anatomy. *Journal of Biomechanical Engineering* 104, 304–310. doi:10.1115/1.3138363

Chen, J., Zhong, S., Kang, E., and Qiao, H. (2019). Realizing human-like manipulation with a musculoskeletal system and biologically inspired control scheme. *Neurocomputing* 339, 116–129. doi:10.1016/j.neucom.2018.12.069

Chou, C.-P. and Hannaford, B. (1996). Measurement and modeling of McKibben pneumatic artificial muscles. *IEEE Transactions on Robotics and Automation* 12, 90–102. doi:10.1109/70.481753

729 Delp, S., Loan, J., Hoy, M., Zajac, F., Topp, E., and Rosen, J. (1990). An interactive graphics-based model
730 of the lower extremity to study orthopaedic surgical procedures. *IEEE Transactions on Biomedical*
731 *Engineering* 37, 757–767. doi:10.1109/10.102791
732 Delp, S. L., Anderson, F. C., Arnold, A. S., Loan, P., Habib, A., John, C. T., et al. (2007). OpenSim:
733 Open-Source Software to Create and Analyze Dynamic Simulations of Movement. *IEEE Transactions*
734 *on Biomedical Engineering* 54, 1940–1950. doi:10.1109/TBME.2007.901024
735 Goldsmith, C. A., Quinn, R. D., and Szczecinski, N. S. (2021). Investigating the role of low level
736 reinforcement reflex loops in insect locomotion. *Bioinspiration & Biomimetics* 16, 065008. doi:10.1088/
737 1748-3190/ac28ea
738 He, J. and Gao, F. (2020). Mechanism, Actuation, Perception, and Control of Highly Dynamic Mul-
739 tilegged Robots: A Review. *Chinese Journal of Mechanical Engineering* 33, 79. doi:10.1186/
740 s10033-020-00485-9
741 Hitzmann, A., Masuda, H., Ikemoto, S., and Hosoda, K. (2018). Anthropomorphic musculoskeletal 10
742 degrees-of-freedom robot arm driven by pneumatic artificial muscles. *Advanced Robotics* 32, 865–878.
743 doi:10.1080/01691864.2018.1494040
744 Hoy, M., Zajac, F., and Gordon, M. (1990). A musculoskeletal model of the human lower extremity: the
745 effect of muscle, tendon, and moment arm on the moment-angle relationship of musculotendon actuators
746 at the hip, knee, and ankle. *Journal of Biomechanics* 23, 157–169
747 Hunt, A., Graber-Tilton, A., and Quinn, R. (2017). Modeling length effects of braided pneumatic
748 actuators. In *Volume 5A: 41st Mechanisms and Robotics Conference* (Cleveland, OH: ASME), vol.
749 58172, V05AT08A008. doi:10.1115/detc2017-67458
750 Ijspeert, A. J. (2020). Amphibious and Sprawling Locomotion: From Biology to Robotics and
751 Back. *Annual Review of Control, Robotics, and Autonomous Systems* 3, 173–193. doi:10.1146/
752 annurev-control-091919-095731. _eprint: https://doi.org/10.1146/annurev-control-091919-095731
753 Kurumaya, S., Suzumori, K., Nabae, H., and Wakimoto, S. (2016). Musculoskeletal lower-limb robot
754 driven by multifilament muscles. *ROBOMECH Journal* 3, 18. doi:10.1186/s40648-016-0061-3
755 Leibach, R. (2020). *Development of a Tunable Compliance Energy Return Actuator*. Ph.D. thesis, Case
756 Western Reserve University
757 Liu, X., Rosendo, A., Ikemoto, S., Shimizu, M., and Hosoda, K. (2018). Robotic investigation on effect of
758 stretch reflex and crossed inhibitory response on bipedal hopping. *Journal of the Royal Society Interface*
759 15, 20180024. doi:10.1098/rsif.2018.0024. Place: London WOS:000428576200018
760 Madden, J., Vandesteeg, N., Anquetil, P., Madden, P., Takshi, A., Pytel, R., et al. (2004). Artificial muscle
761 technology: physical principles and naval prospects. *IEEE Journal of Oceanic Engineering* 29, 706–728.
762 doi:10.1109/JOE.2004.833135. Conference Name: IEEE Journal of Oceanic Engineering
763 Martens, M. and Boblan, I. (2017). Modeling the Static Force of a Festo Pneumatic Muscle Actuator: A
764 New Approach and a Comparison to Existing Models. *Actuators* 6, 33. doi:10.3390/act6040033
765 Menegaldo, L. L., de Toledo Fleury, A., and Weber, H. I. (2004). Moment arms and musculotendon
766 lengths estimation for a three-dimensional lower-limb model. *Journal of Biomechanics* 37, 1447–1453.
767 doi:10.1016/j.jbiomech.2003.12.017
768 Millard, M., Uchida, T., Seth, A., and Delp, S. L. (2013). Flexing Computational Muscle: Modeling
769 and Simulation of Musculotendon Dynamics. *Journal of Biomechanical Engineering* 135, 021005.
770 doi:10.1115/1.4023390
771 Morrow, C., Bolen, B., and Hunt, A. (2020). Optimization of Artificial Muscle Placements for a Humanoid
772 Bipedal Robot. In *Biomimetic and Biohybrid Systems* (Springer, Cham), vol. 12413 of *Lecture Notes in*
773 *Computer Science*, 257–269. doi:10.1007/978-3-030-64313-3_25

Romero, N., Rua Ramirez, E., and Moreno, G. (2020). FOUR-BAR MECHANISM'S ROTATION CENTER DETERMINATION USING NATURAL COORDINATES. *Journal of Engineering and Applied Sciences* 15

Schilling, M., Paskarbeit, J., Ritter, H., Schneider, A., and Cruse, H. (2022). From Adaptive Locomotion to Predictive Action Selection – Cognitive Control for a Six-Legged Walker. *IEEE Transactions on Robotics* 38, 666–682. doi:10.1109/TRO.2021.3106832. Conference Name: IEEE Transactions on Robotics

Seth, A., Hicks, J. L., Uchida, T. K., Habib, A., Dembia, C. L., Dunne, J. J., et al. (2018). OpenSim: Simulating musculoskeletal dynamics and neuromuscular control to study human and animal movement. *PLOS Computational Biology* 14, e1006223. doi:10.1371/journal.pcbi.1006223

[Dataset] Seth, A., Sherman, M., Reinbolt, J. A., and Delp, S. L. (2011). OpenSim: a musculoskeletal modeling and simulation framework for in silico investigations and exchange.

Sherman, M. A., Seth, A., and Delp, S. L. (2013). What is a Moment Arm? Calculating Muscle Effectiveness in Biomechanical Models Using Generalized Coordinates. In *Volume 7B: 9th International Conference on Multibody Systems, Nonlinear Dynamics, and Control* (Portland, Oregon, USA: American Society of Mechanical Engineers), V07BT10A052. doi:10.1115/DETC2013-13633

Shin, H., Ikemoto, S., and Hosoda, K. (2018). Constructive understanding and reproduction of functions of gluteus medius by using a musculoskeletal walking robot. *Advanced Robotics* 32, 202–214. doi:10.1080/01691864.2018.1434015

Steele, A. G. (2018). *Biomimetic Design and Construction of a Bipedal Walking Robot*. Thesis, Portland State University, Portland, Oregon, USA

Steele, A. G., Hunt, A., and Etoundi, A. C. (2017). Development of a Bio-inspired Knee Joint Mechanism for a Bipedal Robot. In *Biomimetic and Biohybrid Systems* (Springer, Cham), Lecture Notes in Computer Science, 418–427. doi:10.1007/978-3-319-63537-8_35

Steele, A. G., Hunt, A., and Etoundi, A. C. (2018). Biomimetic Knee Design to Improve Joint Torque and Life for Bipedal Robotics. In *Towards Autonomous Robotic Systems*, eds. M. Giuliani, T. Assaf, and M. E. Giannaccini (Cham: Springer International Publishing), 91–102

Suderman, B. L. and Vasavada, A. N. (2012). Moving muscle points provide accurate curved muscle paths in a model of the cervical spine. *Journal of Biomechanics* 45, 400–404. doi:10.1016/j.jbiomech.2011.10.014

Suzumori, K. and Faudzi, A. A. (2018). Trends in hydraulic actuators and components in legged and tough robots: a review. *Advanced Robotics* 32, 458–476. doi:10.1080/01691864.2018.1455606. _eprint: https://doi.org/10.1080/01691864.2018.1455606

Sárosi, J. (2012). New approximation algorithm for the force of fluidic muscles. In *Applied Computational Intelligence and Informatics (SACI), 2012 7th IEEE International Symposium on* (IEEE), 229–233

Thelen, D. G. (2003). Adjustment of Muscle Mechanics Model Parameters to Simulate Dynamic Contractions in Older Adults. *Journal of Biomechanical Engineering* 125, 70–77. doi:10.1115/1.1531112

Tondu, B. (2022). Static Modeling of Pneumatic Artificial Muscle Actuator. In *2022 IEEE 5th International Conference on Soft Robotics (RoboSoft)*. 364–369. doi:10.1109/RoboSoft54090.2022.9762146

Veale, A. J. and Xie, S. Q. (2016). Towards compliant and wearable robotic orthoses: A review of current and emerging actuator technologies. *Medical Engineering & Physics* 38, 317–325. doi:10.1016/j.medengphy.2016.01.010

Wu, G. and Cavanagh, P. (1995). FOR STANDARDIZATION IN THE REPORTING OF KINEMATIC DATA. *Journal Biomechanics* 28, 1257–1261. doi:0021-9290(95)00017-8

819 Wu, G., Siegler, S., Allard, P., Kirtley, C., Leardini, A., Rosenbaum, D., et al. (2002). ISB recommendation
820 on definitions of joint coordinate system of various joints for the reporting of human joint motion—part
821 I: ankle, hip, and spine. *Journal of Biomechanics* 35, 543–548. doi:10.1016/S0021-9290(01)00222-6
822 Yamaguchi, G. T. and Zajac, F. E. (1989). A planar model of the knee joint to characterize the knee
823 extensor mechanism. *Journal of Biomechanics* 22, 1–10. doi:10.1016/0021-9290(89)90179-6
824 Young, F., Rode, C., Hunt, A., and Quinn, R. (2019). Analyzing Moment Arm Profiles in a Full-Muscle
825 Rat Hindlimb Model. *Biomimetics* 4, 10. doi:10.3390/biomimetics4010010