



Air Power Company

Block-9/4GF, Pocket-7, Sector-82,

G.B Nagar Noida 201 304. INDIA

GST No : 09BAZPK5978D2ZT

E mail : airpowercompanynoida15@gmail.com

Calibration Certificate for Oxygen Analyzer

Ref- M/s DYNAMIC TRANSMISSION LTD.

02.07.2026

PLOT NO 119-120, ROHTAK

DESCRIPTION OF GOODS	: Oxygen Gas Analyzer : MVS OXYGEN ANALZER
NAME PLATE PARTICULARS	: Model No-MVS, 220VAC
QUANTITY	: 01 Set.
RANGE	: 00.00% TO 21.00% Oxygen
RESOLUTION	: 99.99%
OPERATING TEMPERATURE	: 0-50 degree C
TYPE OF DISPLAY	: Digital
RECORDER OUTPUT	: Yes, 4-20 mA DC for >10.0% Oxygen
ALARM RELAY OUTPUT	: Yes, 0-10.0% Oxygen over 10 Turns of Set Alarm Dial.
TECHNICAL PARAMETERS	: 1. Flow Circuit checked for flow passage, & Supply ± 12 VDC for Amplifier, Alarm PCB and 5V DC for DPM. 2. Checked + 12V DC for Alarm Power Side.
CALIBRATION	: Loaded Sensor in Analyzer MVS-O2 Passed IOLAR High Purity N ₂ (400ml/min outside setting) through Cell. The reading shows 00.00% O ₂ . Closed Nitrogen supply and adjusted By Span/Cal Dial. Adjusted reading to 20.9%. Changed gas IOLAR Nitrogen and after flushing to get reading show 0.00%
ALARM	: SPAN DIAL READING: 1.00% Checked visual display and relay cut out value. Relay gets actuated as output valve greater than set valve. Found O.K
OUTPUT	: Checked 4-20mA Output over 0-10.0% O ₂
CALIBRATION DONE ON	: July 02, 2026
CALIBRATION DUE ON	: October 02, 2026



For Air Power Company