



CERTIFICATE OF ANALYSIS

1. PRODUCT INFORMATION

Product Name	Helium, Compressed
Grade	Research (RES) — Grade 6.0
Purity	≥ 99.9999% (6N)
Lot Number	HE-2026-001
Cylinder Type	50L Seamless Steel Cylinder
Working Pressure	20 MPa (2,900 psig @ 70°F)
Gas Volume	~354 SCF (~10 m³)
Valve Connection	CGA-580
Date of Analysis	03/17/2026
Shelf Life	2 Years from date of analysis

2. ANALYTICAL TEST RESULTS

All concentrations in ppm by volume (v/v). Analysis performed by Gas Chromatography (GC) per ASTM E260.

Impurities	Specification (Max)	Result
Hydrogen (H ₂)	≤ 0.1 ppm	0.0325 ppm
Oxygen + Argon (O ₂ + Ar)	≤ 0.1 ppm	0.0445 ppm
Nitrogen (N ₂)	≤ 0.5 ppm	0.0557 ppm
Methane (CH ₄)	≤ 0.1 ppm	0.0007 ppm
Carbon Dioxide (CO ₂)	≤ 0.1 ppm	0.0421 ppm
Carbon Monoxide (CO)	≤ 0.1 ppm	0.0152 ppm
Total Impurities (Σ)	≤ 1.0 ppm	0.1907 ppm

3. ANALYSIS EQUIPMENT

Parameters	Instrument
H ₂ , O ₂ +Ar, N ₂ , CH ₄ , CO ₂ , CO	Gas Chromatograph (GC)

4. CERTIFICATION

This certificate is provided by Airocean Gases. The above results confirm that Lot No. HE-2026-001 meets all specifications for Research (RES) Grade 6.0 Helium per CGA G-9.1.

Certified by:

Jiahao Guo

Printed Name / Title

Signature:

Signature / Company Stamp

DISCLAIMER: This Certificate of Analysis is issued solely for Lot No. HE-2026-001. Airocean Gases is not responsible for results obtained from improper storage or handling. Refer to the current Safety Data Sheet (SDS) for safe handling procedures. This document shall not be reproduced except in full.