



# CERTIFICATE OF ANALYSIS

## 1. PRODUCT INFORMATION

|                  |                               |
|------------------|-------------------------------|
| Product Name     | Helium, Compressed            |
| Grade            | Research (RES) — Grade 6.0    |
| Purity           | ≥ 99.9999% (6N)               |
| Lot Number       | HE-2026-001                   |
| Cylinder Type    | 50L Seamless Steel Cylinder   |
| Working Pressure | 20 MPa (2,900 psig @ 70°F)    |
| Gas Volume       | ~354 SCF (~10 m³)             |
| Valve Connection | CGA-580                       |
| Date of Analysis | 03/17/2026                    |
| Shelf Life       | 2 Years from date of analysis |

## 2. ANALYTICAL TEST RESULTS

All concentrations in ppm by volume (v/v). Analysis performed by Gas Chromatography (GC) per ASTM E260.

| Impurities                           | Specification (Max) | Result     |
|--------------------------------------|---------------------|------------|
| Hydrogen (H <sub>2</sub> )           | ≤ 0.1 ppm           | 0.0325 ppm |
| Oxygen + Argon (O <sub>2</sub> + Ar) | ≤ 0.1 ppm           | 0.0445 ppm |
| Nitrogen (N <sub>2</sub> )           | ≤ 0.5 ppm           | 0.0557 ppm |
| Methane (CH <sub>4</sub> )           | ≤ 0.1 ppm           | 0.0007 ppm |
| Carbon Dioxide (CO <sub>2</sub> )    | ≤ 0.1 ppm           | 0.0421 ppm |
| Carbon Monoxide (CO)                 | ≤ 0.1 ppm           | 0.0152 ppm |
| Total Impurities (Σ)                 | ≤ 1.0 ppm           | 0.1907 ppm |

## 3. ANALYSIS EQUIPMENT

| Parameters                                                                                   | Instrument             |
|----------------------------------------------------------------------------------------------|------------------------|
| H <sub>2</sub> , O <sub>2</sub> +Ar, N <sub>2</sub> , CH <sub>4</sub> , CO <sub>2</sub> , CO | Gas Chromatograph (GC) |

## 4. CERTIFICATION

This certificate is provided by Airocean Gases. The above results confirm that Lot No. HE-2026-001 meets all specifications for Ultra High Purity (UHP) Grade 6.0 Helium per CGA G-9.1.

Certified by:

Jiahao Guo

Printed Name / Title

Signature:

Signature / Company Stamp

**DISCLAIMER:** This Certificate of Analysis is issued solely for Lot No. HE-2026-001. Airocean Gases is not responsible for results obtained from improper storage or handling. Refer to the current Safety Data Sheet (SDS) for safe handling procedures. This document shall not be reproduced except in full.