# Simulation and Generative Models for Robust Wearable ECG: Physics, Multiphysics, PINNs and Diffusion, Measured on Real Sensor Noise

Akila Lourdes Miriyala Francis
Long Island University, Brooklyn NY github.com/AKilalours/guardian-drive

Technical report, November 2026

## Abstract

Wearable physiological sensing is limited by scarce labels and by motion artifacts that clinical training data do not contain. We study whether simulation and generative modeling can scale training beyond the available real data, and we measure the answer on real data with real recorded sensor noise. Using single-lead PTB-XL ECG and MIT-BIH NSTDB noise under a leakage-free protocol, we build (i) a differentiable ECG dynamical model calibrated to 3,907 real beats (median $R^2$=0.95); (ii) an electro-mechanical electrode–skin multiphysics model (contact mechanics + electro-quasi-static finite volume, verified to $2 \times 10^{-4}$ against an analytic solution); (iii) a physics-based sensor/artifact model; (iv) an inverse PINN; and (v) a conditional 1-D diffusion model. Sensor-model augmentation raises AUROC under 0 dB real noise from 0.654 to 0.818, exceeding augmentation with real recorded noise (0.807). We diagnose a sampler failure that saturated 40% of diffusion samples and trace it to the terminal signal-to-noise ratio of the cosine schedule. The fix needs no retraining: saturation drops to 0.5%, real-vs-synthetic discriminator accuracy falls from 0.96 to 0.63, and a classifier trained only on generated ECG reaches 0.829 AUROC (real-trained: 0.801). Generated data combined with sensor augmentation gives the best result: 0.838 clean and 0.827 AUROC at 0 dB. We also report where methods fail: the PINN beats filtering only at low SNR, and multiphysics-derived sensor priors did not improve the downstream model.

## 1 Problem and protocol

**Task.** NORM vs. abnormal (MI/STTC/CD/HYP) from one 10-s lead (lead II, 100 Hz, a wearable-like view of PTB-XL [1]); 5,137 records with the official folds: 1–8 train (3,960), 9 validation, 10 test (592). **Noise.** MIT-BIH NSTDB [2] electrode-motion, muscle and baseline-wander recordings. Each recording is split in time: the first half is the only noise any training run may use; the second half only corrupts the test set. **Low-data regime.** A stratified 10% of the training records; the simulator bank, diffusion models and classifier in this regime see only these records. All downstream numbers are mean±std over 3 seeds, with model selection on the clean validation fold.

## 2 Methods

**Differentiable physics simulator.** The McSharry ECG model [3] drives a cardiac phase $\theta(t)$ with an RR process (Mayer + respiratory components) and integrates $\dot{z} = -\sum_i a_i \Delta\theta_i e^{-\Delta\theta_i^2/2b_i^2} - (z - z_0)$. Because the $z$-equation is linear, we integrate it *exactly* for zero-order-hold forcing with an FFT convolution (checked against RK4). This makes the simulator batched and differentiable. We calibrate the 16 parameters of every real median beat in parallel by gradient descent through the simulator, with *variable projection*: amplitudes and offset enter linearly and are solved by ridge least squares, which removes a local minimum in which inverted T-waves collapse instead of flipping (Fig. 1).

**Multiphysics electrode model.** Contact mechanics (asperity contact for dry electrodes, elastic gel and stratum-corneum compression, partial lift-off) sets a thin-layer interface admittance, coupled to a cell-centred finite-volume solve of $\nabla \cdot (\sigma^* \nabla \phi) = 0$ with complex conductivity in layered tissue (Fig. 2). Verification: analytic thin-strip error $2 \times 10^{-4}$, reciprocity $10^{-10}$, grid-convergence spread $7 \times 10^{-5}$.

**Sensor model.** Poisson road/body events drive a damped seat/torso response $d(t)$, which produces an AC-coupled skin-stretch potential, impedance-modulated gain $Z_{in}/(Z_{in} + Z_e(t))$, impedance-mismatch mains pickup (60 Hz aliased to 40 Hz), EMG bursts, electrode lift-off to the rail and ADC effects. It is FFT-based and batched, so it runs inside the training loop.

**Inverse PINN.** A Fourier-feature MLP $s(t)$, a low-order baseline $w(t)$ and learnable ODE parameters minimise $\|y - s - w\|^2 + \lambda \|\dot{s} - F(\theta; \phi) + s\|^2$, with $\dot{s}$ by autograd at stochastic collocation points. Segments are solved in parallel with batched (einsum) weights, warm-started from a hard-constraint differentiable-simulation fit.

**Conditional diffusion.** A 1-D U-Net with FiLM conditioning on diffusion time, diagnosis and heart rate, and self-attention at the bottleneck. Cosine-schedule DDPM [4], EMA, DDIM sampling [5], classifier-free guidance [6]. For the low-data regime we pre-train on 8,000

Table 1: Downstream AUROC under real held-out sensor noise (3 seeds). †original diffusion sampler.

| Training data (100% real) | clean | 6 dB | 0 dB |
|---|---|---|---|
| real only | .812±.009 | .762±.010 | .654±.009 |
| + real NSTDB noise (ref.) | .823±.009 | .820±.008 | .807±.010 |
| + **sensor model** | .833±.004 | .829±.004 | .818±.007 |
| + simulator data | .830±.008 | .808±.015 | .743±.021 |
| + diffusion data† | .835±.005 | .814±.014 | .756±.019 |
| + diffusion + sensor† | .836±.003 | .831±.002 | .827±.007 |
| + **diffusion + sensor (fixed)** | **.838**±.001 | **.834**±.003 | **.827**±.003 |
| *10% real* | | | |
| real only | .834±.002 | .794±.011 | .705±.030 |
| + sensor model | .809±.003 | .803±.006 | .790±.007 |
| + sim-pretrained diff. + sensor | .806±.008 | .804±.007 | .787±.005 |

Table 2: Generative fidelity vs. held-out real test data. FED: Fréchet distance on classifier embeddings; Disc.: real-vs-synthetic discriminator accuracy (0.5 ideal); TSTR: train on synthetic, test on real.

| Source | FED↓ | LSD dB↓ | Disc.↓ | TSTR↑ |
|---|---|---|---|---|
| real (train split) | 4.4 | 0.20 | 0.50 | 0.802 |
| calibrated simulator | 6.6 | 8.80 | 0.95 | 0.808 |
| *original sampler* | | | | |
| diffusion, full real | 72.9 | 5.44 | 0.96 | 0.819 |
| diffusion, 10% real | 107.3 | 1.80 | 0.93 | 0.812 |
| diffusion, sim-pre. + 10% | 43.2 | 3.19 | 0.86 | 0.808 |
| *fixed sampler (leading timesteps, selected on validation)* | | | | |
| diffusion, full real | 1.3 | 0.34 | 0.63 | 0.829 |
| diffusion, sim-pre. + 10% | 2.0 | 2.34 | 0.71 | 0.825 |

calibrated-simulator records and fine-tune on the 10% real subset.

# 3 Results

**Robustness (Table 1).** The physics-based sensor model is the largest single effect. Under 0 dB real noise it adds +0.164 AUROC with full data, more than augmenting with the real recorded noise itself, and in the 10% regime it adds +0.085. Diffusion data improves clean and moderate-noise accuracy. The combination is best on every test condition. In the 10% regime, augmentation trades ∼0.025 clean AUROC for robustness.

**Generative fidelity (Table 2).** All synthetic sources carry label information: TSTR AUROC matches or exceeds training on real data (0.801). The calibrated simulator matches the embedding statistics closely (FED 6.6) but not the spectrum (LSD 8.8 dB). With 10% of the data and the original sampler, simulation pre-training cut the diffusion model's FED 2.5× (107→43) and lowered discriminator accuracy (0.93→0.86). Heart-rate conditioning lands within 9.6 bpm of the request. **Sampler diagnosis.** About 40% of the original samples were pinned to the ±12 clip rail. Guidance was the obvious suspect, but the failure persisted at $w$=0 (39%). The cause is the sampler: DDIM started at $t$=$T$−1, where the cosine schedule gives $\bar{\alpha} \approx 2.4 \times 10^{-9}$, so the $\hat{x}_0$ estimate amplifies $\epsilon$-errors $\sim 2 \times 10^4$. Starting at the standard leading timestep ($\bar{\alpha} \approx 9 \times 10^{-4}$) with per-sample dynamic thresholding [7], selected on the validation fold, reduces saturation to 0.5% and validation FED from 49.4 to 0.84, with no retraining (Fig. 3). On test, discriminator accuracy falls from 0.96 to 0.63 and TSTR rises to 0.829. With the fixed sampler, diffusion data plus sensor augmentation reach 0.838±.001 clean and 0.827±.003 at 0 dB. The pre-training comparison above used the original sampler and was not repeated, so it is indicative only.

**PINN.** On simulated data with known parameters (6 dB sensor-model noise), the PINN recovers P/T-wave strength within 4–9% and timing within 0.03 rad, and raises output SNR from 9.0 to 11.7 dB. QRS parameters are poorly identifiable at 100 Hz (R strength error 35%). On real ECG with real noise, the physics prior wins at 0 dB (7.7 vs. 6.3 dB for band-pass). At 6 and 12 dB band-pass wins (13.9 / 21.0 dB vs. 12.2 / 16.3 dB), because a five-event single-template ODE cannot represent all real morphology.

**Multiphysics.** The model predicts 21.5 kΩ (wet) vs. 443 kΩ (dry) two-electrode impedance at 10 Hz, near-zero pressure sensitivity for gel but $\kappa \approx 0.4$ for dry electrodes, and lift-off sensitivity $\kappa \approx 0.6$ for both. These fall partly outside the sensor model's hand-set priors. Widening the priors to the multiphysics range, however, did *not* improve the classifier (0.801±.013 vs. 0.818±.007 at 0 dB). The augmentation benefit is robust to the exact priors.

# 4 Limitations

PTB-XL is resting clinical ECG, and NSTDB noise was recorded on stationary subjects. The multiphysics model is 2-D, uses nominal literature tissue properties and quasi-static contact mechanics, and is written in Python/SciPy rather than COMSOL. Results use one dataset and one task.

# 5 Reproducibility

The script `scripts/run_sensing_gen.sh` downloads the data and reproduces every number (about 4–5 h on an Apple M-series laptop). Results are written to `outputs/sensing_gen/`, and the repository has 17 unit tests and CI.

# References


[1] P. Wagner et al., PTB-XL, a large publicly available ECG dataset, *Scientific Data*, 2020.

[2] G. Moody, W. Muldrow, R. Mark, A noise stress test for arrhythmia detectors, *Computers in Cardiology*, 1984.

[3] P. McSharry et al., A dynamical model for generating synthetic ECG signals, *IEEE TBME*, 2003.

[4] J. Ho, A. Jain, P. Abbeel, Denoising diffusion probabilistic models, NeurIPS 2020.





Figure 1: Left: calibrated ODE vs. held-out real median beats. Right: PINN denoising of real ECG under real 6 dB NSTDB noise.





Figure 2: Multiphysics electrode model: EQS potential in layered tissue, impedance spectra, and motion sensitivity.


[5] J. Song, C. Meng, S. Ermon, Denoising diffusion implicit models, ICLR 2021.

[6] J. Ho, T. Salimans, Classifier-free diffusion guidance, 2022.

[7] C. Saharia et al., Photorealistic text-to-image diffusion models with deep language understanding, NeurIPS 2022.





Figure 3: Randomly drawn (not cherry-picked) real test ECG (top) and diffusion samples with the fixed sampler (middle: full data; bottom: simulation-pretrained, 10% real).