October 03, 2018 Service Request No:R1808738

Ms. Alene Onion
New York State DEC
625 Broadway
Albany, NY 12233-3502

Laboratory Results for: LCI 2018

Dear Ms.Onion,

Enclosed are the results of the sample(s) submitted to our laboratory September 12, 2018
For your reference, these analyses have been assigned our service request number **R1808738.**

All analyses were performed according to our laboratory's quality assurance program. The test results meet requirements of the NELAP standards except as noted in the case narrative report. All results are intended to be considered in their entirety, and ALS Environmental is not responsible for use of less than the complete report. Results apply only to the items submitted to the laboratory for analysis and individual items (samples) analyzed, as listed in the report. The measurement uncertainty of the results included in this report is within that expected when using the prescribed method(s) for analysis of these samples, and represented by Laboratory Control Sample control limits. Any events, such as QC failures, which may add to the uncertainty are explained in the report narrative.

Please contact me if you have any questions. My extension is 7472. You may also contact me via email at Janice.Jaeger@alsglobal.com.

Respectfully submitted,

ALS Group USA, Corp. dba ALS Environmental

Janice Jaeger
Project Manager

CC: Jason Fagel

ADDRESS 1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
PHONE +1 585 288 5380 | FAX +1 585 288 8475
ALS Group USA, Corp.
dba ALS Environmental

Narrative Documents

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

ALS Environmental

Client: New York State DEC
Project: LCI 2018
Sample Matrix: Water

Service Request: R1808738
Date Received: 09/12/2018

CASE NARRATIVE

All analyses were performed consistent with the quality assurance program of ALS Environmental. This report contains analytical results for samples designated for Tier IV, validation deliverables including all summary forms and associated raw data. Analytical procedures performed by the lab are validated in accordance with NELAC standards. Any parameters that are not included in the lab's NELAC accreditation are identified on a "Non-Certified Analytes" report in the Miscellaneous Forms Section of this report. Individual analytical results requiring further explanation are flagged with qualifiers and/or discussed below. The flags are explained in the Report Qualifiers and Definitions page in the Miscellaneous Forms section of this report.

Sample Receipt:

Twenty six water samples were received for analysis at ALS Environmental on 09/12/2018. Any discrepancies noted upon initial sample inspection are noted on the cooler receipt and preservation form included in this data package. The samples were received in good condition and consistent with the accompanying chain of custody form. Samples are refrigerated at 6°C upon receipt at the lab except for aqueous samples designated for metals analyses, which are stored at room temperature.

Metals:

No significant anomalies were noted with this analysis.

General Chemistry:

No significant anomalies were noted with this analysis.

Approved by _____ Date 10/03/2018

Sample Receipt Information

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

SAMPLE CROSS-REFERENCE

SAMPLE #	CLIENT SAMPLE ID	DATE	TIME
R1808738-001	18LHB309	9/11/2018	0930
R1808738-002	18LHB309 Diss	9/11/2018	0930
R1808738-003	18LHB317	9/11/2018	1120
R1808738-004	18LHB317 Diss	9/11/2018	1120
R1808738-005	18LHB318	9/11/2018	1125
R1808738-006	18LHB318 Diss	9/11/2018	1125
R1808738-007	18LHB305	9/11/2018	1300
R1808738-008	18LHB305 Diss	9/11/2018	1300
R1808738-009	18LHB323	9/11/2018	1415
R1808738-010	18LHB323 Diss	9/11/2018	1415
R1808738-011	18LHB321	9/11/2018	0912
R1808738-012	18LHB321 Diss	9/11/2018	0912
R1808738-013	18LHB322	9/11/2018	0925
R1808738-014	18LHB322 Diss	9/11/2018	0925
R1808738-015	18LHB319	9/11/2018	1145
R1808738-016	18LHB319 Diss	9/11/2018	1145
R1808738-017	18LHB320	9/11/2018	1206
R1808738-018	18LHB320 Diss	9/11/2018	1206
R1808738-019	18LHB311	9/11/2018	1351
R1808738-020	18LHB311 Diss	9/11/2018	1351
R1808738-021	18LHB312	9/11/2018	1400
R1808738-022	18LHB312 Diss	9/11/2018	1400
R1808738-023	18LISO60	9/11/2018	1324
R1808738-024	18LISO60 Diss	9/11/2018	1324
R1808738-025	18LISO62	9/11/2018	1530
R1808738-026	18LISO62 Diss	9/11/2018	1530

CHAIN OF CUSTODY

New York State Department of Environmental Conservation – Division of Water

Project Name: LCI	**Project Number:** LCI2018	**NYSDEC SDG:**
Sampler Collector: Sara Gonzalez	**Sampler Signature:** Sara M Gonzalez	**Sampler Phone No.:** 845-216-9575
Project Manager: Alene Onion	**X Report to Project Manager** **Report to:**	**☐ Bill to Project Manager** **Bill to:** Jason Fagel
Address: 625 Broadway, 4th Floor Albany, NY 12233-3502	**Address:**	**Address:** 625 Broadway, 4th Floor Albany, NY 12233-3502
Phone: (518) 402-8166	**Phone:**	**Phone:** 518-402-8156
Email: alene.onion@dec.ny.gov	**Email:**	**Email:** Jason.fagel@dec.ny.gov

Matrix Codes:
WW = Wastewater
GW = Groundwater
AW = Ambient Water
SE = Sediment
SL = Sludge
T = Tissue
O = Other _____

Preservative Codes:
0 = Cool to < 6°C
1 = HCL
2 = HNO_3
3 = H_2SO_4
4 = NaOH
5 = Zn. Acetate
6 = MeOH
7 = NaHSO4
8 = Other_____

Analyses Ordered (list)

NYSDEC LCI Sample ID	Collection Date	Collection Time	Matrix Code	No. of Containers	3: TP, NH4, NOx, TKN	3: TP, NH4, NOx, TKN, NO3 (ANC)	3: Dissolved TOP4	2: Fe, Mn, As,	2: Ca, Mg, Na, K (ANC)	2: Fe, Mn, As, Ca, Mg, Na, K	0: Color	3: TOC	3: DOC (ANC)	0: Alkalinity	0: SO4 & UV-254	0: SO4, Cl (ANC)	0: SO4; Cl, UV-254	0: Chlorophyll a | Vol (ml)		Location Info
18LHB309	09/11	9:30	AW	8	X		X	X			X		X	X	X			X	250	Chodikee L, epi
18LHB317	09/11	11:20	AW	7	X		X	X			X		X	X	X			X	250	Goshen R, epi
18LHB318	09/11	11:25	AW	6	X		X	X			X		X		X					Goshen R, hypo
18LHB305	09/11	13:00	AW	6	X		X				X	X		X				X	150	Binnewater L, epi
18LHB323	09/11	14:15	AW	8	X		X	X			X		X	X	X			X	250	Mill Pond, epi

Special Analysis Instructions:

Relinquished by Sampler:	Date:	Time:	Received by:	Date:	Time:
Sara Gonzalez	09/11	3:30pm			
Relinquished by:	**Date:**	**Time:**	**Received by:**	**Date:**	**Time:**
Relinquished by:	**Date:**	**Time:**	**Received by Laboratory:**	**Date:**	**Time:**
				9/14/18	0930

Laboratory Receipt Notes:

Sample T...
Properly ...
Samples ...

R1808738 5
New York State DEC
LCI 2018

CHAIN OF CUSTODY

New York State Department of Environmental Conservation – Division of Water

Project Name: LCI	**Project Number:** LCI2018	**NYSDEC SDG:**
Sampler Collector: Carrie Buetow	**Sampler Signature:** Carrie Buetow	**Sampler Phone No.:** 5184784975
Project Manager: Alene Onion	**X Report to Project Manager** **Report to:**	☐ **Bill to Project Manager** **Bill to:** Jason Fagel
Address: 625 Broadway, 4th Floor Albany, NY 12233-3502	**Address:**	**Address:** 625 Broadway, 4th Floor Albany, NY 12233-3502
Phone: (518) 402-8166	**Phone:**	**Phone:** 518-402-8156
Email: alene.onion@dec.ny.gov	**Email:**	**Email:** Jason.fagel@dec.ny.gov

Matrix Codes:
WW = Wastewater
GW = Groundwater
AW = Ambient Water
SE = Sediment
SL = Sludge
T = Tissue
O = Other _____

Preservative Codes:
0 = Cool to < 6°C
1 = HCL
2 = HNO_3
3 = H_2SO_4
4 = NaOH
5 = Zn. Acetate
6 = MeOH
7 = NaHSO4
8 = Other_____

Analyses Ordered (list)

NYSDEC LCI Sample ID	Collection Date	Collection Time	Matrix Code	No. of Containers	3 TP, NH4, NOx, TKN	3 ANC TP, NH4, NOx, TKN, NO3	Dissolved TOP4	2 Fe, Mn, As,	ANC Ca, Mg, Na, K	Fe, Mn, As, Ca, Mg, Na, K	0 Color	3 TOC	ANC DOC	0 Alkalinity	SO4 & UV-254	ANC SO4, Cl	SO4, Cl, UV-254	0 Chlorophyll a | Vol (ml)		Location Info
18 LHB 321	09/11/18	9:12	AW	6	X		X				X	X		X				X	250mL	Kinderhook, epi
18 LHB 322	09/11/18	9:25	AW	4	X		X				X	X								Kinderhook, hypo
18 LHB 319	09/11/18	11:45	AW	7	X		X				X		X	X	X			X	500mL	Hollister, epi
18 LHB 320	09/11/18	12:06	AW	6	X		X	X			X		X		X					Hollister, hypo
18 LHB 311	09/11/18	13:51	AW	7	X		X				X		X	X	X			X	250mL	Cossaitie epi
18 LHB 312	09/11/18	14:00	AW	6	X		X	X			X		X		X					Cossaitie hypo

Special Analysis Instructions:

Relinquished by Sampler: Carrie Buetow	**Date:** 9/11/18	**Time:** 4:00 pm	**Received by:**	**Date:**	**Time:**
Relinquished by:	**Date:**	**Time:**	**Received by:**	**Date:**	**Time:**
Relinquished by:	**Date:**	**Time:**	**Received by Laboratory:**	**Date:** 9/12/18	**Time:** 0930

Laboratory Receipt Notes:

Sample
Proper
Sample

R1808738 5
New York State DEC
LCI 2018

CHAIN OF CUSTODY

Page 1 of 1

New York State Department of Environmental Conservation – Division of Water

Project Name: LCI	**Project Number:** LCI2018	**NYSDEC SDG:**
Sampler Collector: Alene Onion	**Sampler Signature:** Alene Onion	**Sampler Phone No.:** 518 402 8166
Project Manager: Alene Onion	**X Report to Project Manager** **Report to:**	☐ **Bill to Project Manager** **Bill to:** Jason Fagel
Address: 625 Broadway, 4th Floor Albany, NY 12233-3502	**Address:**	**Address:** 625 Broadway, 4th Floor Albany, NY 12233-3502
Phone: (518) 402-8166	**Phone:**	**Phone:** 518-402-8156
Email: alene.onion@dec.ny.gov	**Email:**	**Email:** Jason.fagel@dec.ny.gov

Matrix Codes:
WW = Wastewater
GW = Groundwater
AW = Ambient Water
SE = Sediment
SL = Sludge
T = Tissue
O = Other _____

Preservative Codes:
0 = Cool to < 6°C
1 = HCL
2 = HNO_3
3 = H_2SO_4
4 = NaOH
5 = Zn. Acetate
6 = MeOH
7 = NaHSO4
8 = Other_____

Analyses Ordered (list)

NYSDEC LCI Sample ID	Collection Date	Collection Time	Matrix Code	No. of Containers	3: TP, NH4, NOx, TKN	3: ANC TP, NH4, NOx, TKN, NO3	3: Dissolved TOP4	2: Fe, Mn, As,	2: ANC Ca, Mg, Na, K	2: Fe, Mn, As, Ca, Mg, Na, K	0: Color	3: TOC	3: ANC DOC	0: Alkalinity	0: SO4 & UV-254	0: ANC SO4, Cl	0: SO4, Cl, UV-254	0: Chlorophyll a | Vol (ml)		Location Info
18LIS060	9/11/18	1324	AW	6	X		X				X	X		X				X		Fresh Pond (epi)
~~18LIS061~~			~~AW~~	~~4~~	~~X~~		~~X~~				~~X~~	~~X~~								~~Fresh Pond (hypo)~~
18LIS062	9/11/18	1530	AW	6	X		X				X	X		X				X		Big Reed Pond (epi)
~~18LIS063~~			~~AW~~	~~4~~	~~X~~		~~X~~				~~X~~	~~X~~								~~Big Reed Pond (hypo)~~

Special Analysis Instructions:

Relinquished by Sampler: Alene Onion	**Date:** 9/11/18	**Time:** 5:37	**Received by:**	**Date:**	**Time:**	**Laboratory Receipt Notes:**
Relinquished by:	**Date:**	**Time:**	**Received by:**	**Date:**	**Time:**	**Sample Temp.:** °C
Relinquished by:	**Date:**	**Time:**	**Received by Laboratory:**	**Date:** 9/12/18	**Time:** 0930	**Properly Pre** **Samples Int**

R1808738 5
New York State DEC
LCI 2018

Cooler Receipt and Preservation Check Form

R1808738 5
New York State DEC
LCI 2018

Project/Client LCI Folder Number

Cooler received on 9/12/18 by: e

COURIER: ALS (UPS) FEDEX VELOCITY CLIENT

1	Were **Custody seals** on outside of cooler?	(Y) N	5a	**Perchlorate** samples have required headspace?	Y N (NA)
2	**Custody papers** properly completed (ink, signed)?	(Y) N	5b	Did VOA vials, Alk or Sulfide have sig* **bubbles**?	Y (N) NA
3	Did all bottles arrive in good **condition** (unbroken)?	(Y) N	6	Where did the bottles originate?	(ALS/ROC) CLIENT
4	Circle: (Wet Ice) **Dry Ice Gel packs** present?	(Y) N	7	Soil VOA received as: Bulk Encore 5035set	(NA)

8. Temperature Readings Date: 9/12/18 Time: 0933 ID: IR#7 (IR#10) From: Temp Blank (Sample Bottle)

Observed Temp (°C)	4.6	3.4	2.2				
Correction Factor (°C)	+0.4	+0.4	+0.4				
Corrected Temp (°C)	5.0	3.6	2.6				
Temp from:Type of bottle	Centrifuge tube	Centrifuge tube	Centrifuge tube				
Within 0-6°C?	(Y) N	(Y) N	(Y) N	Y N	Y N	Y N	Y N
If <0°C, were samples frozen?	Y N	Y N	Y N	Y N	Y N	Y N	Y N

If out of Temperature, note packing/ice condition: _____ Ice melted Poorly Packed (described below) Same Day Rule

&Client Approval to Run Samples: _____ Standing Approval Client aware at drop-off Client notified by: _____

All samples held in storage location: R-002 by e on 9/12/18 at 0940
5035 samples placed in storage location: _____ by _____ on _____ at _____

Cooler Breakdown/Preservation Check**: Date: 9/13/18 Time: 0825 by: e

9. Were all bottle labels complete (*i.e.* analysis, preservation, etc.)? (YES) NO
10. Did all bottle labels and tags agree with custody papers? YES (NO)
11. Were correct containers used for the tests indicated? (YES) NO
12. Were 5035 vials acceptable (no extra labels, not leaking)? YES NO (N/A)
13. Air Samples: Cassettes / Tubes Intact with MS? Canisters Pressurized Tedlar® Bags Inflated (N/A)

pH	Lot of test paper	Reagent	Preserved? Yes	Preserved? No	Lot Received	Exp	Sample ID Adjusted	Vol. Added	Lot Added	Final pH
≥12		NaOH								
≤2	209318	HNO_3	✓		1117092					
≤2		H_2SO_4	✓		218B0571, 192169					
<4		$NaHSO_4$								
5-9		For 608pest			No=Notify for 3day					
Residual Chlorine (-)		For CN, Phenol, 625, 608pest, 522			If +, contact PM to add $Na_2S_2O_3$ (625, 608, CN), ascorbic (phenol).					
		$Na_2S_2O_3$								
		ZnAcetate	-	-			**VOAs and 1664 Not to be tested before analysis. Otherwise, all bottles of all samples with chemical preservatives are checked (not just representatives).			
		HCl	**	**						

Bottle lot numbers: W133128, 8-072-001, 071611-214A0

Explain all Discrepancies/ Other Comments:

UV-254 bottles
bottle = 18LHB317 COC = 18LHB318
times match
bottle = 18LHB319 COC = 18LHB320
times match

CLRES	BULK
DO	FLDT
HPROD	HGFB
HTR	LL3541
PH	SUB
SO3	MARRS
ALS	REV

Labels secondary reviewed by: e
PC Secondary Review: JMS 9/13/18 *significant air bubbles: VOA > 5-6 mm : WC >1 in. diameter

Miscellaneous Forms

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

REPORT QUALIFIERS AND DEFINITIONS

U	Analyte was analyzed for but not detected. The sample quantitation limit has been corrected for dilution and for percent moisture, unless otherwise noted in the case narrative.
J	Estimated value due to either being a Tentatively Identified Compound (TIC) or that the concentration is between the MRL and the MDL. Concentrations are not verified within the linear range of the calibration. For DoD: concentration >40% difference between two GC columns (pesticides/Arclors).
B	Analyte was also detected in the associated method blank at a concentration that may have contributed to the sample result.
E	Inorganics- Concentration is estimated due to the serial dilution was outside control limits.
E	Organics- Concentration has exceeded the calibration range for that specific analysis.
D	Concentration is a result of a dilution, typically a secondary analysis of the sample due to exceeding the calibration range or that a surrogate has been diluted out of the sample and cannot be assessed.
*	Indicates that a quality control parameter has exceeded laboratory limits. Under the ñNotesò column of the Form I, this qualifier denotes analysis was performed out of Holding Time.
H	Analysis was performed out of hold time for tests that have an ñimmediateò hold time criteria.
#	Spike was diluted out.
+	Correlation coefficient for MSA is <0.995.
N	Inorganics- Matrix spike recovery was outside laboratory limits.
N	Organics- Presumptive evidence of a compound (reported as a TIC) based on the MS library search.
S	Concentration has been determined using Method of Standard Additions (MSA).
W	Post-Digestion Spike recovery is outside control limits and the sample absorbance is <50% of the spike absorbance.
P	Concentration >40% difference between the two GC columns.
C	Confirmed by GC/MS
Q	DoD reports: indicates a pesticide/Aroclor is not confirmed (Ó100% Difference between two GC columns).
X	See Case Narrative for discussion.
MRL LOQ	Method Reporting Limit. Also known as: Limit of Quantitation (LOQ) The lowest concentration at which the method analyte may be reliably quantified under the method conditions.
MDL	Method Detection Limit. A statistical value derived from a study designed to provide the lowest concentration that will be detected 99% of the time. Values between the MDL and MRL are estimated (see J qualifier).
LOD	Limit of Detection. A value at or above the MDL which has been verified to be detectable.
ND	Non-Detect. Analyte was not detected at the concentration listed. Same as U qualifier.

Rochester Lab ID # for State Certifications[1]

Connecticut ID # PH0556	Maine ID #NY0032	Pennsylvania ID# 68-786
Delaware Approved	New Hampshire ID # 2941	Rhode Island ID # 158
DoD ELAP #65817	New York ID # 10145	Virginia #460167
Florida ID # E87674	North Carolina #676	

[1] Analyses were performed according to our laboratoryôs NELAP-approved quality assurance program and any applicable state or agency requirements. The test results meet requirements of the current NELAP/TNI standards or state or agency requirements, where applicable, except as noted in the case narrative. Since not all analyte/method/matrix combinations are offered for state/NELAC accreditation, this report may contain results which are not accredited. For a specific list of accredited analytes, contact the laboratory or go to https://www.alsglobal.com/locations/americas/north-america/usa/new-york/rochester-environmental

ALS Laboratory Group

Acronyms

ASTM	American Society for Testing and Materials
A2LA	American Association for Laboratory Accreditation
CARB	California Air Resources Board
CAS Number	Chemical Abstract Service registry Number
CFC	Chlorofluorocarbon
CFU	Colony-Forming Unit
DEC	Department of Environmental Conservation
DEQ	Department of Environmental Quality
DHS	Department of Health Services
DOE	Department of Ecology
DOH	Department of Health
EPA	U. S. Environmental Protection Agency
ELAP	Environmental Laboratory Accreditation Program
GC	Gas Chromatography
GC/MS	Gas Chromatography/Mass Spectrometry
LUFT	Leaking Underground Fuel Tank
M	Modified
MCL	Maximum Contaminant Level is the highest permissible concentration of a substance allowed in drinking water as established by the USEPA.
MDL	Method Detection Limit
MPN	Most Probable Number
MRL	Method Reporting Limit
NA	Not Applicable
NC	Not Calculated
NCASI	National Council of the Paper Industry for Air and Stream Improvement
ND	Not Detected
NIOSH	National Institute for Occupational Safety and Health
PQL	Practical Quantitation Limit
RCRA	Resource Conservation and Recovery Act
SIM	Selected Ion Monitoring
TPH	Total Petroleum Hydrocarbons
tr	Trace level is the concentration of an analyte that is less than the PQL but greater than or equal to the MDL.

Client: New York State DEC **Service Request:** R1808738
Project: LCI 2018/LCI2018

Non-Certified Analytes

Certifying Agency: New York Department of Health

Method	Matrix	Analyte
SM 5910 B	Water	UV254
SM20 10200 H	Water	Chlorophyll A

Client: New York State DEC **Service Request:** R1808738
Project: LCI 2018/LCI2018

Sample Name: 18LHB309 **Date Collected:** 09/11/18
Lab Code: R1808738-001 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
300.0		AMOSES
351.2	NSMITH	CWOODS
353.2		MROGERSON
365.1	MROGERSON	MROGERSON
ASTM D6919-09		AMOSES
SM 2120 B-2001(2011)		SCYMBAL
SM 2320 B-1997(2011)		CWOODS
SM 5910 B		MROGERSON
SM20 10200 H		NSMITH

Sample Name: 18LHB309 Diss **Date Collected:** 09/11/18
Lab Code: R1808738-002 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
365.1	KWONG	GNITAJOUPPI
SM 5310 C-2000(2011)		CWOODS

Sample Name: 18LHB317 **Date Collected:** 09/11/18
Lab Code: R1808738-003 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
300.0		AMOSES
351.2	NSMITH	CWOODS
353.2		MROGERSON
365.1	MROGERSON	MROGERSON
ASTM D6919-09		AMOSES
SM 2120 B-2001(2011)		SCYMBAL
SM 2320 B-1997(2011)		CWOODS
SM 5910 B		MROGERSON
SM20 10200 H		NSMITH

Client: New York State DEC **Service Request:** R1808738
Project: LCI 2018/LCI2018

Sample Name: 18LHB317 Diss **Date Collected:** 09/11/18
Lab Code: R1808738-004 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
365.1	KWONG	GNITAJOUPPI
SM 5310 C-2000(2011)		CWOODS

Sample Name: 18LHB318 **Date Collected:** 09/11/18
Lab Code: R1808738-005 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
300.0		AMOSES
351.2	NSMITH	CWOODS
353.2		MROGERSON
365.1	MROGERSON	MROGERSON
ASTM D6919-09		AMOSES
SM 2120 B-2001(2011)		SCYMBAL
SM 5910 B		MROGERSON

Sample Name: 18LHB318 Diss **Date Collected:** 09/11/18
Lab Code: R1808738-006 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
365.1	KWONG	GNITAJOUPPI
SM 5310 C-2000(2011)		CWOODS

Sample Name: 18LHB305 **Date Collected:** 09/11/18
Lab Code: R1808738-007 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
351.2	NSMITH	CWOODS
353.2		MROGERSON

Client: New York State DEC **Service Request:** R1808738
Project: LCI 2018/LCI2018

Sample Name: 18LHB305 **Date Collected:** 09/11/18
Lab Code: R1808738-007 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
365.1	MROGERSON	MROGERSON
ASTM D6919-09		AMOSES
SM 2120 B-2001(2011)		SCYMBAL
SM 2320 B-1997(2011)		CWOODS
SM 5310 C-2000(2011)		CWOODS
SM20 10200 H		NSMITH

Sample Name: 18LHB305 Diss **Date Collected:** 09/11/18
Lab Code: R1808738-008 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
365.1	KWONG	GNITAJOUPPI

Sample Name: 18LHB323 **Date Collected:** 09/11/18
Lab Code: R1808738-009 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
300.0		AMOSES
351.2	NSMITH	CWOODS
353.2		MROGERSON
365.1	MROGERSON	MROGERSON
ASTM D6919-09		AMOSES
SM 2120 B-2001(2011)		SCYMBAL
SM 2320 B-1997(2011)		CWOODS
SM 5910 B		MROGERSON
SM20 10200 H		NSMITH

Client: New York State DEC **Service Request:** R1808738
Project: LCI 2018/LCI2018

Sample Name: 18LHB323 Diss **Date Collected:** 09/11/18
Lab Code: R1808738-010 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
365.1	KWONG	GNITAJOUPPI
SM 5310 C-2000(2011)		CWOODS

Sample Name: 18LHB321 **Date Collected:** 09/11/18
Lab Code: R1808738-011 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
351.2	NSMITH	CWOODS
353.2		MROGERSON
365.1	KWONG	GNITAJOUPPI
ASTM D6919-09		AMOSES
SM 2120 B-2001(2011)		SCYMBAL
SM 2320 B-1997(2011)		CWOODS
SM 5310 C-2000(2011)		CWOODS
SM20 10200 H		NSMITH

Sample Name: 18LHB321 Diss **Date Collected:** 09/11/18
Lab Code: R1808738-012 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
365.1	KWONG	GNITAJOUPPI

Sample Name: 18LHB322 **Date Collected:** 09/11/18
Lab Code: R1808738-013 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
351.2	NSMITH	CWOODS
353.2		MROGERSON

Client: New York State DEC **Service Request:** R1808738
Project: LCI 2018/LCI2018

Sample Name: 18LHB322 **Date Collected:** 09/11/18
Lab Code: R1808738-013 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
365.1	KWONG	GNITAJOUPPI
ASTM D6919-09		AMOSES
SM 2120 B-2001(2011)		SCYMBAL
SM 5310 C-2000(2011)		CWOODS

Sample Name: 18LHB322 Diss **Date Collected:** 09/11/18
Lab Code: R1808738-014 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
365.1	KWONG	GNITAJOUPPI

Sample Name: 18LHB319 **Date Collected:** 09/11/18
Lab Code: R1808738-015 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
300.0		AMOSES
351.2	NSMITH	CWOODS
353.2		MROGERSON
365.1	KWONG	GNITAJOUPPI
ASTM D6919-09		AMOSES
SM 2120 B-2001(2011)		SCYMBAL
SM 2320 B-1997(2011)		CWOODS
SM 5910 B		MROGERSON
SM20 10200 H		NSMITH

Client: New York State DEC **Service Request:** R1808738
Project: LCI 2018/LCI2018

Sample Name: 18LHB319 Diss **Date Collected:** 09/11/18
Lab Code: R1808738-016 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
365.1	KWONG	GNITAJOUPPI
SM 5310 C-2000(2011)		CWOODS

Sample Name: 18LHB320 **Date Collected:** 09/11/18
Lab Code: R1808738-017 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
300.0		AMOSES
351.2	NSMITH	CWOODS
353.2		MROGERSON
365.1	KWONG	GNITAJOUPPI
ASTM D6919-09		AMOSES
SM 2120 B-2001(2011)		SCYMBAL
SM 5910 B		MROGERSON

Sample Name: 18LHB320 Diss **Date Collected:** 09/11/18
Lab Code: R1808738-018 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
365.1	KWONG	GNITAJOUPPI
SM 5310 C-2000(2011)		CWOODS

Sample Name: 18LHB311 **Date Collected:** 09/11/18
Lab Code: R1808738-019 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
300.0		AMOSES
351.2	NSMITH	CWOODS

Client: New York State DEC **Service Request:** R1808738
Project: LCI 2018/LCI2018

Sample Name: 18LHB311 **Date Collected:** 09/11/18
Lab Code: R1808738-019 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
353.2		MROGERSON
365.1	KWONG	GNITAJOUPPI
ASTM D6919-09		AMOSES
SM 2120 B-2001(2011)		SCYMBAL
SM 2320 B-1997(2011)		CWOODS
SM 5910 B		MROGERSON
SM20 10200 H		NSMITH

Sample Name: 18LHB311 Diss **Date Collected:** 09/11/18
Lab Code: R1808738-020 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
365.1	KWONG	GNITAJOUPPI
SM 5310 C-2000(2011)		CWOODS

Sample Name: 18LHB312 **Date Collected:** 09/11/18
Lab Code: R1808738-021 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
300.0		AMOSES
351.2	NSMITH	CWOODS
353.2		MROGERSON
365.1	KWONG	GNITAJOUPPI
ASTM D6919-09		AMOSES
SM 2120 B-2001(2011)		SCYMBAL
SM 5910 B		MROGERSON

Client: New York State DEC **Service Request:** R1808738
Project: LCI 2018/LCI2018

Sample Name: 18LHB312 Diss **Date Collected:** 09/11/18
Lab Code: R1808738-022 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
365.1	KWONG	GNITAJOUPPI
SM 5310 C-2000(2011)		CWOODS

Sample Name: 18LISO60 **Date Collected:** 09/11/18
Lab Code: R1808738-023 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
351.2	NSMITH	CWOODS
353.2		MROGERSON
365.1	KWONG	GNITAJOUPPI
ASTM D6919-09		AMOSES
SM 2120 B-2001(2011)		SCYMBAL
SM 2320 B-1997(2011)		CWOODS
SM 5310 C-2000(2011)		CWOODS
SM20 10200 H		NSMITH

Sample Name: 18LISO60 Diss **Date Collected:** 09/11/18
Lab Code: R1808738-024 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
365.1	KWONG	GNITAJOUPPI

Sample Name: 18LISO62 **Date Collected:** 09/11/18
Lab Code: R1808738-025 **Date Received:** 09/12/18
Sample Matrix: Water

Analysis Method	Extracted/Digested By	Analyzed By
351.2	NSMITH	CWOODS
353.2		MROGERSON

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018		

Sample Name:	18LISO62	**Date Collected:**	09/11/18
Lab Code:	R1808738-025	**Date Received:**	09/12/18
Sample Matrix:	Water		

Analysis Method	Extracted/Digested By	Analyzed By
365.1	KWONG	GNITAJOUPPI
ASTM D6919-09		AMOSES
SM 2120 B-2001(2011)		SCYMBAL
SM 2320 B-1997(2011)		CWOODS
SM 5310 C-2000(2011)		CWOODS
SM20 10200 H		NSMITH

Sample Name:	18LISO62 Diss	**Date Collected:**	09/11/18
Lab Code:	R1808738-026	**Date Received:**	09/12/18
Sample Matrix:	Water		

Analysis Method	Extracted/Digested By	Analyzed By
365.1	KWONG	GNITAJOUPPI

The preparation methods associated with this report are found in these tables unless discussed in the case narrative.

Water/Liquid Matrix

Analytical Method	Preparation Method
200.7	200.2
200.8	200.2
6010C	3005A/3010A
6020A	ILM05.3
9014 Cyanide Reactivity	SW846 Ch7, 7.3.4.2
9034 Sulfide Reactivity	SW846 Ch7, 7.3.4.2
9034 Sulfide Acid Soluble	9030B
9056A Bomb (Halogens)	5050A
9066 Manual Distillation	9065
SM 4500-CN-E Residual Cyanide	SM 4500-CN-G
SM 4500-CN-E WAD Cyanide	SM 4500-CN-I

Solid/Soil/Non-Aqueous Matrix

Analytical Method	Preparation Method
6010C	3050B
6020A	3050B
6010C TCLP (1311) extract	3005A/3010A
6010 SPLP (1312) extract	3005A/3010A
7196A	3060A
7199	3060A
9056A Halogens/Halides	5050
300.0 Anions/ 350.1/ 353.2/ SM 2320B/ SM 5210B/ 9056A Anions	DI extraction

For analytical methods not listed, the preparation method is the same as the analytical method reference.

Sample Results

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

Metals

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

METALS

- 1 -

INORGANIC ANALYSIS DATA PACKAGE

Client:	New York State DEC	**Service Request:**	LCI0910
Project No.:	R1808738	**Date Collected:**	9/11/2018
Project Name:		**Date Received:**	9/12/2018
Matrix:	WATER	**Units:**	ug/L
		Basis:	

Sample Name: **18LHB309** **Lab Code:** **R1808738-001**

Analyte	Analysis Method	PQL	MDL	Dil. Factor	Result	C	Q
Arsenic	**200.8**	**1.0**	**0.39**	**1.0**	**1.2**		
Iron	**200.7**	**100**	**13.0**	**1.0**	**470**		
Manganese	**200.7**	**10.0**	**1.7**	**1.0**	**479**		

% Solids: **0.0**

Comments:

METALS

- 1 -

INORGANIC ANALYSIS DATA PACKAGE

Client:	New York State DEC	**Service Request:**	LCI0910
Project No.:	R1808738	**Date Collected:**	9/11/2018
Project Name:		**Date Received:**	9/12/2018
Matrix:	WATER	**Units:**	ug/L
		Basis:	

Sample Name: **18LHB318** **Lab Code:** **R1808738-005**

Analyte	Analysis Method	PQL	MDL	Dil. Factor	Result	C	Q
Arsenic	**200.8**	**1.0**	**0.39**	**1.0**	**1.2**		
Iron	**200.7**	**100**	**13.0**	**1.0**	**1270**		
Manganese	**200.7**	**10.0**	**1.7**	**1.0**	**339**		

% Solids: **0.0**

Comments:

METALS

- 1 -

INORGANIC ANALYSIS DATA PACKAGE

Client:	New York State DEC	**Service Request:**	LCI0910
Project No.:	R1808738	**Date Collected:**	9/11/2018
Project Name:		**Date Received:**	9/12/2018
Matrix:	WATER	**Units:**	ug/L
		Basis:	

Sample Name: **18LHB323** **Lab Code:** **R1808738-009**

Analyte	Analysis Method	PQL	MDL	Dil. Factor	Result	C	Q
Arsenic	**200.8**	**1.0**	**0.39**	**1.0**	**0.88**	**J**	
Iron	**200.7**	**100**	**13.0**	**1.0**	**608**		
Manganese	**200.7**	**10.0**	**1.7**	**1.0**	**80.9**		

% Solids: **0.0**

Comments:

METALS

- 1 -

INORGANIC ANALYSIS DATA PACKAGE

Client:	New York State DEC	**Service Request:**	LCI0910
Project No.:	R1808738	**Date Collected:**	9/11/2018
Project Name:		**Date Received:**	9/12/2018
Matrix:	WATER	**Units:**	ug/L
		Basis:	

Sample Name: **18LHB320** **Lab Code:** **R1808738-017**

Analyte	Analysis Method	PQL	MDL	Dil. Factor	Result	C	Q
Arsenic	**200.8**	**1.0**	**0.39**	**1.0**	**0.59**	**J**	
Iron	**200.7**	**100**	**13.0**	**1.0**	**401**		
Manganese	**200.7**	**10.0**	**1.7**	**1.0**	**1230**		

% Solids: **0.0**

Comments:

METALS

- 1 -

INORGANIC ANALYSIS DATA PACKAGE

Client:	New York State DEC	**Service Request:**	LCI0910
Project No.:	R1808738	**Date Collected:**	9/11/2018
Project Name:		**Date Received:**	9/12/2018
Matrix:	WATER	**Units:**	ug/L
		Basis:	

Sample Name: **18LHB312** **Lab Code:** **R1808738-021**

Analyte	Analysis Method	PQL	MDL	Dil. Factor	Result	C	Q
Arsenic	**200.8**	**1.0**	**0.39**	**1.0**	**1.1**		
Iron	**200.7**	**100**	**13.0**	**1.0**	**1600**		
Manganese	**200.7**	**10.0**	**1.7**	**1.0**	**1180**		

% Solids: **0.0**

Comments:

General Chemistry

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 09:30
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB309	**Basis:**	NA
Lab Code:	R1808738-001		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Alkalinity, Total as CaCO3	SM 2320 B-1997(2011)	**104**	mg/L	2.0	1	09/17/18 20:50	NA	
Ammonia as Nitrogen, undistilled	ASTM D6919-09	**0.500**	mg/L	0.0050	1	09/18/18 23:53	NA	
Chlorophyll A	SM20 10200 H	**17.2**	ug/L	0.64	4	09/25/18 10:35	NA	
Color, True	SM 2120 B-2001(2011)	**53.0**	ColorUnits	1.0	1	09/12/18 14:00	NA	
Nitrate+Nitrite as Nitrogen	353.2	**0.0078**	mg/L	0.0020	1	09/17/18 19:07	NA	
Nitrogen, Total Kjeldahl (TKN)	351.2	**1.54**	mg/L	0.10	1	09/24/18 18:39	09/21/18	
pH of Color Analysis	SM 2120 B-2001(2011)	**7.48**	pH Units	-	1	09/15/18 10:00	NA	*
Phosphorus, Total	365.1	**0.0986**	mg/L	0.0050	1	09/21/18 17:16	09/19/18	
Sulfate	300.0	**5.0**	mg/L	2.0	10	09/17/18 20:09	NA	
UV254	SM 5910 B	**0.255**	cm-1	-	1	09/12/18 22:40	NA	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 09:30
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB309 Diss	**Basis:**	NA
Lab Code:	R1808738-002		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Carbon, Dissolved Organic (DOC)	SM 5310 C-2000(2011)	**8.7**	mg/L	1.0	1	09/13/18 21:06	NA	
Phosphorus, Dissolved	365.1	**0.0271**	mg/L	0.0050	1	09/28/18 10:20	09/24/18	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 11:20
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB317	**Basis:**	NA
Lab Code:	R1808738-003		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Alkalinity, Total as CaCO3	SM 2320 B-1997(2011)	**43.6**	mg/L	2.0	1	09/17/18 20:55	NA	
Ammonia as Nitrogen, undistilled	ASTM D6919-09	**0.147**	mg/L	0.0050	1	09/19/18 00:09	NA	
Chlorophyll A	SM20 10200 H	**18.1**	ug/L	0.64	4	09/25/18 10:35	NA	
Color, True	SM 2120 B-2001(2011)	**24.0**	ColorUnits	1.0	1	09/12/18 14:00	NA	
Nitrate+Nitrite as Nitrogen	353.2	**0.0129**	mg/L	0.0020	1	09/17/18 19:12	NA	
Nitrogen, Total Kjeldahl (TKN)	351.2	**0.95**	mg/L	0.10	1	09/24/18 18:39	09/21/18	
pH of Color Analysis	SM 2120 B-2001(2011)	**7.25**	pH Units	-	1	09/15/18 10:00	NA	*
Phosphorus, Total	365.1	**0.0307**	mg/L	0.0050	1	09/21/18 17:17	09/19/18	
Sulfate	300.0	**8.0**	mg/L	2.0	10	09/17/18 20:14	NA	
UV254	SM 5910 B	**0.0850**	cm-1	-	1	09/12/18 22:40	NA	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 11:20
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB317 Diss	**Basis:**	NA
Lab Code:	R1808738-004		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Carbon, Dissolved Organic (DOC)	SM 5310 C-2000(2011)	**4.9**	mg/L	1.0	1	09/13/18 22:51	NA	
Phosphorus, Dissolved	365.1	**0.0116**	mg/L	0.0050	1	09/28/18 10:23	09/24/18	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 11:25
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB318	**Basis:**	NA
Lab Code:	R1808738-005		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Ammonia as Nitrogen, undistilled	ASTM D6919-09	**0.165**	mg/L	0.0050	1	09/19/18 00:25	NA	
Color, True	SM 2120 B-2001(2011)	**33.0**	ColorUnits	1.0	1	09/12/18 14:00	NA	
Nitrate+Nitrite as Nitrogen	353.2	**0.0160**	mg/L	0.0020	1	09/17/18 19:13	NA	
Nitrogen, Total Kjeldahl (TKN)	351.2	**0.82**	mg/L	0.10	1	09/24/18 18:40	09/21/18	
pH of Color Analysis	SM 2120 B-2001(2011)	**7.54**	pH Units	-	1	09/15/18 10:00	NA	*
Phosphorus, Total	365.1	**0.0441**	mg/L	0.0050	1	09/21/18 17:19	09/19/18	
Sulfate	300.0	**8.2**	mg/L	2.0	10	09/17/18 20:19	NA	
UV254	SM 5910 B	**0.0880**	cm-1	-	1	09/12/18 22:40	NA	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 11:25
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB318 Diss	**Basis:**	NA
Lab Code:	R1808738-006		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Carbon, Dissolved Organic (DOC)	SM 5310 C-2000(2011)	**4.6**	mg/L	1.0	1	09/13/18 23:12	NA	
Phosphorus, Dissolved	365.1	**0.0096**	mg/L	0.0050	1	09/28/18 10:24	09/24/18	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 13:00
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB305	**Basis:**	NA
Lab Code:	R1808738-007		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Alkalinity, Total as CaCO3	SM 2320 B-1997(2011)	**30.0**	mg/L	2.0	1	09/17/18 20:59	NA	
Ammonia as Nitrogen, undistilled	ASTM D6919-09	**0.0053**	mg/L	0.0050	1	09/19/18 00:41	NA	
Carbon, Total Organic (TOC)	SM 5310 C-2000(2011)	**10.0**	mg/L	1.0	1	09/14/18 01:38	NA	
Chlorophyll A	SM20 10200 H	**106**	ug/L	11	40	09/25/18 10:35	NA	
Color, True	SM 2120 B-2001(2011)	**110**	ColorUnits	5.0	5	09/12/18 14:00	NA	
Nitrate+Nitrite as Nitrogen	353.2	**0.0036**	mg/L	0.0020	1	09/17/18 19:14	NA	
Nitrogen, Total Kjeldahl (TKN)	351.2	**1.37**	mg/L	0.10	1	09/24/18 18:42	09/21/18	
pH of Color Analysis	SM 2120 B-2001(2011)	**7.02**	pH Units	-	5	09/15/18 10:00	NA	*
Phosphorus, Total	365.1	**0.0417**	mg/L	0.0050	1	09/21/18 17:20	09/19/18	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 13:00
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB305 Diss	**Basis:**	NA
Lab Code:	R1808738-008		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Phosphorus, Dissolved	365.1	**0.0191**	mg/L	0.0050	1	09/28/18 10:26	09/24/18	

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 14:15
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB323	**Basis:**	NA
Lab Code:	R1808738-009		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Alkalinity, Total as CaCO3	SM 2320 B-1997(2011)	**54.8**	mg/L	2.0	1	09/17/18 21:04	NA	
Ammonia as Nitrogen, undistilled	ASTM D6919-09	**0.0676**	mg/L	0.0050	1	09/19/18 00:57	NA	
Chlorophyll A	SM20 10200 H	**4.74**	ug/L	0.16	1	09/25/18 10:35	NA	
Color, True	SM 2120 B-2001(2011)	**90.0**	ColorUnits	5.0	5	09/12/18 14:00	NA	
Nitrate+Nitrite as Nitrogen	353.2	**0.109**	mg/L	0.0020	1	09/17/18 19:16	NA	
Nitrogen, Total Kjeldahl (TKN)	351.2	**0.83**	mg/L	0.10	1	09/24/18 18:43	09/21/18	
pH of Color Analysis	SM 2120 B-2001(2011)	**7.47**	pH Units	-	5	09/15/18 10:00	NA	*
Phosphorus, Total	365.1	**0.0751**	mg/L	0.0050	1	09/21/18 17:21	09/19/18	
Sulfate	300.0	**6.2**	mg/L	2.0	10	09/17/18 20:24	NA	
UV254	SM 5910 B	**0.288**	cm-1	-	1	09/12/18 22:40	NA	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 14:15
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB323 Diss	**Basis:**	NA
Lab Code:	R1808738-010		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Carbon, Dissolved Organic (DOC)	SM 5310 C-2000(2011)	**9.3**	mg/L	1.0	1	09/13/18 23:32	NA	
Phosphorus, Dissolved	365.1	**0.0408**	mg/L	0.0050	1	09/28/18 10:27	09/24/18	

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 09:12
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB321	**Basis:**	NA
Lab Code:	R1808738-011		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Alkalinity, Total as CaCO3	SM 2320 B-1997(2011)	**66.4**	mg/L	2.0	1	09/17/18 21:14	NA	
Ammonia as Nitrogen, undistilled	ASTM D6919-09	**0.0461**	mg/L	0.0050	1	09/19/18 01:14	NA	
Carbon, Total Organic (TOC)	SM 5310 C-2000(2011)	**6.1**	mg/L	1.0	1	09/14/18 03:22	NA	
Chlorophyll A	SM20 10200 H	**20.5**	ug/L	0.64	4	09/25/18 10:35	NA	
Color, True	SM 2120 B-2001(2011)	**32.0**	ColorUnits	1.0	1	09/12/18 14:00	NA	
Nitrate+Nitrite as Nitrogen	353.2	**0.0197**	mg/L	0.0020	1	09/17/18 19:17	NA	
Nitrogen, Total Kjeldahl (TKN)	351.2	**0.69**	mg/L	0.10	1	09/24/18 18:44	09/21/18	
pH of Color Analysis	SM 2120 B-2001(2011)	**7.55**	pH Units	-	1	09/15/18 10:00	NA	*
Phosphorus, Total	365.1	**0.0272**	mg/L	0.0050	1	09/28/18 11:59	09/24/18	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 09:12
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB321 Diss	**Basis:**	NA
Lab Code:	R1808738-012		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Phosphorus, Dissolved	365.1	**0.0054**	mg/L	0.0050	1	09/28/18 11:34	09/24/18	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 09:25
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB322	**Basis:**	NA
Lab Code:	R1808738-013		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Ammonia as Nitrogen, undistilled	ASTM D6919-09	**1.54**	mg/L	0.0050	1	09/19/18 03:22	NA	
Carbon, Total Organic (TOC)	SM 5310 C-2000(2011)	**5.0**	mg/L	1.0	1	09/14/18 04:25	NA	
Color, True	SM 2120 B-2001(2011)	**52.0**	ColorUnits	1.0	1	09/12/18 14:00	NA	
Nitrate+Nitrite as Nitrogen	353.2	**0.0074**	mg/L	0.0020	1	09/17/18 19:18	NA	
Nitrogen, Total Kjeldahl (TKN)	351.2	**1.87**	mg/L	0.10	1	09/24/18 18:51	09/21/18	
pH of Color Analysis	SM 2120 B-2001(2011)	**7.51**	pH Units	-	1	09/15/18 10:00	NA	*
Phosphorus, Total	365.1	**0.117**	mg/L	0.025	5	09/28/18 12:59	09/24/18	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 09:25
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB322 Diss	**Basis:**	NA
Lab Code:	R1808738-014		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Phosphorus, Dissolved	365.1	**0.084**	mg/L	0.025	5	09/28/18 10:31	09/24/18	

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 11:45
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB319	**Basis:**	NA
Lab Code:	R1808738-015		

Inorganic Parameters

Analyte Name	Analysis Method	Result		Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Alkalinity, Total as CaCO3	SM 2320 B-1997(2011)	**66.0**		mg/L	2.0	1	09/17/18 21:24	NA	
Ammonia as Nitrogen, undistilled	ASTM D6919-09	0.0050	U	mg/L	0.0050	1	09/19/18 03:38	NA	
Chlorophyll A	SM20 10200 H	**10.6**		ug/L	0.32	4	09/25/18 10:35	NA	
Color, True	SM 2120 B-2001(2011)	**42.0**		ColorUnits	1.0	1	09/12/18 14:00	NA	
Nitrate+Nitrite as Nitrogen	353.2	0.0020	U	mg/L	0.0020	1	09/17/18 19:20	NA	
Nitrogen, Total Kjeldahl (TKN)	351.2	**0.51**		mg/L	0.10	1	09/24/18 18:53	09/21/18	
pH of Color Analysis	SM 2120 B-2001(2011)	**7.53**		pH Units	-	1	09/15/18 10:00	NA	*
Phosphorus, Total	365.1	**0.0266**		mg/L	0.0050	1	09/28/18 12:04	09/24/18	
Sulfate	300.0	**5.5**		mg/L	2.0	10	09/17/18 20:30	NA	
UV254	SM 5910 B	**0.181**		cm-1	-	1	09/12/18 22:40	NA	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 11:45
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB319 Diss	**Basis:**	NA
Lab Code:	R1808738-016		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Carbon, Dissolved Organic (DOC)	SM 5310 C-2000(2011)	**6.3**	mg/L	1.0	1	09/13/18 23:53	NA	
Phosphorus, Dissolved	365.1	**0.0079**	mg/L	0.0050	1	09/28/18 10:32	09/24/18	

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 12:06
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB320	**Basis:**	NA
Lab Code:	R1808738-017		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Ammonia as Nitrogen, undistilled	ASTM D6919-09	**0.0592**	mg/L	0.0050	1	09/19/18 03:54	NA	
Color, True	SM 2120 B-2001(2011)	**27.0**	ColorUnits	1.0	1	09/12/18 14:00	NA	
Nitrate+Nitrite as Nitrogen	353.2	**0.0035**	mg/L	0.0020	1	09/17/18 19:24	NA	
Nitrogen, Total Kjeldahl (TKN)	351.2	**0.73**	mg/L	0.10	1	09/24/18 18:54	09/21/18	
pH of Color Analysis	SM 2120 B-2001(2011)	**7.43**	pH Units	-	1	09/15/18 10:00	NA	*
Phosphorus, Total	365.1	**0.0759**	mg/L	0.0050	1	09/28/18 12:05	09/24/18	
Sulfate	300.0	**6.1**	mg/L	2.0	10	09/17/18 20:35	NA	
UV254	SM 5910 B	**0.112**	cm-1	-	1	09/12/18 22:40	NA	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 12:06
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB320 Diss	**Basis:**	NA
Lab Code:	R1808738-018		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Carbon, Dissolved Organic (DOC)	SM 5310 C-2000(2011)	**4.5**	mg/L	1.0	1	09/14/18 00:14	NA	
Phosphorus, Dissolved	365.1	**0.0307**	mg/L	0.0050	1	09/28/18 10:33	09/24/18	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 13:51
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB311	**Basis:**	NA
Lab Code:	R1808738-019		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Alkalinity, Total as CaCO3	SM 2320 B-1997(2011)	**18.0**	mg/L	2.0	1	09/17/18 21:33	NA	
Ammonia as Nitrogen, undistilled	ASTM D6919-09	**0.0290**	mg/L	0.0050	1	09/19/18 04:10	NA	
Chlorophyll A	SM20 10200 H	**27.9**	ug/L	0.64	4	09/25/18 10:35	NA	
Color, True	SM 2120 B-2001(2011)	**43.0**	ColorUnits	1.0	1	09/12/18 14:00	NA	
Nitrate+Nitrite as Nitrogen	353.2	**0.0040**	mg/L	0.0020	1	09/17/18 19:25	NA	
Nitrogen, Total Kjeldahl (TKN)	351.2	**0.91**	mg/L	0.10	1	09/24/18 18:54	09/21/18	
pH of Color Analysis	SM 2120 B-2001(2011)	**7.53**	pH Units	-	1	09/15/18 10:00	NA	*
Phosphorus, Total	365.1	**0.0273**	mg/L	0.0050	1	09/28/18 12:06	09/24/18	
Sulfate	300.0	**5.4**	mg/L	2.0	10	09/17/18 20:40	NA	
UV254	SM 5910 B	**0.178**	cm-1	-	1	09/12/18 22:40	NA	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 13:51
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB311 Diss	**Basis:**	NA
Lab Code:	R1808738-020		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Carbon, Dissolved Organic (DOC)	SM 5310 C-2000(2011)	**7.1**	mg/L	1.0	1	09/14/18 00:56	NA	
Phosphorus, Dissolved	365.1	**0.0076**	mg/L	0.0050	1	09/28/18 10:34	09/24/18	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 14:00
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB312	**Basis:**	NA
Lab Code:	R1808738-021		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Ammonia as Nitrogen, undistilled	ASTM D6919-09	**0.144**	mg/L	0.0050	1	09/19/18 04:26	NA	
Color, True	SM 2120 B-2001(2011)	**58.0**	ColorUnits	1.0	1	09/12/18 14:00	NA	
Nitrate+Nitrite as Nitrogen	353.2	**0.0073**	mg/L	0.0020	1	09/17/18 19:27	NA	
Nitrogen, Total Kjeldahl (TKN)	351.2	**0.85**	mg/L	0.10	1	09/24/18 19:14	09/21/18	
pH of Color Analysis	SM 2120 B-2001(2011)	**7.04**	pH Units	-	1	09/15/18 10:00	NA	*
Phosphorus, Total	365.1	**0.0417**	mg/L	0.0050	1	09/28/18 12:07	09/24/18	
Sulfate	300.0	**4.7**	mg/L	2.0	10	09/17/18 20:45	NA	
UV254	SM 5910 B	**0.213**	cm-1	-	1	09/12/18 22:40	NA	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 14:00
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LHB312 Diss	**Basis:**	NA
Lab Code:	R1808738-022		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Carbon, Dissolved Organic (DOC)	SM 5310 C-2000(2011)	**6.1**	mg/L	1.0	1	09/14/18 01:17	NA	
Phosphorus, Dissolved	365.1	**0.0125**	mg/L	0.0050	1	09/28/18 10:36	09/24/18	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 13:24
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LISO60	**Basis:**	NA
Lab Code:	R1808738-023		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Alkalinity, Total as CaCO3	SM 2320 B-1997(2011)	**17.6**	mg/L	2.0	1	09/17/18 21:45	NA	
Ammonia as Nitrogen, undistilled	ASTM D6919-09	**0.0058**	mg/L	0.0050	1	09/19/18 06:34	NA	
Carbon, Total Organic (TOC)	SM 5310 C-2000(2011)	**13.4**	mg/L	4.0	4	09/14/18 11:13	NA	
Chlorophyll A	SM20 10200 H	**12.0**	ug/L	0.64	4	09/25/18 10:35	NA	
Color, True	SM 2120 B-2001(2011)	**80.0**	ColorUnits	5.0	5	09/12/18 14:00	NA	
Nitrate+Nitrite as Nitrogen	353.2	0.0020 U	mg/L	0.0020	1	09/17/18 19:28	NA	
Nitrogen, Total Kjeldahl (TKN)	351.2	**1.55**	mg/L	0.10	1	09/24/18 19:32	09/21/18	
pH of Color Analysis	SM 2120 B-2001(2011)	**7.37**	pH Units	-	5	09/15/18 10:00	NA	*
Phosphorus, Total	365.1	**0.0643**	mg/L	0.0050	1	09/28/18 12:08	09/24/18	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 13:24
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LISO60 Diss	**Basis:**	NA
Lab Code:	R1808738-024		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Phosphorus, Dissolved	365.1	**0.0085**	mg/L	0.0050	1	09/28/18 10:37	09/24/18	

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 15:30
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LISO62	**Basis:**	NA
Lab Code:	R1808738-025		

Inorganic Parameters

Analyte Name	Analysis Method	Result		Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Alkalinity, Total as CaCO3	SM 2320 B-1997(2011)	**2.0**		mg/L	2.0	1	09/17/18 21:48	NA	
Ammonia as Nitrogen, undistilled	ASTM D6919-09	**0.0057**		mg/L	0.0050	1	09/19/18 06:51	NA	
Carbon, Total Organic (TOC)	SM 5310 C-2000(2011)	**7.7**		mg/L	2.0	2	09/14/18 11:34	NA	
Chlorophyll A	SM20 10200 H	**7.72**		ug/L	0.64	4	09/25/18 10:35	NA	
Color, True	SM 2120 B-2001(2011)	**75.0**		ColorUnits	5.0	5	09/12/18 14:00	NA	
Nitrate+Nitrite as Nitrogen	353.2	0.0020	U	mg/L	0.0020	1	09/17/18 19:29	NA	
Nitrogen, Total Kjeldahl (TKN)	351.2	**0.53**		mg/L	0.10	1	09/24/18 19:33	09/21/18	
pH of Color Analysis	SM 2120 B-2001(2011)	**6.21**		pH Units	-	5	09/15/18 10:00	NA	*
Phosphorus, Total	365.1	**0.0273**		mg/L	0.0050	1	09/28/18 12:09	09/24/18	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18 15:30
Sample Matrix:	Water	**Date Received:**	09/12/18 09:30
Sample Name:	18LISO62 Diss	**Basis:**	NA
Lab Code:	R1808738-026		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Phosphorus, Dissolved	365.1	**0.0144**	mg/L	0.0050	1	09/28/18 10:38	09/24/18	

QC Summary Forms

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

Metals

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

METALS

-3-

BLANKS

Contract: R1808738

Lab Code: _____ Case No.: _____ SAS No.: _____ SDG NO.: LCI0910

Preparation Blank Matrix (soil/water): WATER

Preparation Blank Concentration Units (ug/L, ppt, or mg/kg): UG/L

Analyte	Initial Calib. Blank ug/L	C	Continuing Calibration Blank ug/L 1	C	2	C	3	C	Preparation Blank	C	M
Arsenic	0.39	U	0.39	U	0.39	U	0.39	U	0.39	U	MS
Iron	13.00	U	13.00	U	13.00	U	13.00	U	13.000	U	P
Manganese	1.70	U	1.70	U	1.70	U	1.70	U	1.700	U	P

Comments:

METALS

-3-

BLANKS

Contract: R1808738

Lab Code: _____ Case No.: _____ SAS No.: _____ SDG NO.: LCI0910

Preparation Blank Matrix (soil/water): WATER

Preparation Blank Concentration Units (ug/L, ppt, or mg/kg): UG/L

Analyte	Initial Calib. Blank ug/L	C	Continuing Calibration Blank ug/L 1	C	2	C	3	C	Preparation Blank	C	M
Arsenic			0.39	U							MS
Iron			13.00	U	13.00	U					P
Manganese			1.70	U	1.70	U					P

Comments:

ALS Environmental

METALS

-5A-

SPIKE SAMPLE RECOVERY

SAMPLE NO.

18LHB312S

Contract: R1808738

Lab Code: _____ Case No.: _____ SAS No.: _____ SDG NO.: LCI0910

Matrix (soil/water): WATER Level (low/med): LOW

% Solids for Sample: 0.0

Concentration Units (ug/L or mg/kg dry weight): UG/L

Analyte	Control Limit %R	Spiked Sample Result (SSR)	C	Sample Result (SR)	C	Spike Added (SA)	%R	Q	M
Iron	70 - 130	2520.00		1600.00		1000.0	92		P
Manganese	70 - 130	1630.00		1180.00		500.0	90		P

Comments:

METALS

-5A-

SPIKE SAMPLE RECOVERY

SAMPLE NO.

18LHB312SD

Contract: R1808738

Lab Code: _____ Case No.: _____ SAS No.: _____ SDG NO.: LCI0910

Matrix (soil/water): WATER Level (low/med): LOW

% Solids for Sample: 0.0

Concentration Units (ug/L or mg/kg dry weight): UG/L

Analyte	Control Limit %R	Spiked Sample Result (SSR)	C	Sample Result (SR)	C	Spike Added (SA)	%R	Q	M
Iron	70 - 130	2490.00		1600.00		1000.0	89		P
Manganese	70 - 130	1610.00		1180.00		500.0	86		P

Comments:

METALS
-6-
DUPLICATES

SAMPLE NO.: 18LHB312SD

Contract: R1808738

Lab Code: _____ Case No.: _____ SAS No.: _____ SDG NO.: LCI0910

Matrix (soil/water): WATER Level (low/med): LOW

% Solids for Sample: 0.0 % Solids for Duplicate: 0.0

Concentration Units (ug/L or mg/kg dry weight): UG/L

Analyte	Control Limit	Sample (S)	C	Duplicate (D)	C	RPD	Q	M
Iron		2520.00		2490.00		1		P
Manganese		1630.00		1610.00		1		P

Comments: _____

METALS

-7-

LABORATORY CONTROL SAMPLE

Contract: R1808738

Lab Code: _____ Case No.: _____ SAS No.: _____ SDG NO.: LCI0910

Solid LCS Source: _____

Aqueous LCS Source: ACCUSTANDARD

Analyte	Aqueous (ug/L			Solid (mg/K				
	True	Found	%R	True	Found	C	Limits	%R
Arsenic	20.0	21.1	106					
Iron	1000	971	97					
Manganese	500	505	101					

Comments: _____

General Chemistry

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	NA
Sample Matrix:	Water	**Date Received:**	NA
Sample Name:	Method Blank	**Basis:**	NA
Lab Code:	R1808738-MB1		

Inorganic Parameters

Analyte Name	Analysis Method	Result		Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Alkalinity, Total as CaCO3	SM 2320 B-1997(2011)	2.0	U	mg/L	2.0	1	09/17/18 18:37	NA	
Ammonia as Nitrogen, undistilled	ASTM D6919-09	0.0050	U	mg/L	0.0050	1	09/18/18 20:25	NA	
Carbon, Dissolved Organic (DOC)	SM 5310 C-2000(2011)	1.0	U	mg/L	1.0	1	09/13/18 13:39	NA	
Carbon, Total Organic (TOC)	SM 5310 C-2000(2011)	1.0	U	mg/L	1.0	1	09/13/18 22:30	NA	
Chlorophyll A	SM20 10200 H	4.0	U	ug/L	4.0	1	09/25/18 10:35	NA	
Color, True	SM 2120 B-2001(2011)	**1.0**		ColorUnits	1.0	1	09/12/18 14:00	NA	
Nitrate+Nitrite as Nitrogen	353.2	0.0020	U	mg/L	0.0020	1	09/17/18 18:50	NA	
Nitrogen, Total Kjeldahl (TKN)	351.2	0.10	U	mg/L	0.10	1	09/24/18 18:25	09/21/18	
Phosphorus, Dissolved	365.1	0.0050	U	mg/L	0.0050	1	09/28/18 10:17	09/24/18	
Phosphorus, Total	365.1	0.0050	U	mg/L	0.0050	1	09/21/18 16:47	09/19/18	
Sulfate	300.0	0.20	U	mg/L	0.20	1	09/17/18 18:51	NA	
UV254	SM 5910 B	**0.00**		cm-1	-	1	09/12/18 22:40	NA	

Analytical Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	NA
Sample Matrix:	Water	**Date Received:**	NA
Sample Name:	Method Blank	**Basis:**	NA
Lab Code:	R1808738-MB2		

Inorganic Parameters

Analyte Name	Analysis Method	Result		Units	MRL	Dil.	Date Analyzed	Date Extracted	Q
Alkalinity, Total as CaCO3	SM 2320 B-1997(2011)	2.0	U	mg/L	2.0	1	09/17/18 20:44	NA	
Ammonia as Nitrogen, undistilled	ASTM D6919-09	0.0050	U	mg/L	0.0050	1	09/19/18 02:50	NA	
Carbon, Dissolved Organic (DOC)	SM 5310 C-2000(2011)	1.0	U	mg/L	1.0	1	09/13/18 22:30	NA	
Carbon, Total Organic (TOC)	SM 5310 C-2000(2011)	1.0	U	mg/L	1.0	1	09/14/18 10:20	NA	
Nitrogen, Total Kjeldahl (TKN)	351.2	0.10	U	mg/L	0.10	1	09/24/18 18:47	09/21/18	
Phosphorus, Total	365.1	0.0050	U	mg/L	0.0050	1	09/28/18 11:45	09/24/18	

ALS Group USA, Corp.
dba ALS Environmental

QA/QC Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18
Sample Matrix:	Water	**Date Received:**	09/12/18
		Date Analyzed:	09/17/18

Duplicate Matrix Spike Summary
Nitrate+Nitrite as Nitrogen

Sample Name:	18LHB309	**Units:**	mg/L
Lab Code:	R1808738-001	**Basis:**	NA
Analysis Method:	353.2		

		Matrix Spike R1808738-001MS			Duplicate Matrix Spike R1808738-001DMS					
Analyte Name	**Sample Result**	**Result**	**Spike Amount**	**% Rec**	**Result**	**Spike Amount**	**% Rec**	**% Rec Limits**	**RPD**	**RPD Limit**
Nitrate+Nitrite as Nitrogen	0.0078	0.521	0.500	103	0.524	0.500	103	75-125	<1	20

Results flagged with an asterisk (*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

Client: New York State DEC
Project: LCI 2018/LCI2018
Sample Matrix: Water

Service Request:R1808738
Date Collected:09/11/18
Date Received:09/12/18
Date Analyzed:09/13/18 - 09/28/18

Duplicate Matrix Spike Summary
General Chemistry Parameters

Sample Name: 18LHB309 Diss
Lab Code: R1808738-002

Units:mg/L
Basis:NA

			Matrix Spike				Duplicate Matrix Spike				
			R1808738-002MS				R1808738-002DMS				
Analyte Name	**Method**	**Sample Result**	**Result**	**Spike Amount**	**% Rec**	**Result**	**Spike Amount**	**% Rec**	**% Rec Limits**	**RPD**	**RPD Limit**
Carbon, Dissolved Organic (DOC)	SM 5310 C-2000(2011)	8.7	20.0	10.0	112	19.5	10.0	108	75-125	2	20
Phosphorus, Dissolved	365.1	0.0271	0.0482	0.0250	85	0.0500	0.0250	92	75-125	4	20

Results flagged with an asterisk (*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18
Sample Matrix:	Water	**Date Received:**	09/12/18
		Date Analyzed:	09/14/18

Duplicate Matrix Spike Summary
Carbon, Total Organic (TOC)

Sample Name:	18LHB305	**Units:**	mg/L
Lab Code:	R1808738-007	**Basis:**	NA
Analysis Method:	SM 5310 C-2000(2011)		

		Matrix Spike R1808738-007MS			Duplicate Matrix Spike R1808738-007DMS					
Analyte Name	**Sample Result**	**Result**	**Spike Amount**	**% Rec**	**Result**	**Spike Amount**	**% Rec**	**% Rec Limits**	**RPD**	**RPD Limit**
Carbon, Total Organic (TOC)	10.0	22.0	10.0	120	21.6	10.0	116	75-125	2	20

Results flagged with an asterisk (*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

ALS Group USA, Corp.
dba ALS Environmental

QA/QC Report

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18
Sample Matrix:	Water	**Date Received:**	09/12/18
		Date Analyzed:	09/14/18

Duplicate Matrix Spike Summary
Carbon, Total Organic (TOC)

Sample Name:	18LHB321	**Units:**	mg/L
Lab Code:	R1808738-011	**Basis:**	NA
Analysis Method:	SM 5310 C-2000(2011)		

		Matrix Spike R1808738-011MS			Duplicate Matrix Spike R1808738-011DMS					
Analyte Name	**Sample Result**	**Result**	**Spike Amount**	**% Rec**	**Result**	**Spike Amount**	**% Rec**	**% Rec Limits**	**RPD**	**RPD Limit**
Carbon, Total Organic (TOC)	6.1	17.3	10.0	113	16.6	10.0	106	75-125	4	20

Results flagged with an asterisk (*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

Client:	New York State DEC	**Service Request:**	R1808738
Project:	LCI 2018/LCI2018	**Date Collected:**	09/11/18
Sample Matrix:	Water	**Date Received:**	09/12/18
		Date Analyzed:	09/24/18
		Date Extracted:	09/21/18

Duplicate Matrix Spike Summary
Nitrogen, Total Kjeldahl (TKN)

Sample Name:	18LHB322	**Units:**	mg/L
Lab Code:	R1808738-013	**Basis:**	NA
Analysis Method:	351.2		
Prep Method:	Method		

		Matrix Spike R1808738-013MS			**Duplicate Matrix Spike** R1808738-013DMS					
Analyte Name	**Sample Result**	**Result**	**Spike Amount**	**% Rec**	**Result**	**Spike Amount**	**% Rec**	**% Rec Limits**	**RPD**	**RPD Limit**
Nitrogen, Total Kjeldahl (TKN)	1.87	4.23	2.50	95	4.14	2.50	91	75-125	2	20

Results flagged with an asterisk (*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

Client: New York State DEC
Project: LCI 2018/LCI2018
Sample Matrix: Water

Service Request:R1808738
Date Collected:09/11/18
Date Received:09/12/18
Date Analyzed:09/17/18 - 09/28/18

Duplicate Matrix Spike Summary
General Chemistry Parameters

Sample Name: 18LISO62
Lab Code: R1808738-025

Units:mg/L
Basis:NA

			Matrix Spike			Duplicate Matrix Spike					
			R1808738-025MS			R1808738-025DMS					
Analyte Name	**Method**	**Sample Result**	**Result**	**Spike Amount**	**% Rec**	**Result**	**Spike Amount**	**% Rec**	**% Rec Limits**	**RPD**	**RPD Limit**
Nitrate+Nitrite as Nitrogen	353.2	0.0020 U	0.523	0.500	105	0.520	0.500	104	75-125	<1	20
Nitrogen, Total Kjeldahl (TKN)	351.2	0.53	2.77	2.50	90	2.74	2.50	89	75-125	<1	20
Phosphorus, Total	365.1	0.0273	0.0510	0.0250	94	0.0508	0.0250	94	75-125	<1	20

Results flagged with an asterisk (*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

ALS Group USA, Corp.

dba ALS Environmental

QA/QC Report

Client:	New York State DEC	**Service Request:**	R1808738
Project	LCI 2018/LCI2018	**Date Collected:**	09/11/18
Sample Matrix:	Water	**Date Received:**	09/12/18
		Date Analyzed:	09/12/18

Replicate Sample Summary

General Chemistry Parameters

Sample Name:	18LHB309	**Units:**	cm-1
Lab Code:	R1808738-001	**Basis:**	NA

Analyte Name	Analysis Method	MRL	Sample Result	Duplicate Sample R1808738-001DUP Result	Average	RPD	RPD Limit
UV254	SM 5910 B	-	0.255	0.257	0.256	<1	20

Results flagged with an asterisk (*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

Client:	New York State DEC	**Service Request:**	R1808738
Project	LCI 2018/LCI2018	**Date Collected:**	09/11/18
Sample Matrix:	Water	**Date Received:**	09/12/18
		Date Analyzed:	09/17/18

Replicate Sample Summary
General Chemistry Parameters

Sample Name:	18LHB321	**Units:**	mg/L
Lab Code:	R1808738-011	**Basis:**	NA

Analyte Name	Analysis Method	MRL	Sample Result	Duplicate Sample R1808738-011DUP Result	Average	RPD	RPD Limit
Alkalinity, Total as CaCO3	SM 2320 B-1997(2011)	2.0	66.4	66.0	66.2	<1	20

Results flagged with an asterisk (*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

Client:	New York State DEC	**Service Request:**	R1808738
Project	LCI 2018/LCI2018	**Date Collected:**	09/11/18
Sample Matrix:	Water	**Date Received:**	09/12/18
		Date Analyzed:	09/17/18

Replicate Sample Summary
General Chemistry Parameters

Sample Name:	18LHB319	**Units:**	mg/L
Lab Code:	R1808738-015	**Basis:**	NA

Analyte Name	Analysis Method	MRL	Sample Result	Duplicate Sample R1808738-015DUP Result	Average	RPD	RPD Limit
Alkalinity, Total as CaCO3	SM 2320 B-1997(2011)	2.0	66.0	66.0	66.0	<1	20

Results flagged with an asterisk (*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

Client:	New York State DEC	**Service Request:**	R1808738
Project	LCI 2018/LCI2018	**Date Collected:**	09/11/18
Sample Matrix:	Water	**Date Received:**	09/12/18
		Date Analyzed:	09/12/18

Replicate Sample Summary
General Chemistry Parameters

Sample Name:	18LISO62	**Units:**	ColorUnits
Lab Code:	R1808738-025	**Basis:**	NA

Analyte Name	Analysis Method	MRL	Sample Result	Duplicate Sample R1808738-025DUP Result	Average	RPD	RPD Limit
Color, True	SM 2120 B-2001(2011)	5.0	75.0	75.0	75.0	<1	5

Results flagged with an asterisk (*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

Client:	New York State DEC	**Service Request:**	R1808738
Project	LCI 2018/LCI2018	**Date Collected:**	09/11/18
Sample Matrix:	Water	**Date Received:**	09/12/18
		Date Analyzed:	09/15/18

Replicate Sample Summary
General Chemistry Parameters

Sample Name:	18LISO62	**Units:**	pH Units
Lab Code:	R1808738-025	**Basis:**	NA

Analyte Name	Analysis Method	MRL	Sample Result	Duplicate Sample R1808738-025DUP Result	Average	RPD	RPD Limit
pH of Color Analysis	SM 2120 B-2001(2011)	-	6.21	6.21	6.21	<1	20

Results flagged with an asterisk (*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

Client: New York State DEC
Project: LCI 2018/LCI2018
Sample Matrix: Water

Service Request: R1808738
Date Analyzed: 09/13/18 - 09/28/18

Lab Control Sample Summary
General Chemistry Parameters

Units:mg/L
Basis:NA

Lab Control Sample
R1808738-LCS1

Analyte Name	Analytical Method	Result	Spike Amount	% Rec	% Rec Limits
Alkalinity, Total as CaCO3	SM 2320 B-1997(2011)	20.8	20.0	104	70-130
Ammonia as Nitrogen, undistilled	ASTM D6919-09	0.529	0.500	106	70-130
Carbon, Dissolved Organic (DOC)	SM 5310 C-2000(2011)	10.1	10.0	101	70-130
Carbon, Total Organic (TOC)	SM 5310 C-2000(2011)	9.62	10.0	96	70-130
Nitrate+Nitrite as Nitrogen	353.2	0.522	0.500	104	70-130
Nitrogen, Total Kjeldahl (TKN)	351.2	2.31	2.50	92	70-130
Phosphorus, Dissolved	365.1	0.0227	0.0250	91	70-130
Phosphorus, Total	365.1	0.0247	0.0250	99	70-130
Sulfate	300.0	1.92	2.00	96	70-130

Client: New York State DEC
Project: LCI 2018/LCI2018
Sample Matrix: Water

Service Request: R1808738
Date Analyzed: 09/14/18 - 09/28/18

Lab Control Sample Summary
General Chemistry Parameters

Units:mg/L
Basis:NA

Lab Control Sample
R1808738-LCS2

Analyte Name	Analytical Method	Result	Spike Amount	% Rec	% Rec Limits
Alkalinity, Total as CaCO3	SM 2320 B-1997(2011)	20.8	20.0	104	70-130
Ammonia as Nitrogen, undistilled	ASTM D6919-09	0.500	0.500	100	70-130
Carbon, Dissolved Organic (DOC)	SM 5310 C-2000(2011)	9.6	10.0	96	70-130
Carbon, Total Organic (TOC)	SM 5310 C-2000(2011)	10.1	10.0	101	70-130
Nitrogen, Total Kjeldahl (TKN)	351.2	2.34	2.50	94	70-130
Phosphorus, Total	365.1	0.0242	0.0250	97	70-130