

Four clinical ECG sources

PTB-XL | Georgia | Chapman | Ningbo

77,333 records -> 66,379 retained, 13 harmonised labels

Native 500 Hz ECG

quality indicators measured
before decimation

Anti-aliased resampling to 100 Hz

model input: 12 x 1000

RQ1 Generalisation heterogeneity

leave-one-source-out: train on three sources, test on the
fourth, four rotations. Source, architecture and
interaction effects on external AUPRC

held-out predictions

RQ2 Characterising transfer failure

diagnosis-level performance, prevalence shift,
labelling convention and signal quality

externally evaluated models

RQ3 Reliability under transfer

does confidence rank errors consistently
across the same unseen sources?

quality descriptors