

Four clinical ECG sources

PTB-XL | Georgia | Chapman | Ningbo

77,333 records -> 66,379 retained, 13 harmonised labels

Native 500 Hz ECG

quality indicators measured
before decimation

Anti-aliased resampling to 100 Hz

model input: 12 x 1000

RQ1 External generalisation

LOSO evaluation of six compact encoders

held-out predictions

RQ2 Reliability under source and quality shift

quality characterisation, dAUPRC, and selective reliability

externally evaluated models

RQ3 Reliability-preserving embedded deployment

int8 preservation and STM32F411 operating points

quality descriptors