

# Methods and Applications in Fluorescence



## PAPER

### OPEN ACCESS

RECEIVED  
17 December 2019

REVISED  
3 March 2020

ACCEPTED FOR PUBLICATION  
1 April 2020

PUBLISHED  
14 April 2020

Original content from this work may be used under the terms of the [Creative Commons Attribution 4.0 licence](#).

Any further distribution of this work must maintain attribution to the author(s) and the title of the work, journal citation and DOI.



## Resolution of 4 components in the same pixel in FLIM images using the phasor approach

Alexander Vallmitjana<sup>1,4</sup> , Alexander Dvornikov<sup>1,4</sup>, Belen Torrado<sup>1,4</sup> , David M Jameson<sup>2</sup> , Suman Ranjit<sup>1,3,5</sup> and Enrico Gratton<sup>1,5</sup>

<sup>1</sup> Laboratory for Fluorescence Dynamics, Department of Biomedical Engineering, University of California, Irvine, CA, United States of America

<sup>2</sup> Department of Cell and Molecular Biology, University of Hawaii at Manoa, Hawaii, HI, United States of America

<sup>3</sup> Department of Biochemistry and Molecular & Cellular Biology, Georgetown University, Washington DC, United States of America

<sup>4</sup> These authors contributed equally to this work.

<sup>5</sup> Authors to whom any correspondence should be addressed.

E-mail: [suman.ranjit@georgetown.edu](mailto:suman.ranjit@georgetown.edu) and [egratton22@gmail.com](mailto:egratton22@gmail.com)

**Keywords:** fluorescence, phasor plot, multicomponents

### Abstract

In several cellular systems, the phasor FLIM approach has shown the existence of more than 2 components in the same pixel, a typical example being free and bound NADH. In order to properly quantify the concentrations and the spatial distributions of fluorescence components associated with different molecular species we developed a general method to resolve 3 and 4 components in the same pixel using the phasor approach. The method is based on the law of linear combination of components valid after transformation of the decay curves to phasors for each pixel in the image. In principle, the linear combination rule is valid for an arbitrary number of components. For 3 components we use only the phasor position for the first harmonic, which has a small error, while for 4 components we need the phasor location at higher harmonics that have intrinsically more noise. As a result of the noise in the higher harmonics, caused by limited photon statistics, we are able to use linear algebra to resolve 4 components given the position of the phasors of 4 independent components in mixtures of dyes and 3 components for dyes in cellular systems.

### 1. Introduction

Fluorescence lifetime imaging microscopy (FLIM) is a popular imaging technique increasingly applied in diverse areas of biological research [1–5]. However, quantitative multicomponent analysis of fluorescence lifetime data in the same pixel is not widely used. Yet this analysis could enable one to separate different components in the same pixel and quantify their amounts as chemical species change with time. In this case, the time range available will depend on the time needed to acquire a FLIM image with sufficient precision to perform analysis of multiple components in the same pixel. This concept is different from the common method based on the phasor approach to identify components that are spatially separated in the FLIM image. Here we focus on a general method to quantify 3 and 4 components in the same pixel. In other words, we aim at producing an image of the different components rather than a spatially resolved image of different lifetime values.

In many research labs, FLIM imaging is done using the popular time-domain instrumentation which is mostly limited to analysis of two components—either by exponential fitting of fluorescence decay or by phasor analysis [3–10]. Resolving multiple components in the same pixel from data collected in the time correlated single photon counting (TCSPC) approach involves fitting the intensity decays with multiple exponentials and then associating the exponential coefficients to their molecular fractions. This fitting approach is difficult to perform when the photon statistics are poor and when the lifetime components are in a similar time range. However, in the phasor method which can be applied to time-domain as well as frequency-domain data, this particular difficulty of the time-domain approach can be, in part, avoided. In this work we consider the case in which about 50 to 100 photons are collected in a typical pixel and show that the phasor approach can provide a natural and robust way to resolve multi-components (3 and 4) in a

single pixel [11]. To demonstrate our point, we will make use of the linear combination property of phasors. This approach simplifies the problem when the decay of a molecular species is not a single exponential, which occurs frequently with fluorescent proteins or for autofluorescence in which molecular species can decay as a sum of multiple exponentials. In such cases, a molecular species decays with a multi-exponential decay; but still it has a unique phasor that can be used to identify that molecular species.

For this analysis we assume that we know the phasor positions of individual molecular species either because we can measure them in independent samples or because there are identifiable pixels in an image where we have a pure chemical species (though not necessarily a single exponential decay). When such is the case, the application of the rule of linear combination of phasors reduces the problem to finding the parameters of the linear combination that represent the fractional intensity of each component at any given pixel. This rule assumes that the molecular species are not interacting. If they interact, then the linear combination rule will not apply and molecular species with different phasor values should appear in the analysis and their positions can be identified in the phasor plot and then in the image [12].

In this article we first describe the rule of linear combination of phasors. Then we use simulations to determine how the amount of noise in a pixel affects the resolution of the phasor in the same pixel for 3 and 4 components. Next we perform measurements of mixture of dyes of different chemical compositions and lifetimes to verify that the results of the simulation are reflected in the actual measurements. This comparison will show whether there are important sources of noise in the data that have been neglected in the simulations. Finally, we analyze biological samples which presumably could contain pixels with 3 components contributing to the overall fluorescence intensity in a pixel [8]. In our experiments we use dyes that paint the DNA of cells with the idea that in the same pixel we may observe the emission of several dyes. In this case we use 3 different dyes of different lifetimes and calculate, in each pixel, the relative contribution of the dyes to the overall fluorescence. We tried using 4 dyes to paint the nucleus of cells, but we have no evidence that we could have 4 different lifetimes in the same pixel. Instead, in the case of solutions we were sure to have 4 lifetime components in the same pixel.

## 2. Methods

### 2.1. Algorithms and mathematics

This following section describes the algorithms and equations used in this work. These equations have been published previously and are repeated here for the reader's convenience [11]. In the phasor approach, the fluorescence and harmonic signals collected from each point of the image are transformed to the phasor

space using the following transformations [6, 8].

$$g_{i,j}(\omega) = \int_0^T I_{i,j}(t) \cdot \cos(n\omega t) dt / \int_0^T I_{i,j}(t) dt \quad (1)$$

$$s_{i,j}(\omega) = \int_0^T I_{i,j}(t) \cdot \sin(n\omega t) dt / \int_0^T I_{i,j}(t) dt \quad (2)$$

where,  $I_{i,j}(t)$  is the intensity decay,  $g_{i,j}(\omega)$  and  $s_{i,j}(\omega)$  are the X and Y coordinates in the phasor plot, respectively;  $\omega$  is the angular repetition associated with the frequency ( $f$ ) of the laser ( $\omega = 2\pi f$ ),  $T$  is the period of the laser repetition frequency and  $n$  is the harmonic frequency;  $i, j$  are the X and Y coordinates of a pixel of the original microscopy image, respectively. In frequency-domain measurements, the transformation to the phasor plot uses the following relations,

$$g_{i,j}(\omega) = m_{i,j} \cdot \cos(\phi_{i,j}) \quad (3)$$

$$s_{i,j}(\omega) = m_{i,j} \cdot \sin(\phi_{i,j}) \quad (4)$$

where,  $g_{i,j}(\omega)$  and  $s_{i,j}(\omega)$  are again the X and Y coordinates of the phasor plot, respectively, and  $m_{i,j}$  and  $\phi_{i,j}$  are the modulation and phase at the image pixel  $i, j$ . The longer lifetime is represented by the larger phase angle in the phasor plot. The distribution of phasor points originating from FLIM measurements appears on (for mono-exponential decays, points 1, 2, 3 and 4 in figure 1) or inside (for multi-exponential decays) the universal circle (figure 1). The linear combinations are shown by the blue lines (figure 1). If each component has a multi-exponential decay, its location will be represented by phasors not in the universal circle, but the principle of linear combination remains valid.

For decays described by sum of exponentials (2 in the derivation below), the intensity decay can be expressed as,

$$I(t) = A_1 \cdot e^{-t/\tau_1} + A_2 \cdot e^{-t/\tau_2} \quad (5)$$

The calculated phasor coordinates from this decay are given by,

$$g(\omega) = \left( A_1 \tau_1 \frac{1}{1 + (\omega \tau_1)^2} + A_2 \tau_2 \frac{1}{1 + (\omega \tau_2)^2} \right) / (A_1 \tau_1 + A_2 \tau_2) \quad (6)$$

$$s(\omega) = \left( A_1 \tau_1 \frac{\omega \tau_1}{1 + (\omega \tau_1)^2} + A_2 \tau_2 \frac{\omega \tau_2}{1 + (\omega \tau_2)^2} \right) / (A_1 \tau_1 + A_2 \tau_2) \quad (7)$$

These expressions can be simplified using the definition of fractional intensity:


$$f_n = A_n \cdot \tau_n / \sum_n A_n \cdot \tau_n \quad (8)$$

The phasor coordinates of the mixture are given by,

$$g(\omega) = f_1 \frac{1}{1 + (\omega \tau_1)^2} + f_2 \frac{1}{1 + (\omega \tau_2)^2} \quad (9)$$

$$s(\omega) = f_1 \frac{\omega \tau_1}{1 + (\omega \tau_1)^2} + f_2 \frac{\omega \tau_2}{1 + (\omega \tau_2)^2} \quad (10)$$

For multiple exponential decays the law of phasor addition holds and thus,



$$g(\omega) = \sum_n f_n \cdot g_n(\omega) \quad (11)$$

$$s(\omega) = \sum_n f_n \cdot s_n(\omega) \quad (12)$$

The fractions are normalized using the following equation:

$$\sum_n f_n = 1 \quad (13)$$

Equations (11)–(13) are the system of equations that are used in this paper to solve for the unknown fractions  $f_n$ .

## 2.2. Brief introduction to phasors

The final results (equations (11)–(13)) give the law of phasor addition, which was originally written by G. Weber [13]. Although Weber did not use the term phasor nor did he propose a graphical-geometrical interpretation, his work inspired the modern phasor approach. The modern phasor plot was first described in the review by Jameson *et al* [14] and the vector addition property of phasors was first used to correct phase and modulation (frequency domain) data for background signals [15]. The field was dormant for some time, although Brand's group implemented the phasor approach to following calcium binding [16]; they did not, however, use the term 'phasor' to describe their method. The new millennium witnessed a resurgence of the phasor approach. In 2004, Clayton, Hanley and Vermeer [17] discussed the application of the graphical phasor approach (not termed as such) using a single light modulation frequency in FLIM systems. In this work the plots shown had the axes labeled as A and B. In 2005 Hanley and Clayton followed up this work with a paper formally naming their graphs AB plots [18]. They applied this approach to cell data, using FLIM, to study the phosphorylation of the epidermal growth factor. In an important paper in 2005 [19], Redford and Clegg carefully reviewed the mathematics underlying the phasor approach, which they termed 'polar plots', and pointed out the advantages of such plots in FLIM (although they did not work with cells but rather carried out work *in vitro* in a fast mixing chamber). In 2007, Colyer *et al* [20], described a mathematical model and physical implementation for a low cost digital frequency

domain FLIM (DFD-FLIM) system. They also presented an error analysis that showed the precise parameters for maximizing the quality of lifetime acquisition, and presented phasor plots on some systems. In 2008, Digman *et al* [6] applied the phasor approach to FLIM studies on live cells with particular attention to its use in dealing with autofluorescence as well as FRET. Since these early works, dozens of FLIM studies on cells have utilized the phasor approach (see for example [1, 3, 8, 11, 21–24], to list just a few). In addition to FLIM work on cells, numerous papers have appeared using phasors to study various systems in cuvettes including studies on pyrene [25], fluorescent beads [24], protein systems [26–28], nucleic acids [29] and membranes [30, 31].

## 2.3. The phasor approach to resolving multiple species in the same pixel

In the case of 3 and 4 components, there exists a unique exact algebraic solution to the system of equations (11)–(13) which provides the three (or four) fractions ( $f_1, f_2$  and  $f_3$ ) for every point in the phasor plot under reasonable experimental error commonly obtained in FLIM [20] analysis. Hence, we focus here on the experimental resolution of 4 components in the same pixel. If the phasor point is displaced from the expected solution due to noise, we will still get the fractional intensity contribution of the components that give the specific point in the phasor plot. Notice that this observation simply means that we are propagating the error of the phasor location (due to statistics) to the components fractional intensities. This procedure can be done by error propagation rules in the linear system we are considering here, no matter the number of components, provided that these components are independent. In any case this approach is a simple algebra problem. The 4-component case will give a quadrilateral shape in the phasor plot (figure 1, identified by phasor points 1, 2, 3 and 4). In this case the unknowns are the 4 fractional components. Equations (11)–(13) and the system of equations cannot be resolved since we have 3 equations and 4 unknowns. To obtain more than 4 parameters we make use of the

phasor transformation into higher harmonics, with each harmonic producing a new pair of equations (the equations for G and S for the selected harmonics). We will use only the first harmonic for 3 component analysis. Using the second harmonic for 4 component analysis, the system of equation is determined for 4 fractional intensities and can be solve by linear algebra methods.

The 4 equations for the linear combination of harmonics are written below:

$$f_1 g_1^{h1} + f_2 g_2^{h1} + f_3 g_3^{h1} + f_4 g_4^{h1} = G^{h1} \quad (14)$$

$$f_1 s_1^{h1} + f_2 s_2^{h1} + f_3 s_3^{h1} + f_4 s_4^{h1} = S^{h1} \quad (15)$$

$$f_1 g_1^{h2} + f_2 g_2^{h2} + f_3 g_3^{h2} + f_4 g_4^{h2} = G^{h2} \quad (16)$$

$$f_1 s_1^{h2} + f_2 s_2^{h2} + f_3 s_3^{h2} + f_4 s_4^{h2} = S^{h2} \quad (17)$$

Where  $f_1$ ,  $f_2$ ,  $f_3$  and  $f_4$  are the unknown fractional intensities to be retrieved,  $g_n^{h1}$  and  $s_n^{h1}$  are the S and G components of the first harmonic for the 4 phasor positions of the pure components and  $g_n^{h2}$  and  $s_n^{h2}$  are the S and G components of the second harmonic, respectively. The terms in the right side of equations (14)–(17),  $G^{h1}$ ,  $S^{h1}$ ,  $G^{h2}$  and  $S^{h2}$  are the harmonics experimentally measured first and second harmonics at the pixel where we are attempting to find the 4 fractions  $f_n$ . Note that this simple linear algebra system could not be solved if the harmonics were not orthogonal to one another. Since phasor coordinates of higher orders are orthogonal, the new pair of equations for each harmonics provides a solvable system, since it would be a linear combination and the determinant of the system would be different from zero. In other words, the system of equations (14)–(17), is determined using linear algebra methods to give the 4 fractional intensities  $f_n$ . The algorithm is applied for each pixel of the image. Of course we assume that the harmonics of the pure components can be measured with a given precision. In the following we study, using simulations, the effect of noise in a pixel and its' effect on the precision of the fractional intensities at the pixel.

The algorithm we describe is super-fast. An image  $256 \times 256$  can be resolved in 4 components in less than 1 s.

#### 2.4. Preparation of the 4-components dye mixtures.

9(10H)-Acridanone, Rhodamine 110, NADH, POPOP and DMF were purchased from Sigma-Aldrich (St. Louis, MO). Solutions were prepared by dissolving the solid in 1:1 mixture of DMF and water and diluting them to the proper concentration ( $\sim 10$ – $20 \mu\text{M}$ ) to yield approximately the same signal level. The individual dye solutions were then mixed together to create mixtures used for the experiments. These 4 dyes display a phasor value close to the universal semicircle under our experimental conditions. The fluorescence lifetimes were measured to be 12, 3.8, 1.4 and 0.6 ns for 9(10H)-Acridanone, Rhodamine 110, POPOP and NADH respectively.

Data acquisition and analysis was performed using the Globals for Images-SimFCS software (G-Soft Inc. Champaign, IL-USA) and the DIVER microscope described in [30]. The microscope is connected to a 320 Fast-FLIMbox (ISS, Inc., Champaign-Illinois, USA) to allow FLIM measurements and phasor analysis. The FLIMbox utilizes Spartan 6 FPGA and allows multifrequency data analysis up to 8-th harmonics with a fundamental frequency of 80 MHz. The description of the electronics used in the FLIMbox was previously presented (Colyer *et al* [20] 20). This electronics makes use of the photon counting frequency-domain method. The output consists of a series of harmonics of the decay. In this work we use the first and second harmonics, but higher harmonics are provided by the electronics up to about 7 harmonics. As shown in Colyer *et al* (Ref [20]), the error for each of the coordinate for each of the harmonics is proportional to the inverse square-root of the number of photons in each phasor harmonics.

#### 2.5. Cell culture and staining with nuclear probes


NIH3T3 cells were cultured in Dulbecco's Modified Eagle Medium (DMEM, Thermo Fisher Scientific) supplemented with 10% Fetal Bovine Serum (FBS, Sigma-Aldrich), penicillin ( $100 \text{ U ml}^{-1}$ ) and streptomycin ( $100 \mu\text{g ml}^{-1}$ ) at  $37^\circ\text{C}$  in a 5%  $\text{CO}_2$  incubator. Twenty four hours prior to FLIM imaging, the cells were plated in fibronectin (Sigma-Aldrich) covered glass bottom dishes (Mattek). To improve dye staining in live cells, we permeabilized the cells with 0.1% TritonX-100 for 10 min at  $4^\circ\text{C}$ , washing twice before and after the treatment with 1X Dulbecco's Phosphate Buffer Solution (DPBS) [31]. Triple staining of the cells was performed sequentially. First, cells were incubated 5 min at  $37^\circ\text{C}$  with Rose Bengal (Sigma-Aldrich, dye content  $\geq 80\%$ , working solution  $10 \mu\text{g ml}^{-1}$ ), and washed twice with DPBS. After, cells were incubated at least 5 min with Acridine Orange (Sigma-Aldrich,  $13 \text{ ng ml}^{-1}$ ) and NucBlue™ Live ReadyProbes™ Reagent (Thermo Fisher Scientific, working solution was a 1/30th dilution of manufacturer's recommended 1 drop/1.5 ml). All solutions were prepared in DMEM phenol-free, high glucose, HEPES medium (Thermo Fisher Scientific).

### 3. Results

#### 3.1. Simulations of a mixture of 4 components

Simulation of 22 randomly selected points in the phasor plot were performed for 4 exponential components (figure 4). Only 4 points were in the universal semicircle (the pure species), the other 18 where in the interior of the semicircle. The 4 points in the semicircle have lifetimes of 12.5 ns, 4 ns, 1.8 ns and 0.4 ns. These values match the experimental mixtures describe in the 'mixtures' section.





**Figure 2.** Simulated data with noise added (0.1 Gaussian deviation for the  $g$  and  $s$  coordinates) for the recovery of 4 of the fractional intensity of 4 components with different amounts of each component. The average of the fraction is recovered exactly, but the average deviation depends on the added noise. Blue diamonds indicate the relation between the level of noise of the  $G$  and  $S$  coordinates and the deviation of the recovered fractions for simulation of 3 components. The red squares are the averaged deviations for the recovery of the fraction of 4 components. The equation for the lines are given in the figure.


We simulated images of  $256 \times 256$  pixels with an average intensity of 50 counts at each point of the image (figure 2(A)). We simulated a total of 21 images ( $256 \times 256$  pixels) with the fractional intensities, equal across the whole image (table 1). The values shown in the table are before normalization.

For each point generated in the simulations we get values of  $G$  and  $S$  at two harmonics (80 and 160 MHz in our example). For the simulations we know the corresponding fractional intensities of the components that have generated the point. We then added noise to the  $G$  and  $S$  phasor coordinates of the point according to a Gaussian distribution with a given standard deviation, gradually increasing the amount of noise as shown in figure 2(B). As shown in Colyer *et al* [20], this is the standard deviation of the noise affecting each pixel of each harmonics evaluated. Hence, each of the pixels in the  $256 \times 256$  image has random noise added to the phasor coordinates  $G$  and  $S$ . We then apply the linear algebra algorithm that recovers the fractional intensities. Our expectation is that increasing the noise on the phasor components will decrease the precision of the recovered fractional intensities as shown in figure 2(B). For each simulated point of the image we measure the mean deviation of the recovered value of the fractional intensities with respect to the input values of the fractional intensities

**Table 1.** Recovered values for intensity fractions for the 21 images tested.

	$f1$	$f2$	$f3$	$f4$	
Image 1	0.1	0.1	0.1	0.1	(all equal fractions)
Image 2	0.2	0.1	0.1	0.1	(fraction 1 is increasing)
Image 3	0.4	0.1	0.1	0.1	
Image 4	0.6	0.1	0.1	0.1	
Image 5	0.8	0.1	0.1	0.1	
Image 6	0.1	0.2	0.1	0.1	(fraction 2 is increasing)
Image 7	0.1	0.4	0.1	0.1	
Image 8	0.1	0.6	0.1	0.1	
Image 9	0.1	0.8	0.1	0.1	
Image 10	0.1	0.1	0.2	0.1	(fraction 3 is increasing)
Image 11	0.1	0.1	0.4	0.1	
Image 12	0.1	0.1	0.6	0.1	
Image 13	0.1	0.1	0.8	0.1	
Image 14	0.1	0.1	0.1	0.2	(fraction 4 is increasing)
Image 15	0.1	0.1	0.1	0.4	
Image 16	0.1	0.1	0.1	0.6	
Image 17	0.1	0.1	0.1	0.8	
Image 18	1	0.1	0.1	0.1	(One fraction is 1)
Image 19	0.1	1	0.1	0.1	
Image 20	0.1	0.1	1	0.1	
Image 21	0.1	0.1	0.1	1	

used to calculate a given point. The simulation was done for a system composed of 3 and 4 components. The results are shown in figure 2(B), where the mean standard deviation of the recovered fractional



intensities is plotted against the noise added in units of G and S. The blue points are for 3-component simulation and the red points are for the 4-component simulation. For a relatively large range of errors the average error of the recovered fractional intensities is linear and, of course, the error in the fraction recovered is smaller for 3 components than for 4 components. An example of simulated data with noise is shown in figure 2(A) where the noise level was 0.1 (Gaussian with unity standard deviation) in the G and S coordinates for each harmonics for a 4 components simulation.

### 3.2. Simulation of linear combination of 4 components

The simulated data according the above linear combinations are shown in figure 3(A). For each image and each component, we show the difference between the input data and the recovered data in figure 3(A). All data are recovered relatively well. We note that this is a 4 component analysis and that the original image contains very few counts.

For the simulated mixture of four components noise was added to the value of G and S as shown in figure 3(A). In each panel of figure 3(B), the X axis is the simulated fractional intensity of each component and the Y axis is the recovered fractional intensity using the proposed algorithm. The correlation coefficient of each plot is always above 0.94 and the maximum deviation is less than 10%. Note the maximal deviation is obtained for the fraction F3 as well as the lowest correlation coefficient (0.94). The correlation coefficient as well as the maximum deviation is largely independent on the data noise when noise is added to the data from 0.02 (in terms of coordinates g and s) to 0.16 as shown in figure 3(A). The lack of apparent dependence on noise is due to the way data are simulated; in this case  $256 \times 256$  pixels are simulated and then the phasor is averaged after the calculation for the

whole image. If we had simulated a smaller number of pixels, such as ROI of  $16 \times 16$ , the noise will have become more important.

### 3.3. Experimental confirmation of the simulations using mixtures of pure components


We then constructed a system of these 4 dyes that we mixed to reproduce the conditions used for the simulations. The results are shown in figure 4. The mixture at the center of the 4 green lines was obtained adding 4 components of equal fractional intensity. The individual components were added to the mixture until the solution contained one pure component. These points, which are on the universal semicircle, are indicated with a colored circle in in figure 4.

### 3.4. Result of components analysis for the biological samples


We measured a total of 47 cells. The phasor plot of all cells together is shown in figure 5(A). The fractional intensity of component 1: Acridine Orange (red circle), component 2: NucBlue (green circle) and component 3: Rose Bengal (light blue circle) for dyes with only one component present are shown in figure 5(B). Representative images of 3 cells with the addition of 3 dyes (Acridine Orange F1, NucBlue F2 and Rose Bengal F3) are shown in figure 5(C). All these dyes stain the nucleus and make it fluorescent. Figure 5(D) shows the results from the three component analysis of 16 cells out of 47 measured and in each case the fractional intensity calculations can be obtained.

## 4. Conclusions

In this paper we present a simple algebraic algorithm to resolve the fraction of four components in mixtures of dyes and of 3 components for dyes in cells. The



**Figure 4.** Experimental data of solutions of 22 mixtures of 4 components at different relative concentrations. The central point (crossing of the green lines) is the mixture of 4 component. This mixture was diluted by adding one of the four components. The color circles indicate the pure single exponential species on the universal semicircle. 9(10H)-Acridanone, Rhodamine 110, POPOP, and NADH phasor positions are shown by red, green, blue and orange circles, respectively. The phasor plot was measured at 80 and 160 MHz. The size of the clouds indicates the deviation of the measured image of the solution. Each image has  $256 \times 256$  pixels. The average intensity in a pixel as about 50 counts and 10 frames were acquired per each sample.



**Figure 5.** (A). Phasor plot of the 47 cell samples. The colored circles indicate in the phasor plot the phasor position for the first harmonic at 80 MHz of Acridine Orange (red circle), NucBlue (green circle) and Rose Bengal (light blue circle). (B). The average decomposition of the 9 cell samples with only one component using the 3 component analysis. The color code in this figure corresponds to the color used for the 3 components in figure 5(A). (C). Representative images of 3 cells with the addition of 3 dyes (Acridine Orange F1, NucBlue F2 and Rose Bengal F3). Mostly, the nucleus of the cell is visible since the dyes stain the nucleus. The normalized recovered images of the fractions of the 3 dyes is shown in columns 2–4 labeled F1, F2 and F3 together with the total intensity in column 1. For each cell of the 3 cells, the fraction average is shown at right. (D). The results from the three component analysis of 16 cells out of 47 measured and in each case the fractional intensity calculations can be obtained for each of the four components.

algorithm is fast. All 47 images were analyzed at the single pixel level in about 50 s. The quantification of relative amount of the dye or any other fluorescent molecules in cells could be used to classify diseases or the response of cells to different stimuli. The distribution of the pixels histograms for each cell is another parameter that can be used to classify cell. Multiplexing several dyes in a sample and determine quantitatively their abundance is a crucial capability for modern diagnostics techniques. The phasor approach in this work adds to the already powerful methods allowed by the phasor analysis in which parts of a sample can be recognized by different lifetimes. This work opens another level of analysis by determining the number and abundance of molecular species in a single pixel.

For the algorithm to work for 4 components one needs data at 2 harmonics, which is a simple matter using the high frequency FLIMbox that has a bandwidth of about 1 GHz. Harmonics 2 and 3 have very little noise as shown by the resolution of mixtures of 4 components. We note that in addition to the lifetime phasor, one may acquire the spectral phasor for the same sample at the same time, which can also be used to resolve multiple components [32].

## Author contributions

The manuscript was written through contributions of all authors. All authors have given approval to the final version of the manuscript.

## Funding sources

Work supported in part by grants NIH P41-GM103540 (EG), NIH P50-GM076516 (EG) and NIH R21-MH119516 (DMJ).

## ORCID iDs

Alexander Vallmitjana  <https://orcid.org/0000-0002-7149-9385>

Belen Torrado  <https://orcid.org/0000-0002-0679-3422>

David M Jameson  <https://orcid.org/0000-0002-9627-123X>

Suman Ranjit  <https://orcid.org/0000-0003-3058-6332>

Enrico Gratton  <https://orcid.org/0000-0002-6450-7391>

## References

- [1] Aguilar-Arnal L, Ranjit S, Stringari C, Orozco-Solis R, Gratton E and Sassone-Corsi P 2016 Spatial dynamics of SIRT1 and the subnuclear distribution of NADH species *Proc. Natl Acad. Sci. USA* **113** 12715–20
- [2] Lakowicz J R, Szmacinski H, Nowaczyk K and Johnson M L 1992 Fluorescence lifetime imaging of free and protein-bound NADH *Proc. Natl Acad. Sci. USA* **89** 1271–5
- [3] Ranjit S, Dobrinskikh E, Montford J, Dvornikov A, Lehman A, Orlicky D J, Nemenoff R, Gratton E, Levi M and Furgeson S 2016 Label-free fluorescence lifetime and second harmonic generation imaging microscopy improves quantification of experimental renal fibrosis *Kidney Int.* **90** 1123–8
- [4] Skala M C, Riching K M, Bird D K, Gendron-Fitzpatrick A, Eickhoff J, Eliceiri K W, Keely P J and Ramanujam N 2007 *In vivo* multiphoton fluorescence lifetime imaging of protein-bound and free nicotinamide adenine dinucleotide in normal and precancerous epithelia *J. Biomed. Opt.* **12** 024014
- [5] Skala M C, Riching K M, Gendron-Fitzpatrick A, Eickhoff J, Eliceiri K W, White J G and Ramanujam N 2007 *In vivo* multiphoton microscopy of NADH and FAD redox states, fluorescence lifetimes, and cellular morphology in precancerous epithelia *Proc. Natl Acad. Sci. USA* **104** 19494–9
- [6] Digman M A, Caiola V R, Zamai M and Gratton E 2008 The phasor approach to fluorescence lifetime imaging analysis *Biophys. J.* **94** L14–6
- [7] Niesner R, Peker B, Schlusche P and Gericke K H 2004 Noniterative biexponential fluorescence lifetime imaging in the investigation of cellular metabolism by means of NAD(P)H autofluorescence *Chem. Phys. Chem.* **5** 1141–9
- [8] Ranjit S, Malacrida L, Stakic M and Gratton E 2019 Determination of the metabolic index using the fluorescence lifetime of free and bound nicotinamide adenine dinucleotide using the phasor approach *J. Biophotonics* **12** e201900156
- [9] Talbot C B, Lagarto J, Warren S, Neil M A, French P M and Dunsby C 2015 Correction approach for delta function convolution model fitting of fluorescence decay data in the case of a monoexponential reference fluorophore *J. Fluoresc.* **25** 1169–82
- [10] Warren S C, Margineanu A, Alibhai D, Kelly D J, Talbot C, Alexandrov Y, Munro I, Katan M, Dunsby C and French P M 2013 Rapid global fitting of large fluorescence lifetime imaging microscopy datasets *PLoS One* **8** e70687
- [11] Ranjit S, Datta R, Dvornikov A and Gratton E 2019 Multicomponent analysis of phasor plot in a single pixel to calculate changes of metabolic trajectory in biological systems *J. Phys. Chem. A* **123** 9865–73
- [12] Digman M and Gratton E 2014 The phasor approach to fluorescence lifetime imaging: exploiting phasor linear properties *Fluorescence Lifetime Spectroscopy and Imaging*. (Boca Raton, FL: CRC Press) 10 pp 235–48 (<https://taylorfrancis.com/books/e/9780429193750/chapters/10.1201/b17018-14>) 9780429193750
- [13] Weber G 1981 Resolution of the fluorescence lifetimes in a heterogeneous system by phase and modulation measurements *J. Phys. Chem.* **85** 949–53
- [14] Jameson D M, Gratton E and Hall R 1984 The measurement and analysis of heterogeneous emissions by multifrequency phase and modulation fluorometry *App. Spec. Rev.* **20** 55–106
- [15] Reinhart G D, Marzola P, Jameson D M and Gratton E 1991 A method for on-line background subtraction in frequency domain fluorometry *J. Fluoresc.* **1** 153–62
- [16] Hirshfield K M, Toptygin D, Packard B S and Brand L 1993 Dynamic fluorescence measurements of two-state systems: applications to calcium-chelating probes *Anal. Bio. Chem.* **209** 209–18
- [17] Clayton A H, Hanley Q S and Verveer P J 2004 Graphical representation and multicomponent analysis of single-frequency fluorescence lifetime imaging microscopy data *J. Microsc.* **213** 1–5
- [18] Hanley Q S and Clayton A H 2005 AB-plot assisted determination of fluorophore mixtures in a fluorescence lifetime microscope using spectra or quenchers *J. Microsc.* **218** (Pt 1) 62–7
- [19] Redford G I and Clegg R M 2005 Polar plot representation for frequency-domain analysis of fluorescence lifetimes *J. Fluoresc.* **15** 805–15
- [20] Colyer R A, Lee C and Gratton E 2008 A novel fluorescence lifetime imaging system that optimizes photon efficiency *Microsc. Res. Tech.* **71** 201–13



- [21] Golfetto O, Hinde E and Gratton E 2013 Laurdan fluorescence lifetime discriminates cholesterol content from changes in fluidity in living cell membranes *Biophys. J.* **104** 1238–47
- [22] Malacrida L and Gratton E 2018 LAURDAN fluorescence and phasor plots reveal the effects of a H<sub>2</sub>O<sub>2</sub> bolus in NIH-3T3 fibroblast membranes dynamics and hydration *Free Radic. Biol. Med.* **128** 144–56
- [23] Sun Y, Day R N and Periasamy A 2011 Investigating protein-protein interactions in living cells using fluorescence lifetime imaging microscopy *Nat. Protoc.* **6** 1324–40
- [24] Leben R, Ostendorf L, van Koppen S, Rakhymzhan A, Hauser A E, Radbruch H and Niesner R A 2018 Phasor-based endogenous NAD(P)H Fluorescence lifetime imaging unravels specific enzymatic activity of neutrophil granulocytes preceding NETosis *Int. J. Mol. Sci.* **19** 1018
- [25] Martelo L, Fedorov A and Berberan-Santos M N 2015 Phasor representation of monomer-excimer kinetics: general results and application to pyrene *J. Phys. Chem. B* **119** 15023–9
- [26] Stefl M, James N G, Ross J A and Jameson D M 2011 Applications of phasors to *in vitro* time-resolved fluorescence measurements *Analytical Bio. Chemistry* **410** 62–9
- [27] James N G, Ross J A, Stefl M and Jameson D M 2011 Applications of phasor plots to *in vitro* protein studies *Analytical Bio. Chemistry* **410** 70–6
- [28] Montecinos-Franjola F, James N G, Concha-Marambio L, Brunet J E, Lagos R, Monasterio O and Jameson D M 2014 Single tryptophan mutants of FtsZ: nucleotide binding/exchange and conformational transitions *Biochim. Biophys. Acta* **1844** 1193–200
- [29] Buscaglia R, Jameson D M and Chaires J B 2012 G-quadruplex structure and stability illuminated by 2-aminopurine phasor plots *Nucleic Acids Res.* **40** 4203–15
- [30] Dvornikov A, Malacrida L and Gratton E 2019 The DIVER microscope for imaging in scattering media *Methods Protoc* **2** 53
- [31] van de Ven A L, Adler-Storthz K and Richards-Kortum R R 2009 Delivery of optical contrast agents using Triton-X100, part 2: enhanced mucosal permeation for the detection of cancer biomarkers *Journal of Biomedical Optics* **14** 1–13
- [32] Dvornikov A and Gratton E 2018 Hyperspectral imaging in highly scattering media by the spectral phasor approach using two filters *Biomed. Opt. Express* **9** 3503–11