


Comparative EIS Kinetics Report

Model Structure: R0-p(R1-p(R2,CPE2),CPE1)

Used data: altered data, first 0 points removed and points with $freq > 1e+05$ and $freq < 1e-01$ removed



Element	Unit	Bounds	Cauvel.I.control_48H	Cauvel.I.control_72H
R_0	Ω	$[1.0e+00, 1.0e+03]$	$4.06e+01 \pm 1.8e-01$	$3.02e+01 \pm 1.2e+01$
R_1	Ω	$[1.0e+00, 1.0e+08]$	$4.65e+03 \pm 1.7e+02$	$1.06e+01 \pm 1.2e+01$
R_2	Ω	$[1.0e+00, 1.0e+08]$	$4.68e+03 \pm 2.9e+03$	$5.73e+03 \pm 3.7e+02$
Q_2	$\Omega^{-1} \text{ sec}^a$	$[1.0e-09, 1.0e-03]$	$8.82e-04 \pm 9.7e-05$	$9.71e-05 \pm 1.6e-05$
n_2	-	$[5.0e-01, 1.0e+00]$	$1.00e+00 \pm 1.6e-01$	$8.48e-01 \pm 1.7e-02$
Q_1	$\Omega^{-1} \text{ sec}^a$	$[1.0e-09, 1.0e-03]$	$1.11e-04 \pm 1.9e-06$	$4.36e-05 \pm 1.1e-05$
n_1	-	$[5.0e-01, 1.0e+00]$	$8.37e-01 \pm 3.4e-03$	$5.00e-01 \pm 1.8e-01$
χ^2	-	-	$5.016e-04$	$7.331e-04$

Analysis Results: 48H




Figure 1: EIS Fit Results for Cauvel.L_control_48H

Analysis Results: 72H




Figure 2: EIS Fit Results for Cauvel.L_control.72H