Title:			
Datasource:	SCE-CHEM-C00759_local	Created:	30/07/2009 15:25:49 by admin
Location:	ICS1100\2_Data\AK_171123.SEQ	Last Update:	17/11/2023 08:45:09 by Dionex

Blank Run Subtraction: No Blank Run Subtraction

Detection Table:

No.	Ret. Time [min]	Param. Name	Param. Value	Channel
1	0.000	Minimum Area	0.01 "[Signal]*min"	All Channels
2	0.000	Inhibit Integration	On	All Channels
3	2.678	Inhibit Integration	Off	All Channels

Title:				
Datasource:	SCE-CHEM-C00759_local		Created:	30/07/2009 15:25:49 by admin
Location:	ICS1100\2_Data\AK_171123.SEQ		Last Update:	17/11/2023 08:45:09 by Dionex

Peak Table:

Use Recently Detected Retention Times: Off
Peak Retention Time Determination: Absolute
Dead time:
Delay Time of 2'nd Detector: <None>
Delay Time of 3'rd Detector: <None>

No.	Peak Name	Ret.Time	Window	Standard	Int.Type	Cal.Type	Peak Type	Group	Amount Calibration 1	Amount Calibration 2
1	Fluoride	3.113 min	0.200 AG	External	Area	LOff	Auto		0.000000	0.200000
2	Chloride	3.670 min	0.218 AG	External	Area	LOff	Auto		0.000000	1.000000
3	Nitrite	4.180 min	0.200 AG	External	Area	LOff	Auto		0.000000	1.000000
4	Bromide	4.760 min	0.150 AG	External	Area	LOff	Auto		0.000000	1.000000
5	Nitrate	5.400 min	0.300 AG	External	Area	LOff	Auto		0.000000	1.000000
6	Phosphate	6.420 min	0.202 AG	External	Area	LOff	Auto		0.000000	2.000000
7	*Sulfate*	*7.180 min*	*0.247 AG*	*External*	*Area*	*LOff*	*Auto*		*0.000000*	*1.000000*

Title:
Datasource: SCE-CHEM-C00759_local   Created: 30/07/2009 15:25:49 by admin
Location: ICS1100\2_Data\AK_171123.SEQ   Last Update: 17/11/2023 08:45:09 by Dionex

Peak Table:

Use Recently Detected Retention Times: Off
Peak Retention Time Determination: Absolute
Dead time:
Delay Time of 2'nd Detector: <None>
Delay Time of 3'rd Detector: <None>

No.	Peak Name	Ret.Time	Amount Calibration 3	Amount Calibration 4	Amount Calibration 5	Amount Calibration 6	Amount Calibration 7	Comment
1	Fluoride	3.113 min	0.400000	1.000000	2.000000	4.000000	20.000000	Autogenerated
2	Chloride	3.670 min	2.000000	5.000000	10.000000	20.000000	100.000000	Autogenerated
3	Nitrite	4.180 min	2.000000	5.000000	10.000000	20.000000	100.000000	Autogenerated
4	Bromide	4.760 min	2.000000	5.000000	10.000000	20.000000	100.000000	Autogenerated
5	Nitrate	5.400 min	2.000000	5.000000	10.000000	20.000000	100.000000	Autogenerated
6	Phosphate	6.420 min	4.000000	10.000000	20.000000	40.000000	200.000000	Autogenerated
7	*Sulfate*	*7.180 min*	*2.000000*	*5.000000*	*10.000000*	*20.000000*	*100.000000*	*Autogenerated*

Title:			
Datasource:	SCE-CHEM-C00759_local	Created:	30/07/2009 15:25:49 by admin
Location:	ICS1100\2_Data\AK_171123.SEQ	Last Update:	17/11/2023 08:45:09 by Dionex

Amount Table:

Dimension of Amounts:
Reference volume for amounts: Use inject volume of first standard
Number of Amount Columns: 7
Sample column used for amount column assignment: Sample Name

No.	Peak Name	Ret.Time	Resp.Fact.	Amount Calibration 1	Amount Calibration 2	Amount Calibration 3	Amount Calibration 4	Amount Calibration 5	Amount Calibration 6
1	Fluoride	3.113 min	1.000000	0.000000	0.200000	0.400000	1.000000	2.000000	4.000000
2	Chloride	3.670 min	1.000000	0.000000	1.000000	2.000000	5.000000	10.000000	20.000000
3	Nitrite	4.180 min	1.000000	0.000000	1.000000	2.000000	5.000000	10.000000	20.000000
4	Bromide	4.760 min	1.000000	0.000000	1.000000	2.000000	5.000000	10.000000	20.000000
5	Nitrate	5.400 min	1.000000	0.000000	1.000000	2.000000	5.000000	10.000000	20.000000
6	Phosphate	6.420 min	1.000000	0.000000	2.000000	4.000000	10.000000	20.000000	40.000000
7	*Sulfate*	*7.180 min*	*1.000000*	*0.000000*	*1.000000*	*2.000000*	*5.000000*	*10.000000*	*20.000000*

Title:

| Datasource: | SCE-CHEM-C00759_local | Created: | 30/07/2009 15:25:49 by admin |
| Location: | ICS1100\2_Data\AK_171123.SEQ | Last Update: | 17/11/2023 08:45:09 by Dionex |

Amount Table:

Dimension of Amounts:
Reference volume for amounts: Use inject volume of first standard
Number of Amount Columns: 7
Sample column used for amount column assignment: Sample Name

No.	Peak Name	Ret.Time	Amount Calibration 7	Comment
1	Fluoride	3.113 min	20.000000	Autogenerated
2	Chloride	3.670 min	100.000000	Autogenerated
3	Nitrite	4.180 min	100.000000	Autogenerated
4	Bromide	4.760 min	100.000000	Autogenerated
5	Nitrate	5.400 min	100.000000	Autogenerated
6	Phosphate	6.420 min	200.000000	Autogenerated
7	*Sulfate*	*7.180 min*	*100.000000*	*Autogenerated*

Title:

| Datasource: | SCE-CHEM-C00759_local | | Created: | 30/07/2009 15:25:49 by admin |
| Location: | ICS1100\2_Data\AK_171123.SEQ | | Last Update: | 17/11/2023 08:45:09 by Dionex |

Calibration:

Calibration Mode: Total
Auto Recalibrate: On
Curve Fitting Model: Normal
Dual-Column Separate Calibration: Off

No.	Enabled	Name	Smp.No.	Pos.	Inj. Vol.	Weight	ISTD Amount	Dil. Factor	Inj. Date/Time
1	▨	Calibration 1	1	99	25.0	1.0000	1.0000	1.0000	17/11/2023 08:19:15
2	▨	Calibration 2	2	100	25.0	1.0000	1.0000	1.0000	17/11/2023 08:29:40
3	⊠	Calibration 3	3	101	25.0	1.0000	1.0000	1.0000	17/11/2023 08:40:05
4	⊠	Calibration 4	4	102	25.0	1.0000	1.0000	1.0000	17/11/2023 08:50:38
5	⊠	Calibration 5	5	103	25.0	1.0000	1.0000	1.0000	17/11/2023 09:01:09
6	⊠	Calibration 6	6	104	25.0	1.0000	1.0000	1.0000	17/11/2023 09:11:45

Title:
Datasource: SCE-CHEM-C00759_local   Created: 30/07/2009 15:25:49 by admin
Location: ICS1100\2_Data\AK_171123.SEQ   Last Update: 17/11/2023 08:45:09 by Dionex

---

Calibration:

Calibration Mode: Total
Auto Recalibrate: On
Curve Fitting Model: Normal
Dual-Column Separate Calibration: Off

No.	Enabled	Name	Sample Comment	Calib. Comment
1	▨	Calibration 1		Ok
2	▨	Calibration 2		Ok
3	⊠	Calibration 3		Ok
4	⊠	Calibration 4		Ok
5	⊠	Calibration 5		Ok
6	⊠	Calibration 6		Ok