

Journal Pre-proof



Quantitative computed tomography analysis for stratifying the severity of Coronavirus Disease 2019

Cong Shen, Nan Yu, Shubo Cai, Jie Zhou, Jieixin Sheng, Kang Liu, Heping Zhou, Youmin Guo, Gang Niu

PII: S2095-1779(20)30226-4

DOI: <https://doi.org/10.1016/j.jpha.2020.03.004>

Reference: JPHA 528

To appear in: *Journal of Pharmaceutical Analysis*

Received Date: 27 February 2020

Revised Date: 4 March 2020

Accepted Date: 4 March 2020

Please cite this article as: C. Shen, N. Yu, S. Cai, J. Zhou, J. Sheng, K. Liu, H. Zhou, Y. Guo, G. Niu, Quantitative computed tomography analysis for stratifying the severity of Coronavirus Disease 2019, *Journal of Pharmaceutical Analysis* (2020), doi: <https://doi.org/10.1016/j.jpha.2020.03.004>.

This is a PDF file of an article that has undergone enhancements after acceptance, such as the addition of a cover page and metadata, and formatting for readability, but it is not yet the definitive version of record. This version will undergo additional copyediting, typesetting and review before it is published in its final form, but we are providing this version to give early visibility of the article. Please note that, during the production process, errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

© 2020 Xi'an Jiaotong University. Production and hosting by Elsevier B.V. All rights reserved.

A Quantitative Computed Tomography Analysis for the severity stratification of Coronavirus Disease 2019

Cong Shen¹, Nan Yu², Shubo Cai³, Jie Zhou³, Jiexin Sheng⁴, Kang Liu⁵, Heping

Zhou⁶, Youmin Guo^{1,*}, Gang Niu^{1,*}

¹Department of Medical Imaging, The First Affiliated Hospital of Xi'an Jiaotong University, Xi'an, 710061, Shaanxi, China

²Department of Radiology, The Affiliated Hospital of Shaanxi University of Traditional Chinese Medicine, 712000, Xianyang, Shaanxi, China

³Department of Radiology, Xi'an Chest Hospital, Xi'an, 710100, Shaanxi, China

⁴Department of Radiology, Hanzhong Central Hospital, Hanzhong, 723000, Shaanxi, China

⁵Department of CT&MR Imaging, Weinan Central Hospital, 714000, Weinan, Shaanxi, China

⁶Department of Radiology, Ankang Central Hospital, Ankang, 725000, Shaanxi, China

*Corresponding Authors: YouminGuo, MD, Department of Medical Imaging, The First Affiliated Hospital of Xi'an Jiaotong University, 277 Yanta West Road, Xi'an, Shaanxi China.

E-mail addresses: guoyoumin163@sina.com


Gang Niu, PhD, Department of Medical Imaging, The First Affiliated Hospital of Xi'an Jiaotong University, 277 Yanta West Road, Xi'an, 710061, Shaanxi, China.

E-mail addresses: niugang369@xjtu.edu.cn


Conflicts of interest statement

The authors report no conflicts of interest.


visually scored by radiologists




Computer-aided detected lesion




Correlations of the lesion percentage evaluated by two scoring systems



Correlations of the lesion density evaluated by two scoring systems



The distribution of lesions evaluated by two scoring systems



Quantitative Computed Tomography Analysis for Stratifying the Severity of Coronavirus Disease 2019

Abstract

Purpose: To examine the feasibility of using a computer tool for stratifying the severity of Coronavirus Disease 2019 (COVID-19) based on computed tomography (CT) images. **Materials and methods:** We retrospectively examined 44 confirmed COVID-19 cases. All cases were evaluated separately by radiologists (visually) and through an in-house computer software. The degree of lesions was visually scored by the radiologist, as follows, for each of the 5 lung lobes: 0, no lesion present; 1, < 1/3 involvement; 2, >1/3 and <2/3 involvement; and 3, > 2/3 involvement. Lesion density was assessed based on the proportion of ground-glass opacity (GGO), consolidation and fibrosis of the lesions. The parameters obtained using the computer tool included lung volume (mL), lesion volume (mL), lesion percentage (%), and mean lesion density (HU) of the whole lung, right lung, left lung, and each lobe. The scores obtained by the radiologists and quantitative results generated by the computer software were tested for correlation. A Chi-square test was used to test the consistency of radiologist- and computer-derived lesion percentage in the right/left lung, upper/lower lobe, and each of the 5 lobes. **Result:** The results showed a strong to moderate correlation between lesion percentage scores obtained by radiologists and the computer software (r ranged from 0.7679 to 0.8373, $P < 0.05$), and a moderate correlation between the proportion of GGO and mean lesion density ($r = -0.5894$, $P < 0.05$), and proportion of consolidation and mean lesion density ($r = 0.6282$, $P < 0.05$). Computer-aided quantification showed a statistical significant higher lesion percentage for lower lobes than that assessed by the radiologists ($\chi^2 = 8.160$, $P = 0.004$). **Conclusions:** Our experiments demonstrated that the computer tool could reliably and accurately assess the severity and distribution of pneumonia on CT scans.

Keywords: quantitative computed tomography (QCT); Coronavirus Disease 2019 (COVID-19); severity stratification

1. Introduction

The Coronavirus Disease 2019 (COVID-19) has spread from Wuhan city, Hubei province, China [1-3] to many other areas in China and to other countries. Radiological examinations, especially thin slice chest computer tomography (CT), have become vital for early diagnosis and the assessment of the disease course [4-6] and have demonstrated excellent performance in visualizing the features of COVID-19. Imaging can reveal the following[5]: (1) number of lesions (often more than three); (2) lesion's size (patchy, large block, nodular, lumpy, etc.); (3) lesion density (ground glass density, paving stones-like change, consolidation, fibrosis, etc.); (4) lesion distribution (sub-pleural or along the bronchial vascular bundles); and (5) other concomitant signs (air-bronchogram, rare pleural effusion and mediastinal lymph node enlargement, etc.). However, it is not easy for a human expert to visually assess the extent of the disease and its progress over time.

An efficient and accurate assessment method is urgently needed due to the rapid increase in computed tomography (CT) examinations and the need for accurate assessment or staging of COVID-19 pneumonia [7, 8]. Based on the number of lung lobes involved (5 lobes, score 1-5 for each lobe, total range: 0 [none] to 25), a recent study that focused on the dynamic changes in COVID-19 pneumonia[9] showed that lung tissue involvement peaked at approximately 10 days from the onset of symptoms. Owing to the rapid increase in the number of patients, repeated examinations, and the rapid progress of this disease, efficient and accurate assessment is warranted. Hence, it is necessary to develop novel solutions to improve diagnosis performance and efficiency.

In this study, we validated the performance of a newly developed computer tool that aims to quantitatively assess COVID-19 pneumonia using CT images by comparing the results obtained by radiologists and the computer tool.

2. Materials and methods

2.1 Patients

In this study, initial CT images from 44 patients with COVID-19 confirmed between January 22, 2020 and February 7, 2020 were reviewed retrospectively. Diagnostic criteria for COVID-19 were based on the diagnosis and treatment protocols from the National Health Commission of the People's Republic of China [10]. The confirmed cases met the following criteria: (1) history of travel to Wuhan and its circumjacent area or other communities with confirmed COVID-19 cases within the past 14 days; (2) contact with others with confirmed COVID-19 (positive result on nucleic acid testing) within the past 14 days; (3) contact with someone from Wuhan and its circumjacent area or other communities wherein fever or respiratory symptoms have been reported within the past 14 days; or (4) onset of symptoms after involvement in a public gathering. Patients with severe respiratory dyspnea (respiratory rate >30 breaths/min), low SpO₂ (<93%) at rest, and PaO₂/FiO₂ ≤ 300 mmHg who required oxygen treatment or mechanical ventilation were defined to have severe disease.

All CT images were assessed by a radiologist and processed with the in-house computer software. The percentage of lesion to lung volume, lesion density, and lesion distribution were evaluated and compared (**Fig. 1**).

The study was approved by the institutional board of each participating center.

2.2 Acquisition of CT images

CT images were collected using multi-detector CT (MX 16, Phillips, Cleveland, Netherlands; BrightSpeed, Siemens, Erlangen, Germany; SOMATOM Perspective, Siemens, Erlangen, Germany; Optima CT680 Series, GE MEDICAL SYSTEMS, America; Aquilion, TOSHIBA, Japan; Emotion 16, Siemens, Erlangen, Germany). All patients were scanned in the supine position with their breath held at the end of inspiration. The field of view was set from the apex to the base of the lungs. The tube voltage and the current were 120 kV and 30–140 mA, respectively. All data were reconstructed using a standard reconstruction kernel. The reconstruction matrix was 512×512, and the slice thickness of reconstructed sections was between 0.625 mm and 1 mm. Images were viewed at window settings optimized for the assessment of

the lung parenchyma (width 1500 HU; level -500 HU).

2.3 Severity evaluation

The severity of pneumonia was evaluated by radiologists and the computer tool separately. The evaluation of the indices of COVID-19 included the degree of lesions, mean lesion density, and the distribution of the lesions.

2.3.1 Severity evaluation by radiologist

As described by Pan et al. [9] the percentage of lesion to lung size was evaluated in particular lobes. This study extended beyond previously described methods [9] by visually scoring each of the 5 lobes (i.e., right upper lobe, right middle lobe, right lower lobe, left upper lobe and left lower lobe) on a scale of 0-3 (0: no lesion, 1: <1/3 of the lobe volume involved, 2: >1/3 and < 2/3 of the lobe volume involved, 3: >2/3 of the lobe volume involved). This was done to decrease inter-observer inconsistency. The total CT score for each case was the sum of the score for the 5 lobes, with a maximum possible score of $5 \times 3 = 15$. Two examples of CT scoring are shown in **Fig. 2**. The radiologist-defined lesion percentages of the whole lung, right upper lobe, right middle lobe, right lower lobe, left upper lobe, and left lower lobe were recorded as RLP_{WL} , RLP_{RUL} , RLP_{RML} , RLP_{RLL} , RLP_{LUL} , and RLP_{LLL} , respectively.

The evaluation of lesion density manually was based on the proportion of three major CT signs, that is, ground glass opacity (GGO), consolidation, and fibrosis, which were judged according to the international standard nomenclature defined by the Fleischner Society glossary [11] and peer-reviewed literature on viral pneumonia[8,12]. The composition of each CT sign was evaluated on a scale of 0-10, with a sum score of 10. The proportion of GGO, consolidation, and fibrosis was recorded as P_{GGO} , $P_{consolidation}$, and $P_{fibrosis}$, respectively.

Two radiologists with more than 5years of thoracic-imaging analysis experience evaluated the severity of images in a double blind manner. Any disagreement between the two radiologists was resolved by another, more experienced, radiologist.

2.3.2 Computerized quantification of lesions

A computerized quantitative approach was used to evaluate the severity of COVID-19. The scheme consisted of four primary phases: (1) segmentation of the lung [13, 14] and 5 lobes [15]; (2) segmentation of the pulmonary vessels [16,17]; (3) subtraction of pulmonary vessels from the lung region; and (4) the detection of pneumonia (**Fig. 3**). The details of the last two steps have been presented elsewhere.

The lesion region was segmented based on thresholds and adaptive region growing [18]. The results of segmentation were reviewed by a radiologist with more than 10 years' experience. False positives were deleted, and false negatives were added manually. The computerized parameters included the lung volume (mL), lesion volume (mL), the ratio of lesion volume to that of the corresponding lung or lobes (%), and mean lesion density (HU) of the whole lung, right lung, left lung, and each of the 5 lobes. The segmentation and three-dimensional reconstruction of COVID-19 lesions at the early stage, the progressing stage and the severe stage are shown in **Fig. 4**.

2.4 Statistical analysis

Statistical analyses were performed using IBM SPSS Statistics Software (version21; IBM, New York, USA). Discrete variables were presented as the number of cases unless otherwise specified. Continuous data were presented as mean \pm standard deviation (minimum-maximum). The independent *t* test was used for the comparison of patients with and without severe disease. Correlations between the radiologist-defined CT score of lesion percentage and lesion percentage quantified using the computer tool, and between the radiologist-defined GGO/consolidation proportion and mean lesion density quantified using the computer tool were evaluated by Pearson correlation. Strong correlation was defined as $r > 0.8$, moderate correlation as $r < 0.8$ and > 0.5 , and mild correlation as $r < 0.5$. A Chi-square test was used to test the consistency of lesion distribution results obtained by the radiologists and the computer software. A *P*-value of < 0.05 was defined as statistically significant for all results.

3. Results

3.1 General lung characteristics of the enrolled patients

We examined a total of 44 patients (21 male and 23 female) in the study (**Table 1**). The CT scores for the whole lung and each individual lobe were significantly larger in those with severe disease than in those without severe disease ($P < 0.05$). The lung volume quantified using the computer tool was significantly lower in those with severe disease than in those without severe disease ($P < 0.05$) for the whole lung, right lung, and right lower lobe. The lesion volume quantified using the computer tool was significantly larger in the group with severe disease ($P < 0.05$). The lesion volume to lung/lobe volume percentage (%) quantified using the computer tool was significantly higher in the group with severe disease ($P < 0.05$); however, this did not hold true for the right middle lobe.

3.2 Correlation between CT scores and parameters quantified using the computer tool

The Pearson correlation analyses showed a moderate to strong correlation between lesion percentage determined by radiologists and the computer software (r ranged from 0.7679 to 0.8373, all $P < 0.05$) (**Fig. 5**). A moderate negative correlation was observed between the proportion of GGO and mean lesion density ($r = -0.5894$, $P < 0.05$), and a moderate positive correlation was observed between the proportion of consolidation and mean lesion density ($r = 0.6282$, $P < 0.05$) (**Fig. 6**).

3.3 Distribution of lesions determined by the radiologist and computer tool

The distribution of lesions in the right and left side of the lungs observed by the radiologists was similar to that obtained using the computer software (**Table 2**, $\chi^2 = 0.988$, $df = 1$, $P = 0.320$). However, the distribution of lower lobe lesions obtained through computer-aided quantification significantly differed from the results obtained by radiologists (**Table 3**, $\chi^2 = 8.160$, $df = 1$, $P = 0.004$). The distribution of lesions among the five lobes observed by the radiologists was similar to that obtained through computer-aided quantification (**Table 4**, $\chi^2 = 8.423$, $df = 4$, $P = 0.077$). The

distribution of lesions was the highest in the right lower lobe, followed by the left lower lobe, left upper lobe, right upper lobe, and right middle lobe.

4. Discussion

In this study, a quantitative CT analysis for stratifying COVID-19 cases by severity was established. The CT scores for the whole lung and each individual lung lobe were significantly higher in severe cases than in non-severe cases. The lung volume of the whole lung, right lung, and the right lower lobe quantified using the computer tool was significantly lower in severe cases than in non-severe cases and the lesion volume quantified using the computer tool was also significantly higher in the severe group. The percentage of the lesion volume to the corresponding lung or lobe volume quantified using the computer tool was significantly higher in the severe group, although this was not true for the right middle lobe. Our results showed a strong or moderate correlation between the lesion percentage obtained by radiologists and the computer software, a negative correlation between the proportion of GGO and mean lesion density, and a moderate positive correlation between the proportion of consolidation and mean lesion density.

The results showed that as pneumonia progresses, the functional lung volume decreases. This was likely caused by the swelling of infected lung tissue and filling of alveoli with exudate, leading to a partial loss of lung function [19]. Because of this, the lesion percentage may be an important biomarker to examine in future studies.

A recent study on viral pneumonia [20] showed that the inter-observer reliability of CT scans was poor to slight when determining the presence of intra-lobular reticulation, distribution of consolidation, and GGO. In this study, the lesions were segmented by a computer tool using a standard algorithm, and only a small part of lesion analysis (false positives and false negatives) was performed by a radiologist. This method made it much easier for us to assess the lesions. CT signs, GGO, consolidation, and fibrosis manifest differently over the course of the disease [21] and have different density ranges. This study shows that the proportion of consolidation increases with mean lesion density, and the proportion of GGOs, decreases with mean

lesion density.

However, the distribution of lower lobe lesions obtained by computer-aided quantification significantly differed from radiologist-derived results ($\chi^2 = 8.160$, df = 1, $P = 0.004$). This could be due to the smaller volume and weight of the upper lobes than of the lower lobes and may have resulted in the under-or overestimation of the volume of lesions in the bilateral lower or upper lobes, respectively. The lesion distribution in the right and left side obtained by the radiologist and computer software showed no significant difference, nor did the distribution of lesions among the five lobes. The severity of the involvement in the 5 lobes was similar to that reported in a recent study [9], which showed that the most commonly involved lung segments, in order, were the dorsal segment of the right lower lobe, the posterior basal segment of the right lower lobe, the lateral basal segment of the right lower lobe, the dorsal segment of the left lower lobe, and the posterior basal segment of the left lower lobe.

The limitations of this study included the retrospective nature of the study, selection bias (a lack of severe COVID-19 cases), small sample size, and evaluation bias in the radiologist-defined CT score. In the future, examining the correlation between quantitative CT parameters and clinical symptoms and laboratory indices would be useful for guiding clinical decision-making.

In summary, computer-aided quantification is an accurate, easy, and feasible way to stratify COVID-2019 cases according to severity.

Fig. 1. Study flowchart

Fig.2. Examples of severity evaluation by a radiologist

(A, B): Images from a 60-year-old female patient with confirmed COVID-19. A pure GGO lesion can be seen in the left upper lobe (A, red arrow). As the lesion occupied < 1/3 of the left upper lobe, the lesion percentage score was 1. Another GGO can be seen in the right lower lobe (B, red arrow); this also had a lesion percentage score of 1. The proportion of GGO, consolidation, and fibrosis were 10, 0, and 0, respectively.

(C, D): Images from a 66-year-old female patient with confirmed COVID-19. GGO (C, red arrow; D, red arrow) and bilateral multifocal consolidation were observed (C, green arrow; D,

green arrow). The scores of GGO, consolidation, and fibrosis were 3, 7, and 0, respectively. GGO, ground glass opacity.

Fig. 3. Illustration of lesion identification

The first step was the segmentation of the bilateral lung, and the results are displayed a three-dimensional model (shown in A, the right lung is colored green and the left lung is colored blue). The second step was the segmentation of pulmonary vessels (shown in B, the vessels are colored blue). After the subtraction of the pulmonary vessels from the lung regions, the fourth and final step was the segmentation of pneumonia lesions (shown in C). The red irregular nodular shapes were observed as a result of this lesion segmentation.

Fig. 4. Illustration of lesion segmentation by the computer software

(A, B, C): Images from a 60 year-old female patient with early stage COVID-19. Multifocal pure ground glass opacity (GGO) can be seen in the bilateral lung (A, red arrows). All GGO lesions were segmented, as shown on the same coronal view (B, within red line). All lesions (C, red nodular) can be seen on the three-dimensional reconstruction view.

(D, E, F): Images from a 40-year-old female patient in the progressing stage of COVID-19. Multifocal GGOs (D, red arrows), patchy consolidation (D, green arrow), and fibrosis (D, blue arrows) can be observed. All lesions were segmented, as shown on the same coronal view (E, within red line). F represents the three-dimensional reconstruction of the lesion.

(G, H, I): Images from a 50-year-old male patient in the severe stage of COVID-19. Diffused GGO (G, red arrows) and consolidation (G, green arrow) can be seen in the bilateral lungs. All lesions were segmented, as shown on the same coronal view (H, red color). I represent the three-dimensional reconstruction of the lesion.

Fig. 5. Correlation of lesion percentage derived by radiologists and the computer software

The scatter diagram (A-H, plotted with red reversed triangles) shows the correlation between lesion percentages assessed by the radiologist and the computer software for the whole lung (A), right lung (B), left lung (C), right upper lobe (D), right middle lobe (E), right lower lobe (F), left upper lobe (G), and left lower lobe (H).

WL, whole lung; RL, right lung; LL, left lung; RUL, right upper lobe; RML, right middle lobe; LUL, left upper lobe; LLL, left lower lobe.

RLP, radiologist-assessed lesion percentage; CLP, computer-assessed lesion percentage

Fig. 6. Correlation of lesion density derived by radiologists and the computer software

The scatter diagram (A-B, plotted with red filled circles) shows a moderate negative correlation between the proportion of GGO and mean lesion density (shown in A), and a moderate positive correlation between the proportion of consolidation and mean lesion density (shown in B).

P_{GGO} , proportion of ground glass opacity to the overall lesion;

$P_{consolidation}$, proportion of consolidation to the overall lesion

Acknowledgments

The authors would like to express their gratitude to all emergency services, nurses, doctors, and other hospital staff for their efforts to combat the COVID-19 outbreak.

Conflicts of interest

The author declared that there are no conflicts of interest.

References

- [1] N. Zhu, D.Y. Zhang, W.L. Wang, et al., A novel coronavirus from patients with pneumonia in China, 2019, *N. Engl. J. Med.* 382(2020) 727-733.
- [2] P. Zhou, X.L. Yang, X.G. Wang, et al., A pneumonia outbreak associated with a new coronavirus of probable bat origin, *Nature* (2020), <https://doi.org/10.1038/s41586-020-2012-7> [Epub ahead of print]
- [3] C. Huang, Y. Wang, X. Li, et al., Clinical features of patients infected with 2019 novel coronavirus in Wuhan, China, *Lancet* 395(2020)497-506.
- [4] Z.Y. Zu, M.D. Jiang, P.P. Xu, et al., Coronavirus disease 2019 (COVID-19): A perspective from China, *Radiology* 2020, DOI:10.1148/radiol.2020200490 [Epub ahead of print]
- [5] Y.H. Jin, L. Cai, Z.S. Cheng, et al., A rapid advice guideline for the diagnosis and treatment of 2019 novel coronavirus (2019-nCoV) infected pneumonia (standard version), *Mil. Med. Res.* 2020, DOI :10.1186/s40779-020-0233-6 [Epub ahead of print]
- [6] X. Xie, Z. Zhong, W. Zhao, et al., Chest CT for typical 2019-nCoV pneumonia: Relationship to negative RT-PCR testing, *Radiology* 2020, DOI: 10.1148/radiol.2020200343[Epub ahead of print]
- [7] D. Wang, B. Hu, C. Hu, et al., Clinical characteristics of 138 hospitalized patients with 2019 novel coronavirus-infected pneumonia in Wuhan, China, *JAMA* 2020, DOI: 10.1001/jama.2020.1585 [Epub ahead of print]
- [8] Y. Pan, H. Guan, S. Zhou, et al., Initial CT findings and temporal changes in patients with the novel coronavirus pneumonia (2019-nCoV): a study of 63 patients in Wuhan, China, *Eur. Radiol.* 2020, DOI: 10.1007/s00330-020-06731-x [Epub ahead of print]
- [9] F. Pan, T. Ye, P. Sun, et al., Time course of lung changes on chest CT during recovery from 2019 novel coronavirus (COVID-19) pneumonia, *Radiology* 2020, DOI: 10.1148/radiol.2020200370 [Epub ahead of print]
- [10] National Health Commission of the People's Republic of China, Diagnosis and treatment protocols of Coronavirus Disease 2019 (trial version 6), 2020, <http://www.nhc.gov.cn/xcs/zchengwj/202002/8334a8326dd94d329df351d7da8aefc2.shtml>.
- [11] D.M. Hansell, A.A. Bankier, H. MacMahon, et al., Fleischner Society: glossary of terms for thoracic imaging, *Radiology* 246(2008) 697-722.

- [12] H.J. Koo, S. Lim, J. Choe, et al., Radiographic and CT features of viral pneumonia, Radiographics 38(2018)719-739.
- [13] E.G. Chan, J.R. Landreneau, M.J. Schuchert, et al., Preoperative (3-dimensional) computed tomography lung reconstruction before anatomic segmentectomy or lobectomy for stage I non-small cell lung cancer, J. Thorac. Cardiovasc. Surg. 150(2015) 523-528.
- [14] J. Pu, J. Roos, C.A. Yi, et al., Adaptive border marching algorithm: automatic lung segmentation on chest CT images, Comput. Med. Imaging Graph. 32(2008) 452-462.
- [15] J. Pu, B. Zheng, J.K. Leader, et al., Pulmonary lobe segmentation in CT examinations using implicit surface fitting, IEEE Trans. Med. Imaging 28(2009) 1986-1996.
- [16] S. Gu, C. Fuhrman, X. Meng, et al., Computerized identification of airway wall in CT examinations using a 3D active surface evolution approach, Med. Image Anal. 17(2013)283-296.
- [17] X. Wang, J.K. Leader, R. Wang, et al., Vasculature surrounding a nodule: A novel lung cancer biomarker, Lung Cancer 114(2017) 38-43.
- [18] B.C. Lassen, C. Jacobs, J.M. Kuhnigk, et al., Robust semi-automatic segmentation of pulmonary subsolid nodules in chest computed tomography scans, Phys. Med. Biol. 60(2015)1307-1323.
- [19] S.Tian, W.Hu, L.Niu, et al., Pulmonary pathology of early phase 2019 novel coronavirus (COVID-19) pneumonia in two patients with lung cancer, J Thorac Oncol, 2020, DOI: 10.1016/j.jtho.2020.02.010 [Epub ahead of print]
- [20] T. Ishiguro, Y. Kobayashi, R. Uozumi, et al., Viral pneumonia requiring differentiation from acute and progressive diffuse interstitial lung diseases, Int. Med. 58 (2019)3509-3519.
- [21] J. Wu, X. Wu, W. Zeng, et al., Chest CT findings in patients with corona virus disease 2019 and its relationship with clinical features, Invest. Radiol. 2020, DOI: 10.1097/RLI.0000000000000670 [Epub ahead of print]

Table 2 The lesion distribution of right/left side observed by radiologist and obtained using computer tool

Items	Right lung	Left lung	Sum
Total RLP	207 (57.82%)	151(42.18%)	100%
Total CLP	1282.53 (51.06%)	1229.4 (48.94%)	100%
Summary	108.88%	91.12%	200%

RLP, radiologist-assessed lesion percentage;

CLP, computer-assessed lesion percentage.

Table 3 The lesion distribution of upper/lower lung observed by radiologist and obtained using the computer tool

Items	Upper lobes†	Lower lobes‡	Sum
Total RLP	188 (52.51%)	170(47.49%)	100%
Total CLP	834.86 (33.24%)	1677.1 (66.76%)	100%
Summary	85.75%	114.25%	200%

†Upper lobes were the general name of the right upper lobe, the right middle lobe, and the left upper lobe.

‡Lower lobes were the general name of the right lower lobe and the left lower lobe.

RLP, radiologist-assessed lesion percentage;

CLP, computer-assessed lesion percentage.

Table 4 The lesion distribution among five lobes observed by radiologist and obtained using computer tool

Items	RUL	RML	RLL	LUL	LLL	Sum
Total RLP	73 (20.39%)	47(13.13%)	87(24.30%)	68(18.99%)	83(23.18%)	100%
Total CLP	288.35 (11.48%)	166.46 (6.62%)	827.72 (32.95%)	380.05 (15.12%)	849.34 (33.81%)	100%
Summary	31.87%	19.75%	57.25%	34.11%	56.99%	200%

RLP, radiologist-assessed lesion percentage;

CLP, computer-assessed lesion percentage.

RUL, right upper lobe; RML, right middle lobe; RLL, right lower lobe; LUL, left upper lobe; LLL, left lower lobe.

Table 1 General characteristics and comparison of quantitative CT parameters between non-severe and severe patient groups

Items	Non-severe (n=35)	Severe (n=9)	T/χ ²	p
Gender(M/F)	18/17	4/5	-	0.698*
Age	52.17 ± 12.55	57.25 ± 13.38	-1.029	0.310
RLP _{WL}	7.14 ± 3.14	12.00 ± 2.12	4.377	<0.001
RLP _{RUL}	1.43 ± 0.81	2.56 ± 0.73	3.775	<0.001
RLP _{RML}	0.94 ± 0.76	1.56 ± 0.88	2.079	0.044
RLP _{RLL}	1.77 ± 0.91	2.78 ± 0.44	4.729	<0.001
RLP _{LUL}	1.31 ± 0.93	2.44 ± 0.88	3.277	0.002
RLP _{LLL}	1.69 ± 1.05	2.67 ± 0.50	4.027	<0.001
P _{GGO}	6.44 ± 2.92	7.13 ± 2.10	-0.621	0.538
P _{consolidation}	2.56 ± 2.65	2.25 ± 1.67	0.312	0.757
P _{fibrosis}	1.00 ± 1.07	0.63 ± 0.74	0.939	0.353
LV _{WL} (mL)	3953 ± 1210	3158 ± 647.1	2.677	0.013
LV _{RL} (mL)	2130 ± 636.5	1702 ± 331.2	2.784	0.010
LV _{LL} (mL)	1823 ± 593.4	1456 ± 334.1	1.772	0.084
LV _{RUL} (mL)	875.2 ± 292.4	720.5 ± 150.0	1.293	0.203
LV _{RML} (mL)	403.5 ± 112.3	362.2 ± 151.6	0.892	0.378
LV _{RLL} (mL)	879.9 ± 313.6	632.2 ± 141.4	3.493	0.002
LV _{LUL} (mL)	1023 ± 310.9	853.2 ± 225.7	1.534	0.132
LV _{LLL} (mL)	798.8 ± 312.5	584.1 ± 198.0	1.953	0.058
LeV _{WL} (mL)	259.7 ± 198.4	863.6 ± 460.8	-3.841	0.004
LeV _{RL} (mL)	130.1 ± 99.46	431.3 ± 171.2	-5.064	0.001
LeV _{LL} (mL)	127.8 ± 134.8	428.0 ± 301.1	-2.917	0.017
LeV _{RUL} (mL)	32.39 ± 39.14	120.5 ± 76.43	-3.347	0.008
LeV _{RML} (mL)	8.19 ± 15.33	37.86 ± 26.50	-4.410	0.000
LeV _{RLL} (mL)	94.18 ± 76.62	226.9 ± 123.5	-4.057	0.000
LeV _{LUL} (mL)	49.28 ± 70.39	189.7 ± 139.7	-2.920	0.017
LeV _{LLL} (mL)	81.06 ± 85.58	219.1 ± 159.6	-2.054	0.033
CLP _{WL} (%)	7.28 ± 5.75	27.18 ± 13.47	-4.330	0.002
CLP _{RL} (%)	6.91 ± 6.15	25.16 ± 8.97	-5.755	0.000
CLP _{LL} (%)	8.10 ± 8.78	29.73 ± 19.92	-3.178	0.012
CLP _{RUL} (%)	4.09 ± 4.83	16.11 ± 9.66	-3.618	0.006
CLP _{RML} (%)	2.26 ± 4.29	9.67 ± 9.63	-2.248	0.052
CLP _{RLL} (%)	13.51 ± 13.16	39.42 ± 19.84	-3.714	0.004
CLP _{LUL} (%)	5.55 ± 8.80	20.63 ± 14.41	-2.997	0.014
CLP _{LLL} (%)	13.22 ± 15.25	42.97 ± 30.19	-2.864	0.018
MLD _{WL} (HU)	-414.1 ± 106.8	-348.6 ± 88.0	-1.694	0.098
MLD _{RL} (HU)	-409.5 ± 141.9	-346.1 ± 73.97	-1.289	0.204
MLD _{LL} (HU)	-411.7 ± 176.1	-346.7 ± 101.0	-1.057	0.297
MLD _{RUL} (HU)	-403.3 ± 139.0	-374.2 ± 87.99	-0.596	0.554
MLD _{RML} (HU)	-371.1 ± 194.1	-341.2 ± 271.5	0.376	0.709
MLD _{RLL} (HU)	-421.2 ± 125.6	-333.9 ± 89.30	-1.954	0.057

$\text{MLD}_{\text{LUL}}(\text{HU})$	-452.1 ± 110.3	-378.0 ± 122.0	-1.761	0.085
$\text{MLD}_{\text{LLL}}(\text{HU})$	-431.4 ± 101.6	-337.6 ± 102.8	-2.466	0.018

LV, lung volume; LeV, lesion volume; MLD, mean lesion density; GGO, ground glass opacity; WL, whole lung; RL, right lung; LL, left lung; RUL, right upper lobe; RML, right middle lobe; LUL, left upper lobe; LLL, left lower lobe.


P_{GGO} , $P_{\text{consolidation}}$, P_{fibrosis} were the proportion of GGO, consolidation and fibrosis to the general lesion of one patient.


RLP, radiologist-assessed lesion percentage.


CLP, computer-assessed lesion percentage.


p values less than 0.05 were bolded.


*Fisher's test














Highlights:

1. Lesion percentage quantified using the computer tool correlated highly with visually scored lesions.
2. Lesion density quantified using the computer tool correlated with the visually scored proportion of ground glass opacity and consolidation.
3. The computer tool could reliably and accurately assess the severity and distribution of pneumonia.