
Calibration Table

Calib. Data Modified : Thursday, April 24, 2025 6:00:25 PM

Rel. Reference Window : 5.000 %
Abs. Reference Window : 0.000 min
Rel. Non-ref. Window : 5.000 %
Abs. Non-ref. Window : 0.000 min
Uncalibrated Peaks : not reported
Partial Calibration : Yes, identified peaks are recalibrated
Correct All Ret. Times: No, only for identified peaks

Curve Type : Linear
Origin : Included
Weight : Equal

Recalibration Settings:
Average Response : Average all calibrations
Average Retention Time: Floating Average New 75%

Calibration Report Options :
Printout of recalibrations within a sequence:
Calibration Table after Recalibration
Normal Report after Recalibration
If the sequence is done with bracketing:
Results of first cycle (ending previous bracket)

Signal 1: DAD1 A, Sig=290,4 Ref=360,100

RetTime [min]	Lvl Sig	Amount [ug/ml]	Area	Amt/Area	Ref Grp Name
2.438	1	10.00000	296.16348	3.37651e-2	B6
	2	25.00000	620.81195	4.02698e-2	
	3	50.00000	1286.75647	3.88574e-2	
	4	70.00000	1737.21973	4.02943e-2	
	5	100.00000	2530.36230	3.95200e-2	

Peak Sum Table

No Entries in table
