


==== Shimadzu LabSolutions Calibration Curve =====

ID# : 1
 Name : Cafeina
 Quantitative Method : External Standard
 Function : $f(x)=2,80960e+007*x+164028$
 Rr1=0,9999933 Rr2=0,9999866 RSS=1,555477e+009
 MeanRF: 2,885337e+007 RFSD: 5,134828e+005 RFRSD: 1,779628
 FitType : Linear
 ZeroThrough : Not Through
 Weighted Regression : None
 Offset Correction : Off
 Detector Name : Detector A



#	Conc.(Ratio)	MeanArea	Area
1	0,1	2948718	2948718
2	0,2	5807986	5807986
3	0,4	11415203	11415203
4	0,6	17009026	17009026