

Assay name: pH-metric high logP Analyst: Pion Instrument ID: Assay ID: 18B-27016 T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

### pH-metric Result

logP (XH +) -5.60 ±1.55 (n=48) logP (neutral X) 2.09 ±0.01 (n=48)

# 18B-27016 Points 1 to 20

M11 octanol concentration factor 0.990 Carbonate 0.0000 mM Acidity error 0.67760 mM

#### 18B-27016 Points 21 to 40

M11\_octanol concentration factor 0.815 Carbonate 0.0822 mM 0.32858 mM Acidity error

#### 18B-27016 Points 41 to 62

M11 octanol concentration factor 0.703 Carbonate 0.1416 mM Acidity error 0.69165 mM

# Warnings and errors

Errors None

Warnings One or more logP values out of range

#### Sample logD and percent species

рН	M11_octanol	M11_octanol	M11_octanol	M11_octanol	M11_octanol	Comment
	logD	M11_octanolH	M11_octanol	M11_octanolH*	M11_octanol*	
1.000	-0.80	86.28 %	0.11 %	0.00 %	13.61 %	
1.200	-0.60	79.87 %	0.16 %	0.00 %	19.97 %	Stomach pH
2.000	0.19	38.61 %	0.50 %	0.00 %	60.90 %	
3.000	1.15	5.92 %	0.76 %	0.00 %	93.32 %	
4.000	1.84	0.62 %	0.81 %	0.00 %	98.57 %	
5.000	2.06	0.06 %	0.81 %	0.00 %	99.13 %	
6.000	2.08	0.01 %	0.81 %	0.00 %	99.18 %	
6.500	2.09	0.00 %	0.81 %	0.00 %	99.19 %	
7.000	2.09	0.00 %	0.81 %	0.00 %	99.19 %	
7.400	2.09	0.00 %	0.81 %	0.00 %	99.19 %	Blood pH
8.000	2.09	0.00 %	0.81 %	0.00 %	99.19 %	
9.000	2.09	0.00 %	0.81 %	0.00 %	99.19 %	
10.000	2.09	0.00 %	0.81 %	0.00 %	99.19 %	
11.000	2.09	0.00 %	0.81 %	0.00 %	99.19 %	
12.000	2.09	0.00 %	0.81 %	0.00 %	99.19 %	



Sample name: M11\_octanol Assay name:

pH-metric high logP

Assay ID: 18B-27016 Filename:


C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

Experiment start time: 2/27/2018 10:54:30 PM

Pion Analyst: Instrument ID:

T312060








Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

# **Graphs** (continued)





Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

# pH-metric high logP Titration 1 of 3 18B-27016 Points 1 to 20

#### Overall results

RMSD 0.163
Average ionic strength 0.158 M
Average temperature 25.0°C
Partition ratio 0.0123 : 1

Analyte concentration range 3203.9 µM to 3307.2 µM

Total points considered 17 of 20

### Warnings and errors

Errors None Warnings None


### Four-Plus parameters


### Titrants

#### Sample

M11\_octanol concentration factor 0.990
Base pKa 1 3.89
logP (XH +) -0.58
logP (neutral X) 2.04

#### Sample graphs








Assay ID:

Sample name: M11\_octanol Experiment start time: 2/27/2018 10:54:30 PM

Assay name: pH-metric high logP Analyst: Pion Instrument ID: T312060 18B-27016

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

# Sample graphs (continued)



### Sample logD and percent species


рН	M11_octanol logD	M11_octanol M11_octanolH	M11_octanol M11_octanol	M11_octanol M11_octanolH*	M11_octanol M11_octanol*	Comment
1.000	-0.40	99.38 %	0.13 %	0.32 %	0.17 %	
1.200	-0.31	99.21 %	0.20 %	0.32 %	0.27 %	Stomach pH
2.000	0.22	96.77 %	1.25 %	0.31 %	1.67 %	·
3.000	1.10	76.66 %	9.88 %	0.25 %	13.22 %	
4.000	1.79	24.90 %	32.08 %	0.08 %	42.94 %	
5.000	2.01	3.21 %	41.38 %	0.01 %	55.40 %	
6.000	2.03	0.33 %	42.61 %	0.00 %	57.05 %	
6.500	2.04	0.10 %	42.71 %	0.00 %	57.18 %	
7.000	2.04	0.03 %	42.74 %	0.00 %	57.22 %	
7.400	2.04	0.01 %	42.75 %	0.00 %	57.24 %	Blood pH
8.000	2.04	0.00 %	42.76 %	0.00 %	57.24 %	·
9.000	2.04	0.00 %	42.76 %	0.00 %	57.24 %	
10.000	2.04	0.00 %	42.76 %	0.00 %	57.24 %	
11.000	2.04	0.00 %	42.76 %	0.00 %	57.24 %	
12.000	2.04	0.00 %	42.76 %	0.00 %	57.24 %	


# **Carbonate and acidity**



Carbonate 0.000 mM Acidity error 0.678 mM

# Other graphs








Assay name: pH-metric high logP Analyst: **Pion** Instrument ID: T312060 Assay ID: 18B-27016 Filename:

C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

# Other graphs (continued)





Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

# pH-metric high logP Titration 2 of 3 18B-27016 Points 21 to 40

#### Overall results

RMSD 0.275
Average ionic strength 0.163 M
Average temperature 25.0°C
Partition ratio 0.0582 : 1

Analyte concentration range 2864.7 µM to 2955.2 µM

Total points considered 15 of 20

### Warnings and errors

Errors None Warnings None

## Four-Plus parameters


Alpha 0.130 2/27/2018 10:54:30 PM C:\Sirius\_T3\HCl18B27.t3r S 0.9970 2/27/2018 10:54:30 PM C:\Sirius\_T3\HCl18B27.t3r jH 0.8 2/27/2018 10:54:30 PM C:\Sirius\_T3\HCl18B27.t3r jOH -0.4 2/27/2018 10:54:30 PM C:\Sirius\_T3\HCl18B27.t3r


# Titrants

#### Sample

M11\_octanol concentration factor 0.815
Base pKa 1 3.89
logP (XH +) -0.58
logP (neutral X) 2.15

#### Sample graphs








Assay ID:

Sample name: M11\_octanol Experiment start time: 2/27/2018 10:54:30 PM

Assay name: pH-metric high logP Analyst: Pion Instrument ID: T312060 18B-27016

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r


# Sample graphs (continued)




### Sample logD and percent species


рН	M11_octanol logD	M11_octanol M11_octanolH	M11_octanol M11_octanol	M11_octanol M11_octanolH*	M11_octanol M11_octanol*	
1.000	-0.35	97.35 %	0.13 %		1.03 %	
1.200	-0.26	96.70 %	0.20 %	1.48 %	1.63 %	Stomach pH
2.000	0.31	88.17 %	1.14 %	1.35 %	9.35 %	•
3.000	1.21	45.36 %	5.84 %	0.69 %	48.10 %	
4.000	1.90	7.75 %	9.98 %	0.12 %	82.15 %	
5.000	2.12	0.83 %	10.74 %	0.01 %	88.41 %	
6.000	2.15	0.08 %	10.82 %	0.00 %	89.09 %	
6.500	2.15	0.03 %	10.83 %	0.00 %	89.14 %	
7.000	2.15	0.01 %	10.83 %	0.00 %	89.16 %	
7.400	2.15	0.00 %	10.83 %	0.00 %	89.16 %	Blood pH
8.000	2.15	0.00 %	10.83 %	0.00 %	89.17 %	
9.000	2.15	0.00 %	10.83 %	0.00 %	89.17 %	
10.000	2.15	0.00 %	10.83 %	0.00 %	89.17 %	
11.000	2.15	0.00 %	10.83 %	0.00 %	89.17 %	
12.000	2.15	0.00 %	10.83 %	0.00 %	89.17 %	

# **Carbonate and acidity**



# Other graphs








Experiment start time: 2/27/2018 10:54:30 PM Sample name: M11\_octanol

Assay name: pH-metric high logP Analyst: **Pion** Instrument ID: T312060 Assay ID: 18B-27016 Filename:

C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

# Other graphs (continued)





Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

# pH-metric high logP Titration 3 of 3 18B-27016 Points 41 to 62

### Overall results

RMSD 0.122
Average ionic strength 0.169 M
Average temperature 25.0°C
Partition ratio 0.1640 : 1

Analyte concentration range 2436.6 µM to 2507.4 µM

Total points considered 16 of 22

### Warnings and errors

Errors None Warnings None

## Four-Plus parameters

Alpha 0.130 2/27/2018 10:54:30 PM C:\Sirius\_T3\HCl18B27.t3r S 0.9970 2/27/2018 10:54:30 PM C:\Sirius\_T3\HCl18B27.t3r jH 0.8 2/27/2018 10:54:30 PM C:\Sirius\_T3\HCl18B27.t3r jOH -0.4 2/27/2018 10:54:30 PM C:\Sirius\_T3\HCl18B27.t3r


#### Titrants


0.50 M HCI 0.993513 2/27/2018 10:54:30 PM C:\Sirius\_T3\HCl18B27.t3r 0.50 M KOH 0.999845 2/27/2018 10:54:30 PM C:\Sirius\_T3\KOH18B27.t3r

#### Sample

M11\_octanol concentration factor 0.703
 Base pKa 1 3.89
 logP (XH +) -0.58
 logP (neutral X) 2.09

#### Sample graphs








Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r


# Sample graphs (continued)




## Sample logD and percent species

рН	M11_octanol				M11_octanol	Comment
	logD	M11_octanolH	M11_octanol	M11_octanolH*	M11_octanol*	
1.000	-0.37	93.40 %	0.12 %	4.03 %	2.45 %	
1.200	-0.29	92.02 %	0.19 %	3.97 %	3.82 %	Stomach pH
2.000	0.26	75.86 %	0.98 %	3.27 %	19.89 %	·
3.000	1.16	26.36 %	3.40 %	1.14 %	69.10 %	
4.000	1.84	3.50 %	4.51 %	0.15 %	91.83 %	
5.000	2.06	0.36 %	4.67 %	0.02 %	94.96 %	
6.000	2.09	0.04 %	4.68 %	0.00 %	95.28 %	
6.500	2.09	0.01 %	4.68 %	0.00 %	95.30 %	
7.000	2.09	0.00 %	4.68 %	0.00 %	95.31 %	
7.400	2.09	0.00 %	4.68 %	0.00 %	95.31 %	Blood pH
8.000	2.09	0.00 %	4.68 %	0.00 %	95.32 %	·
9.000	2.09	0.00 %	4.68 %	0.00 %	95.32 %	
10.000	2.09	0.00 %	4.68 %	0.00 %	95.32 %	
11.000	2.09	0.00 %	4.68 %	0.00 %	95.32 %	
12.000	2.09	0.00 %	4.68 %	0.00 %	95.32 %	

### Carbonate and acidity



# Other graphs






Experiment start time: 2/27/2018 10:54:30 PM Sample name: M11\_octanol

**Pion** Assay name: pH-metric high logP Analyst: Instrument ID: T312060 Assay ID: 18B-27016 Filename:

C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

# Other graphs (continued)





Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

# **Assay Model**

Settings	Value	Date/Time changed	Imported from
Sample name	M11_octanol	2/27/2018 4:54:30 PM	User entered value
Sample by	Weight		Default value
Sample weight	0.001100 g	2/27/2018 6:40:56 PM	User entered value
Formula weight	211.22 g/mol	2/27/2018 4:54:30 PM	User entered value
Solubility	Unknown		Default value
Molecular weight	211.22	2/27/2018 4:54:30 PM	User entered value
Individual pKa ionic environments	No		Default value
Number of pKas	1	2/27/2018 4:54:30 PM	User entered value
Sample is a	Base	2/27/2018 4:54:30 PM	User entered value
pKa 1	3.89	2/27/2018 4:54:30 PM	User entered value
logp (XH +)	-0.58	2/27/2018 4:54:54 PM	User entered value
logP (neutral X)	1.24	2/27/2018 4:54:49 PM	User entered value

### **Events**

Event	Water	Acid	Base	Octanol	рН	dpH/dt	pH R-squared	pH SD	dpH/dt time
•									
Data point 3									
Data point 4								0.00045	10.0 s
Data point 5									
Data point 6	1.50000 mL	0.05167 mL	0.04052 mL	0.01999 mL	2.984	-0.01107	0.79804	0.00061	10.0 s
Data point 7	1.50000 mL	0.05167 mL	0.04304 mL	0.01999 mL	3.186	-0.00468	0.88101	0.00025	10.5 s
Data point 8								0.00069	10.0 s
Data point 9	1.50000 mL	0.05167 mL	0.04643 mL	0.01999 mL	3.547	-0.00399	0.79959	0.00022	10.5 s
Data point 10								0.00049	10.0 s
Data point 11	1.50000 mL	0.05167 mL	0.04965 mL	0.01999 mL	3.906	0.00461	0.20532	0.00050	10.5 s
Data point 12							0.89585	0.00082	10.0 s
Data point 13	1.50000 mL	0.05167 mL	0.05205 mL	0.01999 mL	4.458	-0.01283	0.78922	0.00071	10.5 s
Data point 14								0.00063	10.0 s
Data point 15									
									Timed out at
= <sub> </sub>	• • • • • •	•	•	•			• • • • • • • • • • • • • • • • • • • •		59.5 s
Data point 19	1.50000 mL	0.05167 mL	0.05367 mL	0.01999 mL	8.647	-0.01738	0.86596	0.00092	
•									
									Timed out at
Data point 30	1.50000 IIIL	U. 10042 IIIL	0.10310111	0.10000 IIIL	1.102	-0.00014	0.30470	0.0021-	59.5 s
Data point 37	1.50000 mL	0.10842 mL	0.10917 mL	0.10000 mL	7.684	-0.05926	0.99557	0.00293	Timed out at
	Initial pH = 6.35 Data point 1 Data point 2 Data point 3 Data point 4 Data point 5 Data point 6 Data point 7 Data point 8 Data point 10 Data point 11 Data point 12 Data point 13 Data point 15 Data point 15 Data point 15 Data point 15 Data point 16 Data point 17 Data point 18  Data point 19 Data point 18  Data point 20 Data point 21 Data point 21 Data point 22 Data point 23 Data point 24 Data point 24 Data point 25 Data point 25 Data point 26 Data point 27 Data point 28 Data point 29 Data point 30 Data point 31 Data point 32 Data point 33 Data point 34 Data point 35 Data point 36	Initial pH = 6.35     Data point 1     Data point 2     Data point 3     Data point 4     Data point 5     Data point 6     Data point 7     Data point 8     Data point 10     Data point 11     Data point 12     Data point 13     Data point 14     Data point 10     Data point 11     Data point 12     Data point 13     Data point 14     Data point 15     Data point 15     Data point 16     Data point 17     Data point 18     Data point 18     Data point 19     Data point 10     Data point 10     Data point 11     Data point 12     Data point 13     Data point 14     Data point 15     Data point 16     Data point 17     Data point 18     Data point 19     Data point 20     Data point 21     Data point 21     Data point 23     Data point 24     Data point 25     Data point 26     Data point 27     Data point 28     Data point 29     Data point 30     Data point 31     Data point 31     Data point 32     Data point 34     Data point 35     Data point 36     Data point 36     Data point 37     Data point 38     Data point 39     Data point 30     Data point 31     Data point 34     Data point 35     Data point 36     Data point 36     Data point 37     Data point 38     Data point 39     Data point 30     Data point 31     Data point 31     Data point 33     Data point 34     Data point 35     Data point 36     Data point 36     Data point 37     Data point 38     Data point 39     Data point 30     Data point 31     Data point 31     Data point 34     Data point 35     Data point 36     Data point 36	Initial pH = 6.35	Initial pH = 6.35	Initial pH = 6.35				

59.5 s

# **Assay Events**



Sample name: M11\_octanol Experiment start time: 2/27/2018 10:54:30 PM

Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

1:06:25.0 Assay volumes 1.50000 mL 0.16747 mL 0.16980 mL 0.30000 mL

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

# **Events (continued)**

Time	Event	Water	Acid	Base	Octanol	рН	dpH/dt	pH R-squared	pH SD	dpH/dt
40.07.4	Data naint 20	1 F0000 mal	0.40040	0.40007	0.40000	0.244	0.00444	0.07440	0.00400	time
42:37.4	Data point 38	1.50000 ML	0.10842 ML	0.10927 IIIL	0.10000 mL	8.344	-0.02444	0.97413	0.00122	
44:13.0	Data point 20	1 50000 ml	0 10012 ml	0 10026 ml	0.10000 mL	0 601	0.01020	0.94203	0.00098	at 59.5 s
44.13.0 45:33.1	Data point 39 Data point 40		0.10842 IIIL 0.10842 mL		0.10000 mL		-0.01930	0.69862	0.00098	
46:50.9	Data point 41				0.30000 mL			0.09802	0.00036	
40.30.9 47:37.4	Data point 42		0.16747 mL		0.30000 mL		-0.00011	0.60425	0.00030	
47.37.4 48:13.6	Data point 43				0.30000 mL			0.54028	0.00007	
48:49.2	Data point 44				0.30000 mL			0.68600	0.00049	
49:25.4	Data point 45		0.16747 mL		0.30000 mL			0.44516	0.00022	10.0 s
50:00.9	Data point 46				0.30000 mL			0.33641	0.00023	
50:36.4	Data point 47				0.30000 mL			0.30029	0.00092	
50:50. <del>4</del> 51:11.9	Data point 48				0.30000 mL			0.44143	0.00032	
51:47.9	Data point 49				0.30000 mL			0.27260	0.00093	
52:23.4	Data point 50				0.30000 mL		-0.00230	0.16102	0.00028	
53:09.1	Data point 51				0.30000 mL			0.36746	0.00061	10.0 s
53:44.5	Data point 52				0.30000 mL			0.71656	0.00093	
54:30.2	Data point 53				0.30000 mL			0.27004	0.00082	
55:13.3	Data point 54	1.50000 mL	0.16747 mL	0.16921 mL	0.30000 mL	6.006	-0.01875	0.96219	0.00094	37.5 s
56:21.2	Data point 55	1.50000 mL	0.16747 mL	0.16931 mL	0.30000 mL	6.652	-0.05803	0.99479	0.00287	Timed out
	•									at 59.5 s
57:56.8	Data point 56	1.50000 mL	0.16747 mL	0.16938 mL	0.30000 mL	7.209	-0.06064	0.98448	0.00302	Timed out
	·									at 59.5 s
59:27.4	Data point 57	1.50000 mL	0.16747 mL	0.16943 mL	0.30000 mL	7.445	-0.06310	0.99629	0.00312	Timed out
										at 59.5 s
1:01:03.0	Data point 58	1.50000 mL	0.16747 mL	0.16950 mL	0.30000 mL	8.044	-0.04412	0.95409	0.00223	Timed out
										at 59.5 s
1:02:38.8	Data point 59			0.16957 mL	0.30000 mL	8.352	-0.01985	0.97791	0.00099	51.0 s
1:04:10.8	•		0.16747 mL		0.30000 mL	-	-0.01671	0.79899	0.00092	24.0 s
	Data point 61				0.30000 mL			0.88409	0.00096	
1:05:56.8	Data point 62	1.50000 mL	0.16747 mL	0.16980 mL	0.30000 mL	9.010	-0.01788	0.90579	0.00093	19.0 s



Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius T3\Mehtap\20180227 exp27 logP T3-2\18B-27016 M11 octanol pH-metric high logP.t3r

Filename: C:\Sirius_T3\Meh	tap\20180227_exp2	7_logP_T3-2\18E	3-27016_M11_octano	l_pH-metric high logF
Assay Settings				
Setting	Value	Original Value	Date/Time changed	Imported from
General Settings				
Analyst name	Pion			
Standard Experiment Settings				
Number of titrations	3			
Minimum pH	2.000			
Maximum pH	9.000			
pH step between points of	0.200			
Minimum titrant addition	0.00002 mL			
Maximum titrant addition	0.10000 mL			
Argon flow rate	100%			
Start titration using	Cautious pH adjust			
Advanced General Settings				
Detect turbidity using	None			
Collect turbidity sensor data	No			
Collect UV spectra	No			
Stir after titrant addition for	5 seconds			
For titrant addition, stir at	10%			
Titrant Pre-Dose				
Titrant pre-dose	None			
Assay Medium				
ISA water volume	1.50 mL			
Water added	Automatic			
Partition solvent type	Octanol			
Partition volume	0.020 mL			
Partition solvent added	Automatic			
After partition addition, stir for	1 seconds			
Sample Sonication	1 00001140			
Sonicate	Yes			
Adjust pH for sonication	No			
Sonicate for	300 seconds			
After sonication stir for	5 seconds			
Sample Dissolution	0 00001100			
Perform a dissolution stage	Yes			
	To start pH			
Stir to dissolve for	120 seconds			
For dissolution, stir at	10%			
Carbonate purge	10 70			
Perform a carbonate purge	No			
Temperature Control	NO			
Wait for temperature	Yes			
	25.0°C			
Required start temperature Acceptable deviation	0.5°C			
Time to wait	60 seconds			
	50%			
Stir speed of	30%			
Titration 1	المطعامة المسادة			
Titrate from	Low to high pH			
Adjust to start pH	Yes			
After pH adjust stir for	30 seconds			
Stir to allow partitioning for	15 seconds			
Stirrer speed for partitioning	50%			
Titration 2	Laura blak D			
Titrate from	Low to high pH			
Add additional water	0.00 mL			
Additional partition solvent volume				
Additional partition solvent added	Automatic			

Reported at: 2/28/2018 12:07:05 PM

After pH adjust stir for

Stir to allow partitioning for

Stirrer speed for partitioning

Additional partition solvent added Automatic

30 seconds

15 seconds

55%



Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

# Assay Settings (continued)

Value	Original Value	Date/Time changed	Imported from
	_	_	
Low to high pH			
0.00 mL			
0.200 mL			
Automatic			
30 seconds			
15 seconds			
60%			
No			
0 seconds			
20 points			
0.50 seconds			
0.00100 dpH/dt			
60 seconds			
	Low to high pH 0.00 mL 0.200 mL Automatic 30 seconds 15 seconds 60% No 0 seconds 20 points 0.50 seconds 0.00100 dpH/dt	Low to high pH 0.00 mL 0.200 mL Automatic 30 seconds 15 seconds 60%  No 0 seconds 20 points 0.50 seconds 0.00100 dpH/dt	Low to high pH 0.00 mL 0.200 mL Automatic 30 seconds 15 seconds 60%  No 0 seconds 20 points 0.50 seconds 0.00100 dpH/dt

# Calibration Settings

Setting	Value	Date/Time changed	Imported from
Four-Plus alpha	0.130	2/27/2018 10:54:30 PM	C:\Sirius_T3\HCl18B27.t3r
Four-Plus S	0.9970	2/27/2018 10:54:30 PM	C:\Sirius_T3\HCl18B27.t3r
Four-Plus jH	8.0	2/27/2018 10:54:30 PM	C:\Sirius_T3\HCl18B27.t3r
Four-Plus jOH	-0.4	2/27/2018 10:54:30 PM	C:\Sirius_T3\HCl18B27.t3r
Base concentration factor	1.000	2/27/2018 10:54:30 PM	C:\Sirius_T3\KOH18B27.t3r
Acid concentration factor	0.994	2/27/2018 10:54:30 PM	C:\Sirius_T3\HCl18B27.t3r

# Instrument Settings

Setting Instrument owner Instrument ID Instrument type Software version	Value Merck T312060 T3 Simulator 1.1.3.0	Batch Id	Install date
Dispenser module Dispenser 0 Syringe volume Firmware version	Water 2.5 mL 1.2.1(r2)	T3DM1200361	3/31/2009 5:24:52 AM 3/31/2009 5:25:05 AM
Titrant Dispenser 2 Syringe volume Firmware version	Water (0.15 M KCI) Acid 0.5 mL 1.2.1(r2)	02-06-2018	2/27/2018 10:05:59 AM 3/31/2009 5:25:11 AM
Titrant Dispenser 1 Syringe volume Firmware version	Acid (0.5 M HCI) Base 0.5 mL 1.2.1(r2)	02-27-2018	2/27/2018 10:27:22 AM 3/31/2009 5:25:21 AM
Titrant Dispenser 5 Syringe volume Firmware version Distribution valve 5	Base (0.5 M KOH) Cosolvent 2.5 mL 1.2.1(r2) Distribution Valve	9/22/2017	2/27/2018 10:21:22 AM 3/31/2009 5:26:24 AM 3/31/2009 5:28:19 AM
Firmware version Port A Port B Dispenser 3 Syringe volume	1.1.3 Methanol (80%, 0.15 M KCI) Cyclohexane Buffer 0.5 mL	09-26-17 11-01-17	2/7/2018 9:42:01 AM 2/27/2018 10:37:57 AM 8/3/2010 5:05:16 AM
Firmware version Titrant Dispenser 6	1.2.1(r2) Dodecane Octanol	2018/01/31	1/31/2018 12:26:26 PM 10/22/2010 10:52:43 AM



Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

# Instrument Settings (continued)

Syringe volume   Firmware version   1.2.1(r2)   1.21(r2)	- · · ·			
Firmware version   1.2.1(r2)   Octanol   1.31-2018   2/27/2018 9.59:35 AM   Titrator   Titrator   1.17 Al1DI2DO2 Stepper 2   1.17 Al1DI2DO2 Stepper 2   1.17 Al1DI2DO2 Stepper 2   1.11 Al1DI2DO2 Stepper 3   1/23/2018 2:01:00 PM   1.11   1.1	Setting	Value	Batch Id	Install date
Titrath   Catalor   Cata				
Tirator			01 21 2010	2/27/2019 0:50:25 AM
Horizontal axis firmware version		Octanoi		
Vertical axis firmware version		1 17 ΔΙ1DΙ2DO2 Stenner 2	1311111200101	3/3 1/2009 3.24.17 AW
Chassis I/O firmware version				
Probe				
Electrode				
EO calibration			T3E0923	1/23/2018 2:01:00 PM
Filling solution   Liquids   Wash 1	E0 calibration			
Liquids   Wash 1			KCL097	
Wash 2   0.5% Trition X-100 in H20   2/27/2018 9:50:01 AM   Buffer position 1   PHT Wash   PH 7   2/27/2018 9:50:04 AM   2/27/2018 9:50:04 AM   2/27/2018 9:50:04 AM   2/27/2018 9:50:06 AM   2/27/2018 9:50:06 AM   2/27/2018 9:55:02 AM   2/27/2018 9:50:02 AM   2/27/2018 9:50	Liquids			
Buffer position 1   Buffer position 2   PH7 Wash   PH7   Wash   PH7   Wash   PH7   Wash   PH7   Wash   PH7   Wash   Wash water   9.2e+003 mL   02-27-2018   9:55:12 AM   Wash water   9.2e+003 mL   02-27-2018   9:55:12 AM   Wash water   6.2e+003 mL   02-27-2018   9:55:12 AM   Wash water   6.2e+003 mL   07-4811   11/28/2017   136:29 AM   Wash water   07-4811   11/28/2017   0:35:13 AM   Wash water   07-4811   10196   Wavelength coefficient A0   183.333   Wavelength coefficient A1   2.21568   Wavelength coefficient A2   0.000289308   Wavelength coefficient A2   0.000289308   Wavelength coefficient A2   0.000289308   Wash water   0.000289	Wash 1	50% IPA:50% Water		2/27/2018 9:49:58 AM
Buffer position 2   Storage position	Wash 2	0.5% Trition X-100 in H20		2/27/2018 9:50:01 AM
Storage position   Wash water   9.2e+003 mL   02-27-2018   2/27/2018 9:55:12 AM   Wash water   9.2e+003 mL   02-27-2018   2/27/2018 9:54:39 AM   Waster   6.2e+003 mL   11/28/2017 10:36:29 AM   8/5/2010 6:35:13 AM   3/5/2010 6:35:13 AM   3/5				
Wash water         9.2e+003 mL         02-27-2018         2/27/2018 9:54:39 AM           Waste         6.2e+003 mL         11/28/2017 10:36:29 AM           Temperature controller         3/31/2009 5:24:45 AM           Turbidity detector         074811         11/23/2010 11:22:28 AM           Spectrometer         074811         11/23/2010 11:22:28 AM           Dip probe         4074811         10196           Wavelength coefficient A0         183.333         11/23/2010 11:22:28 AM           Wavelength coefficient A1         2.21568         11/23/2010 11:22:28 AM           Calibrated on Integration time         2/27/2018 10:40:38 AM         11/23/2010 11:22:28 AM           Calibrated on Integration time         40         10         11/23/2010 11:22:28 AM           Autoloader         40         2/27/2018 10:40:38 AM         11/23/2010 11:22:28 AM           Left-right axis firmware version         1.17 AI 1DI2DO2 Stepper 2         1.7 AI 1DI2DO2 Stepper 2           Front-back axis firmware version         1.17 AI 1DI2DO2 Stepper 2         1.7 AI 1DI2DO2 Stepper 2           Chassis I/O firmware version         1.17 AI 1DI2DO2 Stepper 2         1.7 AI 1DI2DO2 Stepper 2           Chassis I/O firmware version         1.17 AI 1DI2DO2 Stepper 2         1.1 AI 1DI2DO2 Stepper 2           Chassis I/O firmware version		pH 7		
Waste         6.2e+003 mL         11/28/2017 10:36:29 AM           Temperature controller         8/5/2010 6:35:13 AM           Turbidity detector         074811         11/23/2010 11:22:28 AM           Spectrometer         074811         11/23/2010 11:22:28 AM           Dip probe         1096         11/23/2010 11:22:28 AM           Wavelength coefficient A0         183.333         11/23/2010 11:22:28 AM           Wavelength coefficient A2         -0.000289308         11/23/2010 11:22:28 AM           Total lamp lit time         110:52:56         11/23/2010 11:22:28 AM           Calibrated on Integration time         40         11/23/2010 11:22:28 AM           Scans averaged         10         10         11/23/2010 11:22:28 AM           Left-right axis firmware version         1.17 AI DI2DO2 Stepper 2         11/23/2010 11:22:28 AM           Left-right axis firmware version         1.17 AI DI2DO2 Stepper 2         11/23/2010 11:22:28 AM           Vertical axis firmware version         1.17 AI DI2DO2 Stepper 2         11/10/2015 9:34:13 AM           Left-right axis firmware version         1.17 AI DI2DO2 Stepper 2         1.17 AI DI2DO2 Stepper 2           Chassis I/O firmware version         1.17 AI DI2DO2 Stepper 2         1.17 AI DI2DO2 Stepper 2           Chassis I/O firmware version         1.17 AI DI2DO2 Stepper 2				
Temperature controller			02-27-2018	
Turibidity detector		6.2e+003 mL		
Spectromèter         074811         11/23/2010 11:22:28 AM           Dip probe         10196           Wavelength coefficient A0         183.333           Wavelength coefficient A1         2.21568           Wavelength coefficient A2         -0.000289308           Total lamp lit time         110:52:56           Calibrated on         2/27/2018 10:40:38 AM           Integration time         40           Scans averaged         10           Autoloader         T3AL 1200345           Left-right axis firmware version         1.17 Al1DI2DO2 Stepper 2           Front-back axis firmware version         1.17 Al1DI2DO2 Stepper 2           Vertical axis firmware version         1.17 Al1DI2DO2 Stepper 2           Chassis I/O firmware version         1.17 Al1DI2DO2 Stepper 2           Chassis I/O firmware version         1.17 Al1DI2DO2 Stepper 2           Alternate iteration position         Reference position           Alternate reference position         Reference position           Maximum alternate vial volume         3.50           Automatic action idle period         5 minute(s)           Titrant tube volume         1.3 mL           Syringe flush count         5 s           Flowing wash stir duration         5 s           Solvent wash				
Dip probe			074044	
Wavelength coefficient A0         183.333           Wavelength coefficient A1         2.21588           Wavelength coefficient A2         -0.000289308           Total lamp lit time         110.52:56         11/23/2010 11:22:28 AM           Calibrated on         2/277/2018 10:40:38 AM         11/23/2010 11:22:28 AM           Integration time         40         5.25.26         11/23/2010 11:22:28 AM           Autoloader         T3AL1200345         11/10/2015 9:34:13 AM           Left-right axis firmware version         1.17 Al1DI2DO2 Stepper 2         11/10/2015 9:34:13 AM           Vertical axis firmware version         1.17 Al1DI2DO2 Stepper 2         11/10/2015 9:34:13 AM           Vertical axis firmware version         1.17 Al1DI2DO2 Stepper 2         11/10/2015 9:34:13 AM           Vertical axis firmware version         1.17 Al1DI2DO2 Stepper 2         11/10/2015 9:34:13 AM           Vertical axis firmware version         1.17 Al1DI2DO2 Stepper 2         11/10/2015 9:34:13 AM           Vertical axis firmware version         1.17 Al1DI2DO2 Stepper 2         11/10/2015 9:34:13 AM           Vertical axis firmware version         1.17 Al1DI2DO2 Stepper 2         11/10/2015 9:34:13 AM           Vertical axis firmware version         1.17 Al1DI2DO2 Stepper 1         11/10/2015 9:34:13 AM           Interval axis firmware version         1.17 Al1DI2DO2 St				11/23/2010 11:22:28 AM
Wavelength coefficient A1 Wavelength coefficient A2 Total lamp lit time Calibrated on Integration time Scans averaged Autoloader Left-right axis firmware version Chassis I/O firmware version Alternate titration position Alternate titration position Alternate reference position Maximum standard vial volume Automatic action idle period Titrant tube volume Automatic action idle period Titrant tube volume Syringe flush count Flowing wash stir duration Solvent wash stir speed Solvent wash stir duration Solvent wash stir duration Solvent wash stir duration E0 calibration maximum standard deviation E0 calibration titre wash stir duration E0 calibration buffer wash stir speed E0 calibration buffer wash stir duration E1 calibration buffer wash stir speed E0 calibration buffer wash stir duration E0 calibration buffer wash stir speed E0 calibration buffer wash stir duration E0 calibration		102 222	10190	
Wavelength coefficient A2         -0.000289308           Total lamp lit time         110:52:56         11/23/2010 11:22:28 AM           Calibrated on         2/27/2018 10:40:38 AM         11/23/2010 11:22:28 AM           Scans averaged         40         3           Autoloader         T3AL1200345         11/10/2015 9:34:13 AM           Left-right axis firmware version         1.17 Al1Dl2DO2 Stepper 2         11/10/2015 9:34:13 AM           Front-back axis firmware version         1.17 Al1Dl2DO2 Stepper 2         1.17 Al1Dl2DO2 Stepper 2           Chassis I/O firmware version         1.17 Al1Dl2DO2 Stepper 2         1.17 Al1Dl2DO2 Stepper 2           Chassis I/O firmware version         1.11 Al1Dl0DO4 Norgren I/O         1.11 Al1Dl2DO2 Stepper 2           Chassis I/O firmware version         1.11 Al1Dl0DO4 Norgren I/O         1.11 Al1Dl2DO2 Stepper 2           Chassis I/O firmware version         1.11 Al1Dl0DO4 Norgren I/O         1.11 Al1Dl2DO2 Stepper 2           Chassis I/O firmware version         1.11 Al1Dl0DO4 Norgren I/O         1.11 Al1Dl2DO2 Stepper 2           Chassis I/O firmware version         1.11 Al1Dl0DO4 Norgren I/O         1.11 Al1Dl2DO2 Stepper 2           Chassis I/O firmware version         7.11 Al1Dl2DO2 Stepper 2         1.11 Al1Dl2DO2 Stepper 2           Left-right axis firmware version         7.11 Al1Dl2DO2 Stepper 2         1.11 Al1Dl2DO2 Stepper 2 </td <td></td> <td></td> <td></td> <td></td>				
Total lamp lit time Calibrated on 2/27/2018 10:40:38 AM Integration time Scans averaged 10  Autoloader T3AL1200345 11/10/2015 9:34:13 AM Left-right axis firmware version 1.17 Al1DI2DO2 Stepper 2 Front-back axis firmware version 1.17 Al1DI2DO2 Stepper 2 Vertical axis firmware version 1.17 Al1DI2DO2 Stepper 2 Vertical axis firmware version 1.17 Al1DI2DO2 Stepper 2 Chassis I/O firmware version 1.11 Al1DI2DO2 Stepper 2 Alternate reference position Alternate reference position Maximum standard vial volume 3.50 mL Automatic action idle period 5 minute(s) Titrant tube volume 1.3 mL Syringe flush count 9.5 s Solvent wash stir duration 5 s Solvent wash stir speed 30% Surfactant wash stir speed 30% E0 calibration minimum number of points E0 calibration stir duration 5 s E0 calibration preparation stir speed 30% E0 calibration preparation stir speed 5 s E0 calibration buffer wash stir speed 30% E0 calibration buffer wash stir speed 30% E0 calibration buffer wash stir speed 30%				
Calibrated on Integration time 40   Scans averaged 10   Autoloader				11/23/2010 11·22·28 AM
Integration time Scans averaged Autoloader Left-right axis firmware version Left-right axis firmware version 1.17 Al1Dl2DO2 Stepper 2 Front-back axis firmware version 1.17 Al1Dl2DO2 Stepper 2 Vertical axis firmware version 1.17 Al1Dl2DO2 Stepper 2 Chassis I/O firmware version 1.11 Al1Dl2DO2 Norgren I/O Configuration Alternate itration position Alternate reference position Maximum standard vial volume Aximum alternate vial volume 3.50 mL Automatic action idle period Titrant tube volume 1.3 mL Syringe flush count Flowing wash pump volume 1.3 mL Syringe flush count Flowing wash stir duration Solvent wash stir duration 5 s Flowing wash stir speed 30% Solvent wash stir duration 5 s Solvent wash stir duration 5 s Solvent wash stir duration 5 s Coalibration minimum number of points Co calibration maximum standard deviation Co calibration stir duration Co calibration stir duration Co calibration preparation stir speed Co calibration buffer wash stir duration Co calibration buffer wash stir duration Co calibration buffer wash stir duration Co calibration buffer wash stir speed Co calibration buffer wash stir duration Co calibration buffer wash stir speed Co calibration buffer wash stir spee				11/25/2010 11.22.20 AW
Scans averaged Autoloader Left-right axis firmware version Front-back axis firmware version Vertical axis firmware version Alternate itiration position Alternate itiration position Alternate reference position Maximum standard vial volume Automatic action idle period Titratt tube volume Syringe flush count Syringe flush count Flowing wash stir speed Solvent wash stir speed Solvent wash stir speed Solvent wash stir speed Solvent wash stir speed Calibration minimum number of points Calibration preparation stir speed Calibration preparation stir speed Calibration buffer wash stir speed Calibration buffer wa				
Autoloader Left-right axis firmware version 1.17 Al1Dl2DO2 Stepper 2 Front-back axis firmware version 1.17 Al1Dl2DO2 Stepper 2 Vertical axis firmware version 1.17 Al1Dl2DO2 Stepper 2 Chassis I/O firmware version 1.17 Al1Dl2DO2 Stepper 2 Chassis I/O firmware version 1.11 Al1Dl0DO4 Norgren I/O Configuration Alternate titration position Alternate reference position Maximum standard vial volume Alternate reference position Maximum alternate vial volume 25.00 mL Automatic action idle period 5 minute(s) Titrant tube volume 1.3 mL Syringe flush count Syringe flush count Flowing wash pump volume 1.3 mL Syringe wash stir duration Solvent wash stir geed 30% Solvent wash stir geed 30% Solvant wash stir duration 5 s Surfactant wash stir duration Calibration minimum number of points Calibration minimum number of points Calibration minimum number of points Calibration timeout period Calibration timeout period Calibration preparation stir speed Calibration preparation stir speed Calibration preparation stir speed Calibration buffer wash stir duration S s				
Left-right axis firmware version 1.17 Al1Dl2DO2 Stepper 2 Front-back axis firmware version 1.17 Al1Dl2DO2 Stepper 2 Vertical axis firmware version 1.17 Al1Dl2DO2 Stepper 2 Chassis I/O firmware version 1.17 Al1Dl2DO2 Stepper 2 Chassis I/O firmware version 1.11 Al1DlDO4 Norgren I/O Configuration Alternate titration position Reference position Alternate reference position Reference position Maximum standard vial volume 3.50 mL Maximum alternate vial volume 25.00 mL Automatic action idle period 5 minute(s) Titratt tube volume 1.3 mL Syringe flush count 3.50 Flowing wash pump volume 20.0 mL Flowing wash stir duration 5 s Flowing wash stir speed 30% Solvent wash stir speed 30% Solvent wash stir speed 30% Solvent wash stir turation 5 s Solvent wash stir turation 5 s Surfactant wash stir speed 30% E0 calibration minimum number of points 10 E0 calibration maximum standard deviation 5 s E0 calibration stir duration 5 s E0 calibration preparation stir speed 30% E0 calibration buffer wash stir duration 5 s E0 calibration buffer wash stir duration 5 s E0 calibration buffer wash stir duration 5 s E0 calibration buffer wash stir speed 30%	Autoloader		T3AL1200345	11/10/2015 9:34:13 AM
Vertical axis firmware version Chassis I/O firmware version Alternate itration position Alternate reference position Maximum standard vial volume Maximum alternate vial volume Automatic action idle period Titrat tube volume Syringe flush count Flowing wash pump volume Flowing wash stir duration Solvent wash stir speed Solvent wash stir speed Solvent wash stir speed Coalibration minimum number of points E0 calibration preparation stir speed Coalibration preparation stir speed Calibration buffer wash stir duration E0 calibration buffer wash stir speed Calibration buffer wash stir speed Calibration buffer wash stir duration E0 calibration buffer wash stir speed Calibration buffer wash stir duration E0 calibration buffer wash stir speed E0 calibration buffer wash stir speed E0 calibration buffer wash stir duration E0 calibration buffer wash stir speed E0 calibration buffer wash stir duration E0 calibration buffer wash stir speed E0 calibration buffer wash stir duration E0 calibration buffer wash stir speed E0 calibration buffer wash stir duration E0 calibration buffer wash stir duratio	Left-right axis firmware version	1.17 Al1Dl2DO2 Stepper 2		
Chassis I/O firmware version  Configuration  Alternate titration position Alternate reference position Maximum standard vial volume Maximum alternate vial volume Automatic action idle period Titration position  Automatic action idle period Titratin tube volume Syringe flush count Flowing wash pump volume Plowing wash stir duration Solvent wash stir duration Solvent wash stir duration Solvent wash stir speed Solvent wash stir speed Solvent wash stir speed Surfactant wash stir speed Coalibration minimum number of points E0 calibration stir duration E0 calibration stir duration E0 calibration stir duration E0 calibration preparation stir speed E0 calibration preparation stir speed E0 calibration puffer wash stir speed E0 calibration buffer wash stir duration E0 calibration buffer wash stir speed	Front-back axis firmware version	1.17 Al1Dl2DO2 Stepper 2		
Alternate titration position Alternate reference position Alternate reference position Alternate reference position Maximum standard vial volume Automatic action idle period Titrant tube volume Automatic action idle period Titrant tube volume Syringe flush count Flowing wash pump volume Flowing wash stir duration Flowing wash stir duration Solvent wash stir speed Solvent wash stir speed Solvent wash stir speed Surfactant wash stir speed Goalibration minimum number of points E0 calibration stir duration E0 calibration preparation stir speed Coalibration preparation stir speed Solvent wash stir duration F1 owing wash stir speed Surfactant wash stir speed Surfactant wash stir speed Solvent wash stir duration Solvent wash stir speed Solvent was				
Alternate titration position Alternate reference position Maximum standard vial volume Maximum standard vial volume Automatic action idle period Titrant tube volume Automatic action idle period Titrant tube volume Syringe flush count Flowing wash pump volume Flowing wash stir duration Flowing wash stir speed Solvent wash stir speed Solvent wash stir speed Solvent wash stir duration Surfactant wash stir duration Surfactant wash stir speed Calibration minimum number of points E0 calibration stir duration E0 calibration preparation stir speed C0 calibration puffer wash stir duration S0% C0 calibration puffer wash stir duration S0% C0 calibration preparation stir speed C0 calibration buffer wash stir duration S0% C0 calibration buffer wash stir speed C0 cali		1.11 Al1Dl0DO4 Norgren I/O		
Alternate reference position Maximum standard vial volume Maximum alternate vial volume Automatic action idle period Titrant tube volume Syringe flush count Syringe flush count Flowing wash pump volume Flowing wash stir duration Flowing wash stir speed Solvent wash stir speed Surfactant wash stir speed Calibration maximum standard deviation E0 calibration stir duration E0 calibration turation E0 calibration preparation stir speed Calibration buffer wash stir duration E0 calibration buffer wash stir speed				
Maximum standard vial volume  Maximum alternate vial volume  Automatic action idle period  Titrant tube volume  Syringe flush count  Flowing wash pump volume  Flowing wash stir duration  Flowing wash stir speed  Solvent wash stir speed  Solvent wash stir speed  Solvent wash stir speed  Surfactant wash stir speed  Calibration minimum number of points  EO calibration maximum standard deviation  EO calibration stir duration  EO calibration preparation stir speed  EO calibration preparation stir speed  EO calibration preparation stir speed  EO calibration buffer wash stir duration  EO calibration buffer wash stir speed				
Maximum alternate vial volume Automatic action idle period 5 minute(s) Titrant tube volume 1.3 mL Syringe flush count Syringe				
Automatic action idle period Titrant tube volume 1.3 mL Syringe flush count Syringe fl				
Titrant tube volume  Syringe flush count  Syringe flush count  Syringe flush count  Solvent wash pump volume  Solvent wash stir speed  Solvent wash stir duration  Solvent wash stir speed  Solvent wash stir speed  Sourfactant wash stir duration  Surfactant wash stir speed  Calibration minimum number of points  Calibration timeout period  Calibration stir duration  Solvent wash stir speed  Solvent wash stir speed  Solvent wash stir speed  Outlier of points  Colibration maximum standard deviation  Colibration timeout period  Colibration stir duration  Solvent wash stir speed  Solvent wash stir duration  Solvent wash stir speed  Solvent wash				
Syringe flush count Flowing wash pump volume Flowing wash stir duration Flowing wash stir speed Solvent wash stir speed Surfactant wash stir duration Surfactant wash stir speed Surfactant wash stir speed Surfactant wash stir speed Surfactant wash stir speed Solvent wash stir duration Solvent wash stir speed Solvent wash stir speed Solvent wash stir duration Solvent wash stir speed				
Flowing wash pump volume Flowing wash stir duration Flowing wash stir speed Solvent wash stir speed Solvent wash stir speed Solvent wash stir speed Solvent wash stir speed Surfactant wash stir duration Surfactant wash stir speed Solvent wash stir speed Solvent wash stir duration Solvent wash stir speed Solvent wash stir duration Solvent wash stir speed Solvent wash stir speed Solvent wash stir duration Solvent wash stir du				
Flowing wash stir duration 5 s Flowing wash stir speed 30% Solvent wash stir duration 5 s Solvent wash stir speed 30% Surfactant wash stir duration 5 s Surfactant wash stir speed 30% E0 calibration minimum number of points 10 E0 calibration maximum standard deviation E0 calibration timeout period 60 s E0 calibration stir duration 5 s E0 calibration preparation stir speed 30% E0 calibration buffer wash stir duration 5 s E0 calibration buffer wash stir speed 30%				
Flowing wash stir speed 30% Solvent wash stir duration 5 s Solvent wash stir speed 30% Surfactant wash stir duration 5 s Surfactant wash stir speed 30% E0 calibration minimum number of points 10 E0 calibration maximum standard deviation E0 calibration timeout period 60 s E0 calibration stir duration 5 s E0 calibration preparation stir speed 30% E0 calibration buffer wash stir duration 5 s E0 calibration buffer wash stir speed 30%				
Solvent wash stir duration 5 s Solvent wash stir speed 30% Surfactant wash stir duration 5 s Surfactant wash stir speed 30% E0 calibration minimum number of points 10 E0 calibration maximum standard deviation E0 calibration timeout period 60 s E0 calibration stir duration 5 s E0 calibration preparation stir speed 30% E0 calibration buffer wash stir duration 5 s E0 calibration buffer wash stir speed 30%				
Solvent wash stir speed 30% Surfactant wash stir duration 5 s Surfactant wash stir speed 30% E0 calibration minimum number of points 10 E0 calibration maximum standard deviation E0 calibration timeout period 60 s E0 calibration stir duration 5 s E0 calibration preparation stir speed 30% E0 calibration buffer wash stir duration 5 s E0 calibration buffer wash stir speed 30%				
Surfactant wash stir duration 5 s Surfactant wash stir speed 30% E0 calibration minimum number of points 10 E0 calibration maximum standard deviation 0.01500 E0 calibration timeout period 60 s E0 calibration stir duration 5 s E0 calibration preparation stir speed 30% E0 calibration buffer wash stir duration 5 s E0 calibration buffer wash stir speed 30%				
E0 calibration minimum number of points 10 E0 calibration maximum standard deviation 0.01500 E0 calibration timeout period 60 s E0 calibration stir duration 5 s E0 calibration preparation stir speed 30% E0 calibration buffer wash stir duration 5 s E0 calibration buffer wash stir speed 30%	·	5 s		
E0 calibration maximum standard deviation 0.01500 E0 calibration timeout period 60 s E0 calibration stir duration 5 s E0 calibration preparation stir speed 30% E0 calibration buffer wash stir duration 5 s E0 calibration buffer wash stir speed 30%	Surfactant wash stir speed	30%		
E0 calibration timeout period 60 s E0 calibration stir duration 5 s E0 calibration preparation stir speed 30% E0 calibration buffer wash stir duration 5 s E0 calibration buffer wash stir speed 30%	·			
E0 calibration stir duration 5 s E0 calibration preparation stir speed 30% E0 calibration buffer wash stir duration 5 s E0 calibration buffer wash stir speed 30%				
E0 calibration preparation stir speed 30% E0 calibration buffer wash stir duration 5 s E0 calibration buffer wash stir speed 30%				
E0 calibration buffer wash stir duration 5 s E0 calibration buffer wash stir speed 30%				
E0 calibration buffer wash stir speed 30%				
Eu calibration reading Still Speed U70				
	Lo calibration reading Still Speed	U /0		



Assay name: pH-metric high logP Analyst: Pion Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

## Instrument Settings (continued)

Setting Spectrometer calibration stir duration	<b>Value</b> 5 s	Batch Id	Install date
Spectrometer calibration stir speed	30%		
Spectrometer calibration wash pump volume Spectrometer calibration wash stir duration	20.0 mL 5 s		
Spectrometer calibration wash stir speed	30%		
Overhead dispense height	10000		

#### Refinement Settings

Setting	Value	Default value
Turbidity detection method	None	None
Turbidity wavelength to assess	500.0 nm	500.0 nm
Turbidity maximum absorbance	0.100	0.100
Turbidity probe threshold	50.00	50.00

#### Experiment Log

[2:37]	Air aan	created	for	مte/۱۸	er (0.15 M KCI)	
ر ۲۰.۵ کے	ı Ali yap	Cicalcu	101	vvalc	(0.15 W KCI)	
[2:38]	l Air gan	created	for	Acid	(0.5 M HCI)	
12.001	m qup	oi catca	101	, ,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,	(0.0 101 1 101)	

- [2:38] Air gap created for Base (0.5 M KOH)
- [2:39] Air gap released for Water (0.15 M KCI)
- [2:42] Titrator arm moved over Titration position
- [2:42] Titration 1 of 3
- [2:42] Adding initial titrants
- [2:42] Automatically add 1.50000 mL of water
- [3:08] Dispensed 1.500000 mL of Water (0.15 M KCI)
- [3:12] Titrator arm moved over Drain
- [8:53] Titrator arm moved to Titration position
- [8:53] Argon flow rate set to 100
- [8:53] Stirrer speed set to 10
- [8:58] Automatically add 0.02000 mL of Octanol
- [8:59] Dispensed 0.019991 mL of Octanol
- [9:00] Initial pH = 6.35
- [9:00] Iterative adjust 6.35 -> 2.00
- [9:00] pH 6.35 -> 2.00
- [9:02] Air gap released for Acid (0.5 M HCI)
- [9:02] Dispensed 0.051670 mL of Acid (0.5 M HCI)
- [9:08] Holding pH 2.00
- [11:08] Stirrer speed set to 0
- [11:08] Stirrer speed set to 50
- [11:08] Iterative adjust 1.97 -> 2.00
- [11:08] pH 1.97 -> 2.00
- [11:08] Air gap released for Base (0.5 M KOH)
- [11:09] Dispensed 0.003010 mL of Base (0.5 M KOH)
- [11:59] Stirrer speed set to 0
- [12:10] Datapoint id 1 collected
- [12:10] Stirrer speed set to 50
- [12:15] pH 2.02 -> 2.22
- [12:15] Using cautious pH adjust
- [12:15] Dispensed 0.007573 mL of Base (0.5 M KOH)
- [12:20] Stepping pH = 2.11
- [12:20] Dispensed 0.005950 mL of Base (0.5 M KOH)
- [12:25] Stepping pH = 2.20
- [12:25] Dispensed 0.001411 mL of Base (0.5 M KOH)
- [12:31] Stepping pH = 2.22
- [12:46] Stirrer speed set to 0
- [12:56] Datapoint id 2 collected
- [12:56] Charge balance equation is out by 1.4%
- [12:56] Stirrer speed set to 50



Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

- [13:01] pH 2.22 -> 2.42
- [13:01] Using charge balance adjust
- [13:02] Dispensed 0.009525 mL of Base (0.5 M KOH)
- [13:22] Stirrer speed set to 0
- [13:32] Datapoint id 3 collected
- [13:32] Charge balance equation is out by -4.9%
- [13:32] Stirrer speed set to 50
- [13:37] pH 2.42 -> 2.62
- [13:37] Using charge balance adjust
- [13:37] Dispensed 0.006115 mL of Base (0.5 M KOH)
- [13:57] Stirrer speed set to 0
- [14:07] Datapoint id 4 collected
- [14:07] Charge balance equation is out by -14.8%
- [14:07] Stirrer speed set to 50
- [14:12] pH 2.60 -> 2.80
- [14:12] Using charge balance adjust
- [14:13] Dispensed 0.004186 mL of Base (0.5 M KOH)
- [14:33] Stirrer speed set to 0
- [14:43] Datapoint id 5 collected
- [14:43] Charge balance equation is out by 7.6%
- [14:43] Stirrer speed set to 50
- [14:48] pH 2.82 -> 3.02
- [14:48] Using charge balance adjust
- [14:49] Dispensed 0.002752 mL of Base (0.5 M KOH)
- [15:09] Stirrer speed set to 0
- [15:19] Datapoint id 6 collected
- [15:19] Charge balance equation is out by -19.4%
- [15:19] Stirrer speed set to 50
- [15:24] pH 2.99 -> 3.19
- [15:24] Using cautious pH adjust
- [15:24] Dispensed 0.001058 mL of Base (0.5 M KOH)
- [15:29] Stepping pH = 3.06
- [15:29] Dispensed 0.001152 mL of Base (0.5 M KOH)
- [15:34] Stepping pH = 3.16
- [15:35] Dispensed 0.000306 mL of Base (0.5 M KOH)
- [15:40] Stepping pH = 3.19
- [15:55] Stirrer speed set to 0
- [16:05] Datapoint id 7 collected
- [16:05] Charge balance equation is out by -18.8%
- [16:05] Stirrer speed set to 50
- [16:10] pH 3.19 -> 3.39
- [16:10] Using cautious pH adjust
- [16:10] Dispensed 0.000870 mL of Base (0.5 M KOH)
- [16:16] Stepping pH = 3.28
- [16:16] Dispensed 0.000706 mL of Base (0.5 M KOH)
- [16:21] Stepping pH = 3.36
- [16:21] Dispensed 0.000235 mL of Base (0.5 M KOH)
- [16:26] Stepping pH = 3.38
- [16:41] Stirrer speed set to 0
- [16:51] Datapoint id 8 collected
- [16:51] Charge balance equation is out by -2.8%
- [16:51] Stirrer speed set to 50
- [16:56] pH 3.38 -> 3.58
- [16:56] Using charge balance adjust
- [16:56] Dispensed 0.001576 mL of Base (0.5 M KOH)
- [17:17] Stirrer speed set to 0
- [17:27] Datapoint id 9 collected
- [17:27] Charge balance equation is out by -14.9%
- [17:27] Stirrer speed set to 50



pH-metric high logP Assay name: Analyst: Pion Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

- [17:32] pH 3.55 -> 3.75
- [17:32] Using charge balance adjust
- [17:32] Dispensed 0.001529 mL of Base (0.5 M KOH)
- [17:53] Stirrer speed set to 0
- [18:03] Datapoint id 10 collected
- [18:03] Charge balance equation is out by -21.5%
- [18:03] Stirrer speed set to 50
- [18:08] pH 3.71 -> 3.91
- [18:08] Using cautious pH adjust
- [18:08] Dispensed 0.000729 mL of Base (0.5 M KOH)
- [18:13] Stepping pH = 3.79
- [18:13] Dispensed 0.000729 mL of Base (0.5 M KOH)
- [18:18] Stepping pH = 3.89
- [18:18] Dispensed 0.000118 mL of Base (0.5 M KOH)
- [18:23] Stepping pH = 3.90
- [18:23] Dispensed 0.000118 mL of Base (0.5 M KOH)
- [18:28] Stepping pH = 3.91
- [18:44] Stirrer speed set to 0
- [18:54] Datapoint id 11 collected
- [18:54] Charge balance equation is out by -16.0%
- [18:54] Stirrer speed set to 50
- [18:59] pH 3.91 -> 4.11
- [18:59] Using cautious pH adjust
- [18:59] Dispensed 0.000682 mL of Base (0.5 M KOH)
- [19:04] Stepping pH = 4.01
- [19:04] Dispensed 0.000517 mL of Base (0.5 M KOH)
- [19:10] Stepping pH = 4.11
- [19:25] Stirrer speed set to 0
- [19:35] Datapoint id 12 collected
- [19:35] Charge balance equation is out by 12.1%
- [19:35] Stirrer speed set to 50
- [19:40] pH 4.12 -> 4.32
- [19:40] Using charge balance adjust
- [19:40] Dispensed 0.001199 mL of Base (0.5 M KOH)
- [20:00] Stirrer speed set to 0
- [20:11] Datapoint id 13 collected
- [20:11] Charge balance equation is out by 71.5%
- [20:11] Stirrer speed set to 50
- [20:16] pH 4.46 -> 4.66
- [20:16] Using cautious pH adjust
- [20:16] Dispensed 0.000400 mL of Base (0.5 M KOH)
- [20:21] Stepping pH = 4.58
- [20:21] Dispensed 0.000188 mL of Base (0.5 M KOH)
- [20:26] Stepping pH = 4.68
- [20:41] Stirrer speed set to 0
- [20:51] Datapoint id 14 collected
- [20:51] Charge balance equation is out by 26.5%
- [20:51] Stirrer speed set to 50
- [20:56] pH 4.73 -> 4.93
- [20:56] Using cautious pH adjust
- [20:56] Dispensed 0.000259 mL of Base (0.5 M KOH)
- [21:02] Stepping pH = 4.86
- [21:02] Dispensed 0.000118 mL of Base (0.5 M KOH)
- [21:07] Stepping pH = 4.98
- [21:22] Stirrer speed set to 0
- [21:34] Datapoint id 15 collected
- [21:34] Charge balance equation is out by 28.5%
- [21:34] Stirrer speed set to 50
- [21:39] pH 5.06 -> 5.26



pH-metric high logP Assay name: Analyst: Pion Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

- [21:39] Using cautious pH adjust
- [21:40] Dispensed 0.000141 mL of Base (0.5 M KOH)
- [21:45] Stepping pH = 5.13
- [21:45] Dispensed 0.000141 mL of Base (0.5 M KOH)
- [21:50] Stepping pH = 5.52
- [22:05] Stirrer speed set to 0
- [22:23] Datapoint id 16 collected
- [22:23] Charge balance equation is out by -5.6%
- [22:23] Stirrer speed set to 50
- [22:29] pH 5.70 -> 5.90
- [22:29] Using charge balance adjust
- [22:29] Dispensed 0.000094 mL of Base (0.5 M KOH)
- [22:49] Stirrer speed set to 0
- [23:07] Datapoint id 17 collected
- [23:07] Charge balance equation is out by -30.7%
- [23:07] Stirrer speed set to 50
- [23:12] pH 5.93 -> 6.13
- [23:12] Using cautious pH adjust
- [23:12] Dispensed 0.000047 mL of Base (0.5 M KOH)
- [23:17] Stepping pH = 5.95
- [23:17] Dispensed 0.000094 mL of Base (0.5 M KOH)
- [23:22] Stepping pH = 7.00
- [23:37] Stirrer speed set to 0
- [24:37] Datapoint id 18 collected
- [24:37] Charge balance equation is out by -81.0%
- [24:37] Stirrer speed set to 50
- [24:42] pH 8.28 -> 8.48
- [24:42] Using cautious pH adjust
- [24:42] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [24:48] Stepping pH = 8.30
- [24:48] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [24:53] Stepping pH = 8.31
- [24:53] Dispensed 0.000047 mL of Base (0.5 M KOH)
- [24:58] Stepping pH = 8.37
- [24:58] Dispensed 0.000047 mL of Base (0.5 M KOH)
- [25:03] Stepping pH = 8.51
- [25:18] Stirrer speed set to 0
- [26:03] Datapoint id 19 collected
- [26:03] Charge balance equation is out by -1,164.1%
- [26:03] Stirrer speed set to 50
- [26:08] pH 8.73 -> 8.93
- [26:08] Using cautious pH adjust
- [26:08] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [26:13] Stepping pH = 8.74
- [26:13] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [26:18] Stepping pH = 8.75
- [26:18] Dispensed 0.000094 mL of Base (0.5 M KOH)
- [26:23] Stepping pH = 9.13
- [26:38] Stirrer speed set to 0
- [27:07] Datapoint id 20 collected
- [27:07] Charge balance equation is out by -543.2%
- [27:07] Titration 2 of 3
- [27:07] Adding initial titrants
- [27:07] Automatically add 0.08000 mL of Octanol
- [27:09] Dispensed 0.080009 mL of Octanol
- [27:09] Stirrer speed set to 10
- [27:10] Stirrer speed set to 55
- [27:10] Iterative adjust 9.27 -> 2.00
- [27:10] pH 9.27 -> 2.00



Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

- [27:12] Dispensed 0.054798 mL of Acid (0.5 M HCI)
- [27:17] pH 2.02 -> 2.00
- [27:17] Dispensed 0.001952 mL of Acid (0.5 M HCI)
- [28:07] Stirrer speed set to 0
- [28:17] Datapoint id 21 collected
- [28:17] Stirrer speed set to 55
- [28:22] pH 1.97 -> 2.17
- [28:22] Using cautious pH adjust
- [28:23] Dispensed 0.009102 mL of Base (0.5 M KOH)
- [28:28] Stepping pH = 2.07
- [28:28] Dispensed 0.005833 mL of Base (0.5 M KOH)
- [28:33] Stepping pH = 2.14
- [28:33] Dispensed 0.002516 mL of Base (0.5 M KOH)
- [28:39] Stepping pH = 2.17
- [28:54] Stirrer speed set to 0
- [29:04] Datapoint id 22 collected
- [29:04] Charge balance equation is out by 4.1%
- [29:04] Stirrer speed set to 55
- [29:09] pH 2.17 -> 2.37
- [29:09] Using charge balance adjust
- [29:10] Dispensed 0.011571 mL of Base (0.5 M KOH)
- [29:30] Stirrer speed set to 0
- [29:40] Datapoint id 23 collected
- [29:40] Charge balance equation is out by 4.1%
- [29:40] Stirrer speed set to 55
- [29:45] pH 2.39 -> 2.59
- [29:45] Using charge balance adjust
- [29:46] Dispensed 0.007173 mL of Base (0.5 M KOH)
- [30:06] Stirrer speed set to 0
- [30:16] Datapoint id 24 collected
- [30:16] Charge balance equation is out by -13.1%
- [30:16] Stirrer speed set to 55
- [30:21] pH 2.57 -> 2.77
- [30:21] Using charge balance adjust
- [30:21] Dispensed 0.004915 mL of Base (0.5 M KOH)
- [30:42] Stirrer speed set to 0
- [30:52] Datapoint id 25 collected
- [30:52] Charge balance equation is out by -15.1%
- [30:52] Stirrer speed set to 55
- [30:57] pH 2.75 -> 2.95
- [30:57] Using cautious pH adjust
- [30:57] Dispensed 0.001764 mL of Base (0.5 M KOH)
- [31:02] Stepping pH = 2.81
- [31:02] Dispensed 0.002093 mL of Base (0.5 M KOH)
- [31:07] Stepping pH = 2.92
- [31:07] Dispensed 0.000470 mL of Base (0.5 M KOH)
- [31:12] Stepping pH = 2.95
- [31:27] Stirrer speed set to 0
- [31:37] Datapoint id 26 collected
- [31:37] Charge balance equation is out by -22.6%
- [31:37] Stirrer speed set to 55
- [31:43] pH 2.95 -> 3.15
- [31:43] Using cautious pH adjust
- [31:43] Dispensed 0.001294 mL of Base (0.5 M KOH)
- [31:48] Stepping pH = 3.06
- [31:48] Dispensed 0.000753 mL of Base (0.5 M KOH)
- [31:53] Stepping pH = 3.12
- [31:53] Dispensed 0.000306 mL of Base (0.5 M KOH)
- [31:58] Stepping pH = 3.15



Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

- [32:13] Stirrer speed set to 0
- [32:23] Datapoint id 27 collected
- [32:23] Charge balance equation is out by 9.1%
- [32:23] Stirrer speed set to 55
- [32:28] pH 3.15 -> 3.35
- [32:28] Using charge balance adjust
- [32:29] Dispensed 0.002093 mL of Base (0.5 M KOH)
- [32:49] Stirrer speed set to 0
- [32:59] Datapoint id 28 collected
- [32:59] Charge balance equation is out by -11.4%
- [32:59] Stirrer speed set to 55
- [33:04] pH 3.34 -> 3.54
- [33:04] Using charge balance adjust
- [33:04] Dispensed 0.001858 mL of Base (0.5 M KOH)
- [33:24] Stirrer speed set to 0
- [33:34] Datapoint id 29 collected
- [33:34] Charge balance equation is out by 3.5%
- [33:34] Stirrer speed set to 55
- [33:39] pH 3.55 -> 3.75
- [33:39] Using charge balance adjust
- [33:39] Dispensed 0.001670 mL of Base (0.5 M KOH)
- [34:00] Stirrer speed set to 0
- [34:10] Datapoint id 30 collected
- [34:10] Charge balance equation is out by 67.8%
- [34:10] Stirrer speed set to 55
- [34:15] pH 3.89 -> 4.09
- [34:15] Using cautious pH adjust
- [34:15] Dispensed 0.000659 mL of Base (0.5 M KOH)
- [34:20] Stepping pH = 4.11
- [34:35] Stirrer speed set to 0
- [34:46] Datapoint id 31 collected
- [34:46] Charge balance equation is out by 50.0%
- [34:46] Stirrer speed set to 55
- [34:51] pH 4.14 -> 4.34
- [34:51] Using cautious pH adjust
- [34:51] Dispensed 0.000494 mL of Base (0.5 M KOH)
- [34:56] Stepping pH = 4.42
- [35:11] Stirrer speed set to 0
- [35:22] Datapoint id 32 collected
- [35:22] Charge balance equation is out by 50.0%
- [35:22] Stirrer speed set to 55
- [35:27] pH 4.45 -> 4.65
- [35:27] Using cautious pH adjust
- [35:27] Dispensed 0.000306 mL of Base (0.5 M KOH)
- [35:32] Stepping pH = 4.69
- [35:47] Stirrer speed set to 0
- [35:59] Datapoint id 33 collected
- [35:59] Charge balance equation is out by 50.0%
- [35:59] Stirrer speed set to 55
- [36:04] pH 4.85 -> 5.05
- [36:04] Using cautious pH adjust
- [36:04] Dispensed 0.000141 mL of Base (0.5 M KOH)
- [36:09] Stepping pH = 4.91
- [36:09] Dispensed 0.000165 mL of Base (0.5 M KOH)
- [36:14] Stepping pH = 5.80
- [36:29] Stirrer speed set to 0
- [37:21] Datapoint id 34 collected
- [37:21] Charge balance equation is out by -5.2%
- [37:21] Stirrer speed set to 55



Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

- [37:27] pH 6.56 -> 6.76
- [37:27] Using charge balance adjust
- [37:27] Dispensed 0.000047 mL of Base (0.5 M KOH)
- [37:47] Stirrer speed set to 0
- [38:45] Datapoint id 35 collected
- [38:45] Charge balance equation is out by -43.3%
- [38:45] Stirrer speed set to 55
- [38:51] pH 6.84 -> 7.04
- [38:51] Using cautious pH adjust
- [38:51] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [38:56] Stepping pH = 6.89
- [38:56] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [39:01] Stepping pH = 6.95
- [39:01] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [39:06] Stepping pH = 7.16
- [39:21] Stirrer speed set to 0
- [40:21] Datapoint id 36 collected
- [40:21] Charge balance equation is out by -119.4%
- [40:21] Stirrer speed set to 55
- [40:26] pH 7.34 -> 7.54
- [40:26] Using cautious pH adjust
- [40:26] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [40:31] Stepping pH = 7.43
- [40:31] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [40:36] Stepping pH = 7.48
- [40:36] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [40:42] Stepping pH = 7.57
- [40:57] Stirrer speed set to 0
- [41:57] Datapoint id 37 collected
- [41:57] Charge balance equation is out by -407.6%
- [41:57] Stirrer speed set to 55
- [42:02] pH 7.81 -> 8.01
- [42:02] Using cautious pH adjust
- [42:02] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [42:07] Stepping pH = 7.83
- [42:07] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [42:12] Stepping pH = 7.87
- [42:12] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [42:17] Stepping pH = 7.91
- [42:17] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [42:22] Stepping pH = 8.04
- [42:37] Stirrer speed set to 0
- [43:37] Datapoint id 38 collected
- [43:37] Charge balance equation is out by -1,031.4%
- [43:37] Stirrer speed set to 55
- [43:42] pH 8.38 -> 8.58
- [43:42] Using cautious pH adjust
- [43:42] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [43:48] Stepping pH = 8.40
- [43:48] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [43:53] Stepping pH = 8.43
- [43:53] Dispensed 0.000047 mL of Base (0.5 M KOH)
- [43:58] Stepping pH = 8.58
- [44:13] Stirrer speed set to 0
- [44:58] Datapoint id 39 collected
- [44:58] Charge balance equation is out by -519.7%
- [44:58] Stirrer speed set to 55
- [45:03] pH 8.70 -> 8.90
- [45:03] Using cautious pH adjust



Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

- [45:03] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [45:08] Stepping pH = 8.70
- [45:08] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [45:13] Stepping pH = 8.71
- [45:13] Dispensed 0.000094 mL of Base (0.5 M KOH)
- [45:18] Stepping pH = 9.07
- [45:33] Stirrer speed set to 0
- [45:48] Datapoint id 40 collected
- [45:48] Charge balance equation is out by -554.5%
- [45:48] Titration 3 of 3
- [45:48] Adding initial titrants
- [45:48] Automatically add 0.20000 mL of Octanol
- [45:53] Dispensed 0.200000 mL of Octanol
- [45:53] Stirrer speed set to 10
- [45:54] Stirrer speed set to 60
- [45:54] Iterative adjust 9.28 -> 2.00
- [45:54] pH 9.28 -> 2.00
- [45:55] Dispensed 0.057761 mL of Acid (0.5 M HCI)
- [46:00] pH 2.01 -> 2.00
- [46:01] Dispensed 0.001294 mL of Acid (0.5 M HCI)
- [46:51] Stirrer speed set to 0
- [47:01] Datapoint id 41 collected
- [47:01] Stirrer speed set to 60
- [47:06] pH 1.97 -> 2.17
- [47:06] Using cautious pH adjust
- [47:06] Dispensed 0.009901 mL of Base (0.5 M KOH)
- [47:11] Stepping pH = 2.05
- [47:12] Dispensed 0.007926 mL of Base (0.5 M KOH)
- [47:17] Stepping pH = 2.13
- [47:17] Dispensed 0.003057 mL of Base (0.5 M KOH)
- [47:22] Stepping pH = 2.17
- [47:37] Stirrer speed set to 0
- [47:48] Datapoint id 42 collected
- [47:48] Charge balance equation is out by -5.4%
- [47:48] Stirrer speed set to 60
- [47:53] pH 2.17 -> 2.37
- [47:53] Using charge balance adjust
- [47:53] Dispensed 0.012488 mL of Base (0.5 M KOH)
- [48:14] Stirrer speed set to 0
- [48:24] Datapoint id 43 collected
- [48:24] Charge balance equation is out by -4.2%
- [48:24] Stirrer speed set to 60
- [48:29] pH 2.37 -> 2.57
- [48:29] Using charge balance adjust
- [48:29] Dispensed 0.008184 mL of Base (0.5 M KOH)
- [48:49] Stirrer speed set to 0
- [49:00] Datapoint id 44 collected
- [49:00] Charge balance equation is out by -4.0%
- [49:00] Stirrer speed set to 60
- [49:05] pH 2.57 -> 2.77
- [49:05] Using charge balance adjust
- [49:05] Dispensed 0.005503 mL of Base (0.5 M KOH)
- [49:25] Stirrer speed set to 0
- [49:35] Datapoint id 45 collected
- [49:35] Charge balance equation is out by -14.9%
- [49:35] Stirrer speed set to 60
- [49:40] pH 2.74 -> 2.94
- [49:40] Using charge balance adjust
- [49:41] Dispensed 0.004022 mL of Base (0.5 M KOH)



pH-metric high logP **Pion** Assay name: Analyst: Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

- [50:01] Stirrer speed set to 0
- [50:11] Datapoint id 46 collected
- [50:11] Charge balance equation is out by -5.7%
- [50:11] Stirrer speed set to 60
- [50:16] pH 2.94 -> 3.14
- [50:16] Using charge balance adjust
- [50:16] Dispensed 0.003057 mL of Base (0.5 M KOH)
- [50:36] Stirrer speed set to 0
- [50:46] Datapoint id 47 collected
- [50:46] Charge balance equation is out by 0.9%
- [50:46] Stirrer speed set to 60
- [50:52] pH 3.14 -> 3.34
- [50:52] Using charge balance adjust
- [50:52] Dispensed 0.002446 mL of Base (0.5 M KOH)
- [51:12] Stirrer speed set to 0
- [51:22] Datapoint id 48 collected
- [51:22] Charge balance equation is out by 52.1%
- [51:22] Stirrer speed set to 60
- [51:27] pH 3.45 -> 3.65
- [51:27] Using cautious pH adjust
- [51:28] Dispensed 0.000917 mL of Base (0.5 M KOH)
- [51:33] Stepping pH = 3.65
- [51:48] Stirrer speed set to 0
- [51:58] Datapoint id 49 collected
- [51:58] Charge balance equation is out by 50.0%
- [51:58] Stirrer speed set to 60
- [52:03] pH 3.65 -> 3.85
- [52:03] Using cautious pH adjust
- [52:03] Dispensed 0.000729 mL of Base (0.5 M KOH)
- [52:08] Stepping pH = 3.89
- [52:23] Stirrer speed set to 0
- [52:33] Datapoint id 50 collected
- [52:33] Charge balance equation is out by 50.0%
- [52:33] Stirrer speed set to 60
- [52:38] pH 3.90 -> 4.10
- [52:38] Using cautious pH adjust
- [52:39] Dispensed 0.000517 mL of Base (0.5 M KOH)
- [52:44] Stepping pH = 4.06
- [52:44] Dispensed 0.000094 mL of Base (0.5 M KOH)
- [52:49] Stepping pH = 4.08
- [52:49] Dispensed 0.000071 mL of Base (0.5 M KOH)
- [52:54] Stepping pH = 4.12
- [53:09] Stirrer speed set to 0
- [53:19] Datapoint id 51 collected
- [53:19] Charge balance equation is out by 33.8%
- [53:19] Stirrer speed set to 60
- [53:24] pH 4.14 -> 4.34
- [53:24] Using cautious pH adjust
- [53:24] Dispensed 0.000353 mL of Base (0.5 M KOH)
- [53:29] Stepping pH = 4.47
- [53:44] Stirrer speed set to 0
- [53:54] Datapoint id 52 collected
- [53:54] Charge balance equation is out by 50.0%
- [53:54] Stirrer speed set to 60
- [54:00] pH 4.52 -> 4.72
- [54:00] Using cautious pH adjust
- [54:00] Dispensed 0.000165 mL of Base (0.5 M KOH)
- [54:05] Stepping pH = 4.67
- [54:05] Dispensed 0.000047 mL of Base (0.5 M KOH)



Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

- [54:10] Stepping pH = 4.69
- [54:10] Dispensed 0.000047 mL of Base (0.5 M KOH)
- [54:15] Stepping pH = 4.78
- [54:30] Stirrer speed set to 0
- [54:43] Datapoint id 53 collected
- [54:43] Charge balance equation is out by 22.1%
- [54:43] Stirrer speed set to 60
- [54:48] pH 4.89 -> 5.09
- [54:48] Using cautious pH adjust
- [54:48] Dispensed 0.000071 mL of Base (0.5 M KOH)
- [54:53] Stepping pH = 4.94
- [54:53] Dispensed 0.000118 mL of Base (0.5 M KOH)
- [54:58] Stepping pH = 6.12
- [55:13] Stirrer speed set to 0
- [55:51] Datapoint id 54 collected
- 55:51] Charge balance equation is out by -21.1%
- [55:51] Stirrer speed set to 60
- [55:56] pH 6.13 -> 6.33
- [55:56] Using cautious pH adjust
- [55:56] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [56:01] Stepping pH = 6.16
- [56:01] Dispensed 0.000071 mL of Base (0.5 M KOH)
- [56:06] Stepping pH = 6.51
- EC.041 Ctimer and act to
- [56:21] Stirrer speed set to 0
- [57:21] Datapoint id 55 collected
- [57:21] Charge balance equation is out by -82.3%
- [57:21] Stirrer speed set to 60
- [57:26] pH 6.85 -> 7.05
- [57:26] Using cautious pH adjust
- [57:26] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [57:31] Stepping pH = 6.94
- [57:31] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [57:37] Stepping pH = 7.01
- [57:37] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [57:42] Stepping pH = 7.10
- [57:57] Stirrer speed set to 0
- [58:57] Datapoint id 56 collected
- [58:57] Charge balance equation is out by -92.6%
- [58:57] Stirrer speed set to 60
- [59:02] pH 7.25 -> 7.45
- [59:02] Using cautious pH adjust
- [59:02] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [59:07] Stepping pH = 7.33
- [59:07] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [59:12] Stepping pH = 7.49
- [59:27] Stirrer speed set to 0
- [1:00:27] Datapoint id 57 collected
- [1:00:27] Charge balance equation is out by -136.9%
- [1:00:27] Stirrer speed set to 60
- [1:00:32] pH 7.57 -> 7.77
- [1:00:32] Using cautious pH adjust
- [1:00:32] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [1:00:38] Stepping pH = 7.65
- [1:00:38] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [1:00:43] Stepping pH = 7.74
- [1:00:43] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [1:00:48] Stepping pH = 7.91
- [1:01:03] Stirrer speed set to 0
- [1:02:03] Datapoint id 58 collected



Assay name: pH-metric high logP Analyst: Pion
Assay ID: 18B-27016 Instrument ID: T312060

Filename: C:\Sirius\_T3\Mehtap\20180227\_exp27\_logP\_T3-2\18B-27016\_M11\_octanol\_pH-metric high logP.t3r

- [1:02:03] Charge balance equation is out by -478.6%
- [1:02:03] Stirrer speed set to 60
- [1:02:08] pH 8.11 -> 8.31
- [1:02:08] Using cautious pH adjust
- [1:02:08] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [1:02:13] Stepping pH = 8.12
- [1:02:13] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [1:02:18] Stepping pH = 8.20
- [1:02:19] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [1:02:24] Stepping pH = 8.30
- [1:02:39] Stirrer speed set to 0
- [1:03:30] Datapoint id 59 collected
- [1:03:30] Charge balance equation is out by -552.5%
- [1:03:30] Stirrer speed set to 60
- [1:03:35] pH 8.41 -> 8.61
- [1:03:35] Using cautious pH adjust
- [1:03:35] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [1:03:40] Stepping pH = 8.43
- [1:03:40] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [1:03:45] Stepping pH = 8.46
- [1:03:45] Dispensed 0.000047 mL of Base (0.5 M KOH)
- [1:03:50] Stepping pH = 8.58
- [1:03:51] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [1:03:56] Stepping pH = 8.71
- [1:04:11] Stirrer speed set to 0
- [1:04:35] Datapoint id 60 collected
- [1:04:35] Charge balance equation is out by -623.7%
- [1:04:35] Stirrer speed set to 60
- [1:04:40] pH 8.77 -> 8.97
- [1:04:40] Using cautious pH adjust
- [1:04:40] Dispensed 0.000024 mL of Base (0.5 M KOH)
- [1:04:45] Stepping pH = 8.78
- [1:04:45] Dispensed 0.000047 mL of Base (0.5 M KOH)
- [1:04:50] Stepping pH = 8.84
- [1:04:50] Dispensed 0.000047 mL of Base (0.5 M KOH)
- [1:04:55] Stepping pH = 8.97
- [1:05:11] Stirrer speed set to 0
- [1:05:37] Datapoint id 61 collected
- [1:05:37] Charge balance equation is out by -302.6%
- [1:05:37] Stirrer speed set to 60
- [1:05:42] pH 9.03 -> 9.05
- [1:05:42] Using cautious pH adjust
- [1:05:57] Stirrer speed set to 0
- [1:06:16] Datapoint id 62 collected
- [1:06:16] Charge balance equation is out by 100.0%
- [1:06:16] Argon flow rate set to 0
- [1:06:20] Titrator arm moved over Titration position