

CERTIFICATION OF CONFORMANCE

Title Page of Calibration Certificate Documentation

CUSTOMER:

Massachusetts Institute of Technology
350 Brookline St
Cambridge, MA 02139
UNITED STATES

PURCHASE ORDER #: 1201469


PCB ORDER #: 703025

QTY	ITEM	DESCRIPTION
1	352C22	ACCELEROMETER LW521801
1	378B02	377B02 WITH 426E01 (TEDS) 00173314
1	480B21	3 CHANNEL BATTERY SIGNAL CONDITIONER 00041481
2	070A02	SCOPE INPUT ADAPTOR

Notes:

1. This document certifies that the subject item(s) have been manufactured, repaired (if applicable), tested, or inspected in accordance with referenced purchase order and conform(s) to applicable specifications per PCB Quality Policy Manual.
2. Equipment used in validation is traceable to NIST and appropriate records are on file. The results stated on the calibration certificate relate only to the item referenced on the calibration certificate.
3. Calibrations comply with ISO 17025:2017 and ANSI/NCSL Z540.3-2006 except as noted on associated calibration certificate(s). Order placement is an acknowledgement and acceptance of measurement capability. Product is compliant with specification if measured value is within or equal to the specification tolerance. Product is not compliant with specification if measured value falls outside the specification tolerance. When a statement of conformity appears on a calibration certificate only the specification values listed are covered by the conformity statement. Out of tolerance conditions are noted on the calibration certificate. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
4. Calibrations are performed using processes having a test uncertainty ratio (TUR) of four or more times greater than the unit calibrated, unless otherwise noted on the calibration certificate. Calibration at 4:1 TUR provides reasonable confidence that the instrument is within product specifications.

Logistics Associate:



Date: 03/05/25

PCB PIEZOTRONICS, INC.

3425 WALDEN AVENUE, DEPEW, NY 14043 UNITED STATES

Phone: (716) 684-0001, Fax: (716) 684-0987

CERTIFICATION OF CONFORMANCE

Title Page of Calibration Certificate Documentation

CUSTOMER:


Massachusetts Institute of Technology
350 Brookline St
Cambridge, MA 02139
UNITED STATES

PURCHASE ORDER #: 1201469**PCB ORDER #:** 703025

QTY	ITEM	DESCRIPTION
1	086E80	IMPULSE HAMMER 00052563

Notes:

1. This document certifies that the subject item(s) have been manufactured, repaired (if applicable), tested, or inspected in accordance with referenced purchase order and conform(s) to applicable specifications per PCB Quality Policy Manual.
2. Equipment used in validation is traceable to NIST and appropriate records are on file. The results stated on the calibration certificate relate only to the item referenced on the calibration certificate.
3. Calibrations comply with ISO 17025:2017 and ANSI/NCCL Z540.3-2006 except as noted on associated calibration certificate(s). Order placement is an acknowledgement and acceptance of measurement capability. Product is compliant with specification if measured value is within or equal to the specification tolerance. Product is not compliant with specification if measured value falls outside the specification tolerance. When a statement of conformity appears on a calibration certificate only the specification values listed are covered by the conformity statement. Out of tolerance conditions are noted on the calibration certificate. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
4. Calibrations are performed using processes having a test uncertainty ratio (TUR) of four or more times greater than the unit calibrated, unless otherwise noted on the calibration certificate. Calibration at 4:1 TUR provides reasonable confidence that the instrument is within product specifications

Logistics Associate:**Date:** 03/27/25**PCB PIEZOTRONICS, INC.**

3425 WALDEN AVENUE, DEPEW, NY 14043 UNITED STATES

Phone: (716) 684-0001, Fax: (716) 684-0987

CERTIFICATION OF CONFORMANCE

Title Page of Calibration Certificate Documentation

CUSTOMER:

Massachusetts Institute of Technology
350 Brookline St
Cambridge, MA 02139
UNITED STATES

PURCHASE ORDER #: 1201469

PCB ORDER #: 703025

QTY	ITEM	DESCRIPTION
1	352A73	ACCELEROMETER LW521679

Notes:

1. This document certifies that the subject item(s) have been manufactured, repaired (if applicable), tested, or inspected in accordance with referenced purchase order and conform(s) to applicable specifications per PCB Quality Policy Manual.
2. Equipment used in validation is traceable to NIST and appropriate records are on file. The results stated on the calibration certificate relate only to the item referenced on the calibration certificate.
3. Calibrations comply with ISO 17025:2017 and ANSI/NCSL Z540.3-2006 except as noted on associated calibration certificate(s). Order placement is an acknowledgement and acceptance of measurement capability. Product is compliant with specification if measured value is within or equal to the specification tolerance. Product is not compliant with specification if measured value falls outside the specification tolerance. When a statement of conformity appears on a calibration certificate only the specification values listed are covered by the conformity statement. Out of tolerance conditions are noted on the calibration certificate. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
4. Calibrations are performed using processes having a test uncertainty ratio (TUR) of four or more times greater than the unit calibrated, unless otherwise noted on the calibration certificate. Calibration at 4:1 TUR provides reasonable confidence that the instrument is within product specifications

Logistics Associate:

Sheena W. Kinski

Date: 03/20/25

PCB PIEZOTRONICS, INC.

3425 WALDEN AVENUE, DEPEW, NY 14043 UNITED STATES

Phone: (716) 684-0001, Fax: (716) 684-0987

CERTIFICATION OF CONFORMANCE

Title Page of Calibration Certificate Documentation

CUSTOMER:


Massachusetts Institute of Technology
350 Brookline St
Cambridge, MA 02139
UNITED STATES

PURCHASE ORDER #: 1275582**PCB ORDER #: 737427**

QTY	ITEM	DESCRIPTION
2	378C10	377C10 WITH 426B03 (TEDS) 00175491, 00175492
2	002C10	10-32 PLUG TO BNC PLUG, 10FT

Notes:

1. This document certifies that the subject item(s) have been manufactured, repaired (if applicable), tested, or inspected in accordance with referenced purchase order and conform(s) to applicable specifications per PCB Quality Policy Manual.
2. Equipment used in validation is traceable to NIST and appropriate records are on file. The results stated on the calibration certificate relate only to the item referenced on the calibration certificate.
3. Calibrations comply with ISO 17025:2017 and ANSI/NCCL Z540.3-2006 except as noted on associated calibration certificate(s). Order placement is an acknowledgement and acceptance of measurement capability. Product is compliant with specification if measured value is within or equal to the specification tolerance. Product is not compliant with specification if measured value falls outside the specification tolerance. When a statement of conformity appears on a calibration certificate only the specification values listed are covered by the conformity statement. Out of tolerance conditions are noted on the calibration certificate. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
4. Calibrations are performed using processes having a test uncertainty ratio (TUR) of four or more times greater than the unit calibrated, unless otherwise noted on the calibration certificate. Calibration at 4:1 TUR provides reasonable confidence that the instrument is within product specifications

Logistics Associate:**Date:** 07/24/25**PCB PIEZOTRONICS, INC.**

3425 WALDEN AVENUE, DEPEW, NY 14043 UNITED STATES

Phone: (716) 684-0001, Fax: (716) 684-0987



CERTIFICATION OF CONFORMANCE

Title Page of Calibration Certificate Documentation

CUSTOMER:

Massachusetts Institute of Technology
350 Brookline St
Cambridge, MA 02139
UNITED STATES

PURCHASE ORDER #: 1201469

PCB ORDER #: 703025

QTY	ITEM	DESCRIPTION
1	086C01	HAMMER 00052635

Notes:

1. This document certifies that the subject item(s) have been manufactured, repaired (if applicable), tested, or inspected in accordance with referenced purchase order and conform(s) to applicable specifications per PCB Quality Policy Manual.
2. Equipment used in validation is traceable to NIST and appropriate records are on file. The results stated on the calibration certificate relate only to the item referenced on the calibration certificate.
3. Calibrations comply with ISO 17025:2017 and ANSI/NCSL Z540.3-2006 except as noted on associated calibration certificate(s). Order placement is an acknowledgement and acceptance of measurement capability. Product is compliant with specification if measured value is within or equal to the specification tolerance. Product is not compliant with specification if measured value falls outside the specification tolerance. When a statement of conformity appears on a calibration certificate only the specification values listed are covered by the conformity statement. Out of tolerance conditions are noted on the calibration certificate. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
4. Calibrations are performed using processes having a test uncertainty ratio (TUR) of four or more times greater than the unit calibrated, unless otherwise noted on the calibration certificate. Calibration at 4:1 TUR provides reasonable confidence that the instrument is within product specifications

Logistics Associate:

Date: 02/27/25

PCB PIEZOTRONICS, INC.

3425 WALDEN AVENUE, DEPEW, NY 14043 UNITED STATES

Phone: (716) 684-0001, Fax: (716) 684-0987

CERTIFICATION OF CONFORMANCE

Title Page of Calibration Certificate Documentation

CUSTOMER:

Massachusetts Institute of Technology
350 Brookline St
Cambridge, MA 02139
UNITED STATES

PURCHASE ORDER #: 1201469

PCB ORDER #: 703025

QTY	ITEM	DESCRIPTION
2	002T10	BNC PLUG TO BNC PLUG, 10FT
1	002T20	BNC PLUG TO BNC PLUG, 20FT
1	080A24	PETRO WAX

Notes:

1. This document certifies that the subject item(s) have been manufactured, repaired (if applicable), tested, or inspected in accordance with referenced purchase order and conform(s) to applicable specifications per PCB Quality Policy Manual.
2. Equipment used in validation is traceable to NIST and appropriate records are on file. The results stated on the calibration certificate relate only to the item referenced on the calibration certificate.
3. Calibrations comply with ISO 17025:2017 and ANSI/NC SL Z540.3-2006 except as noted on associated calibration certificate(s). Order placement is an acknowledgement and acceptance of measurement capability. Product is compliant with specification if measured value is within or equal to the specification tolerance. Product is not compliant with specification if measured value falls outside the specification tolerance. When a statement of conformity appears on a calibration certificate only the specification values listed are covered by the conformity statement. Out of tolerance conditions are noted on the calibration certificate. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
4. Calibrations are performed using processes having a test uncertainty ratio (TUR) of four or more times greater than the unit calibrated, unless otherwise noted on the calibration certificate. Calibration at 4:1 TUR provides reasonable confidence that the instrument is within product specifications

Logistics Associate:

Shawn Walmsley

Date: 02/14/25

PCB PIEZOTRONICS, INC.

3425 WALDEN AVENUE, DEPEW, NY 14043 UNITED STATES

Phone: (716) 684-0001, Fax: (716) 684-0987

CERTIFICATION OF CONFORMANCE

Title Page of Calibration Certificate Documentation

CUSTOMER:


Massachusetts Institute of Technology
350 Brookline St
Cambridge, MA 02139
UNITED STATES

PURCHASE ORDER #: 1345810**PCB ORDER #: 749834**

QTY	ITEM	DESCRIPTION
2	002C10	10-32 PLUG TO BNC PLUG, 10FT
2	378C10	377C10 WITH 426B03 (TEDS) 00178033, 00178034

Notes:

1. This document certifies that the subject item(s) have been manufactured, repaired (if applicable), tested, or inspected in accordance with referenced purchase order and conform(s) to applicable specifications per PCB Quality Policy Manual.
2. Equipment used in validation is traceable to NIST and appropriate records are on file. The results stated on the calibration certificate relate only to the item referenced on the calibration certificate.
3. Calibrations comply with ISO 17025:2017 and ANSI/NCSS Z540.3-2006 except as noted on associated calibration certificate(s). Order placement is an acknowledgement and acceptance of measurement capability. Product is compliant with specification if measured value is within or equal to the specification tolerance. Product is not compliant with specification if measured value falls outside the specification tolerance. When a statement of conformity appears on a calibration certificate only the specification values listed are covered by the conformity statement. Out of tolerance conditions are noted on the calibration certificate. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
4. Calibrations are performed using processes having a test uncertainty ratio (TUR) of four or more times greater than the unit calibrated, unless otherwise noted on the calibration certificate. Calibration at 4:1 TUR provides reasonable confidence that the instrument is within product specifications.

Logistics Associate:**Date:** 01/02/26**- ISO 9001 Certified / ISO 17025 Accredited / AS9100 Certified -****PCB PIEZOTRONICS, INC.**

3425 WALDEN AVENUE, DEPEW, NY 14043 UNITED STATES

Phone: (716) 684-0001, Fax: (716) 684-0987



CERTIFICATION OF CONFORMANCE

Title Page of Calibration Certificate Documentation

CUSTOMER:

Massachusetts Institute of Technology
350 Brookline St
Cambridge, MA 02139
UNITED STATES

PURCHASE ORDER #: 1382051

PCB ORDER #: 756076

QTY	ITEM	DESCRIPTION
1	356A03	TRIAXIAL ACCELEROMETER LW458895
1	378B02	377B02 WITH 426E01 (TEDS) 00179741
1	086E80	IMPULSE HAMMER 00054385

Notes:

1. This document certifies that the subject item(s) have been manufactured, repaired (if applicable), tested, or inspected in accordance with referenced purchase order and conform(s) to applicable specifications per PCB Quality Policy Manual.
2. Equipment used in validation is traceable to NIST and appropriate records are on file. The results stated on the calibration certificate relate only to the item referenced on the calibration certificate.
3. Calibrations comply with ISO 17025:2017 and ANSI/NCCL Z540.3-2006 except as noted on associated calibration certificate(s). Order placement is an acknowledgement and acceptance of measurement capability. Product is compliant with specification if measured value is within or equal to the specification tolerance. Product is not compliant with specification if measured value falls outside the specification tolerance. When a statement of conformity appears on a calibration certificate only the specification values listed are covered by the conformity statement. Out of tolerance conditions are noted on the calibration certificate. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
4. Calibrations are performed using processes having a test uncertainty ratio (TUR) of four or more times greater than the unit calibrated, unless otherwise noted on the calibration certificate. Calibration at 4:1 TUR provides reasonable confidence that the instrument is within product specifications

Logistics Associate:

Date: 05/28/26

- ISO 9001 Certified / ISO 17025 Accredited / AS9100 Certified -

PCB PIEZOTRONICS, INC.

3425 WALDEN AVENUE, DEPEW, NY 14043 UNITED STATES

Phone: (716) 684-0001, Fax: (716) 684-0987