

مؤسسة المهند للأجهزة الطبية
AL MOHANAD MEDICAL EQUIPMENT EST.



جودة أفضل لحياة أفضل
Good Quality for Better Life

تقرير الصيانة
Maintenance Report

No. 02695

Work Order No.: 1
Equipment: Accent MC240
Manufacturer: Corning Diagnostic
Machine No.: 1

Date: 29-11-2022
Location: Yamba Medical Center
Model: complex
Serial No.: BC7-15001095

Eng's Action

Requested By:

checked the device. Informed the doctor about the device
Informed about the device working. Device working well.
Tracked down to replace the lamp.
Please use - Acid Wash Detergent for washing probe.

The following Parts are

☐ Required

☐ Changed

Part #	Description	Qty.
	[Acid Wash / Detergent]	

Job Completed on 29 / 11 / 2022 At am/pm

Repair Time : From to

Travel Time : From to

Eng. Name / Sig.

- ☒ Work Completed, Unit Left Operational.
☐ Work Un Completed, Parts Required.

This Job is Done

Approved By Eng.

Clinic / Lab Incharge د. أحمد

Name & Date : أخصائي أول مختبرات