

مؤسسة المهندس للأجهزة الطبية  
AL MOHANAD MEDICAL EQUIPMENT EST.



جودة أفضل لحياة أفضل  
Good Quality for Better Life

تقرير الصيانة  
Maintenance Report

No. 02685

Work Order No.: 1  
Equipment : Nexgen PTS  
Manufacturer : Charles River Lab-  
Machine No. : 1

Date : 13-11-2022  
Location : King Fahad Hospital  
De Baha  
Model :  
Serial No.: 20251114

Eng's. Action

Requested By:

Calibrated the device and performed the qualification.  
Passed the kinetic noise spe test and Aural tests.  
Device working well. Trained the staff about the  
device. Next PPM - November 2023

The following Parts are

☐ Required

☐ Changed

Part #	Description	Qty.



Job Completed on 13 / 11 / 2022 At

am/pm

Repair Time :

From

to

Travel Time :

From

to

Eng. Name / Sig.

☒ Work Completed, Unit Left Operational.

☐ Work Un Completed, Parts Required.

This Job is Done

Clinic / Lab Incharge

Name & Date :

Approved By Eng.

Dr. Ismaeel Moustaf