

مؤسسة المهند للأجهزة الطبية  
AL MOHANAD MEDICAL EQUIPMENT EST.



جودة أفضل لحياة أفضل  
Good Quality for Better Life

تقرير الصيانة  
Maintenance Report

No. 02694

Work Order No.:  
Equipment: Arient M5240  
Manufacturer: Corning Diagnostic  
Machine No.: 1

Date: 29-11-2022  
Location: Kaya Almadinah medical complex  
Model:  
Serial No.: BC7-11000887

Eng's. Action

Requested By:

checked the device. The Dc and results are good.  
Performed PPM on the device.  
Device working well. Recommend to replace the  
filter and cuvette after 1 year of installation.

The following Parts are

☐ Required

☐ Changed

Part #	Description	Qty.

Job Completed on 29/11/22 At am/pm

Repair Time : From to

Travel Time : From to

Eng. Name / Sig.

☒ Work Completed, Unit Left Operational.

☐ Work Un Completed, Parts Required.

This Jop is Done

Clinic / Lab Incharge

Name & Date :

Approved By Eng.