

مؤسسة المهندس للأجهزة الطبية
AL MOHANAD MEDICAL EQUIPMENT EST.



جودة أفضل لحياة أفضل
Good Quality for Better Life

تقرير الصيانة
Maintenance Report

No. 02686

Work Order No.: 1
Equipment: Arient M 240
Manufacturer: Cormac Diagnostic
Machine No.: 1

Date: 13-11-2022
Location: Hashboul Medical Clinic
Model: chem
Serial No.: BC7-11000866

Eng's. Action

Requested By:

checked the device. Found issue in Cuvettes. Need to replace the part. The cuvettes and filter should be changed every 6-12 months according to manufacturer. Fixed the issue. Device working.

The following Parts are

☒ Required

☐ Changed

Part #	Description	Qty.
	Reaction Cuvette (8pc)	1
	Filter	2
	Lamp 20w [Always keep spare stock]	

Job Completed on 13 / 11 / 22 At am/pm

Repair Time : From to

Travel Time : From to

Eng. Name / Sig.

☒ Work Completed, Unit Left Operational.

☐ Work Un Completed, Parts Required.

This Job is Done

Clinic / Lab Incharge

Approved By Eng.

Name & Date :