

مؤسسة المهند للأجهزة الطبية  
AL MOHANAD MEDICAL EQUIPMENT EST.



جودة أفضل لحياة أفضل  
Good Quality for Better Life

تقرير الصيانة  
Maintenance Report

No. 02684

Work Order No.: 1  
Equipment: Nexgen PTS  
Manufacturer: Charles River Laboratories  
Machine No.: 1

Date: 08-11-2022  
Location: Regional Laboratory 2  
Central Board Bank-  
Tajf  
Model: 21431210  
Serial No.: 21431210

Eng's. Action

Requested By:

Installed the device at the lab including all accessories.  
Performed Qualification IQ, OQ, PQ on the device and handed  
the staff on the machine. Performed sample run and  
informed about the test procedure and device working.  
Machine working well.

The following Parts are

☒ Required

☐ Changed

Part #	Description	Qty.

Job Completed on 08 / 11 / 22 At am/pm

Repair Time : From to

Travel Time : From to

Eng. Name / Sig.

☒ Work Completed, Unit Left Operational.

☐ Work Un Completed, Parts Required.

This Jop is Done

Clinic / Lab Incharge

Name & Date :

Amani Muhendis  
Lab Specialist  
5420780

Approved By Eng.

8-11-2022