

# PROTOTYPE TEST REPORT

*CRITICAL NOTICE: This is a prototype test report generated according to the configured OIML test-report structure. It is NOT an officially approved legal certificate.*

## 1. General Information

**Report Number:** REP-20260827181652

**Date of Issue:** 2026-08-27 12:46:52 UTC

**Testing Laboratory:** MetraSure Demo Laboratory (SIH26035)

**Technician:** Tech



## 2. Instrument Identification

|                         |                                   |                                   |         |
|-------------------------|-----------------------------------|-----------------------------------|---------|
| <b>Manufacturer:</b>    | Mettler Toledo                    | <b>Accuracy Class:</b>            | III     |
| <b>Model / Type:</b>    | ICS689                            | <b>Verification Interval (e):</b> | 0.01 kg |
| <b>Serial Number:</b>   | SN-9821-MT                        | <b>Max Capacity:</b>              | 30.0 kg |
| <b>Instrument Type:</b> | Non-Automatic Weighing Instrument | <b>Min Capacity:</b>              | 0.2 kg  |

## 3. Overall Result

**FINAL VERDICT:** PASS

4. Detailed Measurement Results

Test Module: Accuracy

| Measurement ID | Calculated Error | MPE Limit  | Result | Standard Rule    |
|----------------|------------------|------------|--------|------------------|
| Seq #1         | 0.0010 kg        | ±0.0100 kg | PASS   | OIML R-76 (2006) |

Test Module: Repeatability

| Measurement ID | Calculated Error | MPE Limit  | Result | Standard Rule    |
|----------------|------------------|------------|--------|------------------|
| Seq #1         | 0.0100 kg        | ±0.0100 kg | PASS   | OIML R-76 (2006) |

Test Module: Eccentricity

| Measurement ID | Calculated Error | MPE Limit  | Result | Standard Rule    |
|----------------|------------------|------------|--------|------------------|
| Seq #1         | 0.0000 kg        | ±0.0100 kg | PASS   | OIML R-76 (2006) |
| Seq #2         | 0.0100 kg        | ±0.0100 kg | PASS   | OIML R-76 (2006) |
| Seq #3         | 0.0000 kg        | ±0.0100 kg | PASS   | OIML R-76 (2006) |
| Seq #4         | 0.0010 kg        | ±0.0100 kg | PASS   | OIML R-76 (2006) |
| Seq #5         | 0.0000 kg        | ±0.0100 kg | PASS   | OIML R-76 (2006) |

Test Module: Zero

| Measurement ID | Calculated Error | MPE Limit  | Result | Standard Rule    |
|----------------|------------------|------------|--------|------------------|
| Seq #1         | 0.0010 kg        | ±0.0025 kg | PASS   | OIML R-76 (2006) |

Test Module: Tare

| Measurement ID | Calculated Error | MPE Limit  | Result | Standard Rule    |
|----------------|------------------|------------|--------|------------------|
| Seq #1         | 0.0001 kg        | ±0.0100 kg | PASS   | OIML R-76 (2006) |

\_\_\_\_\_  
Technician Signature  
Tech

\_\_\_\_\_  
Reviewing Officer Signature  
System Administrator