

PROTOTYPE TEST REPORT

CRITICAL NOTICE: This is a prototype test report generated according to the configured OIML test-report structure. It is NOT an officially approved legal certificate.

1. General Information

Report Number: REP-20260905001509

Date of Issue: 2026-09-04 18:45:09 UTC

Testing Laboratory: MetraSure Demo Laboratory (SIH26035)

Technician: Tech



2. Instrument Identification

Manufacturer:	Mettler Torado	Accuracy Class:	III
Model / Type:	ICS689	Verification Interval (e):	0.01 kg
Serial Number:	SN-DEMO-202^	Max Capacity:	30.0 kg
Instrument Type:	NAWI	Min Capacity:	0.2 kg

3. Detailed Measurement Results

Test Module: Tare

Measurement ID	Calculated Error	MPE Limit	Result	Standard Rule
Seq #1	0.0030 kg	±0.0100 kg	PASS	OIML R-76 (2006)
Seq #2	0.0020 kg	±0.0100 kg	PASS	OIML R-76 (2006)

Test Module: Zero

Measurement ID	Calculated Error	MPE Limit	Result	Standard Rule
Seq #1	0.0010 kg	±0.0025 kg	PASS	OIML R-76 (2006)

Test Module: Eccentricity

Measurement ID	Calculated Error	MPE Limit	Result	Standard Rule
Seq #1	0.0010 kg	±0.0100 kg	PASS	OIML R-76 (2006)
Seq #2	0.0020 kg	±0.0100 kg	PASS	OIML R-76 (2006)
Seq #3	0.0030 kg	±0.0100 kg	PASS	OIML R-76 (2006)
Seq #4	0.0020 kg	±0.0100 kg	PASS	OIML R-76 (2006)
Seq #5	0.0010 kg	±0.0100 kg	PASS	OIML R-76 (2006)

Test Module: Repeatability

Measurement ID	Calculated Error	MPE Limit	Result	Standard Rule
Seq #1	0.0030 kg	±0.0100 kg	PASS	OIML R-76 (2006)

Test Module: Accuracy

Measurement ID	Calculated Error	MPE Limit	Result	Standard Rule
Seq #1	0.0030 kg	±0.0100 kg	PASS	OIML R-76 (2006)
Seq #2	0.0050 kg	±0.0100 kg	PASS	OIML R-76 (2006)

4. Overall Result

FINAL VERDICT: PASS

Technician Signature

Tech

Reviewing Officer Signature

System Administrator