

S01 - Nonconformity record

Contract DEMO-SUP-26; reference date 2026-10-01.

Item: Ultra low temperature laboratory freezer; quantity in this record: 2 devices.

Two freezers require alarm setting correction; one calibration certificate is missing. Final acceptance remains blocked.

Reference the actual receipt number generated by EPM. This file does not contain a forged signature or official stamp.

Evidence handling: upload this file to the matching workflow record. Preserve the generated reference and the decision history in EPM.