CITY OF MONROE
PLANNING & URBAN DEVELOPMENT
Community Development

P.O. Box 123
Monroe, LA 71201-0123
office: 318-329-2256
fax: 318-329-4151

Performance Characteristic Sheet

EFFECTIVE DATE: **February 1, 2022**

MANUFACTURER AND MODEL:

Make: ***SciAps***
Models: ***Model X-550***
X-Ray Source: ***Rhodium (Rh) or Gold (Au) Anode***

FIELD OPERATION GUIDANCE

ACTION LEVEL SETTING:

1.0 mg/cm²

OPERATING PARAMETERS:

Timed mode: fixed 10-second reading.

Quick mode: variable-time reading (approximately 2-6 seconds).

XRF CALIBRATION CHECK LIMITS:

0.8 to 1.2 mg/cm² (inclusive) on NIST SRM 2579 (1.02 mg/cm²)/NIST SRM 2573, or equivalent

SUBSTRATE CORRECTION:

Not applicable

INCONCLUSIVE RANGE OR THRESHOLD:

Au Anode (quick) **READING DESCRIPTION**	**SUBSTRATE**	**THRESHOLD (mg/cm²)**
Results not corrected for substrate bias on any substrate	Brick Concrete Drywall Metal Plaster Wood	1.0 1.0 1.0 1.0 1.0 1.0
Rh Anode (Timed or Quick), Au Anode (Timed) **READING DESCRIPTION**	**SUBSTRATE**	**THRESHOLD (mg/cm²)**
Results not corrected for substrate bias on any substrate	Brick Concrete Drywall Metal Plaster Wood	0.9 0.9 0.9 0.9 0.9 0.9

CITY OF MONROE
PLANNING & URBAN DEVELOPMENT
Community Development

P.O. Box 123
Monroe, LA 71201-0123
office: 318-329-2256
fax: 318-329-4151



Certificate of Training

Donta Anderson

Of City of Monroe has completed the SciAps instrument operator and Safety Training for the

X550 Pb HUD/EPA Lead Paint Analyzer by Mark Robert

SciAps

Instrument Operator Training and Safety Training for the X550 Pb HUD/EPA Lead Paint Analyzer

I confirm that the above-named individual has received the training listed on this certificate.

Mark Robert
Name

December 23, 2025
Date

Certified Trainer
Title

I certify that I have received the stated training and understand the content presented.

Donta Anderson
Name

December 23, 2025
Date

CITY OF MONROE
PLANNING & URBAN DEVELOPMENT
Community Development

P.O. Box 123
Monroe, LA 71201-0123
office: 318-329-2256
fax: 318-329-4151

STATE OF LOUISIANA

DEPARTMENT OF ENVIRONMENTAL QUALITY

certifies that

Donta K Anderson

Has complied with all requirements of the Louisiana Department of Environmental Quality and is authorized to perform the duties of

Lead Inspector

Accreditation No. DI249483 AI No. 249483

Date of Issuance December 16, 2025 Expiration December 16, 2026

Failure to comply with all applicable provisions of La. R.S. 30:2025.E. (1)(a) and La. R.S. 30:2025.F. (2)(a) may result in civil and/or criminal enforcement actions by the State.

Charles Finley
Public Participation & Permit Support Division
Office of Environmental Services

STATE OF LOUISIANA

DEPARTMENT OF ENVIRONMENTAL QUALITY

certifies that

Donta K Anderson

Has complied with all requirements of the Louisiana Department of Environmental Quality and is authorized to perform the duties of

Lead Risk Assessor

Accreditation No. DR249483 AI No. 249483

Date of Issuance December 16, 2025 Expiration December 16, 2026

Failure to comply with all applicable provisions of La. R.S. 30:2025.E. (1)(a) and La. R.S. 30:2025.F. (2)(a) may result in civil and/or criminal enforcement actions by the State.

Charles Finley
Public Participation & Permit Support Division
Office of Environmental Services

CITY OF MONROE
PLANNING & URBAN DEVELOPMENT
Community Development

P.O. Box 123
Monroe, LA 71201-0123
office: 318-329-2256
fax: 318-329-4151

LEAD SAFETY GLOSSARY

Abatement (Permanent Solutions) - The process of permanently removing lead hazards. This includes removing lead paint, replacing old windows or doors, or covering lead-contaminated soil with asphalt or concrete. (Note: Grass and sod are only temporary fixes).

Bare Soil - Any dirt on the property that isn't covered by grass, plants, or pavement. This includes sand in sandboxes.

Chewable & Impact Surfaces

- **Chewable:** Interior or exterior surfaces (like windowsills) that a child could bite.
- **Impact:** Surfaces that get hit or bumped repeatedly, such as doors and door frames.
- **Friction:** Surfaces that rub together, like window tracks or stair treads, which can create lead dust through wear and tear.

Deteriorated Paint - Paint that is in poor condition. This includes paint that is peeling, chipping, cracking, flaking, or turning into a chalky powder.

Dripline / Foundation Area - The ground within 3 feet of the outside walls of a building.

Interim Controls (Temporary Solutions) - Actions taken to temporarily reduce lead exposure. These include specialized cleaning, basic repairs, repainting, and regular monitoring to make sure the paint stays in good condition.

Lead-Based Paint Hazard - Any condition that could result in lead exposure. This includes damaged or peeling paint, lead dust on interior surfaces, or lead-contaminated soil that may negatively impact health.

Understanding Measurements
To keep families safe, experts measure lead in very small amounts.

UNIT	MEASUREMENT FOR	COMPARISON
Microgram (*ug*)	Dust	$1/1,000,000$ of a gram. (Extremely tiny).
***ug/ft*^{2}**	Dust on Surfaces	Used to measure lead dust on floors or sills.
ppm	Soil / Bulk Paint	"Parts per million." Used for dirt and thick paint samples.
ppb	Drinking Water	"Parts per billion." Used for water testing.
ug/dL	Blood Levels	Used to measure lead in a person's bloodstream.

Safety Standards

- **Lead-Based Paint:** Defined as any paint containing **1.0 *mg/cm*^{2}** or more of lead.
- **Soil (Play Areas):** Hazards start at **400 ppm**.
- **Soil (Rest of Yard):** Hazards start at **1,200 ppm**.

CITY OF MONROE
PLANNING & URBAN DEVELOPMENT
Community Development

P.O. Box 123
Monroe, LA 71201-0123
office: 318-329-2256
fax: 318-329-4151

ADDITIONAL RESOURCES

National Lead Information Center: 1-800-424-LEAD | epa.gov/lead

Centers for Disease Control (CDC): 1-800-CDC-INFO | cdc.gov/lead

Consumer Product Safety Commission: 1-800-638-2772 | cpsc.gov

Louisiana Department of Health: ldh.la.gov