



GT_Plate1_Multiplex_12_11_2024.pcrd

3/27/2025 9:43 AM

Report Information

User: BioRad/admin
Data File Name: GT_Plate1_Multiplex_12_11_2024.pcrd
Data File Path: C:\Users\CSMTSS1\Documents\Tortoises\Plates\Multi_SCNAG_Mt
Well Group Name: All Wells
Report Differs from Last Save: No

Run Setup

Run Information

Run Date: 12/11/2024 12:48 PM
Run User: admin
Run Type: User-defined
Plate File: Elgaria_Multiplex.prcl.pltd
ID:
Notes:
Sample Volume: 20
Temperature Control Mode: Calculated
Lid Temperature: 105
Base Serial Number: 795BR20744
Optical Head Serial Number: 795BR20744

Protocol

- 1: 95.0°C for 3:00
- 2: 95.0°C for 0:15
- 3: 62.0°C for 0:30
Plate Read
- 4: GOTO 2, 44 more times

Plate Display


	1	2	3	4	5	6	7	8	9	10	11	12
A	Std-1 0 VIC STD1	Std-1 0 VIC STD1	Unk-1 0 VIC 11.3_A	Unk-1 0 VIC 11.3_A	Unk-9 0 VIC 9.1_D	Unk-9 0 VIC 9.1_D	Unk-17 0 VIC 5.1_C	Unk-17 0 VIC 5.1_C	Unk-25 0 VIC 7.5_C	Unk-25 0 VIC 7.5_C	Unk-33 0 VIC 9.3_A	Unk-33 0 VIC 9.3_A
B	Std-2 0 VIC STD2	Std-2 0 VIC STD2	Unk-2 0 VIC 3.1_B	Unk-2 0 VIC 3.1_B	Unk-10 0 VIC 5.3_B	Unk-10 0 VIC 5.3_B	Unk-18 0 VIC 15.3_A	Unk-18 0 VIC 15.3_A	Unk-26 0 VIC 13.2_B	Unk-26 0 VIC 13.2_B	Unk-34 0 VIC 17.5_C	Unk-34 0 VIC 17.5_C
C	Std-3 0 VIC STD3	Std-3 0 VIC STD3	Unk-3 0 VIC 9.1_C	Unk-3 0 VIC 9.1_C	Unk-11 0 VIC 15.2_C	Unk-11 0 VIC 15.2_C	Unk-19 0 VIC 3.3_C	Unk-19 0 VIC 3.3_C	Unk-27 0 VIC 10.6_A	Unk-27 0 VIC 10.6_A	Unk-35 0 VIC 9.4_D	Unk-35 0 VIC 9.4_D
D	Std-4 0 VIC STD4	Std-4 0 VIC STD4	Unk-4 0 VIC 8.1_A	Unk-4 0 VIC 8.1_A	Unk-12 0 VIC 10.6_D	Unk-12 0 VIC 10.6_D	Unk-20 0 VIC 9.4_B	Unk-20 0 VIC 9.4_B	Unk-28 0 VIC 5.4_D	Unk-28 0 VIC 5.4_D	Unk-36 0 VIC 9.5_B	Unk-36 0 VIC 9.5_B


Plate Display

	1	2	3	4	5	6	7	8	9	10	11	12
E	Std-5 0 VIC STD5	Std-5 0 VIC STD5	Unk-5 0 VIC 12.2_A	Unk-5 0 VIC 12.2_A	Unk-13 0 VIC 9.2_C	Unk-13 0 VIC 9.2_C	Unk-21 0 VIC 13.3_B	Unk-21 0 VIC 13.3_B	Unk-29 0 VIC 2.3_C	Unk-29 0 VIC 2.3_C	Unk-37 0 VIC 10.4_D	Unk-37 0 VIC 10.4_D
F	Std-6 0 VIC STD6	Std-6 0 VIC STD6	Unk-6 0 VIC 15.5_A	Unk-6 0 VIC 15.5_A	Unk-14 0 VIC 14.5_D	Unk-14 0 VIC 14.5_D	Unk-22 0 VIC 15.2_A	Unk-22 0 VIC 15.2_A	Unk-30 0 VIC 17.3_D	Unk-30 0 VIC 17.3_D	Unk-38 0 VIC 5.2_C	Unk-38 0 VIC 5.2_C
G	*Unk 0 VIC STD7	*Unk 0 VIC STD7	Unk-7 0 VIC 3.1_C	Unk-7 0 VIC 3.1_C	Unk-15 0 VIC 17.3_A	Unk-15 0 VIC 17.3_A	Unk-23 0 VIC 10.6_B	Unk-23 0 VIC 10.6_B	Unk-31 0 VIC 2.4_B	Unk-31 0 VIC 2.4_B	Unk-39 0 VIC 12.4_A	Unk-39 0 VIC 12.4_A
H	NTC-1 0 VIC NEG	NTC-1 0 VIC NEG	Unk-8 0 VIC 7.1_D	Unk-8 0 VIC 7.1_D	Unk-16 0 VIC 13.4_D	Unk-16 0 VIC 13.4_D	Unk-24 0 VIC 17.2_B	Unk-24 0 VIC 17.2_B	Unk-32 0 VIC 3.2_B	Unk-32 0 VIC 3.2_B	Unk-40 0 VIC 7.4_D	Unk-40 0 VIC 7.4_D

Quantification

Step #: 3
Analysis Mode: Fluorophore
Baseline Setting: Baseline Subtracted Curve Fit
Cq Determination: Single Threshold
Baseline Method:
FAM: Auto Calculated
VIC: Auto Calculated
Threshold Setting:
FAM: 150.00, User Defined
VIC: 150.00, User Defined





Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
A01	FAM	0	Std-01	STD1	16.67	16.66	0.010	3.000E+07	7.477	3.00E+07	0.00E+00
A02	FAM	0	Std-01	STD1	16.65	16.66	0.010	3.000E+07	7.477	3.00E+07	0.00E+00
A03	FAM	0	Unkn-01	11.3_A	22.50	22.51	0.015	5.650E+05	5.752	5.61E+05	5.67E+03
A04	FAM	0	Unkn-01	11.3_A	22.52	22.51	0.015	5.569E+05	5.746	5.61E+05	5.67E+03
A05	FAM	0	Unkn-09	9.1_D	23.05	23.05	0.002	3.865E+05	5.587	3.86E+05	6.40E+02
A06	FAM	0	Unkn-09	9.1_D	23.06	23.05	0.002	3.856E+05	5.586	3.86E+05	6.40E+02
A07	FAM	0	Unkn-17	5.1_C	22.92	23.14	0.313	4.227E+05	5.626	3.67E+05	7.81E+04
A08	FAM	0	Unkn-17	5.1_C	23.36	23.14	0.313	3.122E+05	5.494	3.67E+05	7.81E+04
A09	FAM	0	Unkn-25	7.5_C	22.69	22.75	0.078	4.940E+05	5.694	4.76E+05	2.54E+04
A10	FAM	0	Unkn-25	7.5_C	22.80	22.75	0.078	4.580E+05	5.661	4.76E+05	2.54E+04
A11	FAM	0	Unkn-33	9.3_A	22.14	22.19	0.062	7.191E+05	5.857	6.98E+05	2.97E+04
A12	FAM	0	Unkn-33	9.3_A	22.23	22.19	0.062	6.772E+05	5.831	6.98E+05	2.97E+04
B01	FAM	0	Std-02	STD2	20.11	20.09	0.029	3.000E+06	6.477	3.00E+06	0.00E+00
B02	FAM	0	Std-02	STD2	20.07	20.09	0.029	3.000E+06	6.477	3.00E+06	0.00E+00
B03	FAM	0	Unkn-02	3.1_B	21.30	21.28	0.041	1.276E+06	6.106	1.30E+06	3.67E+04
B04	FAM	0	Unkn-02	3.1_B	21.25	21.28	0.041	1.328E+06	6.123	1.30E+06	3.67E+04
B05	FAM	0	Unkn-10	5.3_B	22.36	22.39	0.053	6.222E+05	5.794	6.07E+05	2.21E+04
B06	FAM	0	Unkn-10	5.3_B	22.43	22.39	0.053	5.909E+05	5.772	6.07E+05	2.21E+04
B07	FAM	0	Unkn-18	15.3_A	23.42	23.34	0.106	3.008E+05	5.478	3.17E+05	2.29E+04
B08	FAM	0	Unkn-18	15.3_A	23.27	23.34	0.106	3.333E+05	5.523	3.17E+05	2.29E+04
B09	FAM	0	Unkn-26	13.2_B	22.70	22.74	0.068	4.933E+05	5.693	4.78E+05	2.22E+04
B10	FAM	0	Unkn-26	13.2_B	22.79	22.74	0.068	4.619E+05	5.665	4.78E+05	2.22E+04
B11	FAM	0	Unkn-34	17.5_C	22.88	22.86	0.026	4.356E+05	5.639	4.41E+05	7.81E+03
B12	FAM	0	Unkn-34	17.5_C	22.84	22.86	0.026	4.467E+05	5.650	4.41E+05	7.81E+03
C01	FAM	0	Std-03	STD3	23.48	23.48	0.011	3.000E+05	5.477	3.00E+05	0.00E+00
C02	FAM	0	Std-03	STD3	23.47	23.48	0.011	3.000E+05	5.477	3.00E+05	0.00E+00
C03	FAM	0	Unkn-03	9.1_C	22.33	22.24	0.125	6.335E+05	5.802	6.74E+05	5.73E+04
C04	FAM	0	Unkn-03	9.1_C	22.15	22.24	0.125	7.146E+05	5.854	6.74E+05	5.73E+04

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
C05	FAM	0	Unkn-11	15.2_C	22.79	22.78	0.017	4.612E+05	5.664	4.65E+05	5.56E+03
C06	FAM	0	Unkn-11	15.2_C	22.77	22.78	0.017	4.690E+05	5.671	4.65E+05	5.56E+03
C07	FAM	0	Unkn-19	3.3_C	24.47	24.27	0.283	1.468E+05	5.167	1.70E+05	3.27E+04
C08	FAM	0	Unkn-19	3.3_C	24.07	24.27	0.283	1.931E+05	5.286	1.70E+05	3.27E+04
C09	FAM	0	Unkn-27	10.6_A	23.32	23.26	0.084	3.210E+05	5.506	3.35E+05	1.92E+04
C10	FAM	0	Unkn-27	10.6_A	23.21	23.26	0.084	3.481E+05	5.542	3.35E+05	1.92E+04
C11	FAM	0	Unkn-35	9.4_D	23.77	23.59	0.253	2.362E+05	5.373	2.69E+05	4.63E+04
C12	FAM	0	Unkn-35	9.4_D	23.42	23.59	0.253	3.016E+05	5.479	2.69E+05	4.63E+04
D01	FAM	0	Std-04	STD4	26.78	26.75	0.030	3.000E+04	4.477	3.00E+04	0.00E+00
D02	FAM	0	Std-04	STD4	26.73	26.75	0.030	3.000E+04	4.477	3.00E+04	0.00E+00
D03	FAM	0	Unkn-04	8.1_A	23.86	23.86	0.010	2.231E+05	5.348	2.22E+05	1.57E+03
D04	FAM	0	Unkn-04	8.1_A	23.87	23.86	0.010	2.209E+05	5.344	2.22E+05	1.57E+03
D05	FAM	0	Unkn-12	10.6_D	22.21	22.00	0.297	6.867E+05	5.837	8.01E+05	1.61E+05
D06	FAM	0	Unkn-12	10.6_D	21.79	22.00	0.297	9.148E+05	5.961	8.01E+05	1.61E+05
D07	FAM	0	Unkn-20	9.4_B	23.67	23.63	0.058	2.528E+05	5.403	2.60E+05	1.03E+04
D08	FAM	0	Unkn-20	9.4_B	23.59	23.63	0.058	2.674E+05	5.427	2.60E+05	1.03E+04
D09	FAM	0	Unkn-28	5.4_D	23.27	23.25	0.030	3.320E+05	5.521	3.37E+05	7.01E+03
D10	FAM	0	Unkn-28	5.4_D	23.23	23.25	0.030	3.419E+05	5.534	3.37E+05	7.01E+03
D11	FAM	0	Unkn-36	9.5_B	22.10	22.05	0.068	7.410E+05	5.870	7.66E+05	3.55E+04
D12	FAM	0	Unkn-36	9.5_B	22.00	22.05	0.068	7.912E+05	5.898	7.66E+05	3.55E+04
E01	FAM	0	Std-05	STD5	30.13	30.05	0.105	3.000E+03	3.477	3.00E+03	0.00E+00
E02	FAM	0	Std-05	STD5	29.98	30.05	0.105	3.000E+03	3.477	3.00E+03	0.00E+00
E03	FAM	0	Unkn-05	12.2_A	21.23	20.69	0.750	1.347E+06	6.129	2.06E+06	1.01E+06
E04	FAM	0	Unkn-05	12.2_A	20.16	20.69	0.750	2.781E+06	6.444	2.06E+06	1.01E+06
E05	FAM	0	Unkn-13	9.2_C	23.04	23.08	0.052	3.893E+05	5.590	3.80E+05	1.36E+04
E06	FAM	0	Unkn-13	9.2_C	23.12	23.08	0.052	3.701E+05	5.568	3.80E+05	1.36E+04
E07	FAM	0	Unkn-21	13.3_B	23.24	23.42	0.250	3.398E+05	5.531	3.03E+05	5.15E+04
E08	FAM	0	Unkn-21	13.3_B	23.59	23.42	0.250	2.669E+05	5.426	3.03E+05	5.15E+04
E09	FAM	0	Unkn-29	2.3_C	22.36	22.38	0.032	6.194E+05	5.792	6.10E+05	1.32E+04
E10	FAM	0	Unkn-29	2.3_C	22.41	22.38	0.032	6.007E+05	5.779	6.10E+05	1.32E+04
E11	FAM	0	Unkn-37	10.4_D	22.75	22.85	0.135	4.740E+05	5.676	4.45E+05	4.10E+04
E12	FAM	0	Unkn-37	10.4_D	22.94	22.85	0.135	4.161E+05	5.619	4.45E+05	4.10E+04
F01	FAM	0	Std-06	STD6	33.76	33.61	0.207	3.000E+02	2.477	3.00E+02	0.00E+00
F02	FAM	0	Std-06	STD6	33.47	33.61	0.207	3.000E+02	2.477	3.00E+02	0.00E+00
F03	FAM	0	Unkn-06	15.5_A	20.21	19.69	0.741	2.694E+06	6.430	4.10E+06	1.99E+06
F04	FAM	0	Unkn-06	15.5_A	19.16	19.69	0.741	5.511E+06	6.741	4.10E+06	1.99E+06
F05	FAM	0	Unkn-14	14.5_D	22.64	22.81	0.239	5.115E+05	5.709	4.59E+05	7.47E+04
F06	FAM	0	Unkn-14	14.5_D	22.98	22.81	0.239	4.059E+05	5.608	4.59E+05	7.47E+04
F07	FAM	0	Unkn-22	15.2_A	23.09	23.12	0.049	3.775E+05	5.577	3.69E+05	1.23E+04
F08	FAM	0	Unkn-22	15.2_A	23.16	23.12	0.049	3.600E+05	5.556	3.69E+05	1.23E+04
F09	FAM	0	Unkn-30	17.3_D	23.24	23.32	0.115	3.394E+05	5.531	3.21E+05	2.53E+04
F10	FAM	0	Unkn-30	17.3_D	23.41	23.32	0.115	3.036E+05	5.482	3.21E+05	2.53E+04
F11	FAM	0	Unkn-38	5.2_C	23.05	22.91	0.198	3.876E+05	5.588	4.29E+05	5.79E+04
F12	FAM	0	Unkn-38	5.2_C	22.77	22.91	0.198	4.695E+05	5.672	4.29E+05	5.79E+04
G03	FAM	0	Unkn-07	3.1_C	22.62	22.53	0.126	5.195E+05	5.716	5.53E+05	4.76E+04
G04	FAM	0	Unkn-07	3.1_C	22.44	22.53	0.126	5.868E+05	5.768	5.53E+05	4.76E+04
G05	FAM	0	Unkn-15	17.3_A	21.58	21.63	0.080	1.059E+06	6.025	1.02E+06	5.59E+04
G06	FAM	0	Unkn-15	17.3_A	21.69	21.63	0.080	9.799E+05	5.991	1.02E+06	5.59E+04
G07	FAM	0	Unkn-23	10.6_B	23.75	23.48	0.388	2.398E+05	5.380	2.94E+05	7.71E+04

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
G08	FAM	0	Unkn-23	10.6_B	23.20	23.48	0.388	3.488E+05	5.543	2.94E+05	7.71E+04
G09	FAM	0	Unkn-31	2.4_B	23.34	23.18	0.219	3.177E+05	5.502	3.55E+05	5.30E+04
G10	FAM	0	Unkn-31	2.4_B	23.03	23.18	0.219	3.927E+05	5.594	3.55E+05	5.30E+04
G11	FAM	0	Unkn-39	12.4_A	22.61	22.61	0.003	5.240E+05	5.719	5.23E+05	1.22E+03
G12	FAM	0	Unkn-39	12.4_A	22.61	22.61	0.003	5.223E+05	5.718	5.23E+05	1.22E+03
H01	FAM	0	NTC-01	NEG	N/A	0.00	0.000	N/A	N/A	0.00E+00	0.00E+00
H02	FAM	0	NTC-01	NEG	N/A	0.00	0.000	N/A	N/A	0.00E+00	0.00E+00
H03	FAM	0	Unkn-08	7.1_D	22.31	22.23	0.109	6.414E+05	5.807	6.77E+05	5.03E+04
H04	FAM	0	Unkn-08	7.1_D	22.16	22.23	0.109	7.125E+05	5.853	6.77E+05	5.03E+04
H05	FAM	0	Unkn-16	13.4_D	23.35	23.38	0.036	3.154E+05	5.499	3.10E+05	7.67E+03
H06	FAM	0	Unkn-16	13.4_D	23.40	23.38	0.036	3.046E+05	5.484	3.10E+05	7.67E+03
H07	FAM	0	Unkn-24	17.2_B	22.98	23.18	0.282	4.057E+05	5.608	3.57E+05	6.84E+04
H08	FAM	0	Unkn-24	17.2_B	23.38	23.18	0.282	3.090E+05	5.490	3.57E+05	6.84E+04
H09	FAM	0	Unkn-32	3.2_B	23.21	22.99	0.319	3.463E+05	5.539	4.09E+05	8.83E+04
H10	FAM	0	Unkn-32	3.2_B	22.76	22.99	0.319	4.712E+05	5.673	4.09E+05	8.83E+04
H11	FAM	0	Unkn-40	7.4_D	22.79	22.84	0.066	4.621E+05	5.665	4.48E+05	2.03E+04
H12	FAM	0	Unkn-40	7.4_D	22.88	22.84	0.066	4.334E+05	5.637	4.48E+05	2.03E+04
A01	VIC		Std-01	STD1	18.20	18.18	0.026	3.000E+07	7.477	3.00E+07	0.00E+00
A02	VIC		Std-01	STD1	18.16	18.18	0.026	3.000E+07	7.477	3.00E+07	0.00E+00
A03	VIC		Unkn-01	11.3_A	28.01	27.95	0.086	3.543E+04	4.549	3.70E+04	2.19E+03
A04	VIC		Unkn-01	11.3_A	27.89	27.95	0.086	3.853E+04	4.586	3.70E+04	2.19E+03
A05	VIC		Unkn-09	9.1_D	29.09	29.09	0.002	1.684E+04	4.226	1.69E+04	2.64E+01
A06	VIC		Unkn-09	9.1_D	29.09	29.09	0.002	1.688E+04	4.227	1.69E+04	2.64E+01
A07	VIC		Unkn-17	5.1_C	28.81	28.97	0.230	2.047E+04	4.311	1.84E+04	2.91E+03
A08	VIC		Unkn-17	5.1_C	29.13	28.97	0.230	1.636E+04	4.214	1.84E+04	2.91E+03
A09	VIC		Unkn-25	7.5_C	28.91	28.91	0.002	1.908E+04	4.281	1.91E+04	3.07E+01
A10	VIC		Unkn-25	7.5_C	28.90	28.91	0.002	1.913E+04	4.282	1.91E+04	3.07E+01
A11	VIC		Unkn-33	9.3_A	28.05	28.05	0.004	3.439E+04	4.536	3.44E+04	8.64E+01
A12	VIC		Unkn-33	9.3_A	28.05	28.05	0.004	3.451E+04	4.538	3.44E+04	8.64E+01
B01	VIC		Std-02	STD2	21.51	21.47	0.052	3.000E+06	6.477	3.00E+06	0.00E+00
B02	VIC		Std-02	STD2	21.43	21.47	0.052	3.000E+06	6.477	3.00E+06	0.00E+00
B03	VIC		Unkn-02	3.1_B	27.45	27.41	0.050	5.221E+04	4.718	5.35E+04	1.84E+03
B04	VIC		Unkn-02	3.1_B	27.38	27.41	0.050	5.481E+04	4.739	5.35E+04	1.84E+03
B05	VIC		Unkn-10	5.3_B	28.62	28.73	0.161	2.330E+04	4.367	2.16E+04	2.39E+03
B06	VIC		Unkn-10	5.3_B	28.85	28.73	0.161	1.992E+04	4.299	2.16E+04	2.39E+03
B07	VIC		Unkn-18	15.3_A	29.41	29.32	0.129	1.349E+04	4.130	1.44E+04	1.28E+03
B08	VIC		Unkn-18	15.3_A	29.23	29.32	0.129	1.530E+04	4.185	1.44E+04	1.28E+03
B09	VIC		Unkn-26	13.2_B	28.95	28.88	0.106	1.852E+04	4.268	1.95E+04	1.42E+03
B10	VIC		Unkn-26	13.2_B	28.80	28.88	0.106	2.052E+04	4.312	1.95E+04	1.42E+03
B11	VIC		Unkn-34	17.5_C	28.13	28.13	0.004	3.262E+04	4.514	3.26E+04	8.62E+01
B12	VIC		Unkn-34	17.5_C	28.14	28.13	0.004	3.250E+04	4.512	3.26E+04	8.62E+01
C01	VIC		Std-03	STD3	25.05	25.05	0.001	3.000E+05	5.477	3.00E+05	0.00E+00
C02	VIC		Std-03	STD3	25.05	25.05	0.001	3.000E+05	5.477	3.00E+05	0.00E+00
C03	VIC		Unkn-03	9.1_C	27.98	27.86	0.172	3.606E+04	4.557	3.93E+04	4.64E+03
C04	VIC		Unkn-03	9.1_C	27.74	27.86	0.172	4.262E+04	4.630	3.93E+04	4.64E+03
C05	VIC		Unkn-11	15.2_C	29.04	29.02	0.019	1.744E+04	4.242	1.76E+04	2.35E+02
C06	VIC		Unkn-11	15.2_C	29.01	29.02	0.019	1.778E+04	4.250	1.76E+04	2.35E+02
C07	VIC		Unkn-19	3.3_C	30.14	29.90	0.343	8.175E+03	3.912	9.80E+03	2.30E+03
C08	VIC		Unkn-19	3.3_C	29.65	29.90	0.343	1.142E+04	4.058	9.80E+03	2.30E+03

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
C09	VIC		Unkn-27	10.6_A	29.26	29.20	0.083	1.502E+04	4.177	1.57E+04	8.96E+02
C10	VIC		Unkn-27	10.6_A	29.14	29.20	0.083	1.629E+04	4.212	1.57E+04	8.96E+02
C11	VIC		Unkn-35	9.4_D	29.38	29.19	0.255	1.383E+04	4.141	1.58E+04	2.76E+03
C12	VIC		Unkn-35	9.4_D	29.01	29.19	0.255	1.774E+04	4.249	1.58E+04	2.76E+03
D01	VIC		Std-04	STD4	28.37	28.38	0.006	3.000E+04	4.477	3.00E+04	0.00E+00
D02	VIC		Std-04	STD4	28.38	28.38	0.006	3.000E+04	4.477	3.00E+04	0.00E+00
D03	VIC		Unkn-04	8.1_A	28.17	28.18	0.017	3.167E+04	4.501	3.14E+04	3.59E+02
D04	VIC		Unkn-04	8.1_A	28.20	28.18	0.017	3.116E+04	4.494	3.14E+04	3.59E+02
D05	VIC		Unkn-12	10.6_D	28.67	28.50	0.240	2.247E+04	4.352	2.54E+04	4.19E+03
D06	VIC		Unkn-12	10.6_D	28.33	28.50	0.240	2.839E+04	4.453	2.54E+04	4.19E+03
D07	VIC		Unkn-20	9.4_B	29.71	29.73	0.030	1.098E+04	4.041	1.08E+04	2.22E+02
D08	VIC		Unkn-20	9.4_B	29.75	29.73	0.030	1.067E+04	4.028	1.08E+04	2.22E+02
D09	VIC		Unkn-28	5.4_D	29.10	29.02	0.114	1.675E+04	4.224	1.77E+04	1.39E+03
D10	VIC		Unkn-28	5.4_D	28.94	29.02	0.114	1.872E+04	4.272	1.77E+04	1.39E+03
D11	VIC		Unkn-36	9.5_B	29.16	29.12	0.058	1.600E+04	4.204	1.65E+04	6.57E+02
D12	VIC		Unkn-36	9.5_B	29.08	29.12	0.058	1.693E+04	4.229	1.65E+04	6.57E+02
E01	VIC		Std-05	STD5	31.63	31.55	0.107	3.000E+03	3.477	3.00E+03	0.00E+00
E02	VIC		Std-05	STD5	31.48	31.55	0.107	3.000E+03	3.477	3.00E+03	0.00E+00
E03	VIC		Unkn-05	12.2_A	27.86	27.50	0.509	3.928E+04	4.594	5.19E+04	1.78E+04
E04	VIC		Unkn-05	12.2_A	27.14	27.50	0.509	6.452E+04	4.810	5.19E+04	1.78E+04
E05	VIC		Unkn-13	9.2_C	28.67	28.72	0.074	2.246E+04	4.351	2.17E+04	1.10E+03
E06	VIC		Unkn-13	9.2_C	28.78	28.72	0.074	2.090E+04	4.320	2.17E+04	1.10E+03
E07	VIC		Unkn-21	13.3_B	29.03	29.23	0.287	1.756E+04	4.245	1.54E+04	3.03E+03
E08	VIC		Unkn-21	13.3_B	29.43	29.23	0.287	1.328E+04	4.123	1.54E+04	3.03E+03
E09	VIC		Unkn-29	2.3_C	28.54	28.55	0.022	2.464E+04	4.392	2.44E+04	3.77E+02
E10	VIC		Unkn-29	2.3_C	28.57	28.55	0.022	2.411E+04	4.382	2.44E+04	3.77E+02
E11	VIC		Unkn-37	10.4_D	28.53	28.66	0.184	2.470E+04	4.393	2.27E+04	2.87E+03
E12	VIC		Unkn-37	10.4_D	28.79	28.66	0.184	2.064E+04	4.315	2.27E+04	2.87E+03
F01	VIC		Std-06	STD6	34.69	34.86	0.232	3.000E+02	2.477	3.00E+02	0.00E+00
F02	VIC		Std-06	STD6	35.02	34.86	0.232	3.000E+02	2.477	3.00E+02	0.00E+00
F03	VIC		Unkn-06	15.5_A	26.49	26.09	0.565	1.013E+05	5.006	1.39E+05	5.26E+04
F04	VIC		Unkn-06	15.5_A	25.69	26.09	0.565	1.758E+05	5.245	1.39E+05	5.26E+04
F05	VIC		Unkn-14	14.5_D	28.85	28.96	0.148	1.986E+04	4.298	1.85E+04	1.89E+03
F06	VIC		Unkn-14	14.5_D	29.06	28.96	0.148	1.719E+04	4.235	1.85E+04	1.89E+03
F07	VIC		Unkn-22	15.2_A	29.34	29.29	0.065	1.418E+04	4.152	1.46E+04	6.58E+02
F08	VIC		Unkn-22	15.2_A	29.25	29.29	0.065	1.511E+04	4.179	1.46E+04	6.58E+02
F09	VIC		Unkn-30	17.3_D	29.05	29.09	0.059	1.727E+04	4.237	1.68E+04	6.78E+02
F10	VIC		Unkn-30	17.3_D	29.14	29.09	0.059	1.631E+04	4.212	1.68E+04	6.78E+02
F11	VIC		Unkn-38	5.2_C	28.66	28.67	0.014	2.268E+04	4.356	2.25E+04	2.14E+02
F12	VIC		Unkn-38	5.2_C	28.68	28.67	0.014	2.238E+04	4.350	2.25E+04	2.14E+02
G03	VIC		Unkn-07	3.1_C	28.45	28.40	0.062	2.621E+04	4.418	2.70E+04	1.16E+03
G04	VIC		Unkn-07	3.1_C	28.36	28.40	0.062	2.785E+04	4.445	2.70E+04	1.16E+03
G05	VIC		Unkn-15	17.3_A	27.58	27.62	0.052	4.761E+04	4.678	4.64E+04	1.67E+03
G06	VIC		Unkn-15	17.3_A	27.66	27.62	0.052	4.525E+04	4.656	4.64E+04	1.67E+03
G07	VIC		Unkn-23	10.6_B	29.73	29.50	0.323	1.082E+04	4.034	1.28E+04	2.83E+03
G08	VIC		Unkn-23	10.6_B	29.28	29.50	0.323	1.482E+04	4.171	1.28E+04	2.83E+03
G09	VIC		Unkn-31	2.4_B	29.38	29.20	0.249	1.382E+04	4.141	1.57E+04	2.69E+03
G10	VIC		Unkn-31	2.4_B	29.02	29.20	0.249	1.762E+04	4.246	1.57E+04	2.69E+03
G11	VIC		Unkn-39	12.4_A	28.53	28.62	0.121	2.469E+04	4.393	2.33E+04	1.94E+03

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
G12	VIC		Unkn-39	12.4_A	28.71	28.62	0.121	2.194E+04	4.341	2.33E+04	1.94E+03
H01	VIC		NTC-01	NEG	36.80	37.75	1.344	N/A	N/A	N/A	N/A
H02	VIC		NTC-01	NEG	38.70	37.75	1.344	N/A	N/A	N/A	N/A
H03	VIC		Unkn-08	7.1_D	28.47	28.39	0.106	2.586E+04	4.413	2.73E+04	2.00E+03
H04	VIC		Unkn-08	7.1_D	28.32	28.39	0.106	2.868E+04	4.458	2.73E+04	2.00E+03
H05	VIC		Unkn-16	13.4_D	29.03	29.09	0.085	1.752E+04	4.244	1.68E+04	9.82E+02
H06	VIC		Unkn-16	13.4_D	29.15	29.09	0.085	1.613E+04	4.208	1.68E+04	9.82E+02
H07	VIC		Unkn-24	17.2_B	29.61	29.69	0.113	1.177E+04	4.071	1.12E+04	8.69E+02
H08	VIC		Unkn-24	17.2_B	29.77	29.69	0.113	1.054E+04	4.023	1.12E+04	8.69E+02
H09	VIC		Unkn-32	3.2_B	29.34	29.20	0.202	1.419E+04	4.152	1.57E+04	2.18E+03
H10	VIC		Unkn-32	3.2_B	29.05	29.20	0.202	1.727E+04	4.237	1.57E+04	2.18E+03
H11	VIC		Unkn-40	7.4_D	28.59	28.72	0.187	2.381E+04	4.377	2.18E+04	2.81E+03
H12	VIC		Unkn-40	7.4_D	28.85	28.72	0.187	1.984E+04	4.297	2.18E+04	2.81E+03

Bar Chart

Normalized expression analysis is not possible, either because no target is assigned as a reference or the selected target(s) is not a

Target Names

Name	Full Name	Reference	Auto Efficiency	Efficiency
0	0	False	Yes	98.1%

Sample Names

Name	Full Name	Control
10.4_D	10.4_D	No
10.6_A	10.6_A	No
10.6_B	10.6_B	No
10.6_D	10.6_D	No
11.3_A	11.3_A	No
12.2_A	12.2_A	No
12.4_A	12.4_A	No
13.2_B	13.2_B	No
13.3_B	13.3_B	No
13.4_D	13.4_D	No
14.5_D	14.5_D	No
15.2_A	15.2_A	No
15.2_C	15.2_C	No
15.3_A	15.3_A	No
15.5_A	15.5_A	No
17.2_B	17.2_B	No
17.3_A	17.3_A	No
17.3_D	17.3_D	No
17.5_C	17.5_C	No
2.3_C	2.3_C	No
2.4_B	2.4_B	No

Sample Names

Name	Full Name	Control
3.1_B	3.1_B	No
3.1_C	3.1_C	No
3.2_B	3.2_B	No
3.3_C	3.3_C	No
5.1_C	5.1_C	No
5.2_C	5.2_C	No
5.3_B	5.3_B	No
5.4_D	5.4_D	No
7.1_D	7.1_D	No
7.4_D	7.4_D	No
7.5_C	7.5_C	No
8.1_A	8.1_A	No
9.1_C	9.1_C	No
9.1_D	9.1_D	No
9.2_C	9.2_C	No
9.3_A	9.3_A	No
9.4_B	9.4_B	No
9.4_D	9.4_D	No
9.5_B	9.5_B	No
STD1	STD1	No
STD2	STD2	No
STD3	STD3	No
STD4	STD4	No
STD5	STD5	No
STD6	STD6	No

Gene Expression - Bar Chart Data

Target	Sample	Control	Expression	Expression SEM	Corrected Expression SEM	Mean Cq	Cq SEM	P-Value
0	10.4_D		N/A	N/A	N/A	22.85	0.09545	N/A
0	10.6_A		N/A	N/A	N/A	23.26	0.05942	N/A
0	10.6_B		N/A	N/A	N/A	23.48	0.27409	N/A
0	10.6_D		N/A	N/A	N/A	22.00	0.20983	N/A
0	11.3_A		N/A	N/A	N/A	22.51	0.01046	N/A
0	12.2_A		N/A	N/A	N/A	20.69	0.53061	N/A
0	12.4_A		N/A	N/A	N/A	22.61	0.00241	N/A
0	13.2_B		N/A	N/A	N/A	22.74	0.04811	N/A
0	13.3_B		N/A	N/A	N/A	23.42	0.17662	N/A
0	13.4_D		N/A	N/A	N/A	23.38	0.02560	N/A
0	14.5_D		N/A	N/A	N/A	22.81	0.16915	N/A
0	15.2_A		N/A	N/A	N/A	23.12	0.03457	N/A
0	15.2_C		N/A	N/A	N/A	22.78	0.01236	N/A
0	15.3_A		N/A	N/A	N/A	23.34	0.07494	N/A
0	15.5_A		N/A	N/A	N/A	19.69	0.52371	N/A
0	17.2_B		N/A	N/A	N/A	23.18	0.19928	N/A
0	17.3_A		N/A	N/A	N/A	21.63	0.05673	N/A
0	17.3_D		N/A	N/A	N/A	23.32	0.08143	N/A

Gene Expression - Bar Chart Data

Target	Sample	Control	Expression	Expression SEM	Corrected Expression SEM	Mean Cq	Cq SEM	P-Value
0	17.5_C		N/A	N/A	N/A	22.86	0.01833	N/A
0	2.3_C		N/A	N/A	N/A	22.38	0.02239	N/A
0	2.4_B		N/A	N/A	N/A	23.18	0.15509	N/A
0	3.1_B		N/A	N/A	N/A	21.28	0.02921	N/A
0	3.1_C		N/A	N/A	N/A	22.53	0.08920	N/A
0	3.2_B		N/A	N/A	N/A	22.99	0.22525	N/A
0	3.3_C		N/A	N/A	N/A	24.27	0.20032	N/A
0	5.1_C		N/A	N/A	N/A	23.14	0.22159	N/A
0	5.2_C		N/A	N/A	N/A	22.91	0.14029	N/A
0	5.3_B		N/A	N/A	N/A	22.39	0.03778	N/A
0	5.4_D		N/A	N/A	N/A	23.25	0.02154	N/A
0	7.1_D		N/A	N/A	N/A	22.23	0.07693	N/A
0	7.4_D		N/A	N/A	N/A	22.84	0.04681	N/A
0	7.5_C		N/A	N/A	N/A	22.75	0.05535	N/A
0	8.1_A		N/A	N/A	N/A	23.86	0.00730	N/A
0	9.1_C		N/A	N/A	N/A	22.24	0.08806	N/A
0	9.1_D		N/A	N/A	N/A	23.05	0.00171	N/A
0	9.2_C		N/A	N/A	N/A	23.08	0.03706	N/A
0	9.3_A		N/A	N/A	N/A	22.19	0.04397	N/A
0	9.4_B		N/A	N/A	N/A	23.63	0.04086	N/A
0	9.4_D		N/A	N/A	N/A	23.59	0.17888	N/A
0	9.5_B		N/A	N/A	N/A	22.05	0.04799	N/A
0	STD1		N/A	N/A	N/A	16.66	0.00691	N/A
0	STD2		N/A	N/A	N/A	20.09	0.02066	N/A
0	STD3		N/A	N/A	N/A	23.48	0.00780	N/A
0	STD4		N/A	N/A	N/A	26.75	0.02119	N/A
0	STD5		N/A	N/A	N/A	30.05	0.07450	N/A
0	STD6		N/A	N/A	N/A	33.61	0.14608	N/A

QC Parameters

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Negative control with a Cq less than	38	True		False	
NTC with a Cq less than	38	True	VIC:H1.	False	
NRT with a Cq less than	38	True		False	
Positive control with a Cq greater than	30	True		False	

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Unknown without a Cq	N/A	True		False	
Standard without a Cq	N/A	True		False	
Efficiency greater than	110.0	True			
Efficiency less than	90.0	True			
Std Curve R^2 less than	0.980	True			
Replicate group Cq Std Dev greater than	0.20	True	FAM:A7, A8, C7, C8, C11, C12, D5, D6, E3, E4, E7, E8, F1, F2, F3, F4, F5, F6, G7, G8, G9, G10, H7, H8, H9, H10. VIC:A7, A8, C7, C8, C11, C12, D5, D6, E3, E4, E7, E8, F1, F2, F3, F4, G7, G8, G9, G10, H1, H2, H9, H10.	False	