



GT_Plate2_Multiplex_12_12_2024.prcl.pcrd

3/27/2025 9:45 AM

Report Information

User: BioRad/admin
Data File Name: GT_Plate2_Multiplex_12_12_2024.prcl.pcrd
Data File Path: C:\Users\CSMTSS1\Documents\Tortoises\Plates\Multi_SCNAG_Mt
Well Group Name: All Wells
Report Differs from Last Save: No

Run Setup

Run Information

Run Date: 12/12/2024 2:19 PM
Run User: admin
Run Type: User-defined
Plate File: Elgaria_Multiplex.prcl.pltd
ID:
Notes:
Sample Volume: 20
Temperature Control Mode: Calculated
Lid Temperature: 105
Base Serial Number: 795BR20744
Optical Head Serial Number: 795BR20744

Protocol

- 1: 95.0°C for 3:00
- 2: 95.0°C for 0:15
- 3: 62.0°C for 0:30
Plate Read
- 4: GOTO 2, 44 more times

Plate Display


	1	2	3	4	5	6	7	8	9	10	11	12
A	Std-1 0 VIC STD1	Std-1 0 VIC STD1	Unk-1 0 VIC 15.3_D	Unk-1 0 VIC 15.3_D	Unk-9 0 VIC 5.5_A	Unk-9 0 VIC 5.5_A	Unk-17 0 VIC 6.3_D	Unk-17 0 VIC 6.3_D	Unk-25 0 VIC 10.1_C	Unk-25 0 VIC 10.1_C	Unk-33 0 VIC 15.1_C	Unk-33 0 VIC 15.1_C
B	Std-2 0 VIC STD2	Std-2 0 VIC STD2	Unk-2 0 VIC 10.1_A	Unk-2 0 VIC 10.1_A	Unk-10 0 VIC 3.3_B	Unk-10 0 VIC 3.3_B	Unk-18 0 VIC 5.1_A	Unk-18 0 VIC 5.1_A	Unk-26 0 VIC 10.4_B	Unk-26 0 VIC 10.4_B	Unk-34 0 VIC 11.2_D	Unk-34 0 VIC 11.2_D
C	Std-3 0 VIC STD3	Std-3 0 VIC STD3	Unk-3 0 VIC 7.1_A	Unk-3 0 VIC 7.1_A	Unk-11 0 VIC 14.5_C	Unk-11 0 VIC 14.5_C	Unk-19 0 VIC 7.5_D	Unk-19 0 VIC 7.5_D	Unk-27 0 VIC 10.5_D	Unk-27 0 VIC 10.5_D	Unk-35 0 VIC 5.3_A	Unk-35 0 VIC 5.3_A
D	Std-4 0 VIC STD4	Std-4 0 VIC STD4	Unk-4 0 VIC 13.3_D	Unk-4 0 VIC 13.3_D	Unk-12 0 VIC 12.2_D	Unk-12 0 VIC 12.2_D	Unk-20 0 VIC 3.4_C	Unk-20 0 VIC 3.4_C	Unk-28 0 VIC 10.4_C	Unk-28 0 VIC 10.4_C	Unk-36 0 VIC 15.1_B	Unk-36 0 VIC 15.1_B


Plate Display

	1	2	3	4	5	6	7	8	9	10	11	12
E	Std-5 0 VIC STD5	Std-5 0 VIC STD5	Unk-5 0 VIC 8.1_C	Unk-5 0 VIC 8.1_C	Unk-13 0 VIC 3.1_A	Unk-13 0 VIC 3.1_A	Unk-21 0 VIC 13.2_C	Unk-21 0 VIC 13.2_C	Unk-29 0 VIC 11.3_B	Unk-29 0 VIC 11.3_B	Unk-37 0 VIC 15.3_B	Unk-37 0 VIC 15.3_B
F	Std-6 0 VIC STD6	Std-6 0 VIC STD6	Unk-6 0 VIC 5.3_D	Unk-6 0 VIC 5.3_D	Unk-14 0 VIC 6.4_C	Unk-14 0 VIC 6.4_C	Unk-22 0 VIC 15.4_A	Unk-22 0 VIC 15.4_A	Unk-30 0 VIC 17.5_B	Unk-30 0 VIC 17.5_B	Unk-38 0 VIC 5.4_B	Unk-38 0 VIC 5.4_B
G			Unk-7 0 VIC 9.6_D	Unk-7 0 VIC 9.6_D	Unk-15 0 VIC 9.4_C	Unk-15 0 VIC 9.4_C	Unk-23 0 VIC 5.5_B	Unk-23 0 VIC 5.5_B	Unk-31 0 VIC 6.4_A	Unk-31 0 VIC 6.4_A	Unk-39 0 VIC 2.2_B	Unk-39 0 VIC 2.2_B
H	NTC-1 0 VIC NEG	NTC-1 0 VIC NEG	Unk-8 0 VIC 2.2_D	Unk-8 0 VIC 2.2_D	Unk-16 0 VIC 12.5_B	Unk-16 0 VIC 12.5_B	Unk-24 0 VIC 9.5_A	Unk-24 0 VIC 9.5_A	Unk-32 0 VIC 17.2_A	Unk-32 0 VIC 17.2_A	Unk-40 0 VIC 17.2_C	Unk-40 0 VIC 17.2_C

Quantification

Step #: 3
Analysis Mode: Fluorophore
Baseline Setting: Baseline Subtracted Curve Fit
Cq Determination: Single Threshold
Baseline Method:
FAM: Auto Calculated
VIC: Auto Calculated
Threshold Setting:
FAM: 150.00, User Defined
VIC: 150.00, User Defined





Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
A01	FAM	0	Std-01	STD1	16.74	16.75	0.015	3.000E+07	7.477	3.00E+07	0.00E+00
A02	FAM	0	Std-01	STD1	16.76	16.75	0.015	3.000E+07	7.477	3.00E+07	0.00E+00
A03	FAM	0	Unkn-01	15.3_D	22.69	22.53	0.227	5.125E+05	5.710	5.77E+05	9.08E+04
A04	FAM	0	Unkn-01	15.3_D	22.37	22.53	0.227	6.409E+05	5.807	5.77E+05	9.08E+04
A05	FAM	0	Unkn-09	5.5_A	22.76	22.78	0.030	4.890E+05	5.689	4.82E+05	1.01E+04
A06	FAM	0	Unkn-09	5.5_A	22.81	22.78	0.030	4.747E+05	5.676	4.82E+05	1.01E+04
A07	FAM	0	Unkn-17	6.3_D	23.17	23.25	0.118	3.692E+05	5.567	3.49E+05	2.87E+04
A08	FAM	0	Unkn-17	6.3_D	23.33	23.25	0.118	3.286E+05	5.517	3.49E+05	2.87E+04
A09	FAM	0	Unkn-25	10.1_C	23.12	23.09	0.044	3.805E+05	5.580	3.89E+05	1.19E+04
A10	FAM	0	Unkn-25	10.1_C	23.06	23.09	0.044	3.974E+05	5.599	3.89E+05	1.19E+04
A11	FAM	0	Unkn-33	15.1_C	22.20	22.17	0.037	7.246E+05	5.860	7.38E+05	1.88E+04
A12	FAM	0	Unkn-33	15.1_C	22.15	22.17	0.037	7.511E+05	5.876	7.38E+05	1.88E+04
B01	FAM	0	Std-02	STD2	20.25	20.24	0.014	3.000E+06	6.477	3.00E+06	0.00E+00
B02	FAM	0	Std-02	STD2	20.23	20.24	0.014	3.000E+06	6.477	3.00E+06	0.00E+00
B03	FAM	0	Unkn-02	10.1_A	22.00	22.04	0.058	8.296E+05	5.919	8.07E+05	3.25E+04
B04	FAM	0	Unkn-02	10.1_A	22.08	22.04	0.058	7.836E+05	5.894	8.07E+05	3.25E+04
B05	FAM	0	Unkn-10	3.3_B	22.60	22.56	0.058	5.462E+05	5.737	5.62E+05	2.25E+04
B06	FAM	0	Unkn-10	3.3_B	22.52	22.56	0.058	5.780E+05	5.762	5.62E+05	2.25E+04
B07	FAM	0	Unkn-18	5.1_A	22.38	22.35	0.038	6.390E+05	5.806	6.51E+05	1.73E+04
B08	FAM	0	Unkn-18	5.1_A	22.32	22.35	0.038	6.635E+05	5.822	6.51E+05	1.73E+04
B09	FAM	0	Unkn-26	10.4_B	23.15	23.14	0.015	3.745E+05	5.573	3.77E+05	3.89E+03
B10	FAM	0	Unkn-26	10.4_B	23.13	23.14	0.015	3.800E+05	5.580	3.77E+05	3.89E+03
B11	FAM	0	Unkn-34	11.2_D	23.30	23.29	0.004	3.372E+05	5.528	3.38E+05	9.11E+02
B12	FAM	0	Unkn-34	11.2_D	23.29	23.29	0.004	3.385E+05	5.530	3.38E+05	9.11E+02
C01	FAM	0	Std-03	STD3	23.39	23.39	0.002	3.000E+05	5.477	3.00E+05	0.00E+00
C02	FAM	0	Std-03	STD3	23.39	23.39	0.002	3.000E+05	5.477	3.00E+05	0.00E+00
C03	FAM	0	Unkn-03	7.1_A	22.12	22.04	0.112	7.636E+05	5.883	8.08E+05	6.27E+04
C04	FAM	0	Unkn-03	7.1_A	21.96	22.04	0.112	8.523E+05	5.931	8.08E+05	6.27E+04

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
C05	FAM	0	Unkn-11	14.5_C	22.56	22.53	0.032	5.646E+05	5.752	5.74E+05	1.29E+04
C06	FAM	0	Unkn-11	14.5_C	22.51	22.53	0.032	5.828E+05	5.766	5.74E+05	1.29E+04
C07	FAM	0	Unkn-19	7.5_D	23.01	23.04	0.041	4.120E+05	5.615	4.04E+05	1.16E+04
C08	FAM	0	Unkn-19	7.5_D	23.07	23.04	0.041	3.956E+05	5.597	4.04E+05	1.16E+04
C09	FAM	0	Unkn-27	10.5_D	22.92	22.63	0.408	4.375E+05	5.641	5.46E+05	1.53E+05
C10	FAM	0	Unkn-27	10.5_D	22.34	22.63	0.408	6.537E+05	5.815	5.46E+05	1.53E+05
C11	FAM	0	Unkn-35	5.3_A	22.64	22.51	0.185	5.311E+05	5.725	5.84E+05	7.47E+04
C12	FAM	0	Unkn-35	5.3_A	22.38	22.51	0.185	6.368E+05	5.804	5.84E+05	7.47E+04
D01	FAM	0	Std-04	STD4	26.86	26.94	0.115	3.000E+04	4.477	3.00E+04	0.00E+00
D02	FAM	0	Std-04	STD4	27.02	26.94	0.115	3.000E+04	4.477	3.00E+04	0.00E+00
D03	FAM	0	Unkn-04	13.3_D	23.37	23.25	0.168	3.213E+05	5.507	3.50E+05	4.09E+04
D04	FAM	0	Unkn-04	13.3_D	23.13	23.25	0.168	3.791E+05	5.579	3.50E+05	4.09E+04
D05	FAM	0	Unkn-12	12.2_D	22.23	21.69	0.764	7.070E+05	5.849	1.10E+06	5.60E+05
D06	FAM	0	Unkn-12	12.2_D	21.15	21.69	0.764	1.499E+06	6.176	1.10E+06	5.60E+05
D07	FAM	0	Unkn-20	3.4_C	22.69	22.70	0.011	5.133E+05	5.710	5.10E+05	4.07E+03
D08	FAM	0	Unkn-20	3.4_C	22.71	22.70	0.011	5.075E+05	5.705	5.10E+05	4.07E+03
D09	FAM	0	Unkn-28	10.4_C	22.90	22.83	0.091	4.451E+05	5.649	4.66E+05	2.93E+04
D10	FAM	0	Unkn-28	10.4_C	22.77	22.83	0.091	4.866E+05	5.687	4.66E+05	2.93E+04
D11	FAM	0	Unkn-36	15.1_B	21.90	21.88	0.023	8.906E+05	5.950	9.01E+05	1.42E+04
D12	FAM	0	Unkn-36	15.1_B	21.87	21.88	0.023	9.106E+05	5.959	9.01E+05	1.42E+04
E01	FAM	0	Std-05	STD5	30.18	30.11	0.102	3.000E+03	3.477	3.00E+03	0.00E+00
E02	FAM	0	Std-05	STD5	30.04	30.11	0.102	3.000E+03	3.477	3.00E+03	0.00E+00
E03	FAM	0	Unkn-05	8.1_C	23.16	22.91	0.348	3.717E+05	5.570	4.47E+05	1.07E+05
E04	FAM	0	Unkn-05	8.1_C	22.67	22.91	0.348	5.231E+05	5.719	4.47E+05	1.07E+05
E05	FAM	0	Unkn-13	3.1_A	22.62	22.51	0.152	5.409E+05	5.733	5.84E+05	6.16E+04
E06	FAM	0	Unkn-13	3.1_A	22.40	22.51	0.152	6.280E+05	5.798	5.84E+05	6.16E+04
E07	FAM	0	Unkn-21	13.2_C	23.44	23.53	0.129	3.050E+05	5.484	2.87E+05	2.56E+04
E08	FAM	0	Unkn-21	13.2_C	23.62	23.53	0.129	2.687E+05	5.429	2.87E+05	2.56E+04
E09	FAM	0	Unkn-29	11.3_B	23.36	23.36	0.011	3.220E+05	5.508	3.24E+05	2.52E+03
E10	FAM	0	Unkn-29	11.3_B	23.35	23.36	0.011	3.256E+05	5.513	3.24E+05	2.52E+03
E11	FAM	0	Unkn-37	15.3_B	22.37	22.41	0.059	6.428E+05	5.808	6.25E+05	2.54E+04
E12	FAM	0	Unkn-37	15.3_B	22.45	22.41	0.059	6.068E+05	5.783	6.25E+05	2.54E+04
F01	FAM	0	Std-06	STD6	33.43	33.30	0.182	3.000E+02	2.477	3.00E+02	0.00E+00
F02	FAM	0	Std-06	STD6	33.17	33.30	0.182	3.000E+02	2.477	3.00E+02	0.00E+00
F03	FAM	0	Unkn-06	5.3_D	23.11	23.09	0.028	3.833E+05	5.584	3.89E+05	7.51E+03
F04	FAM	0	Unkn-06	5.3_D	23.07	23.09	0.028	3.939E+05	5.595	3.89E+05	7.51E+03
F05	FAM	0	Unkn-14	6.4_C	22.72	22.68	0.054	5.043E+05	5.703	5.18E+05	1.95E+04
F06	FAM	0	Unkn-14	6.4_C	22.64	22.68	0.054	5.319E+05	5.726	5.18E+05	1.95E+04
F07	FAM	0	Unkn-22	15.4_A	22.52	22.53	0.014	5.798E+05	5.763	5.76E+05	5.59E+03
F08	FAM	0	Unkn-22	15.4_A	22.54	22.53	0.014	5.719E+05	5.757	5.76E+05	5.59E+03
F09	FAM	0	Unkn-30	17.5_B	22.61	22.67	0.087	5.447E+05	5.736	5.22E+05	3.16E+04
F10	FAM	0	Unkn-30	17.5_B	22.73	22.67	0.087	5.001E+05	5.699	5.22E+05	3.16E+04
F11	FAM	0	Unkn-38	5.4_B	22.11	22.12	0.007	7.677E+05	5.885	7.65E+05	3.94E+03
F12	FAM	0	Unkn-38	5.4_B	22.12	22.12	0.007	7.621E+05	5.882	7.65E+05	3.94E+03
G03	FAM	0	Unkn-07	9.6_D	22.93	23.08	0.206	4.350E+05	5.639	3.95E+05	5.63E+04
G04	FAM	0	Unkn-07	9.6_D	23.22	23.08	0.206	3.554E+05	5.551	3.95E+05	5.63E+04
G05	FAM	0	Unkn-15	9.4_C	23.28	23.34	0.084	3.416E+05	5.534	3.28E+05	1.91E+04
G06	FAM	0	Unkn-15	9.4_C	23.40	23.34	0.084	3.146E+05	5.498	3.28E+05	1.91E+04
G07	FAM	0	Unkn-23	5.5_B	22.76	22.94	0.251	4.881E+05	5.689	4.35E+05	7.55E+04

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
G08	FAM	0	Unkn-23	5.5_B	23.12	22.94	0.251	3.813E+05	5.581	4.35E+05	7.55E+04
G09	FAM	0	Unkn-31	6.4_A	23.47	23.51	0.064	2.999E+05	5.477	2.91E+05	1.29E+04
G10	FAM	0	Unkn-31	6.4_A	23.56	23.51	0.064	2.816E+05	5.450	2.91E+05	1.29E+04
G11	FAM	0	Unkn-39	2.2_B	21.53	21.61	0.114	1.155E+06	6.063	1.09E+06	8.63E+04
G12	FAM	0	Unkn-39	2.2_B	21.69	21.61	0.114	1.033E+06	6.014	1.09E+06	8.63E+04
H01	FAM	0	NTC-01	NEG	N/A	0.00	0.000	N/A	N/A	0.00E+00	0.00E+00
H02	FAM	0	NTC-01	NEG	N/A	0.00	0.000	N/A	N/A	0.00E+00	0.00E+00
H03	FAM	0	Unkn-08	2.2_D	23.61	23.73	0.179	2.717E+05	5.434	2.50E+05	3.10E+04
H04	FAM	0	Unkn-08	2.2_D	23.86	23.73	0.179	2.279E+05	5.358	2.50E+05	3.10E+04
H05	FAM	0	Unkn-16	12.5_B	22.61	22.59	0.020	5.448E+05	5.736	5.50E+05	7.64E+03
H06	FAM	0	Unkn-16	12.5_B	22.58	22.59	0.020	5.556E+05	5.745	5.50E+05	7.64E+03
H07	FAM	0	Unkn-24	9.5_A	23.42	23.37	0.074	3.090E+05	5.490	3.21E+05	1.64E+04
H08	FAM	0	Unkn-24	9.5_A	23.32	23.37	0.074	3.322E+05	5.521	3.21E+05	1.64E+04
H09	FAM	0	Unkn-32	17.2_A	21.88	22.10	0.316	9.039E+05	5.956	7.83E+05	1.71E+05
H10	FAM	0	Unkn-32	17.2_A	22.33	22.10	0.316	6.628E+05	5.821	7.83E+05	1.71E+05
H11	FAM	0	Unkn-40	17.2_C	21.76	21.72	0.052	9.839E+05	5.993	1.01E+06	3.65E+04
H12	FAM	0	Unkn-40	17.2_C	21.68	21.72	0.052	1.036E+06	6.015	1.01E+06	3.65E+04
A01	VIC		Std-01	STD1	18.97	18.96	0.018	3.000E+07	7.477	3.00E+07	0.00E+00
A02	VIC		Std-01	STD1	18.94	18.96	0.018	3.000E+07	7.477	3.00E+07	0.00E+00
A03	VIC		Unkn-01	15.3_D	29.46	29.29	0.245	2.327E+04	4.367	2.64E+04	4.45E+03
A04	VIC		Unkn-01	15.3_D	29.11	29.29	0.245	2.956E+04	4.471	2.64E+04	4.45E+03
A05	VIC		Unkn-09	5.5_A	28.85	28.84	0.007	3.550E+04	4.550	3.56E+04	1.65E+02
A06	VIC		Unkn-09	5.5_A	28.84	28.84	0.007	3.574E+04	4.553	3.56E+04	1.65E+02
A07	VIC		Unkn-17	6.3_D	29.72	29.83	0.145	1.940E+04	4.288	1.81E+04	1.81E+03
A08	VIC		Unkn-17	6.3_D	29.93	29.83	0.145	1.683E+04	4.226	1.81E+04	1.81E+03
A09	VIC		Unkn-25	10.1_C	29.86	29.76	0.145	1.764E+04	4.247	1.90E+04	1.90E+03
A10	VIC		Unkn-25	10.1_C	29.66	29.76	0.145	2.033E+04	4.308	1.90E+04	1.90E+03
A11	VIC		Unkn-33	15.1_C	28.17	28.19	0.021	5.653E+04	4.752	5.60E+04	7.95E+02
A12	VIC		Unkn-33	15.1_C	28.20	28.19	0.021	5.540E+04	4.744	5.60E+04	7.95E+02
B01	VIC		Std-02	STD2	22.53	22.52	0.016	3.000E+06	6.477	3.00E+06	0.00E+00
B02	VIC		Std-02	STD2	22.51	22.52	0.016	3.000E+06	6.477	3.00E+06	0.00E+00
B03	VIC		Unkn-02	10.1_A	28.39	28.38	0.022	4.869E+04	4.687	4.92E+04	7.55E+02
B04	VIC		Unkn-02	10.1_A	28.36	28.38	0.022	4.976E+04	4.697	4.92E+04	7.55E+02
B05	VIC		Unkn-10	3.3_B	28.94	28.95	0.019	3.345E+04	4.524	3.31E+04	4.26E+02
B06	VIC		Unkn-10	3.3_B	28.96	28.95	0.019	3.284E+04	4.516	3.31E+04	4.26E+02
B07	VIC		Unkn-18	5.1_A	28.78	28.81	0.046	3.728E+04	4.572	3.65E+04	1.16E+03
B08	VIC		Unkn-18	5.1_A	28.84	28.81	0.046	3.565E+04	4.552	3.65E+04	1.16E+03
B09	VIC		Unkn-26	10.4_B	29.29	29.26	0.038	2.622E+04	4.419	2.67E+04	7.06E+02
B10	VIC		Unkn-26	10.4_B	29.23	29.26	0.038	2.722E+04	4.435	2.67E+04	7.06E+02
B11	VIC		Unkn-34	11.2_D	29.25	29.21	0.057	2.698E+04	4.431	2.78E+04	1.09E+03
B12	VIC		Unkn-34	11.2_D	29.17	29.21	0.057	2.853E+04	4.455	2.78E+04	1.09E+03
C01	VIC		Std-03	STD3	25.70	25.72	0.027	3.000E+05	5.477	3.00E+05	0.00E+00
C02	VIC		Std-03	STD3	25.74	25.72	0.027	3.000E+05	5.477	3.00E+05	0.00E+00
C03	VIC		Unkn-03	7.1_A	29.02	28.96	0.082	3.159E+04	4.500	3.29E+04	1.86E+03
C04	VIC		Unkn-03	7.1_A	28.90	28.96	0.082	3.422E+04	4.534	3.29E+04	1.86E+03
C05	VIC		Unkn-11	14.5_C	29.03	29.02	0.021	3.127E+04	4.495	3.16E+04	4.61E+02
C06	VIC		Unkn-11	14.5_C	29.00	29.02	0.021	3.192E+04	4.504	3.16E+04	4.61E+02
C07	VIC		Unkn-19	7.5_D	29.68	29.67	0.011	2.004E+04	4.302	2.02E+04	1.53E+02
C08	VIC		Unkn-19	7.5_D	29.66	29.67	0.011	2.026E+04	4.307	2.02E+04	1.53E+02

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
C09	VIC		Unkn-27	10.5_D	29.76	29.50	0.370	1.894E+04	4.277	2.30E+04	5.82E+03
C10	VIC		Unkn-27	10.5_D	29.24	29.50	0.370	2.716E+04	4.434	2.30E+04	5.82E+03
C11	VIC		Unkn-35	5.3_A	28.71	28.50	0.291	3.918E+04	4.593	4.56E+04	9.11E+03
C12	VIC		Unkn-35	5.3_A	28.29	28.50	0.291	5.207E+04	4.717	4.56E+04	9.11E+03
D01	VIC		Std-04	STD4	29.28	29.26	0.039	3.000E+04	4.477	3.00E+04	0.00E+00
D02	VIC		Std-04	STD4	29.23	29.26	0.039	3.000E+04	4.477	3.00E+04	0.00E+00
D03	VIC		Unkn-04	13.3_D	29.52	29.39	0.174	2.238E+04	4.350	2.45E+04	2.93E+03
D04	VIC		Unkn-04	13.3_D	29.27	29.39	0.174	2.653E+04	4.424	2.45E+04	2.93E+03
D05	VIC		Unkn-12	12.2_D	28.37	27.98	0.545	4.946E+04	4.694	6.68E+04	2.46E+04
D06	VIC		Unkn-12	12.2_D	27.60	27.98	0.545	8.424E+04	4.926	6.68E+04	2.46E+04
D07	VIC		Unkn-20	3.4_C	29.72	29.63	0.127	1.949E+04	4.290	2.08E+04	1.82E+03
D08	VIC		Unkn-20	3.4_C	29.54	29.63	0.127	2.206E+04	4.344	2.08E+04	1.82E+03
D09	VIC		Unkn-28	10.4_C	29.65	29.60	0.075	2.038E+04	4.309	2.12E+04	1.10E+03
D10	VIC		Unkn-28	10.4_C	29.55	29.60	0.075	2.193E+04	4.341	2.12E+04	1.10E+03
D11	VIC		Unkn-36	15.1_B	27.95	27.97	0.019	6.584E+04	4.819	6.52E+04	8.74E+02
D12	VIC		Unkn-36	15.1_B	27.98	27.97	0.019	6.461E+04	4.810	6.52E+04	8.74E+02
E01	VIC		Std-05	STD5	32.57	32.47	0.140	3.000E+03	3.477	3.00E+03	0.00E+00
E02	VIC		Std-05	STD5	32.37	32.47	0.140	3.000E+03	3.477	3.00E+03	0.00E+00
E03	VIC		Unkn-05	8.1_C	30.03	29.96	0.095	1.572E+04	4.197	1.65E+04	1.08E+03
E04	VIC		Unkn-05	8.1_C	29.89	29.96	0.095	1.725E+04	4.237	1.65E+04	1.08E+03
E05	VIC		Unkn-13	3.1_A	28.35	28.27	0.124	4.998E+04	4.699	5.32E+04	4.53E+03
E06	VIC		Unkn-13	3.1_A	28.18	28.27	0.124	5.639E+04	4.751	5.32E+04	4.53E+03
E07	VIC		Unkn-21	13.2_C	29.26	29.30	0.065	2.678E+04	4.428	2.60E+04	1.16E+03
E08	VIC		Unkn-21	13.2_C	29.35	29.30	0.065	2.514E+04	4.400	2.60E+04	1.16E+03
E09	VIC		Unkn-29	11.3_B	30.16	30.14	0.027	1.438E+04	4.158	1.46E+04	2.74E+02
E10	VIC		Unkn-29	11.3_B	30.12	30.14	0.027	1.477E+04	4.169	1.46E+04	2.74E+02
E11	VIC		Unkn-37	15.3_B	28.27	28.33	0.085	5.292E+04	4.724	5.08E+04	2.99E+03
E12	VIC		Unkn-37	15.3_B	28.39	28.33	0.085	4.869E+04	4.687	5.08E+04	2.99E+03
F01	VIC		Std-06	STD6	35.67	35.63	0.052	3.000E+02	2.477	3.00E+02	0.00E+00
F02	VIC		Std-06	STD6	35.59	35.63	0.052	3.000E+02	2.477	3.00E+02	0.00E+00
F03	VIC		Unkn-06	5.3_D	29.58	29.54	0.063	2.136E+04	4.330	2.20E+04	9.56E+02
F04	VIC		Unkn-06	5.3_D	29.50	29.54	0.063	2.271E+04	4.356	2.20E+04	9.56E+02
F05	VIC		Unkn-14	6.4_C	29.21	29.20	0.010	2.767E+04	4.442	2.78E+04	1.93E+02
F06	VIC		Unkn-14	6.4_C	29.20	29.20	0.010	2.794E+04	4.446	2.78E+04	1.93E+02
F07	VIC		Unkn-22	15.4_A	29.12	29.14	0.029	2.946E+04	4.469	2.90E+04	5.81E+02
F08	VIC		Unkn-22	15.4_A	29.16	29.14	0.029	2.864E+04	4.457	2.90E+04	5.81E+02
F09	VIC		Unkn-30	17.5_B	29.20	29.25	0.067	2.783E+04	4.445	2.70E+04	1.24E+03
F10	VIC		Unkn-30	17.5_B	29.30	29.25	0.067	2.608E+04	4.416	2.70E+04	1.24E+03
F11	VIC		Unkn-38	5.4_B	28.00	27.97	0.038	6.373E+04	4.804	6.49E+04	1.69E+03
F12	VIC		Unkn-38	5.4_B	27.95	27.97	0.038	6.613E+04	4.820	6.49E+04	1.69E+03
G03	VIC		Unkn-07	9.6_D	29.30	29.35	0.080	2.605E+04	4.416	2.51E+04	1.39E+03
G04	VIC		Unkn-07	9.6_D	29.41	29.35	0.080	2.408E+04	4.382	2.51E+04	1.39E+03
G05	VIC		Unkn-15	9.4_C	29.40	29.47	0.106	2.429E+04	4.385	2.31E+04	1.68E+03
G06	VIC		Unkn-15	9.4_C	29.55	29.47	0.106	2.191E+04	4.341	2.31E+04	1.68E+03
G07	VIC		Unkn-23	5.5_B	29.79	29.81	0.029	1.860E+04	4.269	1.83E+04	3.69E+02
G08	VIC		Unkn-23	5.5_B	29.83	29.81	0.029	1.808E+04	4.257	1.83E+04	3.69E+02
G09	VIC		Unkn-31	6.4_A	29.27	29.34	0.094	2.650E+04	4.423	2.53E+04	1.64E+03
G10	VIC		Unkn-31	6.4_A	29.40	29.34	0.094	2.418E+04	4.383	2.53E+04	1.64E+03
G11	VIC		Unkn-39	2.2_B	27.85	27.92	0.097	7.058E+04	4.849	6.74E+04	4.51E+03

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
G12	VIC		Unkn-39	2.2_B	27.99	27.92	0.097	6.421E+04	4.808	6.74E+04	4.51E+03
H01	VIC		NTC-01	NEG	N/A	0.00	0.000	N/A	N/A	0.00E+00	0.00E+00
H02	VIC		NTC-01	NEG	N/A	0.00	0.000	N/A	N/A	0.00E+00	0.00E+00
H03	VIC		Unkn-08	2.2_D	30.06	30.21	0.200	1.534E+04	4.186	1.40E+04	1.92E+03
H04	VIC		Unkn-08	2.2_D	30.35	30.21	0.200	1.262E+04	4.101	1.40E+04	1.92E+03
H05	VIC		Unkn-16	12.5_B	29.12	29.18	0.096	2.953E+04	4.470	2.82E+04	1.86E+03
H06	VIC		Unkn-16	12.5_B	29.25	29.18	0.096	2.690E+04	4.430	2.82E+04	1.86E+03
H07	VIC		Unkn-24	9.5_A	30.34	30.24	0.140	1.267E+04	4.103	1.36E+04	1.31E+03
H08	VIC		Unkn-24	9.5_A	30.14	30.24	0.140	1.453E+04	4.162	1.36E+04	1.31E+03
H09	VIC		Unkn-32	17.2_A	28.37	28.54	0.236	4.926E+04	4.692	4.42E+04	7.18E+03
H10	VIC		Unkn-32	17.2_A	28.71	28.54	0.236	3.911E+04	4.592	4.42E+04	7.18E+03
H11	VIC		Unkn-40	17.2_C	27.88	27.81	0.104	6.935E+04	4.841	7.31E+04	5.26E+03
H12	VIC		Unkn-40	17.2_C	27.73	27.81	0.104	7.678E+04	4.885	7.31E+04	5.26E+03

Bar Chart

Normalized expression analysis is not possible, either because no target is assigned as a reference or the selected target(s) is not a

Target Names

Name	Full Name	Reference	Auto Efficiency	Efficiency
0	0	False	Yes	100.4%

Sample Names

Name	Full Name	Control
10.1_A	10.1_A	No
10.1_C	10.1_C	No
10.4_B	10.4_B	No
10.4_C	10.4_C	No
10.5_D	10.5_D	No
11.2_D	11.2_D	No
11.3_B	11.3_B	No
12.2_D	12.2_D	No
12.5_B	12.5_B	No
13.2_C	13.2_C	No
13.3_D	13.3_D	No
14.5_C	14.5_C	No
15.1_B	15.1_B	No
15.1_C	15.1_C	No
15.3_B	15.3_B	No
15.3_D	15.3_D	No
15.4_A	15.4_A	No
17.2_A	17.2_A	No
17.2_C	17.2_C	No
17.5_B	17.5_B	No
2.2_B	2.2_B	No

Sample Names

Name	Full Name	Control
2.2_D	2.2_D	No
3.1_A	3.1_A	No
3.3_B	3.3_B	No
3.4_C	3.4_C	No
5.1_A	5.1_A	No
5.3_A	5.3_A	No
5.3_D	5.3_D	No
5.4_B	5.4_B	No
5.5_A	5.5_A	No
5.5_B	5.5_B	No
6.3_D	6.3_D	No
6.4_A	6.4_A	No
6.4_C	6.4_C	No
7.1_A	7.1_A	No
7.5_D	7.5_D	No
8.1_C	8.1_C	No
9.4_C	9.4_C	No
9.5_A	9.5_A	No
9.6_D	9.6_D	No
STD1	STD1	No
STD2	STD2	No
STD3	STD3	No
STD4	STD4	No
STD5	STD5	No
STD6	STD6	No

Gene Expression - Bar Chart Data

Target	Sample	Control	Expression	Expression SEM	Corrected Expression SEM	Mean Cq	Cq SEM	P-Value
0	10.1_A		N/A	N/A	N/A	22.04	0.04103	N/A
0	10.1_C		N/A	N/A	N/A	23.09	0.03124	N/A
0	10.4_B		N/A	N/A	N/A	23.14	0.01050	N/A
0	10.4_C		N/A	N/A	N/A	22.83	0.06404	N/A
0	10.5_D		N/A	N/A	N/A	22.63	0.28874	N/A
0	11.2_D		N/A	N/A	N/A	23.29	0.00274	N/A
0	11.3_B		N/A	N/A	N/A	23.36	0.00793	N/A
0	12.2_D		N/A	N/A	N/A	21.69	0.54044	N/A
0	12.5_B		N/A	N/A	N/A	22.59	0.01413	N/A
0	13.2_C		N/A	N/A	N/A	23.53	0.09104	N/A
0	13.3_D		N/A	N/A	N/A	23.25	0.11906	N/A
0	14.5_C		N/A	N/A	N/A	22.53	0.02280	N/A
0	15.1_B		N/A	N/A	N/A	21.88	0.01604	N/A
0	15.1_C		N/A	N/A	N/A	22.17	0.02586	N/A
0	15.3_B		N/A	N/A	N/A	22.41	0.04141	N/A
0	15.3_D		N/A	N/A	N/A	22.53	0.16070	N/A
0	15.4_A		N/A	N/A	N/A	22.53	0.00987	N/A
0	17.2_A		N/A	N/A	N/A	22.10	0.22312	N/A

Gene Expression - Bar Chart Data

Target	Sample	Control	Expression	Expression SEM	Corrected Expression SEM	Mean Cq	Cq SEM	P-Value
0	17.2_C		N/A	N/A	N/A	21.72	0.03681	N/A
0	17.5_B		N/A	N/A	N/A	22.67	0.06148	N/A
0	2.2_B		N/A	N/A	N/A	21.61	0.08026	N/A
0	2.2_D		N/A	N/A	N/A	23.73	0.12644	N/A
0	3.1_A		N/A	N/A	N/A	22.51	0.10741	N/A
0	3.3_B		N/A	N/A	N/A	22.56	0.04071	N/A
0	3.4_C		N/A	N/A	N/A	22.70	0.00810	N/A
0	5.1_A		N/A	N/A	N/A	22.35	0.02700	N/A
0	5.3_A		N/A	N/A	N/A	22.51	0.13050	N/A
0	5.3_D		N/A	N/A	N/A	23.09	0.01964	N/A
0	5.4_B		N/A	N/A	N/A	22.12	0.00524	N/A
0	5.5_A		N/A	N/A	N/A	22.78	0.02133	N/A
0	5.5_B		N/A	N/A	N/A	22.94	0.17762	N/A
0	6.3_D		N/A	N/A	N/A	23.25	0.08375	N/A
0	6.4_A		N/A	N/A	N/A	23.51	0.04519	N/A
0	6.4_C		N/A	N/A	N/A	22.68	0.03827	N/A
0	7.1_A		N/A	N/A	N/A	22.04	0.07901	N/A
0	7.5_D		N/A	N/A	N/A	23.04	0.02921	N/A
0	8.1_C		N/A	N/A	N/A	22.91	0.24573	N/A
0	9.4_C		N/A	N/A	N/A	23.34	0.05925	N/A
0	9.5_A		N/A	N/A	N/A	23.37	0.05198	N/A
0	9.6_D		N/A	N/A	N/A	23.08	0.14534	N/A
0	STD1		N/A	N/A	N/A	16.75	0.01031	N/A
0	STD2		N/A	N/A	N/A	20.24	0.01014	N/A
0	STD3		N/A	N/A	N/A	23.39	0.00112	N/A
0	STD4		N/A	N/A	N/A	26.94	0.08134	N/A
0	STD5		N/A	N/A	N/A	30.11	0.07210	N/A
0	STD6		N/A	N/A	N/A	33.30	0.12874	N/A

QC Parameters

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Negative control with a Cq less than	38	True		False	
NTC with a Cq less than	38	True		False	
NRT with a Cq less than	38	True		False	
Positive control with a Cq greater than	30	True		False	

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Unknown without a Cq	N/A	True		False	
Standard without a Cq	N/A	True		False	
Efficiency greater than	110.0	True			
Efficiency less than	90.0	True			
Std Curve R^2 less than	0.980	True			
Replicate group Cq Std Dev greater than	0.20	True	FAM:A3, A4, C9, C10, D5, D6, E3, E4, G3, G4, G7, G8, H9, H10. VIC:A3, A4, C9, C10, C11, C12, D5, D6, H9, H10.	False	