



GT_Plate3_Multiplex_12_16_2024.prcl.pcrd

3/27/2025 9:46 AM

Report Information

User: BioRad/admin
Data File Name: GT_Plate3_Multiplex_12_16_2024.prcl.pcrd
Data File Path: C:\Users\CSMTSS1\Documents\Tortoises\Plates\Multi_SCNAG_Mt
Well Group Name: All Wells
Report Differs from Last Save: No

Run Setup

Run Information

Run Date: 12/16/2024 1:46 PM
Run User: admin
Run Type: User-defined
Plate File: Elgaria_Multiplex.prcl.pltd
ID:
Notes:
Sample Volume: 25
Temperature Control Mode: Calculated
Lid Temperature: 105
Base Serial Number: 795BR20744
Optical Head Serial Number: 795BR20744

Protocol

- 1: 95.0°C for 3:00
- 2: 95.0°C for 0:15
- 3: 62.0°C for 0:30
Plate Read
- 4: GOTO 2, 44 more times

Plate Display


	1	2	3	4	5	6	7	8	9	10	11	12
A	Std-1 0 VIC STD1	Std-1 0 VIC STD1	Unk-1 0 VIC 7.2_D	Unk-1 0 VIC 7.2_D	Unk-9 0 VIC 2.3_D	Unk-9 0 VIC 2.3_D	Unk-17 0 VIC 5.4_C	Unk-17 0 VIC 5.4_C	Unk-25 0 VIC 8.1_B	Unk-25 0 VIC 8.1_B	Unk-33 0 VIC 2.2_A	Unk-33 0 VIC 2.2_A
B	Std-2 0 VIC STD2	Std-2 0 VIC STD2	Unk-2 0 VIC 10.2_A	Unk-2 0 VIC 10.2_A	Unk-10 0 VIC 9.3_C	Unk-10 0 VIC 9.3_C	Unk-18 0 VIC 10.2_D	Unk-18 0 VIC 10.2_D	Unk-26 0 VIC 13.4_C	Unk-26 0 VIC 13.4_C	Unk-34 0 VIC 10.5_B	Unk-34 0 VIC 10.5_B
C	Std-3 0 VIC STD3	Std-3 0 VIC STD3	Unk-3 0 VIC 7.5_A	Unk-3 0 VIC 7.5_A	Unk-11 0 VIC 10.2_B	Unk-11 0 VIC 10.2_B	Unk-19 0 VIC 7.2_B	Unk-19 0 VIC 7.2_B	Unk-27 0 VIC 14.5_B	Unk-27 0 VIC 14.5_B	Unk-35 0 VIC 11.2_C	Unk-35 0 VIC 11.2_C
D	Std-4 0 VIC STD4	Std-4 0 VIC STD4	Unk-4 0 VIC 13.1_A	Unk-4 0 VIC 13.1_A	Unk-12 0 VIC 17.5_D	Unk-12 0 VIC 17.5_D	Unk-20 0 VIC 6.2_A	Unk-20 0 VIC 6.2_A	Unk-28 0 VIC 5.5_D	Unk-28 0 VIC 5.5_D	Unk-36 0 VIC 11.2_A	Unk-36 0 VIC 11.2_A


Plate Display

	1	2	3	4	5	6	7	8	9	10	11	12
E	Std-5 0 VIC STD5	Std-5 0 VIC STD5	Unk-5 0 VIC 17.4_D	Unk-5 0 VIC 17.4_D	Unk-13 0 VIC 14.5_A	Unk-13 0 VIC 14.5_A	Unk-21 0 VIC 15.1_D	Unk-21 0 VIC 15.1_D	Unk-29 0 VIC 5.1_D	Unk-29 0 VIC 5.1_D	Unk-37 0 VIC 17.3_B	Unk-37 0 VIC 17.3_B
F	Std-6 0 VIC STD6	Std-6 0 VIC STD6	Unk-6 0 VIC 7.4_A	Unk-6 0 VIC 7.4_A	*Unk-14 0 VIC 5.6_C	*Unk-14 0 VIC 5.6_C	Unk-22 0 VIC 7.3_B	Unk-22 0 VIC 7.3_B	Unk-30 0 VIC 5.3_C	Unk-30 0 VIC 5.3_C		
G			Unk-7 0 VIC 3.1_D	Unk-7 0 VIC 3.1_D	Unk-15 0 VIC 13.4_A	Unk-15 0 VIC 13.4_A	Unk-23 0 VIC 12.2_B	Unk-23 0 VIC 12.2_B	Unk-31 0 VIC 13.1_C	Unk-31 0 VIC 13.1_C	Unk-39 0 VIC 8.2_C	Unk-39 0 VIC 8.2_C
H	NTC-1 0 VIC NEG	NTC-1 0 VIC NEG	Unk-8 0 VIC 17.3_C	Unk-8 0 VIC 17.3_C	Unk-16 0 VIC 15.2_D	Unk-16 0 VIC 15.2_D	Unk-24 0 VIC 6.2_C	Unk-24 0 VIC 6.2_C	Unk-32 0 VIC 5.2_B	Unk-32 0 VIC 5.2_B	Unk-40 0 VIC 15.2_B	Unk-40 0 VIC 15.2_B

Quantification

Step #: 3
Analysis Mode: Fluorophore
Baseline Setting: Baseline Subtracted Curve Fit
Cq Determination: Single Threshold
Baseline Method:
FAM: Auto Calculated
VIC: Auto Calculated
Threshold Setting:
FAM: 150.00, User Defined
VIC: 150.00, User Defined





Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
A01	FAM	0	Std-01	STD1	16.58	16.57	0.012	3.000E+07	7.477	3.00E+07	0.00E+00
A02	FAM	0	Std-01	STD1	16.56	16.57	0.012	3.000E+07	7.477	3.00E+07	0.00E+00
A03	FAM	0	Unkn-01	7.2_D	21.50	21.47	0.037	9.464E+05	5.976	9.64E+05	2.46E+04
A04	FAM	0	Unkn-01	7.2_D	21.45	21.47	0.037	9.812E+05	5.992	9.64E+05	2.46E+04
A05	FAM	0	Unkn-09	2.3_D	22.01	22.01	0.007	6.661E+05	5.824	6.68E+05	3.10E+03
A06	FAM	0	Unkn-09	2.3_D	22.00	22.01	0.007	6.704E+05	5.826	6.68E+05	3.10E+03
A07	FAM	0	Unkn-17	5.4_C	22.68	22.68	0.011	4.212E+05	5.625	4.19E+05	3.19E+03
A08	FAM	0	Unkn-17	5.4_C	22.69	22.68	0.011	4.167E+05	5.620	4.19E+05	3.19E+03
A09	FAM	0	Unkn-25	8.1_B	20.85	20.84	0.019	1.479E+06	6.170	1.49E+06	1.99E+04
A10	FAM	0	Unkn-25	8.1_B	20.82	20.84	0.019	1.507E+06	6.178	1.49E+06	1.99E+04
A11	FAM	0	Unkn-33	2.2_A	21.98	21.64	0.480	6.789E+05	5.832	8.81E+05	2.86E+05
A12	FAM	0	Unkn-33	2.2_A	21.30	21.64	0.480	1.084E+06	6.035	8.81E+05	2.86E+05
B01	FAM	0	Std-02	STD2	19.75	19.80	0.064	3.000E+06	6.477	3.00E+06	0.00E+00
B02	FAM	0	Std-02	STD2	19.84	19.80	0.064	3.000E+06	6.477	3.00E+06	0.00E+00
B03	FAM	0	Unkn-02	10.2_A	21.00	21.28	0.394	1.335E+06	6.125	1.12E+06	3.01E+05
B04	FAM	0	Unkn-02	10.2_A	21.56	21.28	0.394	9.097E+05	5.959	1.12E+06	3.01E+05
B05	FAM	0	Unkn-10	9.3_C	21.67	21.66	0.023	8.412E+05	5.925	8.51E+05	1.32E+04
B06	FAM	0	Unkn-10	9.3_C	21.64	21.66	0.023	8.599E+05	5.934	8.51E+05	1.32E+04
B07	FAM	0	Unkn-18	10.2_D	21.16	21.15	0.011	1.198E+06	6.079	1.20E+06	8.77E+03
B08	FAM	0	Unkn-18	10.2_D	21.14	21.15	0.011	1.211E+06	6.083	1.20E+06	8.77E+03
B09	FAM	0	Unkn-26	13.4_C	21.68	21.65	0.041	8.342E+05	5.921	8.51E+05	2.42E+04
B10	FAM	0	Unkn-26	13.4_C	21.62	21.65	0.041	8.684E+05	5.939	8.51E+05	2.42E+04
B11	FAM	0	Unkn-34	10.5_B	21.64	21.66	0.021	8.573E+05	5.933	8.49E+05	1.22E+04
B12	FAM	0	Unkn-34	10.5_B	21.67	21.66	0.021	8.401E+05	5.924	8.49E+05	1.22E+04
C01	FAM	0	Std-03	STD3	23.13	23.15	0.031	3.000E+05	5.477	3.00E+05	0.00E+00
C02	FAM	0	Std-03	STD3	23.17	23.15	0.031	3.000E+05	5.477	3.00E+05	0.00E+00
C03	FAM	0	Unkn-03	7.5_A	21.40	21.36	0.062	1.012E+06	6.005	1.04E+06	4.43E+04
C04	FAM	0	Unkn-03	7.5_A	21.32	21.36	0.062	1.075E+06	6.031	1.04E+06	4.43E+04

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
C05	FAM	0	Unkn-11	10.2_B	21.09	21.08	0.010	1.258E+06	6.100	1.26E+06	8.67E+03
C06	FAM	0	Unkn-11	10.2_B	21.07	21.08	0.010	1.270E+06	6.104	1.26E+06	8.67E+03
C07	FAM	0	Unkn-19	7.2_B	21.59	21.56	0.045	8.904E+05	5.950	9.10E+05	2.81E+04
C08	FAM	0	Unkn-19	7.2_B	21.53	21.56	0.045	9.301E+05	5.969	9.10E+05	2.81E+04
C09	FAM	0	Unkn-27	14.5_B	21.40	21.41	0.008	1.013E+06	6.005	1.01E+06	5.67E+03
C10	FAM	0	Unkn-27	14.5_B	21.41	21.41	0.008	1.005E+06	6.002	1.01E+06	5.67E+03
C11	FAM	0	Unkn-35	11.2_C	21.21	21.24	0.055	1.159E+06	6.064	1.13E+06	4.30E+04
C12	FAM	0	Unkn-35	11.2_C	21.28	21.24	0.055	1.099E+06	6.041	1.13E+06	4.30E+04
D01	FAM	0	Std-04	STD4	26.44	26.46	0.024	3.000E+04	4.477	3.00E+04	0.00E+00
D02	FAM	0	Std-04	STD4	26.48	26.46	0.024	3.000E+04	4.477	3.00E+04	0.00E+00
D03	FAM	0	Unkn-04	13.1_A	21.35	21.29	0.079	1.052E+06	6.022	1.09E+06	5.95E+04
D04	FAM	0	Unkn-04	13.1_A	21.24	21.29	0.079	1.136E+06	6.055	1.09E+06	5.95E+04
D05	FAM	0	Unkn-12	17.5_D	21.59	21.57	0.021	8.922E+05	5.950	9.02E+05	1.33E+04
D06	FAM	0	Unkn-12	17.5_D	21.56	21.57	0.021	9.111E+05	5.960	9.02E+05	1.33E+04
D07	FAM	0	Unkn-20	6.2_A	22.03	21.99	0.043	6.590E+05	5.819	6.73E+05	2.00E+04
D08	FAM	0	Unkn-20	6.2_A	21.96	21.99	0.043	6.874E+05	5.837	6.73E+05	2.00E+04
D09	FAM	0	Unkn-28	5.5_D	21.23	21.23	0.006	1.142E+06	6.058	1.14E+06	4.76E+03
D10	FAM	0	Unkn-28	5.5_D	21.24	21.23	0.006	1.135E+06	6.055	1.14E+06	4.76E+03
D11	FAM	0	Unkn-36	11.2_A	21.43	21.48	0.069	9.956E+05	5.998	9.63E+05	4.60E+04
D12	FAM	0	Unkn-36	11.2_A	21.52	21.48	0.069	9.306E+05	5.969	9.63E+05	4.60E+04
E01	FAM	0	Std-05	STD5	29.54	29.69	0.214	3.000E+03	3.477	3.00E+03	0.00E+00
E02	FAM	0	Std-05	STD5	29.84	29.69	0.214	3.000E+03	3.477	3.00E+03	0.00E+00
E03	FAM	0	Unkn-05	17.4_D	22.03	21.94	0.126	6.582E+05	5.818	7.01E+05	6.05E+04
E04	FAM	0	Unkn-05	17.4_D	21.85	21.94	0.126	7.438E+05	5.871	7.01E+05	6.05E+04
E05	FAM	0	Unkn-13	14.5_A	22.21	22.19	0.026	5.801E+05	5.763	5.87E+05	1.03E+04
E06	FAM	0	Unkn-13	14.5_A	22.17	22.19	0.026	5.947E+05	5.774	5.87E+05	1.03E+04
E07	FAM	0	Unkn-21	15.1_D	21.48	21.44	0.058	9.594E+05	5.982	9.87E+05	3.94E+04
E08	FAM	0	Unkn-21	15.1_D	21.40	21.44	0.058	1.015E+06	6.007	9.87E+05	3.94E+04
E09	FAM	0	Unkn-29	5.1_D	21.35	21.36	0.016	1.051E+06	6.022	1.04E+06	1.15E+04
E10	FAM	0	Unkn-29	5.1_D	21.37	21.36	0.016	1.035E+06	6.015	1.04E+06	1.15E+04
E11	FAM	0	Unkn-37	17.3_B	21.60	21.61	0.017	8.816E+05	5.945	8.74E+05	1.01E+04
E12	FAM	0	Unkn-37	17.3_B	21.63	21.61	0.017	8.673E+05	5.938	8.74E+05	1.01E+04
F01	FAM	0	Std-06	STD6	33.60	33.38	0.323	3.000E+02	2.477	3.00E+02	0.00E+00
F02	FAM	0	Std-06	STD6	33.15	33.38	0.323	3.000E+02	2.477	3.00E+02	0.00E+00
F03	FAM	0	Unkn-06	7.4_A	21.79	21.83	0.044	7.728E+05	5.888	7.57E+05	2.29E+04
F04	FAM	0	Unkn-06	7.4_A	21.86	21.83	0.044	7.404E+05	5.869	7.57E+05	2.29E+04
F07	FAM	0	Unkn-22	7.3_B	21.88	21.91	0.055	7.308E+05	5.864	7.12E+05	2.68E+04
F08	FAM	0	Unkn-22	7.3_B	21.95	21.91	0.055	6.929E+05	5.841	7.12E+05	2.68E+04
F09	FAM	0	Unkn-30	5.3_C	23.14	23.17	0.052	3.066E+05	5.487	2.99E+05	1.06E+04
F10	FAM	0	Unkn-30	5.3_C	23.21	23.17	0.052	2.915E+05	5.465	2.99E+05	1.06E+04
G03	FAM	0	Unkn-07	3.1_D	23.41	23.36	0.073	2.535E+05	5.404	2.63E+05	1.32E+04
G04	FAM	0	Unkn-07	3.1_D	23.31	23.36	0.073	2.722E+05	5.435	2.63E+05	1.32E+04
G05	FAM	0	Unkn-15	13.4_A	21.60	21.54	0.093	8.810E+05	5.945	9.23E+05	5.89E+04
G06	FAM	0	Unkn-15	13.4_A	21.47	21.54	0.093	9.644E+05	5.984	9.23E+05	5.89E+04
G07	FAM	0	Unkn-23	12.2_B	21.75	21.69	0.079	7.992E+05	5.903	8.31E+05	4.54E+04
G08	FAM	0	Unkn-23	12.2_B	21.63	21.69	0.079	8.633E+05	5.936	8.31E+05	4.54E+04
G09	FAM	0	Unkn-31	13.1_C	21.22	21.23	0.014	1.149E+06	6.060	1.14E+06	1.07E+04
G10	FAM	0	Unkn-31	13.1_C	21.24	21.23	0.014	1.134E+06	6.055	1.14E+06	1.07E+04
G11	FAM	0	Unkn-39	8.2_C	21.23	21.28	0.063	1.136E+06	6.055	1.10E+06	4.81E+04

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
G12	FAM	0	Unkn-39	8.2_C	21.32	21.28	0.063	1.068E+06	6.028	1.10E+06	4.81E+04
H01	FAM	0	NTC-01	NEG	N/A	0.00	0.000	N/A	N/A	0.00E+00	0.00E+00
H02	FAM	0	NTC-01	NEG	N/A	0.00	0.000	N/A	N/A	0.00E+00	0.00E+00
H03	FAM	0	Unkn-08	17.3_C	22.27	22.23	0.048	5.581E+05	5.747	5.71E+05	1.87E+04
H04	FAM	0	Unkn-08	17.3_C	22.20	22.23	0.048	5.846E+05	5.767	5.71E+05	1.87E+04
H05	FAM	0	Unkn-16	15.2_D	21.88	21.92	0.045	7.268E+05	5.861	7.11E+05	2.21E+04
H06	FAM	0	Unkn-16	15.2_D	21.95	21.92	0.045	6.956E+05	5.842	7.11E+05	2.21E+04
H07	FAM	0	Unkn-24	6.2_C	22.41	22.41	0.004	5.054E+05	5.704	5.07E+05	1.55E+03
H08	FAM	0	Unkn-24	6.2_C	22.40	22.41	0.004	5.076E+05	5.706	5.07E+05	1.55E+03
H09	FAM	0	Unkn-32	5.2_B	21.69	21.70	0.004	8.277E+05	5.918	8.26E+05	2.30E+03
H10	FAM	0	Unkn-32	5.2_B	21.70	21.70	0.004	8.244E+05	5.916	8.26E+05	2.30E+03
H11	FAM	0	Unkn-40	15.2_B	21.43	21.43	0.003	9.946E+05	5.998	9.93E+05	2.02E+03
H12	FAM	0	Unkn-40	15.2_B	21.43	21.43	0.003	9.918E+05	5.996	9.93E+05	2.02E+03
A01	VIC		Std-01	STD1	17.94	17.96	0.033	3.000E+07	7.477	3.00E+07	0.00E+00
A02	VIC		Std-01	STD1	17.98	17.96	0.033	3.000E+07	7.477	3.00E+07	0.00E+00
A03	VIC		Unkn-01	7.2_D	27.25	27.22	0.050	4.715E+04	4.673	4.83E+04	1.65E+03
A04	VIC		Unkn-01	7.2_D	27.18	27.22	0.050	4.948E+04	4.694	4.83E+04	1.65E+03
A05	VIC		Unkn-09	2.3_D	27.13	27.15	0.025	5.119E+04	4.709	5.06E+04	8.67E+02
A06	VIC		Unkn-09	2.3_D	27.17	27.15	0.025	4.996E+04	4.699	5.06E+04	8.67E+02
A07	VIC		Unkn-17	5.4_C	28.31	28.33	0.029	2.276E+04	4.357	2.24E+04	4.44E+02
A08	VIC		Unkn-17	5.4_C	28.35	28.33	0.029	2.213E+04	4.345	2.24E+04	4.44E+02
A09	VIC		Unkn-25	8.1_B	26.64	26.60	0.070	7.167E+04	4.855	7.42E+04	3.60E+03
A10	VIC		Unkn-25	8.1_B	26.55	26.60	0.070	7.675E+04	4.885	7.42E+04	3.60E+03
A11	VIC		Unkn-33	2.2_A	27.77	27.52	0.349	3.305E+04	4.519	3.98E+04	9.49E+03
A12	VIC		Unkn-33	2.2_A	27.27	27.52	0.349	4.647E+04	4.667	3.98E+04	9.49E+03
B01	VIC		Std-02	STD2	21.18	21.17	0.012	3.000E+06	6.477	3.00E+06	0.00E+00
B02	VIC		Std-02	STD2	21.17	21.17	0.012	3.000E+06	6.477	3.00E+06	0.00E+00
B03	VIC		Unkn-02	10.2_A	27.17	27.30	0.187	4.985E+04	4.698	4.57E+04	5.89E+03
B04	VIC		Unkn-02	10.2_A	27.44	27.30	0.187	4.152E+04	4.618	4.57E+04	5.89E+03
B05	VIC		Unkn-10	9.3_C	27.78	27.73	0.078	3.265E+04	4.514	3.39E+04	1.82E+03
B06	VIC		Unkn-10	9.3_C	27.67	27.73	0.078	3.522E+04	4.547	3.39E+04	1.82E+03
B07	VIC		Unkn-18	10.2_D	26.33	26.33	0.007	8.920E+04	4.950	8.89E+04	4.15E+02
B08	VIC		Unkn-18	10.2_D	26.34	26.33	0.007	8.862E+04	4.948	8.89E+04	4.15E+02
B09	VIC		Unkn-26	13.4_C	26.75	26.74	0.020	6.646E+04	4.823	6.71E+04	9.10E+02
B10	VIC		Unkn-26	13.4_C	26.73	26.74	0.020	6.775E+04	4.831	6.71E+04	9.10E+02
B11	VIC		Unkn-34	10.5_B	27.23	27.39	0.219	4.769E+04	4.678	4.31E+04	6.49E+03
B12	VIC		Unkn-34	10.5_B	27.54	27.39	0.219	3.852E+04	4.586	4.31E+04	6.49E+03
C01	VIC		Std-03	STD3	24.47	24.51	0.052	3.000E+05	5.477	3.00E+05	0.00E+00
C02	VIC		Std-03	STD3	24.54	24.51	0.052	3.000E+05	5.477	3.00E+05	0.00E+00
C03	VIC		Unkn-03	7.5_A	27.21	27.16	0.069	4.867E+04	4.687	5.04E+04	2.40E+03
C04	VIC		Unkn-03	7.5_A	27.11	27.16	0.069	5.206E+04	4.716	5.04E+04	2.40E+03
C05	VIC		Unkn-11	10.2_B	27.10	27.08	0.034	5.218E+04	4.717	5.31E+04	1.24E+03
C06	VIC		Unkn-11	10.2_B	27.06	27.08	0.034	5.393E+04	4.732	5.31E+04	1.24E+03
C07	VIC		Unkn-19	7.2_B	27.22	27.22	0.010	4.806E+04	4.682	4.83E+04	3.36E+02
C08	VIC		Unkn-19	7.2_B	27.21	27.22	0.010	4.853E+04	4.686	4.83E+04	3.36E+02
C09	VIC		Unkn-27	14.5_B	26.92	26.92	0.002	5.946E+04	4.774	5.94E+04	9.80E+01
C10	VIC		Unkn-27	14.5_B	26.92	26.92	0.002	5.932E+04	4.773	5.94E+04	9.80E+01
C11	VIC		Unkn-35	11.2_C	26.90	26.86	0.056	6.007E+04	4.779	6.18E+04	2.38E+03
C12	VIC		Unkn-35	11.2_C	26.82	26.86	0.056	6.344E+04	4.802	6.18E+04	2.38E+03

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
D01	VIC		Std-04	STD4	27.99	28.03	0.046	3.000E+04	4.477	3.00E+04	0.00E+00
D02	VIC		Std-04	STD4	28.06	28.03	0.046	3.000E+04	4.477	3.00E+04	0.00E+00
D03	VIC		Unkn-04	13.1_A	27.32	27.23	0.127	4.510E+04	4.654	4.81E+04	4.20E+03
D04	VIC		Unkn-04	13.1_A	27.14	27.23	0.127	5.105E+04	4.708	4.81E+04	4.20E+03
D05	VIC		Unkn-12	17.5_D	27.21	27.16	0.072	4.843E+04	4.685	5.02E+04	2.50E+03
D06	VIC		Unkn-12	17.5_D	27.11	27.16	0.072	5.197E+04	4.716	5.02E+04	2.50E+03
D07	VIC		Unkn-20	6.2_A	27.02	27.02	0.005	5.513E+04	4.741	5.53E+04	2.09E+02
D08	VIC		Unkn-20	6.2_A	27.02	27.02	0.005	5.543E+04	4.744	5.53E+04	2.09E+02
D09	VIC		Unkn-28	5.5_D	27.07	27.10	0.046	5.357E+04	4.729	5.24E+04	1.68E+03
D10	VIC		Unkn-28	5.5_D	27.13	27.10	0.046	5.119E+04	4.709	5.24E+04	1.68E+03
D11	VIC		Unkn-36	11.2_A	27.11	27.18	0.101	5.206E+04	4.717	4.96E+04	3.44E+03
D12	VIC		Unkn-36	11.2_A	27.25	27.18	0.101	4.719E+04	4.674	4.96E+04	3.44E+03
E01	VIC		Std-05	STD5	31.02	31.13	0.161	3.000E+03	3.477	3.00E+03	0.00E+00
E02	VIC		Std-05	STD5	31.24	31.13	0.161	3.000E+03	3.477	3.00E+03	0.00E+00
E03	VIC		Unkn-05	17.4_D	27.22	27.16	0.088	4.813E+04	4.682	5.03E+04	3.05E+03
E04	VIC		Unkn-05	17.4_D	27.10	27.16	0.088	5.244E+04	4.720	5.03E+04	3.05E+03
E05	VIC		Unkn-13	14.5_A	27.46	27.48	0.032	4.092E+04	4.612	4.03E+04	8.89E+02
E06	VIC		Unkn-13	14.5_A	27.50	27.48	0.032	3.966E+04	4.598	4.03E+04	8.89E+02
E07	VIC		Unkn-21	15.1_D	27.14	27.12	0.020	5.102E+04	4.708	5.15E+04	7.07E+02
E08	VIC		Unkn-21	15.1_D	27.11	27.12	0.020	5.202E+04	4.716	5.15E+04	7.07E+02
E09	VIC		Unkn-29	5.1_D	26.61	26.60	0.013	7.333E+04	4.865	7.38E+04	6.60E+02
E10	VIC		Unkn-29	5.1_D	26.59	26.60	0.013	7.426E+04	4.871	7.38E+04	6.60E+02
E11	VIC		Unkn-37	17.3_B	27.47	27.52	0.059	4.041E+04	4.607	3.93E+04	1.59E+03
E12	VIC		Unkn-37	17.3_B	27.56	27.52	0.059	3.817E+04	4.582	3.93E+04	1.59E+03
F01	VIC		Std-06	STD6	34.77	34.63	0.186	3.000E+02	2.477	3.00E+02	0.00E+00
F02	VIC		Std-06	STD6	34.50	34.63	0.186	3.000E+02	2.477	3.00E+02	0.00E+00
F03	VIC		Unkn-06	7.4_A	27.34	27.31	0.045	4.426E+04	4.646	4.53E+04	1.40E+03
F04	VIC		Unkn-06	7.4_A	27.28	27.31	0.045	4.624E+04	4.665	4.53E+04	1.40E+03
F07	VIC		Unkn-22	7.3_B	27.30	27.41	0.157	4.574E+04	4.660	4.25E+04	4.60E+03
F08	VIC		Unkn-22	7.3_B	27.52	27.41	0.157	3.924E+04	4.594	4.25E+04	4.60E+03
F09	VIC		Unkn-30	5.3_C	28.15	28.15	0.012	2.542E+04	4.405	2.53E+04	2.06E+02
F10	VIC		Unkn-30	5.3_C	28.16	28.15	0.012	2.513E+04	4.400	2.53E+04	2.06E+02
G03	VIC		Unkn-07	3.1_D	28.40	28.31	0.141	2.127E+04	4.328	2.28E+04	2.21E+03
G04	VIC		Unkn-07	3.1_D	28.21	28.31	0.141	2.440E+04	4.387	2.28E+04	2.21E+03
G05	VIC		Unkn-15	13.4_A	27.32	27.27	0.065	4.495E+04	4.653	4.64E+04	2.09E+03
G06	VIC		Unkn-15	13.4_A	27.23	27.27	0.065	4.791E+04	4.680	4.64E+04	2.09E+03
G07	VIC		Unkn-23	12.2_B	27.32	27.25	0.096	4.500E+04	4.653	4.72E+04	3.12E+03
G08	VIC		Unkn-23	12.2_B	27.18	27.25	0.096	4.942E+04	4.694	4.72E+04	3.12E+03
G09	VIC		Unkn-31	13.1_C	27.05	27.00	0.069	5.417E+04	4.734	5.61E+04	2.67E+03
G10	VIC		Unkn-31	13.1_C	26.95	27.00	0.069	5.795E+04	4.763	5.61E+04	2.67E+03
G11	VIC		Unkn-39	8.2_C	26.94	26.98	0.054	5.856E+04	4.768	5.70E+04	2.13E+03
G12	VIC		Unkn-39	8.2_C	27.01	26.98	0.054	5.554E+04	4.745	5.70E+04	2.13E+03
H01	VIC		NTC-01	NEG	40.41	39.68	1.025	N/A	N/A	N/A	N/A
H02	VIC		NTC-01	NEG	38.96	39.68	1.025	N/A	N/A	N/A	N/A
H03	VIC		Unkn-08	17.3_C	27.44	27.41	0.038	4.136E+04	4.617	4.21E+04	1.11E+03
H04	VIC		Unkn-08	17.3_C	27.39	27.41	0.038	4.293E+04	4.633	4.21E+04	1.11E+03
H05	VIC		Unkn-16	15.2_D	27.16	27.22	0.083	5.006E+04	4.699	4.81E+04	2.74E+03
H06	VIC		Unkn-16	15.2_D	27.28	27.22	0.083	4.618E+04	4.664	4.81E+04	2.74E+03
H07	VIC		Unkn-24	6.2_C	27.86	27.86	0.001	3.098E+04	4.491	3.10E+04	3.00E+01

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
H08	VIC		Unkn-24	6.2_C	27.86	27.86	0.001	3.094E+04	4.490	3.10E+04	3.00E+01
H09	VIC		Unkn-32	5.2_B	27.20	27.24	0.050	4.870E+04	4.687	4.75E+04	1.66E+03
H10	VIC		Unkn-32	5.2_B	27.28	27.24	0.050	4.635E+04	4.666	4.75E+04	1.66E+03
H11	VIC		Unkn-40	15.2_B	27.12	27.12	0.007	5.162E+04	4.713	5.18E+04	2.60E+02
H12	VIC		Unkn-40	15.2_B	27.11	27.12	0.007	5.199E+04	4.716	5.18E+04	2.60E+02

Bar Chart

Normalized expression analysis is not possible, either because no target is assigned as a reference or the selected target(s) is not a

Target Names

Name	Full Name	Reference	Auto Efficiency	Efficiency
0	0	False	Yes	99.1%

Sample Names

Name	Full Name	Control
10.2_A	10.2_A	No
10.2_B	10.2_B	No
10.2_D	10.2_D	No
10.5_B	10.5_B	No
11.2_A	11.2_A	No
11.2_C	11.2_C	No
12.2_B	12.2_B	No
13.1_A	13.1_A	No
13.1_C	13.1_C	No
13.4_A	13.4_A	No
13.4_C	13.4_C	No
14.5_A	14.5_A	No
14.5_B	14.5_B	No
15.1_D	15.1_D	No
15.2_B	15.2_B	No
15.2_D	15.2_D	No
17.3_B	17.3_B	No
17.3_C	17.3_C	No
17.4_D	17.4_D	No
17.5_D	17.5_D	No
2.2_A	2.2_A	No
2.3_D	2.3_D	No
3.1_D	3.1_D	No
5.1_D	5.1_D	No
5.2_B	5.2_B	No
5.3_C	5.3_C	No
5.4_C	5.4_C	No
5.5_D	5.5_D	No
6.2_A	6.2_A	No

Sample Names

Name	Full Name	Control
6.2_C	6.2_C	No
7.2_B	7.2_B	No
7.2_D	7.2_D	No
7.3_B	7.3_B	No
7.4_A	7.4_A	No
7.5_A	7.5_A	No
8.1_B	8.1_B	No
8.2_C	8.2_C	No
9.3_C	9.3_C	No
STD1	STD1	No
STD2	STD2	No
STD3	STD3	No
STD4	STD4	No
STD5	STD5	No
STD6	STD6	No

Gene Expression - Bar Chart Data

Target	Sample	Control	Expression	Expression SEM	Corrected Expression SEM	Mean Cq	Cq SEM	P-Value
0	10.2_A		N/A	N/A	N/A	21.28	0.27831	N/A
0	10.2_B		N/A	N/A	N/A	21.08	0.00704	N/A
0	10.2_D		N/A	N/A	N/A	21.15	0.00748	N/A
0	10.5_B		N/A	N/A	N/A	21.66	0.01472	N/A
0	11.2_A		N/A	N/A	N/A	21.48	0.04901	N/A
0	11.2_C		N/A	N/A	N/A	21.24	0.03910	N/A
0	12.2_B		N/A	N/A	N/A	21.69	0.05607	N/A
0	13.1_A		N/A	N/A	N/A	21.29	0.05593	N/A
0	13.1_C		N/A	N/A	N/A	21.23	0.00961	N/A
0	13.4_A		N/A	N/A	N/A	21.54	0.06564	N/A
0	13.4_C		N/A	N/A	N/A	21.65	0.02925	N/A
0	14.5_A		N/A	N/A	N/A	22.19	0.01804	N/A
0	14.5_B		N/A	N/A	N/A	21.41	0.00578	N/A
0	15.1_D		N/A	N/A	N/A	21.44	0.04101	N/A
0	15.2_B		N/A	N/A	N/A	21.43	0.00209	N/A
0	15.2_D		N/A	N/A	N/A	21.92	0.03194	N/A
0	17.3_B		N/A	N/A	N/A	21.61	0.01187	N/A
0	17.3_C		N/A	N/A	N/A	22.23	0.03366	N/A
0	17.4_D		N/A	N/A	N/A	21.94	0.08878	N/A
0	17.5_D		N/A	N/A	N/A	21.57	0.01516	N/A
0	2.2_A		N/A	N/A	N/A	21.64	0.33966	N/A
0	2.3_D		N/A	N/A	N/A	22.01	0.00476	N/A
0	3.1_D		N/A	N/A	N/A	23.36	0.05159	N/A
0	5.1_D		N/A	N/A	N/A	21.36	0.01133	N/A
0	5.2_B		N/A	N/A	N/A	21.70	0.00286	N/A
0	5.3_C		N/A	N/A	N/A	23.17	0.03652	N/A
0	5.4_C		N/A	N/A	N/A	22.68	0.00783	N/A
0	5.5_D		N/A	N/A	N/A	21.23	0.00429	N/A

Gene Expression - Bar Chart Data

Target	Sample	Control	Expression	Expression SEM	Corrected Expression SEM	Mean Cq	Cq SEM	P-Value
0	6.2_A		N/A	N/A	N/A	21.99	0.03054	N/A
0	6.2_C		N/A	N/A	N/A	22.41	0.00314	N/A
0	7.2_B		N/A	N/A	N/A	21.56	0.03172	N/A
0	7.2_D		N/A	N/A	N/A	21.47	0.02619	N/A
0	7.3_B		N/A	N/A	N/A	21.91	0.03873	N/A
0	7.4_A		N/A	N/A	N/A	21.83	0.03109	N/A
0	7.5_A		N/A	N/A	N/A	21.36	0.04365	N/A
0	8.1_B		N/A	N/A	N/A	20.84	0.01372	N/A
0	8.2_C		N/A	N/A	N/A	21.28	0.04485	N/A
0	9.3_C		N/A	N/A	N/A	21.66	0.01596	N/A
0	STD1		N/A	N/A	N/A	16.57	0.00866	N/A
0	STD2		N/A	N/A	N/A	19.80	0.04491	N/A
0	STD3		N/A	N/A	N/A	23.15	0.02171	N/A
0	STD4		N/A	N/A	N/A	26.46	0.01727	N/A
0	STD5		N/A	N/A	N/A	29.69	0.15148	N/A
0	STD6		N/A	N/A	N/A	33.38	0.22807	N/A

QC Parameters

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Negative control with a Cq less than	38	True		False	
NTC with a Cq less than	38	True		False	
NRT with a Cq less than	38	True		False	
Positive control with a Cq greater than	30	True		False	
Unknown without a Cq	N/A	True		False	
Standard without a Cq	N/A	True		False	
Efficiency greater than	110.0	True			
Efficiency less than	90.0	True			
Std Curve R^2 less than	0.980	True			

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Replicate group Cq Std Dev greater than	0.20	True	FAM:A11, A12, B3, B4, E1, E2, F1, F2. VIC:A11, A12, B11, B12, H1, H2.	False	