



GT_Plate4_Multiplex_12_17_2024.prcl.pcrd

3/27/2025 9:46 AM

Report Information

User: BioRad/admin
Data File Name: GT_Plate4_Multiplex_12_17_2024.prcl.pcrd
Data File Path: C:\Users\CSMTSS1\Documents\Tortoises\Plates\Multi_SCNAG_Mt
Well Group Name: All Wells
Report Differs from Last Save: No

Run Setup

Run Information

Run Date: 12/17/2024 1:05 PM
Run User: admin
Run Type: User-defined
Plate File: Elgaria_Multiplex.prcl.pltd
ID:
Notes:
Sample Volume: 20
Temperature Control Mode: Calculated
Lid Temperature: 105
Base Serial Number: 795BR20744
Optical Head Serial Number: 795BR20744

Protocol

- 1: 95.0°C for 3:00
- 2: 95.0°C for 0:15
- 3: 62.0°C for 0:30
Plate Read
- 4: GOTO 2, 44 more times

Plate Display


	1	2	3	4	5	6	7	8	9	10	11	12
A	Std-1 0 VIC STD1	Std-1 0 VIC STD1	Unk-1 0 VIC 7.1_B	Unk-1 0 VIC 7.1_B	Unk-9 0 VIC 15.5_B	Unk-9 0 VIC 15.5_B	Unk-17 0 VIC 17.2_D	Unk-17 0 VIC 17.2_D	Unk-25 0 VIC 8.2_A	Unk-25 0 VIC 8.2_A	Unk-33 0 VIC 15.3_C	Unk-33 0 VIC 15.3_C
B	Std-2 0 VIC STD2	Std-2 0 VIC STD2	Unk-2 0 VIC 3.2_C	Unk-2 0 VIC 3.2_C	Unk-10 0 VIC 8.2_B	Unk-10 0 VIC 8.2_B	Unk-18 0 VIC 6.5_D	Unk-18 0 VIC 6.5_D	Unk-26 0 VIC 3.2_A	Unk-26 0 VIC 3.2_A	Unk-34 0 VIC 17.4_C	Unk-34 0 VIC 17.4_C
C	Std-3 0 VIC STD3	Std-3 0 VIC STD3	*Unk-3 0 VIC 10.3_A	*Unk-3 0 VIC 10.3_A	Unk-11 0 VIC 11.1_B	Unk-11 0 VIC 11.1_B	Unk-19 0 VIC 9.6_C	Unk-19 0 VIC 9.6_C	Unk-27 0 VIC 6.4_B	Unk-27 0 VIC 6.4_B	Unk-35 0 VIC 12.6_D	Unk-35 0 VIC 12.6_D
D	Std-4 0 VIC STD4	Std-4 0 VIC STD4	*Unk-4 0 VIC 15.4_B	*Unk-4 0 VIC 15.4_B	Unk-12 0 VIC 10.3_B	Unk-12 0 VIC 10.3_B	Unk-20 0 VIC 2.4_C	Unk-20 0 VIC 2.4_C	Unk-28 0 VIC 10.1_D	Unk-28 0 VIC 10.1_D	Unk-36 0 VIC 12.2_C	Unk-36 0 VIC 12.2_C


Plate Display

	1	2	3	4	5	6	7	8	9	10	11	12
E	Std-5 0 VIC STD5	Std-5 0 VIC STD5	Unk-5 0 VIC 2.3_A	Unk-5 0 VIC 2.3_A	Unk-13 0 VIC 9.4_A	Unk-13 0 VIC 9.4_A	Unk-21 0 VIC 6.2_B	Unk-21 0 VIC 6.2_B	Unk-29 0 VIC 10.3_D	Unk-29 0 VIC 10.3_D	Unk-37 0 VIC 3.3_D	Unk-37 0 VIC 3.3_D
F	Std-6 0 VIC STD6	Std-6 0 VIC STD6	Unk-6 0 VIC 11.4_C	Unk-6 0 VIC 11.4_C	Unk-14 0 VIC 2.4_A	Unk-14 0 VIC 2.4_A	Unk-22 0 VIC 3.4_A	Unk-22 0 VIC 3.4_A	Unk-30 0 VIC 12.4_D	Unk-30 0 VIC 12.4_D	Unk-38 0 VIC 7.1_C	Unk-38 0 VIC 7.1_C
G			Unk-7 0 VIC 9.2_A	Unk-7 0 VIC 9.2_A	Unk-15 0 VIC 11.3_D	Unk-15 0 VIC 11.3_D	Unk-23 0 VIC 12.5_C	Unk-23 0 VIC 12.5_C	Unk-31 0 VIC 12.6_B	Unk-31 0 VIC 12.6_B	Unk-39 0 VIC 6.4_D	Unk-39 0 VIC 6.4_D
H	NTC-1 0 VIC NEG	NTC-1 0 VIC NEG	Unk-8 0 VIC 13.4_B	Unk-8 0 VIC 13.4_B	Unk-16 0 VIC 13.1_D	Unk-16 0 VIC 13.1_D	Unk-24 0 VIC 11.1_A	Unk-24 0 VIC 11.1_A	Unk-32 0 VIC 12.6_A	Unk-32 0 VIC 12.6_A	Unk-40 0 VIC 7.2_C	Unk-40 0 VIC 7.2_C

Quantification

Step #: 3
Analysis Mode: Fluorophore
Baseline Setting: Baseline Subtracted Curve Fit
Cq Determination: Single Threshold
Baseline Method:
FAM: Auto Calculated
VIC: Auto Calculated
Threshold Setting:
FAM: 150.00, User Defined
VIC: 150.00, User Defined





Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
A01	FAM	0	Std-01	STD1	16.72	16.77	0.071	3.000E+07	7.477	3.00E+07	0.00E+00
A02	FAM	0	Std-01	STD1	16.82	16.77	0.071	3.000E+07	7.477	3.00E+07	0.00E+00
A03	FAM	0	Unkn-01	7.1_B	21.56	21.47	0.129	1.130E+06	6.053	1.20E+06	1.05E+05
A04	FAM	0	Unkn-01	7.1_B	21.38	21.47	0.129	1.279E+06	6.107	1.20E+06	1.05E+05
A05	FAM	0	Unkn-09	15.5_B	21.99	21.99	0.003	8.420E+05	5.925	8.43E+05	1.81E+03
A06	FAM	0	Unkn-09	15.5_B	21.99	21.99	0.003	8.445E+05	5.927	8.43E+05	1.81E+03
A07	FAM	0	Unkn-17	17.2_D	21.54	21.58	0.053	1.142E+06	6.058	1.11E+06	4.04E+04
A08	FAM	0	Unkn-17	17.2_D	21.62	21.58	0.053	1.085E+06	6.035	1.11E+06	4.04E+04
A09	FAM	0	Unkn-25	8.2_A	21.71	21.45	0.359	1.022E+06	6.009	1.23E+06	2.98E+05
A10	FAM	0	Unkn-25	8.2_A	21.20	21.45	0.359	1.443E+06	6.159	1.23E+06	2.98E+05
A11	FAM	0	Unkn-33	15.3_C	22.05	22.00	0.067	8.109E+05	5.909	8.38E+05	3.84E+04
A12	FAM	0	Unkn-33	15.3_C	21.95	22.00	0.067	8.651E+05	5.937	8.38E+05	3.84E+04
B01	FAM	0	Std-02	STD2	20.12	20.13	0.017	3.000E+06	6.477	3.00E+06	0.00E+00
B02	FAM	0	Std-02	STD2	20.14	20.13	0.017	3.000E+06	6.477	3.00E+06	0.00E+00
B03	FAM	0	Unkn-02	3.2_C	21.73	21.62	0.151	1.010E+06	6.004	1.09E+06	1.12E+05
B04	FAM	0	Unkn-02	3.2_C	21.51	21.62	0.151	1.168E+06	6.067	1.09E+06	1.12E+05
B05	FAM	0	Unkn-10	8.2_B	21.82	21.91	0.126	9.457E+05	5.976	8.92E+05	7.61E+04
B06	FAM	0	Unkn-10	8.2_B	22.00	21.91	0.126	8.380E+05	5.923	8.92E+05	7.61E+04
B07	FAM	0	Unkn-18	6.5_D	22.28	22.27	0.012	6.931E+05	5.841	6.97E+05	5.56E+03
B08	FAM	0	Unkn-18	6.5_D	22.26	22.27	0.012	7.010E+05	5.846	6.97E+05	5.56E+03
B09	FAM	0	Unkn-26	3.2_A	21.11	21.11	0.010	1.539E+06	6.187	1.53E+06	1.09E+04
B10	FAM	0	Unkn-26	3.2_A	21.12	21.11	0.010	1.524E+06	6.183	1.53E+06	1.09E+04
B11	FAM	0	Unkn-34	17.4_C	21.28	21.25	0.051	1.365E+06	6.135	1.40E+06	4.87E+04
B12	FAM	0	Unkn-34	17.4_C	21.21	21.25	0.051	1.434E+06	6.157	1.40E+06	4.87E+04
C01	FAM	0	Std-03	STD3	23.48	23.46	0.022	3.000E+05	5.477	3.00E+05	0.00E+00
C02	FAM	0	Std-03	STD3	23.44	23.46	0.022	3.000E+05	5.477	3.00E+05	0.00E+00
C05	FAM	0	Unkn-11	11.1_B	22.12	22.12	0.009	7.703E+05	5.887	7.73E+05	4.48E+03
C06	FAM	0	Unkn-11	11.1_B	22.11	22.12	0.009	7.766E+05	5.890	7.73E+05	4.48E+03

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
C07	FAM	0	Unkn-19	9.6_C	22.38	22.40	0.040	6.491E+05	5.812	6.37E+05	1.75E+04
C08	FAM	0	Unkn-19	9.6_C	22.43	22.40	0.040	6.243E+05	5.795	6.37E+05	1.75E+04
C09	FAM	0	Unkn-27	6.4_B	21.83	21.81	0.030	9.404E+05	5.973	9.54E+05	1.94E+04
C10	FAM	0	Unkn-27	6.4_B	21.79	21.81	0.030	9.679E+05	5.986	9.54E+05	1.94E+04
C11	FAM	0	Unkn-35	12.6_D	22.08	22.10	0.031	7.939E+05	5.900	7.82E+05	1.64E+04
C12	FAM	0	Unkn-35	12.6_D	22.12	22.10	0.031	7.707E+05	5.887	7.82E+05	1.64E+04
D01	FAM	0	Std-04	STD4	26.87	26.87	0.001	3.000E+04	4.477	3.00E+04	0.00E+00
D02	FAM	0	Std-04	STD4	26.87	26.87	0.001	3.000E+04	4.477	3.00E+04	0.00E+00
D05	FAM	0	Unkn-12	10.3_B	22.28	22.25	0.039	6.923E+05	5.840	7.05E+05	1.86E+04
D06	FAM	0	Unkn-12	10.3_B	22.23	22.25	0.039	7.186E+05	5.856	7.05E+05	1.86E+04
D07	FAM	0	Unkn-20	2.4_C	22.19	22.11	0.111	7.385E+05	5.868	7.80E+05	5.88E+04
D08	FAM	0	Unkn-20	2.4_C	22.03	22.11	0.111	8.217E+05	5.915	7.80E+05	5.88E+04
D09	FAM	0	Unkn-28	10.1_D	21.95	21.98	0.039	8.671E+05	5.938	8.51E+05	2.26E+04
D10	FAM	0	Unkn-28	10.1_D	22.01	21.98	0.039	8.352E+05	5.922	8.51E+05	2.26E+04
D11	FAM	0	Unkn-36	12.2_C	21.78	21.87	0.120	9.701E+05	5.987	9.17E+05	7.46E+04
D12	FAM	0	Unkn-36	12.2_C	21.95	21.87	0.120	8.646E+05	5.937	9.17E+05	7.46E+04
E01	FAM	0	Std-05	STD5	30.44	30.31	0.184	3.000E+03	3.477	3.00E+03	0.00E+00
E02	FAM	0	Std-05	STD5	30.18	30.31	0.184	3.000E+03	3.477	3.00E+03	0.00E+00
E03	FAM	0	Unkn-05	2.3_A	21.73	21.70	0.039	1.007E+06	6.003	1.03E+06	2.75E+04
E04	FAM	0	Unkn-05	2.3_A	21.67	21.70	0.039	1.046E+06	6.020	1.03E+06	2.75E+04
E05	FAM	0	Unkn-13	9.4_A	21.54	21.53	0.021	1.143E+06	6.058	1.16E+06	1.68E+04
E06	FAM	0	Unkn-13	9.4_A	21.51	21.53	0.021	1.167E+06	6.067	1.16E+06	1.68E+04
E07	FAM	0	Unkn-21	6.2_B	21.42	21.47	0.067	1.242E+06	6.094	1.20E+06	5.46E+04
E08	FAM	0	Unkn-21	6.2_B	21.52	21.47	0.067	1.165E+06	6.066	1.20E+06	5.46E+04
E09	FAM	0	Unkn-29	10.3_D	21.67	21.71	0.050	1.049E+06	6.021	1.02E+06	3.44E+04
E10	FAM	0	Unkn-29	10.3_D	21.74	21.71	0.050	9.999E+05	6.000	1.02E+06	3.44E+04
E11	FAM	0	Unkn-37	3.3_D	22.17	22.28	0.151	7.473E+05	5.873	6.97E+05	7.13E+04
E12	FAM	0	Unkn-37	3.3_D	22.38	22.28	0.151	6.465E+05	5.811	6.97E+05	7.13E+04
F01	FAM	0	Std-06	STD6	33.29	33.71	0.605	3.000E+02	2.477	3.00E+02	0.00E+00
F02	FAM	0	Std-06	STD6	34.14	33.71	0.605	3.000E+02	2.477	3.00E+02	0.00E+00
F03	FAM	0	Unkn-06	11.4_C	22.47	22.43	0.056	6.073E+05	5.783	6.24E+05	2.36E+04
F04	FAM	0	Unkn-06	11.4_C	22.40	22.43	0.056	6.407E+05	5.807	6.24E+05	2.36E+04
F05	FAM	0	Unkn-14	2.4_A	22.10	22.06	0.061	7.822E+05	5.893	8.06E+05	3.34E+04
F06	FAM	0	Unkn-14	2.4_A	22.02	22.06	0.061	8.294E+05	5.919	8.06E+05	3.34E+04
F07	FAM	0	Unkn-22	3.4_A	20.91	20.87	0.052	1.762E+06	6.246	1.81E+06	6.33E+04
F08	FAM	0	Unkn-22	3.4_A	20.83	20.87	0.052	1.852E+06	6.268	1.81E+06	6.33E+04
F09	FAM	0	Unkn-30	12.4_D	21.66	21.66	0.003	1.056E+06	6.024	1.06E+06	1.99E+03
F10	FAM	0	Unkn-30	12.4_D	21.66	21.66	0.003	1.059E+06	6.025	1.06E+06	1.99E+03
F11	FAM	0	Unkn-38	7.1_C	21.00	20.95	0.075	1.652E+06	6.218	1.71E+06	8.76E+04
F12	FAM	0	Unkn-38	7.1_C	20.89	20.95	0.075	1.776E+06	6.250	1.71E+06	8.76E+04
G03	FAM	0	Unkn-07	9.2_A	21.97	21.93	0.056	8.533E+05	5.931	8.77E+05	3.36E+04
G04	FAM	0	Unkn-07	9.2_A	21.89	21.93	0.056	9.008E+05	5.955	8.77E+05	3.36E+04
G05	FAM	0	Unkn-15	11.3_D	22.84	22.83	0.002	4.749E+05	5.677	4.75E+05	8.03E+02
G06	FAM	0	Unkn-15	11.3_D	22.83	22.83	0.002	4.760E+05	5.678	4.75E+05	8.03E+02
G07	FAM	0	Unkn-23	12.5_C	22.00	22.00	0.005	8.388E+05	5.924	8.41E+05	2.91E+03
G08	FAM	0	Unkn-23	12.5_C	21.99	22.00	0.005	8.429E+05	5.926	8.41E+05	2.91E+03
G09	FAM	0	Unkn-31	12.6_B	22.03	22.00	0.047	8.209E+05	5.914	8.40E+05	2.68E+04
G10	FAM	0	Unkn-31	12.6_B	21.96	22.00	0.047	8.588E+05	5.934	8.40E+05	2.68E+04
G11	FAM	0	Unkn-39	6.4_D	22.60	22.61	0.016	5.570E+05	5.746	5.53E+05	6.16E+03

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
G12	FAM	0	Unkn-39	6.4_D	22.62	22.61	0.016	5.483E+05	5.739	5.53E+05	6.16E+03
H01	FAM	0	NTC-01	NEG	N/A	0.00	0.000	N/A	N/A	0.00E+00	0.00E+00
H02	FAM	0	NTC-01	NEG	N/A	0.00	0.000	N/A	N/A	0.00E+00	0.00E+00
H03	FAM	0	Unkn-08	13.4_B	22.31	22.28	0.044	6.805E+05	5.833	6.95E+05	2.09E+04
H04	FAM	0	Unkn-08	13.4_B	22.24	22.28	0.044	7.100E+05	5.851	6.95E+05	2.09E+04
H05	FAM	0	Unkn-16	13.1_D	22.03	22.08	0.069	8.206E+05	5.914	7.94E+05	3.74E+04
H06	FAM	0	Unkn-16	13.1_D	22.13	22.08	0.069	7.677E+05	5.885	7.94E+05	3.74E+04
H07	FAM	0	Unkn-24	11.1_A	21.69	21.68	0.014	1.037E+06	6.016	1.04E+06	9.83E+03
H08	FAM	0	Unkn-24	11.1_A	21.67	21.68	0.014	1.051E+06	6.022	1.04E+06	9.83E+03
H09	FAM	0	Unkn-32	12.6_A	21.46	21.43	0.047	1.209E+06	6.082	1.24E+06	3.92E+04
H10	FAM	0	Unkn-32	12.6_A	21.39	21.43	0.047	1.264E+06	6.102	1.24E+06	3.92E+04
H11	FAM	0	Unkn-40	7.2_C	21.43	21.45	0.025	1.232E+06	6.091	1.22E+06	2.09E+04
H12	FAM	0	Unkn-40	7.2_C	21.47	21.45	0.025	1.202E+06	6.080	1.22E+06	2.09E+04
A01	VIC		Std-01	STD1	18.13	18.12	0.018	3.000E+07	7.477	3.00E+07	0.00E+00
A02	VIC		Std-01	STD1	18.11	18.12	0.018	3.000E+07	7.477	3.00E+07	0.00E+00
A03	VIC		Unkn-01	7.1_B	27.50	27.45	0.077	5.213E+04	4.717	5.41E+04	2.81E+03
A04	VIC		Unkn-01	7.1_B	27.39	27.45	0.077	5.611E+04	4.749	5.41E+04	2.81E+03
A05	VIC		Unkn-09	15.5_B	28.34	28.33	0.008	2.967E+04	4.472	2.98E+04	1.53E+02
A06	VIC		Unkn-09	15.5_B	28.33	28.33	0.008	2.988E+04	4.475	2.98E+04	1.53E+02
A07	VIC		Unkn-17	17.2_D	26.99	27.02	0.042	7.383E+04	4.868	7.24E+04	2.06E+03
A08	VIC		Unkn-17	17.2_D	27.05	27.02	0.042	7.092E+04	4.851	7.24E+04	2.06E+03
A09	VIC		Unkn-25	8.2_A	28.17	27.86	0.432	3.322E+04	4.521	4.17E+04	1.20E+04
A10	VIC		Unkn-25	8.2_A	27.56	27.86	0.432	5.018E+04	4.701	4.17E+04	1.20E+04
A11	VIC		Unkn-33	15.3_C	27.41	27.45	0.044	5.531E+04	4.743	5.42E+04	1.60E+03
A12	VIC		Unkn-33	15.3_C	27.48	27.45	0.044	5.305E+04	4.725	5.42E+04	1.60E+03
B01	VIC		Std-02	STD2	21.47	21.45	0.025	3.000E+06	6.477	3.00E+06	0.00E+00
B02	VIC		Std-02	STD2	21.43	21.45	0.025	3.000E+06	6.477	3.00E+06	0.00E+00
B03	VIC		Unkn-02	3.2_C	27.41	27.27	0.201	5.533E+04	4.743	6.12E+04	8.29E+03
B04	VIC		Unkn-02	3.2_C	27.13	27.27	0.201	6.706E+04	4.826	6.12E+04	8.29E+03
B05	VIC		Unkn-10	8.2_B	27.74	27.85	0.150	4.433E+04	4.647	4.14E+04	4.17E+03
B06	VIC		Unkn-10	8.2_B	27.95	27.85	0.150	3.843E+04	4.585	4.14E+04	4.17E+03
B07	VIC		Unkn-18	6.5_D	27.37	27.33	0.054	5.692E+04	4.755	5.84E+04	2.12E+03
B08	VIC		Unkn-18	6.5_D	27.30	27.33	0.054	5.992E+04	4.778	5.84E+04	2.12E+03
B09	VIC		Unkn-26	3.2_A	26.91	26.99	0.113	7.795E+04	4.892	7.40E+04	5.62E+03
B10	VIC		Unkn-26	3.2_A	27.07	26.99	0.113	7.000E+04	4.845	7.40E+04	5.62E+03
B11	VIC		Unkn-34	17.4_C	27.00	26.97	0.042	7.335E+04	4.865	7.48E+04	2.10E+03
B12	VIC		Unkn-34	17.4_C	26.94	26.97	0.042	7.633E+04	4.883	7.48E+04	2.10E+03
C01	VIC		Std-03	STD3	24.96	24.95	0.013	3.000E+05	5.477	3.00E+05	0.00E+00
C02	VIC		Std-03	STD3	24.94	24.95	0.013	3.000E+05	5.477	3.00E+05	0.00E+00
C05	VIC		Unkn-11	11.1_B	27.27	27.31	0.050	6.092E+04	4.785	5.95E+04	2.00E+03
C06	VIC		Unkn-11	11.1_B	27.34	27.31	0.050	5.809E+04	4.764	5.95E+04	2.00E+03
C07	VIC		Unkn-19	9.6_C	27.73	27.79	0.088	4.464E+04	4.650	4.28E+04	2.54E+03
C08	VIC		Unkn-19	9.6_C	27.86	27.79	0.088	4.105E+04	4.613	4.28E+04	2.54E+03
C09	VIC		Unkn-27	6.4_B	27.53	27.43	0.131	5.133E+04	4.710	5.48E+04	4.85E+03
C10	VIC		Unkn-27	6.4_B	27.34	27.43	0.131	5.818E+04	4.765	5.48E+04	4.85E+03
C11	VIC		Unkn-35	12.6_D	28.21	28.31	0.131	3.226E+04	4.509	3.04E+04	2.68E+03
C12	VIC		Unkn-35	12.6_D	28.40	28.31	0.131	2.847E+04	4.454	3.04E+04	2.68E+03
D01	VIC		Std-04	STD4	28.31	28.29	0.028	3.000E+04	4.477	3.00E+04	0.00E+00
D02	VIC		Std-04	STD4	28.27	28.29	0.028	3.000E+04	4.477	3.00E+04	0.00E+00

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
D05	VIC		Unkn-12	10.3_B	28.12	28.14	0.038	3.444E+04	4.537	3.38E+04	8.71E+02
D06	VIC		Unkn-12	10.3_B	28.17	28.14	0.038	3.321E+04	4.521	3.38E+04	8.71E+02
D07	VIC		Unkn-20	2.4_C	28.16	28.08	0.108	3.350E+04	4.525	3.53E+04	2.57E+03
D08	VIC		Unkn-20	2.4_C	28.00	28.08	0.108	3.714E+04	4.570	3.53E+04	2.57E+03
D09	VIC		Unkn-28	10.1_D	27.42	27.41	0.010	5.528E+04	4.743	5.56E+04	3.78E+02
D10	VIC		Unkn-28	10.1_D	27.40	27.41	0.010	5.582E+04	4.747	5.56E+04	3.78E+02
D11	VIC		Unkn-36	12.2_C	27.56	27.64	0.119	5.023E+04	4.701	4.75E+04	3.83E+03
D12	VIC		Unkn-36	12.2_C	27.73	27.64	0.119	4.481E+04	4.651	4.75E+04	3.83E+03
E01	VIC		Std-05	STD5	31.89	31.75	0.204	3.000E+03	3.477	3.00E+03	0.00E+00
E02	VIC		Std-05	STD5	31.60	31.75	0.204	3.000E+03	3.477	3.00E+03	0.00E+00
E03	VIC		Unkn-05	2.3_A	27.29	27.29	0.005	6.021E+04	4.780	6.04E+04	2.10E+02
E04	VIC		Unkn-05	2.3_A	27.28	27.29	0.005	6.050E+04	4.782	6.04E+04	2.10E+02
E05	VIC		Unkn-13	9.4_A	27.07	27.08	0.011	6.977E+04	4.844	6.94E+04	5.26E+02
E06	VIC		Unkn-13	9.4_A	27.09	27.08	0.011	6.903E+04	4.839	6.94E+04	5.26E+02
E07	VIC		Unkn-21	6.2_B	27.61	27.77	0.219	4.837E+04	4.685	4.38E+04	6.46E+03
E08	VIC		Unkn-21	6.2_B	27.92	27.77	0.219	3.923E+04	4.594	4.38E+04	6.46E+03
E09	VIC		Unkn-29	10.3_D	27.13	27.11	0.038	6.690E+04	4.825	6.81E+04	1.76E+03
E10	VIC		Unkn-29	10.3_D	27.08	27.11	0.038	6.939E+04	4.841	6.81E+04	1.76E+03
E11	VIC		Unkn-37	3.3_D	27.47	27.55	0.122	5.344E+04	4.728	5.05E+04	4.14E+03
E12	VIC		Unkn-37	3.3_D	27.64	27.55	0.122	4.758E+04	4.677	5.05E+04	4.14E+03
F01	VIC		Std-06	STD6	34.78	35.14	0.515	3.000E+02	2.477	3.00E+02	0.00E+00
F02	VIC		Std-06	STD6	35.50	35.14	0.515	3.000E+02	2.477	3.00E+02	0.00E+00
F03	VIC		Unkn-06	11.4_C	27.64	27.60	0.065	4.734E+04	4.675	4.89E+04	2.16E+03
F04	VIC		Unkn-06	11.4_C	27.55	27.60	0.065	5.040E+04	4.702	4.89E+04	2.16E+03
F05	VIC		Unkn-14	2.4_A	27.54	27.44	0.134	5.086E+04	4.706	5.43E+04	4.92E+03
F06	VIC		Unkn-14	2.4_A	27.35	27.44	0.134	5.783E+04	4.762	5.43E+04	4.92E+03
F07	VIC		Unkn-22	3.4_A	26.97	26.93	0.062	7.467E+04	4.873	7.69E+04	3.22E+03
F08	VIC		Unkn-22	3.4_A	26.88	26.93	0.062	7.922E+04	4.899	7.69E+04	3.22E+03
F09	VIC		Unkn-30	12.4_D	26.92	26.92	0.004	7.749E+04	4.889	7.74E+04	1.93E+02
F10	VIC		Unkn-30	12.4_D	26.92	26.92	0.004	7.722E+04	4.888	7.74E+04	1.93E+02
F11	VIC		Unkn-38	7.1_C	26.91	26.80	0.144	7.802E+04	4.892	8.38E+04	8.12E+03
F12	VIC		Unkn-38	7.1_C	26.70	26.80	0.144	8.951E+04	4.952	8.38E+04	8.12E+03
G03	VIC		Unkn-07	9.2_A	27.28	27.33	0.060	6.045E+04	4.781	5.88E+04	2.38E+03
G04	VIC		Unkn-07	9.2_A	27.37	27.33	0.060	5.708E+04	4.756	5.88E+04	2.38E+03
G05	VIC		Unkn-15	11.3_D	28.54	28.49	0.081	2.581E+04	4.412	2.68E+04	1.46E+03
G06	VIC		Unkn-15	11.3_D	28.43	28.49	0.081	2.788E+04	4.445	2.68E+04	1.46E+03
G07	VIC		Unkn-23	12.5_C	27.89	27.81	0.108	4.013E+04	4.603	4.23E+04	3.08E+03
G08	VIC		Unkn-23	12.5_C	27.74	27.81	0.108	4.449E+04	4.648	4.23E+04	3.08E+03
G09	VIC		Unkn-31	12.6_B	28.28	28.28	0.003	3.077E+04	4.488	3.08E+04	5.63E+01
G10	VIC		Unkn-31	12.6_B	28.28	28.28	0.003	3.085E+04	4.489	3.08E+04	5.63E+01
G11	VIC		Unkn-39	6.4_D	28.94	28.98	0.057	1.973E+04	4.295	1.92E+04	7.44E+02
G12	VIC		Unkn-39	6.4_D	29.02	28.98	0.057	1.868E+04	4.271	1.92E+04	7.44E+02
H01	VIC		NTC-01	NEG	39.08	39.08	0.000	N/A	N/A	N/A	0.00E+00
H02	VIC		NTC-01	NEG	N/A	0.00	0.000	N/A	N/A	0.00E+00	0.00E+00
H03	VIC		Unkn-08	13.4_B	27.32	27.29	0.042	5.885E+04	4.770	6.01E+04	1.72E+03
H04	VIC		Unkn-08	13.4_B	27.26	27.29	0.042	6.128E+04	4.787	6.01E+04	1.72E+03
H05	VIC		Unkn-16	13.1_D	27.49	27.57	0.108	5.257E+04	4.721	5.00E+04	3.64E+03
H06	VIC		Unkn-16	13.1_D	27.64	27.57	0.108	4.743E+04	4.676	5.00E+04	3.64E+03
H07	VIC		Unkn-24	11.1_A	27.59	27.59	0.005	4.923E+04	4.692	4.91E+04	1.76E+02

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
H08	VIC		Unkn-24	11.1_A	27.59	27.59	0.005	4.898E+04	4.690	4.91E+04	1.76E+02
H09	VIC		Unkn-32	12.6_A	27.70	27.66	0.057	4.572E+04	4.660	4.70E+04	1.80E+03
H10	VIC		Unkn-32	12.6_A	27.62	27.66	0.057	4.827E+04	4.684	4.70E+04	1.80E+03
H11	VIC		Unkn-40	7.2_C	27.39	27.53	0.206	5.640E+04	4.751	5.14E+04	7.14E+03
H12	VIC		Unkn-40	7.2_C	27.68	27.53	0.206	4.630E+04	4.666	5.14E+04	7.14E+03

Bar Chart

Normalized expression analysis is not possible, either because no target is assigned as a reference or the selected target(s) is not a

Target Names

Name	Full Name	Reference	Auto Efficiency	Efficiency
0	0	False	Yes	97.2%

Sample Names

Name	Full Name	Control
10.1_D	10.1_D	No
10.3_B	10.3_B	No
10.3_D	10.3_D	No
11.1_A	11.1_A	No
11.1_B	11.1_B	No
11.3_D	11.3_D	No
11.4_C	11.4_C	No
12.2_C	12.2_C	No
12.4_D	12.4_D	No
12.5_C	12.5_C	No
12.6_A	12.6_A	No
12.6_B	12.6_B	No
12.6_D	12.6_D	No
13.1_D	13.1_D	No
13.4_B	13.4_B	No
15.3_C	15.3_C	No
15.5_B	15.5_B	No
17.2_D	17.2_D	No
17.4_C	17.4_C	No
2.3_A	2.3_A	No
2.4_A	2.4_A	No
2.4_C	2.4_C	No
3.2_A	3.2_A	No
3.2_C	3.2_C	No
3.3_D	3.3_D	No
3.4_A	3.4_A	No
6.2_B	6.2_B	No
6.4_B	6.4_B	No
6.4_D	6.4_D	No

Sample Names

Name	Full Name	Control
6.5_D	6.5_D	No
7.1_B	7.1_B	No
7.1_C	7.1_C	No
7.2_C	7.2_C	No
8.2_A	8.2_A	No
8.2_B	8.2_B	No
9.2_A	9.2_A	No
9.4_A	9.4_A	No
9.6_C	9.6_C	No
STD1	STD1	No
STD2	STD2	No
STD3	STD3	No
STD4	STD4	No
STD5	STD5	No
STD6	STD6	No

Gene Expression - Bar Chart Data

Target	Sample	Control	Expression	Expression SEM	Corrected Expression SEM	Mean Cq	Cq SEM	P-Value
0	10.1_D		N/A	N/A	N/A	21.98	0.02759	N/A
0	10.3_B		N/A	N/A	N/A	22.25	0.02745	N/A
0	10.3_D		N/A	N/A	N/A	21.71	0.03501	N/A
0	11.1_A		N/A	N/A	N/A	21.68	0.00980	N/A
0	11.1_B		N/A	N/A	N/A	22.12	0.00603	N/A
0	11.3_D		N/A	N/A	N/A	22.83	0.00176	N/A
0	11.4_C		N/A	N/A	N/A	22.43	0.03947	N/A
0	12.2_C		N/A	N/A	N/A	21.87	0.08478	N/A
0	12.4_D		N/A	N/A	N/A	21.66	0.00196	N/A
0	12.5_C		N/A	N/A	N/A	22.00	0.00361	N/A
0	12.6_A		N/A	N/A	N/A	21.43	0.03304	N/A
0	12.6_B		N/A	N/A	N/A	22.00	0.03317	N/A
0	12.6_D		N/A	N/A	N/A	22.10	0.02185	N/A
0	13.1_D		N/A	N/A	N/A	22.08	0.04902	N/A
0	13.4_B		N/A	N/A	N/A	22.28	0.03124	N/A
0	15.3_C		N/A	N/A	N/A	22.00	0.04768	N/A
0	15.5_B		N/A	N/A	N/A	21.99	0.00224	N/A
0	17.2_D		N/A	N/A	N/A	21.58	0.03776	N/A
0	17.4_C		N/A	N/A	N/A	21.25	0.03620	N/A
0	2.3_A		N/A	N/A	N/A	21.70	0.02790	N/A
0	2.4_A		N/A	N/A	N/A	22.06	0.04312	N/A
0	2.4_C		N/A	N/A	N/A	22.11	0.07857	N/A
0	3.2_A		N/A	N/A	N/A	21.11	0.00738	N/A
0	3.2_C		N/A	N/A	N/A	21.62	0.10695	N/A
0	3.3_D		N/A	N/A	N/A	22.28	0.10669	N/A
0	3.4_A		N/A	N/A	N/A	20.87	0.03649	N/A
0	6.2_B		N/A	N/A	N/A	21.47	0.04722	N/A
0	6.4_B		N/A	N/A	N/A	21.81	0.02118	N/A

Gene Expression - Bar Chart Data

Target	Sample	Control	Expression	Expression SEM	Corrected Expression SEM	Mean Cq	Cq SEM	P-Value
0	6.4_D		N/A	N/A	N/A	22.61	0.01161	N/A
0	6.5_D		N/A	N/A	N/A	22.27	0.00831	N/A
0	7.1_B		N/A	N/A	N/A	21.47	0.09121	N/A
0	7.1_C		N/A	N/A	N/A	20.95	0.05322	N/A
0	7.2_C		N/A	N/A	N/A	21.45	0.01786	N/A
0	8.2_A		N/A	N/A	N/A	21.45	0.25395	N/A
0	8.2_B		N/A	N/A	N/A	21.91	0.08899	N/A
0	9.2_A		N/A	N/A	N/A	21.93	0.03988	N/A
0	9.4_A		N/A	N/A	N/A	21.53	0.01515	N/A
0	9.6_C		N/A	N/A	N/A	22.40	0.02860	N/A
0	STD1		N/A	N/A	N/A	16.77	0.05027	N/A
0	STD2		N/A	N/A	N/A	20.13	0.01171	N/A
0	STD3		N/A	N/A	N/A	23.46	0.01579	N/A
0	STD4		N/A	N/A	N/A	26.87	0.00101	N/A
0	STD5		N/A	N/A	N/A	30.31	0.12978	N/A
0	STD6		N/A	N/A	N/A	33.71	0.42785	N/A

QC Parameters

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Negative control with a Cq less than	38	True		False	
NTC with a Cq less than	38	True		False	
NRT with a Cq less than	38	True		False	
Positive control with a Cq greater than	30	True		False	
Unknown without a Cq	N/A	True		False	
Standard without a Cq	N/A	True		False	
Efficiency greater than	110.0	True			
Efficiency less than	90.0	True			
Std Curve R^2 less than	0.980	True			

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Replicate group Cq Std Dev greater than	0.20	True	FAM:A9, A10, F1, F2. VIC:A9, A10, B3, B4, E1, E2, E7, E8, F1, F2, H11, H12.	False	