



GT_Plate1_Telomeres_12_11_2024.pcrd

3/27/2025 9:31 AM

Report Information

User: BioRad/admin
Data File Name: GT_Plate1_Telomeres_12_11_2024.pcrd
Data File Path: C:\Users\CSMTSS1\Documents\Tortoises\Plates\Telomeres
Well Group Name: All Wells
Report Differs from Last Save: No

Run Setup

Run Information

Run Date: 12/11/2024 4:12 PM
Run User: admin
Run Type: User-defined
Plate File: Elgaria_Multiplex.prcl.pltd
ID:
Notes:
Sample Volume: 20
Temperature Control Mode: Calculated
Lid Temperature: 105
Base Serial Number: 795BR20744
Optical Head Serial Number: 795BR20744

Protocol

- 1: 95.0°C for 3:00
- 2: 95.0°C for 0:15
- 3: 49.0°C for 0:30
- 4: 55.0°C for 0:30
- 5: GOTO 3, 1 more times
- 6: 95.0°C for 0:15
- 7: 56.0°C for 0:30
- 8: 60.0°C for 0:30
Plate Read
- 9: GOTO 6, 34 more times
- 10: Melt Curve 60.0°C to 95.0°C: Increment 0.5°C 0:05
Plate Read

Plate Display


	1	2	3	4	5	6	7	8	9	10	11	12
A	Std-1 SYBR STD1	Std-1 SYBR STD1	Unk-1 SYBR 11.3_A	Unk-1 SYBR 11.3_A	Unk-9 SYBR 9.1_D	Unk-9 SYBR 9.1_D	Unk-17 SYBR 5.1_C	Unk-17 SYBR 5.1_C	Unk-25 SYBR 7.5_C	Unk-25 SYBR 7.5_C	Unk-33 SYBR 9.3_A	Unk-33 SYBR 9.3_A
B	Std-2 SYBR STD2	Std-2 SYBR STD2	Unk-2 SYBR 3.1_B	Unk-2 SYBR 3.1_B	Unk-10 SYBR 5.3_B	Unk-10 SYBR 5.3_B	Unk-18 SYBR 15.3_A	Unk-18 SYBR 15.3_A	Unk-26 SYBR 13.2_B	Unk-26 SYBR 13.2_B	Unk-34 SYBR 17.5_C	Unk-34 SYBR 17.5_C
C	Std-3 SYBR STD3	Std-3 SYBR STD3	Unk-3 SYBR 9.1_C	Unk-3 SYBR 9.1_C	Unk-11 SYBR 15.2_C	Unk-11 SYBR 15.2_C	Unk-19 SYBR 3.3_C	Unk-19 SYBR 3.3_C	Unk-27 SYBR 10.6_A	Unk-27 SYBR 10.6_A	Unk-35 SYBR 9.4_D	Unk-35 SYBR 9.4_D


Plate Display

	1	2	3	4	5	6	7	8	9	10	11	12
D	Std-4 SYBR STD4	Std-4 SYBR STD4	Unk-4 SYBR 8.1_A	Unk-4 SYBR 8.1_A	*Unk-12 SYBR 10.6_D	*Unk-12 SYBR 10.6_D	Unk-20 SYBR 9.4_B	Unk-20 SYBR 9.4_B	Unk-28 SYBR 5.4_D	Unk-28 SYBR 5.4_D	Unk-36 SYBR 9.5_B	Unk-36 SYBR 9.5_B
E	Unk SYBR STD5	Unk SYBR STD5	Unk-5 SYBR 12.2_A	Unk-5 SYBR 12.2_A	Unk-13 SYBR 9.2_C	Unk-13 SYBR 9.2_C	Unk-21 SYBR 13.3_B	Unk-21 SYBR 13.3_B	Unk-29 SYBR 2.3_C	Unk-29 SYBR 2.3_C	Unk-37 SYBR 10.4_D	Unk-37 SYBR 10.4_D
F	Unk SYBR STD6	Unk SYBR STD6	Unk-6 SYBR 15.5_A	Unk-6 SYBR 15.5_A	Unk-14 SYBR 14.5_D	Unk-14 SYBR 14.5_D	Unk-22 SYBR 15.2_A	Unk-22 SYBR 15.2_A	Unk-30 SYBR 17.3_D	Unk-30 SYBR 17.3_D	Unk-38 SYBR 5.2_C	Unk-38 SYBR 5.2_C
G	Unk SYBR STD7	Unk SYBR STD7	Unk-7 SYBR 3.1_C	Unk-7 SYBR 3.1_C	Unk-15 SYBR 17.3_A	Unk-15 SYBR 17.3_A	Unk-23 SYBR 10.6_B	Unk-23 SYBR 10.6_B	Unk-31 SYBR 2.4_B	Unk-31 SYBR 2.4_B	Unk-39 SYBR 12.4_A	Unk-39 SYBR 12.4_A
H	NTC-1 SYBR NEG	NTC-1 SYBR NEG	Unk-8 SYBR 7.1_D	Unk-8 SYBR 7.1_D	Unk-16 SYBR 13.4_D	Unk-16 SYBR 13.4_D	Unk-24 SYBR 17.2_B	Unk-24 SYBR 17.2_B	Unk-32 SYBR 3.2_B	Unk-32 SYBR 3.2_B	Unk-40 SYBR 7.4_D	Unk-40 SYBR 7.4_D

Quantification

Step #: 8
Analysis Mode: Fluorophore
Baseline Setting: Baseline Subtracted Curve Fit
Cq Determination: Single Threshold
Baseline Method:
SYBR: Auto Calculated
Threshold Setting:
SYBR: 150.00, User Defined





Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
A01	SYBR		Std-01	STD1	17.58	17.44	0.201	3.000E+11	11.477	3.00E+11	0.00E+00
A02	SYBR		Std-01	STD1	17.29	17.44	0.201	3.000E+11	11.477	3.00E+11	0.00E+00
A03	SYBR		Unkn-01	11.3_A	18.16	18.04	0.165	2.032E+11	11.308	2.19E+11	2.24E+10
A04	SYBR		Unkn-01	11.3_A	17.92	18.04	0.165	2.348E+11	11.371	2.19E+11	2.24E+10
A05	SYBR		Unkn-09	9.1_D	18.09	18.04	0.065	2.121E+11	11.327	2.18E+11	8.82E+09
A06	SYBR		Unkn-09	9.1_D	17.99	18.04	0.065	2.246E+11	11.351	2.18E+11	8.82E+09
A07	SYBR		Unkn-17	5.1_C	18.77	18.76	0.016	1.389E+11	11.143	1.40E+11	1.41E+09
A08	SYBR		Unkn-17	5.1_C	18.75	18.76	0.016	1.409E+11	11.149	1.40E+11	1.41E+09
A09	SYBR		Unkn-25	7.5_C	18.59	18.52	0.097	1.557E+11	11.192	1.63E+11	9.75E+09
A10	SYBR		Unkn-25	7.5_C	18.45	18.52	0.097	1.695E+11	11.229	1.63E+11	9.75E+09
A11	SYBR		Unkn-33	9.3_A	18.11	18.12	0.016	2.096E+11	11.321	2.08E+11	2.00E+09
A12	SYBR		Unkn-33	9.3_A	18.13	18.12	0.016	2.068E+11	11.315	2.08E+11	2.00E+09
B01	SYBR		Std-02	STD2	21.23	21.13	0.139	3.000E+10	10.477	3.00E+10	0.00E+00
B02	SYBR		Std-02	STD2	21.03	21.13	0.139	3.000E+10	10.477	3.00E+10	0.00E+00
B03	SYBR		Unkn-02	3.1_B	18.16	18.27	0.157	2.028E+11	11.307	1.90E+11	1.84E+10
B04	SYBR		Unkn-02	3.1_B	18.38	18.27	0.157	1.768E+11	11.248	1.90E+11	1.84E+10
B05	SYBR		Unkn-10	5.3_B	18.08	18.08	0.003	2.124E+11	11.327	2.13E+11	4.21E+08
B06	SYBR		Unkn-10	5.3_B	18.08	18.08	0.003	2.130E+11	11.328	2.13E+11	4.21E+08
B07	SYBR		Unkn-18	15.3_A	18.84	18.87	0.048	1.334E+11	11.125	1.31E+11	3.91E+09
B08	SYBR		Unkn-18	15.3_A	18.90	18.87	0.048	1.279E+11	11.107	1.31E+11	3.91E+09
B09	SYBR		Unkn-26	13.2_B	18.32	18.20	0.179	1.832E+11	11.263	1.99E+11	2.20E+10
B10	SYBR		Unkn-26	13.2_B	18.07	18.20	0.179	2.143E+11	11.331	1.99E+11	2.20E+10
B11	SYBR		Unkn-34	17.5_C	18.23	18.20	0.040	1.938E+11	11.287	1.97E+11	4.86E+09
B12	SYBR		Unkn-34	17.5_C	18.18	18.20	0.040	2.007E+11	11.303	1.97E+11	4.86E+09
C01	SYBR		Std-03	STD3	25.53	25.47	0.085	3.000E+09	9.477	3.00E+09	0.00E+00
C02	SYBR		Std-03	STD3	25.41	25.47	0.085	3.000E+09	9.477	3.00E+09	0.00E+00
C03	SYBR		Unkn-03	9.1_C	18.10	18.10	0.002	2.108E+11	11.324	2.11E+11	2.06E+08
C04	SYBR		Unkn-03	9.1_C	18.10	18.10	0.002	2.105E+11	11.323	2.11E+11	2.06E+08

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
C05	SYBR		Unkn-11	15.2_C	17.97	17.94	0.036	2.285E+11	11.359	2.32E+11	5.22E+09
C06	SYBR		Unkn-11	15.2_C	17.92	17.94	0.036	2.359E+11	11.373	2.32E+11	5.22E+09
C07	SYBR		Unkn-19	3.3_C	19.77	19.75	0.021	7.498E+10	10.875	7.57E+10	1.01E+09
C08	SYBR		Unkn-19	3.3_C	19.74	19.75	0.021	7.640E+10	10.883	7.57E+10	1.01E+09
C09	SYBR		Unkn-27	10.6_A	18.39	18.39	0.002	1.758E+11	11.245	1.76E+11	2.12E+08
C10	SYBR		Unkn-27	10.6_A	18.39	18.39	0.002	1.761E+11	11.246	1.76E+11	2.12E+08
C11	SYBR		Unkn-35	9.4_D	18.27	18.31	0.056	1.889E+11	11.276	1.84E+11	6.41E+09
C12	SYBR		Unkn-35	9.4_D	18.35	18.31	0.056	1.798E+11	11.255	1.84E+11	6.41E+09
D01	SYBR		Std-04	STD4	28.40	28.39	0.016	3.000E+08	8.477	3.00E+08	0.00E+00
D02	SYBR		Std-04	STD4	28.38	28.39	0.016	3.000E+08	8.477	3.00E+08	0.00E+00
D03	SYBR		Unkn-04	8.1_A	18.10	18.06	0.054	2.110E+11	11.324	2.16E+11	7.24E+09
D04	SYBR		Unkn-04	8.1_A	18.02	18.06	0.054	2.212E+11	11.345	2.16E+11	7.24E+09
D07	SYBR		Unkn-20	9.4_B	18.29	18.67	0.549	1.876E+11	11.273	1.52E+11	5.06E+10
D08	SYBR		Unkn-20	9.4_B	19.06	18.67	0.549	1.161E+11	11.065	1.52E+11	5.06E+10
D09	SYBR		Unkn-28	5.4_D	18.37	18.33	0.066	1.775E+11	11.249	1.83E+11	7.50E+09
D10	SYBR		Unkn-28	5.4_D	18.28	18.33	0.066	1.881E+11	11.275	1.83E+11	7.50E+09
D11	SYBR		Unkn-36	9.5_B	18.08	18.12	0.061	2.130E+11	11.328	2.07E+11	7.88E+09
D12	SYBR		Unkn-36	9.5_B	18.17	18.12	0.061	2.019E+11	11.305	2.07E+11	7.88E+09
E01	SYBR		Unkn	STD5	29.14	29.14	0.000	2.257E+08	8.354	2.26E+08	0.00E+00
E02	SYBR		Unkn	STD5	29.22	29.22	0.000	2.148E+08	8.332	2.15E+08	0.00E+00
E03	SYBR		Unkn-05	12.2_A	17.92	17.67	0.348	2.356E+11	11.372	2.78E+11	5.94E+10
E04	SYBR		Unkn-05	12.2_A	17.42	17.67	0.348	3.196E+11	11.505	2.78E+11	5.94E+10
E05	SYBR		Unkn-13	9.2_C	18.09	18.20	0.160	2.117E+11	11.326	1.98E+11	1.95E+10
E06	SYBR		Unkn-13	9.2_C	18.32	18.20	0.160	1.841E+11	11.265	1.98E+11	1.95E+10
E07	SYBR		Unkn-21	13.3_B	18.62	18.66	0.055	1.527E+11	11.184	1.49E+11	5.07E+09
E08	SYBR		Unkn-21	13.3_B	18.70	18.66	0.055	1.455E+11	11.163	1.49E+11	5.07E+09
E09	SYBR		Unkn-29	2.3_C	17.93	17.92	0.018	2.331E+11	11.367	2.35E+11	2.60E+09
E10	SYBR		Unkn-29	2.3_C	17.91	17.92	0.018	2.367E+11	11.374	2.35E+11	2.60E+09
E11	SYBR		Unkn-37	10.4_D	18.31	18.35	0.053	1.845E+11	11.266	1.80E+11	5.96E+09
E12	SYBR		Unkn-37	10.4_D	18.39	18.35	0.053	1.760E+11	11.246	1.80E+11	5.96E+09
F01	SYBR		Unkn	STD6	29.28	29.28	0.000	2.076E+08	8.317	2.08E+08	0.00E+00
F02	SYBR		Unkn	STD6	29.34	29.34	0.000	2.002E+08	8.301	2.00E+08	0.00E+00
F03	SYBR		Unkn-06	15.5_A	17.73	17.69	0.057	2.641E+11	11.422	2.71E+11	9.60E+09
F04	SYBR		Unkn-06	15.5_A	17.65	17.69	0.057	2.776E+11	11.443	2.71E+11	9.60E+09
F05	SYBR		Unkn-14	14.5_D	18.39	18.39	0.008	1.754E+11	11.244	1.76E+11	8.51E+08
F06	SYBR		Unkn-14	14.5_D	18.38	18.39	0.008	1.766E+11	11.247	1.76E+11	8.51E+08
F07	SYBR		Unkn-22	15.2_A	19.55	19.43	0.176	8.559E+10	10.932	9.27E+10	1.01E+10
F08	SYBR		Unkn-22	15.2_A	19.30	19.43	0.176	9.985E+10	10.999	9.27E+10	1.01E+10
F09	SYBR		Unkn-30	17.3_D	18.25	18.30	0.067	1.919E+11	11.283	1.87E+11	7.68E+09
F10	SYBR		Unkn-30	17.3_D	18.34	18.30	0.067	1.811E+11	11.258	1.87E+11	7.68E+09
F11	SYBR		Unkn-38	5.2_C	18.63	18.67	0.061	1.517E+11	11.181	1.48E+11	5.59E+09
F12	SYBR		Unkn-38	5.2_C	18.72	18.67	0.061	1.438E+11	11.158	1.48E+11	5.59E+09
G01	SYBR		Unkn	STD7	29.57	29.57	0.000	1.731E+08	8.238	1.73E+08	0.00E+00
G02	SYBR		Unkn	STD7	29.34	29.34	0.000	2.004E+08	8.302	2.00E+08	0.00E+00
G03	SYBR		Unkn-07	3.1_C	18.60	18.57	0.049	1.542E+11	11.188	1.58E+11	4.82E+09
G04	SYBR		Unkn-07	3.1_C	18.53	18.57	0.049	1.611E+11	11.207	1.58E+11	4.82E+09
G05	SYBR		Unkn-15	17.3_A	18.33	18.47	0.193	1.824E+11	11.261	1.68E+11	2.00E+10
G06	SYBR		Unkn-15	17.3_A	18.60	18.47	0.193	1.541E+11	11.188	1.68E+11	2.00E+10
G07	SYBR		Unkn-23	10.6_B	18.70	18.82	0.162	1.449E+11	11.161	1.35E+11	1.36E+10

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
G08	SYBR		Unkn-23	10.6_B	18.93	18.82	0.162	1.257E+11	11.099	1.35E+11	1.36E+10
G09	SYBR		Unkn-31	2.4_B	17.59	17.93	0.478	2.879E+11	11.459	2.39E+11	6.96E+10
G10	SYBR		Unkn-31	2.4_B	18.27	17.93	0.478	1.894E+11	11.277	2.39E+11	6.96E+10
G11	SYBR		Unkn-39	12.4_A	18.40	18.51	0.150	1.744E+11	11.242	1.64E+11	1.52E+10
G12	SYBR		Unkn-39	12.4_A	18.62	18.51	0.150	1.529E+11	11.185	1.64E+11	1.52E+10
H01	SYBR		NTC-01	NEG	29.56	29.52	0.061	N/A	N/A	N/A	N/A
H02	SYBR		NTC-01	NEG	29.48	29.52	0.061	N/A	N/A	N/A	N/A
H03	SYBR		Unkn-08	7.1_D	18.49	18.46	0.040	1.655E+11	11.219	1.68E+11	4.19E+09
H04	SYBR		Unkn-08	7.1_D	18.43	18.46	0.040	1.714E+11	11.234	1.68E+11	4.19E+09
H05	SYBR		Unkn-16	13.4_D	18.02	18.05	0.043	2.206E+11	11.344	2.16E+11	5.82E+09
H06	SYBR		Unkn-16	13.4_D	18.09	18.05	0.043	2.123E+11	11.327	2.16E+11	5.82E+09
H07	SYBR		Unkn-24	17.2_B	18.88	18.95	0.089	1.295E+11	11.112	1.25E+11	6.87E+09
H08	SYBR		Unkn-24	17.2_B	19.01	18.95	0.089	1.197E+11	11.078	1.25E+11	6.87E+09
H09	SYBR		Unkn-32	3.2_B	18.18	18.17	0.011	2.007E+11	11.302	2.02E+11	1.42E+09
H10	SYBR		Unkn-32	3.2_B	18.16	18.17	0.011	2.027E+11	11.307	2.02E+11	1.42E+09
H11	SYBR		Unkn-40	7.4_D	18.19	18.20	0.017	1.990E+11	11.299	1.98E+11	2.07E+09
H12	SYBR		Unkn-40	7.4_D	18.21	18.20	0.017	1.961E+11	11.293	1.98E+11	2.07E+09

Bar Chart

Expression analysis requires that at least two wells contain sample names, targets, and valid Cqs.

QC Parameters

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Negative control with a Cq less than	38	True		False	
NTC with a Cq less than	38	True	SYBR:H1, H2.	False	
NRT with a Cq less than	38	True		False	
Positive control with a Cq greater than	30	True		False	
Unknown without a Cq	N/A	True		False	
Standard without a Cq	N/A	True		False	
Efficiency greater than	110.0	True			

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Efficiency less than	90.0	True	SYBR		
Std Curve R^2 less than	0.980	True			
Replicate group Cq Std Dev greater than	0.20	True	SYBR:A1, A2, D7, D8, E3, E4, G9, G10.	False	