



GT_Plate3_Telomeres_12_16_2024.pcrd

3/27/2025 9:37 AM

Report Information

User: BioRad/admin
Data File Name: GT_Plate3_Telomeres_12_16_2024.pcrd
Data File Path: C:\Users\CSMTSS1\Documents\Tortoises\Plates\Telomeres
Well Group Name: All Wells
Report Differs from Last Save: No

Run Setup

Run Information

Run Date: 12/16/2024 3:07 PM
Run User: admin
Run Type: User-defined
Plate File: Elgaria_Multiplex.prcl.pltd
ID:
Notes:
Sample Volume: 20
Temperature Control Mode: Calculated
Lid Temperature: 105
Base Serial Number: 795BR20744
Optical Head Serial Number: 795BR20744

Protocol

- 1: 95.0°C for 3:00
- 2: 95.0°C for 0:15
- 3: 49.0°C for 0:30
- 4: 55.0°C for 0:30
- 5: GOTO 3, 1 more times
- 6: 95.0°C for 0:15
- 7: 56.0°C for 0:30
- 8: 60.0°C for 0:30
Plate Read
- 9: GOTO 6, 34 more times
- 10: Melt Curve 60.0°C to 95.0°C: Increment 0.5°C 0:05
Plate Read

Plate Display


	1	2	3	4	5	6	7	8	9	10	11	12
A	Std-1 SYBR STD1	Std-1 SYBR STD1	Unk-1 SYBR 7.2_D	Unk-1 SYBR 7.2_D	Unk-9 SYBR 2.3_D	Unk-9 SYBR 2.3_D	Unk-17 SYBR 5.4_C	Unk-17 SYBR 5.4_C	Unk-25 SYBR 8.1_B	Unk-25 SYBR 8.1_B		
B	Std-2 SYBR STD2	Std-2 SYBR STD2	Unk-2 SYBR 10.2_A	Unk-2 SYBR 10.2_A	Unk-10 SYBR 9.3_C	Unk-10 SYBR 9.3_C	Unk-18 SYBR 10.2_D	Unk-18 SYBR 10.2_D	Unk-26 SYBR 13.4_C	Unk-26 SYBR 13.4_C	Unk-34 SYBR 10.5_B	Unk-34 SYBR 10.5_B
C	Std-3 SYBR STD3	Std-3 SYBR STD3	Unk-3 SYBR 7.5_A	Unk-3 SYBR 7.5_A	Unk-11 SYBR 10.2_B	Unk-11 SYBR 10.2_B	Unk-19 SYBR 7.2_B	Unk-19 SYBR 7.2_B	Unk-27 SYBR 14.5_B	Unk-27 SYBR 14.5_B	Unk-35 SYBR 11.2_C	Unk-35 SYBR 11.2_C


Plate Display

	1	2	3	4	5	6	7	8	9	10	11	12
D	Std-4 SYBR STD4	Std-4 SYBR STD4	Unk-4 SYBR 13.1_A	Unk-4 SYBR 13.1_A	Unk-12 SYBR 17.5_D	Unk-12 SYBR 17.5_D	Unk-20 SYBR 6.2_A	Unk-20 SYBR 6.2_A	Unk-28 SYBR 5.5_D	Unk-28 SYBR 5.5_D	Unk-36 SYBR 11.2_A	Unk-36 SYBR 11.2_A
E	*Std-5 SYBR STD5	*Std-5 SYBR STD5	Unk-5 SYBR 17.4_D	Unk-5 SYBR 17.4_D	Unk-13 SYBR 14.5_A	Unk-13 SYBR 14.5_A	Unk-21 SYBR 15.1_D	Unk-21 SYBR 15.1_D	Unk-29 SYBR 5.1_D	Unk-29 SYBR 5.1_D	Unk-37 SYBR 17.3_B	Unk-37 SYBR 17.3_B
F	*Std-6 SYBR STD6	*Std-6 SYBR STD6	Unk-6 SYBR 7.4_A	Unk-6 SYBR 7.4_A	*Unk-14 SYBR 5.6_C	*Unk-14 SYBR 5.6_C	Unk-22 SYBR 7.3_B	Unk-22 SYBR 7.3_B	Unk-30 SYBR 5.3_C	Unk-30 SYBR 5.3_C	Unk-38 SYBR 17.5_A	Unk-38 SYBR 17.5_A
G	*Std-7 SYBR STD7	*Std-7 SYBR STD7	Unk-7 SYBR 3.1_D	Unk-7 SYBR 3.1_D	Unk-15 SYBR 13.4_A	Unk-15 SYBR 13.4_A	Unk-23 SYBR 12.2_B	Unk-23 SYBR 12.2_B	Unk-31 SYBR 13.1_C	Unk-31 SYBR 13.1_C	Unk-39 SYBR 8.2_C	Unk-39 SYBR 8.2_C
H	NTC-1 SYBR NEG	NTC-1 SYBR NEG	Unk-8 SYBR 17.3_C	Unk-8 SYBR 17.3_C	Unk-16 SYBR 15.2_D	Unk-16 SYBR 15.2_D	Unk-24 SYBR 6.2_C	Unk-24 SYBR 6.2_C	Unk-32 SYBR 5.2_B	Unk-32 SYBR 5.2_B	Unk-40 SYBR 15.2_B	Unk-40 SYBR 15.2_B

Quantification

Step #: 8
Analysis Mode: Fluorophore
Baseline Setting: Baseline Subtracted Curve Fit
Cq Determination: Single Threshold
Baseline Method:
SYBR: Auto Calculated
Threshold Setting:
SYBR: 150.00, User Defined





Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
A01	SYBR		Std-01	STD1	17.17	17.09	0.115	3.000E+11	11.477	3.00E+11	0.00E+00
A02	SYBR		Std-01	STD1	17.01	17.09	0.115	3.000E+11	11.477	3.00E+11	0.00E+00
A03	SYBR		Unkn-01	7.2_D	17.85	17.90	0.075	2.037E+11	11.309	1.96E+11	1.02E+10
A04	SYBR		Unkn-01	7.2_D	17.95	17.90	0.075	1.893E+11	11.277	1.96E+11	1.02E+10
A05	SYBR		Unkn-09	2.3_D	17.60	17.58	0.030	2.418E+11	11.384	2.45E+11	5.10E+09
A06	SYBR		Unkn-09	2.3_D	17.56	17.58	0.030	2.490E+11	11.396	2.45E+11	5.10E+09
A07	SYBR		Unkn-17	5.4_C	17.75	17.96	0.296	2.174E+11	11.337	1.90E+11	3.88E+10
A08	SYBR		Unkn-17	5.4_C	18.17	17.96	0.296	1.625E+11	11.211	1.90E+11	3.88E+10
A09	SYBR		Unkn-25	8.1_B	17.39	17.35	0.055	2.804E+11	11.448	2.88E+11	1.10E+10
A10	SYBR		Unkn-25	8.1_B	17.31	17.35	0.055	2.961E+11	11.471	2.88E+11	1.10E+10
B01	SYBR		Std-02	STD2	20.69	20.67	0.030	3.000E+10	10.477	3.00E+10	0.00E+00
B02	SYBR		Std-02	STD2	20.65	20.67	0.030	3.000E+10	10.477	3.00E+10	0.00E+00
B03	SYBR		Unkn-02	10.2_A	17.68	17.62	0.088	2.281E+11	11.358	2.38E+11	1.46E+10
B04	SYBR		Unkn-02	10.2_A	17.56	17.62	0.088	2.486E+11	11.396	2.38E+11	1.46E+10
B05	SYBR		Unkn-10	9.3_C	17.37	17.35	0.040	2.830E+11	11.452	2.89E+11	7.98E+09
B06	SYBR		Unkn-10	9.3_C	17.32	17.35	0.040	2.943E+11	11.469	2.89E+11	7.98E+09
B07	SYBR		Unkn-18	10.2_D	16.86	16.89	0.045	4.049E+11	11.607	3.96E+11	1.24E+10
B08	SYBR		Unkn-18	10.2_D	16.92	16.89	0.045	3.874E+11	11.588	3.96E+11	1.24E+10
B09	SYBR		Unkn-26	13.4_C	17.14	17.12	0.026	3.340E+11	11.524	3.38E+11	6.10E+09
B10	SYBR		Unkn-26	13.4_C	17.10	17.12	0.026	3.427E+11	11.535	3.38E+11	6.10E+09
B11	SYBR		Unkn-34	10.5_B	17.37	17.39	0.029	2.839E+11	11.453	2.80E+11	5.70E+09
B12	SYBR		Unkn-34	10.5_B	17.41	17.39	0.029	2.759E+11	11.441	2.80E+11	5.70E+09
C01	SYBR		Std-03	STD3	24.49	24.39	0.148	3.000E+09	9.477	3.00E+09	0.00E+00
C02	SYBR		Std-03	STD3	24.28	24.39	0.148	3.000E+09	9.477	3.00E+09	0.00E+00
C03	SYBR		Unkn-03	7.5_A	18.00	17.90	0.141	1.835E+11	11.264	1.97E+11	1.93E+10
C04	SYBR		Unkn-03	7.5_A	17.80	17.90	0.141	2.108E+11	11.324	1.97E+11	1.93E+10
C05	SYBR		Unkn-11	10.2_B	17.55	17.45	0.150	2.503E+11	11.398	2.70E+11	2.80E+10
C06	SYBR		Unkn-11	10.2_B	17.34	17.45	0.150	2.899E+11	11.462	2.70E+11	2.80E+10

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
C07	SYBR		Unkn-19	7.2_B	17.39	17.43	0.064	2.809E+11	11.448	2.72E+11	1.22E+10
C08	SYBR		Unkn-19	7.2_B	17.48	17.43	0.064	2.636E+11	11.421	2.72E+11	1.22E+10
C09	SYBR		Unkn-27	14.5_B	17.23	17.15	0.113	3.120E+11	11.494	3.30E+11	2.58E+10
C10	SYBR		Unkn-27	14.5_B	17.07	17.15	0.113	3.486E+11	11.542	3.30E+11	2.58E+10
C11	SYBR		Unkn-35	11.2_C	17.11	17.23	0.166	3.399E+11	11.531	3.14E+11	3.61E+10
C12	SYBR		Unkn-35	11.2_C	17.35	17.23	0.166	2.887E+11	11.461	3.14E+11	3.61E+10
D01	SYBR		Std-04	STD4	26.94	26.90	0.062	3.000E+08	8.477	3.00E+08	0.00E+00
D02	SYBR		Std-04	STD4	26.86	26.90	0.062	3.000E+08	8.477	3.00E+08	0.00E+00
D03	SYBR		Unkn-04	13.1_A	18.32	18.28	0.055	1.472E+11	11.168	1.51E+11	5.79E+09
D04	SYBR		Unkn-04	13.1_A	18.24	18.28	0.055	1.554E+11	11.191	1.51E+11	5.79E+09
D05	SYBR		Unkn-12	17.5_D	17.97	17.91	0.084	1.866E+11	11.271	1.95E+11	1.13E+10
D06	SYBR		Unkn-12	17.5_D	17.86	17.91	0.084	2.025E+11	11.307	1.95E+11	1.13E+10
D07	SYBR		Unkn-20	6.2_A	17.59	17.75	0.233	2.441E+11	11.388	2.19E+11	3.53E+10
D08	SYBR		Unkn-20	6.2_A	17.92	17.75	0.233	1.942E+11	11.288	2.19E+11	3.53E+10
D09	SYBR		Unkn-28	5.5_D	17.47	17.31	0.214	2.655E+11	11.424	2.97E+11	4.38E+10
D10	SYBR		Unkn-28	5.5_D	17.16	17.31	0.214	3.276E+11	11.515	2.97E+11	4.38E+10
D11	SYBR		Unkn-36	11.2_A	17.72	17.84	0.180	2.232E+11	11.349	2.05E+11	2.56E+10
D12	SYBR		Unkn-36	11.2_A	17.97	17.84	0.180	1.870E+11	11.272	2.05E+11	2.56E+10
E03	SYBR		Unkn-05	17.4_D	18.04	18.12	0.110	1.785E+11	11.252	1.69E+11	1.30E+10
E04	SYBR		Unkn-05	17.4_D	18.19	18.12	0.110	1.601E+11	11.204	1.69E+11	1.30E+10
E05	SYBR		Unkn-13	14.5_A	17.93	17.78	0.216	1.924E+11	11.284	2.15E+11	3.21E+10
E06	SYBR		Unkn-13	14.5_A	17.62	17.78	0.216	2.378E+11	11.376	2.15E+11	3.21E+10
E07	SYBR		Unkn-21	15.1_D	17.63	17.56	0.098	2.373E+11	11.375	2.49E+11	1.69E+10
E08	SYBR		Unkn-21	15.1_D	17.49	17.56	0.098	2.612E+11	11.417	2.49E+11	1.69E+10
E09	SYBR		Unkn-29	5.1_D	17.13	17.05	0.113	3.364E+11	11.527	3.56E+11	2.79E+10
E10	SYBR		Unkn-29	5.1_D	16.97	17.05	0.113	3.759E+11	11.575	3.56E+11	2.79E+10
E11	SYBR		Unkn-37	17.3_B	18.15	18.13	0.032	1.648E+11	11.217	1.67E+11	3.67E+09
E12	SYBR		Unkn-37	17.3_B	18.11	18.13	0.032	1.700E+11	11.230	1.67E+11	3.67E+09
F03	SYBR		Unkn-06	7.4_A	17.83	18.03	0.284	2.065E+11	11.315	1.81E+11	3.55E+10
F04	SYBR		Unkn-06	7.4_A	18.23	18.03	0.284	1.563E+11	11.194	1.81E+11	3.55E+10
F07	SYBR		Unkn-22	7.3_B	17.92	17.97	0.067	1.938E+11	11.287	1.88E+11	8.68E+09
F08	SYBR		Unkn-22	7.3_B	18.01	17.97	0.067	1.815E+11	11.259	1.88E+11	8.68E+09
F09	SYBR		Unkn-30	5.3_C	18.33	18.25	0.106	1.458E+11	11.164	1.54E+11	1.13E+10
F10	SYBR		Unkn-30	5.3_C	18.18	18.25	0.106	1.618E+11	11.209	1.54E+11	1.13E+10
F11	SYBR		Unkn-38	17.5_A	18.24	18.14	0.148	1.549E+11	11.190	1.67E+11	1.71E+10
F12	SYBR		Unkn-38	17.5_A	18.03	18.14	0.148	1.791E+11	11.253	1.67E+11	1.71E+10
G03	SYBR		Unkn-07	3.1_D	18.51	18.74	0.334	1.289E+11	11.110	1.11E+11	2.55E+10
G04	SYBR		Unkn-07	3.1_D	18.98	18.74	0.334	9.285E+10	10.968	1.11E+11	2.55E+10
G05	SYBR		Unkn-15	13.4_A	18.02	17.92	0.135	1.811E+11	11.258	1.94E+11	1.81E+10
G06	SYBR		Unkn-15	13.4_A	17.83	17.92	0.135	2.067E+11	11.315	1.94E+11	1.81E+10
G07	SYBR		Unkn-23	12.2_B	17.68	17.69	0.006	2.285E+11	11.359	2.28E+11	1.00E+09
G08	SYBR		Unkn-23	12.2_B	17.69	17.69	0.006	2.270E+11	11.356	2.28E+11	1.00E+09
G09	SYBR		Unkn-31	13.1_C	17.67	17.57	0.147	2.301E+11	11.362	2.48E+11	2.52E+10
G10	SYBR		Unkn-31	13.1_C	17.46	17.57	0.147	2.659E+11	11.425	2.48E+11	2.52E+10
G11	SYBR		Unkn-39	8.2_C	17.70	17.71	0.011	2.255E+11	11.353	2.24E+11	1.72E+09
G12	SYBR		Unkn-39	8.2_C	17.72	17.71	0.011	2.230E+11	11.348	2.24E+11	1.72E+09
H01	SYBR		NTC-01	NEG	28.92	28.54	0.538	N/A	N/A	N/A	N/A
H02	SYBR		NTC-01	NEG	28.16	28.54	0.538	N/A	N/A	N/A	N/A
H03	SYBR		Unkn-08	17.3_C	18.05	18.05	0.006	1.775E+11	11.249	1.77E+11	7.33E+08

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev	Starting Quantity (SQ)	Log Starting Quantity	SQ Mean	SQ Std. Dev
H04	SYBR		Unkn-08	17.3_C	18.05	18.05	0.006	1.765E+11	11.247	1.77E+11	7.33E+08
H05	SYBR		Unkn-16	15.2_D	17.61	17.53	0.126	2.395E+11	11.379	2.55E+11	2.23E+10
H06	SYBR		Unkn-16	15.2_D	17.44	17.53	0.126	2.710E+11	11.433	2.55E+11	2.23E+10
H07	SYBR		Unkn-24	6.2_C	18.19	18.32	0.185	1.602E+11	11.205	1.47E+11	1.89E+10
H08	SYBR		Unkn-24	6.2_C	18.46	18.32	0.185	1.335E+11	11.126	1.47E+11	1.89E+10
H09	SYBR		Unkn-32	5.2_B	17.91	17.77	0.194	1.950E+11	11.290	2.15E+11	2.90E+10
H10	SYBR		Unkn-32	5.2_B	17.64	17.77	0.194	2.360E+11	11.373	2.15E+11	2.90E+10
H11	SYBR		Unkn-40	15.2_B	17.60	17.57	0.039	2.425E+11	11.385	2.47E+11	6.65E+09
H12	SYBR		Unkn-40	15.2_B	17.54	17.57	0.039	2.519E+11	11.401	2.47E+11	6.65E+09

Bar Chart

Expression analysis requires that at least two wells contain sample names, targets, and valid Cqs.

QC Parameters

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Negative control with a Cq less than	38	True		False	
NTC with a Cq less than	38	True	SYBR:H1, H2.	False	
NRT with a Cq less than	38	True		False	
Positive control with a Cq greater than	30	True		False	
Unknown without a Cq	N/A	True		False	
Standard without a Cq	N/A	True		False	
Efficiency greater than	110.0	True			
Efficiency less than	90.0	True			
Std Curve R^2 less than	0.980	True			
Replicate group Cq Std Dev greater than	0.20	True	SYBR:A7, A8, D7, D8, D9, D10, E5, E6, F3, F4, G3, G4, H1, H2.	False	