Lecture 12-1 Image reconstruction

Yuyao Zhang, Xiran Cai PhD

zhangyy8@shanghaitech.edu.cn caixr@shanghaitech.edu.cn

SIST Building 2 302-F/302-C

Course piazza link: piazza.com/shanghaitech.edu.cn/spring2021/cs270spring2021





Outline

- Projection and back-projection
- Radon transform
- Fourier-Slice Theorem
- Filtered back-projection



Projection and back projection










CT scanning methods





Radon transform





$$g(\rho_{j}, \theta_{k}) = \int_{-\infty}^{+\infty} \int_{-\infty}^{+\infty} f(x, y) \delta(x \cos \theta_{k} + y \sin \theta_{k} - \rho_{j}) dx dy$$


$$g(\rho, \theta) = \int_{-\infty}^{+\infty} \int_{-\infty}^{+\infty} f(x, y) \delta(x \cos \theta + y \sin \theta) dx dy$$

$$g(\rho, \theta) = \sum_{x=0}^{M-1} \sum_{y=0}^{N-1} f(x, y) \delta(x \cos \theta + y \sin \theta - \rho)$$



Sinogram

Radon transform $g(\rho,\theta)$ is displayed as an image with ρ and θ as rectilinear coordinates





Back-projection from Sinogram

For a fixed value of rotation θ_k :


$$f_{\theta_k}(x, y) = g(\rho, \theta_k) = g(x \cos \theta_k + y \sin \theta_k, \theta_k)$$


Then a single back-projection image obtained at an angle θ is :

$$f_{\theta}(x,y) = g(x\cos\theta + y\sin\theta,\theta)$$

The reconstructed image is obtained by summing over all the back-projected images:

$$f(x,y) = \sum_{\theta=0}^{\pi} f_{\theta}(x,y)$$







Fourier-Slice Theorem

 \succ The 1D FT of a projection with respect to ρ is:

$$G(\omega, \theta) = \int_{-\infty}^{+\infty} g(\rho, \theta) e^{-j2\pi\omega\rho} d\rho$$


> Then

$$G(\omega, \theta) = \int_{-\infty}^{+\infty} \int_{-\infty}^{+\infty} f(x, y) e^{-j2\pi\omega(x\cos\theta + y\sin\theta)} dx dy$$

$$= \left[\int_{-\infty}^{+\infty} \int_{-\infty}^{+\infty} f(x, y) e^{-j2\pi(ux + vy)} dx dy \right]_{u = \omega \cos \theta; v = \omega \sin \theta}$$



$$G(\omega, \theta) = [F(u, v)]_{u = \omega \cos \theta; v = \omega \sin \theta} = F(\omega \cos \theta, \omega \sin \theta)$$





Filtered back-projection

 \triangleright The 2D IFT of F(u, v) is:

$$f(x,y) = \int_{-\infty}^{+\infty} \int_{-\infty}^{+\infty} F(u,v) e^{j2\pi(ux+vy)} dudv$$


$$= \int_0^{2\pi} \int_0^{\infty} G(\omega, \theta) e^{j2\pi\omega(x\cos\theta + y\sin\theta)} \omega d\omega d\theta$$

$$= \int_0^{\pi} \left[\int_{-\infty}^{+\infty} |\omega| G(\omega, \theta) e^{j2\pi\omega\rho} d\omega \right]_{\rho = x \cos \theta + y \sin \theta} d\theta$$




$$f(x,y) = \int_0^{\pi} [s(\rho) \otimes g(\rho,\theta)]_{\rho=x \cos \theta + y \sin \theta} d\theta$$

Where
$$s(\rho) = IFT(|\omega|), g(\rho, \theta) = IFT(G(\omega, \theta))$$





Filtered back-projection










Fan-Beam based Filtered back-projection





Take home message

- ➤ CT imaging reconstruction is based on accumulating backprojections data directive, while the reconstructed images are blurred.
- The 2D Fourier transform of an image for reconstruction can be obtained by accumulating the Fourier transform of the projections at different angles (Fourier-Slice Theorem).
- > Filtered back-projection can mitigate the blurring effects.

