Organisation Internationale de Métrologie Légale

International Organization of Legal Metrology

COMMENTS TEMPLATE

Template revision date: 2020-01-10
TC9_SC1_P1_N041

TC 9/SC 1/p 1:	**Revision of R 76: *Non-automatic weighing instruments***		
PG vote/comments on 1CD:	TC9_SC1_P1_N035 (part 1), N036 (part 2), N037 (part 3), N038 (part 4), N039 (part 5)		
Circulation date:	30 April 2024	Conveners:	**Closing date for comments: 30 July 2024 at 17:00 CET**
Date comments submitted:	**30 July 2024**	Please type your comments in this form and post it (in Word format) as soon as possible and no later than the closing date using the CD vote and comment page on the OIML website (My access → CD vote & comment).	
PLEASE INSERT THE COUNTRY CODE AND THE PART AND CLAUSE NUMBER IN EACH ROW. PLEASE DO NOT MODIFY THE NUMBER OF COLUMNS IN THE TABLE.			

Instructions for using this template:

The structure of this table allows for the automatic collation of all the comments posted by the participants. However, this is only possible if the following instructions are followed. Please

- do not add any columns to the table,
- do not merge any of the cells,
- add the Country Code in each row,
- fill in the Part number in each row (if the document to be commented has no parts, leave this column blank),
- enter one reference per row in the Clause/Sub clause column. If your comment applies to more than one clause, please repeat the row or make the reference in the Comments column,
- do not embed other tables in the table,
- enter the date on which you make the comments in the heading.

1 **Country code** = ISO 3166 two-letter country code, e.g. CN for China

2 **Type of comment**: ge = general te = technical ed = editorial

Country Code[1]	Part	Clause/ Sub clause	Paragraph / Figure/ Table/	Type of comment[2]	COMMENTS	PROPOSED CHANGE	OBSERVATIONS OF THE CONVENER/PG on each comment submitted
AU				ge	It is not clear whether the scope of the current review is intended to consider technical changes. Or whether it is limited to formatting and structural changes. Australia has concerns with some of the technical changes being proposed. Australia does have technical issues that it wants considered and we have included some technical comments as part of this revision, however we suggest technical changes be deferred to a more substantive review.		
AU	1	3.1.2	Note 1	ed	A non-automatic weighing process allows the operator to take an action (i.e., adjust the unit price) Adjusting the unit price won't influence the weighing result, only the transaction result.	Delete 'adjust the unit price' or reword by changing 'influences the weighing result' to the transaction or measurement result.	
AU	1	5.5.1	1	te	As 5.5.3.3 regarding exception for preset tare has been removed. Now MPEs also apply to calculated net value when a preset tare device is in use. Net value is not calculated net value and tare device is not the same as preset tare device. The statement 'The MPEs apply to the net value when a tare is in use' is not 100% correct.	The MPEs apply to the net value/calculated net value when a tare/preset tare is in use. R76-2 5.6.1 is the test for tare device. Should the similar test to R76-2 5.6.1 to be implemented?	
AU	1	5.5.2		ge	Australia strongly objects to the proposed content of 5.5.2 and the doubling of MPEs for in-service inspections. Australia's legislated in-service MPE's are equal to the value of Verification MPEs. There is no definition of an in-service inspection.	Remove all reference to in-service inspection and in-service MPEs. We suggest that the topic of in-service MPEs be removed from the scope of this revision of OIML R 76 The topic of in-service inspections and MPEs has significant implications at a national level and should be deferred to the next full revision of OIML R 76 to allow for the necessary consultation and analysis processes to occur to support the definition and introduction of in-service MPEs	
AU	1	7.1.10		te	Clause 7.1.3.2, Semi-automatic tare devices, does not permit reduction of tare value, but permits the increase of tare value. In the self-service bring own container case (an increased trend), it is possible for a customer to tare off the empty container first then fill the container with product and tare off the partially filled container, then fill the container again with more product. The customer only pays for the last part of fill.	Add: The action of the tare device must not permit an increase of a value of the tare when using the same container.	

Country Code[1]	Part	Clause/ Sub clause	Paragraph / Figure/ Table/	Type of comment[2]	COMMENTS	PROPOSED CHANGE	OBSERVATIONS OF THE CONVENER/PG on each comment submitted
AU	2	5.1.11		te	But no details of how to interpret the MPE requirements, especially when unloading. The EUT meets MPE of higher range does not mean it could meet the MPE of lowest range.	Add note: when unloading an instrument with multiple ranges the MPE of the instrument at smaller loads uses the MPEs of the highest range reached in loading the instrument before it returns to zero.	
AU	2	5.6.1		ed	Weighing test (R 76-1, 5.5.3.3). 5.5.3.3 has been deleted	Should be 5.5.1	
AU	2	5.10		ge	For Verification testing, the current Australian requirements are 2/3 (66.66%) Max. On a Class III 15 kg capacity instrument as an example this would equate to 10 kg – the OIML standard is 14 kg. Likewise with a pallet scale of 1,500 kg capacity would require 1,000 kg at 2/3 Max – or 1,200 kg at 0.8 Max. A 1,000 kg test using one block is simpler, more efficient and safer than e.g. a 1,000 kg block with 10 x 20 kg blocks sitting on top. At 2/3 capacity, typically the 2nd changeover point (2000e for a 3000 VSI Instrument), therefore determination of any error is made easier also.	**5.10 Repeatability test (R 76-1, 5.5.4)** For type approval two series of weighings shall be performed, one with a load of about 50 % and one with a load close to 100 % of Max. For instruments with Max less than 1 000 kg each series shall consist of ten weighings. In other cases each series shall consist of at least three weighings. Readings shall be taken when the instrument is loaded, and when the unloaded instrument has come to rest between weighings. In the case of a zero deviation between the weighings, the instrument shall be reset to zero, without determining the error at zero. The true zero position need not be determined between the weighings. If the instrument is provided with automatic zero-setting or zero-tracking, it shall be in operation during the test. For verification one series of weighings with 2/3 Max is sufficient. Three weighings on classes III and IV or six weighings on classes I and II are necessary.	
AU	2	12.2.6		te	You agreed with the previous comments that for multiple range instrument, where DR is unknown, the last paragraph should be the condition nLC ≥ 0.5 × Maxr / e1 is satisfied. But CD1.1 is, nLC ≥ 0.4 × Maxr / e1 is satisfied.	The last paragraph is 0.5 instead of 0.4. Also the compatibility check form needs to be updated.	
AU	2	12.2.6		ed	For multi-interval instrument where DR is unknown, the condition nLC ≥ 0.5 × Maxr / e1 is satisfied.	Go back to nLC ≥ Max / e1 is satisfied.	
AU	3	6.2	creep	te	R76-1 5.8.4.1 requires observation at 5 min intervals	Additional rows for 10, 20, 25 min	

<table>
<tr><th>Country Code[1]</th><th>Part</th><th>Clause/ Sub clause</th><th>Paragraph / Figure/ Table/</th><th>Type of comment[2]</th><th>COMMENTS</th><th>PROPOSED CHANGE</th><th>OBSERVATIONS OF THE CONVENER/PG
on each comment submitted</th></tr>
<tr><td>AU</td><td>5</td><td>6</td><td>Table 5</td><td>ge</td><td>Australia does not legislate a reverification period for NAWI’s excepting public weighbridges. Noting this we provide the following as part of our submission.
We strongly disagree with the varying verification periods being based on “intensive use”, especially when 10 weighing’s per week on average is the decisive number. We point out that recording the number of weighings a NAWI is used for would be impractical to regulate.
Consideration could be given to some sort of “Class Based” system as being a fairer and more appropriate option. Generally, Class I, II and smaller capacity Class III instruments are more cared for and in better environments, whereas Class IV and Y (Front End Loaders, etc) are used extensively with many factors that can affect weighing performance. Such factors in weighbridges include, but are not limited to, mud build up around the load cells, build up under the decks and general wear and tear of the load cells and retaining plates. Similar issues can also apply to Pallet scales of up to 3 tonne capacity. Class Y instruments could have hydraulic parts replaced which can affect weighing performance. Mechanical Instruments are more inclined to wear over time and have greater potential of binding and fouling in their mechanisms.
In short, Class I, II, and smaller capacity Class III instruments operate off a single point load cell, everything else has moving parts which can wear or get inhibited in their performance.</td><td>Option 1:<table><tr><td>Class I</td><td>2 years</td></tr><tr><td>Class II</td><td>2 years</td></tr><tr><td>Class III Single Point load cell system</td><td>2 years</td></tr><tr><td>All other Class III</td><td>1 year</td></tr><tr><td>Class IV & Y and mechanical</td><td>1 year</td></tr></table>Option 2:<table><tr><td>Class I</td><td>2 years</td></tr><tr><td>Class II</td><td>2 years</td></tr><tr><td>Class III <= 300 kg</td><td>2 years</td></tr><tr><td>Class III > 300 kg)</td><td>1 year</td></tr><tr><td>Class IV & Y and mechanical</td><td>1 year</td></tr></table></td><td></td></tr>
<tr><td>AU</td><td>5</td><td>6</td><td>Table 5</td><td>ge</td><td>A recommendation should also be inserted that a check / inspection be conducted where weighing instruments have been relocated from the verification location, to ensure that Gravity change or manual handling of this relocation has not altered the weighing instruments metrological performance.</td><td>Add note: Where environment conditions impact the ability for the weighing instrument to maintain usability and display accuracy then a re-verification of less than the recommended verification interval may be required. In such case a Risk assessment may need to be conducted to determine suitable re-verification period</td><td></td></tr>
</table>

Country Code[1]	Part	Clause/ Sub clause	Paragraph / Figure/ Table/	Type of comment[2]	COMMENTS	PROPOSED CHANGE	OBSERVATIONS OF THE CONVENER/PG on each comment submitted
AU	5	7.1		ge	Some Digital Indicators have a Hi Resolution mode (generally x10), permitted with regard to higher capacity instruments, where the Digital Indicator has been endorsed a “Hi Res approved for Verification purposes” in its type approval.	Add note: For instruments that have an approved digital indicator “endorsed to display in hi-resolution mode (auxiliary indication) for the purposes of verification”, testing can be performed using the instrument’s auxiliary indication to determine the error.	
AU	5	7.8		ge	Inappropriate to include Euro-specific link in OIML document. It is noted the link takes the user to an invalid service.		
AU	5	7.8	First and second para.	ge	Is the statement “For class IIII weighing instruments, the effect of local gravity is negligible” always true?	Delete last sentence of first paragraph. Amend second paragraph: For other weighing instruments, being for class 3 and class 4, the following rule can be considered.	
AU	5	7.8		Ge.	Do not agree that gravity value at place of adjustment is “normally written on the nameplate”.	Remove “normally written on the nameplate” from g1	